

JUAN C. ROJAS, M.D. LUIS E. OSTERBERGER, M.D.

VIJAY KUDESIA, M.D., F.A.C.C. DOMINIC A. CUSUMANO III, M.D.

# Cardiovascular Associates, P.C.







27177 LAHSER ROAD, SUITE 103 SOUTHFIELD, MICHIGAN 48034 TELEPHONE: (248) 357-1360 FAX: (248) 357-1745

28295 SCHOENHERR, SUITE C WARREN, MICHIGAN 48088 TELEPHONE: (586) 573-7300

FAX: (586) 573-7315

2300 HAGGERTY ROAD, SUITE 2190 WEST BLOOMFIELD, MICHIGAN 48323

TELEPHONE: (248) 926-9075 FAX: (248) 926-9079



June 26, 2009

MARIA PACIS-BIEDERMAN, D.O., F.A.C.C.

OSWALD BOSTIC, B.S.C., M.D., F.R.C.P., (C)

UNITED STATES NUCLEAR REGULATORY COMMISSION Region III, Materials Licensing Section 2443 Warrenville Road Suite 210 Lisle, IL 60532-4352

Amendment for NRC License #21-15266-01 Re:

Close CVA - Sterling Heights Location

Dear Sir/Madam:

We currently have an amendment request in the possession of the NRC to remove an address of use in West Bloomfield, Michigan.

In addition to the amendment request in process, please amend our NRC license to remove the following location of radioactive use. Due to patient care issues, we ask that you please expedite this request.

Cardiovascular Associates, P.C. 43184 Dequindre Road, Suite 201 Sterling Heights, MI 48314

The results of the close-out are enclosed for your review.

If you have any questions or require additional information please contact our Medical Physics Consultant and Radiation Safety Officer, Cari Dzanbazoff, at (734) 662-3197.

Sincerely.

Juan Rojas, M.D.

Executive Management

Cardiovascular Associates, P.C.

# **Close-out Survey Information**

#### Close-out Location:

Cardiovascular Associates, P.C. 43184 Dequindre Road, Suite 201 Sterling Heights, MI 48314

Date Performed: June 23, 2009

Performed By: Cari Dzanbazoff, Medical Physics Consultants Inc.

Comments: Hot lab equipment and radioactive sealed sources were properly

transferred to the Warren location. Sealed source leak test is enclosed.

No radioactivity had been received at the facility since 5/26/09.

## Instruments:

Wipes for contamination performed with the following well counter:

Manufacturer:

Capintec

Model Number:

Captus 3000

Serial Number:

900271

Efficiency Factor:

2.5 dpm/cpm

Area survey performed with the following survey meter:

Manufacturer:

Ludlum

Type:

GM

Model Number:

Model 14 C\_

Serial Number:

230200

Probe Model:

Pancake Probe

Calibration Date:

07/07/2008

Battery check acceptable:

YES

Operational check acceptable: Constancy check acceptable:

YES YES

Check source value:

8.0 mR/hr

## Radionuclides Used:

### Unsealed

Tc-99m, Tl-201

### Sealed

Ba-133, Co-57, Cs-137

All sealed sources were properly transferred to the Warren location.

Sealed source leak tests enclosed.

Visual Check: The area was checked to ensure that all radioactive waste had been removed. No evidence of radioactive material was noted.

Radiation Level Survey: No area within the lab demonstrated radiation levels in excess of the background reading of 0.02 mR/hr.

Removable Contamination: No area within the lab demonstrated removable contamination in excess of established trigger levels.


Sealed Sources: All sealed sources were properly transferred to the Warren location.

Surveys, Wipes, and Diagrams are enclosed for review.

## **CLOSEOUT DIAGRAM**

# Cardiovascular Associates, P.C. 43184 Dequindre Road, Suite 201 Sterling Heights, MI 48314

# NORTH **Î**



**Exam Room** 

This is a single level facility. There are no specific room numbers for the areas indicated.

# CARDIOVASCULAR ASSOCIATES, P.C. - STERLING HEIGHTS LOCATION WIPE/SURVEY RESULTS

Date: 6/23/2009 Performed By: Cari Dzanbazoff

Survey Meter:Ludium 14CWeil Counter:CapintecSerial Number:230200Model:Captus 3000Calibration Date:7/7/2008Serial Number:900271

		cpm	net dpm	mR/h
1	dosing area	567	128.0	0.03
2	counter	567	128.0	0.03
3	floor	567	128.0	0.03
4	floor	567	128.0	0.03
5	floor	567	128.0	0.03
6	floor	556	100.5	0.03
7	floor	556	100.5	0.03
8	floor	556	100.5	0.03
9	floor	556	100.5	0.03
10	cabinets	556	100.5	0.03
11	desk	531	38.0	0.03
12	floor	531	38.0	0.03
13	floor	531	38.0	0.03
14	camera	531	38.0	0.03
15	camera	531	38.0	0.03
16	equipment	542	65.5	0.03
17	collimators	542	65.5	0.03
18	acquisition	542	65.5	0.03
19	floor	542	65.5	0.03
20	floor	542	65.5	0.03
21	floor	557	103.0	0.03
22	treadmill	557	103.0	0.03
23	exam table	557	103.0	0.03
24	sink	557	103.0	0.03
25	floor	557	103.0	0.03
26	floor	529	33.0	0.03
27	floor	529	33.0	0.03
28	floor	529	33.0	0.03
29	floor	529	33.0	0.03
30	floor	529	33.0	0.03
	background	515.8		0.03

ACTION LEVELS: 1000 dpm/100 cm sq

Efficiency Factor = 2.5 dpm/cpm

Full Spectrum

COMMENTS: No evidence of removable contamination.

## Medical Physics Consultants, Inc.

# Sealed Source Leak Test

Licensee: Cardiovascular Associates, P.C. Sterling Heights

Date: 03/31/09

Performed by:

Aarti Patel

		Calibration	Calibration			
Nuclide	Туре	Activity	Date	Location	M/N	S/N
Cs-137	Vial	202.4 uCi	07/01/05	Hot Lab	IPL	1111-95-17
Current Ac	tivity: 185	5.60uCi				
Ba-133	Vial	266.4 uCi	07/01/05	Hot Lab	IPL	1124-34-3
Current Ac	tivity: 209	9.14uCi				
Co-57	Flood	10 mCi	06/01/05	Hot Lab	NES-392	1121-189
Current Ac	tivity: 0.	279mCi				

Comment: The sources listed above were leak tested using a dry wipe technique and were found to have less than 0.005 uCi removable activity. The following Minimum Detectable Activities are based upon a background at the indicated value. Background was at or below these levels when the above tests were completed. Well Counter: Captus 3000

 Nuclide
 MDA
 Background

 Cs-137
 2.0 x 10 <sup>-4</sup> uCi
 145 counts/1 min

 Ba-133
 8.8 x 10 <sup>-5</sup> uGi
 204 counts/1 min

 Co-57
 1.7 x 10 <sup>-9</sup> uCi
 53 counts/1 min

RADIATION SAFETY OFFICER

#### CARDIOVASCULAR ASSOCIATES, P.C. 27177 LAHSER RD., SUITE 103 SOUTHFIELD, MI 48034



United States Nuclear Regulatory Commission Region III, Materials Licensing Section 2443 Warrenville Road Ste 210 Lisle, 1L 60532-4352