

GE Infrastructure Sensing

Calibration Report 060915A0813

Digital Pressure Indicator

for

Fugro Consultants LP

6100 Hillcroft
Houston, TX 77081

Date of Issue: September 15, 2006

Page 4 of 6
Full Scale: 500 psi gauge

As Found / As Left Calibration Data

Calibration Date: September 15, 2006

Calibration Standard: PC-89, and ,WS-12

Medium: nitrogen

| Applied psi | Uncertainty psi | Displayed psi |
|----------------|--------------------|------------------|
| 0.0 | 0.0E+00 | 0.0 |
| 124.6700 | 1.4E-03 | 124.7 |
| 249.580 | 2.7E-03 | 249.6 |
| 374.330 | 4.1E-03 | 374.4 |
| 499.070 | 5.5E-03 | 499.2 |
| 249.580 | 2.7E-03 | 249.6 |
| 0.0 | 0.0E+00 | 0.1 |

Notes: The instrument was not adjusted.



General Electric Company
10311 Westpark Drive
Houston, TX 77042
USA

T 713 975 0547
F 713 975 6338

GE Infrastructure Sensing

Calibration Report 060915A0813

Digital Pressure Indicator

for

Fugro Consultants LP

6100 Hillcroft

Houston, TX 77081

Date of Issue: September 15, 2006

Page 5 of 6
Full Scale: 250 psi gauge

As Left Calibration Data

Calibration Date: September 15, 2006

Calibration Standard: PC-89, WS-12, and WS-27

Medium: nitrogen

| Applied psi | Uncertainty psi | Displayed psi |
|----------------|--------------------|------------------|
| 0.00 | 0.0E+00 | 0.00 |
| 61.7960 | 6.8E-04 | 61.80 |
| 124.6680 | 1.4E-03 | 124.65 |
| 186.5430 | 2.1E-03 | 186.55 |
| 249.579 | 2.7E-03 | 249.60 |
| 124.6680 | 1.4E-03 | 124.65 |
| 0.00 | 0.0E+00 | 0.00 |

Notes: The instrument was adjusted prior to recording the above data.



General Electric Company
10311 Westpark Drive
Houston, TX 77042
USA

T 713 975 0547
F 713 975 6338

GE Infrastructure Sensing

Calibration Report 060915A0813

Digital Pressure Indicator

for

Fugro Consultants LP

6100 Hillcroft
Houston, TX 77081

Date of Issue: September 15, 2006

Page 6 of 6
Full Scale: 100 psi gauge

As Left Calibration Data

Calibration Date: September 15, 2006

Calibration Standard: PC-67, WS-12, and WS-27

Medium: nitrogen

| Applied psi | Uncertainty psi | Displayed psi |
|----------------|--------------------|------------------|
| 0.00 | 0.0E+00 | 0.00 |
| 24.86100 | 2.5E-04 | 24.84 |
| 49.8390 | 5.0E-04 | 49.84 |
| 74.9860 | 7.5E-04 | 74.98 |
| 99.9310 | 1.0E-03 | 99.96 |
| 49.8390 | 5.0E-04 | 49.82 |
| 0.00 | 0.0E+00 | 0.00 |

Notes: The instrument was adjusted prior to recording the above data.



General Electric Company
10311 Westpark Drive
Houston, TX 77042
USA

T 713 975 0547
F 713 975 6338



**National Voluntary
Laboratory Accreditation Program**



SCOPE OF ACCREDITATION TO ISO/IEC 17025:1999

GE Infrastructure Sensing
10311 Westpark Drive
Houston, TX 77042-5312
Mr. Kenneth A. Kolb
Phone: 713-975-0547 Fax: 713-975-6338
E-mail: kenneth.kolb@ge.com
URL: <http://www.gesensing.com>

CALIBRATION LABORATORIES

NVLAP LAB CODE 200491-0

NVLAP Code: 20/A01 *ANSI/NCSL Z540-1-1994; Part 1* Compliant

MECHANICAL

NVLAP Code: 20/M08
Mass

Calibration of Primary Piston Gauge Masses

| <i>Range</i> | <i>Best Uncertainty (\pm) Relative to Indicated Value ^{note 1}</i> | <i>Remarks</i> |
|----------------|--|---------------------------|
| 1 mg to 17 kg | 5.0×10^{-6} but not less than 0.5 mg | Substitution – Mechanical |
| 1 mg to 1.2 kg | 5.0×10^{-6} but not less than 0.5 mg | Substitution – Electronic |

Calibration of Secondary Piston Gauge Masses

| | | |
|----------------|---|-----------------------------|
| 1 mg to 8.0 kg | 2.0×10^{-5} but not less than 0.5 mg | Substitution – Electronic |
| 1 mg to 1.2 kg | 2.0×10^{-5} but not less than 0.5 mg | Direct Reading - Electronic |
| 1.2 kg to 8 kg | 2.0×10^{-5} but not less than 43 mg | Direct Reading – Electronic |

2006-01-01 through 2006-12-31

Effective dates

For the National Institute of Standards and Technology



National Voluntary Laboratory Accreditation Program



CALIBRATION LABORATORIES

NVLAP LAB CODE 200491-0

THERMODYNAMICS

NVLAP Code: 20/T05

Pressure

Pneumatic Pressure using Primary Piston Gauge ^{note 2}

| Range | Best Uncertainty (\pm) of Reading ^{note 1} | Remarks |
|-----------------------|---|------------------------------|
| -100 kPa to -1.38 kPa | 1.0×10^{-5} but not less than 0.07 Pa | Negative Gauge Mode |
| -16 kPa to 16 kPa | 1.1×10^{-5} but not less than 0.034 Pa | Differential Mode |
| 1.38 kPa to 1.4 MPa | 1.0×10^{-5} but not less than 0.07 Pa | Gauge Mode ^{note 4} |
| 1.4 MPa to 7 MPa | 1.1×10^{-5} but not less than 2.8 Pa | Gauge Mode ^{note 4} |
| 7 MPa to 21 MPa | $1.1 \times 10^{-5} + 1.9 \times 10^{-7}$ per MPa | Gauge Mode |
| 21 MPa to 104 MPa | 3.5×10^{-5} | Gauge Mode |

Pneumatic Effective Area Determination using Primary Piston Gauge ^{note 2}

| Range | Best Uncertainty (\pm) of Reading ^{notes 1, 7} | Remarks |
|----------------------|---|---------|
| 1.38 kPa to 345 kPa | 8.8×10^{-6} | |
| 11.72 kPa to 1.4 MPa | 8.3×10^{-6} | |
| 14 kPa to 7 MPa | $1.0 \times 10^{-5} + 2.4 \times 10^{-7}$ per MPa ^{note 3} | |
| 700 kPa to 21 MPa | $1.0 \times 10^{-5} + 4.8 \times 10^{-7}$ per MPa ^{note 3} | |
| 1.17 MPa to 104 MPa | 3.37×10^{-5} | |

Pneumatic Pressure using Precision Transducer ^{note 2}

| Range | Best Uncertainty (\pm) of Reading ^{note 1} | Remarks |
|--------------------|---|------------------------------|
| 0 Pa to 133 Pa | 0.133 Pa | Absolute Mode |
| -16 kPa to 16 kPa | 5.0×10^{-5} but not less than 0.035 Pa | Differential Mode |
| -100 kPa to 17 MPa | 6.5×10^{-5} but not less than 0.22 Pa | Gauge Mode ^{note 5} |

2006-01-01 through 2006-12-31

Effective dates

W. R. Miller

For the National Institute of Standards and Technology



National Voluntary Laboratory Accreditation Program



CALIBRATION LABORATORIES

NVLAP LAB CODE 200491-0

Pneumatic Effective Area Determination using Precision Transducer ^{note 2}

20 Pa to 17 MPa 7.2×10^{-5} but not less than 0.05 Pa

Pneumatic Deadweight Tester Output Pressure Conformance using Precision Transducer ^{note 2}

| Range | Best Uncertainty (\pm) of Reading ^{notes 1, 8} | Remarks |
|-------|---|---------|
|-------|---|---------|

| | | |
|-----------------|---|--|
| 20 Pa to 17 MPa | 7.5×10^{-5} but not less than 0.053 Pa | |
|-----------------|---|--|

Hydraulic Pressure using Primary Piston Gauge ^{note 1}

| Range | Best Uncertainty (\pm) of Reading ^{notes 1, 6} | Remarks |
|--------------------|---|------------|
| 50 kPa to 7 MPa | 2.5×10^{-5} but not less than 10 Pa | Gauge Mode |
| 7 MPa to 140 MPa | 3.5×10^{-5} | Gauge Mode |
| 14 MPa to 280 MPa | 7.5×10^{-5} | Gauge Mode |
| 280 MPa to 500 MPa | 1.0×10^{-4} | Gauge Mode |

Hydraulic Effective Area Determination using Primary Piston Gauge ^{note 2}

| Range | Best Uncertainty (\pm) of Reading ^{note 1} | Remarks |
|--------------------|---|---------|
| 50 kPa to 7 MPa | 2.31×10^{-5} | |
| 7 MPa to 140 MPa | 3.34×10^{-5} | |
| 140 MPa to 280 MPa | 7.29×10^{-5} | |
| 280 MPa to 500 MPa | 9.80×10^{-5} | |

Hydraulic Effective Area Determination using Secondary Piston Gauge ^{note 2}

70 kPa to 140 MPa 7.2×10^{-5}

2006-01-01 through 2006-12-31

Effective dates

For the National Institute of Standards and Technology



**National Voluntary
Laboratory Accreditation Program**



CALIBRATION LABORATORIES

NVLAP LAB CODE 200491-0

Hydraulic Deadweight Tester Output Pressure Conformance using Secondary Piston Gauge ^{note 2}

70 kPa to 140 MPa

7.5×10^{-5} but not less than 50 Pa

1. Represents an expanded uncertainty using a coverage factor, $k = 2$, at an approximate level of confidence of 95 %.
2. This capability includes on-site calibration service, as limited by influences of operating environment.
3. Component uncertainties are combined in quadrature.
4. For absolute mode, uncertainties increase by $1.33\text{E} + 00$ Pa, combined in quadrature with stated level.
5. For absolute mode, uncertainties increase by $1.88\text{E} + 00$ Pa, combined in quadrature with stated level.
6. For absolute mode, uncertainties increase by $1.31\text{E} + 01$ Pa, combined in quadrature with stated level.
7. Calibration process may include the use of transducers to measure small differential pressures.
8. Conformance evaluation of Deadweight Tester output pressure compared to indicated pressure.

2006-01-01 through 2006-12-31


Effective dates

For the National Institute of Standards and Technology

QUALITY SYSTEM MANUAL

FOR

HOUSTON GEOTECHNICAL LABORATORY



CONTROL #: GEO-1

Fugro Consultants, Inc.

6100 Hillcroft

Houston, Texas 77081

Phone: (713) 369-5400

Fax: (713) 369-5545

Document Revised: July 07, 2007

| CALIBRATION EQUIPMENT OR REFERENCE STANDARDS | | | |
|--|-----------------------------------|----------------|----------------|
| Equipment Name | Calibration Interval | Check Interval | Procedure Used |
| Digital Micrometer & Mechanical Micrometers | 2 years | | Outside Source |
| Force Transducers | 1 year | | Outside Source |
| Metal Specimens | Verify Before Use or After Repair | | HSL-2655 |
| Pressure Gages | 2 years | | Outside Source |
| Set of Gage Blocks | 5 years | | Outside Source |
| Thermometers | 1 year | | Outside Source |
| Torque Transducers | 2 years | | Outside Source |
| Voltmeters/Multimeters (6.5 digit) | 3 years | | Outside Source |

*The term "calibration" is used to maintain consistency with ASTM D 3740. It is taken to mean "verification."

PRE JOB CALIBRATION VERIFICATION

CALIBRATION CERTIFICATE



APPLICANT FGI HOUSTON

Certificate number FC070040

Page 1 of 1

SUBMITTED A Piezo Cone Penetrometer

Manufacturer Fugro Engineers B.V.

Device type CONE, A15F2.5CKE3SW2/ B, 50 bar

Serial number 1701-1788

The device contains an electronic data sheet which contains, amongst others, the characteristics of all the sensors inside the device. The data acquisition system calculates the measured value from these known characteristics. All calibration results are conform the values specified below.

Force calibration

Calibration reference : 548 FRE.001

Procedure : FEBV.CAL.PRO.003 KALIBRATIE KRACHT

Title of channel(s) : Cone and Cone+Fric.

Max. load 150 kN

| Range | Calibration range From to | Sensitivity Deviation | Zero load output |
|-------------------------|------------------------------|--------------------------|---------------------|
| 1 | 0 25 kN | < 0.5 % | < 0.75 kN |
| Calibration uncertainty | | 0.3 % | 0.008 kN |

Pressure test :

| | |
|--|-------|
| Deviation from specified Alpha factor at 2.5 MPa | < 5 % |
|--|-------|

Cone quality control values :

| | |
|---------------------------------------|-------|
| Max. deviation from reference | < 1 % |
| Max. Tip to Sleeve friction Crosstalk | < 1 % |

Pressure calibration

Calibration reference : 3257-0001

Procedure : FEBV.CAL.PRO.004 KALIBRATIE DRUK

Title of channel : Pore 2

Max. load 30 Mpa

| Range | Calibration range From | Sensitivity Deviation | Zero load output |
|---|---------------------------|--------------------------|---------------------|
| 1 | 0 2.5 MPa | < 1.0 % | < 0.002 MPa |
| Calibration uncertainty | | 0.6 % | 0.003 MPa |
| Pore 2 transducer : Kistler 4043A50 SN : 1233109 | | | |

Calibration of the slope sensor

Calibration reference :

Procedure: FEBV.CAL.PRO.006 KALIBRATIE HELLING

Title of channel : Slope x

| Range | Calibration range From | Sensitivity Deviation | Zero load output |
|-------------------------|---------------------------|--------------------------|---------------------|
| 1 | 0 15 deg | < 10 % | < 1.5 deg |
| Calibration uncertainty | | 1 % | 0.5 deg |

Typical values for this type of device

| | | | | | |
|-----------------------|------|-------------------------|-------|-----------------------------|------|
| Cone diameter (mm) | 43.7 | Pore 2 position | 2 | Alpha factor | 0.58 |
| Cone area (square cm) | 15 | Sleeve length (mm) | 144.7 | Cone - Sleeve distance (mm) | 14.4 |
| Sleeve diameter (mm) | 43.9 | Sleeve area (square cm) | 200 | Cone - Pore 2 distance (mm) | 6.0 |

TRACEABILITY: The measurements have been executed using standards for which the traceability to primary and/or (inter)national standards has been demonstrated.


Calibrated by: Hoogendoorn, Raymond

Approved by: Sinjorgo, Gerry

Calibration date: 16-1-2007

Approval date: 16-1-2007

Calibrate before: 16-1-2008



Calibration Verification Certificate



Device Type: Piezo Cone Penetrometer


Device Number: F7.5CKESW2/V 1701-1788

TIP CALIBRATION

Tip area = 15 cm² = 0.0161 ft²

Tip readings in mV

| Load lb | Load Tons | load/area tsf | Tip mV | TIP Cal Factor Mpa |
|------------|--------------|------------------|-----------|--------------------------|
| 0 | 0 | 0 | 0 | 0 |
| 540 | 0.27 | 16.77019 | 32.1 | 50.0287963 |
| 5110 | 2.555 | 158.6957 | 303.4 | 50.0883413 |
| 7580 | 3.79 | 235.4037 | 449.8 | 50.1165401 |
| 10020 | 5.01 | 311.1801 | 596.3 | 49.9728639 |
| 15020 | 7.51 | 466.4596 | 892.5 | 50.0487269 |
| 20005 | 10.0025 | 621.2733 | 1183.2 | 50.2819048 |




FRICTION CALIBRATION

Sleeve area = 200 cm² = 0.2153 ft²

Friction readings in mV

| Load lb | Load Tons | load/area tsf | Friction mV | Friction Cal Factor Mpa |
|------------|--------------|------------------|----------------|-------------------------------|
| 0 | 0 | 0 | 0 | 0 |
| 620 | 0.31 | 1.439851 | 279.8 | 0.49278463 |
| 1990 | 0.995 | 4.621458 | 884.6 | 0.50028711 |
| 3520 | 1.76 | 8.17464 | 1562.3 | 0.50106193 |
| 5010 | 2.505 | 11.63493 | 2212.2 | 0.5036473 |
| 7510 | 3.755 | 17.44078 | 3308.9 | 0.50474203 |




PORE PRESSURE TRANSDUCER CALIBRATION

Serial : 1233109

Pore Pressure readings in mV

| Pressure psi | Pressure tsf | Transducer readings mV | P.Pres. Cal Factor Mpa |
|-----------------|-----------------|------------------------------|------------------------------|
| 0 | 0 | 0 | 0 |
| 75 | 5.4 | 206.7 | 2.50173026 |
| 150 | 10.8 | 413.6 | 2.50052053 |
| 300 | 21.6 | 824.2 | 2.50962215 |



Temperature Calibration (30 - 115 degrees F)

| Temp (deg F) | TIP (mV) | FRIC (mV) | PIEZO (mV) | Deviation | mV | Mpa | % Full Scale |
|--------------|----------|-----------|------------|-----------|-------|----------|--------------|
| 30 | -0.017 | -0.198 | -0.109 | Tip | 0.274 | 0.0137 | 0.0274 |
| 50 | -0.008 | -0.336 | -0.067 | Friction | 3.185 | 0.001593 | 0.3185 |
| 75 | 0.011 | -0.456 | -0.054 | Piezo | 2.593 | 0.006483 | 0.2593 |
| 100 | 0.132 | 2.307 | -2.192 | | | | |
| 115 | 0.257 | 2.987 | -2.647 | | | | |

TIP CALIBRATED BY GEOTAC (A2LA APPROVED) LOAD CELL:

Model 560K, Serial No. 129739

FRICTION CALIBRATED BY INTERFACE (A2LA APPROVED) LOAD CELL :

Model: 1211EX-10K-B, Serial : 113655

PORE PRESSURE TRANSDUCER CALIBRATED BY GE SENSING (AANSI/NCSL APPROVED)

Pressure Indicator Model: UPS3000CC, Serial : A0813

TEMPERATURE CALIBRATED BY HOUSTON PRECISION TYPE K THERMOCOUPLE (A2LA APPROVED)

Model # 8528-40, Serial # C95005824, ID # TD-001

Calibration Verified by: Dennis Stauffer

Date: 11/5/2007

Checked By : Recep Yilmaz



Date: 11/5/2007

HGL Instrument Verification

DATE: 11/5/2007 Instrument No.: ft Location: Houston LVDT?: ☐ Yes; ☒ No
 Type Analysis: ☒ BF; ☐ NonBF; ☒ FS; ☐ Abs. Reading Resolution of Instrument: 0.000001

| Instrument Identification/Data | | Instrument | Standard | Prediction | Abs. Error | Full Scale |
|---|---|-------------|----------|------------|------------|------------|
| Type | Cone Penetrometer_TIP | mV | lb | lb | lb | Error (%) |
| Manufacturer | Fugro | 0.000000 | 0.0 | -13.5 | 13.5 | -0.04 |
| Model Number | F7.5CKESW2/B1 | 32.100000 | 540.0 | 528.5 | 11.5 | -0.03 |
| Serial Number | 1701-1788 (Tip) | 303.400000 | 5110.0 | 5108.8 | 1.2 | 0.00 |
| HGL Instrument Number | ft | 449.800000 | 7580.0 | 7580.5 | 0.5 | 0.00 |
| Excitation (V) | | 596.300000 | 10020.0 | 10053.8 | 33.8 | 0.10 |
| Gain/Span Setting | NA | 892.500000 | 15020.0 | 15054.5 | 34.5 | 0.10 |
| Full Range Output (V) | | 1183.200000 | 20005.0 | 19962.4 | 42.6 | -0.13 |
| Full Range/Capacity (lb) | 33716 | | | | | |
| Date Verified | 11/5/2007 | | | | | |
| Date Due | 11/4/2008 | | | | | |
| Service Status | In Service | | | | | |
| Accept. Abs. or FS Error (%) | 0.1 <input checked="" type="checkbox"/> FS; <input type="checkbox"/> Abs. | | | | | |
| Verification/Standard Equipment | | | | | | |
| Type | Load Cell (A2AL APPROVED) | | | | | |
| Manufacturer | Geotac | | | | | |
| Model Number | 560K | | | | | |
| Serial Number | 129739 | | | | | |
| HGL Instrument Number | | | | | | |
| Date Verified | | | | | | |
| Temperature | °C = °F | | | | | |
| Linear Regression, Uncertainty, & Error Summary | | | | | | |
| Correlation Coeff. (R ²) | 0.999986252 | | | | | |
| Intercept (lb) | -13.45326099 | | | | | |
| Slope (lb/mV) | 16.88287763 | | | | | |
| Verification (Calib.) Factor | 16.88287763 | | | | | |
| Verification Factor Units | lb/mV | | | | | |
| Absolute Zero (V) | | | | | | |
| Floating Zero (V) | | | | | | |
| Combined Uncertainty (lb) | 6.479058372 | | | | | |
| Coverage Factor | 2 | | | | | |
| Expanded Uncertainty (lb) | 12.95811674 | | | | | |
| Max. Abs. or FS Error (%) | -0.13 <input checked="" type="checkbox"/> FS; <input type="checkbox"/> Abs. | | | | | |

MTS ☐ Yes; ☒ No



| Uncertainty Budget Analysis For ft | | | | | | | |
|---|-------------|-----------------|---------|------|-------------------------------|-----------------------------|----------|
| Source of Uncertainty | Value in lb | Distribution | Divisor | Type | Uncertainty (u _i) | u _i ² | Comments |
| Standard's Uncertainty | | N | 2.0000 | B | | | |
| Abs. Error-STDEV ¹ | 6.4791 | N | 1.0000 | A | 6.4791 | 41.9782 | |
| Resolution of Instrument | 0.0000 | R | 3.4641 | B | 0.0000 | 0.0000 | |
| Repeatability ² | | N | 1.0000 | A | | | |
| Resolution of Standard | #N/A | R | 3.4641 | B | | | |
| Combined Uncertainty | 6.4791 | Coverage Factor | | 2 | for 95% confidence level. | | |
| Expanded Uncertainty (Best Measurement Capability) ³ | 12.958 | | | lb | | | |

(1) This equation follows the approach presented by A2LA, not that typically used in uncertainty calculations; i.e., STDEV of the Mean.

(2) This value is unique for type (model) of equipment.

(3) This uncertainty represents an expanded uncertainty expressed as approximately the 95% confidence level using a coverage factor of k=2.

Verified By: RS

Input By: _____

Reviewed By: _____

Checked By: RS

File: _____



Remarks: _____

HGL Instrument Verification

DATE: 11/5/2007 Instrument No.: ft100 Location: LVDT?: ☐ Yes; ☒ No
 Type Analysis: ☒ BF; ☐ NonBF; ☒ FS; ☐ Abs. Reading Resolution of Instrument: 0.000001

| Instrument Identification/Data | | Instrument mv | Standard lb | Prediction lb | Abs. Error lb | Full Scale Error (%) |
|---|--|---------------|-------------|---------------|---------------|----------------------|
| Type | Cone Penetrometer | | | | | |
| Manufacturer | Fugro | 0.000000 | 0.0 | -13.6 | 13.6 | -0.08 |
| Model Number | F7.5CKESW2/B1 | 279.800000 | 620.0 | 621.9 | 1.9 | 0.01 |
| Serial Number | 1701-1788 (Friction) | 884.600000 | 1990.0 | 1995.4 | 5.4 | 0.03 |
| HGL Instrument Number | ft100 | 1562.300000 | 3520.0 | 3534.5 | 14.5 | 0.09 |
| Excitation (V) | | 2212.200000 | 5010.0 | 5010.5 | 0.5 | 0.00 |
| Gain/Span Setting | NA | 3308.900000 | 7510.0 | 7501.2 | 8.8 | -0.05 |
| Full Range Output (V) | | | | | | |
| Full Range/Capacity (lb) | 16858 | | | | | |
| Date Verified | 11/5/2007 | | | | | |
| Date Due | 11/4/2008 | | | | | |
| Service Status | In Service | | | | | |
| Accept. Abs. or FS Error (%) | 0.1 <input checked="" type="checkbox"/> FS; <input type="checkbox"/> Abs. | | | | | |
| Verification/Standard Equipment | | | | | | |
| Type | Load Cell (A2AL APPROVED) | | | | | |
| Manufacturer | INTERFACE | | | | | |
| Model Number | 1211EX-10K-B | | | | | |
| Serial Number | 113655 | | | | | |
| HGL Instrument Number | | | | | | |
| Date Verified | | | | | | |
| Temperature | °C = °F | | | | | |
| Linear Regression, Uncertainty, & Error Summary | | | | | | |
| Correlation Coeff. (R ²) | 0.999987431 | | | | | |
| Intercept (lb) | -13.55356528 | | | | | |
| Slope (lb/mv) | 2.271068817 | | | | | |
| Verification (Calib.) Factor | 2.27106882 | | | | | |
| Verification Factor Units | lb/mv | | | | | |
| Absolute Zero (V) | | | | | | |
| Floating Zero (V) | | | | | | |
| Combined Uncertainty (lb) | 2.398188297 | | | | | |
| Coverage Factor | 2 | | | | | |
| Expanded Uncertainty (lb) | 4.796376595 | | | | | |
| Max. Abs. or FS Error (%) | 0.09 <input checked="" type="checkbox"/> FS; <input type="checkbox"/> Abs. | | | | | |

MTS ☐ Yes; ☒ No



| Uncertainty Budget Analysis For ft100 | | | | | | | |
|---|-------------|--------------|---------|------|-------------------------------|-----------------------------|---------------------------|
| Source of Uncertainty | Value in lb | Distribution | Divisor | Type | Uncertainty (u _i) | u _i ² | Comments |
| Standard's Uncertainty | | N | 2.0000 | B | | | |
| Abs. Error-STDEV ¹ | 2.3982 | N | 1.0000 | A | 2.3982 | 5.7513 | |
| Resolution of Instrument | 0.0000 | R | 3.4641 | B | 0.0000 | 0.0000 | |
| Repeatability ² | | N | 1.0000 | A | | | |
| Resolution of Standard | #N/A | R | 3.4641 | B | | | |
| Combined Uncertainty | 2.3982 | | | | | | |
| Coverage Factor | 2 | | | | | | for 95% confidence level. |
| Expanded Uncertainty (Best Measurement Capability) ³ | 4.796 | | | lb | | | |

(1) This equation follows the approach presented by A2LA, not that typically used in uncertainty calculations; i.e., STDEV of the Mean.

(2) This value is unique for type (model) of equipment.

(3) This uncertainty represents an expanded uncertainty expressed as approximately the 95% confidence level using a coverage factor of k=2.

Verified By: *[Signature]*

Input By: _____

Reviewed By: _____

Checked By: *[Signature]*



File: _____

Remarks: _____

HGL Instrument Verification

DATE: 11/5/2007 Instrument No.: pt Location: Houston LVDT?: ☐ Yes; ☒ No
 Type Analysis: ☒ BF; ☐ NonBF; ☒ FS; ☐ Abs. Reading Resolution of Instrument: 0.000001

| Instrument Identification/Data | | | | Instrument | Standard | Prediction | Abs. Error | Full Scale |
|---|---------------------------|---|-------------------------------|------------|-------------------------------|--|------------|------------|
| Type | | mv | psi | psi | psi | psi | Error (%) | |
| Manufacturer | Fugro | 0.000000 | 0.0 | -0.2 | 0.2 | -0.06 | | |
| Model Number | F7.5CKESW2/B1 | 206.700000 | 75.0 | 75.0 | 0.0 | 0.01 | | |
| Serial Number | 1701-1788-1233109 (Piezo) | 413.600000 | 150.0 | 150.4 | 0.4 | 0.10 | | |
| HGL Instrument Number | pt | 824.200000 | 300.0 | 299.8 | 0.2 | -0.05 | | |
| Excitation (V) | | | | | | | | |
| Gain/Span Setting | NA | | | | | | | |
| Full Range Output (V) | | | | | | | | |
| Full Range/Capacity (psi) | 360 | | | | | | | |
| Date Verified | 11/5/2007 | | | | | | | |
| Date Due | 11/4/2008 | | | | | | | |
| Service Status | In Service | | | | | | | |
| Accept. Abs. or FS Error (%) | 0.25 | <input checked="" type="checkbox"/> FS; | <input type="checkbox"/> Abs. | | | | | |
| Verification/Standard Equipment | | | | | | | | |
| Type | PT(ANSI/NCSL APPROVED) | | | | | | | |
| Manufacturer | Eaton | | | | | | | |
| Model Number | UPS3000CC | | | | | | | |
| Serial Number | A0813 | | | | | | | |
| HGL Instrument Number | | | | | | | | |
| Date Verified | | | | | | | | |
| Temperature | °C = °F | | | | | | | |
| Linear Regression, Uncertainty, & Error Summary | | | | | | | | |
| Correlation Coeff. (R ²) | 0.99999592 | | | | | | | |
| Intercept (psi) | -0.20331044 | | | | | | | |
| Slope (psi/mv) | 0.364010552 | | | | | | | |
| Verification (Calib.) Factor | 0.36401055 | | | | | | | |
| Verification Factor Units | psi/mv | | | | | | | |
| Absolute Zero (V) | | | | | | | | |
| Floating Zero (V) | | | | | | | | |
| Combined Uncertainty (psi) | 0.064150208 | | | | | | | |
| Coverage Factor | 2 | | | | | | | |
| Expanded Uncertainty (psi) | 0.128300415 | | | | | | | |
| Max. Abs. or FS Error (%) | 0.10 | <input checked="" type="checkbox"/> FS; | <input type="checkbox"/> Abs. | | | | | |
| | | | | MTS | <input type="checkbox"/> Yes; | <input checked="" type="checkbox"/> No | | |



| Uncertainty Budget Analysis For pt | | | | | | | |
|---|--------------|-----------------|---------|------|-------------------------------|-----------------------------|----------|
| Source of Uncertainty | Value in psi | Distribution | Divisor | Type | Uncertainty (u _i) | u _i ² | Comments |
| Standard's Uncertainty | | N | 2.0000 | B | | | |
| Abs. Error-STDEV ¹ | 0.0642 | N | 1.0000 | A | 0.0642 | 0.0041 | |
| Resolution of Instrument | 0.0000 | R | 3.4641 | B | 0.0000 | 0.0000 | |
| Repeatability ² | | N | 1.0000 | A | | | |
| Resolution of Standard | #N/A | R | 3.4641 | B | | | |
| Combined Uncertainty | 0.0642 | Coverage Factor | | 2 | for 95% confidence level. | | |
| Expanded Uncertainty (Best Measurement Capability) ³ | 0.128 | | | psi | | | |

(1) This equation follows the approach presented by A2LA, not that typically used in uncertainty calculations; i.e., STDEV of the Mean.

(2) This value is unique for type (model) of equipment.

(3) This uncertainty represents an expanded uncertainty expressed as approximately the 95% confidence level using a coverage factor of k=2.

Verified By: LS

Input By: _____

Reviewed By: _____

Checked By: Ry

File: _____

Remarks: _____

Calibration Verification Certificate



Device Type: Piezo Cone Penetrometer


Device Number: F7.5CKE3SW2/B 1701-0750

TIP CALIBRATION

Tip area = 15 cm² = 0.0161 ft²

Tip readings in mV

| Load lb | Load Tons | load/area tsf | Tip mV | TIP Cal Factor Mpa |
|------------|--------------|------------------|-----------|--------------------------|
| 0 | 0 | 0 | 0 | 0 |
| 530 | 0.265 | 16.45963 | 31.5 | 50.0376197 |
| 5010 | 2.505 | 155.5901 | 298.1 | 49.9812458 |
| 7490 | 3.745 | 232.6087 | 445.2 | 50.0331664 |
| 10020 | 5.01 | 311.1801 | 595.9 | 50.0064083 |
| 15050 | 7.525 | 467.3913 | 895.2 | 49.9974383 |
| 19910 | 9.955 | 618.323 | 1184.3 | 49.9966444 |




FRICITION CALIBRATION

Sleeve area = 200 cm² = 0.2153 ft²

Friction readings in mV

| Load lb | Load Tons | load/area tsf | Friction mV | Friction Cal Factor Mpa |
|------------|--------------|------------------|----------------|-------------------------------|
| 0 | 0 | 0 | 0 | 0 |
| 520 | 0.26 | 1.207617 | 231.3 | 0.49996647 |
| 2490 | 1.245 | 5.782629 | 1107.3 | 0.50008891 |
| 3750 | 1.875 | 8.708778 | 1665.6 | 0.50069555 |
| 5010 | 2.505 | 11.63493 | 2228.3 | 0.50000833 |
| 7510 | 3.755 | 17.44078 | 3340.2 | 0.50001224 |




PORE PRESSURE TRANSDUCER CALIBRATION

Serial : 8091223

Pore Pressure readings in mV

| Pressure psi | Pressure tsf | Transducer readings mV | P.Pres. Cal Factor Mpa |
|-----------------|-----------------|------------------------------|------------------------------|
| 0 | 0 | 0 | 0 |
| 75 | 5.4 | 206.8 | 2.50052053 |
| 150 | 10.8 | 413.5 | 2.50112525 |
| 300 | 21.6 | 825.7 | 2.50506307 |



Temperature Calibration (30 - 115 degrees F)

Temp (deg TIP (mV) FRIC (mV) PIEZO (mV)

| | | | |
|-----|-------|-------|--------|
| 30 | 0.946 | 0.783 | -0.132 |
| 50 | 0.983 | 0.974 | -0.077 |
| 75 | 1.012 | 1.114 | 0.264 |
| 100 | 1.297 | 1.267 | 0.357 |
| 115 | 1.324 | 1.771 | 0.992 |

| Deviation | mV | Mpa | % Full Scale |
|-----------|-------|----------|--------------|
| Tip | 0.378 | 0.0189 | 0.0378 |
| Friction | 0.988 | 0.000494 | 0.0988 |
| Piezo | 1.124 | 0.00281 | 0.1124 |

TIP CALIBRATED BY GEOTAC (A2LA APPROVED) LOAD CELL:

Model 560K, Serial No. 129739

FRICITION CALIBRATED BY INTERFACE (A2LA APPROVED) LOAD CELL :

Model: 1211EX-10K-B, Serial : 113655

PORE PRESSURE TRANSDUCER CALIBRATED BY GE SENSING (AANSI/NCSL APPROVED)

Pressure Indicator Model: UPS3000CC, Serial : A0813

TEMPERATURE CALIBRATED BY HOUSTON PRECISION TYPE K THERMOCOUPLE (A2LA APPROVED)

Model # 8528-40, Serial # C95005824, ID # TD-001

Calibration Verified by: Dennis Stauffer

Date: 4/23/2008

Checked By : Recep Yilmaz



Date: 4/23/2008

HGL Instrument Verification

DATE: 4/23/2008 Instrument No.: ft100 Location: Houston LVDT?: ☐ Yes; ☒ No
 Type Analysis: ☒ BF; ☐ NonBF; ☒ FS; ☐ Abs. Reading Resolution of Instrument: 0.000001

| Instrument Identification/Data | | Instrument mV | Standard lb | Prediction lb | Abs. Error lb | Full Scale Error (%) |
|---|---|---------------|-------------|---------------|---------------|----------------------|
| Type | Cone Penetrometer-TIP | | | | | |
| Manufacturer | Fugro | 0.000000 | 0.0 | 0.9 | 0.9 | 0.01 |
| Model Number | F7.5CKEW2/B | 31.500000 | 530.0 | 530.4 | 0.4 | 0.00 |
| Serial Number | 1701-0750 | 298.100000 | 5010.0 | 5012.5 | 2.5 | 0.01 |
| HGL Instrument Number | ft100 | 445.200000 | 7490.0 | 7485.5 | 4.5 | -0.03 |
| Excitation (V) | | 595.900000 | 10020.0 | 10019.0 | 1.0 | -0.01 |
| Gain/Span Setting | NA | 895.200000 | 15050.0 | 15050.7 | 0.7 | 0.00 |
| Full Range Output (V) | | 1184.300000 | 19910.0 | 19911.0 | 1.0 | 0.01 |
| Full Range/Capacity (lb) | 16858 | | | | | |
| Date Verified | 4/23/2008 | | | | | |
| Date Due | 4/23/2009 | | | | | |
| Service Status | In Service | | | | | |
| Accept. Abs. or FS Error (%) | 0.1 <input checked="" type="checkbox"/> FS; <input type="checkbox"/> Abs. | | | | | |
| Verification/Standard Equipment | | | | | | |
| Type | Load cell (A2LA APPROVED) | | | | | |
| Manufacturer | GEOTAC | | | | | |
| Model Number | 560K | | | | | |
| Serial Number | 129739 | | | | | |
| HGL Instrument Number | | | | | | |
| Date Verified | | | | | | |
| Temperature | °C = °F | | | | | |
| Linear Regression, Uncertainty, & Error Summary | | | | | | |
| Correlation Coeff. (R ²) | 0.999999907 | | | | | |
| Intercept (lb) | 0.877122495 | | | | | |
| Slope (lb/mV) | 16.81173849 | | | | | |
| Verification (Calib.) Factor | 16.81173849 | | | | | |
| Verification Factor Units | lb/mV | | | | | |
| Absolute Zero (V) | | | | | | |
| Floating Zero (V) | | | | | | |
| Combined Uncertainty (lb) | 0.54832333 | | | | | |
| Coverage Factor | 2 | | | | | |
| Expanded Uncertainty (lb) | 1.09664666 | | | | | |
| Max. Abs. or FS Error (%) | -0.03 <input checked="" type="checkbox"/> FS; <input type="checkbox"/> Abs. | | | | | |

MTS ☐ Yes; ☒ No



| Uncertainty Budget Analysis For ft100 | | | | | | | |
|---|-------------|--------------|-----------------|------|-------------------------------|-----------------------------|---------------------------|
| Source of Uncertainty | Value in lb | Distribution | Divisor | Type | Uncertainty (u _i) | u _i ² | Comments |
| Standard's Uncertainty | | N | 2.0000 | B | | | |
| Abs. Error-STDEV ¹ | 0.5483 | N | 1.0000 | A | 0.5483 | 0.3007 | |
| Resolution of Instrument | 0.0000 | R | 3.4641 | B | 0.0000 | 0.0000 | |
| Repeatability ² | | N | 1.0000 | A | | | |
| Resolution of Standard | #N/A | R | 3.4641 | B | | | |
| Combined Uncertainty | 0.5483 | | | | | | |
| Expanded Uncertainty (Best Measurement Capability) ³ | | | Coverage Factor | 2 | | | for 95% confidence level. |
| | | | 1.097 | lb | | | |

(1) This equation follows the approach presented by A2LA, not that typically used in uncertainty calculations; i.e., STDEV of the Mean.

(2) This value is unique for type (model) of equipment.

(3) This uncertainty represents an expanded uncertainty expressed as approximately the 95% confidence level using a coverage factor of k=2.

Verified By: MS

Input By: _____

Reviewed By: _____

Checked By: Ry

File: _____



Remarks: _____

HGL Instrument Verification

DATE: 4/23/2008 Instrument No.: ft100 Location: LVDT?: ☐ Yes; ☒ No
 Type Analysis: ☒ BF; ☐ NonBF; ☒ FS; ☐ Abs. Reading Resolution of Instrument: 0.000001

| Instrument Identification/Data | | | | Instrument mv | Standard lb | Prediction lb | Abs. Error lb | Full Scale Error (%) |
|---|---------------------------|---|-------------------------------|------------------|----------------|------------------|------------------|-------------------------|
| Type | Cone Penetrometer | | | | | | | |
| Manufacturer | Fugro | | | 0.000000 | 0.0 | 0.4 | 0.4 | 0.00 |
| Model Number | F7.5CKE3SW2/B | | | 231.300000 | 520.0 | 520.5 | 0.5 | 0.00 |
| Serial Number | 1701-0750 | | | 1107.300000 | 2490.0 | 2490.1 | 0.1 | 0.00 |
| HGL Instrument Number | ft100 | | | 1666.600000 | 3750.0 | 3747.7 | 2.3 | -0.01 |
| Excitation (V) | | | | 2228.300000 | 5010.0 | 5010.6 | 0.6 | 0.00 |
| Gain/Span Setting | NA | | | 3340.200000 | 7510.0 | 7510.7 | 0.7 | 0.00 |
| Full Range Output (V) | | | | | | | | |
| Full Range/Capacity (lb) | 16858 | | | | | | | |
| Date Verified | 4/23/2008 | | | | | | | |
| Date Due | 4/23/2009 | | | | | | | |
| Service Status | In Service | | | | | | | |
| Accept. Abs. or FS Error (%) | 0.1 | <input checked="" type="checkbox"/> FS; | <input type="checkbox"/> Abs. | | | | | |
| Verification/Standard Equipment | | | | | | | | |
| Type | Load Cell (A2AL APPROVED) | | | | | | | |
| Manufacturer | INTERFACE | | | | | | | |
| Model Number | 1211EX-10K-B | | | | | | | |
| Serial Number | 113655 | | | | | | | |
| HGL Instrument Number | | | | | | | | |
| Date Verified | | | | | | | | |
| Temperature | °C = °F | | | | | | | |
| Linear Regression, Uncertainty, & Error Summary | | | | | | | | |
| Correlation Coeff. (R ²) | 0.99999834 | | | | | | | |
| Intercept (lb) | 0.421964823 | | | | | | | |
| Slope (lb/mv) | 2.248442121 | | | | | | | |
| Verification (Calib.) Factor | 2.24844212 | | | | | | | |
| Verification Factor Units | lb/mv | | | | | | | |
| Absolute Zero (V) | | | | | | | | |
| Floating Zero (V) | | | | | | | | |
| Combined Uncertainty (lb) | 0.319825144 | | | | | | | |
| Coverage Factor | 2 | | | | | | | |
| Expanded Uncertainty (lb) | 0.639650288 | | | | | | | |
| Max. Abs. or FS Error (%) | -0.01 | <input checked="" type="checkbox"/> FS; | <input type="checkbox"/> Abs. | | | | | |

MTS ☐ Yes; ☒ No



| Uncertainty Budget Analysis For ft100 | | | | | | | |
|---|-------------|-----------------|---------|------|-------------------------------|-----------------------------|----------|
| Source of Uncertainty | Value in lb | Distribution | Divisor | Type | Uncertainty (u _i) | u _i ² | Comments |
| Standard's Uncertainty | | N | 2.0000 | B | | | |
| Abs. Error-STDEV ¹ | 0.3198 | N | 1.0000 | A | 0.3198 | 0.1023 | |
| Resolution of Instrument | 0.0000 | R | 3.4641 | B | 0.0000 | 0.0000 | |
| Repeatability ² | | N | 1.0000 | A | | | |
| Resolution of Standard | #N/A | R | 3.4641 | B | | | |
| Combined Uncertainty | 0.3198 | Coverage Factor | | 2 | for 95% confidence level. | | |
| Expanded Uncertainty (Best Measurement Capability) ³ | 0.640 | | | lb | | | |

(1) This equation follows the approach presented by A2LA, not that typically used in uncertainty calculations; i.e., STDEV of the Mean.

(2) This value is unique for type (model) of equipment.

(3) This uncertainty represents an expanded uncertainty expressed as approximately the 95% confidence level using a coverage factor of k=2.

Verified By: MS

Input By: _____

Reviewed By: _____

Checked By: Ry



File: _____

Remarks: _____

HGL Instrument Verification

DATE: 4/23/2008 Instrument No.: pt Location: Houston LVDT?: ☐ Yes; ☒ No
 Type Analysis: ☒ BF; ☐ NonBF; ☒ FS; ☐ Abs. Reading Resolution of Instrument: 0.000001

| Instrument Identification/Data | | | | Instrument | Standard | Prediction | Abs. Error | Full Scale |
|---|------------------------|---|-------------------------------|------------|-------------------------------|--|------------|------------|
| Type | Cone Penetrometer | | | mv | psi | psi | psi | Error (%) |
| Manufacturer | Fugro | | | 0.000000 | 0.0 | -0.1 | 0.1 | -0.03 |
| Model Number | F7.5CKEW2/B | | | 206.800000 | 75.0 | 75.0 | 0.0 | 0.01 |
| Serial Number | 1701-0750-8091223 | | | 413.500000 | 150.0 | 150.1 | 0.1 | 0.04 |
| HGL Instrument Number | pt | | | 825.700000 | 300.0 | 299.9 | 0.1 | -0.02 |
| Excitation (V) | | | | | | | | |
| Gain/Span Setting | NA | | | | | | | |
| Full Range Output (V) | | | | | | | | |
| Full Range/Capacity (psi) | 360 | | | | | | | |
| Date Verified | 4/23/2008 | | | | | | | |
| Date Due | 4/23/2009 | | | | | | | |
| Service Status | In Service | | | | | | | |
| Accept. Abs. or FS Error (%) | 0.25 | <input checked="" type="checkbox"/> FS; | <input type="checkbox"/> Abs. | | | | | |
| Verification/Standard Equipment | | | | | | | | |
| Type | PT(ANSI/NCSL APPROVED) | | | | | | | |
| Manufacturer | Eaton | | | | | | | |
| Model Number | UPS3000CC | | | | | | | |
| Serial Number | A0813 | | | | | | | |
| HGL Instrument Number | | | | | | | | |
| Date Verified | | | | | | | | |
| Temperature | °C = °F | | | | | | | |
| Linear Regression, Uncertainty, & Error Summary | | | | | | | | |
| Correlation Coeff. (R ²) | 0.999999198 | | | | | | | |
| Intercept (psi) | -0.101633254 | | | | | | | |
| Slope (psi/mv) | 0.363351683 | | | | | | | |
| Verification (Calib.) Factor | 0.36335168 | | | | | | | |
| Verification Factor Units | psi/mv | | | | | | | |
| Absolute Zero (V) | | | | | | | | |
| Floating Zero (V) | | | | | | | | |
| Combined Uncertainty (psi) | 0.021757343 | | | | | | | |
| Coverage Factor | 2 | | | | | | | |
| Expanded Uncertainty (psi) | 0.043514685 | | | | | | | |
| Max. Abs. or FS Error (%) | 0.04 | <input checked="" type="checkbox"/> FS; | <input type="checkbox"/> Abs. | | | | | |
| | | | | MTS | <input type="checkbox"/> Yes; | <input checked="" type="checkbox"/> No | | |



| Uncertainty Budget Analysis For pt | | | | | | | |
|---|--------------|--------------|---------|------|-------------------------------|-----------------------------|---------------------------|
| Source of Uncertainty | Value in psi | Distribution | Divisor | Type | Uncertainty (u _i) | u _i ² | Comments |
| Standard's Uncertainty | | N | 2.0000 | B | | | |
| Abs. Error-STDEV ¹ | 0.0218 | N | 1.0000 | A | 0.0218 | 0.0005 | |
| Resolution of Instrument | 0.0000 | R | 3.4641 | B | 0.0000 | 0.0000 | |
| Repeatability ² | | N | 1.0000 | A | | | |
| Resolution of Standard | #N/A | R | 3.4641 | B | | | |
| Combined Uncertainty | 0.0218 | | | | | | |
| Coverage Factor | | | | | 2 | | for 95% confidence level. |
| Expanded Uncertainty (Best Measurement Capability) ³ | 0.044 | | | psi | | | |

(1) This equation follows the approach presented by A2LA, not that typically used in uncertainty calculations; i.e., STDEV of the Mean.

(2) This value is unique for type (model) of equipment.

(3) This uncertainty represents an expanded uncertainty expressed as approximately the 95% confidence level using a coverage factor of k=2.

Verified By: RS

Input By: _____

Reviewed By: _____

Checked By: Ry

File: _____

Remarks: _____

Calibration Verification Certificate



Device Type: Seismograph

Device Manufacturer Geometrics, Inc.


Model Number: ES-3000

Serial Number: 5138

Frequency (Hz)

Input Seismograph

| | |
|---------|------|
| 10.055 | 9.9 |
| 20.037 | 19.9 |
| 30.071 | 29.3 |
| 40.040 | 40.1 |
| 50.061 | 50.6 |
| 60.048 | 60.3 |
| 70.005 | 70.2 |
| 80.011 | 79.9 |
| 90.090 | 90.1 |
| 100.014 | 99.8 |



FREQUENCY CALIBRATED BY INSTEK GOOD WILL INSTRUMENTS (A2LA APPROVED) FREQUENCY COUNTER :

Model: GFC-80101H, Serial No. CF871549

FREQUENCY GENERATED BY EZ DIGITAL, INC (A2LA APPROVED) OSCILLISCOPE WITH BUILT IN FUNCTION GENERATOR


Model: OS-5020G, Serial No.: 3080209

Calibration Verified by: Dennis Stauffer *DS*

Date: 1/22/2008

Checked By : Recep Yilmaz *RY*


Date: 1/22/2008



Geometrics ES-3000 S/N 5138

Frequency (HZ) In: 20.037 Seismograph: 19.9


Date: 1/22/2008



Geometrics ES-3000 S/N 5138

Frequency (HZ) In: 30.071 Seismograph: 29.3


Date: 1/22/2008



Geometrics ES-3000 S/N 5138

Frequency (HZ) In: 40.040 Seismograph: 40.1


Date: 1/22/2008



Geometrics ES-3000 S/N 5138

Frequency (HZ) In: 50.061 Seismograph: 50.6


Date: 1/22/2008



Geometrics ES-3000 S/N 5138

Frequency (HZ) In: 60.048 Seismograph: 60.3


Date: 1/22/2008



Geometrics ES-3000 S/N 5138

Frequency (HZ) In: 70.005 Seismograph: 70.2

Date: 1/22/2008




Geometrics ES-3000 S/N 5138

Frequency (HZ) In: 80.011 Seismograph: 79.9

Date: 1/22/2008


5.0



Geometrics ES-3000 S/N 5138

Frequency (HZ) In: 90.090 Seismograph: 90.1


Date: 1/22/2008



Geometrics ES-3000 S/N 5138

Frequency (HZ) In: 100.014 Seismograph: 99.8

Date: 1/22/2008



POST JOB CALIBRATION VERIFICATION

Calibration Verification Certificate



Device Type: Piezo Cone Penetrometer


Device Number: F7.5CKE3SW2/B 1701-0750

TIP CALIBRATION

Tip area = 15 cm² = 0.0161 ft²

Tip readings in mV

| Load | Load | load/area | Tip | Cal Factor |
|-------|--------|-----------|--------|------------|
| lb | Tons | tsf | mV | Mpa |
| 0 | 0 | 0 | 0 | 0 |
| 540 | 0.27 | 16.77019 | 32.1 | 50.0287963 |
| 5015 | 2.5075 | 155.7453 | 298.3 | 49.9975831 |
| 7510 | 3.755 | 233.2298 | 446.2 | 50.0543352 |
| 10005 | 5.0025 | 310.7143 | 594.3 | 50.0659763 |
| 15010 | 7.505 | 466.1491 | 892.6 | 50.0098021 |
| 19980 | 9.99 | 620.4969 | 1186.6 | 50.0751739 |




FRICTION CALIBRATION

Sleeve area = 200 cm² = 0.2153 ft²

Friction readings in mV

| Load | Load | load/area | Friction | Cal Factor |
|------|--------|-----------|----------|------------|
| lb | Tons | tsf | mV | Mpa |
| 0 | 0 | 0 | 0 | 0 |
| 500 | 0.25 | 1.16117 | 220.3 | 0.50474111 |
| 2520 | 1.26 | 5.852299 | 1110.6 | 0.50461022 |
| 3755 | 1.8775 | 8.72039 | 1666.2 | 0.5011826 |
| 4990 | 2.495 | 11.58848 | 2215.9 | 0.50079912 |
| 7505 | 3.7525 | 17.42917 | 3335.8 | 0.50033843 |




PORE PRESSURE TRANSDUCER CALIBRATION

Serial : 8091223

Pore Pressure readings in mV

| Pressure | Pressure | Transducer | P.Pres. |
|----------|----------|------------|------------|
| psi | tsf | mV | Cal Factor |
| 0 | 0 | 0 | 0 |
| 75 | 5.4 | 206.6 | 2.50294116 |
| 150 | 10.8 | 413.3 | 2.50233557 |
| 300 | 21.6 | 823.4 | 2.51206045 |



Temperature Calibration (30 - 115 degrees F)

| Temp (deg TIP (mV) | FRIC (mV) PIEZO (mV) | Deviation | mV | Mpa | % Full Scale |
|--------------------|----------------------|-----------|--------|----------|--------------|
| 30 | 0.957 | 0.788 | -0.124 | Tip | 0.395 |
| 50 | 0.994 | 0.957 | -0.065 | Friction | 0.000482 |
| 75 | 1.054 | 1.037 | 0.258 | Piezo | 0.002843 |
| 100 | 1.284 | 1.236 | 0.369 | | 0.1137 |
| 115 | 1.352 | 1.752 | 1.013 | | |

TIP CALIBRATED BY GEOTAC (A2LA APPROVED) LOAD CELL:

Model 560K, Serial No. 129739

FRICTION CALIBRATED BY INTERFACE (A2LA APPROVED) LOAD CELL :

Model 560K, Serial No. 129739

PORE PRESSURE TRANSDUCER CALIBRATED BY GE SENSING (AANSI/NCSL APPROVED)

Pressure Indicator Model: UPS3000CC, Serial : A0813

TEMPERATURE CALIBRATED BY HOUSTON PRECISION TYPE K THERMOCOUPLE (A2LA APPROVED)

Model # 8528-40, Serial # C95005824, ID # TD-001

Calibration Verified by: Dennis Stauffer

Date: 6/2/2008



Checked By : Recep Yilmaz

Date: 6/2/2008

HGL Instrument Verification

DATE: 6/2/2008 Instrument No.: #100 Location: Houston LVDT?: ☐ Yes; ☒ No
 Type Analysis: ☒ BF; ☐ NonBF; ☒ FS; ☐ Abs. Reading Resolution of Instrument: 0.000001

| Instrument Identification/Data | | Instrument | Standard | Prediction | Abs. Error | Full Scale |
|---|---------------------------|---|--|------------|------------|------------|
| Type | Cone Penetrometer-TIP | mV | lb | lb | lb | Error (%) |
| Manufacturer | Fugro | 0.000000 | 0.0 | 1.3 | 1.3 | 0.01 |
| Model Number | F7.5CKEW2/B | 32.100000 | 530.0 | 541.0 | 11.0 | 0.07 |
| Serial Number | 1701-0750 | 298.300000 | 5010.0 | 5016.1 | 6.1 | 0.04 |
| HGL Instrument Number | #100 | 446.200000 | 7490.0 | 7502.5 | 12.5 | 0.07 |
| Excitation (V) | | 594.300000 | 10020.0 | 9992.3 | 27.7 | -0.16 |
| Gain/Span Setting | NA | 892.600000 | 15050.0 | 15007.1 | 42.9 | -0.25 |
| Full Range Output (V) | | 1186.600000 | 19910.0 | 19949.6 | 39.6 | 0.23 |
| Full Range/Capacity (lb) | 16858 | | | | | |
| Date Verified | 6/2/2008 | | | | | |
| Date Due | 6/2/2009 | | | | | |
| Service Status | In Service | | | | | |
| Accept. Abs. or FS Error (%) | 0.1 | <input checked="" type="checkbox"/> FS; <input type="checkbox"/> Abs. | | | | |
| Verification/Standard Equipment | | | | | | |
| Type | Load cell (A2LA APPROVED) | | | | | |
| Manufacturer | Geotac | | | | | |
| Model Number | 560K | | | | | |
| Serial Number | 129739 | | | | | |
| HGL Instrument Number | | | | | | |
| Date Verified | | | | | | |
| Temperature | °C = °F | | | | | |
| Linear Regression, Uncertainty, & Error Summary | | | | | | |
| Correlation Coeff. (R ²) | 0.99986128 | | | | | |
| Intercept (lb) | 1.347403615 | | | | | |
| Slope (lb/mV) | 16.81127161 | | | | | |
| Verification (Calib.) Factor | 16.81127161 | | | | | |
| Verification Factor Units | lb/mV | | | | | |
| Absolute Zero (V) | | | | | | |
| Floating Zero (V) | | | | | | |
| Combined Uncertainty (lb) | 6.258542548 | | | | | |
| Coverage Factor | 2 | | | | | |
| Expanded Uncertainty (lb) | 12.5170851 | | | | | |
| Max. Abs. or FS Error (%) | -0.25 | <input checked="" type="checkbox"/> FS; <input type="checkbox"/> Abs. | | | | |
| | | MTS | <input type="checkbox"/> Yes; <input checked="" type="checkbox"/> No | | | |



| Uncertainty Budget Analysis For #100 | | | | | | | |
|---|-------------|--------------|---------|------|-------------------------------|-----------------------------|----------|
| Source of Uncertainty | Value in lb | Distribution | Divisor | Type | Uncertainty (u _i) | u _i ² | Comments |
| Standard's Uncertainty | | N | 2.0000 | B | | | |
| Abs. Error-STDEV ¹ | 6.2585 | N | 1.0000 | A | 6.2585 | 39.1694 | |
| Resolution of Instrument | 0.0000 | R | 3.4641 | B | 0.0000 | 0.0000 | |
| Repeatability ² | | N | 1.0000 | A | | | |
| Resolution of Standard | #N/A | R | 3.4641 | B | | | |
| Combined Uncertainty | 6.2585 | | | | | | |
| Expanded Uncertainty (Best Measurement Capability) ³ | 12.517 | | | | | | |
| Coverage Factor | | | | 2 | for 95% confidence level. | | |

(1) This equation follows the approach presented by A2LA, not that typically used in uncertainty calculations; i.e., STDEV of the Mean.

(2) This value is unique for type (model) of equipment.

(3) This uncertainty represents an expanded uncertainty expressed as approximately the 95% confidence level using a coverage factor of k=2.

Verified By: [Signature]

Input By: _____

Reviewed By: _____

Checked By: [Signature]



Remarks:

HGL Instrument Verification

DATE: 6/2/2008 Instrument No.: ft100 Location: LVDT?: ☐ Yes; ☒ No
 Type Analysis: ☒ BF; ☐ NonBF; ☒ FS; ☐ Abs. Reading Resolution of Instrument: 0.000001

| Instrument Identification/Data | | Instrument | Standard | Prediction | Abs. Error | Full Scale |
|---|---|-------------|----------|------------|------------|------------|
| Type | Cone Penetrometer | mv | lb | lb | lb | Error (%) |
| Manufacturer | Fugro | 0.000000 | 0.0 | 7.3 | 7.3 | 0.04 |
| Model Number | F7.5CKE3SW2/B | 220.300000 | 520.0 | 503.0 | 17.0 | -0.10 |
| Serial Number | 1701-0750 | 1110.600000 | 2490.0 | 2506.3 | 16.3 | 0.10 |
| HGL Instrument Number | ft100 | 1666.200000 | 3750.0 | 3756.5 | 6.5 | 0.04 |
| Excitation (V) | | 2215.900000 | 5010.0 | 4993.4 | 16.6 | -0.10 |
| Gain/Span Setting | NA | 3335.800000 | 7510.0 | 7513.4 | 3.4 | 0.02 |
| Full Range Output (V) | | | | | | |
| Full Range Capacity (lb) | 16858 | | | | | |
| Date Verified | 6/2/2008 | | | | | |
| Date Due | 6/2/2009 | | | | | |
| Service Status | In Service | | | | | |
| Accept. Abs. or FS Error (%) | 0.1 <input checked="" type="checkbox"/> FS; <input type="checkbox"/> Abs. | | | | | |
| Verification/Standard Equipment | | | | | | |
| Type | Load Cell (A2AL APPROVED) | | | | | |
| Manufacturer | Geolac | | | | | |
| Model Number | 560K | | | | | |
| Serial Number | 129739 | | | | | |
| HGL Instrument Number | | | | | | |
| Date Verified | | | | | | |
| Temperature | °C = °F | | | | | |
| Linear Regression, Uncertainty, & Error Summary | | | | | | |
| Correlation Coeff. (R ²) | 0.999976621 | | | | | |
| Intercept (lb) | 7.280791244 | | | | | |
| Slope (lb/mv) | 2.250177247 | | | | | |
| Verification (Calib.) Factor | 2.25017725 | | | | | |
| Verification Factor Units | lb/mv | | | | | |
| Absolute Zero (V) | | | | | | |
| Floating Zero (V) | | | | | | |
| Combined Uncertainty (lb) | 2.492195376 | | | | | |
| Coverage Factor | 2 | | | | | |
| Expanded Uncertainty (lb) | 4.984390751 | | | | | |
| Max. Abs. or FS Error (%) | -0.10 <input checked="" type="checkbox"/> FS; <input type="checkbox"/> Abs. | | | | | |

MTS ☐ Yes; ☒ No



| Uncertainty Budget Analysis For ft100 | | | | | | | |
|---|-------------|--------------|-----------------|------|-------------------------------|-----------------------------|--------------------------|
| Source of Uncertainty | Value in lb | Distribution | Divisor | Type | Uncertainty (u _i) | u _i ² | Comments |
| Standard's Uncertainty | | N | 2.0000 | B | | | |
| Abs. Error-STDEV ¹ | 2.4922 | N | 1.0000 | A | 2.4922 | 6.2110 | |
| Resolution of Instrument | 0.0000 | R | 3.4641 | B | 0.0000 | 0.0000 | |
| Repeatability ² | | N | 1.0000 | A | | | |
| Resolution of Standard | #N/A | R | 3.4641 | B | | | |
| Combined Uncertainty | 2.4922 | | Coverage Factor | 2 | | | for 95% confidence level |
| Expanded Uncertainty (Best Measurement Capability) ³ | 4.984 | | | lb | | | |

(1) This equation follows the approach presented by A2LA, not that typically used in uncertainty calculations; i.e., STDEV of the Mean.

(2) This value is unique for type (model) of equipment.

(3) This uncertainty represents an expanded uncertainty expressed as approximately the 95% confidence level using a coverage factor of k=2.

Verified By:

Input By: _____

Reviewed By: _____

Checked By:

File: _____



Remarks: _____

HGL Instrument Verification

DATE: 6/2/2008 Instrument No.: pt Location: Houston LVDT?: ☐ Yes; ☒ No
 Type Analysis: ☒ BF; ☐ NonBF; ☒ FS; ☐ Abs. Reading Resolution of Instrument: 0.000001

| Instrument Identification/Data | | Instrument | Standard | Prediction | Abs. Error | Full Scale |
|---|--|------------|----------|------------|------------|------------|
| Type | Cone Penetrometer - Piezo | mv | psi | psi | psi | Error (%) |
| Manufacturer | Fugro | 0.000000 | 0.0 | -0.2 | 0.2 | -0.06 |
| Model Number | F7.5CKEW2/B | 206.600000 | 75.0 | 75.1 | 0.1 | 0.02 |
| Serial Number | 1701-0750-8091223 | 413.300000 | 150.0 | 150.4 | 0.4 | 0.10 |
| HGL Instrument Number | pt | 823.400000 | 300.0 | 299.8 | 0.2 | -0.06 |
| Excitation (V) | | | | | | |
| Gain/Span Setting | NA | | | | | |
| Full Range Output (V) | | | | | | |
| Full Range/Capacity (psi) | 360 | | | | | |
| Date Verified | 6/2/2008 | | | | | |
| Date Due | 6/2/2009 | | | | | |
| Service Status | In Service | | | | | |
| Accept. Abs. or FS Error (%) | 0.25 <input checked="" type="checkbox"/> FS; <input type="checkbox"/> Abs. | | | | | |
| Verification/Standard Equipment | | | | | | |
| Type | PT(ANSI/NCSL APPROVED) | | | | | |
| Manufacturer | Eaton | | | | | |
| Model Number | UPS3000CC | | | | | |
| Serial Number | A0813 | | | | | |
| HGL Instrument Number | | | | | | |
| Date Verified | | | | | | |
| Temperature | °C = °F | | | | | |
| Linear Regression, Uncertainty, & Error Summary | | | | | | |
| Correlation Coeff. (R ²) | 0.999995309 | | | | | |
| Intercept (psi) | -0.225296318 | | | | | |
| Slope (psi/mv) | 0.364374132 | | | | | |
| Verification (Calib.) Factor | 0.36437413 | | | | | |
| Verification Factor Units | psi/mv | | | | | |
| Absolute Zero (V) | | | | | | |
| Floating Zero (V) | | | | | | |
| Combined Uncertainty (psi) | 0.064742646 | | | | | |
| Coverage Factor | 2 | | | | | |
| Expanded Uncertainty (psi) | 0.129485291 | | | | | |
| Max. Abs. or FS Error (%) | 0.10 <input checked="" type="checkbox"/> FS; <input type="checkbox"/> Abs. | | | | | |

MTS: ☐ Yes; ☒ No



| Uncertainty Budget Analysis For pt | | | | | | | |
|---|--------------|--------------|---------|------|-------------------------------|-----------------------------|---------------------------|
| Source of Uncertainty | Value in psi | Distribution | Divisor | Type | Uncertainty (u _i) | u _i ² | Comments |
| Standard's Uncertainty | | N | 2.0000 | B | | | |
| Abs. Error-STDEV ¹ | 0.0647 | N | 1.0000 | A | 0.0647 | 0.0042 | |
| Resolution of Instrument | 0.0000 | R | 3.4641 | B | 0.0000 | 0.0000 | |
| Repeatability ² | | N | 1.0000 | A | | | |
| Resolution of Standard | #N/A | R | 3.4641 | B | | | |
| Combined Uncertainty | 0.0647 | | | | | | |
| Expanded Uncertainty (Best Measurement Capability) ³ | 0.129 | | | psi | | | for 95% confidence level. |

(1) This equation follows the approach presented by A2LA, not that typically used in uncertainty calculations; i.e., STDEV of the Mean.

(2) This value is unique for type (model) of equipment.

(3) This uncertainty represents an expanded uncertainty expressed as approximately the 95% confidence level using a coverage factor of k=2.

Verified By: RS Input By: _____ Reviewed By: _____ Checked By: RJ
 File: _____
 Remarks: _____

**FINAL DATA REPORT Rev. 2
GEOTECHNICAL EXPLORATION AND TESTING**

**TURKEY POINT COL PROJECT
FLORIDA CITY, FLORIDA**

October 6, 2008

**VOLUME 2
Appendix D – Geovision Downhole and P-S Logging Report**

Prepared By:

**MACTEC Engineering and Consulting, Inc.
Raleigh, North Carolina**

MACTEC Project No. 6468-07-1950

Prepared For:

**Bechtel Power Corporation
Subcontract No. 25409-102-HC4-CY00-00001**



**DOCUMENTATION OF TECHNICAL REVIEW
SUBCONTRACTOR WORK PRODUCT**

Project Name: Turkey Point COL

Project Number: 6468-07-1950

Project Manager: Scott Auger

Project Principals: Al Tice and Tom McDaniel

The report described below has been prepared by the named subcontractor retained in accordance with the MACTEC QAPD. The work and report have been reviewed by a MACTEC technically qualified person. Comments on the work or report, if any, have been satisfactorily addressed by the subcontractor. The attached report is approved in accordance with section QS-7 of MACTEC's QAPD

The information and data contained in the attached report are hereby released by MACTEC for project use.

REPORT : Report Boring Geophysical Logging. Report 8083-03 rev 0, July 24, 2008_

SUBCONTRACTOR: GeoVision

DATE OF ACCEPTANCE :_ 7-25-08

TECHNICAL REVIEWER:

PROJECT PRINCIPAL

[Handwritten signature]
[Handwritten signature] 7-25-08

DCN TUR - 483



3301 Atlantic Avenue, Raleigh, NC 27604



BORING GEOPHYSICAL LOGGING
BORINGS B-601 (DH), B-604 (DH), B-608 (DH),
B-610 (DH), B-620 (DH), B-640 (DHT), B-701 (DH),
B-704 G (DH), B-708 (DH), B-710 G (DH),
B-720 G (DH) AND B-740 (DHT)

FPL TURKEY POINT COL

Report 8083-03 rev 0

July 24, 2008

DCN TUR 402

FOIA b7C
b7D
6073

BORING GEOPHYSICAL LOGGING
BORINGS B-601 (DH), B-604 (DH), B-608 (DH),
B-610 (DH), B-620 (DH), B-640 (DHT), B-701 (DH),
B-704 G (DH), B-708 (DH), B-710 G (DH),
B-720 G (DH) AND B-740 (DHT)

FPL TURKEY POINT COL

Report 8083-03 rev 0

July 24, 2008

Prepared for:

MACTEC Engineering and Consulting, Inc.

3301 Atlantic Avenue

Raleigh, N. C. 27604

919-831-8000

MACTEC Job number 6468-07-1950

Prepared by

GEOVision Geophysical Services

1124 Olympic Drive

Corona, California 92881

(951) 549-1234

TABLE OF CONTENTS

| | |
|---|----|
| TABLE OF CONTENTS | 3 |
| TABLE OF FIGURES | 4 |
| TABLE OF TABLES | 5 |
| TABLE OF TABLES | 5 |
| INTRODUCTION | 6 |
| SCOPE OF WORK | 6 |
| INSTRUMENTATION | 8 |
| SUSPENSION VELOCITY INSTRUMENTATION | 8 |
| DOWNHOLE VELOCITY INSTRUMENTATION | 11 |
| CALIPER / NATURAL GAMMA INSTRUMENTATION | 13 |
| RESISTIVITY / SPONTANEOUS POTENTIAL / NATURAL GAMMA INSTRUMENTATION | 14 |
| ACOUSTIC TELEVIEWER / BORING DEVIATION INSTRUMENTATION | 16 |
| MEASUREMENT PROCEDURES | 18 |
| SUSPENSION VELOCITY MEASUREMENT PROCEDURES | 18 |
| DOWNHOLE VELOCITY MEASUREMENT PROCEDURES | 19 |
| CALIPER / NATURAL GAMMA MEASUREMENT PROCEDURES | 19 |
| RESISTIVITY / SPONTANEOUS POTENTIAL / NATURAL GAMMA PROCEDURES | 21 |
| ACOUSTIC TELEVIEWER / BORING DEVIATION MEASUREMENT PROCEDURES | 23 |
| DATA ANALYSIS | 24 |
| SUSPENSION VELOCITY ANALYSIS | 24 |
| DOWNHOLE VELOCITY ANALYSIS | 26 |
| CALIPER / NATURAL GAMMA ANALYSIS | 27 |
| RESISTIVITY / SPONTANEOUS POTENTIAL / NATURAL GAMMA ANALYSIS | 27 |
| ACOUSTIC TELEVIEWER / BORING DEVIATION ANALYSIS | 28 |
| RESULTS | 29 |
| SUSPENSION VELOCITY RESULTS | 29 |
| DOWNHOLE VELOCITY RESULTS | 30 |
| CALIPER/ NATURAL GAMMA RESULTS | 31 |
| RESISTIVITY / SPONTANEOUS POTENTIAL / NATURAL GAMMA RESULTS | 31 |
| ACOUSTIC TELEVIEWER / BORING DEVIATION RESULTS | 32 |
| SUMMARY | 33 |
| DISCUSSION OF SUSPENSION VELOCITY RESULTS | 33 |
| DISCUSSION OF DOWNHOLE VELOCITY RESULTS | 34 |
| DISCUSSION OF CALIPER / NATURAL GAMMA RESULTS | 37 |
| DISCUSSION OF RESISTIVITY / SPONTANEOUS POTENTIAL / NATURAL GAMMA RESULTS | 37 |
| DISCUSSION OF ACOUSTIC TELEVIEWER / BORING DEVIATION RESULTS | 38 |
| QUALITY ASSURANCE | 39 |
| SUSPENSION VELOCITY DATA RELIABILITY | 39 |

Table of Figures

| | |
|--|-----|
| Figure 1: Concept illustration of P-S logging system | 48 |
| Figure 2: Downhole P- and S_H -wave energy source (Triple Whammy)..... | 49 |
| Figure 3: Downhole P-wave striking plate with hammer switch..... | 50 |
| Figure 4: Example Calibration Curve for Caliper Probe..... | 51 |
| Figure 5: Example of filtered (1400 Hz lowpass) record..... | 52 |
| Figure 6: Example of unfiltered record | 53 |
| Figure 7: Boring B-640 (DHT), Downhole Vs 100 Hz filtered waveforms and first arrival picks..... | 54 |
| Figure 8: Boring B-601 (DH), Suspension R1-R2 P- and S_H -wave velocities | 55 |
| Figure 9: Boring B-601 (DH) upper section, Caliper, Natural gamma, Resistivity and SP logs..... | 58 |
| Figure 10: Boring B-601 (DH) middle section, Caliper, Natural gamma, Resistivity and SP logs..... | 59 |
| Figure 11: Boring B-601 (DH) lower section, Caliper, Natural gamma, Resistivity and SP logs..... | 60 |
| Figure 12: Boring B-601 (DH), Deviation Projection (dimensions in feet)..... | 61 |
| Figure 13: Boring B-604 (DH), Suspension R1-R2 P- and S_H -wave velocities | 62 |
| Figure 14: Boring B-604 (DH) upper section, Caliper, Natural gamma, Resistivity and SP logs | 64 |
| Figure 15: Boring B-604 (DH) lower section, Caliper, Natural gamma, Resistivity and SP logs | 65 |
| Figure 16: Boring B-604 (DH), Deviation Projection (dimensions in feet)..... | 66 |
| Figure 17: Boring B-608 (DH), Suspension R1-R2 P- and S_H -wave velocities | 67 |
| Figure 18: Boring B-608 (DH) upper section, Caliper, Natural gamma, Resistivity and SP logs | 69 |
| Figure 19: Boring B-608 (DH) lower section, Caliper, Natural gamma, Resistivity and SP logs | 70 |
| Figure 20: Boring B-608 (DH), Deviation Projection (dimensions in feet)..... | 71 |
| Figure 21: Boring B-610 (DH), Suspension R1-R2 P- and S_H -wave velocities | 72 |
| Figure 22: Boring B-610 (DH) upper section, Caliper, Natural gamma, Resistivity and SP logs | 74 |
| Figure 23: Boring B-610 (DH) lower section, Caliper, Natural gamma, Resistivity and SP logs | 75 |
| Figure 24: Boring B-610 (DH), Deviation Projection (dimensions in feet)..... | 76 |
| Figure 25: Boring B-620 (DH), Suspension R1-R2 P- and S_H -wave velocities | 77 |
| Figure 26: Boring B-620 (DH) upper section, Caliper, Natural gamma, Resistivity and SP logs | 79 |
| Figure 27: Boring B-620 (DH) lower section, Caliper, Natural gamma, Resistivity and SP logs | 80 |
| Figure 28: Boring B-620 (DH), Deviation Projection (dimensions in feet)..... | 81 |
| Figure 29: Boring B-701 (DH), Suspension R1-R2 P- and S_H -wave velocities | 82 |
| Figure 30: Boring B-701 (DH) upper section, Caliper, Natural gamma, Resistivity and SP logs | 86 |
| Figure 31: Boring B-701 (DH) upper middle section, Caliper, Natural gamma, Resistivity and SP logs..... | 87 |
| Figure 32: Boring B-701 (DH) lower middle section, Caliper, Natural gamma, Resistivity and SP logs | 88 |
| Figure 33: Boring B-701 (DH) lower section, Caliper, Natural gamma, Resistivity and SP logs | 89 |
| Figure 34: Boring B-701 (DH), Deviation Projection (dimensions in feet)..... | 90 |
| Figure 35: Boring B-704 G (DH), Suspension R1-R2 P- and S_H -wave velocities | 91 |
| Figure 36: Boring B-704 G (DH) upper section, Caliper, Natural gamma, Resistivity and SP logs | 93 |
| Figure 37: Boring B-704 G (DH) lower section, Caliper, Natural gamma, Resistivity and SP logs | 94 |
| Figure 38: Boring B-704 G (DH), Deviation Projection (dimensions in feet)..... | 95 |
| Figure 39: Boring B-708 (DH), Suspension R1-R2 P- and S_H -wave velocities | 96 |
| Figure 40: Boring B-708 (DH) upper section, Caliper, Natural gamma, Resistivity and SP logs | 98 |
| Figure 41: Boring B-708 (DH) lower section, Caliper, Natural gamma, Resistivity and SP logs | 99 |
| Figure 42: Boring B-708 (DH), Deviation Projection (dimensions in feet)..... | 100 |
| Figure 43: Boring B-710 G (DH), Suspension R1-R2 P- and S_H -wave velocities | 101 |
| Figure 44: Boring B-710 G (DH) upper section, Caliper, Natural gamma, Resistivity and SP logs | 103 |
| Figure 45: Boring B-710 G (DH) lower section, Caliper, Natural gamma, Resistivity and SP logs | 104 |
| Figure 46: Boring B-710 G (DH), Deviation Projection (dimensions in feet)..... | 105 |
| Figure 47: Boring B-720 G (DH), Suspension R1-R2 P- and S_H -wave velocities | 106 |
| Figure 48: Boring B-720 G (DH) upper section, Caliper, Natural gamma, Resistivity and SP logs | 108 |
| Figure 49: Boring B-720 G (DH) lower section, Caliper, Natural gamma, Resistivity and SP logs..... | 109 |
| Figure 50: Boring B-720 G (DH), Deviation Projection (dimensions in feet)..... | 110 |

Table of Tables

| | |
|---|-----|
| Table 1. Boring locations and logging dates..... | 40 |
| Table 2. Suspension PS Logging Equipment..... | 40 |
| Table 3. Logging dates and depth ranges | 41 |
| Table 4. Boring Bottom Depths and After Survey Depth Error (ASDE)..... | 44 |
| Table 5. Boring Deviation Data Summary..... | 47 |
| Table 6. Boring B-601 (DH), Suspension R1-R2 depths and P- and S _H -wave velocities..... | 56 |
| Table 7. Boring B-604 (DH), Suspension R1-R2 depths and P- and S _H -wave velocities..... | 63 |
| Table 8. Boring B-608 (DH), Suspension R1-R2 depths and P- and S _H -wave velocities..... | 68 |
| Table 9. Boring B-610 (DH), Suspension R1-R2 depths and P- and S _H -wave velocities..... | 73 |
| Table 10. Boring B-620 (DH), Suspension R1-R2 depths and P- and S _H -wave velocities..... | 78 |
| Table 11. Boring B-701 (DH), Suspension R1-R2 depths and P- and S _H -wave velocities..... | 83 |
| Table 12. Boring B-704 G (DH), Suspension R1-R2 depths and P- and S _H -wave velocities | 92 |
| Table 13. Boring B-708 (DH), Suspension R1-R2 depths and P- and S _H -wave velocities..... | 97 |
| Table 14. Boring B-710 G (DH), Suspension R1-R2 depths and P- and S _H -wave velocities | 102 |
| Table 15. Boring B-720 G (DH), Suspension R1-R2 depths and P- and S _H -wave velocities | 107 |

APPENDICES

| | | |
|-------------------|--|------------|
| APPENDIX A | SUSPENSION VELOCITY MEASUREMENT QUALITY ASSURANCE SUSPENSION SOURCE TO RECEIVER ANALYSIS RESULTS..... | 111 |
| APPENDIX B | DOWNHOLE VELOCITY MEASUREMENT RESULTS..... | 135 |
| APPENDIX C | CALIPER, NATURAL GAMMA, RESISTIVITY, AND SPONTANEOUS POTENTIAL LOGS..... | 150 |
| APPENDIX D | ACOUSTIC TELEVIEWER BASED CALIPER LOGS..... | 203 |
| APPENDIX E | GEOPHYSICAL LOGGING SYSTEMS - NIST TRACEABLE CALIBRATION PROCEDURES AND CALIBRATION RECORDS..... | 373 |
| APPENDIX F | BORING GEOPHYSICAL LOGGING FIELD DATA LOGS..... | 394 |
| APPENDIX G | BORING GEOPHYSICAL LOGGING FIELD MEASUREMENT PROCEDURES..... | 642 |

INTRODUCTION

Boring geophysical measurements were collected in ten uncased and two cased borings located at the Florida Power and Light (FPL) Turkey Point Combined Operating License (COL) Application Project, located near Florida City, Florida. Geophysical data acquisition was performed between March 8 and June 26, 2008 by Robert Steller, Charles Carter, Anthony Martin and Nathan Baldwin of **GEOVision**. Data analysis was performed by Robert Steller and Anthony Martin, and reviewed by John Diehl of GEOVision. Report preparation was performed by Robert Steller and reviewed by John Diehl of GEOVision. The work was performed under subcontract with MACTEC Engineering and Consulting, Inc., (MACTEC) with Stephen Criscenzo serving as the point of contact for MACTEC.

This report describes the field measurements, data analysis, and results of this work.

SCOPE OF WORK

This report presents the results of boring geophysical measurements collected between March 8 and June 26, in twelve borings, as detailed in Table 1. The purpose of these studies was to supplement stratigraphic information obtained during MACTEC's soil and rock sampling program and to acquire shear wave velocities and compressional wave velocities as a function of depth, as a component of the FPL Turkey Point COL Project.

The OYO/Robertson Suspension PS Logging System (Suspension System) was used to obtain in-situ horizontal shear (S_H) and compressional (P) wave velocity measurements in ten borings at 1.6 foot intervals. Measurements followed **GEOVision** Procedure for P-S Suspension Seismic Velocity Logging, revision 1.31. The acquired data was analyzed and a profile of velocity versus depth was produced for both compressional and horizontally polarized shear waves.

The Robertson ELGX and 3ACS probes were used to collect long and short normal resistivity, single point resistance (SPR) Spontaneous Potential (SP), natural gamma and 3 arm mechanical caliper data at 0.05 foot intervals in the ten uncased borings to aid in identification of stratigraphic transitions. Measurement procedures followed these ASTM standards:

- ASTM D5753, "Planning and Conducting Borehole Geophysical Surveys
- ASTM D6167 "Conducting Borehole Geophysical Logging -- Mechanical Caliper"
- ASTM D6274, "Conducting Borehole Geophysical Logging -- Gamma"

The acquired data was combined and a profile of these parameters versus depth was produced.

The Robertson High Resolution Acoustic Televiwer (HiRAT) was used to collect deviation data at 0.04 foot intervals and acoustic televiwer images of the rock section of each boring at 0.008 foot intervals in the ten uncased borings. Measurements followed the **GEOVision** HiRAT Field Procedure, revision 1.0. The acquired data was analyzed and a profile of boring deviation versus depth was produced for each boring, and an image of the rock portions of the uncased borings, with 4 arm caliper dimensions superimposed, was produced.

The Downhole Seismic velocity logging system was used in the two PVC cased borings as a validation of the suspension velocity data collected at this site. In this method, the source remains stationary at the surface, while a single receiver travels down the cased boring at 5 foot intervals. Source energy is transmitted down the soil column from the surface and velocity is calculated from first arrival travel time and receiver depth. Measurements followed **GEOVision** Procedure for Downhole Seismic Velocity Logging, revision 1.1. The acquired data was analyzed and a profile of velocity versus depth was produced for both P- and S_{II} -waves.

A detailed reference for the suspension PS velocity measurement techniques used in this study is:

Guidelines for Determining Design Basis Ground Motions, Report TR-102293,
Electric Power Research Institute, Palo Alto, California, November 1993,
Sections 7 and 8.

INSTRUMENTATION

Suspension Velocity Instrumentation

Suspension velocity measurements were performed in ten uncased nominal 3.88 – 5.0 inch diameter borings using the suspension PS logging system, manufactured by OYO Corporation, and their subsidiary, Robertson Geologging. Components used for these measurements are listed in Table 2. This system directly determines the average velocity of a 3.3 foot high segment of the soil column surrounding the boring of interest by measuring the elapsed time between arrivals of a wave propagating upward through the soil column. The receivers that detect the wave, and the source that generates the wave, are moved as a unit in the boring producing relatively constant amplitude signals at all depths.

The suspension system probe consists of a combined reversible polarity solenoid horizontal shear-wave source (S_H) and compressional-wave source (P), joined to two biaxial receivers by a flexible isolation cylinder, as shown in Figure 2. The separation of the two receivers is 3.3 feet, allowing average wave velocity in the region between the receivers to be determined by inversion of the wave travel time between the two receivers. The total length of the probe as used in these surveys is 19 feet, with the center point of the receiver pair 12.1 feet above the bottom end of the probe.

The probe receives control signals from, and sends the digitized receiver signals to, instrumentation on the surface via an armored 4 conductor cable. The cable is wound onto the drum of a winch and is used to support the probe. Cable travel is measured to provide probe depth data, using a 3.28 foot circumference sheave fitted with a digital rotary encoder.

The entire probe is suspended in the boring by the cable, therefore, source motion is not coupled directly to the boring walls; rather, the source motion creates a horizontally propagating impulsive pressure wave in the fluid filling the boring and surrounding the source. This pressure wave is converted to P and S_H -waves in the surrounding soil and rock as it passes through the

casing and grout annulus and impinges upon the wall of the boring. These waves propagate through the soil and rock surrounding the boring, in turn causing a pressure wave to be generated in the fluid surrounding the receivers as the soil waves pass their location. Separation of the P and S_{II} -waves at the receivers is performed using the following steps:

1. Orientation of the horizontal receivers is maintained parallel to the axis of the source, maximizing the amplitude of the recorded S_{II} -wave signals.
2. At each depth, S_{II} -wave signals are recorded with the source actuated in opposite directions, producing S_{II} -wave signals of opposite polarity, providing a characteristic S_{II} -wave signature distinct from the P-wave signal.
3. The 6.3 foot separation of source and receiver 1 permits the P-wave signal to pass and damp significantly before the slower S_{II} -wave signal arrives at the receiver. In faster soils or rock, the isolation cylinder is extended to allow greater separation of the P- and S_{II} -wave signals.
4. In saturated soils, the received P-wave signal is typically of much higher frequency than the received S_{II} -wave signal, permitting additional separation of the two signals by low pass filtering.
5. Direct arrival of the original pressure pulse in the fluid is not detected at the receivers because the wavelength of the pressure pulse in fluid is significantly greater than the dimension of the fluid annulus surrounding the probe (meter versus centimeter scale), preventing significant energy transmission through the fluid medium.

In operation, a distinct, repeatable pattern of impulses is generated at each depth as follows:

1. The source is fired in one direction producing dominantly horizontal shear with some vertical compression, and the signals from the horizontal receivers situated parallel to the axis of motion of the source are recorded.
2. The source is fired again in the opposite direction and the horizontal receiver signals are recorded.
3. The source is fired again and the vertical receiver signals are recorded. The repeated source pattern facilitates the picking of the P and S_{II} -wave arrivals; reversal of the source changes the polarity of the S_{II} -wave pattern but not the P-wave pattern.

The data from each receiver during each source activation is recorded as a different channel on the recording system. The Suspension PS system has six channels (two simultaneous recording channels), each with a 1024 sample record. The recorded data are displayed as six channels with a common time scale. Data are stored on disk for further processing. Up to 8 sampling sequences can be summed to improve the signal to noise ratio of the signals.

Review of the displayed data on the recorder or computer screen allows the operator to set the gains, filters, delay time, pulse length (energy), sample rate, and summing number to optimize the quality of the data before recording. Verification of the calibration of the Suspension PS digital recorder is performed every twelve months using a NIST traceable frequency source and counter, as outlined in Appendix E.

Downhole Velocity Instrumentation

Downhole velocity measurements were performed in two 2 inch PVC cased borings using a Geostuff BHG-3, 3-component borehole geophone, serial number B-3079. This system orients the downhole geophones parallel to the axis of excitation at the surface, insuring that signals received at the downhole geophones are of maximum amplitude, and are not subject to errors in travel time caused by incorrect phase of first arrival picks, as found with non-orientable downhole probes. The downhole probe consists of horizontal and vertical geophones mounted on a rotatable structure with a fluxgate magnetometer compass sensor. The structure can be pre-set on the surface to match the azimuth of the horizontal geophone axis with the azimuth of the surface shear wave source whenever power is applied to the compass sensor and orientation servo mechanism. The probe receives control signals from, and sends the geophone signals to, instrumentation on the surface via cable. Cable travel is measured to provide probe depth data. The probe is locked into the boring using a motor driven clamp mechanism. The BHGC-4 controller directs the voltages to control the clamping mechanism and orientation mechanism. A meter monitors motor current to indicate the clamping action and force.

A triaxial geophone is placed on the surface adjacent to the boring collar, to record reference waveforms to validate the function of the hammer switch, as well as to monitor shifts in timing due to changes in source coupling to the soil.

The S_{II} -wave energy source consists of an 88-pound elastic band accelerated hammer striking horizontally against the ends of a steel capped traction plank. The traction plank is weighted by placing it beneath the rear end of a truck supported on an air suspension, as shown in Figure 2. The P-wave energy source utilizes the same energy source operating in a vertical orientation, striking an aluminum plate, as shown in Figure 3. A hammer switch mounted on the steel plank caps or aluminum plate is used to provide consistent triggering from each hammer blow. During logging operations, a repeatable pattern of impulses, similar to that produced by the suspension source, is generated at each measurement depth as follows:

1. The plank is struck with the hammer laterally in one direction, producing dominantly horizontal shear with some vertical compression, and the signals generated by the horizontal receivers are recorded. The signals are checked, and repeated (stacked) as needed.
2. The plank is struck in the opposite direction and the horizontal signals are recorded, and stacked as needed.
3. The plate is struck on top, and the signals generated by the vertical receivers are recorded. The repeated source pattern facilitates the picking of the P- and S_{II}-wave arrivals, since the reversal of the source direction changes the polarity of the S_{II}-wave pattern but not that of the P-wave pattern.

The signals from the BHG-3 geophone were recorded on a Geometrics Geode seismograph, controlled by a laptop computer. Geode S/N 3458 was used on both borings. The Geode is a 24-bit exploration seismograph with 113dB dynamic range. Triggered by the hammer switch (see procedure, Appendix G) the seismograph recorded the responses of the borehole and surface sensors. Data was reviewed on the computer screen, and stored internally on hard disk. Multiple hammer blows can be summed to improve the signal-to-noise ratio of the signals. Review of the displayed data on the screen allows the operator to set the gains, filters, sample rate, and summing number in order to optimize the quality of the data before recording to disk for later processing.

Caliper / Natural Gamma Instrumentation

Caliper and natural gamma data were collected using a Model 3ACS 3-leg caliper probe, serial number 5368, manufactured by Robertson Geologging, Ltd. With the short arm configuration used in these surveys, the probe permitted measurement of boring diameters between 1.6 and 16 inches. With this tool, caliper measurements were collected concurrent with measurement of natural gamma emission from the boring walls. The probe was 6.82 feet long, and 1.5 inches in diameter.

This probe is useful in the following studies:

- Measurement of boring diameter and volume
- Location of hard and soft formations
- Location of fissures, caving, pinching and casing damage
- Bed boundary identification
- Strata correlation between borings

The probe receives control signals from, and sends the digitized measurement values to, a Robertson Micrologger II on the surface via an armored 4 conductor cable. The cable is wound onto the drum of a winch and is used to support the probe. Cable travel is measured to provide probe depth data, using a 3.28 foot circumference sheave fitted with a digital rotary encoder. The probe and depth data are transmitted by USB link from the Micrologger unit to a laptop computer where it is displayed and stored on hard disk.

The caliper consists of three arms, each with a toothed quadrant at their base, pivoted in the lower probe body. A toothed rack engages with each quadrant, thus constraining the arms to move together. Linear movement of the rack is converted to opening and closing of the arms. Springs hold the arms open in the operating position. A motor drive is provided to retract the arms, allowing the probe to be lowered into the boring. The rack is coupled to a potentiometer which converts movement into a voltage sensed by the probe's microprocessor.