Project 6468-07-1950 July 24, 2008

CALIBRATION PROCEDURE FOR

GEOVision SEISMIC RECORDER/LOGGER

Reviewed 4/6/06

Objective

The timing/sampling accuracy of seismic recorders or data loggers is required for several GEOVision field procedures including Seismic Refraction, Downhole Seismic Velocity Logging, and P-S Suspension Logging. This procedure describes the method for measuring the timing accuracy of a seismic data logger, such as the OYO Model 170, OYO/Robertson Model 3403, Geometrics Strataview or Geometrics Geode. The objective of this procedure is to verify that the timing accuracy of the recorder is accurate to within 1%.

Frequency of Calibration

The calibration of each GEOVision seismic data logger is twelve (12) months. In the case of rented seismic data loggers, calibration must be performed prior to use.

Test Equipment Required

The following equipment is required. Item #2 must have current NIST traceable calibration.

- 1. Function generator, Krohn Hite 5400B or equivalent
- 2. Frequency counter, HP 5315A or equivalent
- 3. Test cables, from item 1 to item 2, and from item 1 to subject data logger.

Procedure

This procedure is designed to be performed using the accompanying Seismograph Calibration Data Sheet with the same revision number. All data must be entered and the procedure signed by the technician performing the test.

- 1. Record all identification data on the form provided.
- 2. Connect function generator to data logger (such as OYO Model 170) using test cable
- 3. Connect the function generator to the frequency counter using test cable.

Seismic Recorder/Logger Calibration Procedure **GE** *Vision* Revision 1.30 Page 1

- 4. Set up generator to produce a 100.0 Hz, 0.25 volt (amplitude is approximate, modify as necessary to yield less than full scale waveforms on logger display) peak square wave or sine wave. Verify frequency using the counter and initial space on the data sheet.
- 5. Initialize data logger and record a data record of at least 0.1 second using a 100 microsecond or less sample period.
- 6. Measure the recorded square wave frequency by measuring the duration of 9 cycles of data. This measurement can be made using the data logger display device, or by printing out a paper tape. If a paper tape can be printed, the resulting printout must be attached to this procedure. Record the data in the space provided.
- 7. Repeat steps 5 and 6 three more times using separate files.

Criteria

The duration for 9 cycles in any file must be 90.0 milliseconds plus or minus 0.9 milliseconds, corresponding to an average frequency for the nine cycles of 100.0 Hz plus or minus 1 Hz (obtained by dividing 9 cycles by the duration in milliseconds).

If the results are outside this range, the data logger must be marked with a GEOVision REJECT tag until it can be repaired and retested.

If results are acceptable affix label indicating the initials of the person performing the calibration, the date of calibration, and the due date for the next calibration (12 months).

Procedure Approval

Approved by:

CE	Seismic Recorder/Logger Calibration Procedure
Signature	Date
Name	Title
Client Approval (if required):	
Signature	Date
Name	April 6, 2006
John G. Dieni	

MACTEC Engineering and Consulting, Inc. FPL Turkey Point COL Geotechnical Data



A COURSEN ON ITERNALDISON

Metrology 7300 Fenwick Lane Westminster, CA 92683 Phone: 866-723-2257

EDISON ESI[®] Calibration Report

NVLAP Accredited Calibration **GEOVision Geophysical Services** 1151 Pomona Road, Unit P Corona, CA 92882

TEST N	UMBER)
550	393



Manufacturer:	Оуо
Model Number:	3403
Description:	Unit, Suspension Telemetry,
Asset Number:	160023
Serial Number:	160023
PO Number:	8200-080122-01

In Tolerance
In Tolerance
01/25/2008
01/25/2009
12 Months

Remarks:

The UUF (unit under test) was calibrated using the customer's procedure. The UUT was operated by the customer's personnel and data collection was observed by SCE personnel. The UUT was found to be in tolerance to customer supplied specifications. The reference standards used are in compliance with ISO/IEC 17025:1999 and laboratory accreditation criteria established by NIST/NVLAP under the specific scope of accreditation for lab code 105014-0. Frequency is accredited. Measurement uncertainity is 0.2 x E-12 Ifz. Please see attached data. Standanda fitiliand

Standards Outzed						
I.D. No.	Mfg.	Model No.	Description	Cal. Date Due Date		
S1-01252	Hewlett Packard	5335A OPT 010,203040	Counter, Universal,	12/28/2007 06/28/2008		
S1-01347	Hewlett Packard	3325A	Generator, Function, Synthesizer	10/09/2007 04/09/2008		
S1-03686	Fluke	910	Standard, Frequency, Controlled, Gps	01/22/2008 01/22/2009		

Procedure:	Customer	Calibration Performed By			Quality Reviewer:	
Humidity: Test No.:	37% RH 550393	Branson, Craig A	Metrologist	714-895-0714	ChungSterrer	listop
		Naute	Fithe	Phone	Name	Date

This report may not be reproduced, except in full, without written permission of this laboratory. This report may not be used to claim product endorsement by NVLAP or any agency of the US Government. The results stated in this report relate only to the items tested or calibrated. Measurements reported herein are traceable to SI units via national standards maintained by NIST. This calibration is in compliance with NVLAP laboratory accreditation criteria established by NIST/NVLAP under the specific scope of accreditation for lab code 105014-0.

www.edisonmudcats.com

www.edisonmetrology.com



Project 6468-07-1950 July 24, 2008



SEISMOGRAPH CALIBRATION DATA SHEET REV 4/6/06

INSTRUMENT I	DATA							
SYSTEM MER:		MODEL NO	.:	3403				
SERIAL NO.:	ERIAL NO.: 160023			ON DATE:	01/25/2008		******	Arrist-1 (1999)
BY: ROBERT STELLER		DUE DATE:		01/25/2009				
COUNTER MFF	R: HEWLE	ETT PACKARD	MODEL NO	.:	5335A			
SERIAL NO.:	2626A09	881	CALIBRATIC	ON DATE:	12/28/2007			
BY:	SCE #S1	-01252	DUE DATE:		6/28/2008			
FCTN GEN MFI	R: HEWL	ETT PACKARD	MODEL NO	.:	3325A			
SERIAL NO .:	2652A25	647	CALIBRATIC	ON DATE:	10/9/2007			
BY:	SCE #S1	-01347	DUE DATE:		4/9/2008			
SYSTEM SETTI	NGS:							
GAIN:			10					
FILTER:			20 KHZ			·····		
RANGE:			100 MILLISE	C				
DELAY:			0					
STACK: 1 (STD)		1					
PULSE:	·		1.6					
DISPLAY:			NA					
SYSTEM: DATE	= CORR	ECT DATE & TIME	01/25/2008,	12:20PM				
0.25 VOLT PEA AND PRINT WA AND PAPER TA BETWEEN 99.0	K. RECO VEFORM PES, IF A AND 101	RD BOTH ON DISK IS FROM ANALYSIS AVAILABLE, TO THIS .0 HZ.	and paper Utility. At 5 Form. Ave	TAPE, IF A TACH PAP RAGE FRE	VAILABLE.) ER COPIES QUENCY MI	ANALY OF PR JST BE	ZE INTOL E	JT
AS FOUND		100.0	_	AS LEFT	100.0		·····	
		ERECHENCY	TIME FOR	TIME FOR	TIME FO	12.0	AVE	
		THEODENOT	9 CYCLES		CYCLE	3	FP	FO
			Hn	Hr	V	.0		L(x.
SQUARE	301	100.0	90.0	90.0		90.0		100.0
SQUARE	302	100.0	90.0	90.0	[90.0		100.0
SINE	303	100.0	90.0	90.0		90.0		100.0
SINE	304	100.0	90.0	90.0		90.0	<u> </u>	100.0
CALIBRATED BY: ROBERT STELLER 1/25/2008 augusture								
						-		
							Page	2 of 2
	Seismic I	recorder/Logger Calil	pration Data S	Sheet Rev	1.30 4-6-0	6		

i

MACTEC Engineering and Consulting, Inc. FPL Turkey Point COL Geotechnical Data





X SCI THERN CALIFORNI LEDINON -

Metrology 7300 Fenwick Lane Westminster, CA 92683 Phone: 866-723-2257

Manufacturer:

EDISON ESF Calibration Report

NVLAP Accredited Calibration **GEOVision Geophysical Services** 1151 Pomona Road, Unit P Corona, CA 92882

TEST NUMBER	
550394	



Condition As Found:	In Tolerance
Condition As Left;	In Tolerance
Calibration Date:	01/25/2008
Calibration Due Date:	01/25/2009
Calibration Interval:	12 Months

Model Number: 3403 Unit, Suspension Telemetry, Description: Asset Number: 160024 Serial Number: 160024 8200-080122-01 **PO Number:**

Oyo

Remarks:

The UUT (unit under test) was calibrated using the customer's procedure. The UUT was operated by the customer's personnel and data collection was observed by SCE personnel. The UUT was found to be in tolerance to customer supplied specifications. The reference standards used are in compliance with ISO/IEC 17025:1999 and laboratory accreditation criteria established by NIST/NVLAP under the specific scope of accreditation for lab code 105014-0. Frequency is accredited. Measurement uncertainity is $0.2 \times E-12$ Hz. Please see attached data. Standards Litilized

I.D. No.	Mfg	Model No.	Deśćription	Cal. Date	Due Date		
S1-01252	Hewlett Packard	5335A OPT 010,203040	Counter, Universal,	12/28/2007	06/28/2008		
S1-01347	Hewlett Packard	3325A	Generator, Function, Synthesizer	10/09/2007	04/09/2008		
S1-03686	Fluke	910	Standard, Frequency, Controlled, Gps	01/22/2008	01/22/2009		

Procedure:	Customer	Calibration Performed By			Quality Reviewer:	
remperature	20 0				1	
Humidity:	37% RH	» a : 10 ~ 3	N	714 005 0114	OD crist- 1	1 1 .
Test No.:	550394	Branson, Craig A ($/ \sim /$	Metrologist	/14-895-0/14	C. Challence To Summer "	1125708
		Name	Title	Phone	Name	/ Date

This report may not be reproduced, except in full, without written permission of this laboratory. This report may not be used to claim product endorsement by NVLAP or any agency of the US Government. The results stated in this report relate only to the items tested or calibrated. Measurements reported herein are traceable to SI units via national standards maintained by NIST. This calibration is in compliance with NVLAP laboratory accreditation criteria established by NIST/NVLAP under the specific scope of accreditation for lab code 105014-0.

www.edisonmudcats.com

P

www.edisonmetrology.com Page 1 of 2