



SEISMOGRAPH CALIBRATION DATA SHEET REV 4/6/06

INSTRUMENT DATA

SYSTEM MFR: OYO	MODEL NO.: 3403
SERIAL NO.: 160024	CALIBRATION DATE: 01/25/2008
BY: ROBERT STELLER	DUE DATE: 01/25/2009
COUNTER MFR: HEWLETT PACKARD	MODEL NO.: 5335A
SERIAL NO.: 2626A09881	CALIBRATION DATE: 12/28/2007
BY: SCE #S1-01252	DUE DATE: 6/28/2008
FCTN GEN MFR: HEWLETT PACKARD	MODEL NO.: 3325A
SERIAL NO.: 2652A25647	CALIBRATION DATE: 10/9/2007
BY: SCE #S1-01347	DUE DATE: 4/9/2008

SYSTEM SETTINGS:

GAIN:	10
FILTER:	20 KHZ
RANGE:	100 MILLISEC
DELAY:	0
STACK: 1 (STD)	1
PULSE:	1.6
DISPLAY:	NA
SYSTEM: DATE = CORRECT DATE & TIME	01/25/2008, 12:20PM

PROCEDURE:

SET FREQUENCY TO 100.0HZ SQUAREWAVE WITH AMPLITUDE APPROXIMATELY 0.25 VOLT PEAK. RECORD BOTH ON DISK AND PAPER TAPE, IF AVAILABLE. ANALYZE AND PRINT WAVEFORMS FROM ANALYSIS UTILITY. ATTACH PAPER COPIES OF PRINTOUT AND PAPER TAPES, IF AVAILABLE, TO THIS FORM. AVERAGE FREQUENCY MUST BE BETWEEN 99.0 AND 101.0 HZ.

AS FOUND 100.0 AS LEFT 100.0

WAVEFORM	FILE NO	FREQUENCY	TIME FOR 9 CYCLES Hn	TIME FOR 9 CYCLES Hr	TIME FOR 9 CYCLES V	AVERAGE FREQ.
SQUARE	401	100.0	90.0	90.0	90.0	100.0
SQUARE	402	100.0	90.0	90.0	90.0	100.0
SINE	403	100.0	90.0	90.0	90.0	100.0
SINE	404	100.0	90.1	90.0	90.0	100.0

CALIBRATED BY:	ROBERT STELLER	1/25/2008	<i>Rob Steller</i>
	NAME	DATE	SIGNATURE

Page 2 of 2

Seismic recorder/Logger Calibration Data Sheet Rev 1.30 4-6-06
--




Metrology
7300 Fenwick Lane
Westminster, CA 92683
Phone: 866-723-2257

Calibration Report

NVLAP Accredited

Calibration

GEOVision Geophysical Services
1151 Pomona Road, Unit P
Corona, CA 92882



Manufacturer: Geometrics
Model Number: GEODE
Description: Siesmograph,
Asset Number: 3458
Serial Number: 3458
PO Number: 8200-080122-01

Condition As Found: In Tolerance
Condition As Left: In Tolerance
Calibration Date: 01/25/2008
Calibration Due Date: 01/25/2009
Calibration Interval: 12 Months

Remarks:

The UUT (unit under test) was calibrated using the customer's procedure. The UUT was operated by the customer's personnel and data collection was observed by SCE personnel. The UUT was found to be in tolerance to customer supplied specifications. The reference standards used are in compliance with ISO/IEC 17025:1999 and laboratory accreditation criteria established by NIST/NVLAP under the specific scope of accreditation for lab code 105014-0. Frequency is accredited. Measurement uncertainty is 0.2 x 10⁻¹² Hz. Please see attached data.

Standards Utilized

I.D. No.	Mfg.	Model No.	Description	Cal. Date	Due Date
S1-01252	Hewlett Packard	5335A OPT 010,203040	Counter, Universal,	12/28/2007	06/28/2008
S1-01347	Hewlett Packard	3325A	Generator, Function, Synthesizer	10/09/2007	04/09/2008
S1-03686	Fluke	910	Standard, Frequency, Controlled, Gps	01/22/2008	01/22/2009

Procedure: Customer
Temperature: 23° C
Humidity: 37% RH
Test No.: 550389

Calibration Performed By:			Quality Reviewer:	
Branson, Craig A	Metrologist	714-895-0714		
Name	Title	Phone	Name	Date

This report may not be reproduced, except in full, without written permission of this laboratory. This report may not be used to claim product endorsement by NVLAP or any agency of the US Government. The results stated in this report relate only to the items tested or calibrated. Measurements reported herein are traceable to SI units via national standards maintained by NIST. This calibration is in compliance with NVLAP laboratory accreditation criteria established by NIST/NVLAP under the specific scope of accreditation for lab code 105014-0.

www.edisonmetrology.com

www.edisonmetrology.com

Page 1 of 2



SEISMOGRAPH CALIBRATION DATA SHEET REV 4/6/06

INSTRUMENT DATA

SYSTEM MFR: GEOMETRICS	MODEL NO.: GEODE
SERIAL NO.: 3458	CALIBRATION DATE: 01/25/2008
BY: ROBERT STELLER	DUE DATE: 01/25/2009
COUNTER MFR: HEWLETT PACKARD	MODEL NO.: 5335A
SERIAL NO.: 2626A09881	CALIBRATION DATE: 12/28/2007
BY: SCE #S1-01252	DUE DATE: 6/28/2008
FCTN GEN MFR: HEWLETT PACKARD	MODEL NO.: 3325A
SERIAL NO.: 2652A25647	CALIBRATION DATE: 10/9/2007
BY: SCE #S1-01347	DUE DATE: 4/9/2008

SYSTEM SETTINGS:

GAIN:	24 Db
FILTER:	NONE
RANGE:	SAMPLE = 62.5 MICROSEC, RECORD = 0.1 SEC
DELAY:	0
STACK: 1 (STD)	1
PULSE:	NA
DISPLAY:	NA
SYSTEM: DATE = CORRECT DATE & TIME	01/25/2008, 1:10PM

PROCEDURE:

SET FREQUENCY TO 100.0HZ SQUAREWAVE WITH AMPLITUDE APPROXIMATELY 0.25 VOLT PEAK. RECORD BOTH ON DISK AND PAPER TAPE, IF AVAILABLE. ANALYZE AND PRINT WAVEFORMS FROM ANALYSIS UTILITY. ATTACH PAPER COPIES OF PRINTOUT AND PAPER TAPES, IF AVAILABLE, TO THIS FORM. AVERAGE FREQUENCY MUST BE BETWEEN 99.0 AND 101.0 HZ.

AS FOUND 100 AS LEFT 100

WAVEFORM	FILE NO	FREQUENCY	TIME FOR 9 CYCLES Hn	TIME FOR 9 CYCLES Hr	TIME FOR 9 CYCLES V	AVERAGE FREQ.
SQUARE	601	100.0	90	90	90	100
SQUARE	602	100.0	90	90	90	100
SINE	603	100.0	90	90	90	100
SINE	604	100.0	90	90	90	100

CALIBRATED BY:	ROBERT STELLER	1/25/2008	<i>Rob Steller</i>
	NAME	DATE	SIGNATURE

Page 2 of 2

Seismic recorder/Logger Calibration Data Sheet Rev 1.30 4-6-06



**National Voluntary
Laboratory Accreditation Program**



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2005

Southern California Edison Company
7300 Fenwick Lane
Westminster, CA 92683
Ms. Jennifer E. Smith
Phone: 714-895-0133 Fax: 714-895-0781
E-mail: Jennifer.Smith@sce.com
URL: <http://www.edisonmetrology.com>

CALIBRATION LABORATORIES

NVLAP LAB CODE 105014-0

Scope Revised : 2007-04-20

NVLAP Code: 20/A01

ANSI/NCCL Z540-1-1994; Part 1

Compliant

DIMENSIONAL

NVLAP Code: 20/D03
Gage Blocks

Nominal Length in in

Best Uncertainty ($\pm$) in μin ^{note 1}

0.01 to $\leq$ 0.05	1.9
0.05 to $\leq$ 0.1	1.7
0.1 to $\leq$ 1.0	1.2
1.0	1.4
2.0	1.8
3.0	2.2
4.0	2.9
5.0	5.4
6.0	5.6
7.0	5.8
8.0	6.0
10.0	6.8
12.0	7.2
16.0	8.1
20.0	9.4

2007-04-01 through 2008-03-31

Effective dates

Sally S. Bruce

For the National Institute of Standards and Technology



CALIBRATION LABORATORIES

NVLAP LAB CODE 105014-0

Scope Revised : 2007-04-20

Nominal Length in mm

Best Uncertainty ($\pm$) in nm ^{note 1}

0.5 to < 1.0	52
1.0 to < 2.5	44
2.5 to < 25.0	39
25.0	44
50.0	47
75.0	60
100.0	80

NVLAP Code: 20/D11

Spherical Diameter; Ring Gages

Range in inches

Best Uncertainty ($\pm$) in μ in ^{note 1}

Remarks

0.040 to 0.825	10	Comparison to gage blocks
> 0.825 to 1.510	8	Comparison to gage blocks
> 1.510 to 2.510	9	Comparison to gage blocks
> 2.510 to 4.510	14	Comparison to gage blocks
> 4.510 to 6.510	21	Comparison to gage blocks
> 6.510 to 9.010	29	Comparison to gage blocks
> 9.010 to 12.010	40	Comparison to gage blocks
> 12.010 to 13.25	44	Comparison to gage blocks

ELECTROMAGNETICS - DC/LOW FREQUENCY

NVLAP Code: 20/E02

AC Current

Best Uncertainty ($\pm$) in ppm ^{note 1}

Frequency in Hz

Range	10	20	40	400 to 10 k
10 mA	278	209	142	132
20 mA	278	209	142	132
30 mA	278	209	142	132
50 mA	300	202	124	109
100 mA	278	209	142	132

2007-04-01 through 2008-03-31

Sally S. Bruce

Effective dates

For the National Institute of Standards and Technology