Project 6468-07-1950 July 24, 2008





# **CALIBRATION LABORATORIES**

## NVLAP LAB CODE 105014-0

Scope Revised : 2007-04-20

l kg	0.052	Echelon 1
500 g	0.043	Echelon 1
300 g	0.041	Echelon 1
200 g	0.034	Echelon 1
100 g	0.020	Echelon I
50 g	0.013	Echelon I
30 g	0.013	Echelon I
20 g	0.0095	Echelon 1
10 g	0.0073	Echelon I
5 g	0.0048	Echelon I
3 g	0.0038	Echelon I
2 g	0.0029	Echelon I
l g	0.0030	Echelon I
500 mg	0.0017	Echelon I
300 mg	0.0013	Echelon I
200 mg	0.0010	Echelon I
100 mg	0.0009	Echelon I
50 mg	0.0007	Echelon I
30 mg	0.0007	Echelon I
20 mg	0.0005	Echelon 1
10 mg	0.0005	Echelon I
5 mg	0.0006	Echelon I
3 mg	0.0006	Echelon I
2 mg	0.0005	Echelon I
l mg	0.0005	Echelon I
30 kg	56	Echelon II
20 kg	22	Echelon II

Dally S. Buce

2007-04-01 through 2008-03-31

Effective dates

For the National Institute of Standards and Technology

Page 8 of 10

NVLAP-01S (REV: 2004-10-31)

NVLAP LAB CODE 105014-0 Scope Revised : 2007-04-20



## CALIBRATION LABORATORIES

#### THERMODYNAMIC

*NVLAP Code:* 20/T05 Pressure

Range in psi	Best Uncertainty (±) in ppm <sup>note 1</sup>	Remarks
0 to 50	25	Gas
> 50 to 1450	48	Gas
> 1450 to 16 000	90	Gas
> 1000 to 10 000	62	Oil
> 10 000 to 30 000	113	Oil
> 30 000 to 50 000	213	Oil

## NVLAP Code: 20/T07

Resistance Thermometry

	Range	Best Uncertainty (±) in mK <sup>nate 1</sup>	Remarks
77.348 K	-195.80 °C	4,0	BP LN <sub>2</sub> by Comparison
234.3156 K	-38,83 °C	0.7	TPHg
273.16 K	0.01 °C	0.6	TPW
505.078 K	231.93 °C	1.6	FPSn
692.77 K	419.53 °C	2.3	FPZn

Sally S. Bruces

2007-04-01 through 2008-03-31

Effective dates

For the National Institute of Standards and Technology

Page 9 of 10

NVLAP 01S (REV. 2004; 10-31)



Project 6468-07-1950 July 24, 2008



## CALIBRATION LABORATORIES



NT OF A

## NVLAP LAB CODE 105014-0

Scope Revised : 2007-04-20

- Represents an expanded uncertainty using a coverage factor, k = 2, at an approximate level of confidence of 95 %.
- 2. Approximate value. Actual value determined by the test statistics.
- 3. All ACV measurements performed via AC/DC transfer system.
- 4. Uncertainties listed are representative of the laboratory's accredited capabilities within the stated ranges. Accreditation is not limited to only those fixed values shown.
- 5. Dependent upon principle of operation of device being calibrated and its performance relative to standards at the time of the test.
- 6. The equation: uncert. =  $(A + B/mVDC^2)^{0.5}$  (where A = 0.16 and B = 0.013333) is provided in order for potential customers to calculate approximate uncertainties for values down to 1 mV. Example: uncertainty at 1 mVDC would calculate to approximately ±115.47 ppm.
- 7. The laboratory maintains Echelon II capability for ranges 20 kg to 1 mg and separate Echelon III for all ranges.
- 8. Avoirdupois mass calibration services are available by comparison to equivalent metric standards. Uncertainties may be appropriately larger.

2007-04-01 through 2008-03-31

Effective dates

Sally S. Buce

For the National Institute of Standards and Technology

Page 10 of 10

NVLAP-01S (REV: 2004-10-31)

### **GEOVision Borchole Geophysics depth wheel verification**

Performed by Robert Steller on September 23, 2006

	Depth reading in #1	Depth reading out	Depth reading in #2
Depth wheel	100.1 feet	99.95 feet	100,05 feet
S/N 101	(30.51 m)	(30.46 m)	(30.50 m)
500 pulse/revolution			
Circumference = 983mm			
(3225.07 millifeet)		· ·······	
Depth wheel	100.00 feet	100.05 feet	100.00 feet
S/N 102	(30,48) m	(30.50 m)	(30,48) m
500 pulse/revolution			
Circumference 994mm			
(3261.15 millifeet)			
	100.05.0	100.05.0	
Aries winch	100.05 leet	100.05 teet	100.00 feet
200 pulse/revolution	(30.50) m	(30.50 m)	(30.48) m
Circumierence 305.9mm			
Donth whool			
S/N 103			
500 pulse/revolution			
Circumference = 1000mm			
(3.281 feet)			
· · · · · · · · · · · · · · · · · · ·			
Comprobe winch			
500 pulse/revolution			
Circumference 1000mm			
(3.281 feet)			

All measurements taken with a Stanley 100ft flexible stainless steel tape model number 34-130, and a Keeson 300 foot fiberglass tape, both marked in feet, inches and 1/8ths of inches. Enough cable was spooled off of the winch to allow the cable and tape measures to be laid flat on the parking lot surface side-by-side. A permanent marker was used to mark a 100.0 foot interval on the cable, and the marks were also tagged with electrical tape for visibility. The cable was then spooled back onto the winch. When the first mark was at the top of the measuring wheel, a matching permanent mark was placed, and the recording system (Robertson Micrologger) was set to 0.0 feet depth. The cable was spooled in to the second mark, and the distance was recorded. The process was repeated one more time to spool the cable back onto the winch, and the distance was recorded.

Estimated accuracy is of these measurements is  $\pm/-0.1$  foot or  $\pm/-0.03$ m.

## GEOVision Suspension PS probe Receiver 1–Receiver 2 (R1-R2) spacing verification

Performed by Robert Steller on September 23, 2006

	R2 center to R1	R2 center to R1	R1 bottom to source center hanging
	center hanging	center hanging	submerged with 1m isolation tube
	dry	submerged	S/N 280068
Receiver S/N	40.2in	40.0in	76.0in
30086	1.02m	1.02m	1.93m
Receiver S/N	39.8in	39.6in	75.7in
20042	1.01m	1.01m	1.92m
Receiver S/N	40.2in	40.0in	76.0in
12008	1.02m	1.02m	1.93m

All measurements taken with a Lufkin 3.7m flexible steel tape model number HV1034DM, marked in mm and  $100^{th}$  of feet. Probe suspended in 3-inch diameter clear PVC pipe, using chain clamp placed between bottom and center of Receiver 2 hard section (See Figure). Probe "bounced" to establish unrestricted hanging length before measurement. Probe allowed to relax for 5 minutes prior to each measurement. Water level set to submerge bottom of Receiver 2 hard section.. Estimated accuracy due to hysterisis in rubber section approximately +/- 0.01' or +/- 0.003m.



GEOVision Report 8083-03 FPL Turkey Point COL Boring Geophysical Logging rev 0

Page 393 of 673