



SOUTHWEST RESEARCH INSTITUTE®

6220 Culebra Road, P.O. Drawer 28510
Institute Quality Systems
Institute Calibration Laboratory
Phone: 210-522-5215 Fax 210-522-4834



Calibration Laboratory
Certificate #0972-01

Certificate of Calibration

Cost Center: DIV20

Mail Stop: B51

Customer: DON BANNON

Manufacturer/Model: OMEGA / TYPE B

Description: THERMOCOUPLE

Serial Number: 12768

Asset Number: 012768

Procedure: TEMPERATURE PROBES - 5 JUN 06

Work Order: 303087955

Date Issued: 5-Jun-2009

Date Calibrated: 5-Jun-2009

*** Date Due :** 5-Jun-2010

**** Results:** FOUND-LEFT

Temperature: 78°F

Humidity: 38 %

This certificate documents traceability to the National Institute of Standards and Technology (NIST) and the International System of Units (SI). The Laboratory quality system conforms to ISO/IEC 17025, 2005, ANSI/NCSL Z540-1-1994 and relevant requirements of the ISO 9000-2000 standard. This certificate shall not be reproduced, except in full, without the written approval of the Southwest Research Institute Calibration Laboratory. This certificate shall not be used to claim product endorsement by Southwest Research Institute, American Association for Laboratory Accreditation (A2LA) or any agency of the U. S. Government. Results of this calibration relate only to the instrument described above at the time of calibration and does not imply any long term stability of the instrument.

*Determined by the customer, does not imply the instrument will remain within tolerance as any number of factors may cause an out-of-tolerance condition before this date. **Data type found in this certificate or attached measurement report must be interpreted as: Found-left - adjustment and/or repair was not performed, As-found - data is before unit is adjusted and/or repaired, As-left - data is after adjusted and/or repaired was performed. The customer has sole responsibility for determination of in-/out-of-tolerance or compliance/noncompliance.

Measurement uncertainty calculated in accordance with the method described in the ISO "Guide to the Expression of Uncertainty in Measurement" (GUM), for a confidence level of approximately 95 percent using a coverage factor of $k=2$.

Remarks: Cal @ 900, 1000, 1100 °C.

Standards Used

<u>Asset #</u>	<u>Manufacturer</u>	<u>Model</u>	<u>Description</u>	<u>Cal Date</u>	<u>Due Date</u>
010814	HART SCIENTIFIC	1529	THERMOCOUPLE THERMOMETER	17-Nov-2008	17-Nov-2009
013617	HART SCIENTIFIC	5650	THERMOCOUPLE, TYPE S	8-Dec-2008	8-Dec-2009


Walt Hill

Laboratory Manager


Mark Romero

Metrology Technician

Southwest Research Institute
Calibration Laboratory
Measurement Report

Work Order:	303087955	Mfr.	Omega	Technician:	Mark Romero
Asset No.	012768	Model	Type B	Type Data:	Found-left
Serial No.	12768	Type.	Thermocouple	Cal Date:	5-Jun-09

Remarks: Limits taken from ASTM E230-02 and are based on brand new unused thermocouples.

Function/Range	Test Point	TI Reading	Difference	+/- Limit	+/- Uncertainty	Result	% Limit
	°C	°C	°C	°C	°C		
Temperature	900.470	899.411	-1.059	4.5	1.3	Pass	24%
	1000.958	1000.372	-0.586	5.0	1.3	Pass	12%
	1101.205	1101.393	0.188	5.5	1.3	Pass	3%

END OF REPORT