

SOUTHWEST RESEARCH INSTITUTE<sup>®</sup>

6220 Culebra Road, P.O. Drawer 28510 Institute Quality Systems Institute Calibration Laboratory Phone: 210-522-5215 Fax 210-522-4834

## **Certificate of Calibration**



Cost Center:DIV20WMail Stop: B51ICustomer: DON BANNONDateManufacturer/Model: VAISALA / HMP77BDescription: HUMIDITY/TEMPERATURE PROBESerial Number: Y2830017ToAsset Number: 010302Procedure: TEMPERATURE, HUMIDITY, DEW-POINT - 3 APR 08

Work Order: 303088119 Date Issued: 8-Jun-2009 Date Calibrated: 8-Jun-2009 \* Date Due : 8-Jun-2010 \*\* Results: FOUND-LEFT Temperature: 78°F Humid ty: 44 %

This certificate documents traceability to the National Institute of Standards and Technology (NIST) and the International System of Units (SI). The Laboratory quality system conforms to ISO/IEC 17025, 2005, ANSI/NCSL Z540-1-1994 and relevant requirements of the ISO 9000-2000 standard. This certificate shall not be reproduced, except in full, without the written approval of the Southwest Research Institute Calibration Laboratory. This certificate shall not be used to claim product endorsement by Southwest Research Institute, American Association for Laboratory Accreditation (A2LA) or any ager.cy of the U. S. Government. Results of this calibration relate only to the instrument described above at the time of calibration and does not imply any long term stability of the instrument.

\*Determined by the customer, does not imply the instrument will remain within tolerance as any number of factors may cause an out-of-tolerance condition before this date. \*\*Data type found in this certificate or attached measurement report must be interpreted as: Found-left - adjustment and/or repair was not performed, As-found - data is before unit is adjusted and/or repaired, As-left - data is after adjusted and/or repaired was performed. The customer has sole responsibility for determination of in-/out-of-tolerance or compliance/noncompliance.

Measurement uncertainty calculated in accordance with the method described in the ISO "Guide to the Expression of Uncertainty in Measurement" (GUM), for a confidence level of approximately 95 percent using a coverage factor of k=2.

Remarks: CALIBRATED WITH VAISALA MI70 INDICATOR S/N Y2540054.

## Standards Used

<u>Asset #</u>	<u>Manufacturer</u>	<u>Model</u>
006404	THUNDER SCIENTIFIC	2500
012305	HART SCIENTIFIC	1502A
014652	HART SCIENTIFIC	5615

Description	<u>Cal Date</u>	Due Date
HUMIDITY GENERATOR	27-May-2009	27-May-2010
TEMPERATURE READOUT	29-Apr-2009	29-Oct-2009
RTD	29-Apr-2009	29-Oct-2009

Ni Jav

Laboratory Manager

m:\A2LA OCT\_08.rpt

∕ark Romero

Metrology Technician

## Southwest Research Institute Calibration Laboratory Measurement Report

Work Order:	303088119	Mfr.	Vaisala	Technician:	Mark F	Mark Romero	
Asset No.	010302	Model	HMP77B	Type⊨Data:	E Data: Found-left		
Serial No.	Y2830017	Туре.	Humidity/Temp	Cal Date:	8-Jun-09		
Remarks: Calibrated as	s a system with V	aisala MI70. Te	mperature limit in	creased to +/-	1 °C per custodi	an.	
Function/Range	Test Point	TI Reading	Difference	+/- Limit	+/- Uncertaint	y Result	% Limit
Temperature	°C	°C	°C	°C	°C		
at 20% RH	25.17	25.21	0.04	1	0.12	Pass	4%
40% RH	25.20	25.20	0.00			Pass	0%
60% RH	25.21	25.22	0.01			Pass	1%
80% RH	25.24	25.23	-0.01			Pass	1%
	°C	°C	°C				
at 20% RH	49.49	49.11	-0.38			Pass	38%
40% RH	49.50	49.13	-0.37			Pass	37%
60% RH	49.66	49.21	-0.45			Pass	45%
80% RH	49.76	49.45	-0.31			Pass	31%
	°C	°C	°C				
at 20% RH	68.33	67.86	-0.47			Pass	47%
40% RH	68.33	67.85	-0.48			Pass	48%
60% RH	68.35	67.85	-0.50			Pass	50%
80% RH	69.38	69.87	0.49			Pass	49%
Humidity	% RH	% RH	% RH	% RH	% RH		
at 25 °C	20.00	18.69	-1.31	2	0.68	Pass	66%
	40.00	38.76	-1.24			Pass	62%
	60.00	58.63	-1.37			Pass	69%
	80.00	78.66	-1.34			Pass	67%
at 50 °C	20.00	18.77	-1.23			Pass	62%
	40.00	39.45	-0.55			Pass	28%
	60.00	59.74	-0.26			Pass	13%
	80.00	79.43	-0.57			Pass	29%
at 70 °C	20.00	19.46	-0.54			Pass	27%
	40.00	39.83	-0.17			Pass	9%
	60.00	60.21	0.21			Pass	11%
	80.07	79.11	-0.96			Pass	48%
		END	OF REPORT				