Roche

July 14, 2009

J-6

Dennis Lawyer:
U.S. NRC Region I
475 Allendale Road
King of Prussia, PA 19406-1415

Mail Control 143835

03035904

Subject: Corrected Radiation Survey - Laboratory Diagram
Memory Pharmaceutical – termination of License 29-30707-01

Dear Mr. Lawyer

This letter is submitted in reply to your request during our telephone discussion today.

Enclosed is a corrected lab survey diagram, which now identifies the lab as number 1080. No other changes were effected. The number 1058 was inadvertently retained while copying one lab diagram to another.

If you have any questions on this submission, or require additional information, please contact the undersigned directly, or Mr. Jeff Guenther at telephone 973-235-6608.

Sincerely,

Michael. J. Drzyzga
Manager, Site EH&S
Radiation Safety Officer
(V) 973-235-3418
(F) 973-235-4893
(C) 201-310-2943

FAX to 610-337-5269

w:\wpdoc\

Hoffmann-La Roche Inc. 340 Kingsland Street
Nutley, New Jersey 07110-1199

ROCHE> RADIATION SAFETY OFFICE CONTAMINATION SURVEY survform.xls

143835

NMSS/RGN1 MATERIALS-002

1080 corr. 7/14/09 MJ Dougherty

DATE: 7/9/09 LOCATION: Rm ~~1058~~ PRU: WEEK:

RADIATION USERS:

DEPT CODE:

COMMENTS:

Survey Meter Lud 16 w/44-142

Serial # 166343 / 246235

Check Source 4000 cpm

RSO staff: JMS

Resurvey: Date:

Comments:

#	NET DPM H-3	C-14/S-35	P-32	Other
1	<15	<15	<15	
2	26			
3	<15			
4				
5				
6				
7				
8				
9				
10				
11				
12				
13				
14				
15				
16				
17				
18		JMS		
19				
20				



RSO review: MJD Date: 7/16/09

Follow - up:

ROCHE> RADIATION SAFETY OFFICE CONTAMINATION SURVEY survform.xls