



**UNITED STATES
NUCLEAR REGULATORY COMMISSION**

REGION III
2443 WARRENVILLE ROAD, SUITE 210
LISLE, ILLINOIS 60532-4352

JUL 08 2009

James Schweitzer, Ph.D.
Radiation Safety Officer
Radiological & Environmental Management
550 Stadium Drive Mall
West Lafayette, IN 47907-1662


Dear Dr. Schweitzer

This refers to letter dated May 19, 2009, requesting the use of krypton-85 contained in a TSI Aerosol Neutralizer Model 3077A to take air samples from a Beechcraft aircraft. As we discussed on July 3, 2009, this device is generally licensed and you may use this device under your general license. Therefore, there is no need to amend your NRC specific license of broad scope. Your request dated May 19, 2009 is, therefore, voided.

If you have any questions or require clarification on any of the information stated above, you may contact me at (623) 214-5213.

In accordance with 10 CFR 2.390 of the NRC's "Rules of Practice," a copy of this letter and its enclosure will be available electronically for public inspection in the NRC Public Document Room or from the Publicly Available Records (PARS) component of NRC's document system (ADAMS). The NRC's document system is accessible from the NRC Web site at <http://www.nrc.gov/reading-rm/adams.html> (the Public Electronic Reading Room).

Sincerely,


James R. Mullauer, M.H.S.
Health Physicist
Materials Licensing Branch

License No. 13-02812-04
Docket No. 030-00696