



a division of Cardiology Consultants, PA.  
www.CardiacDiagnostic.net

NEW CASTLE COUNTY  
3105 LIMESTONE ROAD  
SUITE 202  
WILMINGTON, DE 19808  
(302) 994-6500  
(302) 994-6922 FAX

3521 SILVERSIDE ROAD  
SUITE 1-A  
WILMINGTON, DE 19810  
(302) 477-6500  
(302) 477-6509 FAX

114 SANDHILL DRIVE  
KETLAY PLAZA  
MIDDLETOWN, DE 19709  
(302) 376-4340  
(302) 376-4350 FAX

KENT COUNTY  
1100 FORREST AVENUE  
DOVER, DE 19904  
(302) 672-1890  
(302) 672-1899 FAX

SUSSEX COUNTY  
550 SOUTH DuPONT PLAZA  
MILFORD, DE 19963  
(302) 422-6050  
(302) 422-6820 FAX

16704 KINGS HIGHWAY  
LEWES, DE 19958  
(302) 644-3822  
(302) 645-1228 FAX

609 ATLANTIC AVENUE  
MILLVILLE, DE 19970  
(302) 645-1233  
(302) 541-8425 FAX

July 15, 2009

Licensing Branch  
Nuclear Materials Safety Section  
Division of Safety and Safeguards  
U.S. N.R.C. – Region 1  
475 Allendale Road  
King of Prussia, PA 19406

Br. 2

2009 JUL 20 PM 12:41

RECEIVED

NRC License No. 07-27897-01

03029521

Dear License Reviewer:

As per the requirement for amending our current license to reflect the new address of our Milford office, 802 North DuPont Highway, Milford, DE, I am enclosing a copy of the close out survey and wipes for our old office address of 550 South DuPont Highway, Milford, DE. We would like an amendment removing our old facility from the license.

The survey and wipes were performed by AnnMarie Messick, NCT, CNMT on July 14, 2009.

Please contact me at 302-452-3462 with any questions. Alternate contacts are Barry Dennenberg, MD (RSO) at 302-421-9721, AnnMarie Messick, NCT, CNMT at 302-994-6500, or Jack Olley of Walter Robinson Associates at 1-800-446-7622.

Thank you for your consideration in this matter.

Sincerely,

Phillip Chapman, Jr.  
Executive Director  
Cardiology Consultants, PA

143930

NW30/RGNI MATER.ALS-002

Cardiac Diagnostic Center- Milford  
Decommission Surveys

Area Surveys

Background: .01mR/hr

Hot Lab:

Lead Cabinet: .01mR/hr

Counter: .01mR/hr

Sink: .01mR/hr

Computer Desk: .01mR/hr

Floor: .01mR/hr

Prep Area:

Floor: .01mR/hr

Sink: .01mR/hr

Counter: .01mR/hr

Stress Test Room:

Counter: .01mR/hr

Sink: .01mR/hr

Floor injection area: .01mR/hr

Floor treadmill area: .01mR/hr

Floor Stretcher area: .01mR/hr

Camera Room:

Floor at camera console: .01mR/hr

Floor at camera: .01mR/hr

Hallway Floor: .01mR/hr

Cardiac Diagnostic Center- Milford  
Decommission Surveys

Area Wipe Surveys

Background: 418 cpm

Hot Lab:

Lead Cabinet: 400 cpm

Counter: 433 cpm

Sink: 438 cpm

Computer Desk: 401 cpm

Floor: 385 cpm

Prep Area:

Floor: 399 cpm

Sink: 447 cpm

Counter: 426 cpm

Stress Test Room:

Counter: 451 cpm

Sink: 430 cpm

Floor injection area: 396 cpm

Floor treadmill area: 449 cpm

Floor Stretcher area: 435 cpm

Camera Room:

Floor at camera console: 415 cpm

Floor at camera: 382 cpm

Hallway Floor: 443 cpm

This is to acknowledge the receipt of your letter/application dated

7/15/09, and to inform you that the initial processing which includes an administrative review has been performed.

Amendment (07-27897-01) There were no administrative omissions. Your application was assigned to a technical reviewer. Please note that the technical review may identify additional omissions or require additional information.

Please provide to this office within 30 days of your receipt of this card

A copy of your action has been forwarded to our License Fee & Accounts Receivable Branch, who will contact you separately if there is a fee issue involved.

Your action has been assigned Mail Control Number 143930.  
When calling to inquire about this action, please refer to this control number.  
You may call us on (610) 337-5398, or 337-5260.