

CONVERSATION RECORD

TIME | DATE  
7-20-2009

VISIT       CONFERENCE       TELEPHONE  
  
 INCOMING  
 OUTGOING

NAME OF PERSON(S) CONTACTED OR IN CONTACT	ORGANIZATION (OFFICE, DEPT ETC.)	TELEPHONE NO.
Bari P. Chase, RSO Lab Manager	Schultz & Summers Engineering, Inc. 4800 West Boulevard Poplar Bluff, MO 63901	573-686-0806 FAX -0550

SUBJECT  
Control No. 317933- Re - Amendment request for renewal of License No. 24-32159-01, by application dated February 3, 2009, regarding moisture/density gauges.

SUMMARY  
In a telephone conversation with Mr. Chase on this date, I noted that the renewal application did not address temporary job site use of the portable gauges. He confirmed that the licensee does use the gauges at temporary job sites and confirmed that it was their intent to request that temporary job site usage continue to be authorized in the subject license as renewed, to which I agreed to do. I also asked whether "and" or "&" should be used in the licensee's name in that the renewal application used "and" while the current license and letter head submitted with the renewal application used the "&". Mr. Chase confirmed that the "&" was in deed correct. We also discussed make and model of gauges and possession limits. I noted that at the time of their NRC inspections they did not possess any Troxler Model 3450 gauges which were listed as a separate line item on their license even though they are part of the Troxler 3400 Series. He confirmed that they have never possessed the model 3450 gauges and that they currently possess two Troxler Model 3440 and two Troxler Model 3440 Plus gauges which both are covered by the Troxler 3400 Series of gauges, as per the Sealed Source and Device Registry . Regarding model of gauges and possession limits to be placed on the license as renewed, Mr. Chase requested authorization to possess a total of six Troxler Model 3400 Series gauges each of which can contain a sealed source of Cs-137 with an activity up to 9 millicuries and also a sealed source of Am-241 with an activity up to 44 millicuries. I agreed to renew the license as noted above.

In accordance with 10 CFR 2.390 of the NRC's "Rules of Practice," a copy of this conversation record will be available electronically for public inspection in the NRC Public Document Room or from the Publicly Available Records (PARS) component of NRC's document system (ADAMS). The NRC's document system is accessible from the NRC Web site at <http://www.nrc.gov/reading-rm/adams.html> (the Public Electronic Reading Room).

ACTION REQUIRED  
License reviewer to consider the above information in evaluation of subject amendment request for license renewal.

NAME OF PERSON DOCUMENTING CONVERSATION	SIGNATURE	DATE
Loren J. Hueter	<i>Loren J. Hueter</i>	7-20-2009

ACTION TAKEN	SIGNATURE	TITLE	DATE