



SOUTHWEST RESEARCH INSTITUTE®

6220 Culebra Road, P.O. Drawer 28510
Institute Quality Systems
Institute Calibration Laboratory
Phone: 210-522-5215 Fax 210-522-4834



Calibration Laboratory
Certificate #0972-01

Certificate of Calibration

Cost Center: DIV20

Mail Stop: B51

Customer: DON BANNON

Manufacturer/Model: FISHER SCIENTIFIC / 15-078-39

Description: DIGITAL THERMOMETER

Serial Number: 41523645

Asset Number: 011675

Procedure: DIGITAL THERMOMETERS (TC/RTD) - 12 JUN 09

Work Order: 303088418

Date Issued: 25-Jun-2009

Date Calibrated: 25-Jun-2009

*** Date Due :** 25-Jun-2010

**** Results:** FOUND-LEFT

Temperature: 74°F

Humidity: 40 %

This certificate documents traceability to the National Institute of Standards and Technology (NIST) and the International System of Units (SI). The Laboratory quality system conforms to ISO/IEC 17025, 2005, ANSI/NCSL Z540-1-1994 and relevant requirements of the ISO 9000-2000 standard. This certificate shall not be reproduced, except in full, without the written approval of the Southwest Research Institute Calibration Laboratory. This certificate shall not be used to claim product endorsement by Southwest Research Institute, American Association for Laboratory Accreditation (A2LA) or any agency of the U. S. Government. Results of this calibration relate only to the instrument described above at the time of calibration and does not imply any long term stability of the instrument.


*Determined by the customer, does not imply the instrument will remain within tolerance as any number of factors may cause an out-of-tolerance condition before this date. **Data type found in this certificate or attached measurement report must be interpreted as: Found-left - adjustment and/or repair was not performed, As-found - data is before unit is adjusted and/or repaired, As-left - data is after adjusted and/or repaired was performed. The customer has sole responsibility for determination of in-/out-of-tolerance or compliance/noncompliance.

Measurement uncertainty calculated in accordance with the method described in the ISO "Guide to the Expression of Uncertainty in Measurement" (GUM), for a confidence level of approximately 95 percent using a coverage factor of $k=2$.

Remarks: Exceeds 70% of limit at 1300 °C. See Measurement Report

Standards Used

<u>Asset #</u>	<u>Manufacturer</u>	<u>Model</u>	<u>Description</u>	<u>Cal Date</u>	<u>Due Date</u>
006413	FLUKE	5520A/SC1100	MULTI-PRODUCT CALIBRATOR	9-Sep-2008	9-Sep-2009


Walt Hill

Laboratory Manager


Bob Trollinger

Metrology Technician

Southwest Research Institute
Calibration Laboratory
Measurement Report

Work Order:	303088418	Mfr:	Fisher Scientific	Technician:	blt
Asset No:	011675	Model:	15-078-39	Type Data:	Found-left
Serial No:	41523645	Type:	Temperature Meter	Cal Date:	25-Jun-09
Remarks: Exceeds 70% of limit. Not adjustable.					

Function/Range	Test Point	TI Reading	Difference	+/- Limit	+/- Uncertainty	Result	% Limit
T1 Type K	°F	°F	°F	°F	°F		
	-58	-57.0	1.0	1.7	0.44	Pass	60%
	355	356	1	3	0.86	Pass	39%
	765	764	1	4	0.86	Pass	26%
	1180	1183	3	5	0.86	Pass	60%
	1995	1998	3	7	1.1	Pass	40%
	°C	°C	°C	°C	°C		
	-50	-49.4	0.6	1.2	0.24	Pass	52%
	220	220	0	2	0.66	Pass	0%
	760	762	2	3	0.66	Pass	61%
	1030	1031	1	4	0.66	Pass	24%
	1300	1296	-4	5	0.75	Pass	82%
T2 Type K	°F	°F	°F	°F	°F		
	-58	-56.9	1.1	1.7	0.44	Pass	66%
	355	356	1	3	0.86	Pass	39%
	765	764	1	4	0.86	Pass	26%
	1180	1183	3	5	0.86	Pass	60%
	1995	1998	3	7	1.1	Pass	40%
	°C	°C	°C	°C	°C		
	-50	-49.6	0.4	1.2	0.24	Pass	35%
	220	219	-1	2	0.66	Pass	60%
	760	762	2	3	0.66	Pass	61%
	1030	1031	1	4	0.66	Pass	24%
	1300	1296	-4	5	0.75	Pass	82%
Difference T1-T2	°F	°F	°F	°F	°F		
	0	0.0	0.0	1.5	0.06	Pass	0%
T1-T2	°C	°C	°C	°C	°C		
	0	0.0	0.0	1.0	0.06	Pass	0%

END OF REPORT

Explanation of Measurement Report Results

“When statements of compliance (Pass/Fail) are made, the uncertainty of measurement shall be taken into account”. Reference ISO/IEC 17025:2005, 5.10.4.2

This explanation is provided to you because the instrument submitted for calibration has one or more of the following results.

Result

Pass – measured value or test is within the \pm limit, in tolerance, with a confidence level of 95 percent.

_____ Pass? – measured value is *within* the \pm limit, but by a margin less than half of the uncertainty interval and has a confidence level of less than 95 percent of being in tolerance. Adjustment is made and the measurement is repeated. If adjustment or repair is not possible or fails to improve the results, then the customer must determine in or out of tolerance.

_____ Fail? – measured value is *outside* the \pm limit, but by a margin less than half of the uncertainty interval and is reported as out of tolerance but it is not possible to state this with a 95 percent confidence level. Adjustment is made and the measurement is repeated. If adjustment or repair is not possible or fails to improve the results, then the customer must determine if out of tolerance action is necessary.

_____ Fail – measured value is *outside* the \pm limits with a 95 percent confidence level. Adjustment is made and the measurement is repeated for As-left data. If adjustment or repair is not possible or fails to improve the results, then the customer must determine if the measured value is in compliance for the intended use.

%Limit

✓ Adjustment is made, if possible, when the As-found measured value is equal to or greater than 70 percent of the \pm limit. If adjustment is not possible or did not lower the As-left reading below 70 percent, the customer shall determine if the instrument is suitable for their requirements.

Type Data

Found-left All measurements were in tolerance and no adjustments or repairs were performed.

As-found One or more measurements were other than Pass or exceeded 70 percent of the \pm limit and adjustment or repairs were performed.

As-left Results of measurements after adjustment or repair.

Uncertainty

Best estimate of the dispersion of the measured value that could be contributed by the; standard, environment, repeatability of the measurement process, characterizes of the instrument being calibrated (i.e. resolution) etc.

Please call extension 5215 for questions or additional information.