



SOUTHWEST RESEARCH INSTITUTE®

6220 Culebra Road, P.O. Drawer 28510
Institute Quality Systems
Institute Calibration Laboratory
Phone: 210-522-5215 Fax 210-522-4834



Calibration Laboratory
Certificate #0972-01

Certificate of Calibration

Cost Center: DIV20

Mail Stop: B51

Customer: DON BANNON

Manufacturer/Model: DURO-SENSE / TYPE K

Description: THERMOCOUPLE

Serial Number: 12397

Asset Number: 012397

Procedure: TEMPERATURE PROBES - 5 JUN 06

Work Order: 303088983

Date Issued: 17-Jul-2009

Date Calibrated: 16-Jul-2009

* Date Due : 16-Jul-2010

** Results: FOUND-LEFT

Temperature: 75°F

Humidity: 50 %

This certificate documents traceability to the National Institute of Standards and Technology (NIST) and the International System of Units (SI). The Laboratory quality system conforms to ISO/IEC 17025, 2005, ANSI/NCSL Z540-1-1994 and relevant requirements of the ISO 9000-2000 standard. This certificate shall not be reproduced, except in full, without the written approval of the Southwest Research Institute Calibration Laboratory. This certificate shall not be used to claim product endorsement by Southwest Research Institute, American Association for Laboratory Accreditation (A2LA) or any agency of the U. S. Government. Results of this calibration relate only to the instrument described above at the time of calibration and does not imply any long term stability of the instrument.

*Determined by the customer, does not imply the instrument will remain within tolerance as any number of factors may cause an out-of-tolerance condition before this date. **Data type found in this certificate or attached measurement report must be interpreted as: Found-left - adjustment and/or repair was not performed, As-found - data is before unit is adjusted and/or repaired, As-left - data is after adjusted and/or repaired was performed. The customer has sole responsibility for determination of in-/out-of-tolerance or compliance/noncompliance.

Measurement uncertainty calculated in accordance with the method described in the ISO "Guide to the Expression of Uncertainty in Measurement" (GUM), for a confidence level of approximately 95 percent using a coverage factor of $k=2$.

Remarks: 3 POINT CAL (25 °, 100 °, 200 °C)

Standards Used

Asset #	Manufacturer	Model	Description	Cal Date	Due Date
009137	HART SCIENTIFIC	1575	SUPER THERMOMETER	15-May-2009	15-Nov-2009
013908	HART SCIENTIFIC	5628	SPRT	20-Feb-2008	20-Feb-2010
015240	HART SCIENTIFIC	2566	TC SCANNER, 12-CHANNEL MODULE	10-Dec-2008	10-Dec-2009


Walt Hill

Laboratory Manager


Mark Romero

Metrology Technician

Southwest Research Institute
Calibration Laboratory
Measurement Report

Work Order:	303088983	Mfr.	Duro-Sense	Technician:	Mark Romero
Asset No.	012397	Model	Type K	Type Data:	Found-left
Serial No.	12397	Type.	Thermocouple	Cal Date:	16-Jul-09

Remarks: Limits taken from ASTM E230-02 and are based on brand new unused thermocouples.

Function/Range	Test Point	TI Reading	Difference	+/- Limit	+/- Uncertainty	Result	% Limit
Temperature	°C	°C	°C	°C	°C		
	25.115	25.014	-0.101	2.2	0.5	Pass	5%
	100.159	100.351	0.192	2.2	0.5	Pass	9%
	200.153	199.919	-0.234	2.2	0.5	Pass	11%

END OF REPORT