



# SOUTHWEST RESEARCH INSTITUTE®

6220 Culebra Road, P.O. Drawer 28510  
Institute Quality Systems  
Institute Calibration Laboratory  
Phone: 210-522-5215 Fax 210-522-4834

## Certificate of Calibration

**Cost Center:** DIV20

**Mail Stop:** B51

**Customer:** DON BANNON

**Manufacturer/Model:** KEITHLEY / 617

**Description:** ELECTROMETER

**Serial Number:** 537418

**Asset Number:** 001044

**Procedure:** KEITHLEY 617 - 14 SEP 06

**Work Order:** 303088969

**Date Issued:** 15-Jul-2009

**Date Calibrated:** 15-Jul-2009

**\* Date Due :** 15-Jan-2010

**\*\* Results:** FOUND-LEFT

**Temperature:** 74°F

**Humidity:** 40 %

This certificate documents traceability to the National Institute of Standards and Technology (NIST) and the International System of Units (SI). The Laboratory quality system conforms to ISO/IEC 17025, 2005, ANSI/NCSL Z540-1-1994 and relevant requirements of the ISO 9000-2000 standard. This certificate shall not be reproduced, except in full, without the written approval of the Southwest Research Institute Calibration Laboratory. This certificate shall not be used to claim product endorsement by Southwest Research Institute or any agency of the U. S. Government. Results of this calibration relate only to the instrument described above at the time of calibration and does not imply any long term stability of the instrument.

\*Determined by the customer, does not imply the instrument will remain within tolerance as any number of factors may cause an out-of-tolerance condition before this date. \*\*Data type found in this certificate or attached measurement report must be interpreted as: Found-left - adjustment and/or repair was not performed, As-found - data is before unit is adjusted and/or repaired, As-left - data is after adjusted and/or repaired was performed. The customer has sole responsibility for determination of in-/out-of-tolerance or compliance/noncompliance.

Measurement uncertainty calculated in accordance with the method described in the ISO "Guide to the Expression of Uncertainty in Measurement" (GUM), for a confidence level of approximately 95 percent using a coverage factor of  $k=2$ .

**Remarks:** Limited Cal: Coulombs, 2 pA, 20 pA and 200 pA not calibrated.

### Standards Used

<u>Asset #</u>	<u>Manufacturer</u>	<u>Model</u>	<u>Description</u>	<u>Cal Date</u>	<u>Due Date</u>
000182	FLUKE	5700A/EP	CALIBRATOR	5-May-2009	5-Aug-2009
000201	FLUKE	5725A	AMPLIFIER	5-May-2009	5-Aug-2009
001022	FLUKE	742A-10K	STANDARD RESISTOR	5-May-2009	5-May-2010
001505	HEWLETT-PACKARD	3458A/OPT 002	MULTIMETER	29-Jan-2009	29-Jan-2010
012977	FLUKE	8508A-7000K	STANDARD RESISTOR	22-Aug-2008	22-Aug-2009
014915	FLUKE	742A-1M	STANDARD RESISTOR	21-Aug-2008	21-Aug-2009
014917	FLUKE	742A-10M	STANDARD RESISTOR	26-Aug-2008	26-Aug-2009
014920	FLUKE	742A-100K	STANDARD RESISTOR	21-Aug-2008	21-Aug-2009

Walt Hill

Laboratory Manager

Bob Trollinger

Metrology Technician

Southwest Research Institute  
Calibration Laboratory  
Measurement Report

Work Order:	303088969	Mfr:	KEITHLEY	Technician:	blt
Asset No:	001044	Model:	617	Cal Date:	15-Jul-09
Serial No:	537418	Type:	ELECTROMETER		
Remarks:					
Coulombs, 2 pA, 20 pA and 200 pA not calibrated.					

Function/Range	Test Point	TI Reading	Difference	+/-Test Limits	+/-Uncertainty	Found/Left
DC Amps	mAmps	mAmps	mAmps	mAmps	mAmps	Result
20 mA	19.000	19.002	0.002	0.030	0.0014	Pass
2 mA	1.9000	1.8998	-0.0002	0.0033	0.00014	Pass
	uAmps	uAmps	uAmps	uAmps	uAmps	
200 uA	190.00	190.02	0.02	0.30	0.029	Pass
20 uA	19.000	19.001	0.001	0.030	0.0019	Pass
2 uA	1.9000	1.8993	-0.0007	0.0033	0.00022	Pass
	nAmps	nAmps	nAmps	nAmps	nAmps	
200 nA	190.00	189.95	-0.05	0.49	0.021	Pass
20 nA	19.000	18.987	-0.013	0.049	0.0017	Pass
2 nA	1.9000	1.9026	0.0026	0.0053	0.00017	Pass
	pAmps	pAmps	pAmps	pAmps	pAmps	
DCV	mVolts	mVolts	mVolts	mVolts	mVolts	
200 mVolt	190.00	189.97	-0.03	0.14	0.012	Pass
	Volts	Volts	Volts	Volts	Volts	
2 Volt	1.9000	1.9000	0.0000	0.0011	0.00012	Pass
20 Volt	19.000	19.000	0.000	0.011	0.0012	Pass
200 Volt	190.00	190.01	0.01	0.14	0.012	Pass
Resistance	GOhm	GOhm	GOhm	GOhm	GOhm	
20 GOhm	1.000	1.00	0.000	0.02	0.020	Pass
2 GOhm	0.9990	1.007	0.0080	0.02	0.00025	Pass
	MOhm	MOhm	MOhm	MOhm	MOhm	
200 MOhm	100.00	100.07	0.07	0.31	0.10	Pass
20 MOhm	10.000	10.001	0.001	0.026	0.0012	Pass
2 MOhm	1.0000	0.9998	-0.0002	0.0026	0.00012	Pass
	kOhm	kOhm	kOhm	kOhm	kOhm	
200 kOhm	100.00	99.99	-0.01	0.26	0.012	Pass
20 kOhm	10.000	10.000	0.000	0.016	0.0017	Pass
2 kOhm	1.0000	1.0005	0.0005	0.00211	0.00010	Pass
Voltage Source	Volts	Volts	Volts	Volts	Volts	
	0.000	-0.032	-0.032	0.050	0.0012	Pass
	1.000	1.015	0.015	0.052	0.0012	Pass
	10.000	10.003	0.003	0.070	0.0012	Pass
	25.000	24.994	-0.006	0.100	0.0015	Pass
	50.000	49.975	-0.025	0.150	0.0015	Pass
	100.000	99.942	-0.058	0.250	0.0012	Pass

END OF REPORT