

# SOUTHWEST RESEARCH INSTITUTE®

6220 Culebra Road, P.O. Drawer 28510  
Institute Quality Systems  
Institute Calibration Laboratory  
Phone: 210-522-5215 Fax 210-522-4834



Calibration Laboratory  
Certificate #0972-01

## Certificate of Calibration

**Cost Center:** DIV20

**Mail Stop:** B51

**Customer:** DON BANNON

**Manufacturer/Model:** TROEMNER / 20G

**Description:** WEIGHT, CLASS 1

**Serial Number:** 66388

**Asset Number:** 009341

**Procedure:** WEIGHTS - 10 DEC 07

**Work Order:** 303088975

**Date Issued:** 16-Jul-2009

**Date Calibrated:** 16-Jul-2009

**\* Date Due :** 16-Jul-2010

**\*\* Results:** FOUND-LEFT

**Temperature:** 68°F

**Humidity:** 41 %

This certificate documents traceability to the National Institute of Standards and Technology (NIST) and the International System of Units (SI). The Laboratory quality system conforms to ISO/IEC 17025, 2005, ANSI/NCCL Z540-1-1994 and relevant requirements of the ISO 9000-2000 standard. This certificate shall not be reproduced, except in full, without the written approval of the Southwest Research Institute Calibration Laboratory. This certificate shall not be used to claim product endorsement by Southwest Research Institute, American Association for Laboratory Accreditation (A2LA) or any agency of the U. S. Government. Results of this calibration relate only to the instrument described above at the time of calibration and does not imply any long term stability of the instrument.

\*Determined by the customer, does not imply the instrument will remain within tolerance as any number of factors may cause an out-of-tolerance condition before this date. \*\*Data type found in this certificate or attached measurement report must be interpreted as: Found-left - adjustment and/or repair was not performed, As-found - data is before unit is adjusted and/or repaired, As-left - data is after adjusted and/or repaired was performed. The customer has sole responsibility for determination of in-/out-of-tolerance or compliance/noncompliance.

Measurement uncertainty calculated in accordance with the method described in the ISO "Guide to the Expression of Uncertainty in Measurement" (GUM), for a confidence level of approximately 95 percent using a coverage factor of  $k=2$ .


**Remarks:** THE CONVENTIONAL MASS VALUE OF THE 20 g WGT IS: 20.000 034 g UNC +/-0.028 mg. CLASS 1 TOL +/-0.07 mg.

### Standards Used

<u>Asset #</u>	<u>Manufacturer</u>	<u>Model</u>	<u>Description</u>	<u>Cal Date</u>	<u>Due Date</u>
007102	TROEMNER	1 MG TO 100G	WEIGHT SET, CLASS E1	22-Dec-2008	22-Dec-2009
012068	SARTORIUS	CC310	MASS COMPARATOR	29-Jan-2009	29-Jan-2010

  
Walt Hill

Laboratory Manager

  
Carlos Mendoza

Metrology Technician