

# U.S. NUCLEAR REGULATORY COMMISSION

## TELEPHONE CONVERSATION RECORD

Date: 8/6/09

Time:

Mail Control or Report No(s). 2009-001 License No(s). 29-02023-06 Docket No(s). 030-32352

Name of Licensee: Ariston Radiology Associates, LLC

Name of Participant(s): Sam Brown, M.D. - RSO

Telephone No. 973-904-0890

Subject: Follow-up to inspection on 6/9/09

(NOTE: This will be used as the Documents Title in ADAMS)

On June 9, 2009 an inspection was done at this facility. A violation was identified for the HDR linearity. The license was not doing the linearity over the range of use. Treatments of up to 500 seconds were being performed and the linearity was done out to 300 seconds. In the licensee's response to the violation they committed to doing the HDR linearity out to 700 seconds. I requested that Dr. Brown fax the results of the linearity that was done following the inspection on June 9, 2009.

Action Required: Wait for additional information

Document Availability:

Publicly Available

Non-Publicly Available

Non-Sensitive

Sensitive - Privacy Act (includes PII)

Immediate Release

Non-Sensitive Copyright

Sensitive - Internal

Normal Release Date: 8/14/2009

Sensitive- Proprietary

Sensitive - Security-Related

Delayed Release Date:

SUNSI Review Completed By: TLWeidner / RA /

Document Accession #: ML092190038