

109 South Petro Avenue Sioux Falls, SD 57107 Phone: 605.330.9060 Toll Free: 800.437.4628 Fax: 877.595.8108

www.DMSHealthTechnologies.com

August 13, 2009

United States Nuclear Regulatory Commission Region IV Attention: Jason M. Razo, Materials Inspector 612 East Lamar Blvd Suite 400 Arlington, TX 76011-4125

RE: NRC License number 40-32477-01

Dear Mr. Razo:

This letter is sent in response to the inspection letter dated August 4, 2009 regarding the inspection conducted on June 22 and 26, 2009 of DMS Health Technologies operating under the NRC license number 40-32477-01.

## Reply to Notice of Violation

1. During the inspection on June 26, 2009, DMS Health Technologies received a package labeled with a Radioactive Yellow II label during working hours, the package was not exempt from the monitoring requirement for radioactive contamination, and the licensee did not perform the required monitoring. Specifically, the package received by DMS Health Technologies contained unit doses of technetium-99m containing approximately 10-20 mCi each in unsealed form. This is a Severity Level IV violation under regulation 10 CFR 20.1906(b) and 10 CFR 20.1907(c).

DMS Health Technologies has contacted the technologist that was involved in this violation by Michelle White on Friday June 26, 2009 to reinforce to him/her the regulations for properly opening packages. DMS Health Technologies will and has conducted an annual Radiation Safety Competency Evaluation which is set to review with the technologist receiving packages according to 10 CFR 20.1906(b) and 10 CFR 20.1907(c). DMS Health Technologies will send out a reminder memorandum with the policy number EC-307 safely Opening Radioactive Material Packages to all Nuclear Medicine and PET technologist under the RAM license number 40-32477-01.

Full compliance of this violation will be achieved by Friday August 21, 2009.





109 South Petro Avenue Sioux Falls, SD 57107 Phone: 605.330.9060 Toll Free: 800.437.4628 Fax: 877.595.8108

www.DMSHealthTechnologies.com

Thank you for your time in making our mobile Nuclear Medicine program in compliance with the United States Nuclear Regulatory Commission's regulations.

If you have any questions or concerns regarding this response, please feel free to contact me at 1-605-357-2610.

Sincerely,

Michelle White

**Radiation Safety Officer** 

Mary Hennings-Frank

Alternate Radiation Safety Officer

Date

Date

Enclosed:

Radiation Safety Competency Evaluation

EC-307 Safely Opening Radioactive Material Packages Policy



Radiation Sa	nety Competency Evaluation	
Performance Review Score Employee Identification		Overall Score: / 3.0
Name:		
Job Title:	inges.	
Evaluation Date:	- 1000 (1000 ) (1000 ) (1000 ) (1000 ) (1000 ) (1000 ) (1000 ) (1000 ) (1000 ) (1000 ) (1000 ) (1000 ) (1000 )	
Announced Visit:		
Evaluated By:		
Regional Operations Manager:		
General Protection Procedures Competent Y - Yes; N - No; N/A - Not Applicable.		
Competency	Comments:	Competent Score Y N N/A /3.0
Technologist follows general rules for safe use and handling of RAM. (ALARA)		
Gloves being used?		000
Technologist is aware of the emergency procedures instructions for spills and where this is posted int he mobile unit.		
Spill kit is available in the work area and cab of vehicle.		
Technologist is able to describe process to clean up a radioactive spill and notification process to supervising manager and RSO.		
Technologist is wearing personal monitoring devices.		999
Syringes and vials containing radioactive material are properly labeled and shielded.		
Facilities, Materials and Equipmen	t	
Competent Y - Yes; N - No; N/A - Not Applicable.	<b>,</b>	Compotent   Seem
Competency	Comments:	Competent Score Y N N/A /3.0
Mobile unit is maintained in an orderly and clean condition.		
Radioactive material is stored properly.	,	
Adequate control to prevent unauthorized individuals from entering restricted areas.		
Adequate controls to prevent access to sources of radiation that is stored in mobile unit.		
Postings		
Competent Y - Yes; N - No; N/A - Not Applicable.		
Competency	Comments:	Competent Score Y N N/A /3.0
Radioactive material license posted.		000
Notice to employees posted.		
Technologist is knowledgeable of the contents of the radioactive material license and notice to employees.		
Technologist permit to practice posted.		
Clearance time and safety procedures for radioactive gas are posted.		
CLIA certificate of waiver posted	1	$\sim 10 \times 10^{-1}$

Use and storage areas are posted "caution radioactive material" and/or "caution radiation area".	j		- 1
Procedure for safety opening packages is posted.		000	$\dashv$
Film badge report is posted.		<u>ŏŏŏ</u>	
Receipt of radioactive material			
Competent Y - Yes; N - No; N/A - Not Applicable.			
Competency	Comments:	Competent Scor	
Incoming packages monitored for radioactive		000	$\neg$
contamination and records maintained.  Incoming packages monitored for external radiation levels			$\dashv$
and records maintained.			
Radioactive shipment receipt stickers used.  Radioactive material is received only in approved			
packages.			
Packages containing radioactive material are labeled properly.			
Waste Disposal			
Competent Y - Yes; N - No; N/A - Not Applicable.		Commotont   Som	
Competency	Comments:	Competent Scot	
Waste disposal policy is followed.		000	
Records of transfer and disposal of radioactive material is being kept, is up to date and is available for review.		000	
No packages above 0.5 mR/hR are being shipped as a		000	$\neg$
limited quantity.			
Transportation			
Competent Y - Yes; N - No; N/A - Not Applicable.		Competent Sco	re
Competency	Comments:	Y N N/A /3.	.0
Shipping papers in cab during transit. Pharmacy and sealed source shipping documents.			
All shipping papers contain emergency response		000	
information. Survey meter in cab of vehicle during transit.			
Ammo cans are blocked and braced properly during			
transit.			
Area Surveys			
Competent Y - Yes; N - No; N/A - Not Applicable.		Competent Sco	ore
Competency	Comments:	Y N N/A /3.	.0
End of day surveys are performed, documented and available for review on site.			
Contamination surveys are conducted in areas where radiopharmaceuticals are routinely prepared for use,		000	
administered and where stored.			
Technologists is familar with action levels.			
Dose rate survey records in millirems per hour.			
Contamination survey records maintained in dpm.	<u> </u>	<u> </u>	
Log Sheets/Dose accountability			
Competent Y - Yes; N - No; N/A - Not Applicable.	T	Competent Sco	ore
Competency	Comments:		3.0
Patent log sheets are completed properly, up to date and maintained on site.			
Patent log sheets are faxed/sent to appropriate destination.		000	
Verification of all doses made prior to injection.		000	
Technologist is familiar with the reporting requirements for		000	
misadministrations.			
Camera quality control  Competent Y - Yes; N - No; N/A - Not Applicable.			
	Comments	Competent Sco	ore
Competency	Comments:	Y N N/A /3	3.0

Camera quality control performed each day prior to injecting patients.		ddd
Camera quality control records being stored.		000
Camera quality control within acceptable limits.		
Well counter		
Competent Y - Yes; N - No; N/A - Not Applicable.		
Competency	Comments:	Competent Score Y N N/A /3.0
Well counter quality control performed each day prior to		000
use. Well counter quality control records available to review on		000
site.  Most recent chi-square and efficiency records stored on		000
mobile unit. Survey Meter		
Competent Y - Yes; N - No; N/A - Not Applicable.		
	Community	Competent Score
Competency	Comments:	Y N N/A /3.0
Appropriate survey instrument is being used.		
Survey meter constancy checked prior to use.		
Survey meter battery checked and is good.	and the second s	
Survey meter daily check records available on site, are up to date and accurate.		
Dose Calibrator		
Competent Y - Yes; N - No; N/A - Not Applicable.		
Competency	Comments:	Competent Score Y N N/A /3.0
Dose calibrator daily quality control performed each day prior to injecting patients.		000
Dose calibrator quarterly quality control records maintained on site and are up to date.		000
Instructions to workers/training		
Competent Y - Yes; N - No; N/A - Not Applicable.		
Competency	Comments:	Competent Score Y N N/A /3.0
Technologist radiation safety and hazardous materials		
training are up to date.		000
Technologist is aware of who the radiation safety officer is and knows how to contact them.		
Radioactive shipment receipt stickers used.	The second secon	000
Glucose Testing PET and PETCT		
Competent Y - Yes; N - No; N/A - Not Applicable.		
Competency	Comments:	Competent Score
•	Comments.	Y N N/A /3.0
Able to perform monitor daily quality control.		
Able to perform cleaning procedure on monitor.		<del></del>
Able to properly test patient glucose level.		$\underline{}$
General Safety Competent Y - Yes; N - No; N/A - Not Applicable.		
		Competent Score
Competency	Comments:	Y N N/A /3.0
There is a process for comparing the patient's current medications with those ordered for the patient while under the care of DMS Imaging. This is documented on the		lddd
patient verification form and completed prior to injection.		
Initial pain assessments are performed as defined by the DMS Imaging policies. This is documented on patient		000
verification form.  Written and verbal post procedure instructions are given		<del></del>
to the patient. The information is consistent with DMS Imaging post procedure policy.		000
It is DMS Imaging policy to maintain a latex free		
environment of care. Is there compliance with this policy? Is the Mobile Unit Life Safety (MULS) checklist posted and		
is the Mobile Offit Life datety (MOLO) checklist posted and		K ) K ) K ) I

	Overall Comments	and the second s
Comments		
Stairs and both railings are properly attached to the mobile coach?		
Is the fire extinguisher inspection up to date?		<u> </u>

	Overall Comments	
Employee		
Evaluator		



Policy:	Safely Opening Radioactive Material Packages
Policy Number:	EC-307
Employees Affected:	Nuclear Medicine and PET Technologists
Approved By:	Radiation Safety Officer, Radiation Safety Committee
Effective Date:	12/31/05
Reviewed Date:	10/29/08
Revised Date:	02/03/09

### Purpose:

The purpose of this policy is to provide information and procedural instructions for safely opening packages containing radioactive material.

# **Policy and Procedure:**

Appropriate safety precautions will be strictly followed at all times. All radioactive shipments will be assumed to be potentially contaminated until proven otherwise. Radioactive shipments will be monitored for external radiation levels and surface contamination within three hours after receipt if received during working hours or not later than three hours from the beginning of the next working day if received after hours. All packages will be opened using the following procedure:

- 1. Put on gloves to prevent hand contamination.
- 2. Visually inspect the package for any sign of damage (e.g. wetness, crushed). If damage is noted, stop procedure and notify the Radiation Safety Officer.
- 3. Measure exposure at one meter from the package surface and record. If the reading is greater than 10 mR/hr, stop the procedure and notify the Radiation Safety Officer.
- 4. Measure the surface exposure rate and record. If the measurement is greater than 200 mR/hr, stop the procedure and notify the Radiation Safety Officer.
- 5. Do a wipe test of the outer surface of the shipping container. Record the results on the sticker. If the outer surface wipe test shows contamination, or there is any damage to the package, a wipe of the external surface of the final source container should also be done. Acceptable limits are less than 2000 dpm/100 cm<sup>2</sup>.
- 6. Open the package with the following precautionary steps:
  - a. Open the outer package (following manufacturer's directions, if supplied) and remove packing slip.
  - b. Open the inner package and verify that the contents within agree with those on the packing slip, compare requisitions, packing slip and vial and/or syringe labels.
  - c. Check the integrity of the final source container (i.e., inspect for breakage of seals or vials, loss of liquid and/or discoloration of packaging materials).
- 7. Check that the shipment does not exceed possession limits and record all appropriate information log book or utilization form.
- 8. Monitor the packing materials and the package for contamination before discarding.

- a. If contaminated, obliterate radiation warning labels before discarding in the regular trash.
- 9. Records of the receipt of Radioactive Materials will be maintained and kept.

#### **Communications:**

Questions, concerns, and recommendations regarding this policy and procedure are welcome and should be directed to the Radiation Safety Committee and/or RSO of DMS Imaging.

# **Review Frequency:**

Every three years or if new changes are needed.