

**U.S. NUCLEAR REGULATORY COMMISSION**

Date: 8-28-09

**TELEPHONE CONVERSATION RECORD**

Time: 10:15am

Mail Control 143800 License No(s). 29-30528-01 Docket No(s). 030-35176  
 or Report No(s).

Name of Licensee: Cardiology Center of North Jersey

Name of Participant(s): Mike Lairmore, Medical Physics Consultation

Telephone No. (201) 693-2277

Subject: Request for Clarification Regarding Modified License Renewal Information Submitted August 27, 2009

(NOTE: This will be used as the Documents Title in ADAMS)

I called Mike Lairmore to clarify issues on the modified license renewal information that was submitted August 27, 2009. For instrument calibrations, Cardiology Center of North Jersey will not be calibrating any survey meters in house. Meters will be sent to MPM Products (Texas Radioactive Materials License Number LOO967). The licensee has a copy of MPM's license and license application (which includes written survey meter calibration procedures). I also clarified that there is no office space in the basement. It is used for storage only.

Action Required: File

**Document Availability:**  Publicly Available  Non-Publicly Available  
 Non-Sensitive  Non-Sensitive Copyright  Sensitive  Sensitive Copyright  
 Immediate Release  Normal Release  Delay Release Date  
 Prepared & SUNSI Review Completed By: Janice Nguyen / RA / Date: 8/28/09