

CME

ENGINEERING

LI

814-443-3344
Fax: 814-444-0365

September 10, 2009

U.S Nuclear Regulatory Commission – Region I
Attention L.A.T.
475 Allendale Road
King of Prussia, PA 19406-1415

RE: Supplementary Information needed for Renewal of
License No. 37-30535-01 Portable Nuclear Device
Reference Document No. 03035224

Dear Sir or Madam:

This letter contains the supplement statements that were inadvertently omitted from our original license renewal application.

CME certified gauge users shall provide routine maintenance and cleaning to our portable nuclear gauges as per manufactures recommendations and only perform service procedures authorized to gauge users by our current license. Services above this level; calibrations, electronics and source access shall be handled by Pennsylvania State and NRC authorized and licensed service providers.

We implement and maintain operating and emergency procedures in accordance with Appendix H of the NUREG-1556 Vol. 1, Rev 1 dated November 2001. All of our certified gauge users have access to a current copy of the procedures and a hard copy is kept inside the gauge transport case and a electronic copy is available on the company server.

If you have any questions or need additional information please call me at 814-443-3344 x 3051 or email dave.blasko@cmemgmt.com.

Thank you.



Dave Blasko
Technical Supervisor / RSO
CME Management LLC,
general partner of CME Engineering LP
cc: Sue Toth / file