CONVERSATION RECORD | TIME | DATE 8-24-2009

O VISIT O CONFERENCE X TELEPHONE

O INCOMING
X OUTGOING

NAME OF PERSON(S) CONTACTED OR IN CONTACT	ORGANIZATION (OFFICE, DEPT.ETC.)	TELEPHONE NO.
Ryan Parker, RSO Vice President	TesTech, Inc. 8534 Yankee Street Dayton, OH 45458	937-435-3200 FAX 937-291-6549

SUBJECT
Control No. 318008- Re - Amendment request for renewal of License No. 34-26553-02, by application dated March 10, 2009, regarding moisture/density gauges.

SUMMARY
In a telephone conversation with Mr. Parker on this date, the following additional items were discussed:
1. It was noted that in the license renewal application dated March 10, 2009, the only requested licensed sources and gauges (as listed in the enclosed license issued to them by the State of Ohio) did not include any CF-252 sealed sources for use in either Troxler or Campbell Pacific Nuclear (CPN) gauges nor any CPN model gauges (for use with Cs-137, Am-241 or Cf-252 sources) as currently authorized on TesTech, Inc. NRC License No. 34-26553-02. The Cf-252 source/devices and the CPN devices need to be addressed in the licensee's response or justify their deletion from the license (i.e. either never possessed or used under License No. 34-26553-02 or when and to whom transferred, a copy of recipients license, a copy of a document from them acknowledging their receipt of the make, model and serial number of the device(s) and a copy of a leak test record, current as of the transfer showing the non-leaking status. Also, provide a statement confirming you never possessed any leaking Cf-252 sealed sources, based on routine leak test data, if that be factual.

In accordance with 10 CFR 2.390 of the NRC's "Rules of Practice," a copy of this conversation record will be available electronically for public inspection in the NRC Public Document Room or from the Publicly Available Records (PARS) component of NRC's document system (ADAMS). The NRC's document system is accessible from the NRC Web site at http://www.nrc.gov/reading-rm/adams.html (the Public Electronic Reading Room).

ACTION REQUIRED
Please provide written response for the above requested information as additional information to Control No. 318008. Please send **response to me by facsimile (630-515-1078)** and **please inform me by telephone (630-829-9829) as soon as response has been sent**. A voice mail confirmation is fine if I am not present. **No follow-up hard copy is needed.**

NAME OF PERSON DOCUMENTING CONVERSATION	SIGNATURE	DATE
Loren J. Hueter	Loren J. Hueter	8-24-2009

ACTION TAKEN

SIGNATURE	TITLE	DATE