



September 22, 2009

Docket No. 030-37336
Control No. 144169

License No. 06-31186-01

Arye Hess
Executive Director, Quality Systems
FUJIFILM Medical Systems USA, Inc.
419 West Avenue
Stamford, CT 06902

SUBJECT: FUJIFILM MEDICAL SYSTEMS USA, INC., LICENSE TERMINATION,
CONTROL NO. 144169

Dear Mr. Hess:

Please find enclosed Amendment No. 1 terminating License No. 06-31186-01 as requested by your NRC Form 314 dated September 16, 2009. This termination is being issued in accordance with the requirements of the applicable NRC License Termination Rule 10 CFR 30.36. Based on our inspection records we understand you possess C-14 check sources in exempt quantities. You are allowed to keep the checked sources as long as you do not exceed the 100 microcurie limit per source. All facilities previously used for licensed activities may be released for unrestricted use.

Current NRC regulations and guidance are included on the NRC's website at www.nrc.gov; select **Nuclear Materials; Medical, Academic, and Industrial Uses of Nuclear Material**; then **Regulations, Guidance, and Communications**. You may also obtain these documents by contacting the Government Printing Office (GPO) toll-free at 1-866-512-1800. The GPO is open from 7:00 a.m. to 6:30 p.m. EST, Monday through Friday (except Federal holidays).

Your cooperation with us is appreciated.

Sincerely,

Original signed by Elizabeth Ullrich

Betsy Ullrich
Senior Health Physicist
Commercial and R&D Branch
Division of Nuclear Materials Safety

Enclosure:
Amendment No. 1

cc:
John Pizzonia, Ph.D., Radiation Safety Officer

DOCUMENT NAME: G:\WordDocs\Current\Lic Cvr Letter\L06-31186-01.144169.doc

SUNSI Review Complete: LROLDAN

After declaring this document "An Official Agency Record" it will be released to the Public.

To receive a copy of this document, indicate in the box: "C" = Copy w/o attach/encl "E" = Copy w/ attach/encl "N" = No copy

OFFICE	DNMS/RI	N	DNMS/RI	N	DNMS/RI		
NAME	L Roldán/LR		E Ullrich/EU				
DATE	9/22/09		9/22/09				

OFFICIAL RECORD COPY