

**U.S. NUCLEAR REGULATORY COMMISSION
TELEPHONE CONVERSATION RECORD**

Date: 10/01/09
Time:

Mail Control or Report No(s). 144102 **License No(s).** 06-00843-03 **Docket No(s).** 03001245

Name of Licensee: St. Vincent's Medical Center
Name of Participant(s): Jerry Randall, RSO
Telephone No. 860-559-9366 (cell); e-mail radrem@aol.com
Subject: Additional information for amendment request
(NOTE: This will be used as the Documents Title in ADAMS)

This record represents communications with Jerry on 10/01/09. I asked the following:
For the old Cs-sources, provide documentation of (a) receipt by an authorized recipient, (b) documentation of the recipient's authorization to receive the sources, and (c) leak test results confirming no leakage at time of transfer. The request said that all but one tube source was transferred and a CDCT1 source was not. Should the license continue to list the CDCT1 source as a brachytherapy source under 35.400 or should a line item for be added for Cs-137 for calibrations? If the latter, please provide the possession limit and indicate who should be listed as non-medical authorized user for calibration.
For the old HDR unit, provide documentation of the transfer, including receipt of the final source by an authorized recipient.
This request adds NARM brachytherapy sources (Pd-103). Are there any other NARM uses or facilities that may need to be added, for example a PET facility?

Action Required: Await response from licensee

Document Availability: **Publicly Available** **Non-Publicly Available**
 Non-Sensitive **Non-Sensitive Copyright** **Sensitive- Proprietary**
 Sensitive – Privacy Act (includes PII) **Sensitive – Internal** **Sensitive – Security-Related**
 Immediate Release **Normal Release Date: 1 week** **Delayed Release Date:**
SUNSI Review Completed By: S. Gabriel **/ RA /** **Document Accession #:** ML092750607