

Br. 1

Luis Rosado, MD
Clinica Cardiovascular de Guaynabo
P.O. Box 965
Guaynabo, PR 00970-0965

October 7, 2009

Regional Administrator, Region II
US NRC, Region I
475 Allendale Road
King of Prussia, PA 19406

RLJ

RE: Change of address, NRC license # 52-31182-01 03037326

Dear Mr. Sir or Madam:

Please amend our license as follows:

Please remove the old location from our license. Calle Carazo #42, Guaynabo Puerto Rico. The close out survey is attached.

If you need any further information, please contact me at (939) 389-2222 or David Rhoe at (787) 245-7248.


Sincerely,


Luis Rosado, MD


REC'D IN LAT _____

144224
NMSS/R6NI MATERIALS-002


Close out Survey
 October 6, 2009
 Calle Carazo #42, Guaynabo, PR



Hot Lab



Camera Area



Treadmill Area

Close out survey
 October 6, 2009
 Calle Carazo #42, Guaynabo, PR

Instruments used

Ludlum Model 14C Sn235068
 Calibrated on 11-7-08
 Background = 0.02 mR/hr
 Check source reading before = 1.8 mR/hr
 Check source reading after = 1.8 mR/hr

Victoreen DWTC 05-578 Sn4742814
 Calibration check on 2-4-09
 Background = 9 cpm
 Efficiency = 9.75
 Trigger = 2200 dpm x 0.0975 = 214 cpm

	Location	cpm	mR/hr
1	Sink - inside	13	0.02
2	Sink - handle	16	0.02
3	Door handle	18	0.02
4	Cabinet inside	10	0.02
5	Floor	13	0.02
6	Floor	15	0.02
7	Floor	13	0.02
8	Floor	7	0.02
9	Light switch	14	0.02
10	Cabinet handle	15	0.02
11	Floor	21	0.02
12	Floor	14	0.02
13	Floor	14	0.02
14	Floor	16	0.02
15	Floor	10	0.02
16	Floor	16	0.02
17	Floor	19	0.02
18	Floor	10	0.02
19	Floor	9	0.02
20	Floor	15	0.02
21	Floor	10	0.02
22	Floor	10	0.02
23	Floor	9	0.02
24	Floor	10	0.02
25	Step	15	0.02
26	Door handle	10	0.02
27	Floor	9	0.02
28	Floor	19	0.02
29	Floor	19	0.02
30	Floor	19	0.02
31	Floor	16	0.02
32	Floor	10	0.02
33	Floor	16	0.02
34	Floor	15	0.02
35	Floor	10	0.02
36	Door handle	9	0.02
37	Light switch	10	0.02
38	Background	9	0.02

This is to acknowledge the receipt of your letter/application dated

10/7/09, and to inform you that the initial processing which includes an administrative review has been performed.

Amendment (52-31182-01) There were no administrative omissions. Your application was assigned to a technical reviewer. Please note that the technical review may identify additional omissions or require additional information.

Please provide to this office within 30 days of your receipt of this card

A copy of your action has been forwarded to our License Fee & Accounts Receivable Branch, who will contact you separately if there is a fee issue involved.

Your action has been assigned **Mail Control Number** 144221.
When calling to inquire about this action, please refer to this control number.
You may call us on (610) 337-5398, or 337-5260.