

The Hospital of Central Connecticut

23 September 2009

USRNC Region I
475 Allendale Rd
King of Prussia PA 19406

Greetings:

Please amend our byproduct materials license number 06-02388-01 as follows:

We wish to change an area of radioactive materials use. We will be moving a nuclear medicine camera from the Women's Center of Southington, Ct. 55 Meriden Avenue, Suite 1D, Southington, Ct. to the Women's Center of Southington, 55 Meriden Avenue, Suite 2E, Southington, Ct.


This facility is located in a building which is wholly owned by Bradley Medical Arts LLC and the suite where the camera is located is leased to THOCC. This building is directly connected to THOCC at 81 Meriden Avenue, Southington CT. The room will be used for imaging only and no administration of Radiopharmaceuticals will occur in the room. However, a Co57 sheet source will be brought to the room for the sole purpose of quality control imaging. Also, there will be occasion to bring a small amount of Tc04- for recalibration after repair as needed. Neither of these two isotopes will be stored at 55 Meriden Ave, Southington CT.

We agree to follow all pertinent regulations and license conditions at this new location, including area monitoring, as applicable. We will ensure that public dose limits in adjacent areas will not be exceeded.

A drawing of the proposed area of use is attached.

Please provide us with EXPEDITED consideration for this license amendment request, as time is short before we will need to utilize this area for patient care.

Sincerely yours,


Clarence Silva

CS/ejl

J-9

2009 OCT -9 PM 12: 27

RECEIVED
REGION 1

03001250

144231
NRC/RCNI MATERIALS-002

Doctors Office

5 ft

Bathroom

4 ft

3 ft

Nurse med
Room

Technologist
Work Area

Changing
Room

7 ft

7 ft

Hallway

Changing
Room

8 ft

Hallway

