

1. Please specify the suite number of where the portable gauges will be stored.
2. Please specify the number of gauges you will have in your possession (i.e., total number of Seaman Nuclear Corporation Models C-200 and C-300 Series portable gauges).
3. For non-routine maintenance or repair operations that require detaching the source or source rod from the gauge, please submit either of the following:
 - A statement that: “We will send the gauge to the manufacturer or other person authorized by NRC or an Agreement State to perform non-routine maintenance or repair operations that require detaching the source or source rod from the gauge.”

OR

- The information listed in NUREG-1556, Vol. 1, Rev. 1, Appendix G supporting a request to perform this work in-house.