

TRANSMISSION VERIFICATION REPORT

TIME : 11/02/2009 20:02
NAME : USNRC RIII
FAX : 6308299782
TEL :
SER. # : 000A7J925774

DATE, TIME 11/02 20:01
FAX NO./NAME 83173517733
DURATION 00:00:54
PAGE(S) 03
RESULT OK
MODE STANDARD
ECM

NRC FORM 386 (RIII)



UNITED STATES
NUCLEAR REGULATORY COMMISSION
REGION III
2443 Warrenville Road, Suite 210
Lisle, Illinois 60532-4352

TELEFAX TRANSMITTAL

DATE: 11/2/09 NUMBER OF PAGES: _____
(including this page)

SEND TO: ANDREA BROWNE, PH.D.

LOCATION: Community Hospitals of Indiana, Inc.

FAX NUMBER: 317-351-7733 **VERIFY BY CALLING SENDER**

FROM: Colleen Carol Casey
(SENDER)

TELEPHONE NUMBER: 630-829-9241 FAX NUMBER: 630-515-1078

If you do not receive the complete fax transmittal, please contact the sender as soon as possible at the telephone number provided above.

MESSAGE

**COLLEEN CAROL CASEY
MATERIALS LICENSING BRANCH
UNITED STATES NUCLEAR REGULATORY COMMISSION**

REGION III
2443 WARRENVILLE ROAD STE 210
LISLE, ILLINOIS 60532-4352
OFFICE: (630)-829-9841 FAX: (630) 515-1078

CONVERSATION RECORD

ACTUALLY FAXED?

(Y) (N)

TIME 8:00 pm CT on DATE

Left VM msg for Ms Browne

Nov. 2, 2009

NAME OF PERSON(S) CONTACTED

ORGANIZATION

TELEPHONE NO.

Andrea Browne, Ph.D., RSO for Community Hospitals of Indiana, Inc. 317-355-5865

Fax: 317-351-7733

SUBJECT

License No.: 13-06009-01

Control No.: **318389**

SUMMARY

We have reviewed your letter dated July 29, 2009, requesting an amendment to your byproduct materials license and find that we need additional information as follows:

Your letter requests the addition of Shawn Hollars, M.A. as an AMP for the use of Iridium-192 in an HDR remote afterloading brachytherapy unit.

However, in reviewing Mr. Hollars' forms 313a (AMP) I noted that he is claiming to have completed one fulltime year of training in medical physics, as required by 10 CFR 35.51(b)(1) during the last year of study prior to receiving his master's degree in physics.

Please note that none of the training described in the alternate pathway can be claimed concurrently.

Please resubmit Mr. Hollars' application when he can demonstrate that he completed one fulltime year of training in medical physics, as required by 10 CFR 35.51(b)(1), independent of the other training described in this section.

Please also be reminded of the provisions in 10 CFR 30.9(a), "Completeness and accuracy of information,"..."(a) Information provided to the Commission by an applicant for a license or by a licensee or information required by statute or by the Commission's regulations, orders, or license conditions to be maintained by the applicant or the licensee shall be complete and accurate in all material respects."


In accordance with 10 CFR 2.390 of the NRC's "Rules of Practice," a copy of this record will be available electronically in the NRC Public Document Room or from the Publicly Available Records (PARS) component of NRC's document system (ADAMS). The NRC's document system is accessible from the NRC Web site at <http://www.nrc.gov/reading-rm/adams.html>.

ACTION REQUIRED

2. As we cannot issue an amendment at this time we are voiding this request in order to enable you to prepare a quality application without time constraints. This is done without prejudice to the resubmission of your request at a later date. Upon receipt of your response we will resume our review. Address your written response to my attention at the above address.

PLEASE NOTE THAT A "VOID" IS AN ADMINISTRATIVE PROCEDURE THAT PUTS YOUR AMENDMENT REQUEST "ON HOLD" (TAKES IT OUT OF OUR ACTIVE CASEWORK DATABASE) UNTIL YOU REACTIVATE IT VIA SUBMISSION OF A WRITTEN RESPONSE. IT "BUYS" YOU TIME TO PREPARE A QUALITY RESPONSE AND IS GENERALLY REGARDED AS A "GOOD THING."

PLEASE DIRECT ANY QUESTIONS YOU MAY HAVE TO ME AT (630) 829-9841 or (800) 522-3025.

| NAME OF PERSON DOCUMENTING CONVERSATION | SIGNATURE | DATE |
|---|---|------------------|
| Colleen Carol Casey |  | November 2, 2009 |
