

**SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION**

1. LICENSEE/LOCATION INSPECTED: <b>Mid-America Calibrations 808 SW Nautica Court Lee's Summit, MO 64082-2329</b>  REPORT NUMBER(S)		2. NRC/REGIONAL OFFICE <b>U.S. Nuclear Regulatory Commission Region III 2443 Warrenville Road Suite 210 Lisle, Illinois 60532-4351</b>	
3. DOCKET NUMBER(S) <b>030-36634</b>	4. LICENSEE NUMBER(S) <b>24-32531-01</b>	5. DATE(S) OF INSPECTION <b>October 29, 2009</b>	

**LICENSEE:**

The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The inspection findings are as follows:

- 1. Based on the inspection findings, no violations were identified.
- 2. Previous violation(s) closed.
- 3. The violation(s), specifically described to you by the inspector as non-cited violations, are not being cited because they were self-identified, non-repetitive, and corrective action was or is being taken, and the remaining criteria in the NRC Enforcement Policy, NUREG-1600, to exercise discretion, were satisfied.

\_\_\_\_\_ Non-Cited Violation(s) was/were discussed involving the following requirement(s) and Corrective Action(s):

- 4. During this inspection certain of your activities, as described below and/or attached, were in violation of NRC requirements and are being cited. This form is a NOTICE OF VIOLATION, which may be subject to posting in accordance with 10 CFR 19.11.  
(Violations and Corrective Actions)

**Licensee's Statement of Corrective Actions for Item 4, above.**

I hereby state that, within 30 days, the actions described by me to the inspector will be taken to correct the violations identified. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201 (corrective steps already taken, corrective steps which will be taken, date when full compliance will be achieved). I understand that no further written response to NRC will be required, unless specifically requested.

Title	Printed Name	Signature	Date
LICENSEE'S REPRESENTATIVE			
NRC INSPECTOR	<b>Robert P. Hays</b>		<b>10/29/2009</b>

**SAFETY INSPECTION REPORT  
AND COMPLIANCE INSPECTION**

1. LICENSEE <b>Mid-America Calibrations</b> REPORT NUMBER(S) <b>2009-001</b>		2. NRC/REGIONAL OFFICE <b>Region III</b> <b>2443 Warrenville Road, Suite 210</b> <b>Lisle, IL 60532</b>	
3. DOCKET NUMBER(S) <b>03036634</b>	4. LICENSE NUMBER(S) <b>24-32531-01</b>	5. DATE(S) OF INSPECTION <b>October 29, 2009</b>	
6. INSPECTION PROCEDURES USED <b>87126 (09/28/05)</b>	7. INSPECTION FOCUS AREAS <b>03.01-03.07</b>		

**SUPPLEMENTAL INSPECTION INFORMATION**

1. PROGRAM CODE(S) <b>03221</b>	2. PRIORITY <b>5</b>	3. LICENSEE CONTACT <b>Karl Artinger, Owner, RSO</b>	4. TELEPHONE NUMBER <b>1-800-488-1391</b>
------------------------------------	-------------------------	---------------------------------------------------------	----------------------------------------------

<input checked="" type="checkbox"/> Main Office Inspection	Next Inspection Date: <b>October 2014</b>
<input type="checkbox"/> Field Office	
<input type="checkbox"/> Temporary Job Site Inspection	

**PROGRAM SCOPE**

The licensee operates a survey instrument calibration and leak test service program staffed with one authorized user who is also the owner and Radiation Safety Officer (RSO). Licensed material was possessed and used in conformance with license requirements.

**Performance Observations**

During the inspection, the licensee's RSO demonstrated/discussed: (1) ANSI calibration procedures; (2) public dose management; (3) leak tests and procedures; (4) receiving and return procedures for survey instruments (5) inventory; (6) dosimetry; (7) recordkeeping; and (8) security of licensed material.