

Cost Center:DIV20Mail Stop:B51Customer:DON BANNONManufacturer/Model:VAISALA / HMP77BDescription:HUMIDITY/TEMPERATURE PROBESerial Number:20750006Asset Number:010700Supplier or Calibration Procedure:TEMPERATURE, HUMIDITY, DEW-POINT - 3 APR 08Work Order:303089335

The above instrument was received for service on July 29, 2009 and found not to meet calibration procedure specifications, or as specified by the cost center. As-found readings are provided in the attached measurement report, or as noted in remarks below, for your review to determine if the instrument is out of tolerance for project requirements and processing in accordance with your cost center quality policy.

Please call extension 5215 if you have any questions or need additional information.

Remarks: Failed humidity tests.



# **Explanation of Measurement Report Results**

"When statements of compliance (Pass/Fail) are made, the uncertainty of measurement shall be taken into account". Reference ISO/IEC 17025:2005, 5.10.4.2

This explanation is provided to you because the instrument submitted for calibration has one or more of the following results.

# <u>Result</u>

Pass – measured value or test is within the  $\pm$  limit, in tolerance, with a confidence level of 95 percent.

- Pass? measured value is *within* the  $\pm$  limit, but by a margin less than half of the uncertainty interval and has a confidence level of less than 95 percent of being in tolerance. Adjustment is made and the measurement is repeated. If adjustment or repair is not possible or fails to improve the results, then the customer must determine in or out of tolerance.
  - Fail? measured value is *outside* the  $\pm$  limit, but by a margin less than half of the uncertainty interval and is reported as out of tolerance but it is not possible to state this with a 95 percent confidence level. Adjustment is made and the measurement is repeated. If adjustment or repair is not possible or fails to improve the results, then the customer must determine if out of tolerance action is necessary.
- Fail measured value is *outside* the  $\pm$  limits with a 95 percent confidence level. Adjustment is made and the measurement is repeated for As-left data. If adjustment or repair is not possible or fails to improve the results, then the customer must determine if the measured value is in compliance for the intended use.

# %Limit

Adjustment is made, if possible, when the As-found measured value is equal to or greater than 70 percent of the  $\pm$  limit. If adjustment is not possible or did not lower the As-left reading below 70 percent, the customer shall determine if the instrument is suitable for their requirements.

# Type Data

Found-left All measurements were in tolerance and no adjustments or repairs were preformed.

- As-found One or more measurements were other than Pass or exceeded 70 percent of the  $\pm$  limit and adjustment or repairs were performed.
- As-left Results of measurements after adjustment or repair.

### **Uncertainty**

Best estimate of the dispersion of the measured value that could be contributed by the; standard, environment, repeatability of the measurement process, characterizes of the instrument being calibrated (i.e. resolution) etc.

Please call extension 5215 for questions or additional information.

#### Southwest Research Institute Calibration Laboratory Measurement Report

• •

.

| Work Order:            | 303089335         | Mfr.:           | Vaisala            |                | Technician:       | Mark Romero |          |
|------------------------|-------------------|-----------------|--------------------|----------------|-------------------|-------------|----------|
| Asset No.:             | 010700            | Model:          | HMP77B             |                | Type Data:        | As-found    |          |
| Serial No.:            | Z0750006          | Туре:           | Humidity/Temp      | Probe          | Cal Date:         | 31-Jul-(    | )9       |
| Remarks: Calibrated as | s a system with V | aisala MI70. Te | mperature limit in | creased to +/- | 1 °C per custodia | an.         |          |
|                        |                   |                 |                    |                |                   |             |          |
| Function/Range         | Test Point        | TI Reading      | Difference         | +/- Limit      | +/- Uncertainty   | Result      | % Limit  |
| Temperature            | °C                | °C              | °C                 | °C             | °C                |             |          |
| at 20% RH              | 25.03             | 25.00           | -0.03              | 1              | 0.12              | Pass        | 3%       |
| 40% RH                 | 25.04             | 25.02           | -0.02              |                |                   | Pass        | 2%       |
| 60% RH                 | 25.09             | 25.01           | -0.08              |                |                   | Pass        | 8%       |
| 80% RH                 | 25.13             | 25.06           | -0.07              |                |                   | Pass        | 7%       |
|                        | °C                | °C              | °C                 |                |                   |             |          |
| at 20% RH              | 49.16             | 49.06           | -0.10              |                |                   | Pass        | 10%      |
| 40% RH                 | 49.21             | 49.12           | -0.09              |                |                   | Pass        | 9%       |
| 60% RH                 | 49.28             | 49.16           | -0.12              |                |                   | Pass        | 12%      |
| 80% RH                 | 49.34             | 49.23           | -0.11              |                |                   | Pass        | 11%      |
|                        | °C                | °C              | °C                 |                |                   |             |          |
| at 20% RH              | 68.11             | 67.48           | -0.63              |                |                   | Pass        | 63%      |
| 40% RH                 | 68,52             | 68.19           | -0.33              |                |                   | Pass        | 33%      |
| 60% RH                 | 68.79             | 68.49           | -0.30              |                |                   | Pass        | 30%      |
| 80% RH                 | 69.09             | 68.70           | -0.39              |                |                   | Pass        | 39%      |
| Humidity               | % RH              | % RH            | % RH               | % RH           | % RH              |             |          |
| at 25 °C               | 20.00             | 22.12           | 2.12               | 2              | 0.68              | Fail?       | 106%     |
|                        | 40.05             | 43.80           | 3.75               | -              |                   | Fail        | 188%     |
|                        | 59.97             | 63.88           | 3.91               |                |                   | Fail        | 196%     |
|                        | 79.95             | 83.70           | 3.75               |                |                   | Fail        | 188%     |
| at 50 °C               | 20.00             | 21.28           | 1 28               |                |                   | Pass        | 64%      |
|                        | 39 99             | 42.96           | 2.97               |                |                   | Fail        | 149%     |
|                        | 60.00             | 63 42           | 3 42               |                |                   | Fail        | 171%     |
|                        | 79.99             | 82.83           | 2.84               |                |                   | Fail        | 142%     |
| at 70 °C               | 19 99             | 20.61           | 0.62               |                |                   | Pass        | 31%      |
|                        | 39.97             | 41 76           | 1 79               |                |                   | Pass?       | 90%      |
|                        | 60.05             | 62.06           | 2 01               |                |                   | Fail?       | 101%     |
|                        | 79 99             | 81.63           | 1 64               |                |                   | Pass        | 82%      |
|                        | . 0.00            | END             | OF REPORT          |                |                   |             | <u> </u> |



SOUTHWEST RESEARCH INSTITUTE<sup>®</sup>

6220 Culebra Road, P.O. Drawer 28510 Institute Quality Systems Institute Calibration Laboratory Phone: 210-522-5215 Fax 210-522-4834

# **Certificate of Calibration**



| Cost Center: DIV20  | Work Order: 303089335                        |
|---|--|
| Mail Stop: B51  | Date Issued: 20-Aug-2009                     |
| Customer: DON BANNON  | Date Calibrated: 20-Aug-2009                 |
| Manufacturer/Model: VAISALA / HMP77B  | * Date Due : 20-Aug-2010                     |
| Description: HUMIDITY/TEMPERATURE PROBE   | ** Results: AS-LEFT                          |
| Serial Number: Z0750006   | Temperature: 75.0 °F                         |
| Asset Number: 010700  | Humidity: 46 %RH                             |
| Procedure: TEMPERATURE, HUMIDITY, DEW-POINT - 3 APR 08  | <b>Barometer:</b> N/A                        |
| This cartificate documents tracephility to the National Institute of Standards and Technology (NIST) and the Inte | motional System of Units (SI) The Laboratory |

This certificate documents traceability to the National Institute of Standards and Technology (NIST) and the International System of Units (SI). The Laboratory quality system conforms to ISO/IEC 17025, 2005, ANSI/NCSL Z540-1-1994 and relevant requirements of the ISO 9000-2000 standard. This certificate shall not be reproduced, except in full, without the written approval of the Southwest Research Institute Calibration Laboratory. This certificate shall not be used to claim product endorsement by Southwest Research Institute, American Association for Laboratory Accreditation (A2LA) or any agency of the U. S. Government. Results of this calibration relate only to the instrument described above at the time of calibration and does not imply any long term stability of the instrument.

\*Determined by the customer, does not imply the instrument will remain within tolerance as any number of factors may cause an out-of-tolerance condition before this date. \*\*Data type found in this certificate or attached measurement report must be interpreted as: Found-left - adjustment and/or repair was not performed, As-found - data is before unit is adjusted and/or repaired, As-left - data is after adjusted and/or repaired was performed. The customer has sole responsibility for determination of in-/out-of-tolerance or compliance/noncompliance.

Measurement uncertainty calculated in accordance with the method described in the ISO "Guide to the Expression of Uncertainty in Measurement" (GUM), for a confidence level of approximately 95 percent using a coverage factor of k=2.

Remarks: Calibrated as a system with Vaisala MI70 SN Z0730015. Limits: ±1 °C/ ±5% RH.

#### Standards Used

| <u>Asset #</u> | Manufacturer       | Model | Description         | Cal Date    | Due Date    |
|----------------|--------------------|-------|---------------------|-------------|-------------|
| 006404         | THUNDER SCIENTIFIC | 2500  | HUMIDITY GENERATOR  | 27-May-2009 | 27-May-2010 |
| 009414         | HART SCIENTIFIC    | 1502A | TEMPERATURE READOUT | 12-Jun-2009 | 12-Dec-2009 |
| 015895         | HART SCIENTIFIC    | 5618B | RTD                 | 12-Jun-2009 | 12-Dec-2009 |

il to Was Walt Hill

ark Romero

Metrology Technician

Laboratory Manager m:VA2LA OCT\_08.rpt

#### Southwest Research Institute Calibration Laboratory Measurement Report

. . .

٠

| Work Order:             | 303089335         | Mfr.:            | Vaisala            |                | Technician:        | Mark R          | omero   |
|-------------------------|-------------------|------------------|--------------------|----------------|--------------------|-----------------|---------|
| Asset No.:              | 010700            | Model:           | HMP77B             |                | Type Data:         | As-left         |         |
| Serial No.:             | Z0750006          | Туре:            | Humidity/Temp      | Probe          | Cal Date:          | Date: 20-Aug-09 |         |
| Remarks: Calibrated as  | s a system with V | aisala MI70. Co  | ntinued pervious   | y assigned ter | mperature limit of | ±1 °C ar        | nd a    |
| "one-time" limit of ±5% | R.H. has been ap  | oplied per custo | dian for immediate | e use.         |                    |                 |         |
| Function/Range          | Test Point        | TI Reading       | Difference         | +/- Limit      | +/- Uncertainty    | Result          | % Limit |
| Temperature             | °C                | °C               | °C                 | °C             | °C                 |                 |         |
| at 20% RH               | 25.06             | 25.05            | -0.01              | 1              | 0.12               | Pass            | 1%      |
| 40% RH                  | 25.11             | 25.08            | -0.03              |                |                    | Pass            | 3%      |
| 60% RH                  | 25.14             | 25.14            | 0.00               |                |                    | Pass            | 0%      |
| 80% RH                  | 25.16             | 25.13            | -0.03              |                |                    | Pass            | 3%      |
|                         | °C                | °C               | °C                 |                |                    |                 |         |
| at 20% RH               | 48.20             | 48.23            | 0.03               |                |                    | Pass            | 3%      |
| 40% RH                  | 48.85             | 48.84            | -0.01              |                |                    | Pass            | 1%      |
| 60% RH                  | 49.10             | 49.00            | -0.10              |                |                    | Pass            | 10%     |
| 80% RH                  |                   |                  | 0.00               |                |                    | Pass            | 0%      |
|                         | °C                | °C               | °C                 |                |                    |                 |         |
| at 20% RH               | 68.28             | 68.30            | 0.02               |                |                    | Pass            | 2%      |
| 40% RH                  | 68.55             | 68.47            | -0.08              |                |                    | Pass            | 8%      |
| 60% RH                  | 68.82             | 68.80            | -0.02              |                |                    | Pass            | 2%      |
| 80% RH                  | 69.08             | 69.11            | 0.03               |                |                    | Pass            | 3%      |
| Humidity                | % RH              | % RH             | % RH               | % RH           | % RH               |                 |         |
| at 25 °C                | 20.01             | 19.91            | -0.10              | 5              | 0.68               | Pass            | 2%      |
|                         | 40.05             | 40.02            | -0.03              |                |                    | Pass            | 1%      |
|                         | 60.00             | 60.07            | 0.07               |                |                    | Pass            | 1%      |
|                         | 80.00             | 76.59            | -3.41              |                |                    | Pass            | 68%     |
| at 50 °C                | 19.98             | 20.10            | 0.12               |                |                    | Pass            | 2%      |
|                         | 39.99             | 39.98            | -0.01              |                |                    | Pass            | 0%      |
|                         | 60.00             | 58.90            | -1.10              |                |                    | Pass            | 22%     |
|                         | 80.00             | 76.58            | -3.42              |                |                    | Pass            | 68%     |
| at 70 °C                | 20.01             | 19.70            | -0.31              |                |                    | Pass            | 6%      |
|                         | 39.98             | 39.78            | -0.20              |                |                    | Pass            | 4%      |
|                         | 60.00             | 58.55            | -1.45              |                |                    | Pass            | 29%     |
|                         | 80.00             | 75.70            | -4.30              |                |                    | Pass            | 86%     |
|                         |                   | END              | OF REPORT          |                |                    |                 |         |