



SOUTHWEST RESEARCH INSTITUTE®

6220 Culebra Road, P.O. Drawer 28510
Institute Quality Systems
Institute Calibration Laboratory
Phone: 210-522-5215 Fax 210-522-4834



Calibration Laboratory
Certificate #0972-01

Certificate of Calibration

Cost Center: DIV20

Mail Stop: B51

Customer: DON BANNON

Manufacturer/Model: VAISALA / MI70

Description: HUMIDITY/ TEMPERATURE METER

Serial Number: Z0730015

Asset Number: 010699

Procedure: TEMPERATURE, HUMIDITY, DEW-POINT - 3 APR 08

Work Order: 303089337

Date Issued: 21-Aug-2009

Date Calibrated: 20-Aug-2009

*** Date Due :** 20-Aug-2010

**** Results:** FOUND-LEFT

Temperature: 75.0 °F

Humidity: 46 %RH

Barometer: N/A

This certificate documents traceability to the National Institute of Standards and Technology (NIST) and the International System of Units (SI). The Laboratory quality system conforms to ISO/IEC 17025, 2005, ANSI/NCSL Z540-1-1994 and relevant requirements of the ISO 9000-2000 standard. This certificate shall not be reproduced, except in full, without the written approval of the Southwest Research Institute Calibration Laboratory. This certificate shall not be used to claim product endorsement by Southwest Research Institute, American Association for Laboratory Accreditation (A2LA) or any agency of the U. S. Government. Results of this calibration relate only to the instrument described above at the time of calibration and does not imply any long term stability of the instrument.

*Determined by the customer, does not imply the instrument will remain within tolerance as any number of factors may cause an out-of-tolerance condition before this date. **Data type found in this certificate or attached measurement report must be interpreted as: Found-left - adjustment and/or repair was not performed, As-found - data is before unit is adjusted and/or repaired, As-left - data is after adjusted and/or repaired was performed. The customer has sole responsibility for determination of in-/out-of-tolerance or compliance/noncompliance.

Measurement uncertainty calculated in accordance with the method described in the ISO "Guide to the Expression of Uncertainty in Measurement" (GUM), for a confidence level of approximately 95 percent using a coverage factor of $k=2$.

Remarks: DISPLAY READOUT. CAL WITH VAISALA PROBES A/N 010700 AND A/N 010701. SEE PROBES FOR DATA.

Standards Used

<u>Asset #</u>	<u>Manufacturer</u>	<u>Model</u>	<u>Description</u>	<u>Cal Date</u>	<u>Due Date</u>
006404	THUNDER SCIENTIFIC	2500	HUMIDITY GENERATOR	27-May-2009	27-May-2010
009414	HART SCIENTIFIC	1502A	TEMPERATURE READOUT	12-Jun-2009	12-Dec-2009
015895	HART SCIENTIFIC	5618B	RTD	12-Jun-2009	12-Dec-2009

Walt Hill

Laboratory Manager

m:\A2LA OCT_08.rpt

Mark Romero

Metrology Technician