



SOUTHWEST RESEARCH INSTITUTE®

6220 Culebra Road, P.O. Drawer 28510
Institute Quality Systems
Institute Calibration Laboratory
Phone: 210-522-5215 Fax 210-522-4834



Calibration Laboratory
Certificate #0972-01

Certificate of Calibration

Cost Center: DIV20

Mail Stop: B51

Customer: DON BANNON

Manufacturer/Model: TROEMNER / 400G

Description: WEIGHT, CLASS 1

Serial Number: 66389

Asset Number: 009342

Procedure: WEIGHTS - 10 DEC 07

Work Order: 303089334

Date Issued: 3-Aug-2009

Date Calibrated: 3-Aug-2009

*** Date Due :** 3-Aug-2010

**** Results:** FOUND-LEFT

Temperature: 68°F

Humidity: 41 %

This certificate documents traceability to the National Institute of Standards and Technology (NIST) and the International System of Units (SI). The Laboratory quality system conforms to ISO/IEC 17025, 2005, ANSI/NCSL Z540-1-1994 and relevant requirements of the ISO 9000-2000 standard. This certificate shall not be reproduced, except in full, without the written approval of the Southwest Research Institute Calibration Laboratory. This certificate shall not be used to claim product endorsement by Southwest Research Institute, American Association for Laboratory Accreditation (A2LA) or any agency of the U. S. Government. Results of this calibration relate only to the instrument described above at the time of calibration and does not imply any long term stability of the instrument.

*Determined by the customer, does not imply the instrument will remain within tolerance as any number of factors may cause an out-of-tolerance condition before this date. **Data type found in this certificate or attached measurement report must be interpreted as: Found-left - adjustment and/or repair was not performed, As-found - data is before unit is adjusted and/or repaired, As-left - data is after adjusted and/or repaired was performed. The customer has sole responsibility for determination of in-/out-of-tolerance or compliance/noncompliance.

Measurement uncertainty calculated in accordance with the method described in the ISO "Guide to the Expression of Uncertainty in Measurement" (GUM), for a confidence level of approximately 95 percent using a coverage factor of $k=2$.

Remarks: THE CONVENTIONAL MASS VALUE OF THE 400 g WGT IS: 400.000 47 g UNC +/-0.49 mg. CLASS 1 TOL +/- 1 mg.

Standards Used

| <u>Asset #</u> | <u>Manufacturer</u> | <u>Model</u> | <u>Description</u> | <u>Cal Date</u> | <u>Due Date</u> |
|----------------|---------------------|--------------|--------------------|-----------------|-----------------|
| 005117 | RICE LAKE | 200G | WEIGHT, CLASS E2 | 16-Apr-2009 | 16-Apr-2010 |
| 007103 | TROEMNER | 200G | WEIGHT, CLASS E1 | 29-Aug-2008 | 29-Aug-2009 |
| 012069 | SARTORIUS | CC1201 | MASS COMPARATOR | 29-Jan-2009 | 29-Jan-2010 |

Walt Hill

Laboratory Manager

Carlos Mendoza

Metrology Technician

Southwest Research Institute
 Calibration Laboratory
 Double Substitution Worksheet
per SOP 4, December 2003

Asset Number: 009342
 Date of Calibration: 8/3/2009
 Accuracy Class: 1

| Nominal | Reading | Unit |
|---|----------|------|
| 400 | | g |
| 1. S | 0.0000 | g |
| 2. X | 0.0008 | g |
| 3. X | 0.0008 | g |
| 4. S | 0.0001 | g |
| ▲ | 0.000750 | g |
| ▲ = d_x $d_x = 0.000750$ g $C_s = 399.9997230$ g $C_x = C_s + d_x$ $C_x = 400.000473$ g | | |