CONVERSATION RECORD

TIME | DATE 10/23/09

O VISIT O CONFERENCE X TELEPHONE

0 INCOMING
X OUTGOING

NAME OF PERSON(S) CONTACTED OR IN CONTACT	ORGANIZATION (OFFICE, DEPT.,ETC.)	TELEPHONE NO.
Dave Schmidt, RSO	Lakeview Diagnostic (Docket No. 03036222)	810-987-3317

SUBJECT

Follow-Up to EN 45455

SUMMARY

Lakeview Diagnostic (licensee) identified that the contaminant associated with the event described in EN 45455 was Tc-99m. Upon return of the package from Port Huron Hospital, the licensee identified a very small spot of contamination that measured 5 mR/hr and 300 cpm/100 square centimeters at the surface. The RSO stated that required pre-shipment package surveys were conducted. The licensee is continuing its investigation of the cause of the event. The licensee agreed to contact the inspector regarding the results of its investigation of the cause of the event.

ACTION REQUIRED

Place in ADAMS.

cc: Tammy Bloomer

Obtain the licensee's results of its investigation of the cause of the event, compare it with the results of Port Huron Hospital's investigation, and recommend next steps to the MIB Chief.

NAME OF PERSON DOCUMENTING CONVERSATION	SIGNATURE	DATE
Bob Gattone	Robert D. Gattone, Jr.	10/23/09

ACTION TAKEN

SIGNATURE	TITLE	DATE