

## NATIONAL SPENT NUCLEAR FUEL PROGRAM ENGINEERING DESIGN FILE

EDF-NSNF- 082

Revision 0

Page 1 of 3

Title: IMPACT TENSILE TESTING OF STAINLESS STEELS AT VARIOUS TEMPERATURES

1. Activity Title: Transportation and Packaging – Canister Basket

2. WBS No. C.B.30.03.02.02.C2

3. Approval	Typed Name	Signature	Date
Preparer	D. K. Morton	<i>D. K. Morton</i>	3/31/08
Preparer	R. K. Blandford	<i>R. K. Blandford</i>	3/31/08
Reviewer	S. D. Snow	<i>S. D. Snow</i>	3/31/08
PSO QE	N. S. MacKay	<i>N. S. MacKay</i>	3/31/08

4. Distribution: (Name and Mail Stop)	<b>Project File</b> B. W. Carlsen, MS 3710; N. S. MacKay, MS 3710; R. K. Blandford, MS 3760; D. K. Morton (10 copies), MS 3760; S.D. Snow, MS 3760	<b>NSNFP File Log No.: 4724.11</b>
--	--	------------------------------------

## Title: IMPACT TENSILE TESTING OF STAINLESS STEELS AT VARIOUS TEMPERATURES

5. Purpose: Stainless steels are used for the construction of numerous spent nuclear fuel or radioactive material containers that may be subjected to high strains and moderate strain rates during accidental drop events. Mechanical characteristics of these base materials and their welds under dynamic loads in the strain rate range of concern (1 to 300 per second) are not well documented. However, research is being performed at the Idaho National Laboratory to quantify these characteristics.

The work presented herein discusses tensile impact testing of dual-marked 304/304L and 316/316L stainless steel material specimens. Both base material and welded material specimens were tested at -20 °F, room temperature, 300 °F, and 600 °F conditions. Utilizing a drop weight impact test machine and 1/4-inch and 1/2-inch thick dog bone-shaped test specimens, a strain rate range of approximately 4 to 40 per second (depending on initial temperature conditions) was achieved. Factors were determined (see table below for typical values) that reflect the amount of increased strain energy the material can absorb due to strain rate effects. Using the factors, elevated true stress-strain curves for these materials at various strain rates and temperatures were generated.

Strain rate (per sec.)	-20 °F	Room Temperature	300 °F	600 °F
304L Stainless Steel				
5	1.333	1.235	1.166	1.043
10	1.361	1.278	1.210	1.094
22	1.428	1.381	1.316	1.217
25	1.445	1.407	1.342	1.247
316L Stainless Steel				
5	1.275	1.265	1.162	1.040
10	1.296	1.281	1.187	1.070
22	1.346	1.321	1.247	1.140
25	1.359	1.331	1.262	1.158

By incorporating the strain rate elevated true stress-strain material curves into an inelastic finite element computer program as the defined material input, significant improvement in the accuracy of the computer analyses was attained. However, additional impact testing is necessary to achieve higher strain rates (up to 300 per second) before complete definition of strain rate effects can be made for accidental drop events and other similar energy-limited impulsive loads.

This research approach, using impact testing and a total energy analysis methodology to quantify strain rate effects, can be applied to many other materials used in government and industry.

The current, principal NSNFP procedures applied to this activity include the following:

- NSNFP Procedure 6.01, *Review and Approval of NSNFP Internal Documents*
- NSNFP Procedure 6.03, *Managing Document Control and Distribution*
- NSNFP Procedure 3.04, *Engineering Documentation*.



# NATIONAL SPENT NUCLEAR FUEL PROGRAM ENGINEERING DESIGN FILE

EDF-NSNF- 082

Revision 0

Page 3 of 3

Title: IMPACT TENSILE TESTING OF STAINLESS STEELS AT VARIOUS TEMPERATURES

Total Attachments:	Attachment Nos.:	No. of pages in each:	Total Pages:
9	Cover sheets, pgs i-x	10	157
	Report, pgs 1-118	118	
	Appendix A, pgs A1-A6	6	
	Appendix B, pgs B1-B2	2	
	Appendix C, pgs C1-C3	3	
	Appendix D, pgs D1-D4	4	
	Appendix E, pgs E1-E2	2	
	Appendix F, pgs F1-F6	6	
	Appendix G, pgs G1-G6	6	

# **IMPACT TENSILE TESTING OF STAINLESS STEELS AT VARIOUS TEMPERATURES**

**WBS Number C.B.30.03.02.02.C2**

**March 2008**

**Authors:**

**D. K. Morton and R. K. Blandford**

Prepared for the  
U. S. Department of Energy  
Assistant Secretary of Environmental Management  
Under DOE Idaho Operations Office  
Contract DE-AC07-05ID14517

*This document was developed and is controlled in accordance with NSNFP procedures.  
Unless noted otherwise, information must be evaluated for adequacy relative to its specific  
use if relied on to support design or decisions important to safety or waste isolation.*

# CONTENTS

<b>1. INTRODUCTION AND PURPOSE .....</b>	<b>1</b>
<b>2. QUALITY ASSURANCE .....</b>	<b>3</b>
2.1 QUALITY REQUIREMENTS .....	3
2.2 SERVICES PROVIDED BY OUTSIDE SUPPLIERS.....	4
2.2.1 <i>Welding and Weld Examination Services for Welded Plates</i> .....	4
2.2.2 <i>Calibration Services</i> .....	5
2.2.3 <i>Computer Support Services for Validated Software</i> .....	5
2.2.4 <i>Confirmatory Dimensional Measuring Services</i> .....	5
2.2.5 <i>Quasi-Static Tensile Testing Services</i> .....	6
2.3 COMPUTER PROGRAM VALIDATION AND CONFIGURATION MANAGEMENT .....	7
2.3.1 <i>I-DEAS Software</i> .....	7
2.3.2 <i>ABAQUS/Explicit Software</i> .....	7
2.3.3 <i>DADISP 2002 Software</i> .....	7
2.3.4 <i>Excel Software</i> .....	8
2.3.5 <i>Digital Motion Analysis Software</i> .....	8
<b>3. REPRESENTATION OF STRESS-STRAIN BEHAVIOR .....</b>	<b>9</b>
3.1 QUASI-STATIC TENSILE TESTING .....	9
3.2 TYPICAL VARIATIONS IN MATERIAL PROPERTIES .....	11
3.3 STRAIN RATE EFFECTS .....	12
3.4 RECOGNITION OF ENERGY DENSITY IN TRUE STRESS-STRAIN CURVES .....	12
<b>4. METHODOLOGY .....</b>	<b>14</b>
<b>5. QUASI-STATIC TENSILE TESTING .....</b>	<b>15</b>
5.1 TEST MATERIAL PROCUREMENT .....	16
5.1.1 <i>Initial Testing</i> .....	16
5.1.2 <i>Additional Testing</i> .....	16
5.2 QUASI-STATIC TEST SPECIMEN GEOMETRIES.....	24
5.2.1 <i>Initial Testing</i> .....	24
5.2.2 <i>Additional Testing</i> .....	25
5.3 QUASI-STATIC TEST PROCEDURE .....	29
5.3.1 <i>Initial Testing</i> .....	29
5.3.2 <i>Additional Testing</i> .....	30
5.4 ACHIEVING TEMPERATURE CONDITIONS .....	33
5.4.1 <i>Initial Testing</i> .....	33
5.4.2 <i>Additional Testing</i> .....	33
5.5 QUASI-STATIC TENSILE TESTING RESULTS.....	35
5.5.1 <i>Initial Testing</i> .....	38
5.5.2 <i>Additional Testing</i> .....	38
<b>6. IMPACT TENSILE TESTING .....</b>	<b>42</b>
6.1 IMPACT TESTING MACHINE .....	42
6.2 TEST SPECIMENS.....	47
6.2.1 <i>Pre-Test Dimensional Measurements</i> .....	52
6.2.2 <i>Final Pre-Test Preparations</i> .....	54
6.3 IMPACT TEST PROCEDURE .....	55
6.4 ACHIEVING IMPACT TEMPERATURE CONDITIONS .....	58
6.4.1 <i>Room Temperature Testing</i> .....	58
6.4.2 <i>Elevated Temperature Testing</i> .....	59
6.4.3 <i>Cold Temperature Testing</i> .....	70
6.5 DATA ACQUISITION.....	73

6.5.1	High-Speed Digital Camera.....	74
6.5.2	Accelerometer and Velocimeter Data .....	75
6.6	PREVIOUS ITM PERFORMANCE.....	76
6.6.1	Test Performance at Different Test Sites .....	76
6.6.2	Longitudinal Versus Transverse Specimen Orientation .....	77
6.7	POST-TEST DIMENSIONAL MEASUREMENTS .....	78
6.7.1	Test Specimen Deformations.....	78
6.7.2	Certified Dimensional Inspector Validations.....	80
6.8	CLOSEOUT CALIBRATION .....	80
<b>7.</b>	<b>TEST RESULTS: DEVELOPMENT OF FACTORS AND STRAIN RATE ELEVATED TRUE STRESS-STRAIN CURVES .....</b>	<b>81</b>
7.1	DEVELOPMENT OF FACTORS.....	81
7.1.1	Development of Spreadsheets.....	83
7.1.2	Checking of Spreadsheets .....	93
7.2	STRAIN RATE ELEVATED TRUE STRESS-STRAIN CURVES.....	95
<b>8.</b>	<b>USING RESULTS: ABAQUS/EXPLICIT ANALYSES USING STRAIN RATE ELEVATED TRUE STRESS-STRAIN CURVES .....</b>	<b>106</b>
<b>9.</b>	<b>OBSERVATIONS FROM TEST RESULTS .....</b>	<b>110</b>
9.1	QUASI-STATIC TENSILE TESTING .....	110
9.2	IMPACT TENSILE TESTING.....	111
9.2.1	Impact Test Anomalies.....	111
9.2.2	Base Versus Welded Material Responses.....	111
9.3	FACTORS AND STRAIN RATE ELEVATED TRUE STRESS-STRAIN CURVES INSIGHTS .....	114
<b>10.</b>	<b>CONCLUSIONS.....</b>	<b>115</b>
<b>11.</b>	<b>REFERENCES.....</b>	<b>116</b>

## APPENDICES

APPENDIX A	– DADiSP INTEGRATION VERIFICATION .....	A1
APPENDIX B	– ENGINEERING SIGNOFFS FOR SPREADSHEET CALCULATIONS ..	B1
APPENDIX C	– INL WELDING PROCEDURE S2.0 .....	C1
APPENDIX D	– REPRESENTATIVE TEST SPECIMEN DATA SHEET .....	D1
APPENDIX E	– DROP WEIGHT COMPARISON .....	E1
APPENDIX F	– INL DIMENSIONAL INSPECTOR QUALIFICATIONS .....	F1
APPENDIX G	– ELECTRONIC FILE INFORMATION .....	G1

## **ACKNOWLEDGMENT**

The authors would like to acknowledge the help, technical expertise, perseverance, and wisdom that Tom E. Rahl (NSNFP test team member and recently retired from the Idaho National Laboratory) provided to this material impact testing effort as well as the friendship shared with the authors.



## ABSTRACT

Stainless steels are used for the construction of numerous spent nuclear fuel or radioactive material containers that may be subjected to high strains and moderate strain rates during accidental drop events. Mechanical characteristics of these base materials and their welds under dynamic loads in the strain rate range of concern (1 to 300 per second) are not well documented. However, research is being performed at the Idaho National Laboratory to quantify these characteristics.

The work presented herein discusses tensile impact testing of dual-marked 304/304L and 316/316L stainless steel material specimens. Both base material and welded material specimens were tested at -20 °F, room temperature, 300 °F, and 600 °F conditions. Utilizing a drop weight impact test machine and 1/4-inch and 1/2-inch thick dog bone-shaped test specimens, a strain rate range of approximately 4 to 40 per second (depending on initial temperature conditions) was achieved. Factors were determined (see table below for typical values) that reflect the amount of increased strain energy the material can absorb due to strain rate effects. Using the factors, elevated true stress-strain curves for these materials at various strain rates and temperatures were generated.

Strain rate (per sec.)	-20 °F	Room Temperature	300 °F	600 °F
304L Stainless Steel				
5	1.333	1.235	1.166	1.043
10	1.361	1.278	1.210	1.094
22	1.428	1.381	1.316	1.217
25	1.445	1.407	1.342	1.247
316L Stainless Steel				
5	1.275	1.265	1.162	1.040
10	1.296	1.281	1.187	1.070
22	1.346	1.321	1.247	1.140
25	1.359	1.331	1.262	1.158

By incorporating the strain rate elevated true stress-strain material curves into an inelastic finite element computer program as the defined material input, significant improvement in the accuracy of the computer analyses was attained. However, additional impact testing is necessary to achieve higher strain rates (up to 300 per second) before complete definition of strain rate effects can be made for accidental drop events and other similar energy-limited impulsive loads.

This research approach, using impact testing and a total energy analysis methodology to quantify strain rate effects, can be applied to many other materials used in government and industry.





## ACRONYMS

ASME	American Society of Mechanical Engineers
ASTM	American Society for Testing and Materials
B&PV	Boiler and Pressure Vessel
CFA	Central Facilities Area (at INL)
CITRC	Critical Infrastructure Test Range Complex (at INL)
CMTR	certified material test report
DOE	U.S. Department of Energy
FY	fiscal year
GV	gauge volume
HRAP	hot rolled, annealed, and pickled
INL	Idaho National Laboratory
IRC	INL Research Center
ITM	Impact Testing Machine
LOP	lack of penetration (pertinent to weld examination)
LVDT	linear variable displacement transducer
NVLAP	National Voluntary Laboratory Accreditation Program
NSNFP	National Spent Nuclear Fuel Program
OAS	overall average strain
OCRWM	Office of Civilian Radioactive Waste Management (DOE-RW)
QA	quality assurance
SNF	spent nuclear fuel
UTS	ultimate tensile strength



# **IMPACT TENSILE TESTING OF STAINLESS STEELS AT VARIOUS TEMPERATURES**

## **1. INTRODUCTION AND PURPOSE**

Most structural design acceptance criteria are stress-based. These stress-based criteria are prudent for normal operating conditions where static loading due to pressure, weight, and thermal loads dominate. Years of experience have repeatedly demonstrated that these criteria (e.g., ASME B31.1 Power Piping Code) have achieved successful designs. However, other component designs become excessively conservative when satisfying these stress-based design criteria for the high stresses typically resulting from low probability, off-normal dynamic events, such as an accidental drop. This conservatism frequently manifests itself as thicker material in an inefficient design. The downside of this inefficiency is that it can significantly elevate life cycle costs and potentially degrade structural performance over the design life of the component.

By extending the knowledge of material responses beyond the typical ‘quasi-static’ uniaxial tensile test behavior, the accuracy of stresses/strains predicted by finite element method (FEM) inelastic analysis computer codes for low probability (but design governing) dynamic events, such as accidental drops, can be improved. Accurate knowledge of the actual responses of components or structures to dynamic loads means improved designs, safer designs, more cost-effective designs, and better overall engineered solutions to the many national infrastructure problems at hand. Examples of dynamic loads include, but are not limited to, radioactive material container drop events, vehicle crashes, vehicle crashes into highway safety barriers or security barriers, and more. These dynamic loads typically yield strain rate responses that are greater than 0.01 per second but less than 300 per second. These moderate strain rates are higher than quasi-static tensile testing rates yet less than strain rates associated with ballistics.

In the past, full-scale component testing was the norm in evaluating the structural integrity of vehicle designs during crashes or the structural integrity and containment capabilities of spent nuclear fuel containers during accidental drop events. Methods for performing nonlinear inelastic analyses were either lacking or very approximate at best. Currently, with improved computational capabilities and software, the trend is to use nonlinear analytical methods, with limited or no actual testing, once the acceptability of the analysis methodology has been established. Improved software and analysis methodologies for performing inelastic, large deformation analyses are now common and offer numerous advantages relative to full-scale component testing, including relatively low cost analytical simulations, ease of evaluating material and design options, elimination of costs associated with actual fabrication, testing, and post-test disposal, etc. In order to rely only on an analytical approach, accurate results from methodologies and software must be demonstrated which in turn mandate a precise definition of inelastic, dynamic material properties (e.g. true stress-strain curves reflecting strain rate effects). Other variables such as temperature, welded material properties, aged material properties, and project specific conditions (if appropriate) must also be considered.

True stress-strain curves at elevated strain rates (in the range of 1 to 300 per second) are not readily available. Additionally, most published material data do not consider variables such as: full uniform strain range; varying strain rates; temperature; specific material composition and specification; welded material properties; or aged material properties (if appropriate), especially in combinations.

Most metals tend to get stronger as the rate of straining increases. Taking advantage of this phenomenon is desirable when designing for low probability dynamic events. The American Society of Mechanical Engineers (ASME) Boiler and Pressure Vessel (B&PV) Code, Section III, Division 1, Appendix F (Reference 1) contains rules for performing inelastic (large deformation) analyses. Paragraph F-1322.3(c) permits the adjustment of the stress-strain curve to include strain rate effects resulting from dynamic behavior (e.g., during drop events). However, these code rules also mandate the justification of that adjusted stress-strain curve. Therefore, it is imperative to perform material impact testing in order to provide the necessary justification for using these strain rate elevated true stress-strain curves for inelastic analyses that invoke this Code. The phrase “strain rate elevated” refers to the fact that the strain rate effects on a stress-strain curve reflecting dynamic loading raises the curve upward (higher strength) when compared to a stress-strain curve reflecting quasi-static conditions.

The development of beyond-yield-strength true stress-strain curves of commonly used construction materials, addressing both strain rate and temperature effects is necessary to achieve accurate analytical results. In order to begin to consider these variables, the Department of Energy’s (DOE) National Spent Nuclear Fuel Program (NSNFP), working with the Office of Civilian Radioactive Waste Management (OCRWM), and the Idaho National Laboratory (INL), has supported an initial effort into the study of strain rate phenomena. This first phase of material impact testing considered 304/304L and 316/316L dual-stamped stainless steel (hereafter referred to as 304L and 316L, respectively). Both base and welded materials at -20 °F, room, 300 °F, and 600 °F temperatures were tested at strain rates between 4 and 40 per second. The strain rates achieved during impact testing were dependent on the temperature of the material as well as the amount of energy imparted to the material. This report addresses the first phase of completed testing. Higher strain rates of interest (40 to 300 per second or more) at the same temperature conditions, using test specimens from the same material heats and welds must be investigated in the future in order to span the entire range of strain rates of interest.

The objective of this research effort was to improve understanding of the strain rate phenomenon by experimentally studying the mechanical properties of candidate materials subjected to impact loading. The purpose of this task was to determine strain rate effects for 304L and 316L stainless steel material under dynamic, impact loading at various cold and hot temperatures. The goal is to ultimately develop true stress-strain curves reflecting various strain rates and temperatures for many materials and provide justification of each strain rate elevated true stress-strain curve. The test data developed can be used to establish an analysis methodology that can then be applied in analytical simulations to more accurately predict the deformation and resulting material straining in the components being evaluated that are subject to dynamic, impulsive loads. The long-term goal is to develop sufficient data to provide clear and distinct guidance regarding impact analysis methodologies and how engineering personnel can perform these analyses and obtain viable results without needing to perform confirmatory testing. This work can also help establish strain-based acceptance criteria for these events.

## 2. QUALITY ASSURANCE

This document was developed and is controlled in accordance with NSNFP procedures. Unless noted otherwise, information must be evaluated for adequacy relative to its specific use if relied on to support design or decisions important to safety or waste isolation.

The NSNFP procedures applied to this activity implement DOE/RW-0333P, *Quality Assurance Requirements and Description* (Reference 2), and are a part of the NSNFP QA Program (Reference 3). The NSNFP QA Program has been assessed and accepted by representatives of the Office of Quality Assurance within the Office of Civilian Radioactive Waste Management for work scope of the NSNFP.

The current, principal NSNFP procedures applied to this activity include the following:

- NSNFP 6.01, *Review and Approval of NSNFP Internal Documents* (Reference 4),
- NSNFP 6.03, *Managing Document Control and Distribution* (Reference 5),
- NSNFP 3.04, *Engineering Documentation* (Reference 6).

### 2.1 Quality Requirements

The material testing effort described in this report was required to satisfy the NSNFP Quality Assurance Program. This required the generation of a test plan, establishment of agreements for quality-affecting activities that the NSNFP could not perform, initiation of laboratory (or scientific) notebook usage during the research effort, and documentation of the material testing effort with a final report.

As identified in NSNFP Procedure 11.01 (Reference 7), a test plan prescribes the requirements, controls, and documentation necessary for testing conducted within the NSNFP. The test plan is to provide background information and objectives, identify test requirements, describe the test methodology, specify the test equipment, software, and procured services, state the applicable acceptance criteria, describe the test procedure, and identify the information necessary for the test documentation. If any quality-affecting services need to be procured, NSNFP Procedure 4.02 (Reference 8) describes the process for acquiring government sector services for the NSNFP, including generating a NSNFP Task Management Agreement document. A Task Management Agreement is written to clarify the NSNFP needs, identify any technical requirements, specify deliverables, and discuss other issues required of the provider of the outside services. The providers sign the Task Management Agreement, clearly indicating their awareness of what is required of them.

Regarding the documentation of technical information during actual impact testing, NSNFP Procedure 3.04 identifies the requirements for creating and using a laboratory notebook. Pertinent impact test data were recorded in the laboratory notebook. Additional test data (such as quasi-static tensile test results or voluminous deformation history data) were generated during the research effort but were placed into separate binders (referenced in the laboratory notebook) and are considered part of the laboratory notebook documentation. During periods of test activity, the laboratory notebook was reviewed at approximately monthly intervals by a

cognizant person (other than the individual filling in the notebook). This laboratory notebook was submitted to the NSNFP to be retained as a quality assurance (QA) record.

The personnel performing the material impact research had ten years or more of engineering experience in structural evaluations and were trained per NSNFP Procedure 2.04 (Reference 9). Software used to perform finite element analysis comparisons to test specimen result met NSNFP 19.01 (Reference 10). Software used to evaluate test data (by performing simple mathematical functions such as adding, multiplying, or integration similar to those capabilities on a handheld calculator) were confirmed with sample problem checks or reviewed like an engineering calculation and reported on in the final documentation (this report). This final test report followed the requirements of NSNFP Procedure 6.01.

Although INL's Consumer Grade – Quality Level 4 process (currently referred to as Commercial Grade – Quality Level 3) was used to procure test specimen materials, ASME Section III approved materials with certified material test reports (CMTRs) were purchased. For confirmation, material samples were tensile tested to verify the CMTR data and to establish the material quasi-static stress-strain curves used by this strain rate research effort.

## **2.2 Services Provided By Outside Suppliers**

Certain services not addressed by the NSNFP QA Program needed to be provided by outside suppliers in order for this investigation to proceed. As such, five different activities (addressed in the following subsections) were considered to be quality-affecting. These activities included: (1) welding and weld examination services for the welded plates; (2) calibrating the measuring devices used to determine pre-impact test specimen measurements and post-impact deformations; (3) providing properly controlled hardware and software to perform computer evaluations to predict resulting test specimen deformations; (4) performing confirmatory dimensional measurements of test specimen gages and post-impact test specimens; and (5) determining quasi-static tensile material property (stress-strain) data. These quality-affecting activities were supplied by the INL as detailed in two NSNFP Task Management Agreement documents, DOE/SNF/TMA-009, Revision 1 (Reference 11) for welding and weld examination services and DOE/SNF/TMA-013, Revision 1 (Reference 12) for the remaining four services. Test Plan DOE/SNF/PP-039, Revision 8 (Reference 13) also discussed these quality-affecting activities.

### **2.2.1 Welding and Weld Examination Services for Welded Plates**

Welding and weld examination services were quality-affecting activities that could not be performed by the NSNFP. Therefore, these activities were performed by a qualified supplier, the INL. This service is discussed in Section 2.3.3 of the Task Management Agreement DOE/SNF/TMA-009. The INL prepared the welded plates (from which test specimens were cut) for dynamic testing in support of Test Plan DOE/SNF/PP-039. The welded plates incorporated the weld joint design anticipated for the Standardized DOE Spent Nuclear Fuel Canisters. Full volumetric radiographic examinations were performed on the completed welds to verify the integrity of the welds prior to dynamic load testing. Examination reports, weld wire material certifications, and the INL Laboratory Notebooks (Reference 14) were submitted to the NSNFP.

### **2.2.2 Calibration Services**

The NSNFP required the services of the INL to initiate or maintain calibration and the associated documentation for specific NSNFP-identified measuring devices, including but not limited to: micrometers, calipers, load cells, accelerometers, etc. Calibration was performed in accordance with INL procedures. The Test Plan (DOE/SNF/PP-039) specified the project-defined minimum accuracy required for particular NSNFP measuring devices.

The basis for acceptance of calibration services included the calibration labels applied directly to each measuring device (or provided to the NSNFP test personnel for those items where affixing a label was not viable) and the applicable calibration sheet(s). These calibration sheets were made available to the NSNFP test personnel as each measuring device completed calibration prior to use. The INL maintains qualification records for all personnel providing calibration services. The NSNFP test personnel retained applicable calibration sheets for NSNFP record keeping.

### **2.2.3 Computer Support Services for Validated Software**

The NSNFP required computer hardware/software services of the INL to provide a computing environment for running the ABAQUS/Explicit software. The ABAQUS/Explicit software was already installed on the identified compute server 'Aurora'. In order to provide adequate documentation in compliance with NSNFP procedures (NSNFP 19.01), the NSNFP performed its own installation test and validation of ABAQUS/Explicit on the hardware and operating system software configuration identified. Any subsequent changes made by the INL that altered the hardware configuration, the operating system software configuration, or the ABAQUS/Explicit software (e.g., adding software updates) would likely nullify the NSNFP software validation. Therefore, the NSNFP requested that the INL maintain a hardware configuration, an operating system configuration, and an ABAQUS/Explicit software configuration (i.e., specific version of the software) for a reasonably long period of time (ideally one year or more). The NSNFP required early notification of any software or hardware configuration changes or scheduled or unexpected maintenance needs that the INL performed so that the impact of any such changes could be evaluated. ABAQUS/Explicit was revalidated prior to continued use after any such configuration changes.

Because the NSNFP performed its own validation and verification of software (e.g., ABAQUS/Explicit) per NSNFP 19.01, compliance with INL procedures associated with software verification and validation were not required. No additional technical requirements beyond those specified by INL procedures, applicable to computer systems operated by the INL are required by the NSNFP.

### **2.2.4 Confirmatory Dimensional Measuring Services**

The NSNFP required the services of the INL to implement the INL Quality Assurance Program requirements for Quality Engineering and Inspection activities. Those activities included dimensional measurements on NSNFP-provided test components or impact test specimens. The Test Plan DOE/SNF/PP-039 provided details and requirements with respect to dimensional measurements to be performed by INL Quality Assurance personnel. Sketches, drawings, data sheets, or specifications were provided by the NSNFP to supplement the Test Plan. Measuring devices used were calibrated by the INL Standards and Calibration Laboratory.



Measurement documentation was provided to the NSNFP test personnel at the completion of the measurement efforts.

The correct dimensional measurement of the test specimens before and after impact testing was significant to the success of this investigation. NSNFP test personnel measured the test specimens before and after impact testing. Pre-test measurements included the use of go/no go gages. Plexiglas templates, referred to as go/no go gages, were developed so that quick, accurate checks of the overall test specimen geometry could be made. These go/no go gages were measured by a qualified INL dimensional inspector and accepted for use by the NSNFP test personnel based on those INL measurements. Pre- and post-test measurements of test specimens were performed by NSNFP test personnel using calibrated calipers. However, to demonstrate the validity of these measurements, post-test measurements of identified impact test specimens were also taken by a qualified INL dimensional inspector.

### **2.2.5 *Quasi-Static Tensile Testing Services***

The NSNFP required the services of the INL to perform material quasi-static tensile testing at varying temperatures in the range of -20 °F to 600 °F. The goal was to obtain sufficient data to adequately plot the quasi-static true stress-strain curve up to test specimen failure. Test Plan DOE/SNF/PP-039 provided details and requirements with respect to this material tensile testing effort. Basic testing requirements were to follow ASTM Standard A 370 (Reference 15), excluding any identified reporting requirements. Sketches, drawings, data sheets, or test requirements were provided by the NSNFP to supplement the Test Plan as appropriate. The tensile test device and any other associated instrumentation providing pertinent data were calibrated by the INL Standards and Calibration Laboratory. NSNFP test personnel or other personnel supporting the NSNFP witnessed the material tensile testing process.

For material initially procured to investigate room temperature strain rate effects in 2004, the quasi-static tensile testing was performed by the Materials Testing Laboratory located at the central facilities of the INL (at CFA-602). The NSNFP Program Applicability Evaluation PAE-010 (Reference 16) provided details of how this INL National Voluntary Laboratory Accreditation Program (NVLAP) qualified supplier would perform this service.

The quasi-static material tensile testing performed in 2006 and 2007 for material procured to investigate strain rate effects at varying temperatures was performed at a laboratory located at the INL Research Center (IRC). This service is discussed in Task Management Agreement DOE/SNF/TMA-013. Since the cognizant laboratory personnel at IRC did not have NVLAP accreditation, personnel associated with the Materials Testing Laboratory provided oversight to assure compliance with ASTM A370 requirements. Material data obtained [continuous force-strain (or displacement) data through test specimen failure] was provided to the NSNFP at the completion of the tensile tests. The equipment used for this quasi-static tensile testing was calibrated through the INL Standards and Calibration Laboratory and documented in the INL Laboratory Notebook LAB-771 (Reference 17) submitted to the NSNFP for record retention.

## **2.3 Computer Program Validation and Configuration Management**

Several activities used in this task to evaluate and operate on test data incorporated computer programs including I-DEAS (Reference 18), ABAQUS/Explicit, Version 6.6-3 (Reference 19), DADiSP 2002 (Reference 20), a collection of EXCEL spreadsheets referred to herein as NSNF/MED/017, Revision 1 (Reference 21), and digital motion analysis software used in conjunction with the high-speed digital camera.

### **2.3.1 I-DEAS Software**

The I-DEAS Master Series solid modeling computer program was used to create the finite element models of the physical impact tests. A solid model was created first, and then used to generate the finite element model. The following is a summary of the configuration management for I-DEAS. I-DEAS was not used for the calculations – only for modeling purposes. Model verification was performed using ABAQUS/Explicit, Version 6.6-3.

Program Used: I-DEAS 10 NX Series

Computer Used: Dell Precision 450, U.S. Govt. ID 374043

### **2.3.2 ABAQUS/Explicit Software**

The computer program ABAQUS/Explicit, a linear and nonlinear finite element analysis software package that is widely used in many industries, was employed to calculate the response of the test specimen model to the impact test events. The ABAQUS/Explicit analytical results were compared to the measured experimental data. Version 6.6-3 was the NSNFP validated version (Reference 22). The following is a summary of the configuration management and validation performed for ABAQUS/Explicit.

Program Used: ABAQUS/Explicit Version: 6.6-3

Computer Used: SGI Model: Altix 4700 (Aurora)

#### **Verification Manual/Test Problem Manual/ Example Manual:**

*ABAQUS Example Problems Manual*, Version 6.6, ABAQUS, Inc.,

Providence, Rhode Island, 2006. (Reference 23)

#### **NSNFP Validation:**

*Software Report For ABAQUS/Explicit Version 6.6-3*, DOE/SNF/REP-107,

Rev. 0, November 2006. (Reference 22)

### **2.3.3 DADiSP 2002 Software**

DADiSP 2002 (hereafter referred to as DADiSP) is a technical data analysis and display program that was used to reduce, manipulate, operate on, and plot test data using mathematical functions in a spreadsheet type environment. Therefore, a complete verification and validation

effort was not deemed necessary. However, insight into the validity of DADiSP results was important. Hence, an effort to verify the specific functions used in DADiSP for this report was completed. Verification of the DADiSP integration functions used in this task is contained in Appendix A.

Program Used: DADiSP 2002

Computer Used: Dell M50 Portable Workstation, U.S. Govt. ID 371849  
Dell Precision 470 Workstation, U.S. Govt. ID 384252

#### **2.3.4 Excel Software**

The collection of Excel spreadsheets used to gather test data, generate stress-strain curves, and calculate the strain rate factors have been processed per the requirements of NSNFP Procedure 19.01. Rather than a verification/validation effort as was done for ABAQUS/Explicit, these spreadsheets were checked just as an engineering hand calculation would be checked. Appendix B identifies the author that generated each spreadsheet and the reviewer of each spreadsheet (including the reviewer's signature).

Software Used: Excel 2003

NSNFP Assigned Identifier: NSNF/MED/017, Revision 1

Computer Used: Dell M50 Portable Workstation, U.S. Govt. ID 371849  
Dell Precision 470 Workstation, U.S. Govt. ID 384252

#### **2.3.5 Digital Motion Analysis Software**

The recording of high-speed digital camera image data and motion analyses were performed with software that was provided with the high-speed digital camera. Only the high-speed digital camera's frame rates could be calibrated (for accurate timing). Since the image data is dependent on each unique camera view, it was not possible to validate the camera software ahead of time. However, the camera digital image was scaled prior to motion analysis to the specimen's measured gauge length between marked points on the test specimen. Additionally, for accuracy checking purposes and following the motion analysis, the final camera determined strain value was compared to the final strain value determined from test specimen measurements made with a calibrated caliper. See Section 6.5.1 for additional details.

Software Used: Photron Motion Tools, Version 1.2.0  
Photron Fastcam Viewer, Version 2.4.3.8

Computer Used: Dell M60 Portable Workstation, U.S. Govt. ID 376246  
Dell Precision 470 Workstation, U.S. Govt. ID 384252

### 3. REPRESENTATION OF STRESS-STRAIN BEHAVIOR

In a one-dimensional quasi-static tensile test, a uniform slender test specimen is stretched (ends displaced at a constant rate) along its long central axis. Such testing follows the requirements of ASTM A 370. The results are plotted as a force versus displacement curve, which is a representation of the performance of the test specimen material. From this force-displacement curve, engineering stress-strain and true stress-strain curves can be produced using classical engineering relationships.

The term quasi-static is sometimes applied to this ‘static’ test since a significant length of time and slow strain rate ( $10^{-5}$  to  $10^{-2}$  per second) are involved in performing the uniaxial tensile test. Strain rate behavior is defined as the difference in strength of a stress-strain curve produced under dynamic loading conditions compared to its quasi-static curve.

#### 3.1 Quasi-Static Tensile Testing

Stress-strain curves are usually presented as:

- ‘Engineering’ stress-strain curves, in which the original specimen cross-sectional area is used to determine stress and the change in length divided by the original length determines strain,
- ‘True’ stress-strain curves, where the instantaneous cross-sectional area of the specimen is used to determine the stress and the strain.

To document a quasi-static tensile test, an engineering stress-strain curve is developed from the load-displacement measurements made during the test on the test specimen (Figure 1, typical for ductile material). The engineering stress,  $S$ , plotted on this curve is the average longitudinal stress in the tensile specimen obtained by dividing the load,  $P$ , by the original specimen area,  $A_0$ . The engineering strain,  $e$ , plotted on the curve is the average linear strain obtained by dividing the change in gauge length,  $\Delta L$ , of the specimen by the original length,  $L_0$ .

$$S = P / A_0$$

$$e = \Delta L / L_0$$

The elastic limit, shown as point B in Figure 1, is the greatest stress the material can withstand without measurable permanent strain remaining after complete release of load. The yield strength, shown as point YS in Figure 1, is the stress required to produce a small, specified amount of inelastic deformation. The usual definition of this property is the offset yield strength determined by the stress corresponding to the intersection of the linear elastic segment of the stress-strain curve offset by a specified strain of 0.2% ( $e = 0.002$ ). The tensile strength, or ultimate strength,  $S_u$ , is the corresponding stress where the maximum load that the material can withstand occurs. This also corresponds to the point where the specimen becomes unstable (onset of necking) and necks down during the remaining course of the tensile test. Necking is the point of rapid, localized reduction of cross-sectional area of a specimen under tensile loading. It is disregarded in calculating engineering stress but is taken into account in determining true stress. Complete fracture (failure point) of the specimen follows necking.




Figure 1. Typical engineering tensile stress-strain curve [from the *Atlas of Stress-Strain Curves* (Reference 24)].

The engineering stress-strain curve does not give the most accurate indication of the deformation characteristics of a material because it is based on the original specimen dimensions that actually change continuously during the test. Also, at the point of ultimate load, necking begins and the cross-sectional area of the specimen decreases rapidly and the load required to continue deformation lessens, as implied in Figure 1. The average stress based on the original area likewise decreases, and produces the downturn in the engineering stress-strain curve beyond the point of maximum load. In reality, the material continues to strain harden to fracture, so that the stress required to produce further deformation should also increase. If the true stress, based on the actual cross-sectional area of the specimen is used, the stress-strain curve increases continuously to fracture. If the strain measurement is also based on instantaneous measurement, the curve obtained is the true stress-strain curve as illustrated in Figure 2. The true stress-strain curve is also known as the flow curve, because it represents the basic plastic-flow characteristics of the material. The true stress is often referred to as the flow stress.

Up to the point of necking, the true stress,  $\sigma_t$ , may be expressed in terms of engineering stress by:

$$\sigma_t = S(e + 1)$$

Up to the onset of necking, the true strain,  $\epsilon_t$ , may be determined from the engineering strain, by:

$$\epsilon_t = \ln(e + 1), \text{ where } \ln \text{ is the natural log.}$$




Figure 2. Comparison of engineering and true stress-strain curves [from the *Atlas of Stress-Strain Curves*].

Beyond the point of maximum load (necking), the true strain is based on the actual current area,  $A$ , and is expressed:

$$\epsilon_t = \ln (A_0/A)$$

and the true stress is based on the load and actual current area:

$$\sigma_t = P/A$$

At the point of fracture, the true strain ( $\epsilon_t$ ) and true stress ( $\sigma_t$ ) are thus expressed:

$$\epsilon_t = \ln (A_0/A_f) \quad \text{and} \quad \sigma_t = P_f/A_f$$

where  $A_f$  is the area at fracture and  $P_f$  is the load at fracture.

The true stress-strain curve beyond the onset of necking (the maximum load or uniform strain limit) is further complicated by the development of radial and hoop stresses in the necking region. The average axial or nominal stress given by  $\sigma_t = P/A$  is not the true equivalent uniaxial stress because the hoop and radial stresses are not zero. Beyond necking, the nominal stress is often corrected to get the true equivalent uniaxial stress using a Bridgman Correction factor (Reference 25), which is dependent upon specimen geometry (corrected shape in Figure 2).

### 3.2 Typical Variations In Material Properties

Reference 26 provides indications (also listed in Table 1 of this report) of the variation in material property data typically achieved in standard engineering tests [i.e., results of identical tests performed on identical specimens fabricated from one specific heat<sup>1</sup> and material

<sup>1</sup> Heat – A unique identifying number assigned to the product of one furnace melt.

specification (plate, bar, etc.)). These typical coefficients of variation basically indicate that even quasi-static tensile testing, which is considered more controllable than dynamic material impact testing, still has a considerable variation in the test data results. Variation is also expected from the dynamic impact tests. However, if dynamic material impact testing can be limited to this same level of variation or less (as evidenced by the accuracy of the analysis comparisons), the impact test results should be considered just as valid and useable as other standard material property test results.

Table 1. Typical Coefficients of Variation.

Variable	Typical Variance (%)
Yield strength of metals	7
Ultimate strength of metals	5
Modulus of elasticity of metals	5
Fracture toughness of metals	15
Tensile strength of welds	10

### 3.3 Strain Rate Effects

Strain rate is simply the time rate of straining. Average strain rates for most quasi-static tensile tests range between  $10^{-2}$  and  $10^{-5}$  per second. Greater strain rates ( $10^{-1}$  to  $10^2$  per second and higher) are considered dynamic tests. The literature generally indicates that steels experience an elevation or increase in strength relative to the quasi-static stress level as the strain rate increases; however, the amount of increase is not well defined for many materials and strain rates. The effects of strain rate are generally expressed as an increase in strength relative to the quasi-static stress-strain curve. This report focuses on determining and justifying true stress-strain curves that reflect strain rates ranging from 4 to 40 per second (nominally) for both 304L and 316L base and welded material at varying temperatures.

### 3.4 Recognition of Energy Density in True Stress-Strain Curves

It is recognized that the “area under the [engineering] stress-strain curve represents energy absorption per unit volume ... of material” (Reference 27). When looking at the associated true stress-strain curve, the significant area under the curve beyond the uniform strain limit might be interpreted as implying that a substantial amount of energy absorption capacity is available. However, this ‘area’ is actually strain energy density (e.g., in.-lb./in.<sup>3</sup>) and the energy absorption capacity is a function of the volume of material straining. Figure 3 shows an engineering stress-strain curve, the corresponding true stress-strain curve, and the uniform strain limit for a representative stainless steel at room temperature. Figure 4 illustrates the volume of stainless steel material being strained during a typical tensile test with respect to the uniform strain limit. Up to the uniform strain limit, the full volume of material in the test specimen gauge (or reduced area) length is being uniformly strained. Past the uniform strain limit, the material starts to neck, involving only a local and greatly reduced amount of material volume contained within the necking region. In fact, the volume of material involved is constantly decreasing as the strain increases. Figure 4 illustrates that the area under the true stress-strain curve beyond the uniform strain limit cannot be used to determine energy absorption capacity without knowing just how much material volume is being locally strained. Viewing the true stress-strain curve from this ‘energy density’ perspective is necessary in understanding that there really is only a

limited amount of energy absorption capacity remaining in the material strained beyond the uniform strain limit. The constantly reducing volume of involved material is also the reason why it is difficult to quantify strain rate effects beyond the uniform strain limit. Since this report is attempting to generate data to support engineering design efforts, limiting the test strains to below the uniform strain limit (the onset of necking) is reasonable.




Figure 3. Representative engineering and true stress-strain curves.




Figure 4. Representation of material volume involved in straining during tensile testing.



## 4. METHODOLOGY

In order to generate a true stress-strain curve at a specified strain rate, classical engineering mechanics (incorporating conservation of energy and momentum methods) are used. A drop weight device is employed to perform dynamic impact testing at increased strain rates. The energy transferred to the test specimen and causing the tensile deformation is determined assuming an inelastic collision of the drop weight and the test specimen holder. Some kinetic energy is lost during the inelastic impact. This loss is accounted for by applying the reduced velocity predicted by conservation of momentum theory to the combined mass of the drop weight and impact driver when determining the total energy imparted to the test specimen. Energy losses due to other causes are considered negligible.

A ‘total impact energy method’ was developed using the concept that the area under a true stress-strain curve (up to the uniform strain limit – the strain at the onset of necking) is equivalent to the amount of energy (strain energy) that the test specimen gauge length volume of material can absorb up to a specific strain level achieved in the material. The total impact energy method develops a strain rate elevated true stress-strain curve by multiplying each stress point (beyond yield) on the quasi-static curve by a constant (referred to as the ‘factor’). This factor is the ratio of the impact energy (as described above) absorbed by the test specimen’s gauge length volume (e.g., in.-lb./in.<sup>3</sup>) divided by the area under the quasi-static true stress-strain curve up to the true strain achieved in the impact test specimen. Elevated true stress-strain curves (reflecting strain rate effects) produced using this methodology are considered valid up to the uniform strain limit of the material. Future testing and analysis efforts may better define the curve between the uniform strain limit and failure.

Incorporating the total impact energy method described above, the strain rate elevated true stress-strain curve development effort used a multi-step process summarized as follows:

1. **Perform Quasi-Static Tensile Testing – Section 5:** Obtain the quasi-static true stress-strain curve(s) of the material(s) under consideration. This is the basis for characterizing the strain rate effects.
2. **Perform Impact Tensile Testing – Section 6:** Determine the response of test specimens subjected to impact tensile testing. This effort provides the test specimen strain, at a specific rate, when subjected to an impact from a known weight dropped from a known height, in a known test apparatus.
3. **Quantify Strain Rate Effects – Section 7:** Calculate the factors, establish a reasonable curve fit, and then generate the associated strain rate elevated true stress-strain curves. This effort invokes the methodology briefly described above. Once the factor (for a specific strain rate at a specific temperature) is calculated, the quasi-static curve is used to generate the elevated true stress-strain curve.
4. **Perform Analyses for Data Validation – Section 8:** Using the strain rate elevated true stress-strain curves as input, perform FEM inelastic analyses of the impact tensile tests to determine the validity of the test data. This effort uses the strain rate data generated to determine its computational effectiveness.

The following sections address each of the above four steps in greater detail.

## 5. QUASI-STATIC TENSILE TESTING

Strain rate effects are best characterized by comparison to quasi-static tensile test results, expressed as true stress-strain curves. Although 304L and 316L stainless steel materials have been studied for many years and by numerous investigators (e.g., References 28-31), relatively little recent, representative data reflecting current commercial supplier practices and dual-stamping is readily available to practitioners. Even less data expressed as true stress-strain relationships out to failure can be found in the literature. Accurate quasi-static stress-strain data were needed for each unique material heat at each temperature to be investigated. To support this moderate strain rate research, considerable quasi-static tensile testing was required. This testing has resulted in a significant amount of basic material data recorded as engineering stress-strain curves and their converted true stress-strain relationships. The results of this quasi-static testing effort, besides establishing an initial basis for the strain rate research, has the added benefit of contributing to the existing data pool for these materials and make the data more readily available to other researchers, engineers, and interested parties. Typical stress-strain values are often of interest for failure analyses and integrity evaluations associated with low probability, extreme loading conditions. Reference 32 provides additional information on this quasi-static testing effort, including resulting material yield strength, ultimate strength, ultimate strain, fracture strength, fracture strain and reduction in area values.

The quasi-static tensile testing discussed herein was conducted in accordance with ASTM procedure A 370 on 304L and 316L stainless steel plate materials at temperatures ranging from -20 °F to 600 °F. Two plate thicknesses, ten different material heats, and both base and welded material were investigated. NSNFP test personnel were present at all quasi-static tensile testing efforts even though INL personnel were performing the actual testing. This provided important background knowledge and confirmed that the testing performed was appropriate.

Efforts to conduct strain rate research began in 2004 and tensile material properties were required of the materials being investigated. Since the initial effort was to consider room temperatures only, quasi-static tensile testing of 304L and 316L specimens was performed by the Materials Testing Laboratory located at the central facilities of the INL (at CFA-602). This laboratory had National Voluntary Laboratory Accreditation Program (NVLAP) credentials. Because this report incorporates some of the impact testing performed in the initial effort, this report will describe the various aspects addressed in the quasi-static tensile testing for that material. This quasi-static tensile testing effort will be referred to as 'Initial Testing'.

After the initial phase of impact testing, funding was provided to impact test base and welded materials at varying temperatures. Because this called for many impact tests and different test specimen geometries, additional plate material was procured and more quasi-static tensile testing was required. This set of quasi-static tensile testing was performed at the INL Research Center (IRC), located in Idaho Falls, because of the need for testing at varying temperatures. Material tensile testing was performed at -20 °F, room, 300 °F, and 600 °F temperatures for both base and welded materials. Because the cognizant laboratory personnel at the IRC did not have NVLAP accreditation, personnel having NVLAP accreditation from the Materials Testing Laboratory provided oversight to assure compliance with ASTM A 370 requirements. This quasi-static tensile testing effort will be referred to as 'Additional Testing'.

## 5.1 Test Material Procurement

INL's Consumer Grade – Quality Level 4 process was used to procure test specimen materials. However, the materials were always stamped as ASME SA-240 (Reference 33) plate and the associated certified material test reports (CMTRs) were obtained. The alloys were procured from various manufacturers as 4-foot by 10-foot plates. The as-received plate material was hot rolled, annealed, and pickled (HRAP finish). For procurement confirmation purposes, the quasi-static tensile testing results were used to verify the CMTR data as well as determine the necessary stress-strain relationships.

### 5.1.1 Initial Testing

The initial material impact testing effort procured two ½-inch thick plates, one 304L and one 316L in FY04. The 304L plate had a heat number of 10W8 and the 316L plate had a heat number of 09T9. This material yielded pertinent impact test data that has been incorporated into this report.

### 5.1.2 Additional Testing

The additional material testing effort started in FY05 with the procurement of four plates of stainless steel material. In order to achieve strain rates below 16 per second, two ¼-inch thick plates were procured. The ¼-inch thick 304L plate had a heat number of 485896 and the 316L plate had a heat number of 48R8. In order to achieve strain rates above 16 per second, two ½-inch thick plates were procured. The ½-inch thick 304L plate had a heat number of 54M7 and the 316L plate had a heat number of 230468. These plates were designated to be used for both base and welded material impact testing. Having base and welded material test specimens from the same material heat would allow for more direct comparison of strain rate responses.

This material was to be handled by many individuals and different organizations while being machined, welded, examined, etc. In order to maintain proper traceability throughout this entire strain rate research effort, it was recognized that each plate material needed to be color coded upon receipt. Both the ¼-inch and the ½-inch 304L plate material were painted bright yellow while the ¼-inch and ½-inch 316L plate material were painted a light blue. Differentiation between the ¼- and ½-inch thick plates was easy so the same color was used for both thicknesses. Figure 5 shows a freshly painted ½-inch thick 304L plate and Figure 6 shows a freshly painted ½-inch thick 316L plate.



Figure 5. Painted ½-inch thick 304L plate.




Figure 6. Painted 1/2-inch thick 316L plate.

From the perspective of this research effort, the welding of the ‘bright yellow and light blue’ plates described above and the associated weld examination process also had to be procured. As previously described in Section 2.2.1, the INL provided these welding and weld examination services. The first step was to cut the 4x10-foot plate into a number of smaller plates that could be more easily handled during and after welding. Water jet cutting was used to accomplish this. Figure 7 pictures the 304L plate cut into 5-1/2-inch wide by 26-inch long (in the direction the plate was rolled at the mill) pieces. After weld preparations (Figure 8) were machined (Figure 9), plate pairs were welded together using a gas tungsten arc welding process (Figure 10) identified as Weld Procedure S2.0 (see Appendix C) from the INL Welding Manual (Reference 34). Unique plate numbers were written (in the format of WXX with a permanent marker) on the upper right hand corner of each welded plate to maintain traceability. For example, as seen in Figure 10, plate W16 was being welded (even though weld machine obscures the ‘W’). Weld wire ER308L was used for the 304L plate material and weld wire ER316L was used for the 316L plate material. As can be seen in Figure 10, a copper backing bar was used as well as a small run out tab at both ends (to get, as much as possible, a full plate length weld up to the plate edges). After completing the weld on the top side (painted side) with two weld passes, the reverse side of the weld was ‘back gouged’ (at an INL machine shop) to prepare that side for welding. Small imperfections at the surface where the copper backing bar was stationed needed to be removed with a small groove. The final single weld pass was made down that groove, completing the welded plate. Figure 11 shows a close-up view of the final weld appearance. The plate welds (also identified with the same unique plate number) were then radiographed and evaluated per INL procedure TPR 4970, Appendix C (Reference 35) to determine their acceptability. TPR 4970, Appendix C invokes ASME B&PV Code, Section III, Class 1 weld radiographic acceptance criteria. Figure 12 shows how the plates were marked for identification and weld examination. The radiograph film length was six inches. Measuring from the top (with the plate number in the upper right hand corner), the plate was marked to identify distances of 4 thru 10 inches and 16 thru 22 inches. These two 6-inch intervals captured the weld length that would become the gauge length for the welded impact test specimens. If additional weld material needed to be used for quasi-static tensile testing, additional radiography was performed on those additional weld lengths. At the completion of the INL effort, all welds passed the examination criteria, as evidenced by Reference 14.






Figure 7. 304L plates cut and weld prepared.




Figure 8. Example weld preparation profiles used on 1/4-inch and 1/2-inch thick plate material.




Figure 9. Close-up of machined weld preparation on 1/2-inch thick plate.




Figure 10. 304L plates being welded.






Figure 11. Close-up of final weld appearance on the welded plates.



Figure 12. Markings showing where radiographs were taken on the welded plate.

During the plate welding effort, NSNFP test personnel maintained two lists that noted the material and thickness of each marked plate, the plate's facility location (for tracking fabrication progress on top side welding, back gouging, reverse side welding, repairs, weld examination, etc.), and if the welded plate was used for quasi-static tensile test specimens, impact test specimens, or placed in storage for future use. Tables 2 and 3 show these 'final status' lists for the 304L and the 316L welded plates respectively. Obviously, during the fabrication and weld examination efforts, these tables were 'living documents', constantly being revised and updated but capturing plate specific details related to the welding and weld examination process. For example, a plate number in the format of WXX indicated that the plate was welded and passed the weld examination process on the first attempt. A plate number in the format of WXX R1 indicated that the welding was completed but did not pass the radiography examination on the first attempt and the weld needed to be repaired. After the weld repair effort was completed, the plate passed the weld examination effort. A plate number in the format of WXX-2 indicates that the initial weld did not pass the weld examination or the wrong weld wire was used and the entire weld needed to be removed and the process started over. In those cases, the welded plate was cut apart and the central portion of the plate (approximately a 1 to 1-1/4-inch wide strip including the weld) was removed. The long edges (that were the outside edges of the previous welded plate) were then weld prepped (using the same geometry as before) and the plates were then welded together again. As the final step, these plates and new welds were radiographed again for final examination. Tables 2 and 3 also indicate that certain 1/4-inch welded plates had an initial indication of a 'lack of penetration' (LOP). However, this finding was questioned by the INL weld engineer and upon reading further radiography efforts and evaluating cut, etched, and polished test coupons (from plate W57), these LOP indications were finally accepted.

The INL submitted their documentation (Reference 14) generated during this welding and weld examination process to the NSNFP. Included in that documentation were the laboratory notebooks, identification of weld process parameters, weld examination results, and the associated radiographs. Notes were made by INL personnel that additional radiography was performed but the reason was unknown. Actually, the additional radiographs were performed at the request of NSNFP test personnel to examine portions of welds that were to be used for quasi-static tensile testing. This data package is in the NSNFP document control center and can be reviewed for more detailed information. Note that the real purpose behind the weld documentation was to establish that the welds being tested (for either quasi-static tensile testing or impact tensile testing) were representative and acceptable with no significant flaws that could adversely affect either test results. Hence, welds meeting the radiographic requirements were mandated. Actual canister closure welds would only be welded from one side but the reverse side or back pass performed for this research effort was necessary to provide sufficient welded material for material testing purposes.

Since the welding of the above plates took a substantial amount of time, more plate material was procured in FY06. This plate material would permit an earlier onset of impact testing with new test specimen geometries. In addition, this material would yield a good estimate of the required amount of drop weight and drop height necessary to achieve the desired strain rates for the limited number of welded plate test specimens. Two 1/4-inch thick plates were procured (the 304L plate had a heat number of 64A1 and the 316L plate had a heat number of 76H3) and two 1/2-inch thick plates were procured (the 304L plate had a heat number of 72K9 and the 316L plate had a heat number of 67K0). These plates were initially referred to as our



‘auxiliary’ plate material but the test data from these plates were later deemed just as significant as any other. It also permitted the incorporation of more material heats into the resulting data collection.

Table 2. 304L welded plate material history.

Plate Identifier	Thickness (in.)	Current Location	Status	Designated Use
W1-2	½	613	Cut/Reweld/B-done/1W-done/TS	D-4-4/TS
W2-2	½	613	Cut/Reweld/Storage	
W3	½	613	Storage	
W4 R1	½	613	Weld Repair/Storage	
W5	½	613	B-done/TS	D-4-4/TS
W6	½	613	B-done/W-done	D-4-4
W7	½	613	B-done/W-done	D-4-4
W8	½	613	B-done/W-done	D-4-4
W9	½	613	B-done/W-done	D-4-4
W10	½	613	B-done/TS	D-4-4/TS
W11 R1	½	613	Weld Repair/Storage	
W12	½	613	Storage	
W13	½	613	Storage	
W14 R1	½	613	Weld Repair/Storage	
W15 R1	½	613	Weld Repair/Storage	
W16	½	613	Storage	
W49	¼	613	Storage	
W50*	¼	613	B-done/TS	D-2-2/TS
W51	¼	613	Storage	
W52*	¼	613	Storage	
W53	¼	613	Storage	
W54	¼	613	B-done/W-done	D-2-2
W55*	¼	613	Only weld specimens cut out	D-2-2
W56*	¼	613	B-done/W-done	D-2-2
W57*	¼	613	Partial cut/B-done/1W-done/TS	*S*/TS/D-2-2
W58	¼	613	Only weld specimens cut out	D-2-2
W59*	¼	613	B-done/W-done	D-2-2
W60-2	¼	613	Cut/Reweld/Storage	
W61	¼	613	Storage	
W62	¼	613	Storage	
W63	¼	613	B-done/W-done	D-2-2
W64*	¼	613	Storage	

Notes: B – Base material W – Weld material TS – Quasi-static tensile test specimen  
D-2-2 and D-4-4 are impact test specimen geometry designations  
\* - Initial lack of penetration (LOP) indication later acceptable  
\*S\* - Special Top Welded D-2-2 cut out with initial lack of penetration (LOP) for test/examination of circular coupons

Table 3. 316L welded plate material history.

Plate Identifier	Thickness (in.)	Current Location	Status	Designated Use
W17	½	613	B-done/TS	D-4-4/TS
W18	½	613	B-done/W-done	D-4-4
W19	½	613	B-done/W-done	D-4-4
W20	½	613	B-done/W-done	D-4-4
W21	½	613	B-done/W-done	D-4-4
W22	½	613	Storage	
W23	½	613	B-done/W-done/TS	D-4-4/TS
W24	½	613	Storage	
W25	½	613	Storage	
W26	½	613	Storage	
W27	½	613	Storage	
W28	½	613	B-done/1W-done/TS	D-4-4/TS
W29	½	613	Storage	
W30	½	613	Storage	
W31	½	613	Storage	
W32 R1	½	613	Weld Repair/Storage	
W33	¼	613	Storage	
W34*	¼	613	B-done/W-done	D-2-2
W35 R1	¼	613	Weld Repair/Storage	
W36*	¼	613	B-done/W-done	D-2-2
W37-2	¼	613	Cut/Reweld/Storage	
W38*	¼	613	Storage	
W39*	¼	613	Only weld specimens cut out	D-2-2
W40*	¼	613	Storage/TS	D-2-2/TS
W41*	¼	613	Only weld specimens cut out	D-2-2
W42*	¼	613	Storage	
W43-2	¼	613	Cut/Reweld/Storage	
W44*	¼	613	B-done/W-done	D-2-2
W45*	¼	613	B-done/W-done	D-2-2
W46	¼	613	Storage	
W47	¼	613	Storage	
W48*	¼	613	B-done/1W-done/TS	D-2-2/TS
Notes: B – Base material W – Weld material TS – Quasi-static tensile test specimen D-2-2 and D-4-4 are impact test specimen geometry designations * - Initial lack of penetration (LOP) indication later acceptable				

## 5.2 Quasi-Static Test Specimen Geometries

The quasi-static tensile test specimens, satisfying the geometry recommendations of ASTM A 370, were either rectangular or round in cross section, depending on the test machine used to perform the tensile testing.

### 5.2.1 Initial Testing

The test specimen geometry used in the initial phase of quasi-static tensile testing was the ‘Sheet-Type Subsize Specimen’ identified in ASTM A 370. Figure 13 illustrates the rectangular cross section test specimen geometry used. Smaller pieces of material from the full 304L and 316L plates were cut out and marked (using a permanent marker) indicating the material type, heat number, and the longitudinal (parallel to) or transverse (perpendicular to) orientation with respect to the mill rolling direction of the plate. These smaller pieces were provided to the INL’s machine shop to fabricate the test specimens. Only one smaller piece at a time was provided to the machine shop in order to assure proper distinction between 304L and 316L material and the proper plate orientation (longitudinal or transverse). Only after each batch of test specimens were received back from the machine shop would another plate be provided for more test specimen fabrication efforts. These specimens, marked with an alphanumeric identifier, material type, heat number, and orientation (transverse or blank indicating longitudinal), did not receive any heat treatments following this machining.




Figure 13. Sheet-type subsize specimen geometry used for initial quasi-static testing.

### 5.2.2 Additional Testing

A slightly different approach from that described for the initial testing was taken to fabricate quasi-static tensile test specimens for the additional testing, including the use of round test specimens. However, only longitudinal oriented (parallel with the plate rolling direction) test specimens were fabricated from these plates. The ‘auxiliary’ base material blanks were water jet cut from the large 4x10-foot plates (see the smaller rectangular cutouts in Figures 14 and 15). Color coding occurred as these water jet cut blanks were removed from these plates. Figures 14 and 15 show the color coding used for both thicknesses of the 304L material (orange) and the 316L material (purple), respectively. The color coding was used for material traceability.




Figure 14. Auxiliary 304L plate with orange paint coding.



Figure 15. Auxiliary 316L plate with purple paint coding.

The base material blanks from the welded plates material (heat numbers 485896, 48R8, 54M7, and 230468) described in the previous section were cut from the remaining 'drop' material produced when cutting the 5-1/2-inch by 26-inch long plates from the 4 by 10-foot plates. These smaller drop plate pieces were marked to establish the longitudinal direction (parallel with the plate's rolling direction) and the material type and heat number. These smaller pieces of drop material from the 304L and 316L plates were then provided to the INL's machine shop to fabricate the test specimens. Only one smaller plate at a time was provided to the machine shop in order to assure proper distinction between 304L and 316L material. Only after each batch of test specimens were received back from the machine shop would another plate be provided for more fabrication efforts.

The welded material blanks were cut from the welded plates. Figures 16 and 17 below show examples of welded material blanks water jet cut from the welded plates. As was done for the base material, the welded blanks (identified by color and thickness) were controlled by NSNFP test personnel so as to maintain full traceability between 304L and 316L welded material during test specimen fabrication.



Figure 16. Both 304L and 316L 1/2-inch thick welded plates showing welded material blanks (attached to impact test specimens, which were later removed).





Figure 17. 316L 1/4-inch thick plate showing welded material blank cut from the plate (also attached to impact test specimen, which was later removed).

With the base and welded material blanks or smaller drop plate pieces (obtained as described above), it was now possible to fabricate the test specimens. The test specimen geometry depended upon the plate thickness. The 1/2-inch thick plate blanks were machined into round, small-sized 0.350 inch diameter test specimens as shown in Figure 18. Specimen blanks from the 1/4-inch thick plates were machined into round, small-sized 0.160-inch diameter specimens as shown in Figure 19. The test specimens were marked (with a unique numerical identifier) to indicate material type, heat number, and whether the test specimen was either base or welded material. The ends of the specimens outside of the gauge length were threaded to match the holders on the tensile test machine. Test specimens were not heat treated following machining.




Figure 18. Round, small-size 0.350-inch diameter quasi-static tensile test specimen used for additional testing of 1/2-inch thick plate.




Figure 19. Round, small-sized 0.160-inch diameter quasi-static tensile test specimen used for additional testing of 1/4-inch thick plate.

### 5.3 Quasi-Static Test Procedure

The quasi-static tensile testing procedures followed the requirements of ASTM A 370 except for the stated reporting requirements. A total of three tensile tests for each material heat (and each orientation if applicable) at each temperature under consideration, for base and welded material as applicable, were performed. The quasi-static tensile testing was performed by INL personnel with NSNFP test personnel in attendance to assure acceptable data was obtained.

#### 5.3.1 Initial Testing

Quasi-static tensile tests were performed at the Materials Testing Laboratory on February 19 and June 3, 2004. Mr. H. C. Bean, laboratory manager, conducted the tests using the Tinius-Olsen tensile test machine (see Figure 20). Test specimen dimensions were measured by the INL laboratory manager prior to tensile testing. NSNFP test personnel also recorded continuous force-displacement data for the entire duration of each test (load initiation through specimen failure). Both longitudinal and transverse oriented test specimens were tensile tested. The flat ends of the test specimens were inserted into plate grippers as seen in Figure 21. The strain rate achieved during the quasi-static tensile testing was approximately  $10^{-3}$  to  $10^{-4}$  per second.



Figure 20. Tinius-Olsen tensile test machine at the Materials Testing Laboratory.





Figure 21. Close-up of inserted test specimen with extensometer attached.

### 5.3.2 Additional Testing

Conventional quasi-static tensile tests were performed at -20 °F, room temperature, 300 °F, and 600 °F using an Instron Model 4505 universal testing machine with a maximum capacity of 22,000 pounds. Prior to tensile testing, each specimen was dimensionally checked. Those measurements were used in the determination of the test specimen's cross sectional area or were within acceptable tolerances to use the nominal cross sectional area. These round specimens were gripped in threaded connectors, aligned using pins and clevises, and loaded by a crosshead displacement that produced strain rates of approximately  $10^{-3}$  to  $10^{-4}$  per second. Force-displacement output was continuously recorded to specimen failure. LabVIEW 7.0 (Reference 36) software was used to record and display the specimen temperature, force-displacement and engineering stress-strain data, and write the data to an Excel file for evaluation.

The -20 °F temperature testing was achieved using an insulated 'coldbox' surrounding the test specimen and extensometer. Liquid nitrogen was fed into the coldbox to achieve the desired initial temperature conditions of -20 °F and to maintain those conditions throughout the tensile test. Figure 22 shows this test setup, including the three thermocouples attached to the test specimen.

For the room temperature tests, displacements were measured over the specimen gauge length using an extensometer. Figure 23 shows a room temperature tensile test setup with the extensometer in place. The room temperature within the laboratory was considered acceptable for temperature conditions.



Figure 22. Cold (-20°F) temperature tensile testing setup.



Figure 23. Room temperature tensile testing setup at IRC.

For the 300 °F and 600 °F tests, an electric furnace enclosure was used. Because the extensometer was not compatible with the temperatures inside the furnace enclosure, two coupled, vertical rods, one on each side of the test specimen and attached to the upper specimen holder, transferred the deformation in the gauge length to an extensometer and linear variable displacement transducer (LVDT) located outside of the furnace. The elevated temperature tensile test setup is shown in Figure 24 with the furnace open to show the specimen with transfer rods, extensometer, and LVDT. Figure 25 shows an alternate furnace setup that was also used.



Figure 24. Elevated temperature tensile testing setup with furnace door open.



Figure 25. Alternate furnace setup used with door closed.



## 5.4 Achieving Temperature Conditions

Due to the importance of achieving the proper temperatures for the quasi-static tensile testing performed, this section provides more details of how and what temperatures conditions were achieved during the quasi-static tensile testing.

### 5.4.1 Initial Testing

The quasi-static tensile testing at the Materials Testing Laboratory was performed in a typical office/laboratory conditioned environment at ambient room conditions, meaning that the temperature was in the 65 °F to 75 °F range per the laboratory manager, Mr. H. C. Bean. This falls within a very close  $71\text{ °F} \pm 6\text{ °F}$  temperature range. Hence, these test specimen temperatures were acceptable.

### 5.4.2 Additional Testing

For the room temperature tests, test specimen temperatures were not directly measured for every test. Recorded data for three different days from the INL Lab Notebook LAB-771 (that documented the quasi-static tensile testing at the IRC) indicates the room temperature ranged from 72 to 74 °F on those days. The engineer in charge of the tensile testing (Mr. R. Lloyd) indicated that the anticipated worst temperature range in the IRC labs would be 68 °F to 77 °F, also within a very close  $71\text{ °F} \pm 6\text{ °F}$  temperature range. These room temperature conditions were acceptable.

Test specimen temperatures other than room conditions (while in a furnace or cold box) were monitored throughout each tensile test cycle using temperature data from three thermocouples attached (thin wire spot-welded) to each test specimen. Figure 26 shows three thermocouples attached at the top, middle, and bottom locations of the gauge length of a typical round test specimen machined from ½-inch thick plate.




Figure 26. Thermocouples attached to a quasi-static tensile test specimen.

Prior to beginning the quasi-static tensile test, the test specimens were allowed to uniformly 'soak' to the designated test temperature. Because of the small size of the test specimens, automated methods of preheating, pre-cooling and holding the specimens at appropriate conditions during the quasi-static tensile test were not employed. The three temperatures from the test specimen were monitored by the laboratory technician and furnace or cold box temperatures were manually adjusted as necessary so that variations in test specimen temperatures were minimized as much as possible.

The temperature data for the three thermocouples were recorded. These three temperatures were averaged and then the maximum and the minimum average temperatures achieved during the quasi-static tensile testing (from test start to test specimen fracture) were established. These data are included in files 1A and 1B on the Report CD-R (Reference 37) included with this report. Laboratory Notebook NSNF/SN/04.01 (hereafter referred to as NSNFP Lab Notebook) Binder Volume 6 (Reference 38) contains a summary of these maximum and minimum average temperatures for the quasi-static tensile tests that were actually used as a basis for establishing the true stress-strain properties for that specific material and temperature conditions. Table 4 summarizes the test specimen temperature ranges measured during the 2006 and 2007 quasi-static tensile testing of 304L and 316L stainless steels performed by the INL for the NSNFP.

Table 4. Worst case average temperature ranges achieved during quasi-static tensile testing.

<b>Desired Test Temperature</b>	<b>Worst Case Average Temperature Range Achieved</b>
-20 °F	+14/-5 °F
300 °F	+13/-8 °F
600 °F	+18/-19 °F

Table 4 temperature ranges are the worst-case ranges for all basis testing completed. Most of the tests had temperatures much closer to the desired test temperature. However, even these worst-case temperature ranges are considered acceptable. The data on the Report CD-R contains the maximum and minimum average temperatures for the entire test interval, including the onset of necking through final rupture. However, the material impact testing effort only used data up to the onset of necking. Therefore, the temperature ranges listed above in Table 4 reflect the averaged maximum and minimum temperature ranges achieved up to the onset of necking.

The upper worst-case temperature range result for the -20 °F testing (+14 °F) appears somewhat high until it is recognized that the test specimens actually get hotter during tensile testing near the middle location due to the material straining. This fact is clearly illustrated by the plot of measured test specimen temperatures for Test #141, a 316L material test (see Figure 27). Even though the environment was kept close to -20 °F (as evidenced by the top and bottom temperatures), the middle temperature rose much higher (due to material straining), significantly affecting the average temperature results. However, this is the physical phenomenon that occurs during actual drop events. The important fact is that the starting (soaked) temperature and the sustained environment (represented by the larger end portions of the test specimen) adequately

reflect the  $-20^{\circ}\text{F}$  conditions. Heat generation also occurs in the higher temperature tensile tests but the temperature values are more removed from zero, making the effects less noticeable.




Figure 27. Data from three thermocouples showing how middle temperature rises due to straining during  $-20^{\circ}\text{F}$  tensile test.

## 5.5 Quasi-Static Tensile Testing Results

A significant number of quasi-static tensile tests (over 160) were performed in order to support the material impact testing effort discussed herein. As previously mentioned, three tensile tests for each material heat (and each plate orientation if applicable) at each temperature under consideration for both base and welded material were performed. One test was chosen from each set of three tests that best represented typical results. Because the number of tests for a particular set of conditions was limited to three, statistical methods were not employed to establish an average or mean result. All three tests were plotted as engineering stress-strain curves and the most representative specimen result was chosen for the applicable set of conditions.

Tables 5 and 6 list the specific quasi-static tensile test specimen identifier (of the three tests performed) chosen to represent that specific material condition. These tables (with background colors matching the material color coding) include both 'initial' and 'additional' testing. Tables 5 and 6 address only the longitudinally oriented test specimen results. In the tables, 'NA' indicates that quasi-static testing was not performed for those specific conditions. As earlier discussed, transverse oriented test specimens (perpendicular to the plate's rolling direction) were also quasi-static tensile tested at the INL Materials Testing Laboratory under room temperature conditions. These transverse test results were performed to help assess any potential response variation in test specimens reflecting varying orientation with respect to the direction that the plate material was rolled at the mill. Section 6.6.2 contains specific

information regarding the comparison of results between longitudinal and transverse oriented quasi-static and impact tensile tests.

Table 5. Identification of 304L tensile tests establishing basis true stress-strain curves.

Material	Thickness	Heat	Type	Temp. (F)	Specimen Used
304L	1/2-inch Thick	10W8	Base	-20	NA
				Room	MT03
				300	NA
				600	NA
			Weld	-20	NA
				Room	NA
				300	NA
				600	NA
		54M7	Base	-20	156
				Room	153
				300	146
				600	151
			Weld	-20	183W
				Room	177W
				300	171W
				600	179W
		72K9	Base	-20	106
				Room	113
				300	109
				600	111
			Weld	-20	NA
				Room	NA
				300	NA
				600	NA
	1/4-inch Thick	64A1	Base	-20	13
				Room	2
				300	5
				600	8
			Weld	-20	NA
				Room	NA
				300	NA
				600	NA
		485896	Base	-20	55
				Room	47
				300	50
				600	53
			Weld	-20	81W
				Room	73W
				300	74W
				600	77W

Table 6. . Identification of 316L tensile tests establishing basis true stress-strain curves.

Material	Thickness	Heat	Type	Temp. (F)	Specimen Used
316L	1/2-inch Thick	09T9	Base	-20	NA
				Room	MT12
				300	NA
				600	NA
			Weld	-20	NA
				Room	NA
				300	NA
				600	NA
		67K0	Base	-20	125
				Room	128
				300	120
				600	123
			Weld	-20	NA
				Room	NA
				300	NA
				600	NA
		230468	Base	-20	141
				Room	138
				300	131
				600	135
			Weld	-20	181W
				Room	167W
				300	161W
				600	164W
	1/4-inch Thick	48R8	Base	-20	42
				Room	33
				300	35
				600	37
			Weld	-20	84W
				Room	63W
				300	66W
				600	68W
		76H3	Base	-20	26
				Room	17
				300	19
				600	22
			Weld	-20	NA
				Room	NA
				300	NA
				600	NA



The quasi-static tensile tests were performed through the entire material response range, out to failure. However, because the impact testing only needed to use data out to the uniform strain limit (the strain at the onset of necking), no Bridgman correction was applied to this data.

### **5.5.1 Initial Testing**

Quasi-static tensile tests were performed on 304L and 316L plate material at the INL Materials Testing Laboratory following the tension test guidelines of ASTM A 370. Three test specimens were tested for each material and MT03 was chosen as typical for the 304L material and MT12 was chosen as typical of the 316L material. The quasi-static engineering and true stress-strain curves developed from these tests are contained on the Report CD-R, files 1A and 1B.

### **5.5.2 Additional Testing**

Quasi-static tensile tests were performed on 304L and 316L plate material at the INL Research Center following the tension test guidelines of ASTM A 370. Three test specimens were tested for each material and Table 5 shows the tests chosen as typical for the 304L material and Table 6 shows the tests chosen as typical of the 316L material. The quasi-static engineering and true stress-strain curves developed from these tests are contained on the Report CD-R, files 1A and 1B (data is too voluminous for incorporation into the pages of this report).

For a typical test sequence, specimens were tensile tested to failure with the specimen beginning to 'neck' after achieving the maximum engineering strength followed shortly by fracture of the specimen. As the neck progressed to failure, non-uniform geometry altered the uniaxial stress state to a complex stress state involving shear components as well as normal stresses. Specimens typically failed in a combination of shear and tensile 'cup and cone' geometry characteristic of ductile materials and is illustrated in Figure 28. The tensile testing of welded material test specimens was performed identical to the base material tests. However, the welded material specimens looked unique when compared to the base material specimens shown in Figure 28. Figure 29 shows a close-up picture of a welded material test specimen where the surface roughness (known as the 'orange peel' effect due to the straining of the irregular grain structure resulting from the welding) can be seen.



Figure 28. 'Cup and cone' type failure at fracture point.



Figure 29. Welded quasi-static tensile test specimen showing 'orange peel' effect.

In order to substantiate the quasi-static stress-strain curves determined herein, an attempt was made to locate similar data published in the literature. A limited amount of comparative data was found for 304 stainless steel and is illustrated in Figure 30 as true stress-strain terminated at or near the uniform stress limit. Figure 30 shows a combination of data from two sources (identified in the legend and most likely from different material heats) taken from Reference 24 at the temperatures indicated. Figure 31 shows the true quasi-static stress-strain curves derived herein from a single material heat (54M7) for the four temperature conditions considered in this investigation. These results are typical and representative of the quasi-static tensile test finding for the 304L plate material. The Figure 31 curves have been terminated at their uniform strain limits.

By comparing Figures 30 and 31, one can see great similarities, even though different material heats are involved. Stress and strain magnitudes are comparable as are the data trends. The true stress-strain curves lower with increasing temperature in both figures. The uniform strain limit reduces with increasing temperatures above room temperature (the 600 °F curve from the published data is from a different source than the other three curves so a definitive conclusion cannot be made there but this conclusion is true for the 300 °F data). The shape of the 304 and 304L true stress-strain curves at cold temperatures take on a unique shape. (For the most part, this phenomenon was not observed in the 316L material tested although a small, but noticeable hump was noted for one 316L base material specimen tested at -20 °F.) Therefore, the quasi-static true stress-strain curves obtained herein are indeed acceptable and representative. The testing methodology used herein was appropriate.

Finally, establishing the uniform strain limits (in terms of true strain) was important in order to establish the acceptable limits for the impact testing. Based on the quasi-static testing performed, Table 7 indicates the uniform strain limits established for both the base and welded materials at the temperatures considered. Background colors match the color coding used for each material.




Figure 30. Stress-strain curves from published 304 data in *Atlas of Stress-Strain Curves*.




Figure 31. 304L stress-strain curves determined from INL testing discussed herein.

Table 7. Uniform strain limits established for base and welded materials.

Thickness (inches)	Material	Heat #	Temp. (°F)	Uniform Strain Limit	
				Base	Weld
1/4	304L	485896	-20	0.38	0.47
			70	0.46	0.39
			300	0.30	0.21
			600	0.25	0.23
		64A1	-20	0.44	-
			70	0.52	-
			300	0.31	-
			600	0.29	-
	316L	48R8	-20	0.51	0.42
			70	0.46	0.40
			300	0.28	0.25
			600	0.27	0.24
		76H3	-20	0.52	-
			70	0.48	-
			300	0.31	-
			600	0.28	-
1/2	304L	54M7	-20	0.36	0.38
			70	0.48	0.37
			300	0.32	0.21
			600	0.29	0.20
		72K9	-20	0.36	-
			70	0.50	-
			300	0.31	-
			600	0.28	-
		10W8	70	0.42	-
	316L	230468	-20	0.40	0.37
			70	0.46	0.35
			300	0.30	0.25
			600	0.28	0.20
		67K0	-20	0.43	-
			70	0.41	-
			300	0.27	-
			600	0.25	-
		09T9	70	0.40	-

## **6. IMPACT TENSILE TESTING**

The data needed to justify the definition of appropriate material responses considering strain rate effects was obtained using an impact test device. This impact test device (identified as the Impact Testing Machine or ITM) is a drop weight tower that drops a large weight from an identified height onto a test fixture that holds the material test specimens. By using a combination of different weights, different drop heights, and different specimen sizes, varying strain levels and strain rates can be achieved in the test specimens.

### **6.1 Impact Testing Machine**

The goal behind the design of the ITM was to create a test device that could be used to investigate strain rate responses of multiple materials (including steels, plastics, etc.) as well as structural impact responses of actual components (e.g., impact limiter material or concrete specimens) small enough to fit within the ITM. Hydraulic-based systems could evaluate constant strain rate effects (with certain limits on strain rates and test specimen sizes) but the representation of an actual drop event or vehicle crash would not be present. Rather than pursuing a costly hydraulic-based test apparatus design with a smaller test specimen capacity, the ITM could test larger-sized material specimens and accurately reflect true impact loads (energy-limited events) with dropped weights nearing 1600 pounds (or more if needed). From the instant of impact during a drop event or crash, energy dissipation occurs. This is a response that is important to reflect during these material impact tests so that proper insights into analysis methodologies can be obtained.

During FY03, the ITM (see Figure 32) was taken from the conceptual design stage to an assembled test apparatus. The basic concept behind the ITM design is that of a falling weight (from heights up to 13 feet) impacting a test specimen resulting in permanent deformation. The falling weight is controlled within a vertical tower, while the loading on the test specimen is controlled by way of a specimen holder or test fixture. The ITM consists of a drop tower (including base plate, structural tube framing and vertically slotted pipe with an attached hoist), a drop weight (sometimes referred to as a 'pig') which was fabricated from 14-inch diameter bar stock, a drop hook (electronic quick release mechanism), and the test fixture. The drop weight (pig) is designed so that cylindrical disks of pre-determined weight can be added to vary the total magnitude of the dropped weight. The ITM is approximately 23 feet tall.

Major components of the tower, the drop weights, and the test specimen fixtures (that hold the test specimens) were fabricated. Figure 33 shows these components, ready for assembly. Figure 34 illustrates the three specimen fixtures or holders (for shear, bending, and tension loading investigations) that were also fabricated. The bending fixture has had limited test use (not discussed herein) and the shear specimen fixture has not yet seen service.

The major components of the drop tower were then assembled. A four-sided enclosure was also fabricated with the goal to confine potential fragments and reduce noise concerns for the NSNFP test personnel. The assembled ITM (see Figure 35) was then located at the INL's Reactor Test Complex, in the Materials Test Reactor building. Later on, the ITM was moved to another location at the INL. See Section 6.6.1 for additional information.






Figure 32. Conceptual design of Impact Testing Machine (ITM).



Figure 33. ITM fabricated components ready for assembly.



Shear specimen fixture



Bending specimen fixture



Tension specimen fixture

Figure 34. Three existing ITM test specimen fixtures.




Figure 35. Installed Impact Testing Machine at the Material Test Reactor building.



Drop weight impact test machines are typically used for compression testing. To carry out a tensile test under moderate strain rates, the ITM incorporated a special tensile holding fixture. The tensile test fixture (previously pictured in Figure 34) is illustrated in Figure 36 with specific components identified. The tensile test specimen fixture consists of a support stand and an impact driver. The support stand is made up of a bottom plate that bolts to the ITM base plate, two vertical legs, and an upper cross-member. The impact driver consists of a top (impact surface) plate, four vertical legs, and a lower cross-member. All structural members of the impact driver and support stand were fabricated from solid bar and plate carbon steel. The impact driver is connected to the support stand through the pinned ends of the dog bone-shaped tensile test specimen. During a test, the dropped weight impacts the impact driver on the tensile test fixture that is centered under the slotted guide pipe. The dropped weight contacts the impact driver, transferring its kinetic energy to the test specimen by way of the lower cross-member. The impact force is applied to the lower end of the tensile test specimen and is reacted through the upper end of the specimen into the upper cross-member of the fixture support stand. The pinned ends on the test specimen provide for pure tension loading of the specimen. [Note that in October-November 2005, the lower cross-member's top surface was notched (approximately 1-1/4-inch wide and 1-1/2-inch deep, both front and back) so that the bottom punch marks on the original test specimens could be visually seen.]




Figure 36. ITM tensile test specimen fixture.

## 6.2 Test Specimens

The initial phase of material impact testing performed in FY04 and continuing into FY05 was performed at room temperature conditions and used a dog bone-shaped test specimen as identified in Figure 37 (geometry ‘A’). Note that the test length includes the gauge length (uniform cross section along entire length) as well as both end transitions. All test specimens were water jet cut from their respective plates (see Section 5.1.1). Color coding, applied when multiple test specimens were cut from the entire plate at one time, was used to differentiate not only 304L or 316L material but was also used to differentiate longitudinal versus transverse oriented test specimens. Figure 38 illustrates the 304L plate already water jet cut and painted. Red indicated longitudinal test specimens, yellow indicated transverse test specimens, and green indicated shear blank specimens (for future use). Figure 39 illustrates the 316L plate already water jet cut and painted. Black indicated longitudinal test specimens, white indicated transverse test specimens, and brown indicated shear blank specimens for future use.




Figure 37. Tensile test specimen ‘A’ for initial material impact testing.

This initial testing allowed the NSNFP test personnel to gain valuable insights regarding the response of this test specimen geometry. Basically, this geometry worked well with strain rates in the 25 per second range but attempts to get strain rates of 10 per second produced low strains that were difficult to measure. Continuing material impact testing performed in FY05 permitted the testing and evaluation of many geometries. The results indicated that the test specimens needed to be long enough to avoid substantial end-effect complications yet be short enough to achieve the highest strain rates possible. Strains also needed to be large enough to be measurable by a high-speed digital camera. Low strains (near the yield strength) were difficult to measure (pixel limitations of the camera) and they did not clearly define a reasonable amount of area beneath the stress-strain curve (energy absorbed by the material). Larger strains were easier to measure and produced reasonable amounts of area beneath the stress-strain curve.



Figure 38. Initial 304L plate water jet cut and test specimens painted.




Figure 39. Initial 316L plate water jet cut and test specimens painted.

The additional geometry testing and associated data evaluation efforts indicated that a minimum gauge length of 3 inches was necessary for a good uniform strain distribution along the gauge length with minimal transition end effects. In addition, with the goal of achieving strain rate tests in the 5 to 10 per second strain rate range, thinner 1/4-inch thick specimens of the same minimum gauge length of 3 inches were deemed necessary. At this time, it was also recognized that a more formal test specimen identification method was needed that would clearly indicate not only the test specimen's generic profile but also the width and thickness of the gauge length. Therefore, an identifier format of 'X-N-N' was adopted. The X would be an alpha character that could specify generic geometry designations or profiles (e.g., A, D, etc), the first N was a numeric value that indicated the gauge width nominally measured in eighths of an inch, and the second N (the last character) was also a numeric character that indicated the gauge thickness also nominally measured in eighths of an inch. Hence, D-4-4 indicated a D profile with a 1/2-inch gauge width and a 1/2-inch gauge thickness. Similarly, A-2-2 indicated an A profile with a 1/4-inch gauge width and a 1/4-inch gauge thickness. Use of the original 'A' geometry identifier was continued in testing records for material impact tests already completed. However, any material impact testing performed after November 2005 using the older 'A' test specimens would refer to those test specimens as having an A-4-4 geometry.

Figures 40 and 41 illustrate the geometries for the 3-inch gauge length test specimens cut from 1/2-inch plate material (D-4-4 specimens) and 1/4-inch plate material (D-2-2 specimens), respectively. The material was identified in Section 5.1.2.




Figure 40. D-4-4 test specimen geometry for 1/2-inch thick plate material.




Figure 41. D-2-2 test specimen geometry for 1/4-inch thick plate material.

Finally, it was also decided to try a longer gauge length for the 1/4-inch thickness plates, similar to the ‘A’ geometry. Therefore, an A-2-2 test specimen geometry was used to evaluate its effectiveness at achieving low strain rates in the 5 per second range. Figure 42 illustrates this test specimen geometry.




Figure 42. A-2-2 test specimen geometry for 1/4-inch thick plate material.



Figures 14 and 15, in addition to illustrating the material blanks cut for quasi-static testing, show how the above impact test specimens were water jet cut from the 'auxiliary' plate material. Figures 16 and 17 have previously provided illustrations on how the welded plates were water jet cut to obtain impact test specimens along with the material necessary for quasi-static tensile testing. Figure 43 provides a clear indication of how six typical material impact test specimens (four base and two welded material test specimens), were water jet cut from a single welded plate. All of the impact test specimens from the 'auxiliary' plates and the welded plates were longitudinally (parallel with the mill rolling direction) oriented.




Figure 43. Material impact test specimens (D-4-4) water jet cut from welded plate W6.

Note that the base material test specimens were ready for pre-test inspection directly after removal from the plates. On the other hand, the welded material test specimens required additional machining to remove the weld crowns and achieve a square cross section along the entire gauge length. Figure 44 illustrates what a typical welded material impact test specimen square cross section would look like. Included in the test material are weld material, base material, the interface, and heat-affected zone material.




Figure 44. Cross-section of welded material test specimen showing weld material (cross-hatched) and base metal (not cross-hatched) within the square test specimen cross-section profile.

### 6.2.1 Pre-Test Dimensional Measurements

Prior to impact testing, the uniformity and acceptability of each test specimen was determined by NSNFP test personnel using specially developed go/no go gages and marking gages followed by detailed measurements with a calibrated caliper.

The first step was to assess the overall geometry of the test specimen. Even though the gauge length dimensions were of primary importance, it was still deemed appropriate to have nearly identical test specimens. This was determined using a go/no go gage. A go/no go gage is a plate with the test specimen maximum and minimum-tolerance shape step-machined into it. The dimensions checked using these go/no go gages were considered ‘important’ in order to achieve overall test specimen uniformity and so the maximum tolerances were established at  $\pm 0.100$  inches. The fabrication of the go/no go gage reflected these tolerances. Figure 45 illustrates this type of profile check go/no go gage with the ‘important’ dimensions blanked out for determination by the INL dimensional inspector. If an actual test specimen dimension is too large, it will not fit into the gage (can’t go past the Level 1, identified on Figure 45). If an actual test specimen dimension is too small, it will fall past the minimum tolerance step when inserted into the gage (past Level 2 – the next step down). A dimensionally acceptable test specimen will fit into the gage but will not fall past the Level 2 step.




Figure 45. Typical go/no go gage for overall geometry checks of impact test specimens.

Special care was taken to examine the test specimen as it rested on Level 2 of the go/no go gage. The test specimen might not fall past Level 2 if the go/no go gage had a minimal number of locations (potentially just two small points) where the test specimen could rest. Therefore, each test specimen was carefully inspected as it rested in the go/no go gage, looking for gaps around the periphery.

After determining if the test specimen's overall geometry was acceptable, the width and thickness of the entire gauge length was checked. This also was determined with a go/no go gage. Figure 46 is the fabrication sketch of one of the go/no go gages used to determine the geometry acceptance of the gauge length. This go/no go gage was inserted onto the gauge length portion of the test specimen as fully as possible (down to the first machined step), moving the gage along the entire length for the dimension being measured. Insertion past the outer surface and down to the first machined step was required or the test specimen would be rejected. If the test specimen could be fully inserted to the first step, the measurement was deemed acceptable. However, insertion beyond the opening of the first step would cause the test specimen to be rejected since this would indicate a too small dimension. Since these measurements were along the gauge length, these measurements were considered 'critical' for accurate strain measurements and so the maximum tolerances were established at  $\pm 0.020$  inches. If acceptable, the test specimen was marked with its unique test specimen identifier.




Figure 46. Typical go/no go gage for width and thickness checks.

Next, as part of the pre-test checking process, pre-defined punch guide holes (also considered critical dimensions) were established with separate marking gages (see Figure 47 with critical dimensions blanked out for determination by a qualified dimensional inspector), allowing each test specimen to be consistently marked with a punch. Finally, after punch marking and identifying those marks with alpha characters starting with 'A' (see Figure 48 for a marking example), calipers with points were then used to measure the pre-test distances between the punch marks. A second pair of calibrated calipers were then used to measure the widths (front and back) and thicknesses at intermediate points (e.g., points B and C for D-2-2 and D-4-4 geometries). Acceptance of test specimen critical width dimensions were based on the average of front and back width values. The distance between the two outer-most punch marks defined the test specimen's gauge length. These punch mark lengths and width and thickness measurements along the inner



gauge length were recorded on the specimen data sheets [see Appendix D for a representative data sheet (page 1 for pre-test data) and see NSNFP Lab Notebook Binder Volumes 2 and 3 for all the actual pre-test specimen data].




Figure 47. Typical punching gage for 'A' (or A-4-4) test specimens.

Fabrication of these go/no go gages and marking gages were not considered a quality-affecting activity since they were dimensionally checked by an INL qualified dimensional inspector after fabrication and prior to use. After the gages were measured by the qualified dimensional inspector, NSNFP test personnel determined whether or not the go/no go gages were within acceptable tolerances ( $\pm 0.100$  inches for important dimensions,  $\pm 0.020$  inches for critical dimensions). Appropriate gages were used for each different test specimen geometry used for final material impact testing. Dimensional inspections of the gages were made by the qualified dimensional inspector following the completion of each phase of testing. Dimensional inspections of the go/no go and marking gages performed at the completion of prior testing phases were considered as the 'prior to use' dimensional inspection for the next testing phase. As part of a final closeout, the INL Standards and Calibration Laboratory measured these go/no go gages. From their independent measurements, it was determined that the gages were still within the acceptable tolerances for both important and critical dimensions.

### 6.2.2 Final Pre-Test Preparations

The final step of the pre-test preparations was to visibly mark each test specimen so that high-speed digital imagery could be used to perform motion analysis of the impact event. Figure 48 illustrates these marking along with the test specimen's identifier in the format of 304L-XXX or 316L-XXX where XXX is the unique test specimen number.




Figure 48. Typical marked material impact test specimens (with close-up picture of gauge length highlighting punch mark identifiers).

### 6.3 Impact Test Procedure

Actual impact testing began once the test specimens were measured and marked to assure acceptable test specimen geometry, that pre-test gauge lengths and cross-sections were measured, and specimen identification was clear and unique. All specimen measurements were taken using calibrated measuring devices (see Report CD-R, File 3). The generic test procedure is very similar for all of the material impact tests performed.

Step 1: First, the desired test weight (or pig) is selected and that weight is loaded into the ITM, above the tensile test fixture. The drop height has already been established (through preliminary analytical estimates, engineering judgment, or from the findings of prior ITM tests) in an attempt to achieve the desired strain rate and strain in the test specimen.

Step 2: The room temperature impact testing process continues with the insertion of a room temperature soaked test specimen into the tensile test fixture. For the other temperature conditions (-20 °F, 300 °F, and 600 °F), the test specimens are soaked to their necessary pre-test temperatures prior to insertion into the tensile test fixture (cold) or prior to impact testing (hot) as discussed in the next Section 6.4.

Step 3: The drop weight is positioned at the predetermined drop height (measured using a tape measure) in order to achieve the desired impact velocity and impact energy. High-speed digital camera settings and lighting are already established but final checks are made to assure operability.

Step 4: When test preparations are complete, data recordings are started [for accelerometers and the velocimeter (if desired)] and the electronic drop hook is tripped, allowing the drop weight to fall under the influence of gravity.

Step 5: Following conclusion of the material impact test, high-speed digital camera data is evaluated to determine if an adequate strain rate has been achieved. Pertinent information concerning the test (date, test number, drop weight used, drop height, temperature, test specimen number, etc.) was recorded in the NSNFP Lab Notebook.

Weight measurements of the drop weights used in the tests were initially determined with 500 and 5000 pound calibrated load cells (see Report CD-R, File 3) and marked with the measured weight (see Figure 49). These marked weights included the weight of the rigging (approximately 5.7 pounds) necessary to interface with the electronic drop hook. After the completion of impact testing, the weights were weighed by the INL Standards and Calibration Laboratory using a device that was significantly more accurate than the load cells. Appendix E provides detailed comparisons of the weight measurements taken using both devices. For the weights used for the impact testing reported herein, the comparisons were within one pound for all individual weights less than 600 pounds and the single remaining largest weight (labeled 'A' and initially weighed at 1097 pounds) measured 10 pounds lighter, resulting in less than a 1% variation. Hence, the test weight data used throughout the testing effort was considered acceptable.




Figure 49. ITM test weights 'B', 'C', and 'A' (left to right) and disk weights #1 through #6 (on 'B' and 'C') along with the tensile (back left on left pallet), shear (front left on right pallet), and bending (front right on right pallet) test fixtures.

Details regarding data acquisition can be found in Section 6.5 and certain aspects of recording the necessary test data did affect the impact test procedure. Accelerometer output from preliminary ITM tests in early 2004 yielded indications of high frequency responses believed associated with metal-to-metal contact of the drop weight and impact plate. Depending on the stiffness of the system and the proximity of the transducers, metal-to-metal impact can excite high frequency (but low energy) transducer resonances or 'ringing'. Subsequent ITM tests, and all tests reported herein, incorporated a ½-inch thick Buna-N-Rubber (Durometer hardness of approximately 60A, somewhat harder than tire treads) pad laid onto the impact surface of the impact driver to reduce the ringing (and to prevent lateral slippage/sliding at the interface). Test results obtained with and without the pad showed no significant change in test specimen strains.

The early ITM tests attempted to use the accelerometer data to determine the strain rate response of each test specimen. However, when applying various filtering techniques to the accelerometer data to reduce higher frequency noise and lower frequency drift, it was noticed that the time content was being altered. Rather than having to address that adverse consequence, high-speed digital camera results were utilized. This device not only visually recorded the impact event but it also permitted the determination of the deformation or strain history, from which the test specimen's strain rate could be calculated. Acceleration data was no longer needed for quality-affecting data purposes. However, acceleration data was still recorded during impact tests in case future insights of specific test responses needed to be ascertained.

A velocimeter was installed on the ITM even though the INL Standards and Calibration Laboratory determined that the device could not be calibrated within their system. The decision was made that the velocimeter could not provide qualified data. However, the velocimeter data could be used for confirmatory purposes. There was an initial concern that the drop weight might possibly rub along the guide pipe, slowing the drop weight and reducing the final impact velocity. The velocimeter data could be used to support the contention that the drop weight (pig) did not lose any measurable velocity during the drop event. The actual ITM installation and alignment was very accurate and the guide pipe was vertically positioned. No rubbing noises could be heard nor measured by the accelerometers during any of the material impact tests. Measurements from the velocimeter (for over 250 impact tests) confirmed that the final impact velocity matched the velocity magnitude calculated using the simple equation of motion for a free fall of a body under the influence of gravity:

$$v = \sqrt{2gh}$$

where:                      v = velocity at impact (in./sec.)  
                                    g = acceleration of gravity (386.4 in./sec.<sup>2</sup>)  
                                    h = drop height (in.).

Even after performing numerous material impact tests, the NSNFP test personnel were vigilant in monitoring data accuracy and constantly looking for improvements. In mid-June 2006, during impact tests using lighter drop weights (less than 800 pounds), a questionable response was noticed in the deformation history. A potential cause of this odd response was thought to be a slight drag from the velocimeter (as its line was rolling out) that might be tipping the drop weight and altering its direct, flat impact onto the test fixture. It appeared that lighter weights were more susceptible to this potential occurrence. Limited testing was performed that indicated there might be some small affects beyond the variability of the material properties. Therefore, the decision was made to immediately stop using the velocimeter in order to eliminate any potential concerns about its possible affect on the impact tests. In order to reevaluate past impact tests, limited testing was performed using the heavier 1097 pound drop weight (labeled 'A'). No significant variation was seen in this comparison impact testing when the velocimeter was and was not attached to the drop weight. Therefore, only past impact tests that used a drop weight equal to or greater than 1097 pounds (that had the velocimeter attached) would be considered valid.

The material impact testing performed (for acceptable impact tests) limited the maximum strains to below the uniform strain limit, such that necking of the test specimens was not expected to occur. This is a strain limit that most component designs would not want to exceed in pure tension due to impending stability concerns if the through-wall strain levels exceeded this value.



The impact test procedure also needed to reflect an adequate number of impact tests to be performed and an acceptable tolerance on the test results obtained. At the onset of material impact testing in 2004, the goal was to achieve three valid impact tests at certain strain rates (called target strain rates) identified in the Test Plan. Typically, multiple impact tests were performed in order to achieve the desired strain rates. This was necessary due to variations in material properties and the nonlinear responses of the test specimens, making initial response predictions difficult. It was hoped that the test specimen inelastic strains would vary significantly between 0.1 and the uniform strain limit for the three tests conducted at each target strain rate. With a reasonable spread in maximum strain achieved for the three tests at the same strain rate, it would then be possible to generate an elevated true stress-strain curve reflecting the average of those three target strain rate tests. Obviously, a tight tolerance on the strain rate was needed in order to generate an elevated true stress-strain curve that reflected the target strain rate. Therefore, it became necessary to define an acceptable tolerance on the strain rates achieved. For target strain rates above 20 per second, a tolerance of  $\pm 2$  per second was established. For target strain rates of 10 per second and below, an initial tolerance of  $\pm 2$  per second was established but was later tightened to  $\pm 1$  per second for the associated impact tests reported herein. The goal was to develop a sufficient number of these elevated true stress-strain curves such that the structural analyst could adequately define material responses for a drop event. While this was adequate for initial impact testing and establishing performance characteristics of the test specimens, numerous impact tests were not being utilized.

After contemplating various approaches to increase the use of all data from valid impact tests, it was decided to still attempt to achieve the target strain rates identified in the Test Plan. For base material, three distinct impact tests were attained for each target strain rate. For welded material (due to a limited number of available test specimens), only two impact tests per target strain rate were attained. Impact testing still had to be performed at some strain rate levels and achieving the target strain rates satisfied the Test Plan requirement. This would also still permit the generation of elevated true stress-strain curves at specific target strain rates if deemed pertinent. However, finding a way to use all of the available valid impact test data was still needed. The NSNFP test personnel determined that if factors (constant values that quantified the strain rate effects for impact tests) could be calculated efficiently, then factor versus strain rate curves could be generated. This would incorporate all valid impact testing results. Hence, this approach was adopted.

## **6.4 Achieving Impact Temperature Conditions**

Room temperature impact testing results and engineering judgment were combined to determine that only the D-2-2 and the D-4-4 test specimen geometries would be needed to achieve the desired strain rates for the cold and elevated temperature impact testing. Limiting the number of test specimen geometries minimized the amount of temperature baseline testing needed to establish the parameters necessary to achieve adequate impact temperature conditions. If this initial assessment proved wrong, additional temperature baseline testing would be performed for additional test specimen geometries. However, the impact testing indeed demonstrated that these two test specimen geometries were sufficient to allow all of the target strain rates to be achieved.

### **6.4.1 Room Temperature Testing**

Considering all of the room temperature material impact testing completed, the room temperatures ranged from 52 °F to 85 °F. The locations where the material impact testing was performed were in two older facilities that had limited temperature controls during the fall, winter,

spring, and summer seasons. However, this range of room temperatures is still considered acceptable. Performing the room temperature impact testing did not require establishing any specific release times for the drop hook. The impact tests were performed when test preparations were complete and the NSNFP test personnel were ready.

#### **6.4.2 Elevated Temperature Testing**

The NSNFP assigned the task of performing strain rate impact testing at elevated temperatures of 300 °F and 600 °F for both base and welded material. The initial question was how to achieve the test specimen heating process. Heating the entire test fixture and test specimen to 600 °F was not considered wise when repeatedly impacting the test fixture (made of A36 carbon steel) with weights up to 1600 pounds. This would probably deform the test fixture too much for future use. Heating the test specimens in an oven and then loading them into the test fixture would be possible but the process of handling a hot test specimen in a short enough time interval to maintain proper temperature conditions did not seem reasonable. Open flame heating was considered too dangerous. Electric heat guns (focused on heating the test specimen itself while in the test fixture) were chosen as the best option.

Multiple heating setups were tested before effective heating processes were proven viable for both the D-4-4 and the D-2-2 test specimens. The D-4-4 and D-2-2 test specimens used for temperature baseline testing had five equally spaced holes for thermocouples to be embedded into the central volume of the test specimen along its gauge length (Figure 50 illustrates the D-2-2 temperature baseline test specimen). Preliminary baseline testing (not documented in the NSNFP Lab Notebook) tried different numbers of electric heat guns, different heat gun positions, different heat gun settings, and different heating time intervals. In addition, it was recognized during the preliminary testing that heat shields or reflectors were needed to provide additional local heating control. Two heat shields (for front and back) were developed for the D-4-4 and D-2-2 specimen testing and another unique pair were created for the D-2-2 specimen testing at 600 °F (see Figure 51 for a picture of each heat shield geometry). Finally, for the D-2-2 baseline specimen testing, it was recognized that the thinner gauge length lost heat much quicker than the D-4-4 specimen. In order to better protect the D-2-2 specimen, a 'cradle' (see Figure 52) was fabricated that would essentially insulate the D-2-2 specimen test length on three sides (while still leaving viewing room for the high-speed camera). Hence, the preliminary testing permitted a large variety of heating methods to be tried and adjusted before the final (or near final) test setup was established. At this stage, documentation of the temperature baseline testing commenced in the NSNFP Lab Notebook.

The goal of the temperature baseline testing was to demonstrate the validity of the proposed heating process for both the 300 °F and 600 °F testing efforts and to demonstrate that it was repeatable. The test process approach required that the test specimens be overheated in order to allow time for test preparations (e.g., removing heating equipment for a clear camera view) to occur while the test specimen was cooling down to the target temperature. Therefore, the heating process also had to determine at what time interval (following the heat guns being turned off) the impact test could commence (see Section 6.4.2.1).




Figure 50. Temperature baseline testing specimen D-2-2 showing thermocouple hole locations.



Figure 51. Heat shields for all test specimens (left) except D-2-2 test specimens at 600 °F (right).





Figure 52. D-2-2 cradle holding an impacted test specimen (upper right insert is of cradle alone).

The final heating configuration consisted of four heat guns. The heat guns were symmetrically placed, two in the front and two in the back, as illustrated in Figure 53. The bottom heat guns had three-inch long nozzle extenders and were positioned into the slot cut into the lower cross-member and pushed fully forward. This alignment centered the directed heat onto a spot approximately 1-inch below the 3-inch wide shoulder transition, onto the large bottom holder area. The top heat guns were positioned so that the nozzles were in contact with the bottom surface of the upper cross-member and pushed fully forward. This alignment centered the directed heat onto a spot approximately ½-inch below the 3-inch wide shoulder transition, onto the gauge length. Once an acceptable heating level was achieved, the barrels on the heat guns were marked to establish that particular heat gun setting (no gradations existed on the heat guns so marking was necessary in order to reset the heat gun settings). Table 8 indicates the heat gun barrel marking used.

Table 8. Heat gun barrel markings used for elevated temperature impact testing.

Test Specimen Geometry		300 °F Testing	600 °F Testing
D-2-2	Top Guns	L-L	J-J
	Bottom Guns		
D-4-4	Top Guns	D-D	E-E
	Bottom Guns	H-H	



Figure 53. Temperature baseline testing setup showing four heat guns in position.

The heat shields were placed on top of the lower cross-member, bridging across the slot cut, and pressed up against the test specimen, front and back. Constant monitoring during the temperature baseline testing (and during the heating process for impact testing) was necessary to make sure these heat shields did not vibrate (due to operation of the heat guns) out of position

during the heating process. The heat shields allowed the heating process to concentrate more heat into the bottom holder areas of the test specimens, minimizing heat loss through those large areas while the notches permitted hot air to rise and bathe the gauge length. Overall, much improved temperature distributions were measured throughout the gauge length with the heat shields in place.

In order to achieve acceptable test specimen temperature tolerances, a very prescriptive heat-up process was used. The settings for the heat guns, marked during baseline testing, were achieved by adjusting to a full power setting and then reducing the setting down to the desired power level marking at the start of every day of elevated temperature testing. As indicated above, five thermocouple readouts (at 1.5 second intervals) were recorded (at three second intervals) for each temperature baseline test performed. Achieving consistent timing of the heating cycle durations was also considered vitally important in order to achieve consistency in the heating process (Section 6.4.2.1 explains the timing aspects in more detail). Every temperature baseline test performed as a basis test is included in NSNFP Lab Notebook Binder Volume 4. As can be seen in Figure 54, a representative D-2-2 test specimen was heated until a steady-state condition was achieved, and then the heat guns were turned off (after 25 minutes of heating as monitored on the digital timer). Thermocouple data was continuously recorded after the heat guns were turned off so that the test specimen cooldown (toward the target impact temperature of 600 °F in this specific case) could be monitored. Figure 55 illustrates the associated enlargement of the cooldown process (also contained in the NSNFP Lab Notebook Binder Volume 4). As can be seen, 18 seconds after the heat guns were turned off, the maximum temperature was 626 °F and the minimum temperature was 578 °F.




Figure 54. Representative temperature baseline test showing heat-up, steady-state, and the cooldown following the heat guns being turned off after 25 minutes of heating.




Figure 55. Representative cooldown with release time of 18 seconds identified by vertical dashed line showing temperature range of 578 to 626 °F.

It was recognized that heat gun settings and time duration alone were not sufficient to justify the acceptable heating of the test specimens for the actual impact testing. Therefore, it was decided that actual temperature measurements of each test specimen before impact testing at pre-established times (see Section 6.4.2.1) near the end of the heating process should be performed to gain further confirmation that each heat-up cycle was acceptable and following the baseline testing parameters. The initial assumption during temperature baseline testing was that the temperature at the center portion of the gauge length would be sufficient to judge the adequacy of the test specimen temperature and the heating process. Early on in the temperature baseline testing, test specimen measurements of the hot spot where the top heat guns were directed onto the gauge length were monitored to determine just how hot those locations became. However, during heat gun failure testing (see Section 6.4.2.3), it was determined that another temperature measurement near the bottom punch mark of the test specimen was needed to assure proper heat gun performance during the heating cycle. Therefore, after the heat gun failure testing, it became routine during temperature baseline testing to measure the temperatures of the test specimen (D-4-4 specimen on the side) or test specimen holder (D-2-2 cradle on the side) at the center and at the bottom punch mark locations. A thermocouple was used for these temperature measurements. The temperature baseline testing established the timing for the measurements and the acceptable ranges for these measured test specimen temperatures [Figure 54 illustrates a bottom temperature was measured first (570 °F) and then a center position (619 °F) temperature measurement was taken]. Table 9 indicates the acceptable ranges of these two test specimen temperatures (based on the multiple temperature baseline tests performed) as well as the ranges of temperatures achieved during actual material impact testing. All material impact tests conformed to the test specimen temperature measurement tolerances established.



Table 9. Measured test specimen temperature ranges during baseline and impact testing.

Data Recorded	Measurement Location	300 °F Testing		600 °F Testing	
		D-2-2	D-4-4	D-2-2	D-4-4
During Baseline Testing	Bottom	270 to 320 °F	280 to 320 °F	540 to 630 °F	570 to 620 °F
	Center	280 to 320 °F	285 to 320 °F	585 to 635 °F	580 to 635 °F
During Impact Testing	Bottom	270 to 314 °F	282 to 321 °F*	545 to 611 °F	572 to 607 °F
	Center	283 to 307 °F	286 to 317 °F	585 to 627 °F	580 to 592 °F
* - The one 321 °F temperature was accepted since heat gun failure was clearly not indicated.					

Note that early on in the impact testing, efforts to measure the D-4-4 test specimens proved difficult. The measured temperatures were unexpectedly low. After various attempts, including performing heat-up testing with the thermocouple instrumented baseline test specimen again, it was recognized that the impact test specimens, since they were water jet cut, had a rougher surface on the sides where the temperature measurements were taken whereas the baseline test specimens had smoother machined surfaces. (For D-2-2 testing, a cradle was used for baseline testing and impact testing and the cradle side surfaces were machined smooth.) Therefore, the side surfaces on the impact test specimens where the temperature measurements were to be taken were locally polished in order to improve the thermocouple contact. This greatly improved the impact test specimen temperature measuring process. For each successful heat-up cycle, the confirmatory bottom and center test specimen temperature measurements made prior to impact testing were recorded and entered into the lab notebook. If the measured temperatures were not acceptable, the heating process was continued (heat guns turned off at the proper time) but the impact test was not performed.

Two other issues were discussed during preliminary temperature testing and a final decision was made to address each in the temperature baseline testing. The heat-up cycle was very time consuming (see Section 6.4.2.1) with a lot of heat being absorbed by the test fixture. It was believed that another indication of repeatability would be the temperatures of the test fixture, taken before and after each heat-up cycle. Although not expected to be a precise indicator of test specimen temperatures, knowing the generic temperature history of the test fixture throughout the heating process could provide additional confirmation of a proper and repeatable process. Section 6.4.2.2 contains additional information on temperature measurements of the test fixture. In an effort to address potential equipment failure, additional testing was performed to determine if a single heat gun failure could be detected. Section 6.4.2.3 describes how the test specimen pre-impact test temperatures were utilized to help detect a potential heat gun failure.

The NSNFP Lab Notebook provides the overall process and insights related to the qualified temperature baseline testing performed and the associated Binder Volume 4 provides the test specific baseline results (each test geometry at each elevated temperature) as well as testing summaries. A full testing day's worth of temperature baseline testing was achieved and then repeated on another day. In an effort to demonstrate full repeatability, the heat guns were reset prior to a second day of temperature baseline testing. Finally, if any problems arose, it was decided early on that the instrumented test specimen used for the temperature baseline testing could be re-used on impact testing days or other days as necessary to gain insights on a continuing acceptable heat-up

process. Those types of confirmation checks were made and each confirmed that the heating process was always within stated tolerances. These confirmatory checks are also documented in the NSNFP Lab Notebook Binder Volume 4.

#### **6.4.2.1 Heating/Cooldown Durations and Trip Hook Release Times**

The following shows the heat-up and cooldown cycle time intervals used during testing. These intervals were determined during the preliminary testing, used throughout temperature baseline testing, and followed during actual material impact testing at elevated temperatures. The time intervals established between heat guns off and heat guns back on allowed not only for the impact test to be performed but also were established in order to allow adequate time for the determination of the impact test strain rate, permit the test fixture to cooldown for handling purposes, and to allow for sufficient time to change drop weights. All times were determined and followed using a digital timer. A continuous timing sequence was used (see below for temperature specific timing intervals) that began at the start of a test day (heat guns on) and ended with the final material impact test for that day.

##### **Event Timing**

0 min.  
12 to 13 min.  
14 to 15 min.  
17 min.  
Specimen dependent  
52 min. or  
1 hr 22 min. (after third test only)

##### **300 °F Impact Tests (Six tests maximum per day)**

Start Heat Guns  
Bottom Punch Measurement (initially the hot spot)  
Centerline Measurement  
Heat Guns Off  
Drop Weight Released  
Heat Guns Back On (repeat above steps for next test)

##### **Event Timing**

First Test of Day:

0 min.  
55 to 56 min.  
57 to 58 min.  
60 min.  
Specimen dependent  
1 hr 45 min.

##### **600 °F Impact Tests (Five tests maximum per day)**

Start Heat Guns  
Bottom Punch Measurement (initially the hot spot)  
Centerline Measurement  
Heat Guns Off  
Drop Weight Released  
Heat Guns On (go to timing below for remaining tests)

Remaining Four Tests of Day:

0 min.  
20 to 21 min.  
22 to 23 min.  
25 min.  
Specimen dependent  
1 hr 15 min.

Start Heat Guns  
Bottom Punch Measurement (initially the hot spot)  
Centerline Measurement  
Heat Guns Off  
Drop Weight Released  
Heat Guns Back On (repeat above steps for next test)

In addition, the drop hook release time needed to be established. The baseline testing established the drop hook release times, which were hoped to be less than 30 seconds from turning off the heat guns in order to minimize thermal gradients across the test specimen thickness. Obviously, the impact time is the most critical time parameter but with different weights resulting in different drop heights, the time between drop hook release and actual impact varied between zero and 1 second. Therefore, if a 1 second long time interval could be determined wherein the worst case temperature values achieved acceptable weighted average temperatures and acceptable maximum and minimum temperatures, then the drop hook should be activated at the start of that 1 second time interval. This is what is documented in NSNFP Lab Notebook Binder Volume 4. The one exception was for the D-4-4 test specimen at 300 °F temperature. The time at which the measured worst case temperature ranges achieved the minimal temperature ranges (22 seconds) was much longer than the other tests (16 to 20 seconds). Therefore, it was decided to not exceed 20 seconds for a trip hook activation time so that any potential thermal gradient effects through the test specimen thickness would be minimized and not become a detrimental issue. Table 10 below summarizes the temperature tolerances achieved during temperature baseline testing for each target temperature for each unique test specimen geometry at the drop hook release time or a one second interval afterwards. These drop hook release times were achieved by the NSNFP test personnel for each material impact test completed.

Table 10. Drop hook release times and associated test specimen temperatures.

Target Temp. (°F)	Test Specimen Geometry					
	D-2-2			D-4-4		
	Drop Hook Release Time (sec.)	Weighted Avg. Temp. Range (°F)	Max/Min Temp. Range (°F)	Drop Hook Release Time (sec.)	Weighted Avg. Temp. Range (°F)	Max/Min Temp. Range (°F)
300	16	+5/-10	+23/-24	20	+12/-7	+18/-12
600	18	+15/+4	+32/-32	18	+8/-7	+28/-28
Note: Temperature ranges are from target temperatures of 300 °F and 600 °F. Hook release time is interval after heat guns are turned off.						

The weighted average temperature covers the entire gauge length whereas the max/min temperature ranges are the highest or lowest temperature regardless of specific location. One important point to recognize is that the minimum temperature range typically reflected the temperature measurements made at the extreme ends (top and bottom locations) of the gauge length. This is reasonable due to the heat loss through the pin connected ends (through the top and bottom cross-members) of the test specimen. Note that Table 9 indicates the worst-case temperature tolerances achieved during all of the numerous temperature baseline tests completed. For most of the tests, the weighted average temperature and maximum and minimum temperatures were much closer to the target test temperatures.



#### 6.4.2.2 Test Fixture Temperatures

Test fixture temperatures were recorded during the temperature baseline testing as well as during the actual material impact tests. Temperature measurements of the upper cross-member (on the top surface centered in front of the slotted opening) and lower cross-member (facing the impact driver, on the right side approximately 2 inches down from the top and 2 inches toward the back) were taken immediately before the heat guns were turned on and approximately one minute after the heat guns were turned off. These temperatures, recorded in the NSNFP Lab Notebook obviously varied somewhat due to variations in the heating process and to varying initial conditions.

Although not considered as significant as the measured test specimen temperatures, this information was considered to provide insights into the overall performance of the heat guns throughout the impact testing effort and to also provide test repeatability insights. Since the start of each day's impact testing could be at a different room temperature than that of the baseline testing, it was decided to look at the temperature changes of the test fixture throughout the multiple heat-up cycles for a full day of testing and see how they compared to the baseline testing data. Variances are those temperature values that were beyond the range of test fixture temperatures established during baseline testing. The variance was considered zero if the change in temperature was within the tolerance achieved during the two days of baseline testing. These variance data are determined in the NSNFP Lab Notebook Binder Volume 4 and are summarized below in Table 11.

Table 11. Test fixture temperature variances.

Impact Test Variances From Temperature Baseline Testing	300 °F		600 °F	
	D-2-2	D-4-4	D-2-2	D-4-4
Largest Single Test Variance	8 °F	11 °F	24 °F	20 °F
Average Variance of Viable Impact Tests	2.4 °F	3.7 °F	9.3 °F	7.0 °F

Initial desires were that the worst case variance from that established during temperature baseline testing would be less than 25 °F. This tolerance was achieved. Hence, the test fixture temperature data also confirms an acceptable heating process was used for the elevated temperature material impact testing.

#### 6.4.2.3 Heat Gun Failure

The possibility of heat gun failure during a heat-up process was investigated. Documentation of that effort and the establishment of acceptable temperature range checks is contained in the temperature baseline testing data. Plots in the NSNFP Lab Notebook Binder Volume 4 show temperature time histories of what happened during 300 °F and 600 °F baseline testing if a specific heat gun failed. This was achieved by following the established temperature baseline heat-up process (using the instrumented test specimens) until reasonable steady state conditions were achieved and then a heat gun was turned off while the remaining three heat guns

continued to heat. This was repeated four times until each heat gun in sequence had been turned off. The 300 °F heat-up process was repeated but it was felt that one test for 600 °F testing was sufficient since the results were similar to the 300 °F testing.

The results show that center test specimen measurements accurately indicated a top front or top back heat gun failure. The center temperatures reduced by 28 to 44 °F for 300 °F testing and 56 to 74 °F for 600 °F testing, a very measurable difference. For a bottom front or bottom back heat gun failure, the center test specimen temperature measurement proved to be inconclusive so measurements of the bottom punch (or base) area were taken and these proved much more insightful. The bottom punch or base temperatures reduced by approximately 42 to 61 °F for 300 °F testing and 70 to 132 °F for 600 °F testing, again a very measurable difference.

It was decided that in place of hot spot temperature checks, test specimen temperature measurements at the bottom punch mark were necessary to assure that none of the four heat guns failed during the heat-up cycle. Therefore, heat gun failure baseline testing was used to establish acceptable ranges for bottom punch (or base) temperature measurements. The heat-up and cooldown baseline testing mandated higher acceptable temperatures than those determined for proper heat gun performance at the center position. In conclusion, these heat gun failure data were used to assure that heat gun failure would be detected and impact testing would therefore be terminated until a replacement heat gun could be incorporated and validated with additional temperature baseline testing.

#### **6.4.2.4 Resulting Test Specimen Temperatures at Impact**

An initial accuracy target of the weighted average test temperature of  $\pm 25$  °F was specified in the Test Plan but it was quickly realized that not only should a reasonable weighted average test specimen temperature be achieved but that a limit on the highest and lowest temperatures must also be imposed. Initial desires for acceptable maximum and minimum temperature ranges (within the test specimen gauge length) were  $\pm 25$  °F for the 300 °F testing and  $\pm 35$  °F for the 600 °F testing. As can be seen in the baseline testing information provided in Table 10, these accuracy goals were met. The worst case extremes (of any of the associated baseline tests) were then used to establish the temperature ranges applicable to the heating process used. However, most of the baseline testing results were actually much closer to the target temperatures than the extremes listed.

It was expected that certain variabilities in test conditions could arise for the many anticipated days of actual material impact testing. Therefore, the previously stated temperature ranges achieved during temperature baseline testing (see Table 10) were rounded up to values that would become the stated temperature tolerances achieved during impact testing. These final temperature tolerances are presented in Table 12.

Table 12. Temperature tolerances for material impact testing.

Target Temperature (°F)	All Test Specimens	
	Weighted Avg. Temperature Range (°F)	Max/Min Temperature Range (°F)
300	+15/-10	$\pm 25$
600	+15/-10	$\pm 35$

In addition to the significant amount of measured temperature data, a thermal analysis of the temperature cooldown of a D-4-4 test specimen was performed to provide more insights into the validity of the measured temperatures and insight on whether significant thermal gradients existed through the thickness during cooldown. This thermal analysis effort, reported in EDF-NSNF-065 (Reference 39) provided the following conclusion:

“The main conclusion, deduced from the comparison of the predictions of the ABAQUS model to the test data for the 300 and 600 °F tests, is that the thermocouples should accurately represent the transient thermal response of the test coupon during the initial cooldown (i.e., the first 20-30 seconds after the heaters are shut off). Thermocouple lag, or delay, effects were not seen in the data. Also, the temperature distribution across the thickness of the coupon test section was determined to be uniform.”

#### **6.4.3 Cold Temperature Testing**

For the -20 °F material impact testing, the decision was made to use a research freezer (with a stated capacity down to -40 °F) to cold-soak the test specimens. Similar to the elevated temperature testing, it was decided to not cold-soak the entire tensile test fixture down to -20 °F. It was decided to cold-soak each test specimen down to a temperature lower than -20 °F and then take the time (while the test specimen was warming) to insert the test specimen into the test fixture and prepare for impact testing.

Preliminary temperature baseline testing determined how much colder than -20 °F was necessary to allow sufficient time for placement of the test specimen into the test fixture and prepare for the impact test. A time interval no longer than one minute was thought to be adequate for impact testing once the test specimen had been removed from the research freezer. Instrumented temperature baseline test specimens [both the D-2-2 (with cradle) and the D-4-4 test specimens as used for the elevated temperature testing] with three thermocouples inserted into the mid-cross section (top punch, bottom punch, and center positions along the gauge length) were used to determine the test specimen temperature responses. Both the D-4-4 and the D-2-2 preliminary temperature results indicated that an initial cold-soak temperature of approximately -33 °F was needed. Preliminary testing also indicated that having just a few specimens in the freezer caused the freezer to cycle frequently because it was difficult to keep just air sufficiently cold. Therefore, a thick stainless steel plate was inserted into the freezer so that this denser mass could help the freezer temperature stabilize. This established a more consistent temperature in the freezer and all test specimens were stacked onto this plate to promote more rapid cooling.

Temperature baseline testing for cold temperatures could now proceed with data being recorded in the NSNFP Lab Notebook. In order to simulate impact testing conditions as much as possible, the three thermocouple leads were taped close to the test specimen body so that the specimen could still be inserted into the test fixture while monitoring and recording the temperatures. Timing was monitored using the same digital timer used for the elevated temperature testing. Timing started with the opening of the research freezer. The instrumented test specimen was removed from the freezer, loaded into the test fixture, a one minute pause interval occurred, and then the specimen was returned to the freezer. Rather than directly holding the test specimen, a tab of duct tape was used to hold the test specimen, thus minimizing localized heat gain. Only test specimens that had cold-soaked at least overnight were used to establish the warm-up timing.

However, other warm-up tests were performed during the remainder of the day once the test specimen had cold-soaked and stabilized to at least a temperature below  $-30^{\circ}\text{F}$ . These tests, using both test specimen geometries, were used to demonstrate the repeatability of the test process. The NSNFP Lab Notebook Binder Volume 4 contains this repeatability data. These repeatability tests (using the weighted averages of the recorded temperatures) were biased to start at the average temperature of the established starting temperature range. These plots (illustrated in Figure 56 for the D-2-2 test specimen repeatability tests biased to  $-33.25^{\circ}\text{F}$ ) show a very tight grouping of curves, confirming a very repeatable warm-up process, especially up to the time that the drop hook was released. The D-4-4 temperature baseline testing results (not shown), reflecting a test specimen with more thermal mass, had an even closer temperature grouping than the D-2-2 results.




Figure 56. Repeatability tests for D-2-2 test specimen all temperatures biased to start at  $-33.25^{\circ}\text{F}$ .

As previously mentioned, only the thoroughly cold-soaked test specimens were tested (first tests of the day) to establish the timing for the impact tests. Warm-up tests 1C, 11C, 24C, and 25C addressed the D-2-2 specimens and warm-up tests 2C, 12C, 21C, and 23C addressed the D-4-4 specimens. Presentation and evaluation of this data is contained in NSNFP Lab Notebook Binder Volume 4. Figure 57 illustrates the warm-up test results for D-2-2 test specimen at test 25C. Again, note that the D-4-4 test results provided even closer temperature grouping due to the larger thermal mass of the thicker test specimen.




Figure 57. Typical D-2-2 warm-up test from overnight cold-soaked conditions.

Table 13 lists the weighted average temperature and the maximum and minimum temperature ranges determined during the temperature baseline testing. In addition, the averages of the weighted average, maximum, and minimum temperatures for each of the four baseline tests were then determined. From these averages of the averages, constant factor adjustments were made to address the probable temperatures that would be achieved at an anticipated minimum starting temperature and at an anticipated maximum starting temperature. These anticipated starting temperatures became the acceptable starting temperature ranges for cold impact testing. For the D-4-4 specimens, the starting acceptable temperature range for impact testing was -31 to -34.5 °F. For the D-2-2 specimens, the acceptable starting temperature range for impact testing was -32 to -34.5 °F.

Table 13. Test specimen cold temperature ranges at impact test.

Temperature Parameter	D-2-2 Specimen	D-4-4 Specimen
Weighted Average Temperature Range at Impact	-21.8 to -23.6 °F	-20.7 to -22.1 °F
Maximum / Minimum Temperature Range at Impact	-19.0 to -24.3 °F	-20.5 to -22.3 °F

Testing was also performed to demonstrate that cradles reused during the test day would not adversely affect the temperatures of D-2-2 test specimens that had cold-soaked overnight. After performing an overnight cold-soaked D-2-2 warm-up test, the cradle was returned to the

freezer no longer than 3 minutes and 15 seconds after initial opening of the freezer for test specimen removal. After waiting 45 minutes, another overnight cold-soaked test specimen was placed into the cradle. After 15 more minutes of cooling, the center temperature of the cold-soaked test specimen was checked. The D-2-2 test specimen was just as cold as it had been after the overnight cold-soak. Therefore, this same minimal timing for cradle cooling was applied if cradles were to be reused during a single test day. Appropriate time intervals were recorded in the NSNFP Lab Notebook for cradles reused during a single day.

#### **6.4.3.1 Trip Hook Release Times**

The established timing interval for the -20 °F material impact testing (from freezer opening to release of drop hook) was 36 seconds for the D-2-2 specimens and 51 seconds for the D-4-4 test specimens. Temperature baseline test data can be found in the NSNFP Lab Notebook Binder Volume 4 for the -20 °F temperature testing efforts.

#### **6.4.3.2 Acceptability of Cold Test Specimen Temperatures at Impact**

Table 14 summarizes the actual measured starting temperature ranges achieved during the -20 °F impact testing performed during 2007. These temperature ranges are based on temperature measurements from an instrumented test specimen that was placed inside the freezer during the entire testing sequence. This starting temperature range satisfies both the D-2-2 and the D-4-4 starting temperature range criteria established during temperature baseline testing. The room temperatures during actual impact tests ranged from 66 to 71 °F, very similar to the room temperatures experienced during baseline testing (68 to 71 °F). In fact, the room temperature during cold impact testing never exceeded the maximum baseline testing room temperature experienced and were within 2 °F of the minimum room temperature experienced during baseline testing. Hence, the room temperature conditions during actual impact testing had no adverse effect on the timing intervals established for the cold test specimens.

Table 14. Initial temperature ranges for cold material impact testing.

<b>Test Specimen Geometry</b>	<b>Initial Temperature Range (°F)</b>
D-2-2 or D-4-4	-32.1 to -33.4 °F

Therefore, the actual temperatures achieved in the test specimens for cold impact testing were expected to have been within the ranges established during temperature baseline testing and shown in Table 13. These narrow temperature ranges were acceptable for impact testing. It is interesting to note that the warmest starting test specimen temperature recorded for impact testing was -32.1 °F (compared to an allowable of -31 °F) so the actual temperature ranges would have been even closer than the temperatures predicted in Table 13.

## **6.5 Data Acquisition**

This section discusses in more detail the types of data recorded during the actual impact tests. Pre- and post-test dimensional measurements of the test specimens are addressed in Sections 6.2 and 6.7, respectively.



### 6.5.1 High-Speed Digital Camera

Strain rates vary with time during accidental drops or crash events. Since the ITM reflects a true impact event, the strain rate starts out at a fairly constant value but then dissipates to zero as the impact energy of the dropped weight is transformed into permanent test specimen deformation.

A Photron Fastcam high-speed digital camera was used to record image data of the impact events. The camera was positioned to look directly at the specimen front surface. Displacement histories for test specimen geometry 'A' (A-4-4) punch marks A through D (see Figure 48) were recorded during a drop event. Punch mark E was not visible to the camera during any impact testing performed prior to cutting the slot into the test fixture's lower cross-member (see Section 6.1). For test specimen geometries A-2-2, D-2-2, and D-4-4 (all tested after November 21, 2005 with the slot already cut), the entire gauge length of the test specimen was visible to the camera. Imaging data were recorded at a frame rate of 3000 frames per second and a resolution of 512 x 1024 pixels for the initial strain rate testing performed in 2004. However, starting in early 2005, the imaging data were recorded at 8000 frames per second at a resolution of 256 x 1024 pixels.

Motion analysis of the digital image data was used to determine displacement histories of the camera visible specimen punch marks. Displacement histories of camera visible punch marks on the test specimen were measured directly from the frame exposures. Engineering strain histories were developed from the recorded displacement histories and were converted to true strain histories using the relationship:

$$\varepsilon_{true} = \ln(1 + \varepsilon_{engineering})$$

For this strain rate research, the impact test strain rate was defined as that rate of straining occurring early in the specimen response stage where the strain rate is nearly constant with time. Figure 58 illustrates a typical strain history data plot used to determine strain rates. For the test shown, the slope was determined to be a line from the origin through a true strain data point of 0.1558 at time 0.006 seconds. This calculates to be a strain rate of  $0.1558/0.006 = 26.0$  per second. Strain rate data were recorded on page 2 of the test data sheets (as shown in the Appendix D sample and contained in the NSNFP Lab Notebook Binder Volumes 2 and 3 for each impact test).

Imagery of the impact event from before impact to after specimen initial rebound was saved to a DVD, typically 1340 frames for each test. (These DVDs are part of the NSNFP Lab Notebook Binder Volume 5 and were submitted to the NSNFP as records.) The displacements of the camera visible punch marks were used to develop strain histories for the specimens by performing motion analysis of the recorded image files. Motion analysis data are contained in the NSNFP Lab Notebook Binder Volume 5 (eight books total). The camera frame rate was calibrated before and after testing to assure accuracy (see Report CD-R, file 3). Motion analysis results (strains) were also compared to test specimen final measured results (strains) to verify accuracy. The camera data were (on average for all 260 impact tests used to calculate factors) within 2.5% of the measured test specimen strains. This indicates that the strain rates derived from the camera data were accurately determined. Section 7.1 contains camera versus measured strain comparisons for every impact test considered herein.




Figure 58. Camera determined true strain history for specimen 304L-40.

### 6.5.2 Accelerometer and Velocimeter Data

During an impact test, test specimen acceleration histories were measured using calibrated 500 g and 5000 g accelerometers mounted on the back side of the impact driver lower cross-member. Acceleration histories were recorded on a Data Acquisition System (DAQ) 2000A at a rate of 5000 samples per second for 20 seconds. Acceleration data were recorded with the goal of having that additional impact data in case of future test specimen response questions. The raw acceleration data were filtered using a Butterworth bandpass filter. The frequency band used was 2Hz – 500Hz. Integration of this acceleration trace was also performed using DADiSP 2002. Both of these results were recorded on page 2 of the test data sheets for both the 304L and 316L test specimens, located in NSNFP Lab Notebook Binder Volumes 2 and 3, respectively.

As previously mentioned in Section 6.3, drop weight velocities were initially measured by the attachment of a velocimeter (a device that could not be calibrated at the INL) to the lower end of the drop weight. Velocity histories were recorded on a Data Acquisition System (DAQ) 2000A at a rate of 5000 samples per second for 20 seconds. Prior to mid-June 2006, velocimeter data was recorded and these results can also be found in the NSNFP Lab Notebook Binder Volumes 2 and 3 for those tests where incorporated.

Accelerometer and velocimeter data were recorded as ‘data of interest’. The data gathered by these devices supported observations made by the NSNFP test personnel and certain aspects of the camera data. However, they were not used in the development of the factors or the strain rate elevated true stress-strain curves.

## 6.6 Previous ITM Performance

Having described the ITM, the test specimens, and the impact test methodology used, a brief review of the two locations where the ITM has been used for material impact testing is appropriate. Also discussed herein is the past testing of both longitudinal and transverse oriented test specimens, the results of which were used to establish how test specimens were to be cut from the plate material.

### 6.6.1 Test Performance at Different Test Sites

The initial placement of the ITM at the Material Test Reactor building (RTC-603) at the INL occurred on November 12, 2003. The MTR building had an overhead crane, more than 30 feet of overhead clearance above the ITM, and a robust floor (consisting of a 12-inch thick concrete slab poured over large concrete encased steel beams) with a load rating of 1250 pounds per square foot. Due to building demolition plans, the ITM was then moved to the INL's Critical Infrastructure Test Range Complex (CITRC) Building 613 in late April, 2005. This facility also has an overhead crane, more than 25 feet of overhead clearance above the ITM, and a 4.75-foot thick reinforced concrete floor situated on bedrock that has a posted allowable floor loading of 2500 lb./ft<sup>2</sup>. Both locations have strong, rigid floors. However, it was decided to perform confirmatory testing of duplicate impact tests in order to determine if any noticeable change in test specimen response could be detected. Comparisons of these confirmatory impact tests appear in Table 15 below for 304L material and Table 16 below for 316L material. Some of the tests reported below are using different material heats and different test specimen geometries (all of the 304L tests excluding 69 and 172) from another research effort but the comparative insights are still valid.

Table 15. Comparison of 304L strain responses for impact tests performed at different locations.

Test Number	Location	Drop Weight (lb.)	Drop Height (in.)	True Strain	Variation (%)
69	MTR	1513	31	0.3242	2.56
172	613			0.3161	
91	MTR	1513	14	0.4116	2.57
176	613			0.4013	
96	MTR	1347	15.75	0.4146	9.45
185	613			0.3788	
100	MTR	1097	20	0.3949	0.92
193	613			0.3913	
104	MTR	790	25.375	0.3406	0.38
198	613			0.3393	
Average					3.18

Table 16. Comparison of 316L strain responses for impact tests performed at different locations.

Test Number	Location	Drop Weight (lb.)	Drop Height (in.)	True Strain	Variation (%)
24	MTR	1513	36	0.3878	2.05
174	613			0.3959	
75	MTR	1347	40	0.3843	2.19
184	613			0.3929	
28	MTR	1097	45	0.3493	0.85
192	613			0.3523	
49	MTR	790	50	0.2661	0.22
197	613			0.2667	
Average					1.33

As can be seen from Tables 15 and 16, the comparisons of identical impact tests at different facilities are extremely close. Even including one set of 304L tests that were within 9.45% (evidencing material variability), the average variations are very low. Therefore, material impact tests performed at either location can be combined together since the ITM responses are deemed to be facility independent.

### 6.6.2 Longitudinal Versus Transverse Specimen Orientation

One of the initial goals of the material impact testing was to compare impact responses of test specimens that were fabricated reflecting longitudinal (parallel to plate rolling) and transverse (perpendicular to plate rolling) orientations. The first effort was to perform “quasi-static” tensile testing which was previously described in Section 5.0. Both 304L (Heat #10W8) and 316L (Heat #09T9) material were tested. The testing consisted of three specimens from each orientation for both materials, resulting in twelve total tests. Table 17 contains the tensile material property results of that testing.

Table 17. Longitudinal and transverse quasi-static tensile test results.

Test ID	Specimen ID	Material	Heat Number	Orientation	Elongation (%)	Ultimate Load (lb.)
03	MT03	304L	10W8	Longitudinal	63	22,100
04	MT04	304L	10W8	Longitudinal	64.5	22,200
23	MT23	304L	10W8	Longitudinal	62	22,000
05	MT05	304L	10W8	Transverse	62.5	22,000
06	MT06	304L	10W8	Transverse	63	22,200
24	MT24	304L	10W8	Transverse	59	21,800
11	MT11	316L	09T9	Longitudinal	61	20,500
12	MT12	316L	09T9	Longitudinal	61	20,500
21	MT21	316L	09T9	Longitudinal	59	20,400
09	MT09	316L	09T9	Transverse	58	20,700
10	MT10	316L	09T9	Transverse	60.5	20,700
22	MT22	316L	09T9	Transverse	58	20,500

As can be seen, the ultimate load and the elongation results indicate very little difference between the longitudinal and the transverse-oriented test specimens for each material.

The next step was to perform impact tests, using the same material as indicated above. Identical test weights and drop heights were used during the actual impact tests, using both the longitudinal and the transverse-oriented test specimens described in Section 6.2. For 304L, the longitudinal specimens were color coded red and the transverse specimens were color coded light yellow. For 316L, the longitudinal specimens were color coded black and the transverse specimens were color coded white. Drop weights and drop heights were chosen to maximize test specimen strains, purposely challenging floor flexibility in order to maximize any potential facility difference. The results of that comparative testing are summarized in Table 18. The results show that the variations between longitudinally-oriented test specimens and transverse-oriented test specimens are insignificant, less than a 3% variation on average. This is well within the range of material property variations previously discussed. Therefore, based on these comparative results of both quasi-static and impact tensile testing, the decision was made to proceed with material impact testing at varying temperatures with longitudinally oriented test specimens only.

## **6.7 Post-Test Dimensional Measurements**

Post-test dimensional measurements were needed to determine the deformation changes so that achieved strains could be calculated for each impact tested specimen.

### **6.7.1 Test Specimen Deformations**

After impact testing, appropriate test specimen deformation measurements were taken by NSNFP test personnel. These measurements included the deformed lengths between the punch marks as well as the reduced widths and thicknesses at the same punch locations measured during pre-test measurements. This characterized the post-impact geometry of each test specimen. These measurements were recorded on page 1 of the test data sheet for pre- and post-impact deformation measurements (as representatively shown in Appendix D and contained in the NSNFP Lab Notebook Binder Volumes 2 and 3). With this post-test dimensional data, strains could be calculated and recorded on page 2 of the test data sheet along with the acceleration and velocity traces previously discussed in Section 6.5.2.

After impact testing, the welded material test specimens exhibited the ‘orange peel’ effect, as did the quasi-static tensile test specimens (see Section 5.5.2). Figure 59 illustrates a welded material test specimen before impact testing (left) and a welded material test specimen after impact testing (right). The rough surface (due to the straining of the irregular grain structure resulting from welding) made post-test width and thickness measurements more difficult to achieve.

[illegible]



### **6.7.2 Certified Dimensional Inspector Validations**

The NSNFP test personnel measured the post-test impact test specimens. As a means of checking the post-test measurements and to add validity to all dimensional measurements performed by NSNFP test personnel, an INL qualified dimensional inspector also measured the deformed test specimens. Page 3 of the test specimen data sheets contains the results of those independent measurements. NSNFP Lab Notebook Binder Volumes 2 and 3 contain the test data sheets and these confirmatory measurements. All dimensions measured by the INL dimensional inspector validated the NSNFP test personnel measurements. None of the post-test measurements differed by more than 0.75% (INL compared to NSNFP test personnel). Appendix F contains qualification documentation for the INL dimensional inspector.

The INL qualified dimensional inspector also measured the go/no go gages periodically or on an as-needed basis when newer devices completed fabrication and needed to be verified before use. This activity was previously discussed in Section 6.2.1. Documentation of the measurements made by the INL qualified dimensional inspector on the go/no go gages is contained in the NSNFP Lab Notebook Binder Volume 1.

## **6.8 Closeout Calibration**

Calibrated measuring and test equipment were used for measuring all quality-affecting data. In order to demonstrate the calibration history of the measuring and test equipment used, the Report CD-R (file 3) contains the INL Calibration and Standards Laboratory calibration sheets for all of the devices used for the time intervals when the device was used for this strain rate research effort. The applicable measuring and test equipment used in this strain rate research effort also received a closeout calibration. The closeout calibration indicated all measuring and test equipment were acceptable.



Figure 59. 'Orange peel' effect in welded material impact test specimen (right).

## **7. TEST RESULTS: DEVELOPMENT OF FACTORS AND STRAIN RATE ELEVATED TRUE STRESS-STRAIN CURVES**

The data needed to justify the definition of material responses considering strain rate effects were obtained using the ITM. By using a combination of different weights and different drop heights, varying strain levels and strain rates were achieved. The energy input in straining the test specimen was known from the input test variables of mass (drop weight and impact driver) and impact velocity (of drop weight). The total test energy also included the work done elongating the test specimen. At a given test temperature, specimen displacement histories were measured. These test inputs and outputs are used to determine test specimen strain, rate of straining, and a 'factor' (derived from the test energy) that can be applied to the appropriate quasi-static true stress-strain curve to generate a corresponding strain rate elevated true stress-strain curve.

A brief literature search (References 40 and 41) indicated that the shape of strain rate elevated true stress-strain curves for stainless steels relative to the quasi-static shape are similar. Two potential options that preserve this curve shape are (1) shifting the curve upward (the addition of a constant to each point) or (2) factoring (multiplication of each stress point by a constant). Both techniques utilize the established shape of the quasi-static curve as determined from tensile tests as the starting point. Since the literature was inconsistent as to a specific shape of strain rate elevated curves, variations in curve shape were initially addressed by evaluating both a factored and a shifted technique (Reference 42). These early evaluation efforts (up to the uniform strain limit) indicated less than a 3% difference between the two techniques. The 'factored' approach was the simplest for structural analysts to apply since the area under the quasi-static stress-strain curve did not have to be determined as it would be if using the shifted approach. One only need to multiply the stress values by the strain rate factor to attain the strain-rate elevated stress-strain curve. Therefore, the 'factored' approach was adopted for this study.

### **7.1 Development of Factors**

Each unique dynamic impact test performed results in a strain rate 'factor' corresponding to the particular test material, heat, specimen geometry, temperature, and strain rate achieved in the test. By performing a sufficient number of impact tests using various weights dropped from a variety of heights, a plot of the strain rate factor versus strain rate can be generated for a particular type of material, heat, and temperature.

A total impact energy approach was used that considers the test specimen strain energy. Strain energy is the energy required to strain (deform) a volume of material a specified amount. It is equivalent to the area under the stress-strain curve up to a specified strain level. Remember that the goal of the ITM testing was to limit strains to below the uniform strain limit. For this approach, the quasi-static true stress-strain curve is determined and the area under that curve up to the strain level of interest is evaluated. The strain level of interest is that strain achieved in an actual impact test specimen at a given strain rate. It is acknowledged that energy losses may exist in the transfer of energy from the drop weight to the specimen. However, losses are believed small. Knowing the energy input to the specimen during the impact test and assuming the shape for the strain rate elevated true stress-strain curve being generated, the corresponding true stress

value can be established by bounding the input energy density (area under the stress-strain curve) to the value of true strain achieved in the impact test.

The energy input to the specimen is determined by applying conservation of momentum theory to the total impact energy assuming an inelastic impact. The total impact energy applied to specimen deformation is calculated as follows:

$$E_{total} = \frac{1}{2}mv^2 + w\Delta$$

where:  $\frac{1}{2}mv^2$  is the kinetic energy of the combined mass (drop weight + impact driver) and  $w\Delta$  is the potential energy due to deformation of the specimen.

$m =$  mass =  $m_{drop\ weight} + m_{impact\ driver}$

$w =$  weight =  $w_{drop\ weight} + w_{impact\ driver}$

$\Delta =$  elongation of test specimen

$v =$  velocity of the combined mass (drop weight + impact driver)

$v =$   $(m_{drop\ weight})(v_{impact}) / m$  (conservation of momentum)

$v_{impact} =$  impact velocity of the drop weight

$= \sqrt{2gh}$  where  $g$  is the acceleration of gravity and  $h$  is the drop height

The total impact energy approach assumes that the area under the strain rate elevated true stress-strain curve (up to the strain value achieved in the test specimen) equals the total impact energy per material volume used to dynamically strain the test specimen during the ITM impact test. The units for total impact energy are in.-lb. The area under a material stress-strain curve has the units of psi (stress) per in./in. (strain), or in.-lb. per cubic inch, or strain energy density. The area under the strain rate elevated true stress-strain curve up to the magnitude of the overall average strain achieved in the test specimen is determined by dividing the total impact energy ( $E_{total}$ ) by the material volume absorbing the impact energy (volume of test specimen material strained).

The volume of material strained during the dynamic impact test was the specimen gauge length volume plus a portion of the transition region volume. Because all strain energy density needed to be addressed, it was also necessary to determine the distribution of strain in the test specimen transition regions. Strains would not be expected to be uniform or of a magnitude equal to the overall average strain in the gauge length. They would likely vary from the overall average strain value at the gauge end of the transition to near zero at the large end of the transition region. The transition region strain distribution and material volumes were determined using dynamic simulations of the ITM impact tests in the computer program ABAQUS/Explicit. Finite element models of the test specimens used and the associated ITM impact driver mass were generated. Test specimen material properties were based on the quasi-static true stress strain curves developed for each particular material and heat. The resulting transition region strain distribution and material volume results from these computer simulations were incorporated into the procedure for determining test specimen strain energy densities and strain rate factors (see Report CD-R, file 2B).

The actual process of determining the factor values involves an iterative integration of the true stress-strain curves. At a given magnitude of overall average strain, the applicable (material,

heat and temperature dependent) quasi-static true stress strain curve must be elevated such that the strain energy density represented by the area under the resulting curve matches the energy density from the impact test (input energy and volume of material strained). In this process the magnitude of overall average strain being used must match the magnitude achieved in the impact test. The factor is then determined as the ratio of the area under the elevated curve to that under the quasi-static curve both determined up to the same magnitude of the overall average strain. The corresponding strain rate is that resulting directly from the impact test being evaluated.

This process was accomplished using the computer program DADiSP. Knowing the quasi-static stress-strain curve shape and the area required under the curve at a given strain level, the factor value can be 'estimated', the resulting stress-strain curve plotted and integrated to see if the area matches that required. An improved estimate is made of the factor value, if necessary, and the process iterates to a final, acceptable value.

### **7.1.1 Development of Spreadsheets**

In practice, a 'reverse' process was employed for quickly determining the factor for a given set of three impact test parameters: (1) the total impact energy ( $E_{\text{total}}$ ), (2) the overall average strain achieved (OAS), and (3) the gauge volume (GV). The total energy ( $E_{\text{total}}$ ) is defined above in Section 7.1. The overall average strain (OAS) is the average true strain through the gauge length resulting from the impact test and is determined from the measured data recorded on the specimen data sheets following the test. The gauge volume (GV) is the volume of material within the gauge length strained as a result of the impact test and is also determined from the test specimen measured and recorded data. These three impact test parameters are presented in Tables 19 through 22 for the 304L impact tests at the temperatures considered and Tables 23 through 26 for the 316L impact tests at the temperatures considered. In these tables, the 'True Strain' column is OAS, the 'Total Energy' column is  $E_{\text{total}}$ , and the 'Test Specimen Volume' column is GV.

Rather than directly calculating the factor for each unique set of test parameters using the iterative integration approach (a time consuming effort), a series of worksheets and spreadsheets were developed for a broad range of specified factors and overall average strains that bounded those expected from the actual impact testing. The iterative integration approach was performed only once for each set of unique material properties. This 'pre-solving' approach saved time. Using this set of worksheets and spreadsheets and a given set of three impact test parameters, the corresponding factor could be quickly 'interpolated' from the bounding results. A discussion of the development of the worksheets and spreadsheets is presented in the following paragraphs.

For a given material type (304L or 316L), heat (base and welded material), and temperature, a series of factored true stress-strain curves is generated using the appropriate digitized quasi-static true stress-strain curve and a range of bounding factors and overall average strains. For this work, factors ranged from 1.0 to 2.0 in six increments and overall average strain values ranged from 0.01 to 0.45 in nine increments. A DADiSP worksheet was employed for this effort. The quasi-static true stress-strain curve is first factored. The area under the resulting true stress-strain curve up to the limit of each overall average strain value is determined for each factor. The resulting set of values ( $6 \times 9 = 54$ ) of 'area under the curve' (units of in.-lb./in.<sup>3</sup>) and corresponding true overall average strain values are input to a Material Curve Excel spreadsheet

(see upper left hand corner of Figure 60). The DADiSP worksheets used for the work performed in this report are contained on the Report CD-R (file 2A).

Table 19. Selected impact test parameters and factors for 304L at -20 °F conditions.

Test ID	Test Specimen Geometry	% Diff.	True Strain Rate (sec. <sup>-1</sup> )	True Strain	Test Specimen Identifier	Material Heat Number	Total Energy (in.-lb.)	Test Specimen Volume (in. <sup>3</sup> )	Factor
432	D-2-2	-0.5	8.9	0.1168	304L-261	64A1	2273	0.1887	1.436
434	D-2-2	-3.2	11.4	0.2377	304L-262	64A1	5066	0.1875	1.214
445	D-2-2	-0.9	7.5	0.1190	304L-289	64A1	2327	0.1878	1.443
436	D-2-2	-2.5	5.5	0.0813	304L-263	64A1	1466	0.1872	1.447
453	D-2-2	-4.1	5.4	0.0613	304L-292	64A1	1179	0.1893	1.582
455	D-2-2	-5.4	11.3	0.1784	304L-293	64A1	3559	0.1878	1.285
449	D-2-2	-3.8	13.4	0.2286	304L-264	64A1	4963	0.1860	1.269
489	D-2-2	-2.6	15.4	0.2648	304L-311	64A1	5962	0.1851	1.231
438	D-4-4	-2.9	23.7	0.1496	304L-235	72K9	14793	0.7599	1.556
442	D-4-4	-3.9	29.5	0.2019	304L-288	72K9	22294	0.7545	1.537
443	D-4-4	-1.6	28.4	0.3209	304L-290	72K9	36872	0.7584	1.252
457	D-4-4	-4.7	19.9	0.1596	304L-294	72K9	17024	0.7635	1.627
459	D-4-4	-0.9	28.9	0.2677	304L-295	72K9	29456	0.7638	1.316
477	D-4-4	-3.4	19.4	0.1203	304L-296	72K9	12643	0.7593	1.794
496	D-4-4	-2.5	35.7	0.3166	304L-298	72K9	37656	0.7569	1.308
451	D-2-2	-1.5	9.7	0.1474	304L-291	485896	2951	0.1872	1.293
462	D-2-2	-5.1	10.0	0.1860	304L-303	485896	4073	0.1911	1.268
466	D-2-2	-4.2	10.9	0.1265	304L-304	485896	2467	0.1914	1.297
470	D-2-2	-1.5	15.3	0.2590	304L-305	485896	5931	0.1803	1.202
485	D-2-2 W	-3.4	10.8	0.1243	304L-308	485896	2457	0.1896	1.437
490	D-2-2 W	-3.6	10.5	0.1451	304L-309	485896	2940	0.1911	1.404
464	D-4-4	-3.2	23.1	0.1825	304L-300	54M7	17128	0.7545	1.422
468	D-4-4	-4.7	22.6	0.1472	304L-301	54M7	13193	0.7620	1.469
481	D-4-4	-6.5	23.5	0.1209	304L-312	54M7	11481	0.7698	1.652
487	D-4-4 W	-2.3	20.8	0.1220	304L-284	54M7	13109	0.7602	1.518
491	D-4-4 W	-1.4	22.5	0.1556	304L-285	54M7	17005	0.7521	1.477

Table 20. Selected impact test parameters and factors for 304L at room temperature conditions.

Test ID	Test Specimen Geometry	% Diff.	True Strain Rate (sec. <sup>-1</sup> )	True Strain	Test Specimen Identifier	Material Heat Number	Total Energy (in.-lb.)	Test Specimen Volume (in. <sup>3</sup> )	Factor
63	A	2.0	21.8	0.2536	304L-37	10W8	32448	1.1399	1.343
64	A	0.7	22.8	0.2753	304L-38	10W8	37278	1.1457	1.376
65	A	-6.1	24.9	0.3054	304L-39	10W8	42171	1.1615	1.335
66	A	-5.9	26.0	0.3438	304L-40	10W8	47602	1.1556	1.287
69	A	-9.4	25.2	0.3242	304L-43	10W8	44528	1.1556	1.305
134	A	-3.1	24.3	0.3971	304L-48	10W8	57402	1.1412	1.284
135	A	-1.7	26.4	0.4138	304L-105	10W8	60274	1.1547	1.257
136	A	1.1	27.5	0.4282	304L-106	10W8	63128	1.1511	1.257
165	A	-3.6	25.8	0.3051	304L-124	10W8	42163	1.1543	1.345
167	A	-4.3	25.9	0.3305	304L-125	10W8	47479	1.1520	1.360
172	A	-2.3	24.9	0.3161	304L-126	10W8	44442	1.1331	1.375
250	A-2-2	-2.3	5.9	0.1191	304L-180	64A1	3023	0.2858	1.307
253	A-2-2	2.7	5.9	0.1501	304L-181	64A1	3928	0.2849	1.275
247	D-2-2	-2.5	12.8	0.2439	304L-183	64A1	4805	0.1920	1.219
249	D-2-2	-0.4	13.3	0.2841	304L-184	64A1	5830	0.1926	1.200
252	D-2-2	-2.0	11.2	0.2797	304L-185	64A1	5678	0.1911	1.203
259	D-2-2	0.0	10.6	0.1908	304L-189	64A1	3616	0.1911	1.274
262	D-2-2	-3.8	10.6	0.2236	304L-190	64A1	4381	0.1899	1.261
294	D-2-2	-4.3	5.3	0.0766	304L-204	64A1	1242	0.1905	1.366
305	D-2-2	-1.6	10.9	0.2745	304L-207	64A1	5310	0.1869	1.179
312	D-2-2	-1.8	9.6	0.1348	304L-209	64A1	2335	0.1869	1.307
246	D-4-4	-0.9	27.2	0.2691	304L-175	72K9	26423	0.7695	1.456
248	D-4-4	-0.8	27.4	0.2696	304L-176	72K9	26589	0.7692	1.462
251	D-4-4	-3.1	26.2	0.3117	304L-177	72K9	31434	0.7680	1.416
254	D-4-4	-2.2	29.2	0.3658	304L-178	72K9	37194	0.7794	1.317
258	D-4-4	-1.1	27.1	0.2785	304L-179	72K9	27874	0.7737	1.458
260	D-4-4	-1.9	27.0	0.2793	304L-186	72K9	29285	0.7734	1.527
261	D-4-4	-2.4	26.8	0.2910	304L-187	72K9	30008	0.7716	1.473
288	D-4-4	-4.3	24.5	0.2163	304L-202	72K9	19624	0.7674	1.456
289	D-4-4	-3.6	23.7	0.2026	304L-201	72K9	17922	0.7653	1.454
290	D-4-4	-3.2	22.1	0.1936	304L-203	72K9	17179	0.7707	1.469
301	D-4-4	-6.3	32.6	0.3329	304L-215	72K9	33299	0.7647	1.373
302	D-4-4	-4.1	29.3	0.3081	304L-216	72K9	29669	0.7572	1.377
313	D-4-4	-4.0	24.0	0.1806	304L-217	72K9	16492	0.7560	1.573
314	D-4-4	2.1	22.4	0.1578	304L-218	72K9	14821	0.7719	1.650
315	D-4-4	-2.6	29.1	0.2380	304L-219	72K9	22427	0.7629	1.472
316	D-4-4	0.8	27.0	0.1629	304L-220	72K9	14788	0.7563	1.611
317	D-4-4	-2.7	21.6	0.1201	304L-221	72K9	9880	0.7608	1.575
348	D-2-2	1.3	10.5	0.1460	304L-225	485896	2531	0.1833	1.191
355	D-2-2	-3.0	9.9	0.1720	304L-244	485896	3299	0.1935	1.203
360	D-2-2	-2.7	10.1	0.2227	304L-245	485896	4299	0.1890	1.155
379	D-2-2 W	-0.1	9.6	0.1621	304L-272	485896	3250	0.1926	1.301
403	D-2-2 W	-3.2	8.2	0.1306	304L-275	485896	2479	0.1800	1.374
413	D-2-2 W	-3.6	8.1	0.1156	304L-276	485896	2430	0.1917	1.469
349	D-4-4	-0.7	23.9	0.1824	304L-226	54M7	14349	0.7317	1.469
354	D-4-4	-1.0	21.9	0.1302	304L-227	54M7	9905	0.7485	1.535
356	D-4-4	0.6	23.0	0.2118	304L-248	54M7	17147	0.7470	1.411
381	D-4-4 W	-1.4	22.4	0.1834	304L-273	54M7	17016	0.7506	1.321
401	D-4-4 W	0.7	20.3	0.1528	304L-274	54M7	14245	0.7377	1.398



Table 21. Selected impact test parameters and factors for 304L at 300 °F conditions.

Test ID	Test Specimen Geometry	% Diff.	True Strain Rate (sec. <sup>-1</sup> )	True Strain	Test Specimen Identifier	Material Heat Number	Total Energy (in.-lb.)	Test Specimen Volume (in. <sup>3</sup> )	Factor
274	D-4-4	1.2	34.7	0.2961	304L-196	72K9	21555	0.7656	1.258
285	D-4-4	-0.4	25.3	0.2785	304L-200	72K9	19928	0.7701	1.256
327	D-4-4	-0.7	27.5	0.2570	304L-208	72K9	18043	0.7692	1.267
329	D-4-4	-3.3	24.1	0.2167	304L-222	72K9	14713	0.7587	1.310
331	D-4-4	-1.7	22.5	0.1876	304L-223	72K9	12697	0.7746	1.336
339	D-4-4	-4.1	19.2	0.1181	304L-229	72K9	7476	0.7530	1.451
340	D-4-4	-0.5	23.9	0.1441	304L-230	72K9	9139	0.7446	1.396
341	D-4-4	0.4	22.6	0.1269	304L-233	72K9	8297	0.7557	1.467
376	D-4-4	-3.4	21.6	0.2118	304L-232	72K9	14463	0.7674	1.312
321	D-2-2	-3.1	10.6	0.1806	304L-210	64A1	2493	0.1863	1.191
324	D-2-2	-0.1	8.5	0.1575	304L-211	64A1	2185	0.1869	1.238
367	D-2-2	-8.4	5.3	0.1238	304L-239	64A1	1684	0.1881	1.282
370	D-2-2	-2.6	5.3	0.1083	304L-240	64A1	1185	0.1869	1.069
383	D-2-2	-2.9	6.7	0.1089	304L-242	64A1	1264	0.1893	1.120
385	D-2-2	-7.6	4.4	0.0770	304L-243	64A1	893	0.1896	1.219
386	D-2-2	-1.8	9.1	0.1861	304L-260	64A1	2542	0.1860	1.171
397	D-2-2	0.6	10.6	0.2003	304L-214	64A1	2757	0.1863	1.154
388	D-2-2	-1.9	10.4	0.2064	304L-246	485896	3103	0.1848	1.164
396	D-2-2	-2.8	9.1	0.1717	304L-268	485896	2639	0.1851	1.245
398	D-2-2	-1.3	10.9	0.1775	304L-270	485896	2640	0.1917	1.156
428	D-2-2 W	-1.7	9.9	0.1761	304L-283	485896	2965	0.1947	1.265
430	D-2-2 W	-0.8	10.9	0.1599	304L-281	485896	2578	0.2019	1.189
377	D-4-4	-2.1	22.5	0.2382	304L-251	54M7	15230	0.7443	1.308
405	D-4-4	-1.7	22.5	0.1936	304L-254	54M7	12203	0.7464	1.374
407	D-4-4	-0.9	21.7	0.1370	304L-255	54M7	7921	0.7368	1.409
512	D-4-4	-1.4	18.6	0.1636	304L-249	54M7	10032	0.7488	1.400
513	D-4-4 W	1.1	21.0	0.1562	304L-286	54M7	12073	0.7548	1.369
516	D-4-4 W	1.1	20.7	0.1917	304L-321	54M7	13898	0.7512	1.242

Table 22. Selected impact test parameters and factors for 304L at 600 °F conditions.

Test ID	Test Specimen Geometry	% Diff.	True Strain Rate (sec. <sup>-1</sup> )	True Strain	Test Specimen Identifier	Material Heat Number	Total Energy (in.-lb.)	Test Specimen Volume (in. <sup>3</sup> )	Factor
333	D-4-4	-0.5	22.9	0.2147	304L-224	72K9	12477	0.7704	1.228
335	D-4-4	-0.3	21.6	0.1332	304L-228	72K9	6913	0.7566	1.292
417	D-4-4	-0.4	21.4	0.2536	304L-234	72K9	14666	0.7593	1.169
509	D-4-4	-3.4	16.8	0.1533	304L-299	72K9	8353	0.7698	1.281
297	D-2-2	3.9	4.7	0.1123	304L-205	64A1	1160	0.1899	1.095
344	D-2-2	-1.9	10.9	0.2481	304L-212	64A1	2966	0.1881	1.014
347	D-2-2	-1.2	8.2	0.1776	304L-213	64A1	2165	0.1875	1.151
361	D-2-2	-11.2	4.6	0.1098	304L-236	64A1	1099	0.1887	1.074
364	D-2-2	-4.4	5.0	0.1065	304L-237	64A1	1041	0.1890	1.056
365	D-2-2	-6.6	9.9	0.2305	304L-238	64A1	2755	0.1860	1.050
410	D-2-2	-3.0	10.7	0.1913	304L-241	64A1	2211	0.1899	1.054
389	D-2-2	-0.7	8.4	0.2006	304L-247	485896	2612	0.1848	1.062
391	D-2-2	-3.2	9.7	0.2505	304L-266	485896	3320	0.1836	1.023
392	D-2-2	-3.5	8.1	0.1931	304L-267	485896	2725	0.1896	1.135
408	D-2-2	-2.4	9.5	0.2456	304L-269	485896	3548	0.1911	1.079
412	D-2-2	-3.5	9.8	0.1476	304L-271	485896	1819	0.1965	1.022
424	D-2-2 W	-4.8	9.2	0.1443	304L-279	485896	1811	0.1944	1.120
506	D-2-2 W	1.4	8.6	0.1805	304L-282	485896	2165	0.1896	1.039
395	D-4-4	-2.3	21.0	0.1514	304L-253	54M7	7759	0.7731	1.262
415	D-4-4	-1.4	21.7	0.2170	304L-258	54M7	12168	0.7587	1.255
419	D-4-4	0.1	21.5	0.2587	304L-259	54M7	14692	0.7599	1.194
423	D-4-4 W	-2.9	20.3	0.1991	304L-278	54M7	12104	0.7299	1.146
507	D-4-4 W	1.1	19.1	0.1978	304L-320	54M7	11591	0.7314	1.103

Table 23. Selected impact test parameters and factors for 316L at -20 °F conditions.

Test ID	Test Specimen Geometry	% Diff.	True Strain Rate (sec. <sup>-1</sup> )	True Strain	Test Specimen Identifier	Material Heat Number	Total Energy (in.-lb.)	Test Specimen Volume (in. <sup>3</sup> )	Factor
437	D-4-4	-2.5	23.7	0.1565	316L-115	67K0	14657	0.7494	1.367
439	D-4-4	-1.9	29.2	0.2145	316L-125	67K0	21705	0.7494	1.350
441	D-4-4	-2.8	20.4	0.1066	316L-126	67K0	9824	0.7599	1.459
444	D-4-4	-2.9	28.3	0.3365	316L-180	67K0	35635	0.7635	1.185
458	D-4-4	0.7	19.7	0.1528	316L-184	67K0	14423	0.7527	1.381
460	D-4-4	-1.9	31.8	0.3011	316L-185	67K0	32198	0.7548	1.262
478	D-4-4	-2.0	16.7	0.1074	316L-186	67K0	9784	0.7551	1.448
497	D-4-4	-1.1	39.1	0.4101	316L-188	67K0	45885	0.7506	1.171
431	D-2-2	-2.7	9.7	0.1104	316L-111	76H3	2252	0.1941	1.349
433	D-2-2	-3.1	8.6	0.1652	316L-160	76H3	3606	0.1902	1.334
435	D-2-2	-5.4	6.5	0.1042	316L-161	76H3	1904	0.1896	1.250
446	D-2-2	-2.9	5.3	0.0770	316L-164	76H3	1445	0.1881	1.378
456	D-2-2	-1.3	10.8	0.1678	316L-183	76H3	3510	0.1884	1.284
448	D-2-2	-5.4	13.0	0.2223	316L-163	76H3	5050	0.1866	1.295
450	D-2-2	-2.1	14.7	0.2702	316L-162	76H3	6220	0.1881	1.218
452	D-2-2	-2.9	11.3	0.1723	316L-181	48R8	3298	0.1830	1.147
465	D-2-2	-5.0	10.1	0.1463	316L-195	48R8	2947	0.1863	1.238
473	D-2-2	-4.5	10.0	0.1934	316L-197	48R8	4115	0.1896	1.194
461	D-2-2	-4.6	10.7	0.1225	316L-194	48R8	2293	0.1836	1.216
475	D-2-2	-3.5	17.2	0.3062	316L-198	48R8	6994	0.1821	1.148
492	D-2-2 W	-3.2	10.0	0.1166	316L-199	48R8	2273	0.1929	1.347
486	D-2-2 W	0.0	10.2	0.1392	316L-174	48R8	2914	0.2007	1.340
463	D-4-4	-5.2	23.7	0.2148	316L-191	230468	18687	0.7554	1.384
467	D-4-4	-4.7	24.5	0.1708	316L-192	230468	13912	0.7485	1.437
471	D-4-4	-6.8	21.5	0.1444	316L-193	230468	11912	0.7641	1.517
482	D-4-4	-5.3	19.4	0.1183	316L-201	230468	9875	0.7482	1.672
484	D-4-4	-1.2	23.1	0.1336	316L-203	230468	10713	0.7374	1.566
488	D-4-4 W	2.0	21.0	0.1306	316L-178	230468	11865	0.7191	1.427
493	D-4-4 W	0.0	22.9	0.1927	316L-179	230468	18578	0.7266	1.376

Table 24. Selected impact test parameters and factors for 316L at room temperature conditions.

Test ID	Test Specimen Geometry	% Diff.	True Strain Rate (sec. <sup>-1</sup> )	True Strain	Test Specimen Identifier	Material Heat Number	Total Energy (in.-lb.)	Test Specimen Volume (in. <sup>3</sup> )	Factor
23	A	-2.9	24.1	0.3743	316L-7	09T9	49097	1.1286	1.292
24	A	-4.3	25.9	0.3878	316L-8	09T9	51928	1.1525	1.274
25	A	-4.0	24.4	0.3788	316L-9	09T9	50302	1.1448	1.284
26	A	-3.5	24.0	0.3710	316L-10	09T9	47867	1.1102	1.295
27	A	0.4	25.8	0.3660	316L-11	09T9	47325	1.1079	1.308
28	A	-5.5	25.2	0.3493	316L-12	09T9	44386	1.1115	1.305
30	A	-5.6	14.0	0.1293	316L-14	09T9	13856	1.1421	1.451
31	A	-8.4	16.9	0.1647	316L-15	09T9	17824	1.1534	1.364
32	A	-2.9	18.4	0.1975	316L-16	09T9	21780	1.1151	1.365
51	A	-6.3	26.2	0.3043	316L-25	09T9	38433	1.1358	1.337
75	A	-4.3	25.8	0.3843	316L-28	09T9	50361	1.1457	1.258
76	A	-2.9	27.2	0.2877	316L-29	09T9	35273	1.1489	1.309
164	A	-0.4	26.5	0.3472	316L-48	09T9	44378	1.1205	1.305
169	A	-1.2	24.3	0.3300	316L-50	09T9	42758	1.1228	1.346
170	A	-1.7	21.8	0.2688	316L-51	09T9	31579	1.0998	1.338
184	A	-1.5	25.8	0.3929	316L-54	09T9	50441	1.0926	1.279
192	A	0.0	27.1	0.3523	316L-55	09T9	44407	1.1142	1.287
229	A	-1.1	20.4	0.2510	316L-61	09T9	29325	1.1016	1.359
239	A-2-2	-5.4	7.3	0.1910	316L-75	76H3	5313	0.2768	1.287
240	A-2-2	-0.4	6.0	0.1508	316L-76	76H3	3936	0.2808	1.267
241	A-2-2	-2.0	4.9	0.1186	316L-77	76H3	2981	0.2786	1.301
237	D-2-2	-3.8	8.5	0.2038	316L-71	76H3	3844	0.1869	1.256
238	D-2-2	1.3	10.6	0.2639	316L-72	76H3	5241	0.1860	1.226
242	D-2-2	0.1	10.3	0.1968	316L-73	76H3	3644	0.1839	1.265
293	D-2-2	3.3	12.0	0.2249	316L-78	76H3	4246	0.1845	1.236
295	D-2-2	-2.6	5.9	0.0812	316L-79	76H3	1262	0.1878	1.280
303	D-2-2	1.3	5.7	0.0908	316L-97	76H3	1513	0.1920	1.313
306	D-2-2	-3.5	9.3	0.1323	316L-98	76H3	2326	0.1929	1.270
230	D-4-4	-3.4	29.2	0.3422	316L-62	67K0	32526	0.7455	1.292
231	D-4-4	-2.7	30.5	0.3532	316L-63	67K0	34796	0.7500	1.316
232	D-4-4	-3.4	28.0	0.2913	316L-64	67K0	28413	0.7536	1.392
233	D-4-4	-2.6	30.2	0.3322	316L-65	67K0	32356	0.7533	1.326
276	D-4-4	-5.5	33.4	0.3017	316L-83	67K0	27125	0.7347	1.298
277	D-4-4	0.9	32.3	0.3031	316L-84	67K0	27131	0.7347	1.290
278	D-4-4	-2.0	29.7	0.3204	316L-85	67K0	28557	0.7353	1.258
279	D-4-4	-3.8	28.9	0.3128	316L-86	67K0	28521	0.7404	1.290
280	D-4-4	3.0	27.6	0.3095	316L-87	67K0	28503	0.7431	1.304
291	D-4-4	-0.8	22.0	0.1811	316L-91	67K0	14550	0.7446	1.344
292	D-4-4	0.0	23.3	0.2032	316L-94	67K0	16757	0.7368	1.348
304	D-4-4	-2.4	29.3	0.3910	316L-95	67K0	36038	0.7404	1.197
307	D-4-4	-1.9	27.5	0.2290	316L-103	67K0	19849	0.7473	1.349
308	D-4-4	0.3	22.8	0.1481	316L-104	67K0	13834	0.7479	1.643
309	D-4-4	-0.9	23.7	0.1828	316L-105	67K0	15226	0.7476	1.385
310	D-4-4	-1.3	22.4	0.1713	316L-106	67K0	14709	0.7446	1.459
311	D-4-4	0.1	29.0	0.2428	316L-107	67K0	21808	0.7524	1.364
318	D-4-4	-2.5	21.7	0.1138	316L-108	67K0	9864	0.7371	1.649
350	D-2-2	-1.9	10.2	0.1491	316L-127	48R8	2701	0.1947	1.285
357	D-2-2	-2.6	9.8	0.1928	316L-128	48R8	3391	0.1917	1.177
359	D-2-2	-3.6	10.4	0.2627	316L-136	48R8	4728	0.1848	1.129
380	D-2-2 W	-1.5	10.8	0.1914	316L-165	48R8	3385	0.1854	1.313
404	D-2-2 W	-2.9	11.9	0.1581	316L-168	48R8	2732	0.1950	1.278
414	D-2-2 W	-6.0	9.6	0.1505	316L-169	48R8	2707	0.1923	1.364
351	D-4-4	-0.4	24.6	0.1973	316L-137	230468	14323	0.7728	1.351
352	D-4-4	1.6	21.4	0.1760	316L-138	230468	12178	0.7467	1.381
353	D-4-4	0.0	20.5	0.1381	316L-139	230468	8624	0.7359	1.361
358	D-4-4	-2.0	21.8	0.2069	316L-142	230468	14671	0.7524	1.331
382	D-4-4 W	0.1	20.4	0.1729	316L-166	230468	14514	0.7422	1.306
402	D-4-4 W	1.3	21.8	0.1470	316L-167	230468	12080	0.7461	1.311

Table 25. Selected impact test parameters and factors for 316L at 300 °F conditions.

Test ID	Test Specimen Geometry	% Diff.	True Strain Rate (sec. <sup>-1</sup> )	True Strain	Test Specimen Identifier	Material Heat Number	Total Energy (in.-lb.)	Test Specimen Volume (in. <sup>3</sup> )	Factor
284	D-4-4	-1.3	23.3	0.2589	316L-88	67K0	17964	0.7467	1.213
286	D-4-4	-0.3	21.9	0.2331	316L-89	67K0	15964	0.7395	1.245
287	D-4-4	-0.8	22.5	0.2313	316L-90	67K0	15490	0.7416	1.217
328	D-4-4	0.3	25.4	0.2271	316L-112	67K0	15385	0.7551	1.216
330	D-4-4	-0.2	21.6	0.1689	316L-113	67K0	11995	0.7587	1.369
337	D-4-4	-2.3	25.4	0.1590	316L-120	67K0	10378	0.7500	1.290
338	D-4-4	-0.9	23.0	0.1258	316L-121	67K0	8495	0.7380	1.432
511	D-4-4	-2.3	16.6	0.1312	316L-189	67K0	8018	0.7554	1.255
320	D-2-2	-0.5	9.7	0.1610	316L-99	76H3	2424	0.1911	1.266
322	D-2-2	-1.8	11.4	0.2156	316L-100	76H3	3224	0.1932	1.156
323	D-2-2	-0.6	10.4	0.1745	316L-101	76H3	2553	0.1929	1.196
366	D-2-2	-2.9	4.0	0.0978	316L-133	76H3	1198	0.1887	1.165
368	D-2-2	-4.4	4.9	0.1144	316L-134	76H3	1628	0.1890	1.307
369	D-2-2	4.2	5.5	0.1043	316L-135	76H3	1302	0.1896	1.166
384	D-2-2	-2.2	4.1	0.0694	316L-158	76H3	862	0.1890	1.274
387	D-2-2	-1.0	9.6	0.1981	316L-159	76H3	2950	0.1872	1.213
325	D-2-2	-4.4	9.0	0.1614	316L-116	48R8	2300	0.1821	1.184
399	D-2-2	-3.7	9.4	0.1776	316L-153	48R8	2640	0.1869	1.176
400	D-2-2	-0.3	10.7	0.2250	316L-154	48R8	3312	0.1878	1.086
429	D-2-2 W	-1.6	10.3	0.1697	316L-177	48R8	2614	0.2010	1.150
495	D-2-2 W	-3.6	11.2	0.2064	316L-200	48R8	2931	0.1899	1.071
342	D-4-4	-2.8	23.6	0.1683	316L-118	230468	9102	0.7491	1.228
378	D-4-4	2.0	21.7	0.2654	316L-145	230468	16128	0.7398	1.208
406	D-4-4	2.1	22.5	0.2216	316L-147	230468	12621	0.7401	1.202
515	D-4-4 W	-2.8	24.8	0.1871	316L-205	230468	12499	0.7257	1.236
514	D-4-4 W	-0.2	20.9	0.2153	316L-206	230468	14014	0.7098	1.197

Table 26. Selected impact test parameters and factors for 316L at 600 °F conditions.

Test ID	Test Specimen Geometry	% Diff.	True Strain Rate (sec. <sup>-1</sup> )	True Strain	Test Specimen Identifier	Material Heat Number	Total Energy (in.-lb.)	Test Specimen Volume (in. <sup>3</sup> )	Factor
296	D-2-2	0.2	5.0	0.0940	316L-92	76H3	965	0.1947	1.017
300	D-2-2	2.2	10.2	0.2384	316L-96	76H3	3144	0.1929	1.023
343	D-2-2	-1.7	9.7	0.2020	316L-102	76H3	2668	0.1941	1.071
345	D-2-2	-5.4	11.7	0.2042	316L-129	76H3	2576	0.1899	1.041
346	D-2-2	-1.3	9.5	0.1734	316L-130	76H3	2149	0.1896	1.073
362	D-2-2	-8.7	5.1	0.1196	316L-131	76H3	1323	0.1902	1.054
363	D-2-2	-10.0	5.3	0.1018	316L-132	76H3	1156	0.1890	1.136
319	D-4-4	-0.6	23.6	0.1780	316L-109	67K0	10028	0.7485	1.171
332	D-4-4	-0.6	20.9	0.1873	316L-119	67K0	10788	0.7506	1.177
334	D-4-4	0.3	23.2	0.1517	316L-114	67K0	8060	0.7593	1.138
418	D-4-4	-1.8	20.7	0.2269	316L-124	67K0	13132	0.7596	1.105
390	D-2-2	-1.4	9.7	0.2292	316L-140	48R8	2982	0.1977	1.034
409	D-2-2	-1.3	9.1	0.2205	316L-155	48R8	2839	0.1875	1.091
411	D-2-2	-1.4	10.7	0.2041	316L-157	48R8	2497	0.1869	1.066
425	D-2-2 W	-0.1	9.0	0.1781	316L-173	48R8	2405	0.1956	1.137
505	D-2-2 W	-2.5	8.3	0.1943	316L-176	48R8	2349	0.1965	0.991
394	D-4-4	0.5	22.4	0.2396	316L-146	230468	12171	0.7461	1.138
416	D-4-4	-1.3	21.9	0.1666	316L-148	230468	7998	0.7281	1.242
420	D-4-4	-0.3	20.5	0.2705	316L-152	230468	13817	0.7206	1.131
508	D-4-4 W	-0.8	19.1	0.2011	316L-204	230468	11139	0.7365	1.053
510	D-4-4 W	-2.2	21.4	0.2010	316L-171	230468	11472	0.7341	1.088



factored data						
strain	Area per curve factor					
	1.20	1.40	1.60	1.80	2.00	1.00
0.01	403	459	515	571	627	347
0.05	2840	3302	3763	4225	4687	2378
0.1	6743	7856	8969	10081	11194	5631
0.200	16922	19731	22540	25349	28158	14113
0.250	23028	26855	30682	34509	38336	19202
0.300	29742	34687	39633	44578	49524	24796
0.350	37034	43195	49356	55517	61678	30873
0.400	44890	52361	59832	67302	74773	37420
0.450	53295	62166	71037	79908	88780	44424

Based on 304L MT153\_TER\_BaseA in LabBook Fdevelop  
70 deg F 304L H# 54M7 Yellow




Figure 60. Excel Material Curve spreadsheet, 304L, heat 54M7, ½-inch base material at room temperature.

Material curve equations are developed next in an Excel spreadsheet using the 'area under the curve' ( $\text{in.-lb./in.}^3$ ) and overall average strain values determined in the DADiSP worksheets for the specified factors and strain limits. These material curves are material type (304L or 316L), heat (base and welded material), and temperature specific. The Material Curve Excel spreadsheet for 304L, heat 54M7, base material, at 70°F is shown in Figure 60. In the Excel spreadsheet, the 54 'area under the curve' values are input versus factor and overall average strain values and are then plotted [area (y coordinate) versus true strain (x coordinate)] and fit with a 2<sup>nd</sup> order polynomial. The equation for the curve is then determined for each factor. This equation provides the relationship between the area under the factored true stress strain curve (energy density) up to a given strain limit. The Material Curve spreadsheets used for the work performed herein are contained on the Report CD-R (file 2D).

The strain distribution through the transition regions was based on the overall average strain in the gauge length and the ABAQUS/Explicit predictions by applying the ratio of the overall average strain to the corresponding value used in the computer simulations. The volume of transition region material strained to these distributions was also based on the ABAQUS/Explicit model that incorporated nominal specimen dimensions. An Excel spreadsheet was used to determine the strain and volume distributions from ABAQUS/Explicit output coordinates and coordinate displacements. The spreadsheets for the specimen geometries used for the work reported herein are contained on the Report CD-R (file 2D).

The resulting material curve equations are input into a final Excel spreadsheet (see Figure 61 for a representative example) to determine the specific factor value for a given strain rate impact test using the overall average strain (OAS), gauge volume (GV), and test total impact energy ( $E_{\text{total}}$ ) parameters. The final spreadsheet used to calculate the elevated strain rate factor for a specific impact test is material type (304L or 316L), heat (base and welded material), temperature, and geometry specific. The spreadsheet is material, heat, and temperature specific because it utilizes the derived material curve equations to determine appropriate energies. The spreadsheet is geometry dependent because it utilizes the ABAQUS/Explicit predicted strain and material volume distributions in the transition regions to determine appropriate energies. Representative spreadsheets for all the materials, temperatures, and specimen geometries used for the work performed herein are provided on the Report CD-R (file 2D). Also contained on the Report CD-R (file 2D) are the calculated factors. NSNFP Lab Notebook Binder Volume 7 also contains the resulting factors along with details of the associated test conditions and parameters. These factors are presented in this report in Tables 19 through 22 for the 304L impact tests at the temperatures considered and Tables 23 through 26 for the 316L impact tests at the temperatures considered.

### **7.1.2 Checking of Spreadsheets**

Once the Excel spreadsheets were developed, they were thoroughly checked in accordance with NSNFP Procedure 19.01 for verification of routines and macros within exempt commercial software. The spreadsheets were assigned the unique identifier NSNF/MED/017, Revision 1 by the NSNFP Document Control Coordinator and were submitted to NSNFP Document Control. Checking and verification of the spreadsheets is documented on the Signature Sheet contained in Appendix B. Verification of the DADiSP routines used in this process is documented in Appendix A.




Figure 61. Excel Spreadsheet for Final Factor Determination, Test #356, 304L, Heat 54M7, ½-Inch Base Material at 70°F.

## 7.2 Strain Rate Elevated True Stress-Strain Curves

For each successful strain rate test, an associated strain rate factor was determined as described in Section 7.1. For each of the materials (304L or 316L) investigated, the factors were plotted versus strain rate and fitted with a linear curve. These plots are shown in Figures 62 through 69 for the two materials and four temperatures tested. The results shown in these figures for a particular material and temperature include all material heats, both base and weld metal, and various specimen geometries (A and D profiles in 1/2-inch and 1/4-inch thicknesses). Given the relatively limited strain rate range studied to date, a linear fit of all of the material and temperature related results was considered to best trend the material data. It is recognized that, at this time, the majority of the linear fit curves do not pass through the coordinate point (0.0, 1.0), representing a factor of one at a strain rate of zero (quasi-static conditions), as would be expected. This may be attributed to a lack of data at higher strain rates and/or a need for a higher order curve fit. With the data currently limited to strain rates of just 40 per second or lower (depending on temperature), a higher order fit is no easier justified and the linear fit is adequate at this time for trending. The data point (0.0, 1.0) was added to the strain rate data when establishing the linear curve fits shown in the figures.

For a particular material (304L or 316L) and chosen temperature, the strain rate elevated true stress-strain curve can be developed up to the uniform strain limit from the factor predicted by the linear curve fit at the strain rate of interest. Using the corresponding material and temperature quasi-static true stress-strain curve, the strain rate elevated curve is developed by multiplying the appropriate factor times the true stress coordinates of the quasi-static true stress-strain curve.

Strain rate elevated true stress-strain curves for 304L and 316L materials developed using the linear curve fit relationships at various strain rates and temperatures are illustrated in Figures 70 through 79 (with strains up to their respective uniform strain limit). The 304L curves shown were based on the quasi-static true stress strain curve for 1/2-inch thick, 304L, heat 54M7, base material at the indicated temperature. The 316L curves shown were based on the quasi-static true stress-strain curve for 1/2-inch thick, 316L, heat 230468, base material at the indicated temperature. The quasi-static true stress-strain curves for all materials, heats, and temperatures can be found on the Report CD-R (files 1A and 1B).

Figures 70 through 74 illustrate the strain rate elevated true stress-strain curves for 304L stainless steel at 0 (quasi-static results), 5, 10, 22, and 25 per second strain rates at -20, room, 300, and 600 °F temperatures respectively. Figures 75 through 79 illustrate the same information but for 316L material.




Figure 62. Factor versus strain rate curve at -20 °F for 304L.




Figure 63. Factor versus strain rate curve at room temperature for 304L.




Figure 64. Factor versus strain rate curve at 300 °F for 304L.




Figure 65. Factor versus strain rate curve 600 °F for 304L.




Figure 66. Factor versus strain rate curve at -20 °F for 316L.




Figure 67. Factor versus strain rate curve at room temperature for 316L.






Figure 68. Factor versus strain rate curve at 300 °F for 316L.




Figure 69. Factor versus strain rate curve at 600 °F for 316L.




Figure 70. 304L heat 54M7 true stress-strain curve at 22/second strain rate at varying temperatures.




Figure 71. 304L heat 54M7 true stress-strain curve at -20°F at varying strain rates.




Figure 72. 304L heat 54M7 true stress-strain curve at room temperature at varying strain rates.




Figure 73. 304L heat 54M7 true stress-strain curve at 300 °F at varying strain rates.




Figure 74. 304L heat 54M7 true stress-strain curve at 600°F at varying strain rates.




Figure 75. 316L heat 230468 true stress-strain curve at 22/second strain rate at varying temperatures.




Figure 76. 316L heat 230468 true stress-strain curve at -20°F at varying strain rates.




Figure 77. 316L heat 230468 true stress-strain curve at room temperature at varying strain rates.




Figure 78. 316L heat 230468 true stress-strain curve at 300 °F at varying strain rates.




Figure 79. 316L heat 230468 true stress-strain curve at 600 °F at varying strain rates.

Although this testing program did not investigate the entire strain rate range of interest (up to 300 per second), certain important insights can be observed by evaluating the findings to date. The strain rate elevated true stress-strain curves (Figures 70 through 79) clearly illustrate that increasing strain rates result in increased (higher strength) true stress-strain curves (factors are positive and increasing with increasing strain rate). It is expected that this trend would continue, possibly to a limit, as higher strain rates are investigated. However, actual magnitudes and rates of change beyond those considered herein cannot be quantified without further testing.

Comparing 304L versus 316L curves, the increased capacity to absorb impact energy appears more significant in the 304L than in the 316L material. Figures 70 and 75 illustrate how the strain rate elevated true stress-strain curves vary with temperature at a strain rate of 22 per second. These two figures both indicate that at the same strain rate, increasing temperature decreases the strain rate gain in strength. This is shown more clearly by the listing of factors (calculated from each appropriate curve fit) in Table 27 where the magnitudes decrease as the temperatures increase.

Table 27. Factors for specified strain rates.

<b>Strain rate (per sec.)</b>	<b>-20 °F</b>	<b>Room Temperature</b>	<b>300 °F</b>	<b>600 °F</b>
<b>304L Stainless Steel</b>				
5	1.333	1.235	1.166	1.043
10	1.361	1.278	1.210	1.094
22	1.428	1.381	1.316	1.217
25	1.445	1.407	1.342	1.247
<b>316L Stainless Steel</b>				
5	1.275	1.265	1.162	1.040
10	1.296	1.281	1.187	1.070
22	1.346	1.321	1.247	1.140
25	1.359	1.331	1.262	1.158



## **8. USING RESULTS: ABAQUS/EXPLICIT ANALYSES USING STRAIN RATE ELEVATED TRUE STRESS-STRAIN CURVES**

With the strain rate factors and strain rate elevated true stress strain curves developed (within the specified strain rate range), it is appropriate to investigate the actual validity of these impact test results. As discussed in Section 1 of this report:

*“The test data developed can be used to establish an analysis methodology that can then be applied in analytical simulations to more accurately predict the deformation and resulting material straining in the components being evaluated that are subject to dynamic, impulsive loads.”*

Forty finite element analyses were performed of various impact tests using the fully dynamic, inelastic analysis software ABAQUS/Explicit, Version 6.6-3. Material properties were input as either non-factored (the quasi-static true stress-strain curve) or factored true stress-strain curves (reflecting strain rate effects as quantified herein using the appropriate curve fit data). The true stress-strain curves input to each analysis reflected the proper material heat (either base or welded material) at the proper temperature, as described in Sections 5 and 6. For the factored true stress-strain input, two strain rate elevated true stress strain curves were input into the finite element model. The first true stress-strain curve had a zero strain rate designated and the second curve had an upper bound (typically 30 per second) strain rate designated. The factors applied were the factors resulting from the curve fits identified on the appropriate factor versus strain rate curves presented in Section 7, even if the factor was not unity at zero strain rate. Linear interpolation of the material properties for strain rates between these two bounding strain rate points was used.

Due to symmetry in two planes, a quarter-model of the test specimen was created (see Figure 80) with mass from the impact driver appropriately applied. Model restraint was provided by fixing the upper cross-member pin in space and by applying plane symmetry boundary restraints to the specimen and lower cross-member pin. To initiate a test simulation, the model mass representing the drop weight was given an initial velocity equal to the drop weight's impact velocity. The analyses were run for a time period of approximately 0.030 seconds, which was determined from the actual testing to be sufficient to capture the full downward motion of the test specimen. Nominal dimensions for the test specimen geometry were used (just like a typical design analysis approach), rather than test specimen specific dimensions based on pre-test measurements.

Table 28 provides a summary of all forty analyses performed, grouped by temperature and strain rate. Percent difference comparisons using non-factored and factored analysis results are made to the actual resulting gauge length axial deformations (for the temperatures considered) at the target strain rates of 10 and 22 per second. In Table 28, the plus values indicate over-prediction and the negative values indicate under-prediction. The Table 28 results clearly indicate that the strain rate adjusted (factored) material input yields more accurate analysis predictions than when just the quasi-static (non-factored) true stress-strain curves are used. Considerable error (over 40%) results when just the quasi-static true stress-strain curves are used, indicating that strain rate effects are real and significant. As expected, all of the analysis results using the non-factored input over-predicted the axial deformation. Without

elevating the true stress-strain curve, a higher strain prediction results in order to equate the impact test's strain energy density with the area under the defined stress-strain curve. The factored material input analyses had estimates above and below the actual deformation, anticipated considering material property and test specimen geometry variabilities.

In general terms, the analysis results show significant improvement for all temperatures except for the 600 °F results. Seven out of the ten 600 °F analyses performed had more accurate results using the factored input but three tests (#391, #508, and #394) had higher percent differences. The percentage variation at 600 °F between non-factored and factored input is reduced due to the lower magnitude of strain rate factors, especially at the strain rates of 5 and 10 per second. These associated factors are less than 1.10, clearly within the variability of plate material properties. Table 29 provides more detailed test and analysis comparison results for each impact test analytically considered, with the shaded portions reflecting the lower strain rate analysis results. Appendix G contains the names and dates of the ABAQUS/Explicit models used for this comparative evaluation. This information is being provided in accordance with NSNFP Procedure 19.03 (Reference 43). The strain rate data quantified herein has indeed provided an improved analysis methodology for dynamic, impulsive events.




Figure 80. ABAQUS/Explicit symmetric model of impact test specimen.

Table 28. Summary of comparisons of analytical results using non-factored and factored input to actual impact test deformations.

Temperature	Percent Difference Comparisons Between Actual Test Results and ABAQUS/Explicit Calculated Results (%)			
	Non-Factored		Factored	
	SR=10	SR=22	SR=10	SR=22
-20 °F	+24.8 to +40.7	+25.2 to +36.9	-6.2 to +10.6	-3.6 to +6.3
Room	+22.6 to +33.3	+21.4 to +34.8	-5.5 to +4.4	-4.5 to +4.0
300 °F	+11.4 to +34.7	+8.9 to +24.1	-7.8 to +12.7	-9.4 to -0.3
600 °F	+2.6 to +12.0	+2.0 to +12.4	-5.7 to +5.9	-7.6 to -1.8
Positive table values indicate over-predicted deformations and negative table values indicate under-predicted deformations, when compared to actual test results. SR – strain rate (sec <sup>-1</sup> )				

Table 29. Comparison of ABAQUS/Explicit predictions using non-factored and factored input to test deformations

Material	Temp. (F)	Strain Rate (sec. <sup>-1</sup> )	Specimen Geom.	Drop Weight (lb.)	Drop Height (in.)	Test No.	Test Specimen No.	Gauge Deform. (in.)	Material Test Used	Without Factors		
										ABAQUS Deform. (in.)	Percent Difference	A Def
304L	-20	9.7		1097	2.5	451	304L-291	0.4775	55			
		8.9		790	3.0	432	304L-261	0.3710	13			
		22.5		1097	17.5	491	304L-285	0.5060	183W	0.6929	36.9%	
		22.6		790	20.0	468	304L-301	0.4760	156	0.6184	29.9%	
		19.9		1097	17.5	457	304L-294	0.5210	106	0.7067	35.6%	
	Room	9.9		1097	2.75	355	304L-244	0.5660	47			
		9.6		790	3.0	312	304L-209	0.4340	2			
		22.4		1097	17.375	381	304L-273	0.6050	177W	0.7701	27.3%	
		23.9		790	21.625	349	304L-226	0.6005	153	0.7652	27.4%	
		22.1		1097	17.5	290	304L-203	0.6415	113	0.8648	34.8%	
	300	10.4		1097	2.375	388	304L-246	0.6865	50			
		10.6		790	3.0	321	304L-210	0.5955	5			
		20.7		1097	14.0	516	304L-321	0.6335	171W	0.7836	23.7%	
		22.5		790	18.188	405	304L-254	0.6420	146	0.7807	21.6%	
		21.6		1097	14.5	376	304L-232	0.7090	109	0.8797	24.1%	
	600	9.7		1097	2.375	391	304L-266	0.8550	53			
		10.7		790	2.5	410	304L-241	0.6325	8			
		19.1		1097	11.5	507	304L-320	0.6560	179W	0.7065	7.7%	
		21.7		790	18.0	415	304L-258	0.7285	151	0.8139	12.4%	
		21.4		1097	14.5	417	304L-234	0.8665	111	0.9289	7.2%	
316L	-20	10.1		1097	2.5	465	316L-195	0.4745	42			
		9.7		790	3.0	431	316L-111	0.3500	26			
		22.9		1097	19.0	493	316L-179	0.6390	181W	0.8003	25.2%	
		21.5		790	18.0	471	316L-193	0.4675	141	0.6218	33.0%	
		19.7		1097	14.75	458	316L-184	0.4965	125	0.6286	26.6%	
	Room	9.8		1097	2.75	357	316L-128	0.6380	33			
		9.3		790	3.0	306	316L-98	0.4250	17			
		20.4		1097	14.75	382	316L-166	0.5675	167W	0.7126	25.6%	
		21.4		790	18.25	352	316L-138	0.5765	138	0.6997	21.4%	
		22.0		1097	14.75	291	316L-91	0.5955	128	0.7335	23.2%	
	300	10.7		1097	2.5	400	316L-154	0.7575	35			
		9.7		790	3.0	320	316L-99	0.5255	19			
		20.9		1097	14.0	514	316L-206	0.7230	161W	0.8118	12.3%	
		22.5		790	18.688	406	316L-147	0.7440	131	0.8103	8.9%	
		21.9		1097	16.0	286	316L-89	0.7865	120	0.9208	17.1%	
	600	9.7		1097	2.125	390	316L-140	0.7735	37			
		9.5		790	2.5	346	316L-130	0.5690	22			
		19.1		1097	11.0	508	316L-204	0.6675	164W	0.6854	2.7%	
		22.4		790	17.875	394	316L-146	0.8120	135	0.8286	2.0%	
		20.7		1097	13.0	418	316L-124	0.7640	123	0.8073	5.7%	

## 9. OBSERVATIONS FROM TEST RESULTS

A total of 260 dynamic impact tensile tests of 304L and 316L stainless steels were used in this report to calculate strain rate factors for the strain rate range of 4 to 40 per second. More than 160 quasi-static tensile tests were also performed to establish stress-strain relationships of the materials used. This section discusses several observations related to this large population of tests.

### 9.1 Quasi-Static Tensile Testing

The quasi-static tensile testing performed for this research effort produced results that were correct, as evidence in Section 5.5. However, additional discussion is needed to more fully explain how the uniform strain limits were determined.

The text book explanation for determining the uniform strain limit of typical steels would indicate that it is the strain at which the tensile stress is at its maximum (ultimate tensile strength or UTS). However, with these stainless steels, the quasi-static engineering curves did not have a significant ‘hump’ shape (as seen in Figure 2) but were relatively flat for a significant strain range. This is illustrated in Figure 81.

Figure 81 illustrates how the maximum UTS value [the red dot that is less than 0.25% higher in strength than the adjacent blue (left) and black (right) square boxes] would lead the text book engineer to pick 0.43 engineering strain as the uniform strain limit. In reality, the onset of **significant necking** is actually closer to the black square box value of 0.47 engineering strain (if not a little higher). Typical accuracy tolerances for these tensile test machines is  $\pm 0.5\%$  of the load (which carries over to stress accuracy). Attempts to indicate that the maximum UTS can be so precisely determined within a few psi is not appropriate. Since this research effort was interested in accurately defining the strain energy density for each material at the onset of significant necking, the higher uniform strain limit value would have been chosen. In Section 5.5, Table 7 lists the uniform strain limits chosen for use in this research effort.




Figure 81. Engineering plot of 304L material exhibiting flat response near UTS.

## **9.2 Impact Tensile Testing**

A number of significant insights were gained from the impact testing performed for factor determination. A more thorough understanding of what strain rates are achievable with various test specimen geometries was a significant step forward for test planning purposes. However, two specific issues, (1) impact test anomalies and (2) base versus welded material responses, are pertinent to this report and are discussed in further detail below.

### **9.2.1 Impact Test Anomalies**

Ten impact tests were able to achieve strains beyond the established uniform strain limit (see Table 7) yet did not exhibit any signs of necking. Discarding those instances where that strain exceedance was 5% or less (attempting to determine a trend by focusing on extreme indicators), only three impact tests remain. With strains exceeding their established uniform strain limits by 8.7%, 8.7%, and 17.3%, these three 'high indicator' 316L tests (#326, #375, and #422) did not provide clear indications that the uniform strain limit increases with increasing strain rate.

On the other hand, there were a total of four impact tests where the strain achieved was below the established uniform strain limit and the test specimen either necked or broke. Again discarding those tests that had strains greater than 95% of their uniform strain limit (attempting to determine a trend by focusing on extreme indicators), two 316L tests remain [#427 (welded) and #469] that necked at strains of 90.8% and 77.5% of their established uniform strain limits. These 'low indicator' tests were surprises from the perspective that the material failed earlier than anticipated, especially Test #469 that necked at  $\frac{3}{4}$  of the uniform strain limit. However, there was no clear indication that the uniform strain limit decreases with increasing strain rate.

These 'high' and 'low' test anomalies highlight the probabilistic nature of testing that should not be unexpected. Even though material variability has an effect on the impact test results and the uniform strain limit, the impact testing performed herein at strain rates below 40 per second did not provide clear indications that the uniform strain limit varies with strain rate.

### **9.2.2 Base Versus Welded Material Responses**

Prior drop testing experience of full-scale Department of Energy spent nuclear fuel canisters (References 44, 45, and 46) indicated there was no significant variation in the deformation responses of the canister wall when impact occurred directly onto a canister weld. A limited number of impact tests of welded material at elevated temperatures were performed prior to the completion of that material's quasi-static tensile testing. These first impact tests used the same drop weight and drop height as was used for the base material. The resulting strain rate responses of the welded material test specimens appeared to be very similar to the base material. However, some of the welded material test specimens necked or broke where the base material test specimens had not. The conclusion reached was that the welded material had a lower uniform strain limit and fracture strain than the base material. After completion of the quasi-static tensile testing, this indeed turned out to be the case.

Figures 82 and 83 show, for 304L and 316L respectively, quasi-static tensile test results of base and welded materials at 300° F temperature conditions. These representative engineering stress-strain plots illustrate that the uniform strain limit and failure strain for the welded material

are lower than the associated base material. Figures 84 and 85 are comparative strain history plots of 304L and 316L base and welded material impact tests performed at -20 °F. These impact tests used the same drop weight, drop height, and test specimen geometry. These plots illustrate how similar the base and welded material were in terms of strain rate response. The plots also show that the welded material absorbs the impact energy with a lower maximum strain than the base material (i.e., welded material is stronger than base material). However, as discussed above, the welds fail at lower strain levels than the base material (i.e., welded material is less ductile than the base material).




Figure 82. Quasi-static tensile test results for 304L base and welded material at 300° F.




Figure 83. Quasi-static tensile test results for 316L base and welded material at 300° F.






Figure 84. Comparison of base and welded 304L material to identical impact tests at -20 °F.




Figure 85. Comparison of base and welded 316L material to identical impact tests at -20 °F.

Based on the strain rate range achieved herein, the welded material test specimens responded very similar to the base material test specimens. Therefore, one would expect the strain rate factors to be similar. This also was the case, as evidenced by the factor versus strain rate curves (Figures 62 through 69). As can be seen, the welds (identified as triangular-shaped data points) are consistent with the base material data. This permits the structural analyst to use the same strain rate factor data for both base and weld materials when incorporating strain rate effects into finite element models. However, the structural analyst must be fully aware that the welds have a lower uniform strain limit and failure strain and this must be correctly incorporated into the acceptance criteria that the structural analyst is employing.

### **9.3 Factors and Strain Rate Elevated True Stress-Strain Curves Insights**

It is important to recognize that the factor versus strain rate curves (Figures 62 through 69) reflect multiple heats and different test specimen geometries. Considering the variability of material properties of all of the test specimens used, the consistency of the curves is quite striking. However, it is too early to engage in discussions on how the formulation of these strain rate test results compare to a characterization of strain rate dependent material behavior (material constitutive equation such as the Cowper-Symonds equation) until strain rate data at higher ranges (up to 300 per second) become available.

## 10. CONCLUSIONS

Strain rate effects in the range of 4 to 40 per second for tensile loading of 304L and 316L stainless steel materials were successfully quantified at the INL using a large drop weight test machine and a total energy analysis approach. By incorporating the strain rate elevated true stress-strain material curves into an inelastic finite element computer program as the defined material input, significant improvement in the accuracy of the computer analyses was attained. However, additional impact testing is necessary to achieve higher strain rates (up to 300 per second) before complete definition of strain rate effects can be made for accidental drop events and other similar energy-limited impulsive loads.

The results of this material impact testing effort can be used to support the development of strain-based acceptance criteria for national codes and standards use. As documented by this research effort, in addition to quantifying the strain rate effects, the following items may prove useful to codes and standards developers: (1) the effects of strain rate decrease with increasing temperature, (2) base and welded materials appear to behave similarly during impact testing, though welded materials have lower uniform strain and failure strain limits, and (3) the uniform strain limits for both welded material and base material do not appear to vary from the values established during quasi-static tensile testing for the strain rate range discussed herein. In addition to the need for impact tensile testing research at higher strain rates, bending response impact testing would be very beneficial in establishing maximum viable thru-wall and surface strain limits.

This research approach, using impact testing and a total energy analysis methodology to quantify strain rate effects, can be applied to many other materials used in government and industry.

## 11. REFERENCES

The references listed below were the most current used for this research task but earlier versions in effect at the time of use were used as appropriate.

1. American Society of Mechanical Engineers, *Boiler and Pressure Vessel Code*, Section III, Division 1, Appendix F, 2007 Edition.
2. Department of Energy, *Quality Assurance Requirements and Description*, DOE/RW-0333P, Revision 17, May 8, 2006.
3. Department of Energy, National Spent Nuclear Fuel Program, *NSNFP Quality Assurance Program Plan*, DOE/SNF/QAPP-001, Revision 3, April 14, 2006.
4. Department of Energy, National Spent Nuclear Fuel Program, *Review and Approval of NSNFP Internal Documents*, NSNFP 6.01, Revision 12, August 1, 2005.
5. Department of Energy, National Spent Nuclear Fuel Program, Program Procedure, *Managing Document Control and Distribution*, NSNFP 6.03, Revision 2, October 22, 2004.
6. Department of Energy, National Spent Nuclear Fuel Program, Program Procedure, *Engineering Documentation*, NSNFP 3.04, Revision 6, May 12, 2007.
7. Department of Energy, National Spent Nuclear Fuel Program, Program Procedure, *Testing*, NSNFP 11.01, Revision 3, May 12, 2007.
8. Department of Energy, National Spent Nuclear Fuel Program, Program Procedure, *Acquiring Government Sector Services*, NSNFP 4.02, Revision 1, October 22, 2004.
9. Department of Energy, National Spent Nuclear Fuel Program, Program Procedure, *Personnel Indoctrination and Proficiency Training*, NSNFP 2.04, Revision 8, June 29, 2007.
10. Department of Energy, National Spent Nuclear Fuel Program, Program Procedure, *Software Control*, NSNFP 19.01, Revision 4, August 1, 2005.
11. Department of Energy, National Spent Nuclear Fuel Program, *Task Management Agreement for the Canister RemoteWeld/Repair System*, DOE/SNF/TMA-009, Revision 1, February, 2005.
12. Department of Energy, National Spent Nuclear Fuel Program, *Task Management Agreement for Transportation & Packaging Tasks to be Performed by the INL M&O Contractor*, DOE/SNF/TMA-013, Revision 1, February 2006.
13. Department of Energy, National Spent Nuclear Fuel Program, *Test Plans for the Department of Energy Spent Nuclear Fuel Canister and Basket Development Project*, DOE/SNF/PP-039, Revision 8, June 2007.
14. INL Welding Laboratory Notebooks #628 (No. 1), #684 (No. 2), and #969 (No.3), "FY2005 Welding and Weld Examination Services for Material Impact Testing" and associated files containing radiographs, radiograph reader sheets, project explanations, etc., (NSNFP Record).
15. ASTM A370-06, 2006, *Standard Test Methods and Definitions for Mechanical Testing of Steel Products*, ASTM International, Inc., West Conshohocken, PA.

16. National Spent Nuclear Fuel Program, *Planning/QA Program Applicability Evaluation*, PAE-010, Revision 4, September 28, 2006.
17. INL Laboratory Notebook LAB 771, "NSNFP Testing Alloy 3xx Stainless Steel", (NSNFP Record).
18. I-DEAS 10 NX Series, Unigraphics Solutions, Inc., 2002.
19. *ABAQUS/Explicit*, Version 6.6-3, ABAQUS, Inc., Providence, RI, 2006.
20. DADiSP 2002, The DADiSP Worksheet 2002, Data Analysis and Display Software, DSP Development Corporation, One Kendall Square, Cambridge, Ma., Version 6.0, NI NK B07, April 11, 2003.
21. Collection of Excel spreadsheets identified under NSNF/MED/017, Revision 1, (NSNFP Record).
22. Department of Energy, National Spent Nuclear Fuel Program, *Software Report For ABAQUS/Explicit Version 6.6-3*, DOE/SNF/REP-107, Revision 0, November 2006.
23. *ABAQUS Example Problems Manual*, Version 6.6, ABAQUS, Inc., Providence, RI, 2006.
24. ASM International, *Atlas of Stress-Strain Curves*, Material Park: ASM International, Second Edition, 2002.
25. P. W. Bridgman, *The Stress Distribution at the Neck of a Tension Specimen*, Transactions of the A. S. M., Twenty-fifth Annual Convention, Chicago, Illinois, October 18-22, 1943.
26. Norman E. Dowling, *Mechanical Behavior of Materials*, Upper Saddle River: Prentice-Hall Inc., Second Edition, 1999, Appendix B.
27. William D. Callister, Jr., *Materials Science and Engineering – An Introduction*, New York: John Wiley & Sons, Inc., 1985.
28. H. E. McCoy, Jr. and R. D. Waddell, *Mechanical Properties of Several Products from a Single Heat of Type 304 Stainless Steel*, Journal of Engineering Materials and Technology, October 1975, p. 343.
29. W. F. Simmons and H. C. Cross, *Report on the Elevated-Temperature Properties of Stainless Steels*, ASTM Spec. Tech. Pub. No. 124, 1952.
30. J. B. Conway, R. H. Stentz, and J. T. Berling, *Fatigue, Tensile, and Relaxation Behavior of Stainless Steels*, Technical Information Center, USAEC, 1975, p. 213.
31. *Bettis Plant Materials Manual*, Westinghouse Electric Corp., Standards Engineering Section, May 1957. As published in *Aerospace Structural Metals Handbook*, Vol. 2, Code 1303, CINDAS/USAF CRDA Handbooks Operation, Purdue University, 1995, P. 13.
32. R. K. Blandford, D. K. Morton, S. D. Snow, and T. E. Rahl, *Tensile Stress-Strain Results for 304L and 316L Stainless Steel Plate at Temperature*, PVP2007-26096, Proceedings of the 2007 ASME Pressure Vessels and Piping Conference, San Antonio, Texas, July 20-26, 2007.
33. American Society of Mechanical Engineers, *Boiler and Pressure Vessel Code*, Section II, Part A, Specification SA-240/SA-240M, 2004 Edition.
34. INL Weld Manual, Volumes 1 and 1A, Welding Procedure Specifications, Revision 48, April 18, 2005.

35. INL Procedure, *Radiographic Examination*, TPR-4970, Revision 1, January 27, 2005.
36. *LabVIEW Software*, Version 7.0, National Instruments, Inc., Austin TX, 2006.
37. Report CD-R (contained herein this report), generated March 2008.
38. NSNFP Laboratory Notebook NSNF/SN/04.01, "NSNFP Material Strain Rate Testing", Volumes I, II, and III and associated binders (Volumes 1 through 8), (NSNFP Record).
39. J. E. Fisher, *Thermal Analysis of NSNFP Strain Rate Test Coupon Heating Tests*, EDF-NSNF-065, Revision 0, August 22, 2006.
40. G. Solomos, et. al., *Strain Rate Effects of Nuclear Steels in Room and Higher Temperatures*, Transactions of SMiRT 16, Paper #1356, Washington, DC, August 2001.
41. Woei-Shyan Lee, et. al., *The Effects of Strain Rate and Welding Current Mode on the Dynamic Impact Behavior of Plasma-Arc-Welded 304L Stainless Steel Weldments*, Metallurgical and Materials Transactions A, Volume 35A, pp. 1501-1515, May 2004.
42. R. K. Blandford and T. E. Rahl, *FY 2004 Material Strain Rate Testing*, EDF-NSNF-042, Revision 2, April 19, 2005.
43. Department of Energy, National Spent Nuclear Fuel Program, Program Procedure, *Control of the Electronic Management of Information*, NSNFP 19.03, Revision 1, October 22, 2004.
44. D. K. Morton, S. D. Snow, and T. E. Rahl, *FY1999 Drop Testing Report for the 18-Inch Standardized DOE SNF Canister*, EDF-NSNF-007, Revision 2, September 5, 2002.
45. D. K. Morton and S. D. Snow, *Drop Testing Representative 24-Inch Diameter Idaho Spent Fuel Project Canisters*, EDF-NSNF-045, Revision 0, January 28, 2005.
46. D. K. Morton and S. D. Snow, *Drop Testing Representative Multi-Canister Overpacks*, EDF-NSNF-047, Revision 0, January 28, 2005.

## **Appendix A**

### **DADiSP Integration Verification**



## A1. OBJECTIVE


The objective of this appendix is to verify the integration function used in the computer code DADiSP for accuracy. Verification will be accomplished by solving two problems using DADiSP and comparing the DADiSP solutions with the known classical solutions.

## A2. KNOWN SOLUTION SAMPLE PROBLEMS

### A2.1 Area of a Triangle

The first sample problem chosen for which a known solution (solved by classical methods) exists is a triangle. The particular one used is shown graphically in Figure A1. The area is calculated using classical methods as follows:

$$Area = \frac{b \times h}{2} = \frac{1 \times 2}{2} = 1$$



**Figure A1. Sample Problem #1 – Triangle.**

Figure A2 is the problem defined in DADiSP with Figure A3 showing the execution of the integration function of DADiSP. Note the final integration value of 1 which agrees with the classical solution.




Figure A2. Problem #1 Definition in DADiSP.





Figure A3. DADiSP Integration Plot of Problem #1.

## A2.2 Area of a Quarter Circle

The second sample problem chosen for which a known solution (solved by classical methods) exists is a quarter circle of radius 2 inches. It is shown graphically in Figure A4. The area is calculated using classical methods as follows:

$$Area = \frac{\pi \times r^2}{4} = \frac{\pi \times 2^2}{4} = 3.14159$$




**Figure A4. Sample Problem #2 – Quarter Circle.**


Figure A5 is the problem defined in DADiSP with Figure A6 showing the execution of the integration function of DADiSP. Note the final integration value of 3.14096 as a percentage of the classical solution is:

$$\frac{3.14096}{3.14159} \times 100 = 99.98 \%$$

The data points of Figure A4 are tabulated in Table A1.



**Figure A5. Problem #2 Definition in DADiSP.**



**Figure A6. DADiSP Integration Plot of Problem #2.**

**Table A1. XY values of Figure A4.**

<b>x-value</b>	<b>y-value</b>
0	0
0.001218	0.069799
0.004872	0.139513
0.010956	0.209057
0.019464	0.278346
0.030384	0.347296
0.043705	0.415823
0.059409	0.483844
0.077477	0.551275
0.097887	0.618034
0.120615	0.68404
0.145632	0.749213
0.172909	0.813473
0.202412	0.876742
0.234105	0.938943
0.267949	1
0.303904	1.059839
0.341925	1.118386
0.381966	1.175571
0.423978	1.231323
0.467911	1.285575
0.51371	1.338261
0.56132	1.389317
0.610683	1.43868
0.661739	1.48629
0.714425	1.532089
0.768677	1.576022
0.824429	1.618034
0.881614	1.658075
0.940161	1.696096
1	1.732051
1.061057	1.765895
1.123258	1.797588
1.186527	1.827091
1.250787	1.854368
1.31596	1.879385
1.381966	1.902113
1.448725	1.922523
1.516156	1.940591
1.584177	1.956295
1.652704	1.969616
1.721654	1.980536
1.790943	1.989044
1.860487	1.995128
1.930201	1.998782
2	2

## **Appendix B**








### **Engineering Signoffs for Spreadsheet Calculations**

## Signature Sheet

### Verification of Excel Routines and Macros Identified in NSNF/MED/017, Rev. 1

The purpose of this signature sheet is to document verification of certain software routines and macros within exempt commercial software as required by NSNFP Procedure 19.01, Revision 4, Section III, A. The applicable routines and macros were generated using Microsoft Office Excel 2003 (11.8169.8172) SP3 software and are uniquely identified by the NSNFP Document Control Coordinator (DCC) as NSNF/MED/017, Revision 1.

The Excel software was used to generate spreadsheets containing equations, routines, and macros used to manipulate, convert, calculate, and present raw strain rate testing related data into meaningful results. The following 'verifier' signatures indicate independent verification by checking, inspection, and/or alternate calculation that the spreadsheet equations, routines, and macros calculate correct results:

IDENTIFIER	PERFORMER	VERIFIER SIGNATURE		DATE
		PRINT	SIGN	
Spreadsheets A_304L Material Static True Stress Strain	T.E. Rahl	R.K. Blandford		3/17/2008
Spreadsheets B_316L Material Static True Stress Strain	T.E. Rahl	R.K. Blandford		3/17/2008
Spreadsheets C_Calibration of Static Test Data	T. E. Rahl	R.K. Blandford		3/17/2008
Spreadsheets D_Strain Rate Specimen Strain Distribution	T. E. Rahl	R.K. Blandford		3/17/2008
Spreadsheets E_Material Curve Equations	T. E. Rahl	R.K. Blandford		3/17/2008
Spreadsheets F_Strain Rate Test Factor Evaluation	T. E. Rahl	R.K. Blandford		3/17/2008
Spreadsheets G_Strain Rate Test Summary Tables	D. K. Morton	R.K. Blandford		3/17/2008



## **Appendix C**

### **INL Welding Procedure S2.0**



IDAHO NATIONAL ENGINEERING LABORATORY  
WELDING PROCEDURE SPECIFICATION

WELDING PROCEDURE SPECIFICATION NUMBER: S2.0  
REVISION NUMBER: 11  
WELDING PROCESS(ES): GTAW

PAGE 1 of 2  
PQR: See Notes \*\*  
DATE: 03/24/92  
TYPE(S): Manual

SCOPE: This welding procedure specification must be used in conjunction with the Idaho National Engineering Laboratory's Welding Manual. This welding procedure specification meets the requirements of the ASME B & PV Code Section IX, 1986 edition up to and including A87 addenda. This specification is to be used for welding austenitic stainless steel.

JOINTS (QW-402)

DETAILS

JOINT DESIGN: Grooves and Fillets  
BACKING MATERIAL (TYPE):  
3XX stainless steel, if required

NONMETALLIC/NONFUSING METAL RETAINERS: None

See INEL Welding Manual

BASE MATERIALS (QW-403)

GENERAL CLASS: 3XX stainless steel

P NO.: B To P NO.: 8

SPECIFICATION TYPE AND GRADE:

TO SPECIFICATION TYPE AND GRADE:

CHEMICAL ANALYSIS AND MECH. PROP.:

TO CHEMICAL ANALYSIS AND MECH. PROP.:

THICKNESS RANGE:

BASE MATERIAL - GROOVE: 0.020 to 8.0 in. (note 2)

FILLET: All

PIPE DIA. RANGE:

BASE MATERIAL - GROOVE: All

FILLET: All

SINGLE PASS THICKNESS LIMIT: less than 1/2 in.

FILLER METALS (QW-404)

SPECIFICATION NO. (SFA): A/SFA 5.9

AWS NO. (CLASS): See note 1

F NO.: 6

A NO.: 8

NOMINAL CHEMICAL COMPOSITION:

SIZE OF FILLER METALS: 0.125 in. max.

DEPOSITED WELD METAL

THICKNESS RANGE:

GROOVE: 0.020 to 8.00 in.

FILLET: All

CONSUMABLE INSERT: None

SUPPLEMENTAL FILLER/POWDER: None

SOLID/TUBULAR/FLUX COVERED (GMAW): n/a

POSITION (QW-405)

POSITION(S) OF GROOVE: All

VERTICAL WELDING PROGRESSION: Up

POSITION(S) OF FILLET: All

PREHEAT (QW-406)

PREHEAT TEMP. (min.): 50 F

INTERPASS TEMP. (max.): 350 F

PREHEAT MAINTENANCE: As required



## WELDING PROCEDURE SPECIFICATION

WELDING PROCEDURE SPECIFICATION NUMBER: S2.0

REVISION NO.: 11

PAGE: 2 of 2

### POSTWELD HEAT TREATMENT (QW-407)

TEMP. RANGE: None

TIME RANGE: None

### GAS (QW-408)

SHIELDING GAS(ES): Argon

COMPOSITION: 100%

FLOW RATE: 10 to 25 cfh

GAS BACKING: Argon

COMPOSITION: 100%

FLOW RATE: 0.5 cfh min.

TRAILING GAS: None

COMPOSITION: None

FLOW RATE: None

### ELECTRICAL CHARACTERISTICS (QW-409)

CURRENT TYPE: Direct

POLARITY: Straight

AMPS: 10 to 200 \*

VOLTS: 8 to 16 \*

TUNGSTEN ELECTRODE SIZE AND TYPE: AWS A5.12 EWth-2 0.125 in. max.

MODE OF TRANSFER FOR GMAW: n/a

ELECTRODE WIRE FEED SPEED RANGE: n/a

PULSING: Optional

### TECHNIQUE (QW-410)

STRING OR WEAVE BEAD: Both

GAS CUP SIZE: 0.375 in. min.

INITIAL AND INTERPASS CLEANING: Clean with alcohol, acetone, or approved cleaners prior to welding. Grinding and/or wire brushing is permitted.

METHOD OF BACKGROUING: Grinding, filing, or machining

OSCILLATION: 0.50 in. max.

CONTACT TUBE TO WORK DISTANCE: n/a

MULTIPLE OR SINGLE PASS (per side): Either

MULTIPLE OR SINGLE ELECTRODES: Single

TRAVEL SPEED (ipm): 1 to 10

PEENING: Not allowed

### ADDITIONAL NOTES:

- 1) Suggested filler materials: For type 304 use 308 filler, 304L use 308L filler, 316 use 316 filler, 316L use 316L filler, 347 or 348 use 347 or 348 filler, 321 use 321 or 347 filler.
- 2) This procedure is qualified for welding metals requiring notch toughness testing in the base metal thickness range of 0.124 to 0.496 in.  
\* If charpy impact requirements are imposed, the maximum joules/in. shall be: pass 1 35.6 kJ/in.; pass 2: 39.3 kJ/in.; pass 3 and greater 43.1 kJ/in.. Welding must be approved by an authorized Welding Engineer.  
\*\* Q225, Q877, Q4144, Q6233, Q6802

COMPANY: EG&G IDAHO, INC.

PREPARED BY:

*Douglas D. Hansen*

DATE:

*3/25/92*

APPROVED BY:

*Mark J. Clark*

DATE:

*3/25/92*


## **Appendix D**

### **Representative Test Specimen Data Sheet**

### Data Sheet – ITM Tensile Test Coupon – Page 1

Coupon ID: 316L-103  
Material Orientation: (L) T (circle one)  
Drop Height: 2 ft 6 in  
Heat Number: 67k0  
Weld Number: n/a

Test Date: 7 / 12 / 06  
Pig Weight: 790 lbs  
Pre-Test Temp: 79 °F  
Caliper Offset: 0.40 in.  
Specimen Geometry: D44



Front Side Shown

Distance	Target Gage Length (inches)	Pre-Test Measured Gage Length (inches)	Pre-Test Actual Gage Length, L* (inches)	Post-Test Measured Gage Length (inches)	Post-Test Actual Gage Length* (inches)	Change in Gage Length, ΔL (inches)	Calipers Calibration ID#
A-B	1.000	.6070	1.0070	.8635	1.2635	.2565	721714
B-C	1.000	.6030	1.0030	.8620	1.2620	.2590	
C-D	1.000	.6110	1.0110	.8660	1.2660	.2550	
A-C	2.000	1.6035	2.0075	2.1200	2.5200	.5165	
B-D	2.000	1.6075	2.0075	2.1275	2.5275	.5200	
A-D	3.000	2.6090	3.0090	3.3835	3.7835	.7745	721714

\*-includes caliper points offset

Pt	Reduced Section Width				Reduced Section Thickness		Reduced Section Area (W x T)		Calipers Calibration ID#
	Pre Front	Pre Back	Post Front	Post Back	Pre	Post	Pre	Post	
B	.4820	.5005	.4305	.4485	.5060	.4495	.2486	.1976	721715
C	.4850	.5000	.4285	.4475	.5065	.4500	.2495	.1971	721715

Pre-measurements performed by: TE Roll  
Pre-measurements checked by: RK Blandford  
Post-measurements performed by: TE Roll  
Post-measurements checked by: SD Snow

Date: 7 / 10 / 06  
Date: 7 / 10 / 06  
Date: 7 / 12 / 06  
Date: 7 / 12 / 06

## Data Sheet – ITM Tensile Test Coupon – Page 2

Coupon ID : 316L-103

Accel FN : test307 071206



Accel. Serial # 38617

Cam FN : Test\_307

True strain rate from camera = 275 sec<sup>-1</sup>

See attached camera data for more details.

Distance	Target Gage Length (inches)	Engr. Strain, $\epsilon_L$ ( $\Delta L/L$ )	True Strain, $\epsilon_t$ $\ln(\Delta L/L + 1)$
A-B	1.000	nc	nc
B-C	1.000	nc	nc
C-D	1.000	nc	nc
A-C	2.000	nc	nc
B-D	2.000	nc	nc
A-D	3.000	.2574	.2290



Comments:

nc – not calculated for auxillary tests

Pt	Engr. Strain, $\epsilon_A$ $\left( \frac{A_{pre} - A_{post}}{A_{post}} \right)$	True Strain, $\epsilon_A$ $\ln \left( \frac{A_{pre}}{A_{post}} \right)$
B	nc	nc
C	nc	nc
ave	.2620	.2327

Measurements performed by: R. Blandford

Date: 07 / 11 / 07


Checked by: D. K. Morton

Date: 07 / 16 / 07

**Qualified Dimensional Inspector Data Sheet – ITM Tensile Test Coupon**  
**Page 3**

Coupon ID : 316L-103

Dimensional Inspector (print): Robert T Moloney



**Front Side Shown**

Distance	Gage Length (inches)	Post-Test Length (inches)	Calipers Calibration ID#
A-D	3.000	3.7795	721714 exp 9-26-07

Dimensional Inspector (sign):

RM

Date:

10 / 12 / 2006

## **Appendix E**

### **Drop Weight Comparison**



1. Weights measured and marked by NSNFP test personnel using calibrated load cells.
2. Shackle weight used for total weight represents average weight of both shackles.
3. Weight not used for qualified data testing.
4. Weights #1 thru #6 include two bolts.
5. Total weight is the sum of the calibration laboratory measurement 'Item Weight' and the 'Shackle & Link Weight.'
6. Nominal measured weight is the Total Weight rounded to the nearest whole number.

## **Appendix F**

### **INL Dimensional Inspector Qualifications**

414.73  
03/10/98  
Rev. 01

## PERSONNEL CERTIFICATION DATA FORM

### SECTION I – Request (To be filled out by Applicant's Supervisor)

Applicant: Robert B. Mohney	S#: 57562	Current certification expires: 10/20/00	Date: 09/20/00
Inspection activity candidate will be performing: Precision Dimensional Inspections.			

### SECTION II – Processing Documentation and Evaluations: (To be filled out by Level II and others as requested)

1. Discipline: Mechanical	Method: Precision Dimensional	Level: II	Inspector
Limits: None			

2. Physical/Vision Examination Ref MCP-535 Appendix I				Applicant's Education Level	
Due Date 08/02/01	A. Near Vision	B. Far Vision	C. Color Discrimination	<input type="checkbox"/> 4 Yr. Degree	<input type="checkbox"/> 2 Yr. Degree
Requirement	20/25 Snellen			<input type="checkbox"/> H.S./Equal	<input type="checkbox"/> Other
Corrected/Uncorrected	Corrected	Corrected	Verify contrast.	Note: See Inspectors Certification File	

3. Training: * Sub Exp. = additional experience permitted to substitute for formal training as specified in Appendix F or G					
MCP-535 Appendix B, F, or G	Class Hr by Education Level			Training Acquired (Ref MCP-535 Appendix D)	
Subject/Topic/Description	H.S.	+ 2 Yr.	Sub Exp.*	Hr.	Description
1. Drawing interpretation	4 hours	---	3 months		Experience continued since previous certification obtained as a Precision Dimensional Inspector.  See Certification File.
2. Dimensional metrology	20 hours	---	6 months		
3. Measuring hand tools	20 hours	---	6 months		
4. Geometric tolerancing	40 hours	---	1 year		
5. Coordinate measuring machine	40 hours	---	1 year		
6. Contour Projector	8 hours		4 months		
7. Optical Measuring Equipment	8 hours		4 months		

3a. OJT/Self Study:					
Type & Activity/Objective Required	Req. Hr	Date	Hr.	Reference Documentation and comments:	
1. Drawing interpretation	10 hours			See Certification File.	
2. Dimensional Metrology	20 hours				
3. Measuring Hand Tools	20 hours				
4. Geometric Tolerancing	40 hours				
5. Coordinate Measuring Machine	80 hours				
6. Contour Projector	10 hours				
7. Optical Measuring Equipment	10 hours				

4. Experience: Ref MCP-535 Appendix A, B, F or G					
Hr. Required		Required Experience		Experience Obtained	
H.S.	+ 2 Yr.	Other	Description	Hr.	Reference/Documentation
1 Year	6 Month			55 Hrs.	Application & Interpretation of Geometric Dimensioning & Tolerancing. (Tech. Doc. Consultants INC.)

5. Examination Results:							
TEST TYPE	INDIVIDUAL TEST SCORES						COMPOSITE SCORE
	Written Examinations & Min Number Questions*			Practical Examinations			
	Gen. / ( )	Spec. / ( )	Other 97/(117)	Demo	Oral	Other	
Description			G/D/&T – F	Door Support		G/D/&T – Q/H	
SCORES			82 %	97 %		96 %	92 %
DATE			01/18/2000	10/10/20000		01/18/2000	

\* The Minimum number of examination questions is reduced when the scope of qualification is LIMITED MCP-535 Appendix A & D. Min # questions/(actual # questions)

414.73  
03/10/98  
Rev. 01

## PERSONNEL CERTIFICATION DATA FORM

6.	Additional Training Required Prior to Re-examination:		Minimum up-date or re-certification training:	
	NONE		Hours	Subject Required
			Hours	Ref. Documentation

7. The "Applicant" is Certified in accordance with MCP-535 to perform the above Discipline – Method.			
Certifying		Effective Date of Certification:	Certification Expiration Date:
Level III: Michael J. Chaffin ,		10/10/2000	10/10/2003
Date: 10/10/2000			

8.	Certificate/Endorsement issued: (Initials & Date)	Entered into Database: (Initials & Date)
----	--	---

414.73  
02/20/2003  
Rev. 04  
Use with MCP-535

## BBWI PERSONNEL CERTIFICATION DATA FORM

### SECTION I – Request (To be filled out by Applicant's Supervisor)

Applicant: Robert B. Mohney	S#: 57562	Current certification expires: 10/10/2003	Date: 09/11/2003
Inspection activity candidate will be performing: Precision Dimensional Inspections.			

### SECTION II – Processing Documentation and Evaluations: (To be filled out by Level II and others as requested.)

1. Discipline: Mechanical	Method: Precision Dimensional	Level: II	Inspector
Limits: NONE			

2. Physical/Vision Examination Ref. MCP-535 Appendix I				Applicant's Education Level	
Due Date: 08/18/2004	A. Near Vision	B. Far Vision	C. Color Discrimination	<input type="checkbox"/> 4 Yr. Degree	<input type="checkbox"/> 2 Yr. Degree
Requirement	20/25 Snellen	N/A	Verify Contrast	<input type="checkbox"/> H.S./Equal	<input type="checkbox"/> Other
Corrected/Uncorrected	Corrected		OK	Note: See Inspectors Certification File	

3. Training: * Sub Exp. = additional experience permitted to substitute for formal training as specified in Appendix F or G					
MCP-535 Appendix B, F, or G		Class Hr. by Education Level		Training Acquired (Ref. MCP-535 Appendix D)	
Subject/Topic/Description	H.S.	+ 2 Yr.	Sub Exp.*	Hr.	Description
1. Drawing Interpretation	4 hours		1 year		Experience continued since previous certification obtained as a Precision Dimensional Inspector.  See Certification File.
2. Dimensional Metrology	20 hours		1 year		
3. Measuring Hand Tools	20 hours		2 years		
4. Geometric Tolerancing	40 hours		1 year		
5. Coordinate Measuring Machine	40 hours		1 year		
6. Contour Projector	8 hours		1 year		
7. Optical Measuring Equipment	8 hours		6 months		

3a. OJT/Self Study:					
Type & Activity/Objective Required	Req. Hr.	Date	Hr.	Reference Documentation and Comments:	
1. Drawing Interpretation	10 hours			See Certification File.	
2. Dimensional Metrology	20 hours				
3. Measuring Hand Tools	20 hours				
4. Geometric Tolerancing	40 hours				
5. Coordinate Measuring Machine	80 hours				
6. Contour Projector	10 hours				
7. Optical Measuring Equipment	10 hours				

4. Experience: Ref. MCP-535 Appendix A, B, F or G					
Hr. Required			Required Experience	Experience Obtained	
H.S.	+ 2 Yr.	Other	Description	Hr.	Reference/Documentation
1 Year	6 Month			55 Hrs.	Application & Interpretation of Geometric Dimensioning & Tolerancing. (Tech. Doc. Consultants INC.)
				40 Hrs.	Gage Mentor (CBT Training)

414.73  
 02/20/2003  
 Rev. 04  
 Use with MCP-535

## BBWI PERSONNEL CERTIFICATION DATA FORM

5. **Examination Results:**

TEST TYPE	INDIVIDUAL TEST SCORES						COMPOSITE SCORE
	Written Examinations & Min Number Questions*			Practical Examinations			
	Gen.25/(30)	Spec. /(\ )	Other /(\ )	Demo	Oral	Other	
Description	Written			Plate/Holes		Practical Demo	88 %
Scores	83 %			90 %		90 %	
Date	9/02/03			9/04/03		9/11/03	

\* The Minimum number of examination questions is reduced when the scope of qualification is LIMITED MCP-535 Appendix A & D. Min # questions/(actual # questions)

6. Additional Training Required Prior to Re-examination:	Minimum up-date or re-certification training:			
None	Hours	Subject Required		Hours
				Ref. Documentation

7. The "Applicant" is Certified in accordance with MCP-535 to perform the above Discipline – Method.

Effective Date of Certification:	Certification Expiration Date:	TRAIN Qualification Code:
09/11/03	10/10/2008	QLMECH02
Michael J. Chaffin		09/11/03
Principle Level III Examiner	Principle Level III Examiner	Date
Print/Type Name	Signature	

414.73  
01/28/2004  
Rev. 05  
Use with MCP-535 & MCP-1309

## PERSONNEL CERTIFICATION DATA FORM

### SECTION I – Request (To be filled out by Applicant's Supervisor)

Applicant: Robert B. Mohney	S#: 57582	Current certification expires: 10/10/08	Date: 10/12/08
-----------------------------	-----------	---	----------------

### SECTION II – Processing Documentation and Evaluations: (To be filled out by Level III)

1. Discipline: Precision Dimensional Inspector, 3 Year Re - Certification Level: II Inspector

2. Physical/Vision Examination Applicant's Education Level

Due Date 08/07	A. Near Vision	B. Far Vision	C. Color Discrimination	<input type="checkbox"/> 4 Yr. Degree	<input type="checkbox"/> 2 Yr. Degree
Requirement	20/25 Snellen		Verify Contrast	<input type="checkbox"/> H.S./Equal	<input type="checkbox"/> Other
Corrected/Uncorrected	Corrected	Un-Corrected	OK	Note:	

3. Training: <sup>1</sup> Sub Exp. = additional experience permitted to substitute for formal training as specified in MCP-535 & MCP1309.

Class Hr. by Education Level		Training Acquired	
Subject/Topic/Description	H.S. + 2 Yr. Sub Exp. <sup>1</sup>	Hr.	Description
See Previous Certifications			See Certification File

3a. OJT/Self Study:

Type & Activity/Objective Required	Req. Hr.	Date	Hr.	Reference Documentation and Comments:
N/A				See Certification File

4. Experience:

Hr. Required			Required Experience		Experience Obtained	
H.S.	+ 2 Yr.	Other	Description		Hr.	Reference/Documentation
			See Previous Certifications			See Certification File

5. Examination Results: <sup>2</sup> The Minimum number of examination questions is reduced when the scope of qualification is LIMITED MCP-535 & MCP1309. Min # questions/(actual # questions)

TEST TYPE	INDIVIDUAL TEST SCORES						COMPOSITE SCORE
	Written Examinations & Min Number Questions <sup>2</sup>			Practical Examinations			
	Gen. I( )	Spec. I( )	Other I( )	Demo	Oral	Other	
Description							N/A
Scores							
Date							

6. Additional Training Required Prior to Re-examination: Minimum up-date or re-certification training:

None	Hours	Subject Required	Hours	Ref. Documentation

7. The "Applicant" is Certified in accordance with MCP-535 & MCP1309 to perform the above Discipline – Method.

Effective Date of Certification: 10/10/08	Certification Expiration Date: 10/10/09	TRAIN Qualification Code: QLMECH02
Michael J. Chaffin Principle Level III Examiner Print/Type Name		10/12/08 Date
_____ Principle Level III Examiner Signature		

## **Appendix G**

### **Electronic File Information**



## **ELECTRONIC FILES**

The following electronic files associated with work performed in this report were generated and stored on 1 Digital Video Disk (DVD) using the +R DVD format. The DVD has the unique title "Electronic Files SR Test Comparisons". All ABAQUS/Explicit input (.inp) and output (.odb) files were transferred from PC (Govt. Property ID: 384252) to the DVD.

After the files were written to the DVD, the files were checked by accessing and visually checking the ABAQUS/Explicit input files and viewing the results in ABAQUS Viewer for the output files. This work was performed by the preparer (R. K. Blandford). The reviewer (S. D. Snow) performed random readability checks of the DVD.

The files were used for the work described in report Section 8 (USING RESULTS: ABAQUS/EXPLICIT ANALYSES USING STRAIN RATE ELEVATED TRUE STRESS-STRAIN CURVES) of this report and are listed below. The "Date Modified" indicates the date when the model file was last changed in any way and then saved to the identified PC hard drive and documents the final configuration as used in the analyses contained in this report.











































The file names on the DVD are descriptive with the associated acronyms describing various model details that were used in the particular input file. The significant acronyms identifying the input files model configuration are:

Typical file title        304L\_70\_T290\_D44\_72K9\_1097\_175\_FR.inp











































304L	material (304L, 316L)
70	temperature (-20, 70, 300, 600°F)
T###	strain rate test number
D44	test specimen geometry (an included W implies a welded material otherwise base material is implied)
72K9	material heat number (72K9, 64A1, 54M7, 485896, 67K0, 230468, 48R8, 76H3)
1097	test drop weight magnitude
175	a number descriptor for the test drop height magnitude




































175	17.5-inches	14	14-inches
3	3-inches	2125	21.25-inches
21625	21.625-inches	250	2.50-inches
275	2.75-inches	18	18-inches
17375	17.375-inches	1150	11.50-inches
1450	14.50-inches	20	20-inches
2375	23.75-inches	1825	18.25-inches
18188	18.188-inches	1475	14.75-inches
16	16-inches	18688	18.688-inches
17875	17.875-inches	13	13-inches
11	11-inches	19	19-inches

FR	a descriptor for the true stress strain curve
	FR or FRY implies a factored true stress-strain curve
	S implies a quasi-static true stress-strain curve (no factors applied)










































Name ^	Size	Type	Date Modified
 304L_70_T290_D44_72K9_1097_175_FR.inp	564 KB	INP File	1/30/2008 12:25 PM
 304L_70_T290_D44_72K9_1097_175_FR.odt	12,287 KB	ODT File	1/30/2008 6:07 PM
 304L_70_T290_D44_72K9_1097_175_S.inp	564 KB	INP File	1/23/2008 8:38 AM
 304L_70_T290_D44_72K9_1097_175_S.odt	12,275 KB	ODT File	1/23/2008 9:27 AM
 304L_70_T312_D22_64A1_790_3_FR.inp	256 KB	INP File	1/30/2008 3:41 PM
 304L_70_T312_D22_64A1_790_3_FR.odt	6,623 KB	ODT File	1/30/2008 5:15 PM
 304L_70_T312_D22_64A1_790_3_S.inp	256 KB	INP File	1/17/2008 3:35 PM
 304L_70_T312_D22_64A1_790_3_S.odt	6,590 KB	ODT File	1/21/2008 10:34 AM
 304L_70_T349_D44_54M7_790_21625_FR.inp	564 KB	INP File	1/30/2008 3:43 PM
 304L_70_T349_D44_54M7_790_21625_FR.odt	12,289 KB	ODT File	1/30/2008 5:13 PM
 304L_70_T349_D44_54M7_790_21625_S.inp	564 KB	INP File	1/23/2008 7:25 AM
 304L_70_T349_D44_54M7_790_21625_S.odt	12,285 KB	ODT File	1/23/2008 8:18 AM
 304L_70_T355_D22_4858_1097_275_FR.inp	257 KB	INP File	1/30/2008 8:35 AM
 304L_70_T355_D22_4858_1097_275_FR.odt	6,542 KB	ODT File	1/30/2008 10:23 AM
 304L_70_T355_D22_4858_1097_275_S.inp	257 KB	INP File	1/17/2008 2:53 PM
 304L_70_T355_D22_4858_1097_275_S.odt	6,596 KB	ODT File	1/21/2008 8:43 AM
 304L_70_T381_D44W_54M7_1097_17375_FR.inp	565 KB	INP File	1/30/2008 12:30 PM
 304L_70_T381_D44W_54M7_1097_17375_FR.odt	12,311 KB	ODT File	1/31/2008 8:50 AM
 304L_70_T381_D44W_54M7_1097_17375_S.inp	565 KB	INP File	1/23/2008 12:35 PM
 304L_70_T381_D44W_54M7_1097_17375_S.odt	12,256 KB	ODT File	1/23/2008 1:55 PM
 304L_300_T321_D22_64A1_790_3_FR.inp	257 KB	INP File	1/30/2008 2:10 PM
 304L_300_T321_D22_64A1_790_3_FR.odt	6,613 KB	ODT File	1/31/2008 10:54 AM
 304L_300_T321_D22_64A1_790_3_S.inp	256 KB	INP File	1/22/2008 3:26 PM
 304L_300_T321_D22_64A1_790_3_S.odt	6,515 KB	ODT File	1/23/2008 7:16 PM
 304L_300_T376_D44_72K9_1097_1450_FR.inp	564 KB	INP File	1/30/2008 2:14 PM
 304L_300_T376_D44_72K9_1097_1450_FR.odt	12,295 KB	ODT File	1/31/2008 12:53 PM
 304L_300_T376_D44_72K9_1097_1450_S.inp	564 KB	INP File	1/22/2008 4:37 PM
 304L_300_T376_D44_72K9_1097_1450_S.odt	12,278 KB	ODT File	1/24/2008 8:22 AM
 304L_300_T388_D22_4858_1097_2375_FR.inp	257 KB	INP File	1/30/2008 2:05 PM
 304L_300_T388_D22_4858_1097_2375_FR.odt	6,602 KB	ODT File	1/31/2008 10:39 AM
 304L_300_T388_D22_4858_1097_2375_S.inp	256 KB	INP File	1/17/2008 4:17 PM
 304L_300_T388_D22_4858_1097_2375_S.odt	6,595 KB	ODT File	1/21/2008 11:40 AM
 304L_300_T405_D44_54M7_790_18188_FR.inp	565 KB	INP File	1/30/2008 2:12 PM
 304L_300_T405_D44_54M7_790_18188_FR.odt	12,267 KB	ODT File	1/31/2008 12:59 PM
 304L_300_T405_D44_54M7_790_18188_S.inp	564 KB	INP File	1/22/2008 4:02 PM
 304L_300_T405_D44_54M7_790_18188_S.odt	12,259 KB	ODT File	1/24/2008 8:31 AM
 304L_300_T516_D44W_54M7_1097_14_FR.inp	564 KB	INP File	1/30/2008 12:32 PM
 304L_300_T516_D44W_54M7_1097_14_FR.odt	12,267 KB	ODT File	1/31/2008 8:51 AM
 304L_300_T516_D44W_54M7_1097_14_S.inp	564 KB	INP File	1/23/2008 3:47 PM
 304L_300_T516_D44W_54M7_1097_14_S.odt	12,284 KB	ODT File	1/28/2008 6:17 PM
 304L_600_T391_D22_4858_1097_2125_FRY.inp	257 KB	INP File	2/13/2008 12:30 PM
 304L_600_T391_D22_4858_1097_2125_FRY.odt	6,611 KB	ODT File	2/14/2008 8:09 AM



Name ▲	Size	Type	Date Modified
 304L_600_T391_D22_4858_1097_2125_SY.inp	257 KB	INP File	2/13/2008 12:31 PM
 304L_600_T391_D22_4858_1097_2125_SY.odb	6,597 KB	ODB File	2/14/2008 8:39 AM
 304L_600_T410_D22_64A1_790_250_FR.inp	257 KB	INP File	1/30/2008 2:18 PM
 304L_600_T410_D22_64A1_790_250_FR.odb	6,566 KB	ODB File	1/31/2008 5:17 PM
 304L_600_T410_D22_64A1_790_250_S.inp	257 KB	INP File	1/23/2008 8:12 AM
 304L_600_T410_D22_64A1_790_250_S.odb	6,596 KB	ODB File	1/24/2008 12:13 PM
 304L_600_T415_D44_54M7_790_18_FR.inp	564 KB	INP File	1/30/2008 2:19 PM
 304L_600_T415_D44_54M7_790_18_FR.odb	12,294 KB	ODB File	1/31/2008 3:35 PM
 304L_600_T415_D44_54M7_790_18_S.inp	564 KB	INP File	1/23/2008 10:17 AM
 304L_600_T415_D44_54M7_790_18_S.odb	12,304 KB	ODB File	1/24/2008 3:59 PM
 304L_600_T417_D44_72K9_1097_1450_FR.inp	564 KB	INP File	1/30/2008 2:22 PM
 304L_600_T417_D44_72K9_1097_1450_FR.odb	12,269 KB	ODB File	1/31/2008 5:28 PM
 304L_600_T417_D44_72K9_1097_1450_S.inp	564 KB	INP File	1/23/2008 10:39 AM
 304L_600_T417_D44_72K9_1097_1450_S.odb	12,289 KB	ODB File	1/24/2008 5:52 PM
 304L_600_T507_D44W_54M7_1097_1150_FR.inp	564 KB	INP File	1/30/2008 2:03 PM
 304L_600_T507_D44W_54M7_1097_1150_FR.odb	12,273 KB	ODB File	1/31/2008 8:44 AM
 304L_600_T507_D44W_54M7_1097_1150_S.inp	564 KB	INP File	1/23/2008 3:02 PM
 304L_600_T507_D44W_54M7_1097_1150_S.odb	12,252 KB	ODB File	1/23/2008 6:42 PM
 304L_-20_T432_D22_64A1_790_3_FR.inp	257 KB	INP File	1/30/2008 11:16 AM
 304L_-20_T432_D22_64A1_790_3_FR.odb	6,632 KB	ODB File	1/30/2008 1:04 PM
 304L_-20_T432_D22_64A1_790_3_S.inp	257 KB	INP File	1/17/2008 2:09 PM
 304L_-20_T432_D22_64A1_790_3_S.odb	6,551 KB	ODB File	1/22/2008 8:06 AM
 304L_-20_T451_D22_4858_1097_250_FR.inp	257 KB	INP File	1/30/2008 11:13 AM
 304L_-20_T451_D22_4858_1097_250_FR.odb	6,639 KB	ODB File	1/30/2008 1:07 PM
 304L_-20_T451_D22_4858_1097_250_S.inp	257 KB	INP File	1/17/2008 12:43 PM
 304L_-20_T451_D22_4858_1097_250_S.odb	6,597 KB	ODB File	1/22/2008 8:32 AM
 304L_-20_T457_D44_72K9_1097_175_FR.inp	565 KB	INP File	1/30/2008 11:58 AM
 304L_-20_T457_D44_72K9_1097_175_FR.odb	12,301 KB	ODB File	1/30/2008 3:28 PM
 304L_-20_T457_D44_72K9_1097_175_S.inp	565 KB	INP File	1/21/2008 1:35 PM
 304L_-20_T457_D44_72K9_1097_175_S.odb	12,263 KB	ODB File	1/22/2008 2:29 PM
 304L_-20_T468_D44_54M7_790_20_FR.inp	565 KB	INP File	1/30/2008 11:57 AM
 304L_-20_T468_D44_54M7_790_20_FR.odb	12,275 KB	ODB File	1/30/2008 3:28 PM
 304L_-20_T468_D44_54M7_790_20_S.inp	565 KB	INP File	1/21/2008 11:32 AM
 304L_-20_T468_D44_54M7_790_20_S.odb	12,277 KB	ODB File	1/22/2008 9:50 AM
 304L_-20_T491_D44W_54M7_1097_175_FR.inp	564 KB	INP File	1/30/2008 12:28 PM
 304L_-20_T491_D44W_54M7_1097_175_FR.odb	12,274 KB	ODB File	1/31/2008 8:28 AM
 304L_-20_T491_D44W_54M7_1097_175_S.inp	564 KB	INP File	1/21/2008 3:06 PM
 304L_-20_T491_D44W_54M7_1097_175_S.odb	12,272 KB	ODB File	1/23/2008 10:58 AM
 316L_70_T291_D44_67K0_1097_1475_FR.inp	564 KB	INP File	1/30/2008 3:10 PM
 316L_70_T291_D44_67K0_1097_1475_FR.odb	12,298 KB	ODB File	2/4/2008 10:59 AM
 316L_70_T291_D44_67K0_1097_1475_S.inp	564 KB	INP File	1/24/2008 11:21 AM
 316L_70_T291_D44_67K0_1097_1475_S.odb	12,271 KB	ODB File	1/25/2008 3:13 PM

Name ▲	Size	Type	Date Modified
 316L_70_T306_D22_76H3_790_3_FR.inp	257 KB	INP File	1/30/2008 3:07 PM
 316L_70_T306_D22_76H3_790_3_FR.odb	6,626 KB	ODB File	2/4/2008 9:10 AM
 316L_70_T306_D22_76H3_790_3_S.inp	257 KB	INP File	1/21/2008 8:18 AM
 316L_70_T306_D22_76H3_790_3_S.odb	6,601 KB	ODB File	1/21/2008 2:56 PM
 316L_70_T352_D44_2304_790_1825_FR.inp	564 KB	INP File	1/30/2008 3:09 PM
 316L_70_T352_D44_2304_790_1825_FR.odb	12,295 KB	ODB File	2/4/2008 10:50 AM
 316L_70_T352_D44_2304_790_1825_S.inp	564 KB	INP File	1/24/2008 10:10 AM
 316L_70_T352_D44_2304_790_1825_S.odb	12,305 KB	ODB File	1/29/2008 6:24 PM
 316L_70_T357_D22_48R8_1097_275_FR.inp	257 KB	INP File	1/30/2008 3:05 PM
 316L_70_T357_D22_48R8_1097_275_FR.odb	6,580 KB	ODB File	2/4/2008 8:44 AM
 316L_70_T357_D22_48R8_1097_275_S.inp	257 KB	INP File	1/25/2008 12:00 PM
 316L_70_T357_D22_48R8_1097_275_S.odb	6,590 KB	ODB File	1/25/2008 12:58 PM
 316L_70_T382_D44W_2304_1097_1475_FR.inp	564 KB	INP File	1/30/2008 3:56 PM
 316L_70_T382_D44W_2304_1097_1475_FR.odb	12,269 KB	ODB File	2/4/2008 12:38 PM
 316L_70_T382_D44W_2304_1097_1475_S.inp	564 KB	INP File	1/24/2008 4:42 PM
 316L_70_T382_D44W_2304_1097_1475_S.odb	12,270 KB	ODB File	1/25/2008 6:09 PM
 316L_300_T286_D44_67K0_1097_16_FR.inp	565 KB	INP File	1/31/2008 8:20 AM
 316L_300_T286_D44_67K0_1097_16_FR.odb	12,280 KB	ODB File	2/5/2008 8:52 AM
 316L_300_T286_D44_67K0_1097_16_S.inp	565 KB	INP File	1/25/2008 11:14 AM
 316L_300_T286_D44_67K0_1097_16_S.odb	12,290 KB	ODB File	1/28/2008 2:55 PM
 316L_300_T320_D22_76H3_790_250_FR.inp	257 KB	INP File	1/31/2008 8:10 AM
 316L_300_T320_D22_76H3_790_250_FR.odb	6,496 KB	ODB File	2/4/2008 5:44 PM
 316L_300_T320_D22_76H3_790_250_S.inp	256 KB	INP File	1/21/2008 8:58 AM
 316L_300_T320_D22_76H3_790_250_S.odb	6,586 KB	ODB File	1/21/2008 4:17 PM
 316L_300_T400_D22_48R8_1097_250_FR.inp	257 KB	INP File	1/31/2008 8:03 AM
 316L_300_T400_D22_48R8_1097_250_FR.odb	6,617 KB	ODB File	2/4/2008 5:01 PM
 316L_300_T400_D22_48R8_1097_250_S.inp	256 KB	INP File	1/25/2008 9:10 AM
 316L_300_T400_D22_48R8_1097_250_S.odb	6,536 KB	ODB File	1/28/2008 10:33 AM
 316L_300_T406_D44_2304_790_18688_FR.inp	564 KB	INP File	1/31/2008 8:18 AM
 316L_300_T406_D44_2304_790_18688_FR.odb	12,279 KB	ODB File	2/4/2008 5:29 PM
 316L_300_T406_D44_2304_790_18688_S.inp	564 KB	INP File	1/25/2008 10:38 AM
 316L_300_T406_D44_2304_790_18688_S.odb	12,294 KB	ODB File	1/28/2008 12:04 PM
 316L_300_T514_D44W_2304_1097_14_FR.inp	564 KB	INP File	1/31/2008 7:52 AM
 316L_300_T514_D44W_2304_1097_14_FR.odb	12,214 KB	ODB File	2/4/2008 2:51 PM
 316L_300_T514_D44W_2304_1097_14_S.inp	564 KB	INP File	1/25/2008 7:53 AM



Name ▲	Size	Type	Date Modified
 316L_300_T514_D44W_2304_1097_14_S.odb	12,243 KB	ODB File	1/28/2008 8:41 AM
 316L_600_T346_D22_76H3_790_250_FR.inp	256 KB	INP File	1/31/2008 8:39 AM
 316L_600_T346_D22_76H3_790_250_FR.odb	6,584 KB	ODB File	2/5/2008 8:20 AM
 316L_600_T346_D22_76H3_790_250_S.inp	256 KB	INP File	1/21/2008 10:55 AM
 316L_600_T346_D22_76H3_790_250_S.odb	6,583 KB	ODB File	1/21/2008 5:26 PM
 316L_600_T390_D22_48R8_1097_2125_FR.inp	257 KB	INP File	1/31/2008 8:25 AM
 316L_600_T390_D22_48R8_1097_2125_FR.odb	6,499 KB	ODB File	2/5/2008 8:16 AM
 316L_600_T390_D22_48R8_1097_2125_S.inp	257 KB	INP File	1/25/2008 1:56 PM
 316L_600_T390_D22_48R8_1097_2125_S.odb	6,592 KB	ODB File	1/28/2008 4:51 PM
 316L_600_T394_D44_2304_790_17875_FR.inp	564 KB	INP File	1/31/2008 8:43 AM
 316L_600_T394_D44_2304_790_17875_FR.odb	12,297 KB	ODB File	2/5/2008 10:38 AM
 316L_600_T394_D44_2304_790_17875_S.inp	564 KB	INP File	1/25/2008 2:03 PM
 316L_600_T394_D44_2304_790_17875_S.odb	12,302 KB	ODB File	1/28/2008 6:07 PM
 316L_600_T418_D44_67K0_1097_13_FR.inp	564 KB	INP File	1/31/2008 8:47 AM
 316L_600_T418_D44_67K0_1097_13_FR.odb	12,277 KB	ODB File	2/5/2008 9:48 AM
 316L_600_T418_D44_67K0_1097_13_S.inp	564 KB	INP File	1/25/2008 2:51 PM
 316L_600_T418_D44_67K0_1097_13_S.odb	12,283 KB	ODB File	1/29/2008 9:01 AM
 316L_600_T508_D44W_2304_1097_11_FR.inp	564 KB	INP File	1/31/2008 7:55 AM
 316L_600_T508_D44W_2304_1097_11_FR.odb	12,267 KB	ODB File	2/4/2008 3:12 PM
 316L_600_T508_D44W_2304_1097_11_S.inp	564 KB	INP File	1/25/2008 8:42 AM
 316L_600_T508_D44W_2304_1097_11_S.odb	12,239 KB	ODB File	1/28/2008 8:19 AM
 316L_-20_T431_D22_76H3_790_3_FR.inp	257 KB	INP File	1/30/2008 2:31 PM
 316L_-20_T431_D22_76H3_790_3_FR.odb	6,532 KB	ODB File	1/31/2008 6:09 PM
 316L_-20_T431_D22_76H3_790_3_S.inp	256 KB	INP File	1/23/2008 3:29 PM
 316L_-20_T431_D22_76H3_790_3_S.odb	6,608 KB	ODB File	1/25/2008 9:02 AM
 316L_-20_T458_D44_67K0_1097_1475_FR.inp	564 KB	INP File	1/30/2008 2:34 PM
 316L_-20_T458_D44_67K0_1097_1475_FR.odb	12,291 KB	ODB File	2/4/2008 9:08 AM
 316L_-20_T458_D44_67K0_1097_1475_S.inp	564 KB	INP File	1/24/2008 8:06 AM
 316L_-20_T458_D44_67K0_1097_1475_S.odb	12,275 KB	ODB File	1/25/2008 11:25 AM
 316L_-20_T465_D22_48R8_1097_250_FR.inp	256 KB	INP File	1/30/2008 2:26 PM
 316L_-20_T465_D22_48R8_1097_250_FR.odb	6,627 KB	ODB File	1/31/2008 6:00 PM
 316L_-20_T465_D22_48R8_1097_250_S.inp	256 KB	INP File	1/23/2008 11:19 AM
 316L_-20_T465_D22_48R8_1097_250_S.odb	6,587 KB	ODB File	1/24/2008 6:18 PM
 316L_-20_T471_D44_2304_790_18_FR.inp	565 KB	INP File	1/30/2008 2:33 PM
 316L_-20_T471_D44_2304_790_18_FR.odb	12,298 KB	ODB File	2/4/2008 8:57 AM
 316L_-20_T471_D44_2304_790_18_S.inp	564 KB	INP File	1/23/2008 4:45 PM
 316L_-20_T471_D44_2304_790_18_S.odb	12,280 KB	ODB File	1/25/2008 8:52 AM
 316L_-20_T493_D44W_2304_1097_19_FR.inp	564 KB	INP File	1/30/2008 3:54 PM
 316L_-20_T493_D44W_2304_1097_19_FR.odb	12,293 KB	ODB File	2/4/2008 12:40 PM
 316L_-20_T493_D44W_2304_1097_19_S.inp	564 KB	INP File	1/24/2008 3:10 PM
 316L_-20_T493_D44W_2304_1097_19_S.odb	12,262 KB	ODB File	1/25/2008 4:41 PM

LABORATORY NOTEBOOK NSNF/SN/04.01  
NSNFP MATERIAL STRAIN RATE TESTING

# **LAB NOTEBOOK BINDER VOLUME 8**

**Strain Rate Testing  
Calibration Information**

# **GO / NO GO GAUGES**

**INL**  
**CALIBRATION INPUT DATA**

1/10/2008

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 725378    Mfr: BEA    Model: NG-1    Noun Name: GO-NOGO GAGE    Serial #:

Calibration Date: 1/10/2008

**ACTION CODE**

**AS FOUND**

Next Cal Due Date: 1/10/2009	1	<input type="checkbox"/>	Acceptance Test	1	<input checked="" type="radio"/>	In Tolerance
Charge Level: 12	2	<input type="checkbox"/>	Special Test	2	<input type="radio"/>	Out of Tolerance >1x <2x
Repair/Adj/etc C.L: 0	3	<input checked="" type="checkbox"/>	Calibration to MFG Specs	3	<input type="radio"/>	Out of Tolerance >2x <3x
Material Amount: 0	4	<input type="checkbox"/>	Clean	4	<input type="radio"/>	Out of Tolerance >3x <5x
Charge Number: 100853GSA	5	<input type="checkbox"/>	Limited Calibration	5	<input type="radio"/>	Out of Tolerance >5x
Cal Work Inst ID: 6166	6	<input type="checkbox"/>	Functional Check	6	<input type="radio"/>	Out of Tolerance-Undetermined
Outside Vendor:	7	<input type="checkbox"/>	Performance Check	7	<input type="checkbox"/>	Inoperative
	8	<input type="checkbox"/>	Modify	8	<input type="checkbox"/>	Damaged
	9	<input type="checkbox"/>	Repair-needs Charge Level	9	<input type="checkbox"/>	Not Used
	10	<input type="checkbox"/>	Other	10	<input type="checkbox"/>	Not Determined
				11	<input type="checkbox"/>	Excessed
Calibrated By: Larry Deming	S#: 39571	Phone: 526-2761		12	<input type="checkbox"/>	Extension

**CALIBRATION STANDARDS USED**

724663	704666						

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

NOMINAL (STD)	OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION		MFG. ACCURACY
	UNITS	AS FOUND (UUT)	

COMMENTS



**INL**  
**CALIBRATION INPUT DATA**

1/10/2008

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 725379      Mfr: BEA      Model: NG-2      Noun Name: GO-NOGO GAGE      Serial #:

Calibration Date: 1/10/2008

**ACTION CODE**

**AS FOUND**

Next Cal Due Date: 1/10/2009	1	<input type="checkbox"/>	Acceptance Test	1	<input checked="" type="radio"/>	In Tolerance
Charge Level: 4	2	<input type="checkbox"/>	Special Test	2	<input type="radio"/>	Out of Tolerance >1x <2x
Repair/Adj/etc C.L.: 0	3	<input checked="" type="checkbox"/>	Calibration to MFG Specs	3	<input type="radio"/>	Out of Tolerance >2x <3x
Material Amount: 0	4	<input type="checkbox"/>	Clean	4	<input type="radio"/>	Out of Tolerance >3x <5x
Charge Number: 100853GSA	5	<input type="checkbox"/>	Limited Calibration	5	<input type="radio"/>	Out of Tolerance >5x
Cal Work Inst ID: 6166	6	<input type="checkbox"/>	Functional Check	6	<input type="radio"/>	Out of Tolerance-Undetermined
Outside Vendor:	7	<input type="checkbox"/>	Performance Check	7	<input type="checkbox"/>	Inoperative
	8	<input type="checkbox"/>	Modify	8	<input type="checkbox"/>	Damaged
	9	<input type="checkbox"/>	Repair-needs Charge Level	9	<input type="checkbox"/>	Not Used
	10	<input type="checkbox"/>	Other	10	<input type="checkbox"/>	Not Determined
				11	<input type="checkbox"/>	Excessed
Calibrated By: Larry Deming      S#: 39571      Phone: 526-2761				12	<input type="checkbox"/>	Extension

**CALIBRATION STANDARDS USED**

724663							

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

NOMINAL (STD)	OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION UNITS	AS FOUND (UUT)	MFG. ACCURACY

**COMMENTS**

**INL  
CALIBRATION INPUT DATA**

1/10/2008

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 725380    Mfr: BEA    Model: NG-3    Noun Name: GO-NOGO GAGE    Serial #:

Calibration Date: 1/10/2008

**ACTION CODE****AS FOUND**

Next Cal Due Date: 1/10/2009	1	<input type="checkbox"/>	Acceptance Test	1	<input checked="" type="radio"/>	In Tolerance
Charge Level: 12	2	<input type="checkbox"/>	Special Test	2	<input type="radio"/>	Out of Tolerance >1x <2x
Repair/Adj/etc C.L: 0	3	<input checked="" type="checkbox"/>	Calibration to MFG Specs	3	<input type="radio"/>	Out of Tolerance >2x <3x
Material Amount: 0	4	<input type="checkbox"/>	Clean	4	<input type="radio"/>	Out of Tolerance >3x <5x
Charge Number: 100853GSA	5	<input type="checkbox"/>	Limited Calibration	5	<input type="radio"/>	Out of Tolerance >5x
Cal Work Inst ID: 6166	6	<input type="checkbox"/>	Functional Check	6	<input type="radio"/>	Out of Tolerance-Undetermined
Outside Vendor:	7	<input type="checkbox"/>	Performance Check	7	<input type="checkbox"/>	Inoperative
	8	<input type="checkbox"/>	Modify	8	<input type="checkbox"/>	Damaged
	9	<input type="checkbox"/>	Repair-needs Charge Level	9	<input type="checkbox"/>	Not Used
	10	<input type="checkbox"/>	Other	10	<input type="checkbox"/>	Not Determined
				11	<input type="checkbox"/>	Excessed
Calibrated By: Larry Deming	S#: 39571	Phone: 526-2761		12	<input type="checkbox"/>	Extension

**CALIBRATION STANDARDS USED**

724663	704666						

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-5.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

NOMINAL (STD)	OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION UNITS	AS FOUND (UUT)	MFG. ACCURACY

**COMMENTS**

# INL CALIBRATION INPUT DATA

1/10/2008

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 725381      Mfr: BEA      Model: NG-4      Noun Name: GO-NOGO GAGE      Serial #:

Calibration Date: 1/10/2008

**ACTION CODE****AS FOUND**

Next Cal Due Date: 1/10/2009	1	<input type="checkbox"/>	Acceptance Test	1	<input checked="" type="radio"/>	In Tolerance
Charge Level: 12	2	<input type="checkbox"/>	Special Test	2	<input type="radio"/>	Out of Tolerance >1x <2x
Repair/Adj/etc C.L: 0	3	<input checked="" type="checkbox"/>	Calibration to MFG Specs	3	<input type="radio"/>	Out of Tolerance >2x <3x
Material Amount: 0	4	<input type="checkbox"/>	Clean	4	<input type="radio"/>	Out of Tolerance >3x <5x
Charge Number: 100853GSA	5	<input type="checkbox"/>	Limited Calibration	5	<input type="radio"/>	Out of Tolerance >5x
Cal Work Inst ID: 6166	6	<input type="checkbox"/>	Functional Check	6	<input type="radio"/>	Out of Tolerance-Undetermined
Outside Vendor:	7	<input type="checkbox"/>	Performance Check	7	<input type="checkbox"/>	Inoperative
	8	<input type="checkbox"/>	Modify	8	<input type="checkbox"/>	Damaged
	9	<input type="checkbox"/>	Repair-needs Charge Level	9	<input type="checkbox"/>	Not Used
	10	<input type="checkbox"/>	Other	10	<input type="checkbox"/>	Not Determined
				11	<input type="checkbox"/>	Excessed
Calibrated By: Larry Deming	S#: 39571	Phone: 526-2761		12	<input type="checkbox"/>	Extension

**CALIBRATION STANDARDS USED**

724663	704666						

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

NOMINAL (STD)	OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION UNITS	AS FOUND (UUT)	MFG. ACCURACY

COMMENTS

**INL**  
**CALIBRATION INPUT DATA**

1/10/2008

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 725382	Mfr: BEA	Model: NG-5	Noun Name: GO-NOGO GAGE	Serial #:
Calibration Date: 1/10/2008	<b>ACTION CODE</b>		<b>AS FOUND</b>	
Next Cal Due Date: 1/10/2009	1	<input type="checkbox"/> Acceptance Test	1	<input type="checkbox"/> In Tolerance
Charge Level: 4	2	<input type="checkbox"/> Special Test	2	<input checked="" type="checkbox"/> Out of Tolerance >1x <2x
Repair/Adj/etc C.L: 0	3	<input checked="" type="checkbox"/> Calibration to MFG Specs	3	<input type="checkbox"/> Out of Tolerance >2x <3x
Material Amount: 0	4	<input type="checkbox"/> Clean	4	<input type="checkbox"/> Out of Tolerance >3x <5x
Charge Number: 100853GSA	5	<input type="checkbox"/> Limited Calibration	5	<input type="checkbox"/> Out of Tolerance >5x
Cal Work Inst ID: 6166	6	<input type="checkbox"/> Functional Check	6	<input type="checkbox"/> Out of Tolerance-Undetermined
Outside Vendor:	7	<input type="checkbox"/> Performance Check	7	<input type="checkbox"/> Inoperative
	8	<input type="checkbox"/> Modify	8	<input type="checkbox"/> Damaged
	9	<input type="checkbox"/> Repair-needs Charge Level	9	<input type="checkbox"/> Not Used
	10	<input type="checkbox"/> Other	10	<input type="checkbox"/> Not Determined
			11	<input type="checkbox"/> Excessed
Calibrated By: Larry Deming	S#: 39571	Phone: 526-2761	12	<input type="checkbox"/> Extension

**CALIBRATION STANDARDS USED**

724663							

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

NOMINAL (STD)	OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION		MFG. ACCURACY
	UNITS	AS FOUND (UUT)	

COMMENTS

**Notification Status : Approved**

[illegible]

**Calibration Remarks :** *(file attachments allowed)*

Measurement just barely out of tolerance.

**Evaluation Required ?:**☒ REQUIRED☐ Not Required-New Equipment☐ Not Required-Other**Evaluation:****Evaluator:** Dana K Morton**Evaluation:** *(file attachments allowed)*

This go/no go gage is still acceptable for use through the measuring procedure used by the team members. Even though certain measurements exceed tolerances, the fact that multiple measurements were taken for a specific portion of the gage and that those other measurements do satisfy the required tolerance allows for the determination of an acceptable go/no go gage.

During gage use, the test specimen is fully inserted (as far as it can go), with a minimal insertion to the first shoulder, approximately a 1/4-inch depth. An acceptable test specimen will insert to this depth but not proceed further. The Cal Lab was requested to make three gap measurements along each of the two gage depths (initial depth checking for too large of a test specimen and the secondary depth checking for too small of a test specimen). The dimensional gap H (in the initial depth) exceeds the tolerance but the other two measurements (F and G) are 0.270 inches, meaning that by fully inserting to the gage shoulder, the test specimen cannot be larger than 0.270 inches, which is acceptable. In a similar fashion, dimensional gaps A and B (in the secondary depth) barely exceed the tolerance but dimensional gap C, measuring 0.230 inches, is acceptable and is located at the very initial opening of the secondary depth, again providing a very clear indication of test specimen acceptance within the stated tolerances when the test specimen cannot proceed past the first shoulder.

This gage was initially measured by an INL qualified dimensional inspector and placed into use in August 2004. The slight over-tolerance was known but follow-on direct measurements were used to double check test specimen acceptability. However, in May 2005, this N-5 gage was replaced with NG-6. NG-5 has not been used since May 2005 and will not be used in the future.

**Selected Approver:** Robert K Blandford

**NOTE:** This evaluation or a separate out -of-tolerance evaluation must be retained with organization quality assurance records. Quality requirements for this evaluation and quality record retention are based on PDD-13450 and LWP-13455.

**Evaluation Approval:****Date:** 01/10/2008 **Approval:** Robert K Blandford**Approval Comments :** I agree.**Comments/Attachments**

Add any comments here and/or use the area for any attachment.

**Document History:**

Created by Larry W Deming on 01/10/2008 07:54:22 AM

Larry W Deming edited document 08:02 AM 01/10/2008

Notification sent: Larry W Deming on Thursday, January 10, 2008.

Sent to user: Dana K Morton: 08:02 AM 01/10/2008

Dana K Morton edited document 12:28 PM 01/10/2008

Submitted for approval to Robert K Blandford: Dana K Morton on Thursday, January 10, 2008

Sent for approval to Robert K Blandford: 12:29 PM 01/10/2008

Evaluation Approval: Robert K Blandford on Thursday, January 10, 2008

## COMMENTS

**INL**  
**CALIBRATION INPUT DATA**

1/10/2008

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 725384    Mfr: BEA    Model: MK-1    Noun Name: PUNCH TEMPLATE    Serial #:

Calibration Date: 1/10/2008

**ACTION CODE****AS FOUND**

Next Cal Due Date: 1/10/2009

1 ☐ Acceptance Test1 ☒ In Tolerance

Charge Level: 4

2 ☐ Special Test2 ☐ Out of Tolerance >1x <2x

Repair/Adj/etc C.L: 0

3 ☒ Calibration to MFG Specs3 ☐ Out of Tolerance >2x <3x

Material Amount: 0

4 ☐ Clean4 ☐ Out of Tolerance >3x <5x

Charge Number:

5 ☐ Limited Calibration5 ☐ Out of Tolerance >5x

Cal Work Inst ID: 6166

6 ☐ Functional Check6 ☐ Out of Tolerance-Undetermined

Outside Vendor:

7 ☐ Performance Check7 ☐ Inoperative8 ☐ Modify8 ☐ Damaged9 ☐ Repair-needs Charge Level9 ☐ Not Used10 ☐ Other10 ☐ Not Determined11 ☐ Exceeded

Calibrated By: Larry Deming

S#: 39571 Phone: 526-2761

12 ☐ Extension**CALIBRATION STANDARDS USED**

724663							

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

**OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION**

NOMINAL (STD)	UNITS	AS FOUND (UUT)	MFG. ACCURACY

**COMMENTS**



**INL**  
**CALIBRATION INPUT DATA**

1/10/2008

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 725385      Mfr: BEA      Model: MK-2      Noun Name: PUNCH TEMPLATE      Serial #:

Calibration Date: 1/10/2008

**ACTION CODE****AS FOUND**

Next Cal Due Date: 1/10/2009	1	<input type="checkbox"/>	Acceptance Test	1	<input checked="" type="radio"/>	In Tolerance
Charge Level: 2	2	<input type="checkbox"/>	Special Test	2	<input type="radio"/>	Out of Tolerance >1x <2x
Repair/Adj/etc C.L: 0	3	<input checked="" type="checkbox"/>	Calibration to MFG Specs	3	<input type="radio"/>	Out of Tolerance >2x <3x
Material Amount: 0	4	<input type="checkbox"/>	Clean	4	<input type="radio"/>	Out of Tolerance >3x <5x
Charge Number:	5	<input type="checkbox"/>	Limited Calibration	5	<input type="radio"/>	Out of Tolerance >5x
Cal Work Inst ID: 6166	6	<input type="checkbox"/>	Functional Check	6	<input type="radio"/>	Out of Tolerance-Undetermined
Outside Vendor:	7	<input type="checkbox"/>	Performance Check	7	<input type="checkbox"/>	Inoperative
	8	<input type="checkbox"/>	Modify	8	<input type="checkbox"/>	Damaged
	9	<input type="checkbox"/>	Repair-needs Charge Level	9	<input type="checkbox"/>	Not Used
	10	<input type="checkbox"/>	Other	10	<input type="checkbox"/>	Not Determined
				11	<input type="checkbox"/>	Excessed
Calibrated By: Larry Deming				12	<input type="checkbox"/>	Extension

S#: 39571 Phone: 526-2761

**CALIBRATION STANDARDS USED**

724663							

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED  
VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

NOMINAL (STD)	OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION UNITS	AS FOUND (UUT)	MFG. ACCURACY

COMMENTS



**INL**  
**CALIBRATION INPUT DATA**

1/10/2008

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 725387      Mfr: BEA      Model: MK-4      Noun Name: DIMENSIONAL CHECK GAGE      Serial #:

Calibration Date: 1/10/2008

**ACTION CODE****AS FOUND**

Next Cal Due Date: 1/10/2009	1	<input type="checkbox"/>	Acceptance Test	1	<input checked="" type="radio"/>	In Tolerance
Charge Level: 4	2	<input type="checkbox"/>	Special Test	2	<input type="radio"/>	Out of Tolerance >1x <2x
Repair/Adj/etc C.L: 0	3	<input checked="" type="checkbox"/>	Calibration to MFG Specs	3	<input type="radio"/>	Out of Tolerance >2x <3x
Material Amount: 0	4	<input type="checkbox"/>	Clean	4	<input type="radio"/>	Out of Tolerance >3x <5x
Charge Number: 100853GSA	5	<input type="checkbox"/>	Limited Calibration	5	<input type="radio"/>	Out of Tolerance >5x
Cal Work Inst ID: 6166	6	<input type="checkbox"/>	Functional Check	6	<input type="radio"/>	Out of Tolerance-Undetermined
Outside Vendor:	7	<input type="checkbox"/>	Performance Check	7	<input type="checkbox"/>	Inoperative
	8	<input type="checkbox"/>	Modify	8	<input type="checkbox"/>	Damaged
	9	<input type="checkbox"/>	Repair-needs Charge Level	9	<input type="checkbox"/>	Not Used
	10	<input type="checkbox"/>	Other	10	<input type="checkbox"/>	Not Determined
				11	<input type="checkbox"/>	Excessed
Calibrated By: Larry Deming				12	<input type="checkbox"/>	Extension

S#: 39571 Phone: 526-2761

**CALIBRATION STANDARDS USED**

724663							

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

**OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION**

NOMINAL (STD)

UNITS

AS FOUND (UUT)

MFG. ACCURACY


COMMENTS

# **ACCELEROMETERS**

# INL CALIBRATION INPUT DATA

11/29/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 716211      Mfr: PCB      Model: 353B03      Noun Name: ACCELEROMETER      Serial #: 38617

Calibration Date: 7/26/1999 12:47:32 PM

**ACTION CODE****AS FOUND**

Next Cal Due Date: 7/19/2000

Charge Level: 0

Repair/Adj/etc C.L: 0

Material Amount: 0

Charge Number: 827257038

Cal Work Inst ID: PCB

Outside Vendor: PCB

1 ☐ Acceptance Test2 ☐ Special Test3 ☒ Calibration to MFG Specs4 ☐ Clean5 ☐ Limited Calibration6 ☐ Functional Check7 ☐ Performance Check8 ☐ Modify9 ☐ Repair-needs Charge Level10 ☐ Other1 ☒ In Tolerance2 ☐ Out of Tolerance >1x <2x3 ☐ Out of Tolerance >2x <3x4 ☐ Out of Tolerance >3x <5x5 ☐ Out of Tolerance >5x6 ☐ Out of Tolerance-Undetermined7 ☐ Inoperative8 ☐ Damaged9 ☐ Not Used10 ☐ Not Determined11 ☐ Excessed12 ☐ Extension

Calibrated By: Mike Stears

S#: 61767 Phone: 526-2761

**CALIBRATION STANDARDS USED**


STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

**OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION**

NOMINAL (STD)

UNITS

AS FOUND (UUT)

MFG. ACCURACY


COMMENTS

**INL  
CALIBRATION INPUT DATA**

11/29/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 716211      Mfr: PCB      Model: 353B03      Noun Name: ACCELEROMETER      Serial #: 38617

Calibration Date: 7/17/2000 4:55:38 PM

**ACTION CODE****AS FOUND**

Next Cal Due Date: 7/12/2001

Charge Level: 1

Repair/Adj/etc C.L: 0

Material Amount: 0

Charge Number: P44300001

Cal Work Inst ID: PCB

Outside Vendor: PCB

1 ☐ Acceptance Test2 ☐ Special Test3 ☒ Calibration to MFG Specs4 ☐ Clean5 ☐ Limited Calibration6 ☐ Functional Check7 ☐ Performance Check8 ☐ Modify9 ☐ Repair-needs Charge Level10 ☐ Other1 ☒ In Tolerance2 ☐ Out of Tolerance >1x <2x3 ☐ Out of Tolerance >2x <3x4 ☐ Out of Tolerance >3x <5x5 ☐ Out of Tolerance >5x6 ☐ Out of Tolerance-Undetermined7 ☐ Inoperative8 ☐ Damaged9 ☐ Not Used10 ☐ Not Determined11 ☐ Excessed12 ☐ Extension

Calibrated By: Mike Stears

S#: 61767 Phone: 526-2761

**CALIBRATION STANDARDS USED**


STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

**OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION****NOMINAL (STD)****UNITS****AS FOUND (UUT)****MFG. ACCURACY**


**COMMENTS**

**INL  
CALIBRATION INPUT DATA**

11/29/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 716211    Mfr: PCB    Model: 353B03    Noun Name: ACCELEROMETER    Serial #: 38617

Calibration Date: 8/19/2003

**ACTION CODE**

**AS FOUND**

Next Cal Due Date: 7/7/2004	1	<input type="checkbox"/>	Acceptance Test	1	<input checked="" type="radio"/>	In Tolerance
Charge Level: 2	2	<input type="checkbox"/>	Special Test	2	<input type="radio"/>	Out of Tolerance >1x <2x
Repair/Adj/etc C.L.: 0	3	<input checked="" type="checkbox"/>	Calibration to MFG Specs	3	<input type="radio"/>	Out of Tolerance >2x <3x
Material Amount: 0	4	<input type="checkbox"/>	Clean	4	<input type="radio"/>	Out of Tolerance >3x <5x
Charge Number: 100348027	5	<input type="checkbox"/>	Limited Calibration	5	<input type="radio"/>	Out of Tolerance >5x
Cal Work Inst ID: PCB	6	<input type="checkbox"/>	Functional Check	6	<input type="radio"/>	Out of Tolerance-Undetermined
Outside Vendor: PCB	7	<input type="checkbox"/>	Performance Check	7	<input type="checkbox"/>	Inoperative
	8	<input type="checkbox"/>	Modify	8	<input type="checkbox"/>	Damaged
	9	<input type="checkbox"/>	Repair-needs Charge Level	9	<input type="checkbox"/>	Not Used
	10	<input type="checkbox"/>	Other	10	<input type="checkbox"/>	Not Determined
				11	<input type="checkbox"/>	Excessed
Calibrated By: Stan Zohner	S#:	58146	Phone: 526-2761	12	<input type="checkbox"/>	Extension

**CALIBRATION STANDARDS USED**


STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-5.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

NOMINAL (STD)	OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION		MFG. ACCURACY
	UNITS	AS FOUND (UUT)	

**COMMENTS**

PCB  
test referral QA# 102401

**INL  
CALIBRATION INPUT DATA**

11/29/2007

NAME: DANA KEITH MORTON

BADGE: 35698

PH: 526-1274

AREA: STC

BLDG: EROB

RM: W2D1

ID Number: 716211 Mfr: PCB Model: 353B03 Noun Name: ACCELEROMETER Serial #: 38617

Calibration Date: 11/11/2004

**ACTION CODE****AS FOUND**

Next Cal Due Date: 10/21/2005	1	<input type="checkbox"/>	Acceptance Test	1	<input checked="" type="radio"/>	In Tolerance
Charge Level: 2	2	<input type="checkbox"/>	Special Test	2	<input type="radio"/>	Out of Tolerance >1x <2x
Repair/Adj/etc C.L.: 0	3	<input checked="" type="checkbox"/>	Calibration to MFG Specs	3	<input type="radio"/>	Out of Tolerance >2x <3x
Material Amount: 0	4	<input type="checkbox"/>	Clean	4	<input type="radio"/>	Out of Tolerance >3x <5x
Charge Number: 100664GSA	5	<input type="checkbox"/>	Limited Calibration	5	<input type="radio"/>	Out of Tolerance >5x
Cal Work Inst ID: PCB	6	<input type="checkbox"/>	Functional Check	6	<input type="radio"/>	Out of Tolerance-Undetermined
Outside Vendor: PCB	7	<input type="checkbox"/>	Performance Check	7	<input type="checkbox"/>	Inoperative
	8	<input type="checkbox"/>	Modify	8	<input type="checkbox"/>	Damaged
	9	<input type="checkbox"/>	Repair-needs Charge Level	9	<input type="checkbox"/>	Not Used
	10	<input type="checkbox"/>	Other	10	<input type="checkbox"/>	Not Determined
				11	<input type="checkbox"/>	Excessed
Calibrated By: Scott Lish	S#: 101141	Phone: 526-2761		12	<input type="checkbox"/>	Extension

**CALIBRATION STANDARDS USED**


STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-5.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

**OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION**

NOMINAL (STD)	UNITS	AS FOUND (UUT)	MFG. ACCURACY

**COMMENTS**

Unit sent to manufacture for calibration.  
QA# 111220



**INL**  
**CALIBRATION INPUT DATA**

11/29/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 716211 Mfr: PCB Model: 353B03 Noun Name: ACCELEROMETER Serial #: 38617

Calibration Date: 11/4/2005

**ACTION CODE****AS FOUND**

Next Cal Due Date: 11/4/2006	1	<input type="checkbox"/>	Acceptance Test	1	<input checked="" type="radio"/>	In Tolerance
Charge Level: 2	2	<input type="checkbox"/>	Special Test	2	<input type="radio"/>	Out of Tolerance >1x <2x
Repair/Adj/etc C.L: 0	3	<input checked="" type="checkbox"/>	Calibration to MFG Specs	3	<input type="radio"/>	Out of Tolerance >2x <3x
Material Amount: 0	4	<input type="checkbox"/>	Clean	4	<input type="radio"/>	Out of Tolerance >3x <5x
Charge Number: 100853GSA	5	<input type="checkbox"/>	Limited Calibration	5	<input type="radio"/>	Out of Tolerance >5x
Cal Work Inst ID: PCB	6	<input type="checkbox"/>	Functional Check	6	<input type="radio"/>	Out of Tolerance-Undetermined
Outside Vendor: PCB	7	<input type="checkbox"/>	Performance Check	7	<input type="checkbox"/>	Inoperative
	8	<input type="checkbox"/>	Modify	8	<input type="checkbox"/>	Damaged
	9	<input type="checkbox"/>	Repair-needs Charge Level	9	<input type="checkbox"/>	Not Used
	10	<input type="checkbox"/>	Other	10	<input type="checkbox"/>	Not Determined
				11	<input type="checkbox"/>	Excessed
Calibrated By: Scott Lish	S#: 101141	Phone: 526-2761		12	<input type="checkbox"/>	Extension

**CALIBRATION STANDARDS USED**


STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-5.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

NOMINAL (STD)	OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION		MFG. ACCURACY
	UNITS	AS FOUND (UUT)	

**COMMENTS**

QA#115894  
S&CL overcheck required.

**INL  
CALIBRATION INPUT DATA**

11/29/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 716211      Mfr: PCB      Model: 353B03      Noun Name: ACCELEROMETER      Serial #: 38617

Calibration Date: 12/14/2006

**ACTION CODE****AS FOUND**

Next Cal Due Date: 11/11/2007	1	<input type="checkbox"/>	Acceptance Test	1	<input checked="" type="radio"/>	In Tolerance
Charge Level: 2	2	<input type="checkbox"/>	Special Test	2	<input type="radio"/>	Out of Tolerance >1x <2x
Repair/Adj/etc C.L: 0	3	<input checked="" type="checkbox"/>	Calibration to MFG Specs	3	<input type="radio"/>	Out of Tolerance >2x <3x
Material Amount: 0	4	<input type="checkbox"/>	Clean	4	<input type="radio"/>	Out of Tolerance >3x <5x
Charge Number: 100853GSA	5	<input type="checkbox"/>	Limited Calibration	5	<input type="radio"/>	Out of Tolerance >5x
Cal Work Inst ID: PCB	6	<input type="checkbox"/>	Functional Check	6	<input type="radio"/>	Out of Tolerance-Undetermined
Outside Vendor: PCB	7	<input type="checkbox"/>	Performance Check	7	<input type="checkbox"/>	Inoperative
	8	<input type="checkbox"/>	Modify	8	<input type="checkbox"/>	Damaged
	9	<input type="checkbox"/>	Repair-needs Charge Level	9	<input type="checkbox"/>	Not Used
	10	<input type="checkbox"/>	Other	10	<input type="checkbox"/>	Not Determined
				11	<input type="checkbox"/>	Excessed
Calibrated By: Scott Lish	S#:	101141	Phone: 526-2761	12	<input type="checkbox"/>	Extension

**CALIBRATION STANDARDS USED**


STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

NOMINAL (STD)	OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION UNITS	AS FOUND (UUT)	MFG. ACCURACY

**COMMENTS**

QA# 122337

S&amp;CL overcheck required.

**INL**  
**CALIBRATION INPUT DATA**

12/12/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 716211      Mfr: PCB      Model: 353B03      Noun Name: ACCELEROMETER      Serial #: 38617

Calibration Date: 12/11/2007

**ACTION CODE**

**AS FOUND**

Next Cal Due Date: 11/26/2008	1	<input type="checkbox"/>	Acceptance Test	1	<input checked="" type="radio"/> In Tolerance
Charge Level: 2	2	<input type="checkbox"/>	Special Test	2	<input type="radio"/> Out of Tolerance >1x <2x
Repair/Adj/etc C.L: 0	3	<input checked="" type="checkbox"/>	Calibration to MFG Specs	3	<input type="radio"/> Out of Tolerance >2x <3x
Material Amount: 0	4	<input type="checkbox"/>	Clean	4	<input type="radio"/> Out of Tolerance >3x <5x
Charge Number: 100853GSA	5	<input type="checkbox"/>	Limited Calibration	5	<input type="radio"/> Out of Tolerance >5x
Cal Work Inst ID: PCB	6	<input type="checkbox"/>	Functional Check	6	<input type="radio"/> Out of Tolerance-Undetermined
Outside Vendor: PCB	7	<input type="checkbox"/>	Performance Check	7	<input type="checkbox"/> Inoperative
	8	<input type="checkbox"/>	Modify	8	<input type="checkbox"/> Damaged
	9	<input type="checkbox"/>	Repair-needs Charge Level	9	<input type="checkbox"/> Not Used
	10	<input type="checkbox"/>	Other	10	<input type="checkbox"/> Not Determined
				11	<input type="checkbox"/> Excessed
Calibrated By: Scott Lish	S#: 101141	Phone: 526-2761		12	<input type="checkbox"/> Extension

**CALIBRATION STANDARDS USED**


STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

NOMINAL (STD)	OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION UNITS	AS FOUND (UUT)	MFG. ACCURACY

**COMMENTS**

Calibrated at PCB  
QA# 141993  
S&CL overcheck required.

**INL  
CALIBRATION INPUT DATA**

11/29/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 716213      Mfr: PCB      Model: 353B03      Noun Name: ACCELEROMETER      Serial #: 38619

Calibration Date: 7/26/1999 12:50:56 PM

**ACTION CODE****AS FOUND**

Next Cal Due Date: 7/26/2000

Charge Level: 0

Repair/Adj/etc C.L.: 0

Material Amount: 0

Charge Number: 827257038

Cal Work Inst ID: PCB

Outside Vendor: PCB

1 ☐ Acceptance Test2 ☐ Special Test3 ☒ Calibration to MFG Specs4 ☐ Clean5 ☐ Limited Calibration6 ☐ Functional Check7 ☐ Performance Check8 ☐ Modify9 ☐ Repair-needs Charge Level10 ☐ Other1 ☒ In Tolerance2 ☐ Out of Tolerance >1x <2x3 ☐ Out of Tolerance >2x <3x4 ☐ Out of Tolerance >3x <5x5 ☐ Out of Tolerance >5x6 ☐ Out of Tolerance-Undetermined7 ☐ Inoperative8 ☐ Damaged9 ☐ Not Used10 ☐ Not Determined11 ☐ Exceeded12 ☐ Extension

Calibrated By: Mike Stears

S#: 61767 Phone: 526-2761

**CALIBRATION STANDARDS USED**


STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)    20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)    23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)    20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)    23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)    20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:    23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

**OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION**

NOMINAL (STD)

UNITS

AS FOUND (UUT)

MFG. ACCURACY


**COMMENTS**

**INL**  
**CALIBRATION INPUT DATA**

11/29/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 716213      Mfr: PCB      Model: 353B03      Noun Name: ACCELEROMETER      Serial #: 38619

Calibration Date: 7/17/2000 4:47:30 PM

**ACTION CODE****AS FOUND**

Next Cal Due Date: 7/12/2001

Charge Level: 1

Repair/Adj/etc C.L: 0

Material Amount: 0

Charge Number: P44300001

Cal Work Inst ID: PCB

Outside Vendor: PCB

1 ☐ Acceptance Test2 ☐ Special Test3 ☒ Calibration to MFG Specs4 ☐ Clean5 ☐ Limited Calibration6 ☐ Functional Check7 ☐ Performance Check8 ☐ Modify9 ☐ Repair-needs Charge Level10 ☐ Other1 ☒ In Tolerance2 ☐ Out of Tolerance >1x <2x3 ☐ Out of Tolerance >2x <3x4 ☐ Out of Tolerance >3x <5x5 ☐ Out of Tolerance >5x6 ☐ Out of Tolerance-Undetermined7 ☐ Inoperative8 ☐ Damaged9 ☐ Not Used10 ☐ Not Determined11 ☐ Excessed12 ☐ Extension

Calibrated By: Mike Stears

S#: 61767 Phone: 526-2761

**CALIBRATION STANDARDS USED**


STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
PhysDim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

NOMINAL (STD)	OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION		MFG. ACCURACY
	UNITS	AS FOUND (UUT)	

**COMMENTS**

**INL**  
**CALIBRATION INPUT DATA**

11/29/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 716213    Mfr: PCB    Model: 353B03    Noun Name: ACCELEROMETER    Serial #: 38619

Calibration Date: 8/19/2003

**ACTION CODE****AS FOUND**

Next Cal Due Date: 7/7/2004	1	<input type="checkbox"/>	Acceptance Test	1	<input checked="" type="radio"/>	In Tolerance
Charge Level: 2	2	<input type="checkbox"/>	Special Test	2	<input type="radio"/>	Out of Tolerance >1x <2x
Repair/Adj/etc C.L.: 0	3	<input checked="" type="checkbox"/>	Calibration to MFG Specs	3	<input type="radio"/>	Out of Tolerance >2x <3x
Material Amount: 0	4	<input type="checkbox"/>	Clean	4	<input type="radio"/>	Out of Tolerance >3x <5x
Charge Number: 100348027	5	<input type="checkbox"/>	Limited Calibration	5	<input type="radio"/>	Out of Tolerance >5x
Cal Work Inst ID: PCB	6	<input type="checkbox"/>	Functional Check	6	<input type="radio"/>	Out of Tolerance-Undetermined
Outside Vendor: PCB	7	<input type="checkbox"/>	Performance Check	7	<input type="checkbox"/>	Inoperative
	8	<input type="checkbox"/>	Modify	8	<input type="checkbox"/>	Damaged
	9	<input type="checkbox"/>	Repair-needs Charge Level	9	<input type="checkbox"/>	Not Used
	10	<input type="checkbox"/>	Other	10	<input type="checkbox"/>	Not Determined
				11	<input type="checkbox"/>	Excessed
Calibrated By: Stan Zohner	S#:	58146	Phone: 526-2761	12	<input type="checkbox"/>	Extension

**CALIBRATION STANDARDS USED**


STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

**OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION**

NOMINAL (STD)	UNITS	AS FOUND (UUT)	MFG. ACCURACY
_____	_____	_____	_____
_____	_____	_____	_____
_____	_____	_____	_____
_____	_____	_____	_____
_____	_____	_____	_____

**COMMENTS**

PCB  
test referral QA# 102401

**INL  
CALIBRATION INPUT DATA**

11/29/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 716213      Mfr: PCB      Model: 353B03      Noun Name: ACCELEROMETER      Serial #: 38619

Calibration Date: 11/11/2004

**ACTION CODE**

**AS FOUND**

Next Cal Due Date: 10/21/2005	1	<input type="checkbox"/>	Acceptance Test	1	<input checked="" type="radio"/>	In Tolerance
Charge Level: 2	2	<input type="checkbox"/>	Special Test	2	<input type="radio"/>	Out of Tolerance >1x <2x
Repair/Adj/etc C.L.: 0	3	<input checked="" type="checkbox"/>	Calibration to MFG Specs	3	<input type="radio"/>	Out of Tolerance >2x <3x
Material Amount: 0	4	<input type="checkbox"/>	Clean	4	<input type="radio"/>	Out of Tolerance >3x <5x
Charge Number: 100664GSA	5	<input type="checkbox"/>	Limited Calibration	5	<input type="radio"/>	Out of Tolerance >5x
Cal Work Inst ID: PCB	6	<input type="checkbox"/>	Functional Check	6	<input type="radio"/>	Out of Tolerance-Undetermined
Outside Vendor: PCB	7	<input type="checkbox"/>	Performance Check	7	<input type="checkbox"/>	Inoperative
	8	<input type="checkbox"/>	Modify	8	<input type="checkbox"/>	Damaged
	9	<input type="checkbox"/>	Repair-needs Charge Level	9	<input type="checkbox"/>	Not Used
	10	<input type="checkbox"/>	Other	10	<input type="checkbox"/>	Not Determined
				11	<input type="checkbox"/>	Excessed
Calibrated By: Scott Lish				12	<input type="checkbox"/>	Extension
S#: 101141			Phone: 526-2761			

**CALIBRATION STANDARDS USED**


STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-5.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

NOMINAL (STD)	OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION UNITS	AS FOUND (UUT)	MFG. ACCURACY

**COMMENTS**

Unit sent to manufacture for calibration.  
QA# 111220

# INL CALIBRATION INPUT DATA

11/29/2007

NAME: DANA KEITH MORTON      BADGE: 35698      PH: 526-1274      AREA: STC      BLDG: EROB      RM: W2D1

ID Number: 716213      Mfr: PCB      Model: 353B03      Noun Name: ACCELEROMETER      Serial #: 38619

Calibration Date: 11/4/2005

## ACTION CODE

## AS FOUND

Next Cal Due Date: 11/4/2006

1 ☐

Acceptance Test

1 ☒ In Tolerance

Charge Level: 2

2 ☐

Special Test

2 ☐ Out of Tolerance >1x <2x

Repair/Adj/etc C.L: 0

3 ☒

Calibration to MFG Specs

3 ☐ Out of Tolerance >2x <3x

Material Amount: 0

4 ☐

Clean

4 ☐ Out of Tolerance >3x <5x

Charge Number: 100853GSA

5 ☐

Limited Calibration

5 ☐ Out of Tolerance >5x

Cal Work Inst ID: PCB

6 ☐

Functional Check

6 ☐ Out of Tolerance-Undetermined

Outside Vendor: PCB

7 ☐

Performance Check

7 ☐ Inoperative8 ☐

Modify

8 ☐ Damaged9 ☐

Repair-needs Charge Level

9 ☐ Not Used10 ☐

Other

10 ☐ Not Determined11 ☐ Excessed

Calibrated By: Scott Lish

S#: 101141 Phone: 526-2761

12 ☐ Extension

## CALIBRATION STANDARDS USED


STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

## LABORATORY TEMPERATURE AND HUMIDITY

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

## OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION

NOMINAL (STD)

UNITS

AS FOUND (UUT)

MFG. ACCURACY


## COMMENTS

QA#115894  
S&CL overcheck required.



**INL  
CALIBRATION INPUT DATA**

11/29/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 716213      Mfr: PCB      Model: 353B03      Noun Name: ACCELEROMETER      Serial #: 38619

Calibration Date: 12/14/2006

**ACTION CODE**

**AS FOUND**

Next Cal Due Date: 11/11/2007	1	<input type="checkbox"/>	Acceptance Test	1	<input checked="" type="radio"/>	In Tolerance
Charge Level: 2	2	<input type="checkbox"/>	Special Test	2	<input type="radio"/>	Out of Tolerance >1x <2x
Repair/Adj/etc C.L: 0	3	<input checked="" type="checkbox"/>	Calibration to MFG Specs	3	<input type="radio"/>	Out of Tolerance >2x <3x
Material Amount: 0	4	<input type="checkbox"/>	Clean	4	<input type="radio"/>	Out of Tolerance >3x <5x
Charge Number: 100853GSA	5	<input type="checkbox"/>	Limited Calibration	5	<input type="radio"/>	Out of Tolerance >5x
Cal Work Inst ID: PCB	6	<input type="checkbox"/>	Functional Check	6	<input type="radio"/>	Out of Tolerance-Undetermined
Outside Vendor: PCB	7	<input type="checkbox"/>	Performance Check	7	<input type="checkbox"/>	Inoperative
	8	<input type="checkbox"/>	Modify	8	<input type="checkbox"/>	Damaged
	9	<input type="checkbox"/>	Repair-needs Charge Level	9	<input type="checkbox"/>	Not Used
	10	<input type="checkbox"/>	Other	10	<input type="checkbox"/>	Not Determined
				11	<input type="checkbox"/>	Excessed
Calibrated By: Scott Lish	S#: 101141	Phone: 526-2761		12	<input type="checkbox"/>	Extension

**CALIBRATION STANDARDS USED**


STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

NOMINAL (STD)	OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION		MFG. ACCURACY
	UNITS	AS FOUND (UUT)	

**COMMENTS**

QA# 116282  
S&CL overcheck required.

**INL**  
**CALIBRATION INPUT DATA**

12/12/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 716213    Mfr: PCB    Model: 353B03    Noun Name: ACCELEROMETER    Serial #: 38619

Calibration Date: 12/11/2007

**ACTION CODE**

**AS FOUND**

Next Cal Due Date: 11/26/2008	1	<input type="checkbox"/>	Acceptance Test	1	<input checked="" type="radio"/> In Tolerance
Charge Level: 2	2	<input type="checkbox"/>	Special Test	2	<input type="radio"/> Out of Tolerance >1x <2x
Repair/Adj/etc C.L: 0	3	<input checked="" type="checkbox"/>	Calibration to MFG Specs	3	<input type="radio"/> Out of Tolerance >2x <3x
Material Amount: 0	4	<input type="checkbox"/>	Clean	4	<input type="radio"/> Out of Tolerance >3x <5x
Charge Number: 100853GSA	5	<input type="checkbox"/>	Limited Calibration	5	<input type="radio"/> Out of Tolerance >5x
Cal Work Inst ID: PCB	6	<input type="checkbox"/>	Functional Check	6	<input type="radio"/> Out of Tolerance-Undetermined
Outside Vendor: PCB	7	<input type="checkbox"/>	Performance Check	7	<input type="checkbox"/> Inoperative
	8	<input type="checkbox"/>	Modify	8	<input type="checkbox"/> Damaged
	9	<input type="checkbox"/>	Repair-needs Charge Level	9	<input type="checkbox"/> Not Used
	10	<input type="checkbox"/>	Other	10	<input type="checkbox"/> Not Determined
				11	<input type="checkbox"/> Exceeded
Calibrated By: Scott Lish    S#: 101141    Phone: 526-2761				12	<input type="checkbox"/> Extension

**CALIBRATION STANDARDS USED**


STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

NOMINAL (STD)	OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION UNITS	AS FOUND (UUT)	MFG. ACCURACY

**COMMENTS**

Calibrated at PCB  
QA# 141994  
S&CL overcheck required.

**INL**  
**CALIBRATION INPUT DATA**

11/29/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 716215    Mfr: PCB    Model: 353B03    Noun Name: ACCELEROMETER    Serial #: 38621

Calibration Date: 7/26/1999 12:53:52 PM

**ACTION CODE****AS FOUND**

Next Cal Due Date: 7/19/2000

Charge Level: 0

Repair/Adj/etc C.L: 0

Material Amount: 0

Charge Number: 827257038

Cal Work Inst ID: PCB

Outside Vendor: PCB

1 ☐ Acceptance Test2 ☐ Special Test3 ☒ Calibration to MFG Specs4 ☐ Clean5 ☐ Limited Calibration6 ☐ Functional Check7 ☐ Performance Check8 ☐ Modify9 ☐ Repair-needs Charge Level10 ☐ Other1 ☒ In Tolerance2 ☐ Out of Tolerance >1x <2x3 ☐ Out of Tolerance >2x <3x4 ☐ Out of Tolerance >3x <5x5 ☐ Out of Tolerance >5x6 ☐ Out of Tolerance-Undetermined7 ☐ Inoperative8 ☐ Damaged9 ☐ Not Used10 ☐ Not Determined11 ☐ Excessed12 ☐ Extension

Calibrated By: Mike Stears

S#: 61767 Phone: 526-2761

**CALIBRATION STANDARDS USED**


STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

NOMINAL (STD)	OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION	AS FOUND (UUT)	MFG. ACCURACY
	UNITS		


**COMMENTS**

**INL  
CALIBRATION INPUT DATA**

11/29/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 716215    Mfr: PCB    Model: 353B03    Noun Name: ACCELEROMETER    Serial #: 38621

Calibration Date: 7/17/2000 4:44:24 PM

**ACTION CODE****AS FOUND**

Next Cal Due Date: 7/12/2001

Charge Level: 1

Repair/Adj/etc C.L.: 0

Material Amount: 0

Charge Number: P44300001

Cal Work Inst ID: PCB

Outside Vendor: PCB

1 ☐ Acceptance Test2 ☐ Special Test3 ☒ Calibration to MFG Specs4 ☐ Clean5 ☐ Limited Calibration6 ☐ Functional Check7 ☐ Performance Check8 ☐ Modify9 ☐ Repair-needs Charge Level10 ☐ Other1 ☒ In Tolerance2 ☐ Out of Tolerance >1x <2x3 ☐ Out of Tolerance >2x <3x4 ☐ Out of Tolerance >3x <5x5 ☐ Out of Tolerance >5x6 ☐ Out of Tolerance-Undetermined7 ☐ Inoperative8 ☐ Damaged9 ☐ Not Used10 ☐ Not Determined11 ☐ Excessed12 ☐ Extension

Calibrated By: Mike Stears

S#: 61767 Phone: 526-2761

**CALIBRATION STANDARDS USED**


STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-5.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

**OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION**

NOMINAL (STD)

UNITS

AS FOUND (UUT)

MFG. ACCURACY


**COMMENTS**

**INL  
CALIBRATION INPUT DATA**

11/29/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 716215    Mfr: PCB    Model: 353B03    Noun Name: ACCELEROMETER    Serial #: 38621

Calibration Date: 8/19/2003

**ACTION CODE**

**AS FOUND**

Next Cal Due Date: 7/7/2004	1	<input type="checkbox"/>	Acceptance Test	1	<input checked="" type="radio"/>	In Tolerance
Charge Level: 2	2	<input type="checkbox"/>	Special Test	2	<input type="radio"/>	Out of Tolerance >1x <2x
Repair/Adj/etc C.L.: 0	3	<input checked="" type="checkbox"/>	Calibration to MFG Specs	3	<input type="radio"/>	Out of Tolerance >2x <3x
Material Amount: 0	4	<input type="checkbox"/>	Clean	4	<input type="radio"/>	Out of Tolerance >3x <5x
Charge Number: 100348027	5	<input type="checkbox"/>	Limited Calibration	5	<input type="radio"/>	Out of Tolerance >5x
Cal Work Inst ID: PCB	6	<input type="checkbox"/>	Functional Check	6	<input type="radio"/>	Out of Tolerance-Undetermined
Outside Vendor: PCB	7	<input type="checkbox"/>	Performance Check	7	<input type="checkbox"/>	Inoperative
	8	<input type="checkbox"/>	Modify	8	<input type="checkbox"/>	Damaged
	9	<input type="checkbox"/>	Repair-needs Charge Level	9	<input type="checkbox"/>	Not Used
	10	<input type="checkbox"/>	Other	10	<input type="checkbox"/>	Not Determined
				11	<input type="checkbox"/>	Excessed
				12	<input type="checkbox"/>	Extension

Calibrated By: Stan Zohner    S#: 58146    Phone: 526-2761

**CALIBRATION STANDARDS USED**


STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-5.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

NOMINAL (STD)	OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION UNITS	AS FOUND (UUT)	MFG. ACCURACY

**COMMENTS**

PCB  
test referral QA# 102401

**INL  
CALIBRATION INPUT DATA**

11/29/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 716215    Mfr: PCB    Model: 353B03    Noun Name: ACCELEROMETER    Serial #: 38621

Calibration Date: 6/21/2005

**ACTION CODE****AS FOUND**

Next Cal Due Date: 5/24/2006	1	<input type="checkbox"/>	Acceptance Test	1	<input checked="" type="radio"/>	In Tolerance
Charge Level: 2	2	<input type="checkbox"/>	Special Test	2	<input type="radio"/>	Out of Tolerance >1x <2x
Repair/Adj/etc C.L: 0	3	<input checked="" type="checkbox"/>	Calibration to MFG Specs	3	<input type="radio"/>	Out of Tolerance >2x <3x
Material Amount: 0	4	<input type="checkbox"/>	Clean	4	<input type="radio"/>	Out of Tolerance >3x <5x
Charge Number: 100853GSA	5	<input type="checkbox"/>	Limited Calibration	5	<input type="radio"/>	Out of Tolerance >5x
Cal Work Inst ID: PCB	6	<input type="checkbox"/>	Functional Check	6	<input type="radio"/>	Out of Tolerance-Undetermined
Outside Vendor: PCB	7	<input type="checkbox"/>	Performance Check	7	<input type="checkbox"/>	Inoperative
	8	<input type="checkbox"/>	Modify	8	<input type="checkbox"/>	Damaged
	9	<input type="checkbox"/>	Repair-needs Charge Level	9	<input type="checkbox"/>	Not Used
	10	<input type="checkbox"/>	Other	10	<input type="checkbox"/>	Not Determined
				11	<input type="checkbox"/>	Excessed
Calibrated By: Scott Lish				12	<input type="checkbox"/>	Extension

**CALIBRATION STANDARDS USED**


STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-5.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

NOMINAL (STD)	OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION UNITS	AS FOUND (UUT)	MFG. ACCURACY

**COMMENTS**

QA# 114602  
S&CL overcheck required.

**INL  
CALIBRATION INPUT DATA**

11/29/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 716215    Mfr: PCB    Model: 353B03    Noun Name: ACCELEROMETER    Serial #: 38621

Calibration Date: 6/1/2006

**ACTION CODE****AS FOUND**

Next Cal Due Date: 5/17/2007	1	<input type="checkbox"/>	Acceptance Test	1	<input checked="" type="radio"/>	In Tolerance
Charge Level: 2	2	<input type="checkbox"/>	Special Test	2	<input type="radio"/>	Out of Tolerance >1x <2x
Repair/Adj/etc C.L.: 0	3	<input checked="" type="checkbox"/>	Calibration to MFG Specs	3	<input type="radio"/>	Out of Tolerance >2x <3x
Material Amount: 0	4	<input type="checkbox"/>	Clean	4	<input type="radio"/>	Out of Tolerance >3x <5x
Charge Number: 100853GSA	5	<input type="checkbox"/>	Limited Calibration	5	<input type="radio"/>	Out of Tolerance >5x
Cal Work Inst ID: PCB	6	<input type="checkbox"/>	Functional Check	6	<input type="radio"/>	Out of Tolerance-Undetermined
Outside Vendor: PCB	7	<input type="checkbox"/>	Performance Check	7	<input type="checkbox"/>	Inoperative
	8	<input type="checkbox"/>	Modify	8	<input type="checkbox"/>	Damaged
	9	<input type="checkbox"/>	Repair-needs Charge Level	9	<input type="checkbox"/>	Not Used
	10	<input type="checkbox"/>	Other	10	<input type="checkbox"/>	Not Determined
				11	<input type="checkbox"/>	Excessed
Calibrated By: Scott Lish				12	<input type="checkbox"/>	Extension

**CALIBRATION STANDARDS USED**


STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-5.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

NOMINAL (STD)	OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION UNITS	AS FOUND (UUT)	MFG. ACCURACY

**COMMENTS**

QA# 117751  
S&CL overcheck required.

**INL  
CALIBRATION INPUT DATA**

11/29/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 716215    Mfr: PCB    Model: 353B03    Noun Name: ACCELEROMETER    Serial #: 38621

Calibration Date: 6/13/2007

**ACTION CODE****AS FOUND**

Next Cal Due Date: 5/24/2008	1	<input type="checkbox"/>	Acceptance Test	1	<input checked="" type="radio"/>	In Tolerance
Charge Level: 2	2	<input type="checkbox"/>	Special Test	2	<input type="radio"/>	Out of Tolerance >1x <2x
Repair/Adj/etc C.L: 0	3	<input checked="" type="checkbox"/>	Calibration to MFG Specs	3	<input type="radio"/>	Out of Tolerance >2x <3x
Material Amount: 0	4	<input type="checkbox"/>	Clean	4	<input type="radio"/>	Out of Tolerance >3x <5x
Charge Number: 100853GSA	5	<input type="checkbox"/>	Limited Calibration	5	<input type="radio"/>	Out of Tolerance >5x
Cal Work Inst ID: PCB	6	<input type="checkbox"/>	Functional Check	6	<input type="radio"/>	Out of Tolerance-Undetermined
Outside Vendor: PCB	7	<input type="checkbox"/>	Performance Check	7	<input type="checkbox"/>	Inoperative
	8	<input type="checkbox"/>	Modify	8	<input type="checkbox"/>	Damaged
	9	<input type="checkbox"/>	Repair-needs Charge Level	9	<input type="checkbox"/>	Not Used
	10	<input type="checkbox"/>	Other	10	<input type="checkbox"/>	Not Determined
				11	<input type="checkbox"/>	Excessed
				12	<input type="checkbox"/>	Extension

Calibrated By: Scott Lish    S#: 101141    Phone: 526-2761

**CALIBRATION STANDARDS USED**


STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

NOMINAL (STD)	OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION UNITS	AS FOUND (UUT)	MFG. ACCURACY

**COMMENTS**

QA# 134165  
S&CL overcheck required.



**INL  
CALIBRATION INPUT DATA**

11/29/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 716333	Mfr: PCB	Model: 350A04	Noun Name: ACCELEROMETER	Serial #: 1427
Calibration Date: 10/21/1999 4:22:01 PM	<b>ACTION CODE</b>		<b>AS FOUND</b>	
Next Cal Due Date:	1 <input type="checkbox"/> Acceptance Test	1 <input type="radio"/> In Tolerance		
Charge Level: 0	2 <input type="checkbox"/> Special Test	2 <input type="radio"/> Out of Tolerance >1x <2x		
Repair/Adj/etc C.L: 0	3 <input type="checkbox"/> Calibration to MFG Specs	3 <input type="radio"/> Out of Tolerance >2x <3x		
Material Amount: 0	4 <input type="checkbox"/> Clean	4 <input type="radio"/> Out of Tolerance >3x <5x		
Charge Number: 572121017	5 <input type="checkbox"/> Limited Calibration	5 <input type="radio"/> Out of Tolerance >5x		
Cal Work Inst ID: PCB	6 <input type="checkbox"/> Functional Check	6 <input type="radio"/> Out of Tolerance-Undetermined		
Outside Vendor: PCB	7 <input type="checkbox"/> Performance Check	7 <input type="checkbox"/> Inoperative		
	8 <input type="checkbox"/> Modify	8 <input type="checkbox"/> Damaged		
	9 <input type="checkbox"/> Repair-needs Charge Level	9 <input type="checkbox"/> Not Used		
	10 <input checked="" type="checkbox"/> Other	10 <input checked="" type="checkbox"/> Not Determined		
		11 <input type="checkbox"/> Excessed		
Calibrated By: Terry Wilde	S#: 57438 Phone: 526-2761	12 <input type="checkbox"/> Extension		

**CALIBRATION STANDARDS USED**


STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

NOMINAL (STD)	OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION		MFG. ACCURACY
	UNITS	AS FOUND (UUT)	

**COMMENTS**

RETURNED BACK TO USER

**INL**  
**CALIBRATION INPUT DATA**

11/29/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 716333	Mfr: PCB	Model: 350A04	Noun Name: ACCELEROMETER	Serial #: 1427
Calibration Date: 7/17/2000 4:57:16 PM	<b>ACTION CODE</b>		<b>AS FOUND</b>	
Next Cal Due Date: 7/12/2001	1 <input type="checkbox"/> Acceptance Test	1 <input checked="" type="radio"/> In Tolerance		
Charge Level: 1	2 <input type="checkbox"/> Special Test	2 <input type="radio"/> Out of Tolerance >1x <2x		
Repair/Adj/etc C.L: 0	3 <input checked="" type="checkbox"/> Calibration to MFG Specs	3 <input type="radio"/> Out of Tolerance >2x <3x		
Material Amount: 0	4 <input type="checkbox"/> Clean	4 <input type="radio"/> Out of Tolerance >3x <5x		
Charge Number: P44300001	5 <input type="checkbox"/> Limited Calibration	5 <input type="radio"/> Out of Tolerance >5x		
Cal Work Inst ID: PCB	6 <input type="checkbox"/> Functional Check	6 <input type="radio"/> Out of Tolerance-Undetermined		
Outside Vendor: PCB	7 <input type="checkbox"/> Performance Check	7 <input type="checkbox"/> Inoperative		
	8 <input type="checkbox"/> Modify	8 <input type="checkbox"/> Damaged		
	9 <input type="checkbox"/> Repair-needs Charge Level	9 <input type="checkbox"/> Not Used		
	10 <input type="checkbox"/> Other	10 <input type="checkbox"/> Not Determined		
		11 <input type="checkbox"/> Excessed		
Calibrated By: Mike Stears	S#: 61767 Phone: 526-2761	12 <input type="checkbox"/> Extension		

**CALIBRATION STANDARDS USED**


STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-5.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

NOMINAL (STD)	OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION UNITS	AS FOUND (UUT)	MFG. ACCURACY

COMMENTS

# **INL** **CALIBRATION INPUT DATA**

11/29/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 716333 Mfr: PCB Model: 350A04 Noun Name: ACCELEROMETER Serial #: 1427

Calibration Date: 8/21/2003 11:44:08 AM

**ACTION CODE****AS FOUND**

Next Cal Due Date: 8/21/2004

Charge Level: 2

Repair/Adj/etc C.L: 0

Material Amount: 0

Charge Number: 100348027

Cal Work Inst ID: PCB

Outside Vendor: PCB

1 ☐ Acceptance Test2 ☐ Special Test3 ☒ Calibration to MFG Specs4 ☐ Clean5 ☐ Limited Calibration6 ☐ Functional Check7 ☐ Performance Check8 ☐ Modify9 ☐ Repair-needs Charge Level10 ☐ Other1 ☒ In Tolerance2 ☐ Out of Tolerance >1x <2x3 ☐ Out of Tolerance >2x <3x4 ☐ Out of Tolerance >3x <5x5 ☐ Out of Tolerance >5x6 ☐ Out of Tolerance-Undetermined7 ☐ Inoperative8 ☐ Damaged9 ☐ Not Used10 ☐ Not Determined11 ☐ Excessed12 ☐ Extension

Calibrated By: Stan Zohner

S#: 58146 Phone: 526-2761

**CALIBRATION STANDARDS USED**


STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED  
VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-5.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

NOMINAL (STD)	OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION UNITS	AS FOUND (UUT)	MFG. ACCURACY

**COMMENTS**

PCB  
QA# 102391

**INL  
CALIBRATION INPUT DATA**

11/29/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 716333    Mfr: PCB    Model: 350A04    Noun Name: ACCELEROMETER    Serial #: 1427

Calibration Date: 6/21/2005

**ACTION CODE**

**AS FOUND**

Next Cal Due Date: 5/24/2006	1	<input type="checkbox"/>	Acceptance Test	1	<input checked="" type="radio"/>	In Tolerance
Charge Level: 2	2	<input type="checkbox"/>	Special Test	2	<input type="radio"/>	Out of Tolerance >1x <2x
Repair/Adj/etc C.L.: 0	3	<input checked="" type="checkbox"/>	Calibration to MFG Specs	3	<input type="radio"/>	Out of Tolerance >2x <3x
Material Amount: 0	4	<input type="checkbox"/>	Clean	4	<input type="radio"/>	Out of Tolerance >3x <5x
Charge Number: 100853GSA	5	<input type="checkbox"/>	Limited Calibration	5	<input type="radio"/>	Out of Tolerance >5x
Cal Work Inst ID: PCB	6	<input type="checkbox"/>	Functional Check	6	<input type="radio"/>	Out of Tolerance-Undetermined
Outside Vendor: PCB	7	<input type="checkbox"/>	Performance Check	7	<input type="checkbox"/>	Inoperative
	8	<input type="checkbox"/>	Modify	8	<input type="checkbox"/>	Damaged
	9	<input type="checkbox"/>	Repair-needs Charge Level	9	<input type="checkbox"/>	Not Used
	10	<input type="checkbox"/>	Other	10	<input type="checkbox"/>	Not Determined
				11	<input type="checkbox"/>	Excessed
Calibrated By: Scott Lish				12	<input type="checkbox"/>	Extension

**CALIBRATION STANDARDS USED**


STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-5.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

NOMINAL (STD)	OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION		MFG. ACCURACY
	UNITS	AS FOUND (UUT)	

**COMMENTS**

QA# 114639  
S&CL overcheck required.

**INL**  
**CALIBRATION INPUT DATA**

11/29/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 716333    Mfr: PCB    Model: 350A04    Noun Name: ACCELEROMETER    Serial #: 1427

Calibration Date: 6/1/2006

**ACTION CODE**

**AS FOUND**

Next Cal Due Date: 5/17/2007	1	<input type="checkbox"/>	Acceptance Test	1	<input checked="" type="radio"/>	In Tolerance
Charge Level: 2	2	<input type="checkbox"/>	Special Test	2	<input type="radio"/>	Out of Tolerance >1x <2x
Repair/Adj/etc C.L: 0	3	<input checked="" type="checkbox"/>	Calibration to MFG Specs	3	<input type="radio"/>	Out of Tolerance >2x <3x
Material Amount: 0	4	<input type="checkbox"/>	Clean	4	<input type="radio"/>	Out of Tolerance >3x <5x
Charge Number: 100853GSA	5	<input type="checkbox"/>	Limited Calibration	5	<input type="radio"/>	Out of Tolerance >5x
Cal Work Inst ID: PCB	6	<input type="checkbox"/>	Functional Check	6	<input type="radio"/>	Out of Tolerance-Undetermined
Outside Vendor: PCB	7	<input type="checkbox"/>	Performance Check	7	<input type="checkbox"/>	Inoperative
	8	<input type="checkbox"/>	Modify	8	<input type="checkbox"/>	Damaged
	9	<input type="checkbox"/>	Repair-needs Charge Level	9	<input type="checkbox"/>	Not Used
	10	<input type="checkbox"/>	Other	10	<input type="checkbox"/>	Not Determined
				11	<input type="checkbox"/>	Excessed
Calibrated By: Scott Lish				12	<input type="checkbox"/>	Extension

**CALIBRATION STANDARDS USED**


STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-5.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

NOMINAL (STD)	OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION		MFG. ACCURACY
	UNITS	AS FOUND (UUT)	

**COMMENTS**

QA# 117750  
S&CL overcheck required.

**INL**  
**CALIBRATION INPUT DATA**

11/29/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 716333    Mfr: PCB    Model: 350A04    Noun Name: ACCELEROMETER    Serial #: 1427

Calibration Date: 6/13/2007

**ACTION CODE****AS FOUND**

Next Cal Due Date: 5/24/2008

Charge Level: 2

Repair/Adj/etc C.L.: 0

Material Amount: 0

Charge Number: 100853GSA

Cal Work Inst ID: PCB

Outside Vendor: PCB

1 ☐

Acceptance Test

2 ☐

Special Test

3 ☒

Calibration to MFG Specs

4 ☐

Clean

5 ☐

Limited Calibration

6 ☐

Functional Check

7 ☐

Performance Check

8 ☐

Modify

9 ☐

Repair-needs Charge Level

10 ☐

Other

1 ☒ In Tolerance2 ☐ Out of Tolerance >1x <2x3 ☐ Out of Tolerance >2x <3x4 ☐ Out of Tolerance >3x <5x5 ☐ Out of Tolerance >5x6 ☐ Out of Tolerance-Undetermined7 ☐ Inoperative8 ☐ Damaged9 ☐ Not Used10 ☐ Not Determined11 ☐ Excessed12 ☐ Extension

Calibrated By: Scott Lish

S#: 101141 Phone: 526-2761

**CALIBRATION STANDARDS USED**


STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

**OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION**

NOMINAL (STD)

UNITS

AS FOUND (UUT)

MFG. ACCURACY


**COMMENTS**

QA# 134166

S&amp;CL overcheck required.

**INL**  
**CALIBRATION INPUT DATA**

11/29/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 717830    Mfr: PCB    Model: 350B04    Noun Name: ACCELEROMETER    Serial #: 6203

Calibration Date: 9/18/2000 1:53:25 PM

**ACTION CODE****AS FOUND**

Next Cal Due Date: 8/15/2001

Charge Level: 1

Repair/Adj/etc C.L: 0

Material Amount: 0

Charge Number: 572152030

Cal Work Inst ID: PCB

Outside Vendor: PCB

1 ☐ Acceptance Test2 ☐ Special Test3 ☒ Calibration to MFG Specs4 ☐ Clean5 ☐ Limited Calibration6 ☐ Functional Check7 ☐ Performance Check8 ☐ Modify9 ☐ Repair-needs Charge Level10 ☐ Other1 ☒ In Tolerance2 ☐ Out of Tolerance >1x <2x3 ☐ Out of Tolerance >2x <3x4 ☐ Out of Tolerance >3x <5x5 ☐ Out of Tolerance >5x6 ☐ Out of Tolerance-Undetermined7 ☐ Inoperative8 ☐ Damaged9 ☐ Not Used10 ☐ Not Determined11 ☐ Excessed12 ☐ Extension

Calibrated By: Mike Stears

S#: 61767 Phone: 526-2761

**CALIBRATION STANDARDS USED**


**STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES**

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

**Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.**

NOMINAL (STD)	OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION		MFG. ACCURACY
	UNITS	AS FOUND (UUT)	

**COMMENTS**

**INL**  
**CALIBRATION INPUT DATA**

11/29/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 717830      Mfr: PCB      Model: 350B04      Noun Name: ACCELEROMETER      Serial #: 6203

Calibration Date: 8/21/2003 12:01:55 PM

**ACTION CODE****AS FOUND**

Next Cal Due Date: 8/21/2004

Charge Level: 2

Repair/Adj/etc C.L: 0

Material Amount: 0

Charge Number: 100348027

Cal Work Inst ID: PCB

Outside Vendor: PCB

1 ☐ Acceptance Test2 ☐ Special Test3 ☒ Calibration to MFG Specs4 ☐ Clean5 ☐ Limited Calibration6 ☐ Functional Check7 ☐ Performance Check8 ☐ Modify9 ☐ Repair-needs Charge Level10 ☐ Other1 ☒ In Tolerance2 ☐ Out of Tolerance >1x <2x3 ☐ Out of Tolerance >2x <3x4 ☐ Out of Tolerance >3x <5x5 ☐ Out of Tolerance >5x6 ☐ Out of Tolerance-Undetermined7 ☐ Inoperative8 ☐ Damaged9 ☐ Not Used10 ☐ Not Determined11 ☐ Excessed12 ☐ Extension

Calibrated By: Stan Zohner

S#: 58146 Phone: 526-2761

**CALIBRATION STANDARDS USED**


STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

**OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION**

NOMINAL (STD)	UNITS	AS FOUND (UUT)	MFG. ACCURACY

**COMMENTS**

PCB  
QA# 102393



**INL**  
**CALIBRATION INPUT DATA**

11/29/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 717830    Mfr: PCB    Model: 350B04    Noun Name: ACCELEROMETER    Serial #: 6203

Calibration Date: 11/11/2004

**ACTION CODE**

**AS FOUND**

Next Cal Due Date: 10/21/2005	1	<input type="checkbox"/>	Acceptance Test	1	<input checked="" type="radio"/>	In Tolerance
Charge Level: 2	2	<input type="checkbox"/>	Special Test	2	<input type="radio"/>	Out of Tolerance >1x <2x
Repair/Adj/etc C.L: 0	3	<input checked="" type="checkbox"/>	Calibration to MFG Specs	3	<input type="radio"/>	Out of Tolerance >2x <3x
Material Amount: 0	4	<input type="checkbox"/>	Clean	4	<input type="radio"/>	Out of Tolerance >3x <5x
Charge Number: 100664GSA	5	<input type="checkbox"/>	Limited Calibration	5	<input type="radio"/>	Out of Tolerance >5x
Cal Work Inst ID: PCB	6	<input type="checkbox"/>	Functional Check	6	<input type="radio"/>	Out of Tolerance-Undetermined
Outside Vendor: PCB	7	<input type="checkbox"/>	Performance Check	7	<input type="checkbox"/>	Inoperative
	8	<input type="checkbox"/>	Modify	8	<input type="checkbox"/>	Damaged
	9	<input type="checkbox"/>	Repair-needs Charge Level	9	<input type="checkbox"/>	Not Used
	10	<input type="checkbox"/>	Other	10	<input type="checkbox"/>	Not Determined
				11	<input type="checkbox"/>	Excessed
Calibrated By: Scott Lish			S#: 101141 Phone: 526-2761	12	<input type="checkbox"/>	Extension

**CALIBRATION STANDARDS USED**


STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

NOMINAL (STD)	OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION		MFG. ACCURACY
	UNITS	AS FOUND (UUT)	

**COMMENTS**

Unit sent to manufacture for calibration.  
QA# 111221

**INL**  
**CALIBRATION INPUT DATA**

11/29/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 717830    Mfr: PCB    Model: 350B04    Noun Name: ACCELEROMETER    Serial #: 6203

Calibration Date: 11/7/2005

**ACTION CODE****AS FOUND**

Next Cal Due Date: 11/7/2006

Charge Level: 2

Repair/Adj/etc C.L.: 0

Material Amount: 0

Charge Number: 100853GSA

Cal Work Inst ID: PCB

Outside Vendor: PCB

1 ☐

Acceptance Test

2 ☐

Special Test

3 ☒

Calibration to MFG Specs

4 ☐

Clean

5 ☐

Limited Calibration

6 ☐

Functional Check

7 ☐

Performance Check

8 ☐

Modify

9 ☐

Repair-needs Charge Level

10 ☐

Other

1 ☒ In Tolerance2 ☐ Out of Tolerance >1x <2x3 ☐ Out of Tolerance >2x <3x4 ☐ Out of Tolerance >3x <5x5 ☐ Out of Tolerance >5x6 ☐ Out of Tolerance-Undetermined7 ☐ Inoperative8 ☐ Damaged9 ☐ Not Used10 ☐ Not Determined11 ☐ Excessed12 ☐ Extension

Calibrated By: Scott Lish

S#: 101141 Phone: 526-2761

**CALIBRATION STANDARDS USED**


STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)

20.0 ° +/-0.3 ° C (40-55% RH)

Electronic STD (106D)

23.0 ° +/-0.5 ° C (30-45% RH)

Dimensional STD (106B)

20.0 ° +/-0.25 ° C (30-45% RH)

Electronic CAL (Lab 112)

23.0 ° +/-1.0 ° C (20-50% RH)

Phys/Dim CAL (Lab 111)

20.0 ° +/-0.5 ° C (20-50% RH)

Remaining S&amp;CL calibration areas:

23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

**OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION****NOMINAL (STD)****UNITS****AS FOUND (UUT)****MFG. ACCURACY**


**COMMENTS**

QA#115895

S&amp;CL overcheck required.

**INL**  
**CALIBRATION INPUT DATA**

11/29/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 717830    Mfr: PCB    Model: 350B04    Noun Name: ACCELEROMETER    Serial #: 6203

Calibration Date: 12/14/2006

**ACTION CODE**

**AS FOUND**

Next Cal Due Date: 11/29/2007	1	<input type="checkbox"/>	Acceptance Test	1	<input checked="" type="radio"/>	In Tolerance
Charge Level: 2	2	<input type="checkbox"/>	Special Test	2	<input type="radio"/>	Out of Tolerance >1x <2x
Repair/Adj/etc C.L: 0	3	<input checked="" type="checkbox"/>	Calibration to MFG Specs	3	<input type="radio"/>	Out of Tolerance >2x <3x
Material Amount: 0	4	<input type="checkbox"/>	Clean	4	<input type="radio"/>	Out of Tolerance >3x <5x
Charge Number: 100853GSA	5	<input type="checkbox"/>	Limited Calibration	5	<input type="radio"/>	Out of Tolerance >5x
Cal Work Inst ID: PCB	6	<input type="checkbox"/>	Functional Check	6	<input type="radio"/>	Out of Tolerance-Undetermined
Outside Vendor: PCB	7	<input type="checkbox"/>	Performance Check	7	<input type="checkbox"/>	Inoperative
	8	<input type="checkbox"/>	Modify	8	<input type="checkbox"/>	Damaged
	9	<input type="checkbox"/>	Repair-needs Charge Level	9	<input type="checkbox"/>	Not Used
	10	<input type="checkbox"/>	Other	10	<input type="checkbox"/>	Not Determined
				11	<input type="checkbox"/>	Excessed
Calibrated By: Scott Lish				12	<input type="checkbox"/>	Extension

**CALIBRATION STANDARDS USED**


STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

NOMINAL (STD)	OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION		MFG. ACCURACY
	UNITS	AS FOUND (UUT)	

**COMMENTS**

QA# 122338  
S&CL overcheck required.

**INL**  
**CALIBRATION INPUT DATA**

12/12/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 717830    Mfr: PCB    Model: 350B04    Noun Name: ACCELEROMETER    Serial #: 6203

			ACTION CODE	AS FOUND
Next Cal Due Date: 11/26/2008	1	<input type="checkbox"/>	Acceptance Test	1 <input checked="" type="radio"/> In Tolerance
Charge Level: 2	2	<input type="checkbox"/>	Special Test	2 <input type="radio"/> Out of Tolerance >1x <2x
Repair/Adj/etc C.L: 0	3	<input checked="" type="checkbox"/>	Calibration to MFG Specs	3 <input type="radio"/> Out of Tolerance >2x <3x
Material Amount: 0	4	<input type="checkbox"/>	Clean	4 <input type="radio"/> Out of Tolerance >3x <5x
Charge Number: 100853GSA	5	<input type="checkbox"/>	Limited Calibration	5 <input type="radio"/> Out of Tolerance >5x
Cal Work Inst ID: PCB	6	<input type="checkbox"/>	Functional Check	6 <input type="radio"/> Out of Tolerance-Undetermined
Outside Vendor: PCB	7	<input type="checkbox"/>	Performance Check	7 <input type="checkbox"/> Inoperative
	8	<input type="checkbox"/>	Modify	8 <input type="checkbox"/> Damaged
	9	<input type="checkbox"/>	Repair-needs Charge Level	9 <input type="checkbox"/> Not Used
	10	<input type="checkbox"/>	Other	10 <input type="checkbox"/> Not Determined
				11 <input type="checkbox"/> Exceeded
Calibrated By: Scott Lish    S#: 101141    Phone: 526-2761				12 <input type="checkbox"/> Extension

**CALIBRATION STANDARDS USED**


STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

NOMINAL (STD)	OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION		MFG. ACCURACY
	UNITS	AS FOUND (UUT)	

**COMMENTS**

Calibrated at PCB  
QA# 141994  
S&CL overcheck required.

**INL**  
**CALIBRATION INPUT DATA**

11/29/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 721278      Mfr: PCB      Model: 350B04      Noun Name: ACCELEROMETER      Serial #: 6657

Calibration Date: 8/21/2003 12:00:38 PM

**ACTION CODE****AS FOUND**

Next Cal Due Date: 8/21/2004

Charge Level: 2

Repair/Adj/etc C.L.: 0

Material Amount: 0

Charge Number: 100348027

Cal Work Inst ID: PCB

Outside Vendor: PCB

1 ☐ Acceptance Test2 ☐ Special Test3 ☒ Calibration to MFG Specs4 ☐ Clean5 ☐ Limited Calibration6 ☐ Functional Check7 ☐ Performance Check8 ☐ Modify9 ☐ Repair-needs Charge Level10 ☐ Other1 ☒ In Tolerance2 ☐ Out of Tolerance >1x <2x3 ☐ Out of Tolerance >2x <3x4 ☐ Out of Tolerance >3x <5x5 ☐ Out of Tolerance >5x6 ☐ Out of Tolerance-Undetermined7 ☐ Inoperative8 ☐ Damaged9 ☐ Not Used10 ☐ Not Determined11 ☐ Excessed12 ☐ Extension

Calibrated By: Stan Zohner

S#: 58146 Phone: 526-2761

**CALIBRATION STANDARDS USED**


STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

**OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION****NOMINAL (STD)****UNITS****AS FOUND (UUT)****MFG. ACCURACY**


**COMMENTS**

PCB  
QA# 102393

**INL**  
**CALIBRATION INPUT DATA**

11/29/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 721278      Mfr: PCB      Model: 350B04      Noun Name: ACCELEROMETER      Serial #: 6657

Calibration Date: 11/11/2004

**ACTION CODE****AS FOUND**

Next Cal Due Date: 10/21/2005	1	<input type="checkbox"/>	Acceptance Test	1	<input checked="" type="radio"/>	In Tolerance
Charge Level: 2	2	<input type="checkbox"/>	Special Test	2	<input type="radio"/>	Out of Tolerance >1x <2x
Repair/Adj/etc C.L: 0	3	<input checked="" type="checkbox"/>	Calibration to MFG Specs	3	<input type="radio"/>	Out of Tolerance >2x <3x
Material Amount: 0	4	<input type="checkbox"/>	Clean	4	<input type="radio"/>	Out of Tolerance >3x <5x
Charge Number: 100664GSA	5	<input type="checkbox"/>	Limited Calibration	5	<input type="radio"/>	Out of Tolerance >5x
Cal Work Inst ID: PCB	6	<input type="checkbox"/>	Functional Check	6	<input type="radio"/>	Out of Tolerance-Undetermined
Outside Vendor: PCB	7	<input type="checkbox"/>	Performance Check	7	<input type="checkbox"/>	Inoperative
	8	<input type="checkbox"/>	Modify	8	<input type="checkbox"/>	Damaged
	9	<input type="checkbox"/>	Repair-needs Charge Level	9	<input type="checkbox"/>	Not Used
	10	<input type="checkbox"/>	Other	10	<input type="checkbox"/>	Not Determined
				11	<input type="checkbox"/>	Excessed
Calibrated By: Scott Lish				12	<input type="checkbox"/>	Extension

**CALIBRATION STANDARDS USED**


STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

NOMINAL (STD)	OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION UNITS	AS FOUND (UUT)	MFG. ACCURACY

**COMMENTS**

Sent to manufacture for calibration.  
QA# 111221

**INL  
CALIBRATION INPUT DATA**

11/29/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 721278    Mfr: PCB    Model: 350B04    Noun Name: ACCELEROMETER    Serial #: 6657

Calibration Date: 11/7/2005

**ACTION CODE****AS FOUND**

Next Cal Due Date: 11/7/2006	1	<input type="checkbox"/>	Acceptance Test	1	<input checked="" type="radio"/>	In Tolerance
Charge Level: 2	2	<input type="checkbox"/>	Special Test	2	<input type="radio"/>	Out of Tolerance >1x <2x
Repair/Adj/etc C.L.: 0	3	<input checked="" type="checkbox"/>	Calibration to MFG Specs	3	<input type="radio"/>	Out of Tolerance >2x <3x
Material Amount: 0	4	<input type="checkbox"/>	Clean	4	<input type="radio"/>	Out of Tolerance >3x <5x
Charge Number: 100853GSA	5	<input type="checkbox"/>	Limited Calibration	5	<input type="radio"/>	Out of Tolerance >5x
Cal Work Inst ID: PCB	6	<input type="checkbox"/>	Functional Check	6	<input type="radio"/>	Out of Tolerance-Undetermined
Outside Vendor: PCB	7	<input type="checkbox"/>	Performance Check	7	<input type="checkbox"/>	Inoperative
	8	<input type="checkbox"/>	Modify	8	<input type="checkbox"/>	Damaged
	9	<input type="checkbox"/>	Repair-needs Charge Level	9	<input type="checkbox"/>	Not Used
	10	<input type="checkbox"/>	Other	10	<input type="checkbox"/>	Not Determined
				11	<input type="checkbox"/>	Excessed
Calibrated By: Scott Lish				12	<input type="checkbox"/>	Extension

S#: 101141 Phone: 526-2761

**CALIBRATION STANDARDS USED**


STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

NOMINAL (STD)	OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION UNITS	AS FOUND (UUT)	MFG. ACCURACY

**COMMENTS**

QA#115895  
S&CL overcheck required.

**INL  
CALIBRATION INPUT DATA**

11/29/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 721278    Mfr: PCB    Model: 350B04    Noun Name: ACCELEROMETER    Serial #: 6657

Calibration Date: 12/14/2006

**ACTION CODE****AS FOUND**

Next Cal Due Date: 11/29/2007	1	<input type="checkbox"/>	Acceptance Test	1	<input checked="" type="radio"/>	In Tolerance
Charge Level: 2	2	<input type="checkbox"/>	Special Test	2	<input type="radio"/>	Out of Tolerance >1x <2x
Repair/Adj/etc C.L: 0	3	<input checked="" type="checkbox"/>	Calibration to MFG Specs	3	<input type="radio"/>	Out of Tolerance >2x <3x
Material Amount: 0	4	<input type="checkbox"/>	Clean	4	<input type="radio"/>	Out of Tolerance >3x <5x
Charge Number: 100853GSA	5	<input type="checkbox"/>	Limited Calibration	5	<input type="radio"/>	Out of Tolerance >5x
Cal Work Inst ID: PCB	6	<input type="checkbox"/>	Functional Check	6	<input type="radio"/>	Out of Tolerance-Undetermined
Outside Vendor: PCB	7	<input type="checkbox"/>	Performance Check	7	<input type="checkbox"/>	Inoperative
	8	<input type="checkbox"/>	Modify	8	<input type="checkbox"/>	Damaged
	9	<input type="checkbox"/>	Repair-needs Charge Level	9	<input type="checkbox"/>	Not Used
	10	<input type="checkbox"/>	Other	10	<input type="checkbox"/>	Not Determined
				11	<input type="checkbox"/>	Excessed
				12	<input type="checkbox"/>	Extension

Calibrated By: Scott Lish    S#: 101141    Phone: 526-2761

**CALIBRATION STANDARDS USED**


STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-5.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

NOMINAL (STD)	OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION UNITS	AS FOUND (UUT)	MFG. ACCURACY

**COMMENTS**

QA# 122338  
S&CL overcheck required.



# **INL** **CALIBRATION INPUT DATA**

12/12/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 721278      Mfr: PCB      Model: 350B04      Noun Name: ACCELEROMETER      Serial #: 6657

Calibration Date: 12/11/2007

**ACTION CODE****AS FOUND**

Next Cal Due Date: 11/26/2008	1	<input type="checkbox"/>	Acceptance Test	1	<input checked="" type="radio"/>	In Tolerance
Charge Level: 2	2	<input type="checkbox"/>	Special Test	2	<input type="radio"/>	Out of Tolerance >1x <2x
Repair/Adj/etc C.L.: 0	3	<input checked="" type="checkbox"/>	Calibration to MFG Specs	3	<input type="radio"/>	Out of Tolerance >2x <3x
Material Amount: 0	4	<input type="checkbox"/>	Clean	4	<input type="radio"/>	Out of Tolerance >3x <5x
Charge Number: 100853GSA	5	<input type="checkbox"/>	Limited Calibration	5	<input type="radio"/>	Out of Tolerance >5x
Cal Work Inst ID: PCB	6	<input type="checkbox"/>	Functional Check	6	<input type="radio"/>	Out of Tolerance-Undetermined
Outside Vendor: PCB	7	<input type="checkbox"/>	Performance Check	7	<input type="checkbox"/>	Inoperative
	8	<input type="checkbox"/>	Modify	8	<input type="checkbox"/>	Damaged
	9	<input type="checkbox"/>	Repair-needs Charge Level	9	<input type="checkbox"/>	Not Used
	10	<input type="checkbox"/>	Other	10	<input type="checkbox"/>	Not Determined
				11	<input type="checkbox"/>	Excessed
Calibrated By: Scott Lish				12	<input type="checkbox"/>	Extension

S#: 101141 Phone: 526-2761

**CALIBRATION STANDARDS USED**


STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

NOMINAL (STD)	OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION UNITS	AS FOUND (UUT)	MFG. ACCURACY

**COMMENTS**

Calibrated at PCB  
QA# 141994  
S&CL overcheck required.

**INL**  
**CALIBRATION INPUT DATA**

11/30/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 722795    Mfr: PCB    Model: 353B03    Noun Name: ACCELEROMETER    Serial #: 85173

Calibration Date: 6/21/2005

**ACTION CODE**

**AS FOUND**

Next Cal Due Date: 5/24/2006	1	<input type="checkbox"/>	Acceptance Test	1	<input checked="" type="radio"/>	In Tolerance
Charge Level: 2	2	<input type="checkbox"/>	Special Test	2	<input type="radio"/>	Out of Tolerance >1x <2x
Repair/Adj/etc C.L.: 0	3	<input checked="" type="checkbox"/>	Calibration to MFG Specs	3	<input type="radio"/>	Out of Tolerance >2x <3x
Material Amount: 0	4	<input type="checkbox"/>	Clean	4	<input type="radio"/>	Out of Tolerance >3x <5x
Charge Number: 100853GSA	5	<input type="checkbox"/>	Limited Calibration	5	<input type="radio"/>	Out of Tolerance >5x
Cal Work Inst ID: PCB	6	<input type="checkbox"/>	Functional Check	6	<input type="radio"/>	Out of Tolerance-Undetermined
Outside Vendor: PCB	7	<input type="checkbox"/>	Performance Check	7	<input type="checkbox"/>	Inoperative
	8	<input type="checkbox"/>	Modify	8	<input type="checkbox"/>	Damaged
	9	<input type="checkbox"/>	Repair-needs Charge Level	9	<input type="checkbox"/>	Not Used
	10	<input type="checkbox"/>	Other	10	<input type="checkbox"/>	Not Determined
				11	<input type="checkbox"/>	Excessed
				12	<input type="checkbox"/>	Extension

Calibrated By: Scott Lish    S#: 101141    Phone: 526-2761

**CALIBRATION STANDARDS USED**


STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-5.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

NOMINAL (STD)	OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION		MFG. ACCURACY
	UNITS	AS FOUND (UUT)	

**COMMENTS**

QA# 114602  
S&CL overcheck required.

**INL**  
**CALIBRATION INPUT DATA**

11/30/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 722795    Mfr: PCB    Model: 353B03    Noun Name: ACCELEROMETER    Serial #: 85173

Calibration Date: 6/1/2006

**ACTION CODE****AS FOUND**

Next Cal Due Date: 5/17/2007	1	<input type="checkbox"/>	Acceptance Test	1	<input checked="" type="radio"/>	In Tolerance
Charge Level: 2	2	<input type="checkbox"/>	Special Test	2	<input type="radio"/>	Out of Tolerance >1x <2x
Repair/Adj/etc C.L.: 0	3	<input checked="" type="checkbox"/>	Calibration to MFG Specs	3	<input type="radio"/>	Out of Tolerance >2x <3x
Material Amount: 0	4	<input type="checkbox"/>	Clean	4	<input type="radio"/>	Out of Tolerance >3x <5x
Charge Number: 100853GSA	5	<input type="checkbox"/>	Limited Calibration	5	<input type="radio"/>	Out of Tolerance >5x
Cal Work Inst ID: PCB	6	<input type="checkbox"/>	Functional Check	6	<input type="radio"/>	Out of Tolerance-Undetermined
Outside Vendor: PCB	7	<input type="checkbox"/>	Performance Check	7	<input type="checkbox"/>	Inoperative
	8	<input type="checkbox"/>	Modify	8	<input type="checkbox"/>	Damaged
	9	<input type="checkbox"/>	Repair-needs Charge Level	9	<input type="checkbox"/>	Not Used
	10	<input type="checkbox"/>	Other	10	<input type="checkbox"/>	Not Determined
				11	<input type="checkbox"/>	Excessed
Calibrated By: Scott Lish				12	<input type="checkbox"/>	Extension

**CALIBRATION STANDARDS USED**


STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-5,-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

NOMINAL (STD)	OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION UNITS	AS FOUND (UUT)	MFG. ACCURACY

**COMMENTS**

QA# 1147751  
S&CL overcheck required.

**INL**  
**CALIBRATION INPUT DATA**

11/30/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 722795    Mfr: PCB    Model: 353B03    Noun Name: ACCELEROMETER    Serial #: 85173

Calibration Date: 6/13/2007

**ACTION CODE**

**AS FOUND**

Next Cal Due Date: 5/24/2008	1	<input type="checkbox"/>	Acceptance Test	1	<input checked="" type="radio"/>	In Tolerance
Charge Level: 2	2	<input type="checkbox"/>	Special Test	2	<input type="radio"/>	Out of Tolerance >1x <2x
Repair/Adj/etc C.L: 0	3	<input checked="" type="checkbox"/>	Calibration to MFG Specs	3	<input type="radio"/>	Out of Tolerance >2x <3x
Material Amount: 0	4	<input type="checkbox"/>	Clean	4	<input type="radio"/>	Out of Tolerance >3x <5x
Charge Number: 100853GSA	5	<input type="checkbox"/>	Limited Calibration	5	<input type="radio"/>	Out of Tolerance >5x
Cal Work Inst ID: PCB	6	<input type="checkbox"/>	Functional Check	6	<input type="radio"/>	Out of Tolerance-Undetermined
Outside Vendor: PCB	7	<input type="checkbox"/>	Performance Check	7	<input type="checkbox"/>	Inoperative
	8	<input type="checkbox"/>	Modify	8	<input type="checkbox"/>	Damaged
	9	<input type="checkbox"/>	Repair-needs Charge Level	9	<input type="checkbox"/>	Not Used
	10	<input type="checkbox"/>	Other	10	<input type="checkbox"/>	Not Determined
				11	<input type="checkbox"/>	Excessed
				12	<input type="checkbox"/>	Extension

Calibrated By: Scott Lish    S#: 101141    Phone: 526-2761

**CALIBRATION STANDARDS USED**


STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-5.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

NOMINAL (STD)	OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION		MFG. ACCURACY
	UNITS	AS FOUND (UUT)	

**COMMENTS**

QA# 134165  
S&CL overcheck required.

# **CALIPERS**

**INL  
CALIBRATION INPUT DATA**

11/29/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 716309      Mfr: FOWLER      Model: 6 INCH      Noun Name: DIGITAL CALIPER      Serial #: S225

Calibration Date: 8/23/1999 12:20:04 PM

**ACTION CODE****AS FOUND**

Next Cal Due Date: 8/23/2000

1 ☐ Acceptance Test1 ☒ In Tolerance

Charge Level: 1

2 ☐ Special Test2 ☐ Out of Tolerance >1x <2x

Repair/Adj/etc C.L: 0

3 ☒ Calibration to MFG Specs3 ☐ Out of Tolerance >2x <3x

Material Amount: 0

4 ☐ Clean4 ☐ Out of Tolerance >3x <5x

Charge Number: 572121017

5 ☐ Limited Calibration5 ☐ Out of Tolerance >5x

Cal Work Inst ID: 3053G

6 ☐ Functional Check6 ☐ Out of Tolerance-Undetermined

Outside Vendor:

7 ☐ Performance Check7 ☐ Inoperative8 ☐ Modify8 ☐ Damaged9 ☐ Repair-needs Charge Level9 ☐ Not Used10 ☐ Other10 ☐ Not Determined11 ☐ Excessed

Calibrated By: Shane Warner

S#: 43896 Phone: 526-2761

12 ☐ Extension**CALIBRATION STANDARDS USED**

700835							

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED  
VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

**OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION**

NOMINAL (STD)	UNITS	AS FOUND (UUT)	MFG. ACCURACY

**COMMENTS**

**INL  
CALIBRATION INPUT DATA**

11/29/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 716309      Mfr: FOWLER      Model: 6 INCH      Noun Name: DIGITAL CALIPER      Serial #: S225

Calibration Date: 1/24/2000 2:01:26 PM

**ACTION CODE**

**AS FOUND**

Next Cal Due Date: 1/24/2001

Charge Level: 1

Repair/Adj/etc C.L.: 0

Material Amount: 0

Charge Number: 572152030

Cal Work Inst ID: 3053G

Outside Vendor:

1 ☐ Acceptance Test

2 ☐ Special Test

3 ☒ Calibration to MFG Specs

4 ☐ Clean

5 ☐ Limited Calibration

6 ☐ Functional Check

7 ☐ Performance Check

8 ☐ Modify

9 ☐ Repair-needs Charge Level

10 ☐ Other

1 ☒ In Tolerance

2 ☐ Out of Tolerance >1x <2x

3 ☐ Out of Tolerance >2x <3x

4 ☐ Out of Tolerance >3x <5x

5 ☐ Out of Tolerance >5x

6 ☐ Out of Tolerance-Undetermined

7 ☐ Inoperative

8 ☐ Damaged

9 ☐ Not Used

10 ☐ Not Determined

11 ☐ Excessed

12 ☐ Extension

Calibrated By: Bob Randall

S#: 34986 Phone: 526-2761

**CALIBRATION STANDARDS USED**

704670	703081	702056	707769				

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-5.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

NOMINAL (STD)	OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION UNITS	AS FOUND (UUT)	MFG. ACCURACY

**COMMENTS**

**INL  
CALIBRATION INPUT DATA**

11/29/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 716309      Mfr: FOWLER      Model: 6 INCH      Noun Name: DIGITAL CALIPER      Serial #: S225

Calibration Date: 11/21/2000 10:51:53 AM

**ACTION CODE****AS FOUND**

Next Cal Due Date: 11/21/2001

Charge Level: 1

Repair/Adj/etc C.L.: 0

Material Amount: 0

Charge Number: 572152020

Cal Work Inst ID: 3053G

Outside Vendor:

1 ☐ Acceptance Test2 ☐ Special Test3 ☒ Calibration to MFG Specs4 ☐ Clean5 ☐ Limited Calibration6 ☐ Functional Check7 ☐ Performance Check8 ☐ Modify9 ☐ Repair-needs Charge Level10 ☐ Other1 ☒ In Tolerance2 ☐ Out of Tolerance >1x <2x3 ☐ Out of Tolerance >2x <3x4 ☐ Out of Tolerance >3x <5x5 ☐ Out of Tolerance >5x6 ☐ Out of Tolerance-Undetermined7 ☐ Inoperative8 ☐ Damaged9 ☐ Not Used10 ☐ Not Determined11 ☐ Excessed12 ☐ Extension

Calibrated By: Bob Randall

S#: 34986 Phone: 526-2761

**CALIBRATION STANDARDS USED**

700835	702056	707769					

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)    20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)    23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)    20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)    23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)    20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:    23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

**OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION**

NOMINAL (STD)

UNITS

AS FOUND (UUT)

MFG. ACCURACY


**COMMENTS**



**INL  
CALIBRATION INPUT DATA**

11/29/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 716309      Mfr: FOWLER      Model: 6 INCH      Noun Name: DIGITAL CALIPER      Serial #: S225

Calibration Date: 8/20/2001 11:11:02 AM

**ACTION CODE****AS FOUND**

Next Cal Due Date: 5/20/2002

Charge Level: 1

Repair/Adj/etc C.L.: 0

Material Amount: 0

Charge Number: 572152010

Cal Work Inst ID: 3053H

Outside Vendor:

1 ☐ Acceptance Test2 ☐ Special Test3 ☒ Calibration to MFG Specs4 ☐ Clean5 ☐ Limited Calibration6 ☐ Functional Check7 ☐ Performance Check8 ☐ Modify9 ☐ Repair-needs Charge Level10 ☐ Other1 ☒ In Tolerance2 ☐ Out of Tolerance >1x <2x3 ☐ Out of Tolerance >2x <3x4 ☐ Out of Tolerance >3x <5x5 ☐ Out of Tolerance >5x6 ☐ Out of Tolerance-Undetermined7 ☐ Inoperative8 ☐ Damaged9 ☐ Not Used10 ☐ Not Determined11 ☐ Excessed12 ☐ Extension

Calibrated By: Terry Wilde

S#: 57438 Phone: 526-2761

**CALIBRATION STANDARDS USED**

718307	703081	707769					

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

NOMINAL (STD)	OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION UNITS	AS FOUND (UUT)	MFG. ACCURACY

**COMMENTS**

**INL**  
**CALIBRATION INPUT DATA**

11/29/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 716309      Mfr: FOWLER      Model: 6 INCH      Noun Name: DIGITAL CALIPER      Serial #: S225

Calibration Date: 6/4/2003 4:24:36 PM

**ACTION CODE****AS FOUND**

Next Cal Due Date: 3/4/2004

1 ☐ Acceptance Test1 ☒ In Tolerance

Charge Level: 2

2 ☐ Special Test2 ☐ Out of Tolerance >1x <2x

Repair/Adj/etc C.L: 0

3 ☒ Calibration to MFG Specs3 ☐ Out of Tolerance >2x <3x

Material Amount: 0

4 ☐ Clean4 ☐ Out of Tolerance >3x <5x

Charge Number: 100348027

5 ☐ Limited Calibration5 ☐ Out of Tolerance >5x

Cal Work Inst ID: 3053J

6 ☐ Functional Check6 ☐ Out of Tolerance-Undetermined

Outside Vendor:

7 ☐ Performance Check7 ☐ Inoperative8 ☐ Modify8 ☐ Damaged9 ☐ Repair-needs Charge Level9 ☐ Not Used10 ☐ Other10 ☐ Not Determined11 ☐ Excessed

Calibrated By: Terry Wilde

S#: 57438 Phone: 526-2761

12 ☐ Extension**CALIBRATION STANDARDS USED**

718307	703081	702056	707769				

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-5.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

**OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION**

NOMINAL (STD)	UNITS	AS FOUND (UUT)	MFG. ACCURACY

**COMMENTS**

**INL  
CALIBRATION INPUT DATA**

11/29/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 716309      Mfr: FOWLER      Model: 6 INCH      Noun Name: DIGITAL CALIPER      Serial #: S225

Calibration Date: 3/9/2004 8:46:37 AM

**ACTION CODE**

**AS FOUND**

Next Cal Due Date: 12/9/2004

Charge Level: 2

Repair/Adj/etc C.L: 0

Material Amount: 3

Charge Number: 530130226

Cal Work Inst ID: 3053J

Outside Vendor:

1 ☐ Acceptance Test

2 ☐ Special Test

3 ☒ Calibration to MFG Specs

4 ☒ Clean

5 ☐ Limited Calibration

6 ☐ Functional Check

7 ☐ Performance Check

8 ☐ Modify

9 ☐ Repair-needs Charge Level

10 ☐ Other

1 ☒ In Tolerance

2 ☐ Out of Tolerance >1x <2x

3 ☐ Out of Tolerance >2x <3x

4 ☐ Out of Tolerance >3x <5x

5 ☐ Out of Tolerance >5x

6 ☐ Out of Tolerance-Undetermined

7 ☐ Inoperative

8 ☐ Damaged

9 ☐ Not Used

10 ☐ Not Determined

11 ☐ Excessed

12 ☐ Extension

Calibrated By: Terry Wilde

S#: 57438 Phone: 526-2761

**CALIBRATION STANDARDS USED**

718307	702056	703081	707769				

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-5.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

NOMINAL (STD)	OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION UNITS	AS FOUND (UUT)	MFG. ACCURACY

**COMMENTS**

REPLACED 2 BATTERIES AT A COST OF \$3.00

# INL CALIBRATION INPUT DATA

11/29/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 716309	Mfr: FOWLER	Model: 6 INCH	Noun Name: DIGITAL CALIPER	Serial #: S225
Calibration Date: 2/8/2005	<b>ACTION CODE</b>		<b>AS FOUND</b>	
Next Cal Due Date: 2/8/2006	1 <input type="checkbox"/>	Acceptance Test	1 <input checked="" type="radio"/>	In Tolerance
Charge Level: 2	2 <input type="checkbox"/>	Special Test	2 <input type="radio"/>	Out of Tolerance >1x <2x
Repair/Adj/etc C.L: 0	3 <input checked="" type="checkbox"/>	Calibration to MFG Specs	3 <input type="radio"/>	Out of Tolerance >2x <3x
Material Amount: 0	4 <input type="checkbox"/>	Clean	4 <input type="radio"/>	Out of Tolerance >3x <5x
Charge Number: 100853GSA	5 <input type="checkbox"/>	Limited Calibration	5 <input type="radio"/>	Out of Tolerance >5x
Cal Work Inst ID: 3053L	6 <input type="checkbox"/>	Functional Check	6 <input type="radio"/>	Out of Tolerance-Undetermined
Outside Vendor:	7 <input type="checkbox"/>	Performance Check	7 <input type="checkbox"/>	Inoperative
	8 <input type="checkbox"/>	Modify	8 <input type="checkbox"/>	Damaged
	9 <input type="checkbox"/>	Repair-needs Charge Level	9 <input type="checkbox"/>	Not Used
	10 <input type="checkbox"/>	Other	10 <input type="checkbox"/>	Not Determined
			11 <input type="checkbox"/>	Excessed
Calibrated By: Terry Wilde	S#: 57438	Phone: 526-2761	12 <input type="checkbox"/>	Extension

## CALIBRATION STANDARDS USED

718307	702056	703081	707769				

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

## LABORATORY TEMPERATURE AND HUMIDITY

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

NOMINAL (STD)	OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION UNITS	AS FOUND (UUT)	MFG. ACCURACY

## COMMENTS

**INL  
CALIBRATION INPUT DATA**

11/29/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 716309	Mfr: FOWLER	Model: 6 INCH	Noun Name: DIGITAL CALIPER	Serial #: S225
-------------------	-------------	---------------	----------------------------	----------------

Calibration Date: 2/2/2006

**ACTION CODE****AS FOUND**

Next Cal Due Date: 2/2/2007

Charge Level: 2

Repair/Adj/etc C.L: 0

Material Amount: 0

Charge Number: 100853GSA

Cal Work Inst ID: 3053L

Outside Vendor:

- |    |                                     |                           |
|----|-------------------------------------|---------------------------|
| 1  | <input type="checkbox"/>            | Acceptance Test           |
| 2  | <input type="checkbox"/>            | Special Test              |
| 3  | <input checked="" type="checkbox"/> | Calibration to MFG Specs  |
| 4  | <input type="checkbox"/>            | Clean                     |
| 5  | <input type="checkbox"/>            | Limited Calibration       |
| 6  | <input type="checkbox"/>            | Functional Check          |
| 7  | <input type="checkbox"/>            | Performance Check         |
| 8  | <input type="checkbox"/>            | Modify                    |
| 9  | <input type="checkbox"/>            | Repair-needs Charge Level |
| 10 | <input type="checkbox"/>            | Other                     |

- |    |                                  |                               |
|----|----------------------------------|-------------------------------|
| 1  | <input checked="" type="radio"/> | In Tolerance                  |
| 2  | <input type="radio"/>            | Out of Tolerance >1x <2x      |
| 3  | <input type="radio"/>            | Out of Tolerance >2x <3x      |
| 4  | <input type="radio"/>            | Out of Tolerance >3x <5x      |
| 5  | <input type="radio"/>            | Out of Tolerance >5x          |
| 6  | <input type="radio"/>            | Out of Tolerance-Undetermined |
| 7  | <input type="checkbox"/>         | Inoperative                   |
| 8  | <input type="checkbox"/>         | Damaged                       |
| 9  | <input type="checkbox"/>         | Not Used                      |
| 10 | <input type="checkbox"/>         | Not Determined                |
| 11 | <input type="checkbox"/>         | Excessed                      |
| 12 | <input type="checkbox"/>         | Extension                     |

Calibrated By: Terry Wilde S#: 57438 Phone: 526-2761

**CALIBRATION STANDARDS USED**

718307	702056	703081	707769				

**STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES**

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-5.0 ° C (20-50% RH)

**Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.**

NOMINAL (STD)	OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION UNITS	AS FOUND (UUT)	MFG. ACCURACY

**COMMENTS**

**INL**  
**CALIBRATION INPUT DATA**

1/31/2008

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 716309      Mfr: FOWLER      Model: 6 INCH      Noun Name: DIGITAL CALIPER      Serial #: S225  
 Next Cal Due Date: 1/24/2009      **ACTION CODE**      **AS FOUND**

Calibration Date: 1/24/2008	1	<input type="checkbox"/>	Acceptance Test	1	<input checked="" type="radio"/>	In Tolerance
Charge Level: 2	2	<input type="checkbox"/>	Special Test	2	<input type="radio"/>	Out of Tolerance >1x <2x
Repair/Adj/etc C.L.: 0	3	<input checked="" type="checkbox"/>	Calibration to MFG Specs	3	<input type="radio"/>	Out of Tolerance >2x <3x
Material Amount: 0	4	<input type="checkbox"/>	Clean	4	<input type="radio"/>	Out of Tolerance >3x <5x
Charge Number: 100853GSA	5	<input type="checkbox"/>	Limited Calibration	5	<input type="radio"/>	Out of Tolerance >5x
Cal Work Inst ID: 30530	6	<input type="checkbox"/>	Functional Check	6	<input type="radio"/>	Out of Tolerance-Undetermined
Outside Vendor:	7	<input type="checkbox"/>	Performance Check	7	<input type="checkbox"/>	Inoperative
	8	<input type="checkbox"/>	Modify	8	<input type="checkbox"/>	Damaged
	9	<input type="checkbox"/>	Repair-needs Charge Level	9	<input type="checkbox"/>	Not Used
	10	<input type="checkbox"/>	Other	10	<input type="checkbox"/>	Not Determined
				11	<input type="checkbox"/>	Excessed
Calibrated By: Terry Wilde	S#: 57438	Phone: 526-2761		12	<input type="checkbox"/>	Extension

**CALIBRATION STANDARDS USED**

718307	702056	703081	707769				

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

NOMINAL (STD)	OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION UNITS	AS FOUND (UUT)	MFG. ACCURACY

COMMENTS

**INL  
CALIBRATION INPUT DATA**

11/29/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 716309      Mfr: FOWLER      Model: 6 INCH      Noun Name: DIGITAL CALIPER      Serial #: S225

Calibration Date: 1/31/2007

**ACTION CODE**

**AS FOUND**

Next Cal Due Date: 1/31/2008	1	<input type="checkbox"/>	Acceptance Test	1	<input checked="" type="radio"/>	In Tolerance
Charge Level: 2	2	<input type="checkbox"/>	Special Test	2	<input type="radio"/>	Out of Tolerance >1x <2x
Repair/Adj/etc C.L: 1	3	<input checked="" type="checkbox"/>	Calibration to MFG Specs	3	<input type="radio"/>	Out of Tolerance >2x <3x
Material Amount: 1.5	4	<input type="checkbox"/>	Clean	4	<input type="radio"/>	Out of Tolerance >3x <5x
Charge Number: 100853GSA	5	<input type="checkbox"/>	Limited Calibration	5	<input type="radio"/>	Out of Tolerance >5x
Cal Work Inst ID: 3053L	6	<input type="checkbox"/>	Functional Check	6	<input type="radio"/>	Out of Tolerance-Undetermined
Outside Vendor:	7	<input type="checkbox"/>	Performance Check	7	<input type="checkbox"/>	Inoperative
	8	<input type="checkbox"/>	Modify	8	<input type="checkbox"/>	Damaged
	9	<input checked="" type="checkbox"/>	Repair-needs Charge Level	9	<input type="checkbox"/>	Not Used
	10	<input type="checkbox"/>	Other	10	<input type="checkbox"/>	Not Determined
				11	<input type="checkbox"/>	Excessed
				12	<input type="checkbox"/>	Extension

Calibrated By: Terry Wilde      S#: 57438      Phone: 526-2761

**CALIBRATION STANDARDS USED**

718307	702056	703081	707769				

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-5.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

NOMINAL (STD)	OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION UNITS	AS FOUND (UUT)	MFG. ACCURACY

**COMMENTS**

REPAIR CHARGE LEVEL OF 1 WAS FOR REPLACING BATTERY AT A COST OF \$1.50

**INL  
CALIBRATION INPUT DATA**

11/30/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 721714      Mfr: STARRETT      Model: 6 INCH      Noun Name: DIGITAL CALIPER      Serial #:

03290036      Calibration Date: 12/15/2003 3:57:03 PM

Next Cal Due Date: 9/15/2004

Charge Level: 2

Repair/Adj/etc C.L: 0

Material Amount: 0

Charge Number: 100664GSB

Cal Work Inst ID: 3053J

Outside Vendor:

**ACTION CODE**

**AS FOUND**

- |    |                                     |                           |    |                                  |                               |
|----|-------------------------------------|---------------------------|----|----------------------------------|-------------------------------|
| 1  | <input type="checkbox"/>            | Acceptance Test           | 1  | <input checked="" type="radio"/> | In Tolerance                  |
| 2  | <input type="checkbox"/>            | Special Test              | 2  | <input type="radio"/>            | Out of Tolerance >1x <2x      |
| 3  | <input checked="" type="checkbox"/> | Calibration to MFG Specs  | 3  | <input type="radio"/>            | Out of Tolerance >2x <3x      |
| 4  | <input type="checkbox"/>            | Clean                     | 4  | <input type="radio"/>            | Out of Tolerance >3x <5x      |
| 5  | <input type="checkbox"/>            | Limited Calibration       | 5  | <input type="radio"/>            | Out of Tolerance >5x          |
| 6  | <input type="checkbox"/>            | Functional Check          | 6  | <input type="radio"/>            | Out of Tolerance-Undetermined |
| 7  | <input type="checkbox"/>            | Performance Check         | 7  | <input type="checkbox"/>         | Inoperative                   |
| 8  | <input type="checkbox"/>            | Modify                    | 8  | <input type="checkbox"/>         | Damaged                       |
| 9  | <input type="checkbox"/>            | Repair-needs Charge Level | 9  | <input type="checkbox"/>         | Not Used                      |
| 10 | <input type="checkbox"/>            | Other                     | 10 | <input type="checkbox"/>         | Not Determined                |
|    |                                     |                           | 11 | <input type="checkbox"/>         | Exceeded                      |
|    |                                     |                           | 12 | <input type="checkbox"/>         | Extension                     |

Calibrated By: Terry Wilde

S#: 57438 Phone: 526-2761

**CALIBRATION STANDARDS USED**

718307	707769	703081	702056				

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-5.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

**OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION**

NOMINAL (STD)	UNITS	AS FOUND (UUT)	MFG. ACCURACY

**COMMENTS**

INITIAL CALIBRATION  
TEST REFERRAL QA# 104353



**INL**  
**CALIBRATION INPUT DATA**

11/30/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 721714    Mfr: STARRETT    Model: 6 INCH    Noun Name: DIGITAL CALIPER    Serial #:

Calibration Date: 11/2/2004

Next Cal Due Date: 11/2/2005

Charge Level: 2

Repair/Adj/etc C.L: 0

Material Amount: 0

Charge Number: 100664GSB

Cal Work Inst ID: 3053K

Outside Vendor:

**ACTION CODE****AS FOUND**

- |    |                                     |                           |    |                                  |                               |
|----|-------------------------------------|---------------------------|----|----------------------------------|-------------------------------|
| 1  | <input type="checkbox"/>            | Acceptance Test           | 1  | <input checked="" type="radio"/> | In Tolerance                  |
| 2  | <input type="checkbox"/>            | Special Test              | 2  | <input type="radio"/>            | Out of Tolerance >1x <2x      |
| 3  | <input checked="" type="checkbox"/> | Calibration to MFG Specs  | 3  | <input type="radio"/>            | Out of Tolerance >2x <3x      |
| 4  | <input type="checkbox"/>            | Clean                     | 4  | <input type="radio"/>            | Out of Tolerance >3x <5x      |
| 5  | <input type="checkbox"/>            | Limited Calibration       | 5  | <input type="radio"/>            | Out of Tolerance >5x          |
| 6  | <input type="checkbox"/>            | Functional Check          | 6  | <input type="radio"/>            | Out of Tolerance-Undetermined |
| 7  | <input type="checkbox"/>            | Performance Check         | 7  | <input type="checkbox"/>         | Inoperative                   |
| 8  | <input type="checkbox"/>            | Modify                    | 8  | <input type="checkbox"/>         | Damaged                       |
| 9  | <input type="checkbox"/>            | Repair-needs Charge Level | 9  | <input type="checkbox"/>         | Not Used                      |
| 10 | <input type="checkbox"/>            | Other                     | 10 | <input type="checkbox"/>         | Not Determined                |
|    |                                     |                           | 11 | <input type="checkbox"/>         | Excessed                      |
|    |                                     |                           | 12 | <input type="checkbox"/>         | Extension                     |

Calibrated By: Terry Wilde    S#: 57438    Phone: 526-2761

**CALIBRATION STANDARDS USED**

718307	702056	703081	707769				

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED  
VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
PhysDim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

**OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION**

NOMINAL (STD)	UNITS	AS FOUND (UUT)	MFG. ACCURACY

**COMMENTS**

**INL  
CALIBRATION INPUT DATA**

11/30/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 721714      Mfr: STARRETT      Model: 6 INCH      Noun Name: DIGITAL CALIPER      Serial #:

03290036      Calibration Date: 10/10/2005

Next Cal Due Date: 10/10/2006

Charge Level: 2

Repair/Adj/etc C.L: 0

Material Amount: 0

Charge Number: 100583GSA

Cal Work Inst ID: 3053L

Outside Vendor:

**ACTION CODE**

- |    |                                     |                           |
|----|-------------------------------------|---------------------------|
| 1  | <input type="checkbox"/>            | Acceptance Test           |
| 2  | <input type="checkbox"/>            | Special Test              |
| 3  | <input checked="" type="checkbox"/> | Calibration to MFG Specs  |
| 4  | <input type="checkbox"/>            | Clean                     |
| 5  | <input type="checkbox"/>            | Limited Calibration       |
| 6  | <input type="checkbox"/>            | Functional Check          |
| 7  | <input type="checkbox"/>            | Performance Check         |
| 8  | <input type="checkbox"/>            | Modify                    |
| 9  | <input type="checkbox"/>            | Repair-needs Charge Level |
| 10 | <input type="checkbox"/>            | Other                     |

**AS FOUND**

- |    |                                  |                               |
|----|----------------------------------|-------------------------------|
| 1  | <input checked="" type="radio"/> | In Tolerance                  |
| 2  | <input type="radio"/>            | Out of Tolerance >1x <2x      |
| 3  | <input type="radio"/>            | Out of Tolerance >2x <3x      |
| 4  | <input type="radio"/>            | Out of Tolerance >3x <5x      |
| 5  | <input type="radio"/>            | Out of Tolerance >5x          |
| 6  | <input type="radio"/>            | Out of Tolerance-Undetermined |
| 7  | <input type="checkbox"/>         | Inoperative                   |
| 8  | <input type="checkbox"/>         | Damaged                       |
| 9  | <input type="checkbox"/>         | Not Used                      |
| 10 | <input type="checkbox"/>         | Not Determined                |
| 11 | <input type="checkbox"/>         | Excessed                      |
| 12 | <input type="checkbox"/>         | Extension                     |

Calibrated By: Terry Wilde      S#: 57438      Phone: 526-2761

**CALIBRATION STANDARDS USED**

718307	702056	703081	707769				

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

**OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION**

NOMINAL (STD)	UNITS	AS FOUND (UUT)	MFG. ACCURACY

**COMMENTS**

**INL  
CALIBRATION INPUT DATA**

11/30/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 721714      Mfr: STARRETT      Model: 6 INCH      Noun Name: DIGITAL CALIPER      Serial #:

03290036

Calibration Date: 9/26/2006

**ACTION CODE**

**AS FOUND**

Next Cal Due Date: 9/26/2007

Charge Level: 2

Repair/Adj/etc C.L.: 0

Material Amount: 0

Charge Number: 100853GSA

Cal Work Inst ID: 3053L

Outside Vendor:

1 ☐

Acceptance Test

2 ☐

Special Test

3 ☒

Calibration to MFG Specs

4 ☐

Clean

5 ☐

Limited Calibration

6 ☐

Functional Check

7 ☐

Performance Check

8 ☐

Modify

9 ☐

Repair-needs Charge Level

10 ☐

Other

1 ☒ In Tolerance

2 ☐ Out of Tolerance >1x <2x

3 ☐ Out of Tolerance >2x <3x

4 ☐ Out of Tolerance >3x <5x

5 ☐ Out of Tolerance >5x

6 ☐ Out of Tolerance-Undetermined

7 ☐ Inoperative

8 ☐ Damaged

9 ☐ Not Used

10 ☐ Not Determined

11 ☐ Excessed

12 ☐ Extension

Calibrated By: Terry Wilde

S#: 57438 Phone: 526-2761

**CALIBRATION STANDARDS USED**

718307	702056	703081	707769				

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

**OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION**

NOMINAL (STD)	UNITS	AS FOUND (UUT)	MFG. ACCURACY

**COMMENTS**

**INL**  
**CALIBRATION INPUT DATA**

11/30/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 721714      Mfr: STARRETT      Model: 6 INCH      Noun Name: DIGITAL CALIPER      Serial #:

Calibration Date: 10/11/2007

Next Cal Due Date: 10/11/2008

Charge Level: 2

Repair/Adj/etc C.L.: 0

Material Amount: 0

Charge Number: 100853GSA

Cal Work Inst ID: 3053N

Outside Vendor:

**ACTION CODE****AS FOUND**

- |    |                                     |                           |    |                                  |                               |
|----|-------------------------------------|---------------------------|----|----------------------------------|-------------------------------|
| 1  | <input type="checkbox"/>            | Acceptance Test           | 1  | <input checked="" type="radio"/> | In Tolerance                  |
| 2  | <input type="checkbox"/>            | Special Test              | 2  | <input type="radio"/>            | Out of Tolerance >1x <2x      |
| 3  | <input checked="" type="checkbox"/> | Calibration to MFG Specs  | 3  | <input type="radio"/>            | Out of Tolerance >2x <3x      |
| 4  | <input type="checkbox"/>            | Clean                     | 4  | <input type="radio"/>            | Out of Tolerance >3x <5x      |
| 5  | <input type="checkbox"/>            | Limited Calibration       | 5  | <input type="radio"/>            | Out of Tolerance >5x          |
| 6  | <input type="checkbox"/>            | Functional Check          | 6  | <input type="radio"/>            | Out of Tolerance-Undetermined |
| 7  | <input type="checkbox"/>            | Performance Check         | 7  | <input type="checkbox"/>         | Inoperative                   |
| 8  | <input type="checkbox"/>            | Modify                    | 8  | <input type="checkbox"/>         | Damaged                       |
| 9  | <input type="checkbox"/>            | Repair-needs Charge Level | 9  | <input type="checkbox"/>         | Not Used                      |
| 10 | <input type="checkbox"/>            | Other                     | 10 | <input type="checkbox"/>         | Not Determined                |
|    |                                     |                           | 11 | <input type="checkbox"/>         | Excessed                      |
|    |                                     |                           | 12 | <input type="checkbox"/>         | Extension                     |

Calibrated By: Terry Wilde      S#: 57438      Phone: 526-2761

**CALIBRATION STANDARDS USED**

718307	705471	703081	707769				

**STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES**

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-5.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

**OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION**

NOMINAL (STD)	UNITS	AS FOUND (UUT)	MFG. ACCURACY
---------------	-------	----------------	---------------


**COMMENTS**

**INL**  
**CALIBRATION INPUT DATA**

12/4/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 721715      Mfr: STARRETT      Model: 6 INCH      Noun Name: DIGITAL CALIPER      Serial #:  
03290062

Calibration Date: 12/15/2003 3:58:14 PM

**ACTION CODE****AS FOUND**

Next Cal Due Date: 9/15/2004

1 ☐ Acceptance Test1 ☒ In Tolerance

Charge Level: 2

2 ☐ Special Test2 ☐ Out of Tolerance >1x <2x

Repair/Adj/etc C.L: 0

3 ☒ Calibration to MFG Specs3 ☐ Out of Tolerance >2x <3x

Material Amount: 0

4 ☐ Clean4 ☐ Out of Tolerance >3x <5x

Charge Number: 100664GSB

5 ☐ Limited Calibration5 ☐ Out of Tolerance >5x

Cal Work Inst ID: 3053J

6 ☐ Functional Check6 ☐ Out of Tolerance-Undetermined

Outside Vendor:

7 ☐ Performance Check7 ☐ Inoperative8 ☐ Modify8 ☐ Damaged9 ☐ Repair-needs Charge Level9 ☐ Not Used10 ☐ Other10 ☐ Not Determined11 ☐ Excessed

Calibrated By: Terry Wilde

S#: 57438 Phone: 526-2761

12 ☐ Extension**CALIBRATION STANDARDS USED**

718307	702056	703081	707769				

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED  
VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

**OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION****NOMINAL (STD)****UNITS****AS FOUND (UUT)****MFG. ACCURACY**


**COMMENTS**

INITIAL CALIBRATION  
TEST REFERRAL QA# 104353

**INL  
CALIBRATION INPUT DATA**

11/30/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 721715      Mfr: STARRETT      Model: 6 INCH      Noun Name: DIGITAL CALIPER      Serial #:  
03290062

Calibration Date: 10/12/2004

**ACTION CODE**

**AS FOUND**

Next Cal Due Date: 10/12/2005	1	<input type="checkbox"/>	Acceptance Test	1	<input checked="" type="radio"/>	In Tolerance
Charge Level: 2	2	<input type="checkbox"/>	Special Test	2	<input type="radio"/>	Out of Tolerance >1x <2x
Repair/Adj/etc C.L: 0	3	<input checked="" type="checkbox"/>	Calibration to MFG Specs	3	<input type="radio"/>	Out of Tolerance >2x <3x
Material Amount: 0	4	<input type="checkbox"/>	Clean	4	<input type="radio"/>	Out of Tolerance >3x <5x
Charge Number: 100664GSA	5	<input type="checkbox"/>	Limited Calibration	5	<input type="radio"/>	Out of Tolerance >5x
Cal Work Inst ID: 3053K	6	<input type="checkbox"/>	Functional Check	6	<input type="radio"/>	Out of Tolerance-Undetermined
Outside Vendor:	7	<input type="checkbox"/>	Performance Check	7	<input type="checkbox"/>	Inoperative
	8	<input type="checkbox"/>	Modify	8	<input type="checkbox"/>	Damaged
	9	<input type="checkbox"/>	Repair-needs Charge Level	9	<input type="checkbox"/>	Not Used
	10	<input type="checkbox"/>	Other	10	<input type="checkbox"/>	Not Determined
				11	<input type="checkbox"/>	Excessed
Calibrated By: Terry Wilde				12	<input type="checkbox"/>	Extension
S#: 57438			Phone: 526-2761			

**CALIBRATION STANDARDS USED**

718307	702056	703081	707769				

**STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES**

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

**OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION**

NOMINAL (STD)	UNITS	AS FOUND (UUT)	MFG. ACCURACY

**COMMENTS**

**INL  
CALIBRATION INPUT DATA**

11/30/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 721715      Mfr: STARRETT      Model: 6 INCH      Noun Name: DIGITAL CALIPER      Serial #:  
03290062

Calibration Date: 10/10/2005	ACTION CODE	AS FOUND
Next Cal Due Date: 10/10/2006	1 <input type="checkbox"/> Acceptance Test	1 <input checked="" type="radio"/> In Tolerance
Charge Level: 2	2 <input type="checkbox"/> Special Test	2 <input type="radio"/> Out of Tolerance >1x <2x
Repair/Adj/etc C.L.: 0	3 <input checked="" type="checkbox"/> Calibration to MFG Specs	3 <input type="radio"/> Out of Tolerance >2x <3x
Material Amount: 0	4 <input type="checkbox"/> Clean	4 <input type="radio"/> Out of Tolerance >3x <5x
Charge Number: 100583GSA	5 <input type="checkbox"/> Limited Calibration	5 <input type="radio"/> Out of Tolerance >5x
Cal Work Inst ID: 3053L	6 <input type="checkbox"/> Functional Check	6 <input type="radio"/> Out of Tolerance-Undetermined
Outside Vendor:	7 <input type="checkbox"/> Performance Check	7 <input type="checkbox"/> Inoperative
	8 <input type="checkbox"/> Modify	8 <input type="checkbox"/> Damaged
	9 <input type="checkbox"/> Repair-needs Charge Level	9 <input type="checkbox"/> Not Used
	10 <input type="checkbox"/> Other	10 <input type="checkbox"/> Not Determined
		11 <input type="checkbox"/> Excessed
Calibrated By: Terry Wilde	S#: 57438 Phone: 526-2761	12 <input type="checkbox"/> Extension

**CALIBRATION STANDARDS USED**

718307	703081	702056	707769				

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

NOMINAL (STD)	OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION UNITS	AS FOUND (UUT)	MFG. ACCURACY

COMMENTS

**INL  
CALIBRATION INPUT DATA**

11/30/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 721715      Mfr: STARRETT      Model: 6 INCH      Noun Name: DIGITAL CALIPER      Serial #:

Calibration Date: 9/26/2006

**ACTION CODE****AS FOUND**

Next Cal Due Date: 9/26/2007

Charge Level: 2

Repair/Adj/etc C.L.: 0

Material Amount: 0

Charge Number: 100853GSA

Cal Work Inst ID: 3053L

Outside Vendor:

1 ☐

Acceptance Test

2 ☐

Special Test

3 ☒

Calibration to MFG Specs

4 ☐

Clean

5 ☐

Limited Calibration

6 ☐

Functional Check

7 ☐

Performance Check

8 ☐

Modify

9 ☐

Repair-needs Charge Level

10 ☐

Other

1 ☒ In Tolerance2 ☐ Out of Tolerance >1x <2x3 ☐ Out of Tolerance >2x <3x4 ☐ Out of Tolerance >3x <5x5 ☐ Out of Tolerance >5x6 ☐ Out of Tolerance-Undetermined7 ☐ Inoperative8 ☐ Damaged9 ☐ Not Used10 ☐ Not Determined11 ☐ Excessed12 ☐ Extension

Calibrated By: Terry Wilde      S#: 57438      Phone: 526-2761

**CALIBRATION STANDARDS USED**

718307	702056	703081	707769				

**STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES**

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

**OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION****NOMINAL (STD)****UNITS****AS FOUND (UUT)****MFG. ACCURACY**


**COMMENTS**



**INL**  
**CALIBRATION INPUT DATA**

11/30/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 721715      Mfr: STARRETT      Model: 6 INCH      Noun Name: DIGITAL CALIPER      Serial #:

03290062      Calibration Date: 10/11/2007

Next Cal Due Date: 10/11/2008

Charge Level: 2

Repair/Adj/etc C.L: 0

Material Amount: 0

Charge Number: 100853GSA

Cal Work Inst ID: 3053N

Outside Vendor:

**ACTION CODE**

- |    |                                     |                           |
|----|-------------------------------------|---------------------------|
| 1  | <input type="checkbox"/>            | Acceptance Test           |
| 2  | <input type="checkbox"/>            | Special Test              |
| 3  | <input checked="" type="checkbox"/> | Calibration to MFG Specs  |
| 4  | <input type="checkbox"/>            | Clean                     |
| 5  | <input type="checkbox"/>            | Limited Calibration       |
| 6  | <input type="checkbox"/>            | Functional Check          |
| 7  | <input type="checkbox"/>            | Performance Check         |
| 8  | <input type="checkbox"/>            | Modify                    |
| 9  | <input type="checkbox"/>            | Repair-needs Charge Level |
| 10 | <input type="checkbox"/>            | Other                     |

**AS FOUND**

- |    |                                  |                               |
|----|----------------------------------|-------------------------------|
| 1  | <input checked="" type="radio"/> | In Tolerance                  |
| 2  | <input type="radio"/>            | Out of Tolerance >1x <2x      |
| 3  | <input type="radio"/>            | Out of Tolerance >2x <3x      |
| 4  | <input type="radio"/>            | Out of Tolerance >3x <5x      |
| 5  | <input type="radio"/>            | Out of Tolerance >5x          |
| 6  | <input type="radio"/>            | Out of Tolerance-Undetermined |
| 7  | <input type="checkbox"/>         | Inoperative                   |
| 8  | <input type="checkbox"/>         | Damaged                       |
| 9  | <input type="checkbox"/>         | Not Used                      |
| 10 | <input type="checkbox"/>         | Not Determined                |
| 11 | <input type="checkbox"/>         | Excessed                      |
| 12 | <input type="checkbox"/>         | Extension                     |

Calibrated By: Terry Wilde      S#: 57438      Phone: 526-2761

**CALIBRATION STANDARDS USED**

718307	705471	703081	707769				

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

**OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION**

NOMINAL (STD)	UNITS	AS FOUND (UUT)	MFG. ACCURACY

**COMMENTS**

# **DATA ACQUISITION**

**INL  
CALIBRATION INPUT DATA**

11/30/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 722721    Mfr: IOTECH    Model: DAQ BOOK 2000A    Noun Name: 16 BIT DAC    Serial#:

238927

Calibration Date: 4/12/2005

**ACTION CODE****AS FOUND**

Next Cal Due Date: 4/12/2006

Charge Level: 8

Repair/Adj/etc C.L: 4

Material Amount: 0

Charge Number: 100853GSA

Cal Work Inst ID: 6054

Outside Vendor:

- |    |                                     |                           |    |                                  |                               |
|----|-------------------------------------|---------------------------|----|----------------------------------|-------------------------------|
| 1  | <input type="checkbox"/>            | Acceptance Test           | 1  | <input type="radio"/>            | In Tolerance                  |
| 2  | <input type="checkbox"/>            | Special Test              | 2  | <input checked="" type="radio"/> | Out of Tolerance >1x <2x      |
| 3  | <input checked="" type="checkbox"/> | Calibration to MFG Specs  | 3  | <input type="radio"/>            | Out of Tolerance >2x <3x      |
| 4  | <input type="checkbox"/>            | Clean                     | 4  | <input type="radio"/>            | Out of Tolerance >3x <5x      |
| 5  | <input type="checkbox"/>            | Limited Calibration       | 5  | <input type="radio"/>            | Out of Tolerance >5x          |
| 6  | <input type="checkbox"/>            | Functional Check          | 6  | <input type="radio"/>            | Out of Tolerance-Undetermined |
| 7  | <input type="checkbox"/>            | Performance Check         | 7  | <input type="checkbox"/>         | Inoperative                   |
| 8  | <input type="checkbox"/>            | Modify                    | 8  | <input type="checkbox"/>         | Damaged                       |
| 9  | <input type="checkbox"/>            | Repair-needs Charge Level | 9  | <input type="checkbox"/>         | Not Used                      |
| 10 | <input type="checkbox"/>            | Other                     | 10 | <input type="checkbox"/>         | Not Determined                |
|    |                                     |                           | 11 | <input type="checkbox"/>         | Excessed                      |
|    |                                     |                           | 12 | <input type="checkbox"/>         | Extension                     |

Calibrated By: Brian Berls    S#: 102182    Phone: 526-2761

**CALIBRATION STANDARDS USED**

216578	701587						

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED  
VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

**OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION**

NOMINAL (STD)	UNITS	AS FOUND (UUT)	MFG. ACCURACY
X1000	9.9 mV	9.952	+0.02875

**COMMENTS**

Initial calibration

Adjusted unit to meet manufacturers specifications and recalibrated unit.

**Notification Status : Approved**

**Area Point of Contact :** Gary D Roberts

If evaluation assistance is desired, call Quality Engineering . If further information is desired regarding the out-of-tolerance condition, call the calibrator listed below .

Calibration Organization : S&CL

Adjusted unit to meet manufacturers specifications and recalibrated unit.

---

**Evaluation:**

**Evaluator:** Tommy E Rahl

**Evaluation:** *(file attachments allowed)*

The x1000 gain of the DBK54 module attached to the DAQBook 2001A has never been used for any data acquisition tasks performed with this system.

**Selected Approver:** Dana K Morton

**NOTE:** This evaluation or a separate out -of-tolerance evaluation must be retained with organization quality assurance records . Quality requirements for this evaluation and quality record retention are based on MCP-2391 and PRD-101.

---

**Evaluation Approval:**

**Date:** 04/14/2005 **Approval:** Dana K Morton

**Approval Comments :** Agree with evaluation.

---

**Comments/Attachments**

Add any comments here and/or use the area for any attachment.

---

**Document History:**

Created by Brian L Berls on 04/12/2005 04:33:44 PM

Brian L Berls edited document 04:36 PM 04/12/2005

Notification sent: Brian L Berls on Tuesday, April 12, 2005.

Sent to user: Tommy E Rahl: 04:36 PM 04/12/2005

Tommy E Rahl edited document 04:44 PM 04/12/2005

Submitted for approval to Dana K Morton : Tommy E Rahl on Tuesday, April 12, 2005

Sent for approval to Dana K Morton: 04:44 PM 04/12/2005

Dana K Morton edited document 12:02 PM 04/14/2005

Evaluation Approval: Dana K Morton on Thursday, April 14, 2005

---

**INL  
CALIBRATION INPUT DATA**

11/30/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 722721      Mfr: IOTECH      Model: DAQ BOOK 2000A      Noun Name: 16 BIT DAC      Serial #:

Calibration Date: 4/17/2006

**ACTION CODE**

**AS FOUND**

Next Cal Due Date: 4/17/2007	1	<input type="checkbox"/>	Acceptance Test	1	<input checked="" type="radio"/>	In Tolerance
Charge Level: 8	2	<input type="checkbox"/>	Special Test	2	<input type="radio"/>	Out of Tolerance >1x <2x
Repair/Adj/etc C.L.: 0	3	<input checked="" type="checkbox"/>	Calibration to MFG Specs	3	<input type="radio"/>	Out of Tolerance >2x <3x
Material Amount: 0	4	<input type="checkbox"/>	Clean	4	<input type="radio"/>	Out of Tolerance >3x <5x
Charge Number: 100853GSA	5	<input type="checkbox"/>	Limited Calibration	5	<input type="radio"/>	Out of Tolerance >5x
Cal Work Inst ID: 6054	6	<input type="checkbox"/>	Functional Check	6	<input type="radio"/>	Out of Tolerance-Undetermined
Outside Vendor:	7	<input type="checkbox"/>	Performance Check	7	<input type="checkbox"/>	Inoperative
	8	<input type="checkbox"/>	Modify	8	<input type="checkbox"/>	Damaged
	9	<input type="checkbox"/>	Repair-needs Charge Level	9	<input type="checkbox"/>	Not Used
	10	<input type="checkbox"/>	Other	10	<input type="checkbox"/>	Not Determined
				11	<input type="checkbox"/>	Excessed
Calibrated By: Brian Berls				12	<input type="checkbox"/>	Extension

S#: 102182 Phone: 526-2761

**CALIBRATION STANDARDS USED**

701587	714891	301893					

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-5.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

**OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION**

NOMINAL (STD)	UNITS	AS FOUND (UUT)	MFG. ACCURACY

**COMMENTS**

**INL  
CALIBRATION INPUT DATA**

11/30/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 722721    Mfr: IOTECH    Model: DAQ BOOK 2000A    Noun Name: 16 BIT DAC    Serial #:

238927    Calibration Date: 4/2/2007

Next Cal Due Date: 4/2/2008

Charge Level: 8

Repair/Adj/etc C.L: 0

Material Amount: 0

Charge Number: 100853GSA

Cal Work Inst ID: 6054

Outside Vendor:

**ACTION CODE**

- |    |                                     |                           |
|----|-------------------------------------|---------------------------|
| 1  | <input type="checkbox"/>            | Acceptance Test           |
| 2  | <input type="checkbox"/>            | Special Test              |
| 3  | <input checked="" type="checkbox"/> | Calibration to MFG Specs  |
| 4  | <input type="checkbox"/>            | Clean                     |
| 5  | <input type="checkbox"/>            | Limited Calibration       |
| 6  | <input type="checkbox"/>            | Functional Check          |
| 7  | <input type="checkbox"/>            | Performance Check         |
| 8  | <input type="checkbox"/>            | Modify                    |
| 9  | <input type="checkbox"/>            | Repair-needs Charge Level |
| 10 | <input type="checkbox"/>            | Other                     |

**AS FOUND**

- |    |                                  |                               |
|----|----------------------------------|-------------------------------|
| 1  | <input checked="" type="radio"/> | In Tolerance                  |
| 2  | <input type="radio"/>            | Out of Tolerance >1x <2x      |
| 3  | <input type="radio"/>            | Out of Tolerance >2x <3x      |
| 4  | <input type="radio"/>            | Out of Tolerance >3x <5x      |
| 5  | <input type="radio"/>            | Out of Tolerance >5x          |
| 6  | <input type="radio"/>            | Out of Tolerance-Undetermined |
| 7  | <input type="checkbox"/>         | Inoperative                   |
| 8  | <input type="checkbox"/>         | Damaged                       |
| 9  | <input type="checkbox"/>         | Not Used                      |
| 10 | <input type="checkbox"/>         | Not Determined                |
| 11 | <input type="checkbox"/>         | Exceeded                      |
| 12 | <input type="checkbox"/>         | Extension                     |

Calibrated By: Donnie Lindsay    S#: 105138    Phone: 526-2761

**CALIBRATION STANDARDS USED**

301893							

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-5.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

**OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION**

NOMINAL (STD)

UNITS

AS FOUND (UUT)

MFG. ACCURACY


**COMMENTS**

S&CL overcheck required.

**INL**  
**CALIBRATION INPUT DATA**

12/11/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 722721    Mfr: IOTECH    Model: DAQ BOOK 2000A    Noun Name: 16 BIT DAC    Serial #: 238927

Calibration Date: 12/5/2007

**ACTION CODE****AS FOUND**

Next Cal Due Date: 12/5/2008

Charge Level: 8

Repair/Adj/etc C.L: 0

Material Amount: 0

Charge Number: 100853GSA

Cal Work Inst ID: 6054

Outside Vendor:

- |    |                                     |                           |    |                                  |                               |
|----|-------------------------------------|---------------------------|----|----------------------------------|-------------------------------|
| 1  | <input type="checkbox"/>            | Acceptance Test           | 1  | <input checked="" type="radio"/> | In Tolerance                  |
| 2  | <input type="checkbox"/>            | Special Test              | 2  | <input type="radio"/>            | Out of Tolerance >1x <2x      |
| 3  | <input checked="" type="checkbox"/> | Calibration to MFG Specs  | 3  | <input type="radio"/>            | Out of Tolerance >2x <3x      |
| 4  | <input type="checkbox"/>            | Clean                     | 4  | <input type="radio"/>            | Out of Tolerance >3x <5x      |
| 5  | <input type="checkbox"/>            | Limited Calibration       | 5  | <input type="radio"/>            | Out of Tolerance >5x          |
| 6  | <input type="checkbox"/>            | Functional Check          | 6  | <input type="radio"/>            | Out of Tolerance-Undetermined |
| 7  | <input type="checkbox"/>            | Performance Check         | 7  | <input type="checkbox"/>         | Inoperative                   |
| 8  | <input type="checkbox"/>            | Modify                    | 8  | <input type="checkbox"/>         | Damaged                       |
| 9  | <input type="checkbox"/>            | Repair-needs Charge Level | 9  | <input type="checkbox"/>         | Not Used                      |
| 10 | <input type="checkbox"/>            | Other                     | 10 | <input type="checkbox"/>         | Not Determined                |
|    |                                     |                           | 11 | <input type="checkbox"/>         | Excessed                      |
|    |                                     |                           | 12 | <input type="checkbox"/>         | Extension                     |

Calibrated By: Donnie Lindsay    S#: 105138    Phone: 526-2761

**CALIBRATION STANDARDS USED**

304399							

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED  
VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

**OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION**

NOMINAL (STD)	UNITS	AS FOUND (UUT)	MFG. ACCURACY

**COMMENTS**



**INL  
CALIBRATION INPUT DATA**

11/30/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 722722    Mfr: PCB    Model: 483A    Noun Name: AMPLIFIER    Serial #: 747

Calibration Date: 4/7/2005

**ACTION CODE****AS FOUND**

Next Cal Due Date: 4/7/2006	1	<input type="checkbox"/>	Acceptance Test	1	<input checked="" type="radio"/>	In Tolerance
Charge Level: 4	2	<input type="checkbox"/>	Special Test	2	<input type="radio"/>	Out of Tolerance >1x <2x
Repair/Adj/etc C.L.: 0	3	<input checked="" type="checkbox"/>	Calibration to MFG Specs	3	<input type="radio"/>	Out of Tolerance >2x <3x
Material Amount: 0	4	<input type="checkbox"/>	Clean	4	<input type="radio"/>	Out of Tolerance >3x <5x
Charge Number: 100108GSA	5	<input type="checkbox"/>	Limited Calibration	5	<input type="radio"/>	Out of Tolerance >5x
Cal Work Inst ID: 6053	6	<input type="checkbox"/>	Functional Check	6	<input type="radio"/>	Out of Tolerance-Undetermined
Outside Vendor:	7	<input type="checkbox"/>	Performance Check	7	<input type="checkbox"/>	Inoperative
	8	<input type="checkbox"/>	Modify	8	<input type="checkbox"/>	Damaged
	9	<input type="checkbox"/>	Repair-needs Charge Level	9	<input type="checkbox"/>	Not Used
	10	<input type="checkbox"/>	Other	10	<input type="checkbox"/>	Not Determined
				11	<input type="checkbox"/>	Excessed
Calibrated By: Brian Berls				12	<input type="checkbox"/>	Extension

**CALIBRATION STANDARDS USED**

301893	290716	278037	701465				

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-5.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

NOMINAL (STD)	OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION		MFG. ACCURACY
	UNITS	AS FOUND (UUT)	

**COMMENTS**

Initial calibration

**INL  
CALIBRATION INPUT DATA**

11/30/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 722722    Mfr: PCB    Model: 483A    Noun Name: AMPLIFIER    Serial #: 747

Calibration Date: 4/17/2006

**ACTION CODE**

**AS FOUND**

Next Cal Due Date: 4/17/2007	1	<input type="checkbox"/>	Acceptance Test	1	<input checked="" type="radio"/>	In Tolerance
Charge Level: 4	2	<input type="checkbox"/>	Special Test	2	<input type="radio"/>	Out of Tolerance >1x <2x
Repair/Adj/etc C.L.: 0	3	<input checked="" type="checkbox"/>	Calibration to MFG Specs	3	<input type="radio"/>	Out of Tolerance >2x <3x
Material Amount: 0	4	<input type="checkbox"/>	Clean	4	<input type="radio"/>	Out of Tolerance >3x <5x
Charge Number: 100853GSA	5	<input type="checkbox"/>	Limited Calibration	5	<input type="radio"/>	Out of Tolerance >5x
Cal Work Inst ID: 6053	6	<input type="checkbox"/>	Functional Check	6	<input type="radio"/>	Out of Tolerance-Undetermined
Outside Vendor:	7	<input type="checkbox"/>	Performance Check	7	<input type="checkbox"/>	Inoperative
	8	<input type="checkbox"/>	Modify	8	<input type="checkbox"/>	Damaged
	9	<input type="checkbox"/>	Repair-needs Charge Level	9	<input type="checkbox"/>	Not Used
	10	<input type="checkbox"/>	Other	10	<input type="checkbox"/>	Not Determined
				11	<input type="checkbox"/>	Excessed
				12	<input type="checkbox"/>	Extension

Calibrated By: Brian Berls    S#: 102182    Phone: 526-2761

**CALIBRATION STANDARDS USED**

718405	301893	290716	701464				

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

**OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION**

NOMINAL (STD)	UNITS	AS FOUND (UUT)	MFG. ACCURACY

**COMMENTS**

**INL  
CALIBRATION INPUT DATA**

11/30/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 722722    Mfr: PCB    Model: 483A    Noun Name: AMPLIFIER    Serial #: 747

Calibration Date: 4/2/2007

**ACTION CODE****AS FOUND**

Next Cal Due Date: 4/2/2008	1	<input type="checkbox"/>	Acceptance Test	1	<input checked="" type="radio"/>	In Tolerance
Charge Level: 4	2	<input type="checkbox"/>	Special Test	2	<input type="radio"/>	Out of Tolerance >1x <2x
Repair/Adj/etc C.L.: 0	3	<input checked="" type="checkbox"/>	Calibration to MFG Specs	3	<input type="radio"/>	Out of Tolerance >2x <3x
Material Amount: 0	4	<input type="checkbox"/>	Clean	4	<input type="radio"/>	Out of Tolerance >3x <5x
Charge Number: 100853GSA	5	<input type="checkbox"/>	Limited Calibration	5	<input type="radio"/>	Out of Tolerance >5x
Cal Work Inst ID: 6053	6	<input type="checkbox"/>	Functional Check	6	<input type="radio"/>	Out of Tolerance-Undetermined
Outside Vendor:	7	<input type="checkbox"/>	Performance Check	7	<input type="checkbox"/>	Inoperative
	8	<input type="checkbox"/>	Modify	8	<input type="checkbox"/>	Damaged
	9	<input type="checkbox"/>	Repair-needs Charge Level	9	<input type="checkbox"/>	Not Used
	10	<input type="checkbox"/>	Other	10	<input type="checkbox"/>	Not Determined
				11	<input type="checkbox"/>	Excessed
Calibrated By: Brian Berls    S#: 102182    Phone: 526-2761				12	<input type="checkbox"/>	Extension

**CALIBRATION STANDARDS USED**

301893	290716	278037					

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-5.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

NOMINAL (STD)	OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION UNITS	AS FOUND (UUT)	MFG. ACCURACY

**COMMENTS**

**INL**  
**CALIBRATION INPUT DATA**

12/13/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 722722    Mfr: PCB    Model: 483A    Noun Name: AMPLIFIER    Serial #: 747

Calibration Date: 12/13/2007

**ACTION CODE****AS FOUND**

Next Cal Due Date: 12/13/2008	1	<input type="checkbox"/>	Acceptance Test	1	<input checked="" type="radio"/> In Tolerance
Charge Level: 4	2	<input type="checkbox"/>	Special Test	2	<input type="checkbox"/> Out of Tolerance >1x <2x
Repair/Adj/etc C.L: 0	3	<input checked="" type="checkbox"/>	Calibration to MFG Specs	3	<input type="checkbox"/> Out of Tolerance >2x <3x
Material Amount: 0	4	<input type="checkbox"/>	Clean	4	<input type="checkbox"/> Out of Tolerance >3x <5x
Charge Number: 100853GSA	5	<input type="checkbox"/>	Limited Calibration	5	<input type="checkbox"/> Out of Tolerance >5x
Cal Work Inst ID: 6053	6	<input type="checkbox"/>	Functional Check	6	<input type="checkbox"/> Out of Tolerance-Undetermined
Outside Vendor:	7	<input type="checkbox"/>	Performance Check	7	<input type="checkbox"/> Inoperative
	8	<input type="checkbox"/>	Modify	8	<input type="checkbox"/> Damaged
	9	<input type="checkbox"/>	Repair-needs Charge Level	9	<input type="checkbox"/> Not Used
	10	<input type="checkbox"/>	Other	10	<input type="checkbox"/> Not Determined
				11	<input type="checkbox"/> Excessed
Calibrated By: Brian Berls    S#: 102182    Phone: 526-2761				12	<input type="checkbox"/> Extension

**CALIBRATION STANDARDS USED**

290716	383171	301893					

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

NOMINAL (STD)	OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION UNITS	AS FOUND (UUT)	MFG. ACCURACY

COMMENTS

# **EXTENSOMETER**

**INL  
CALIBRATION INPUT DATA**

11/30/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 721774      Mfr: EPSILON      Model: 3542      Noun Name: EXTENSOMETER      Serial #: E82572-2"/ 2 INCH

Calibration Date: 1/26/2004 4:32:41 PM

**ACTION CODE**

**AS FOUND**

Next Cal Due Date: 1/26/2005

Charge Level: 8

Repair/Adj/etc C.L.: 0

Material Amount: 0

Charge Number: 100666H27

Cal Work Inst ID: 6021

Outside Vendor:

- |    |                                     |                           |    |                                  |                               |
|----|-------------------------------------|---------------------------|----|----------------------------------|-------------------------------|
| 1  | <input type="checkbox"/>            | Acceptance Test           | 1  | <input checked="" type="radio"/> | In Tolerance                  |
| 2  | <input type="checkbox"/>            | Special Test              | 2  | <input type="radio"/>            | Out of Tolerance >1x <2x      |
| 3  | <input checked="" type="checkbox"/> | Calibration to MFG Specs  | 3  | <input type="radio"/>            | Out of Tolerance >2x <3x      |
| 4  | <input type="checkbox"/>            | Clean                     | 4  | <input type="radio"/>            | Out of Tolerance >3x <5x      |
| 5  | <input type="checkbox"/>            | Limited Calibration       | 5  | <input type="radio"/>            | Out of Tolerance >5x          |
| 6  | <input type="checkbox"/>            | Functional Check          | 6  | <input type="radio"/>            | Out of Tolerance-Undetermined |
| 7  | <input type="checkbox"/>            | Performance Check         | 7  | <input type="checkbox"/>         | Inoperative                   |
| 8  | <input type="checkbox"/>            | Modify                    | 8  | <input type="checkbox"/>         | Damaged                       |
| 9  | <input type="checkbox"/>            | Repair-needs Charge Level | 9  | <input type="checkbox"/>         | Not Used                      |
| 10 | <input type="checkbox"/>            | Other                     | 10 | <input type="checkbox"/>         | Not Determined                |
|    |                                     |                           | 11 | <input type="checkbox"/>         | Excessed                      |
|    |                                     |                           | 12 | <input type="checkbox"/>         | Extension                     |

Calibrated By: Larry Deming

S#: 39571 Phone: 526-2761

**CALIBRATION STANDARDS USED**

350619	718306	719653	715111	718783			

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-5.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

**OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION**

NOMINAL (STD)	UNITS	AS FOUND (UUT)	MFG. ACCURACY

**COMMENTS**

Meets mfg. Tolerance.  
Initial calibration.  
QA#: 104740 104741 104742

**INL**  
**CALIBRATION INPUT DATA**

11/30/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 721774    Mfr: EPSILON    Model: 3542    Noun Name: EXTENSOMETER    Serial #: E82572-2"/ 2 INCH

Calibration Date: 2/8/2005

**ACTION CODE****AS FOUND**

Next Cal Due Date: 2/8/2006

Charge Level: 8

Repair/Adj/etc C.L: 0

Material Amount: 0

Charge Number: 100108GSA

Cal Work Inst ID: 6021

Outside Vendor:

- |    |                                     |                           |    |                                  |                               |
|----|-------------------------------------|---------------------------|----|----------------------------------|-------------------------------|
| 1  | <input type="checkbox"/>            | Acceptance Test           | 1  | <input checked="" type="radio"/> | In Tolerance                  |
| 2  | <input type="checkbox"/>            | Special Test              | 2  | <input type="radio"/>            | Out of Tolerance >1x <2x      |
| 3  | <input checked="" type="checkbox"/> | Calibration to MFG Specs  | 3  | <input type="radio"/>            | Out of Tolerance >2x <3x      |
| 4  | <input type="checkbox"/>            | Clean                     | 4  | <input type="radio"/>            | Out of Tolerance >3x <5x      |
| 5  | <input type="checkbox"/>            | Limited Calibration       | 5  | <input type="radio"/>            | Out of Tolerance >5x          |
| 6  | <input type="checkbox"/>            | Functional Check          | 6  | <input type="radio"/>            | Out of Tolerance-Undetermined |
| 7  | <input type="checkbox"/>            | Performance Check         | 7  | <input type="checkbox"/>         | Inoperative                   |
| 8  | <input type="checkbox"/>            | Modify                    | 8  | <input type="checkbox"/>         | Damaged                       |
| 9  | <input type="checkbox"/>            | Repair-needs Charge Level | 9  | <input type="checkbox"/>         | Not Used                      |
| 10 | <input type="checkbox"/>            | Other                     | 10 | <input type="checkbox"/>         | Not Determined                |
|    |                                     |                           | 11 | <input type="checkbox"/>         | Excessed                      |
|    |                                     |                           | 12 | <input type="checkbox"/>         | Extension                     |

Calibrated By: Larry Deming    S#: 39571    Phone: 526-2761

**CALIBRATION STANDARDS USED**

350619	718306	718416	721053	715111			

**STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES**

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

**Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.**

**OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION****NOMINAL (STD)****UNITS****AS FOUND (UUT)****MFG. ACCURACY**


**COMMENTS**

**INL**  
**CALIBRATION INPUT DATA**

11/30/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 721774    Mfr: EPSILON    Model: 3542    Noun Name: EXTENSOMETER    Serial #: E82572-2"/2 INCH

Calibration Date: 2/8/2006

**ACTION CODE****AS FOUND**

Next Cal Due Date: 2/8/2007	1	<input type="checkbox"/>	Acceptance Test	1	<input checked="" type="radio"/>	In Tolerance
Charge Level: 8	2	<input type="checkbox"/>	Special Test	2	<input type="radio"/>	Out of Tolerance >1x <2x
Repair/Adj/etc C.L: 1	3	<input checked="" type="checkbox"/>	Calibration to MFG Specs	3	<input type="radio"/>	Out of Tolerance >2x <3x
Material Amount: 0	4	<input type="checkbox"/>	Clean	4	<input type="radio"/>	Out of Tolerance >3x <5x
Charge Number: 100853GSA	5	<input type="checkbox"/>	Limited Calibration	5	<input type="radio"/>	Out of Tolerance >5x
Cal Work Inst ID: 6021	6	<input type="checkbox"/>	Functional Check	6	<input type="radio"/>	Out of Tolerance-Undetermined
Outside Vendor:	7	<input type="checkbox"/>	Performance Check	7	<input type="checkbox"/>	Inoperative
	8	<input type="checkbox"/>	Modify	8	<input type="checkbox"/>	Damaged
	9	<input type="checkbox"/>	Repair-needs Charge Level	9	<input type="checkbox"/>	Not Used
	10	<input type="checkbox"/>	Other	10	<input type="checkbox"/>	Not Determined
				11	<input type="checkbox"/>	Excessed
Calibrated By: Larry Deming    S#: 39571    Phone: 526-2761				12	<input type="checkbox"/>	Extension

**CALIBRATION STANDARDS USED**

350619	715111	718783	718416	718306			

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

**OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION****NOMINAL (STD)****UNITS****AS FOUND (UUT)****MFG. ACCURACY**


**COMMENTS**

Instrument was found to be very close to being out of tolerance. Re-ran Travel verses VDC to re-establish values.  
Repair= (Optimized)



# **HIGH-SPEED DIGITAL CAMERA**

**INL  
CALIBRATION INPUT DATA**

12/4/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 376450 Mfr: PHOTRON Model: APX Noun Name: HIGH SPEED DIGITAL

CAMERA Serial #: 118114112

Calibration Date: 4/1/2004

**ACTION CODE****AS FOUND**

Next Cal Due Date:	1	<input type="checkbox"/>	Acceptance Test	1	<input type="radio"/>	In Tolerance
Charge Level: 4	2	<input checked="" type="checkbox"/>	Special Test	2	<input type="radio"/>	Out of Tolerance >1x <2x
Repair/Adj/etc C.L: 0	3	<input type="checkbox"/>	Calibration to MFG Specs	3	<input type="radio"/>	Out of Tolerance >2x <3x
Material Amount: 0	4	<input type="checkbox"/>	Clean	4	<input type="radio"/>	Out of Tolerance >3x <5x
Charge Number: 100664GSA	5	<input type="checkbox"/>	Limited Calibration	5	<input type="radio"/>	Out of Tolerance >5x
Cal Work Inst ID: SPEC	6	<input type="checkbox"/>	Functional Check	6	<input type="radio"/>	Out of Tolerance-Undetermined
Outside Vendor:	7	<input type="checkbox"/>	Performance Check	7	<input type="checkbox"/>	Inoperative
	8	<input type="checkbox"/>	Modify	8	<input type="checkbox"/>	Damaged
	9	<input type="checkbox"/>	Repair-needs Charge Level	9	<input type="checkbox"/>	Not Used
	10	<input checked="" type="checkbox"/>	Other	10	<input checked="" type="checkbox"/>	Not Determined
				11	<input type="checkbox"/>	Excessed
Calibrated By: Scott Lish				12	<input type="checkbox"/>	Extension

**CALIBRATION STANDARDS USED**

700661							

**STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES**

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-5.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

**OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION**

NOMINAL (STD)	UNITS	AS FOUND (UUT)	MFG. ACCURACY

**COMMENTS**

Three runs were taken on the camera to assure repeatability. The test point was off the General Out, Positive Trigger coax connection. The uncertainty analysis was based on the uncertainty of the frequency counter and statistical analysis.

Bechtel B&W Idaho  
Standards and Calibration Laboratories

# Certificate of Calibration

For

Instrument:	<u>High Speed Digital Camera</u>	Submitted by:	<u>Keith Morton</u>
Manufacturer:	<u>Photron</u>	Address:	<u>INEEL (EROB)</u>
Model No.:	<u>APX</u>		
Serial No.:	<u>118114112</u>		
Calibration Recall No.:	<u>376450</u>		

This high-speed camera was compared directly with a frequency counter traceable to the National Institute of Standards and Technology (NIST). A minimum of three runs were made on this high-speed camera to ensure the reproducibility of the data. The certified value is the mean of all measurements.

Calibration Date:	<u>4/01/04</u>	Measurements performed by:	<u>J. Scott Lish</u>
Procedure No.:	<u>NA</u>	Certificate No.:	<u>04-025</u>
Temperature:	<u>23°C</u>	Humidity:	<u>35%RH</u>

Approved by:

William J. Albed  
Metrology Manager)

Bechtel B&W Idaho  
Standards and Calibration Laboratories  
P.O. Box 1625  
Idaho Falls, Idaho 83415-4137  
208.526.2017

# ***Bechtel B&W Idaho***

## **Standards & Calibration Laboratories Bechtel B&W Idaho Idaho Falls, Idaho**

Instrument:	<u>High Speed Digital Camera</u>	Certificate No.:	<u>04-025</u>
Manufacturer:	<u>Photron</u>	Calibration Date:	<u>4/01/04</u>
Model No.:	<u>APX</u>	Serial No.:	<u>118114112</u>
Calibration Recall No.:	<u>376450</u>		

**Traceability:** Values and the associated uncertainties supplied by the Standards and Calibration Laboratory (S&CL) are traceable to one or more of the following:

- A. The values of the units (either base or derived) maintained and disseminated by a national measurement institute such as NIST, NRC, etc.
- B. Intrinsic standards - fundamental physical phenomena
- C. Standards derived by the ratio of self-calibration techniques.

The values of the standard then transferred to other measurement standards by the use of approved procedures that incorporate nationally acceptable techniques thus allowing for an unbroken chain of traceability. These values are checked with round robins and the use of check standards. Finally, the values are transferred to the general-purpose measurement and test equipment using approved procedures and calculated uncertainties using tech note 1297.

**Note 1:** This certificate or report shall not be reproduced, except in full, without the advance written approval of the Idaho National Engineering and Environmental Laboratory (INEEL) Standards and Calibration Laboratories, Idaho Falls, Idaho.

**Note 2:** The as-received condition of the standard, set of standards, or measurement equipment described herein was as expected, unless noted in the body of the certificate or report.

# *Bechtel B&W Idaho*

## Standards & Calibration Laboratories Bechtel B&W Idaho Idaho Falls, Idaho

Instrument: High Speed Digital Camera Certificate No.: 04-025  
Manufacturer: Photron Calibration Date: 04/01/04  
Model No.: APX Serial No.: 118114112  
Calibration Recall No.: 376450

### Frames Per Second

Measured Value	Nominal	Uncertainty
59.99933	60	2 ppm
124.99861	125	2 ppm
249.9972	250	2 ppm
499.9944	500	2 ppm
999.9888	1,000	2 ppm
1,999.977	2,000	2 ppm
2,999.966	3,000	2 ppm
3,999.955	4,000	2 ppm
5,999.933	6,000	2 ppm
7,999.910	8,000	2 ppm
9,999.889	10,000	2 ppm
14,999.833	15,000	2 ppm
19,999.77	20,000	2 ppm
23,999.73	24,000	2 ppm
29,999.66	30,000	2 ppm
39,999.55	40,000	2 ppm
49,999.44	50,000	2 ppm
87,599.02	87,600	2 ppm
99,998.88	100,000	2 ppm
119,998.65	120,000	2 ppm

Note: In the determination of the uncertainties there is no allowance for drift, temperature coefficients, power coefficients, etc. Measured uncertainty is for time of measurement only at a K-factor of two.

Standards Used				
ID No.	Manufacturer	Model No.	Description	Due Date
700661	HP	5316B	Frequency Counter	6/04/04

# INL CALIBRATION INPUT DATA

12/4/2007

NAME: DANA KEITH MORTON      BADGE: 35698      PH: 526-1274      AREA: STC      BLDG: EROB      RM: W2D1

ID Number: 376450      Mfr: PHOTRON      Model: APX      Noun Name: HIGH SPEED DIGITAL  
CAMERA      Serial #: 118114112

Calibration Date: 10/19/2004

## ACTION CODE

## AS FOUND

Next Cal Due Date:	1	<input type="checkbox"/>	Acceptance Test	1	<input type="radio"/>	In Tolerance
Charge Level: 4	2	<input checked="" type="checkbox"/>	Special Test	2	<input type="radio"/>	Out of Tolerance >1x <2x
Repair/Adj/etc C.L: 0	3	<input type="checkbox"/>	Calibration to MFG Specs	3	<input type="radio"/>	Out of Tolerance >2x <3x
Material Amount: 0	4	<input type="checkbox"/>	Clean	4	<input type="radio"/>	Out of Tolerance >3x <5x
Charge Number: 100664GSA	5	<input type="checkbox"/>	Limited Calibration	5	<input type="radio"/>	Out of Tolerance >5x
Cal Work Inst ID: SPEC	6	<input type="checkbox"/>	Functional Check	6	<input type="radio"/>	Out of Tolerance-Undetermined
Outside Vendor:	7	<input type="checkbox"/>	Performance Check	7	<input type="checkbox"/>	Inoperative
	8	<input type="checkbox"/>	Modify	8	<input type="checkbox"/>	Damaged
	9	<input type="checkbox"/>	Repair-needs Charge Level	9	<input type="checkbox"/>	Not Used
	10	<input type="checkbox"/>	Other	10	<input checked="" type="checkbox"/>	Not Determined
				11	<input type="checkbox"/>	Excessed
Calibrated By: Scott Lish      S#: 101141      Phone: 526-2761				12	<input type="checkbox"/>	Extension

## CALIBRATION STANDARDS USED

700661							

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

## LABORATORY TEMPERATURE AND HUMIDITY

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

## OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION

NOMINAL (STD)	UNITS	AS FOUND (UUT)	MFG. ACCURACY

## COMMENTS

Three runs were taken on the camera to assure repeatability. The test point was off the General Out, Positive Trigger coax connection. The uncertainty analysis was based on the uncertainty of the frequency counter and statistical analysis. See history folder for analysis.

Bechtel B&W Idaho  
Standards and Calibration Laboratories

# Certificate of Calibration

For

Instrument:	<u>High Speed Digital Camera</u>	Submitted by:	<u>Keith Morton</u>
Manufacturer:	<u>Photron</u>	Address:	<u>INEEL (EROB)</u>
Model No.:	<u>APX</u>		
Serial No.:	<u>118114112</u>		
Calibration Recall No.:	<u>376450</u>		

This high-speed camera was compared directly with a frequency counter traceable to the National Institute of Standards and Technology (NIST). A minimum of three runs were made on this high-speed camera to ensure the reproducibility of the data. The certified value is the mean of all measurements.

Calibration Date:	<u>10/19/04</u>	Measurements performed by:	<u>J. Scott Lish</u>
Procedure No.:	<u>NA</u>	Certificate No.:	<u>04-034</u>
Temperature:	<u>23°C</u>	Humidity:	<u>35%RH</u>

Approved by: *Mike Stears*  
Metrology Manager)

Bechtel B&W Idaho  
Standards and Calibration Laboratories  
P.O. Box 1625  
Idaho Falls, Idaho 83415-4137  
208.526.2017

# ***Bechtel B&W Idaho***

## **Standards & Calibration Laboratories Bechtel B&W Idaho Idaho Falls, Idaho**

Instrument:	<u>High Speed Digital Camera</u>	Certificate No.:	<u>04-034</u>
Manufacturer:	<u>Photron</u>	Calibration Date:	<u>10/19/04</u>
Model No.:	<u>APX</u>	Serial No.:	<u>118114112</u>
Calibration Recall No.:	<u>376450</u>		

Traceability: Values and the associated uncertainties supplied by the Standards and Calibration Laboratory (S&CL) are traceable to one or more of the following:

- A. The values of the units (either base or derived) maintained and disseminated by a national measurement institute such as NIST, NRC, etc.
- B. Intrinsic standards - fundamental physical phenomena
- C. Standards derived by the ratio of self-calibration techniques.

The values of the standard are then transferred to other measurement standards by the use of approved procedures that incorporate nationally acceptable techniques thus allowing for an unbroken chain of traceability. These values are checked with round robins and the use of check standards. Finally, the values are transferred to the general-purpose measurement and test equipment using approved procedures and calculated uncertainties using tech note 1297.

Note 1: This certificate or report shall not be reproduced, except in full, without the advance written approval of the Idaho National Engineering and Environmental Laboratory (INEEL) Standards and Calibration Laboratories, Idaho Falls, Idaho.

Note 2: The as-received condition of the standard, set of standards, or measurement equipment described herein was as expected, unless noted in the body of the certificate or report.



# *Bechtel B&W Idaho*

## Standards & Calibration Laboratories Bechtel B&W Idaho Idaho Falls, Idaho

Instrument: High Speed Digital Camera Certificate No.: 04-034  
Manufacturer: Photron Calibration Date: 10/19/04  
Model No.: APX Serial No.: 118114112  
Calibration Recall No.: 376450

### Frames Per Second

Measured Value	Nominal	Uncertainty
59.99932	60	3 ppm
124.99859	125	3 ppm
249.9971	250	3 ppm
499.9943	500	3 ppm
999.9888	1,000	3 ppm
1,999.977	2,000	3 ppm
2,999.966	3,000	3 ppm
3,999.955	4,000	3 ppm
5,999.933	6,000	3 ppm
7,999.909	8,000	3 ppm
9,999.887	10,000	3 ppm
14,999.830	15,000	3 ppm
19,999.77	20,000	3 ppm
23,999.72	24,000	3 ppm
29,999.66	30,000	3 ppm
39,999.54	40,000	3 ppm
49,999.43	50,000	3 ppm
87,599.01	87,600	3 ppm
99,998.87	100,000	3 ppm
119,998.63	120,000	3 ppm

Note: In the determination of the uncertainties there is no allowance for drift, temperature coefficients, power coefficients, etc. Measured uncertainty is for time of measurement only at a K-factor of two.

Standards Used				
ID No.	Manufacturer	Model No.	Description	Due Date
700661	HP	5316B	Frequency Counter	12/14/04

# INL CALIBRATION INPUT DATA

12/4/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 376450      Mfr: PHOTRON      Model: APX      Noun Name: HIGH SPEED DIGITAL  
CAMERA      Serial #: 118114112

Calibration Date: 8/3/2005

## ACTION CODE

## AS FOUND

Next Cal Due Date: 8/3/2006	1	<input type="checkbox"/>	Acceptance Test	1	<input checked="" type="radio"/>	In Tolerance
Charge Level: 8	2	<input checked="" type="checkbox"/>	Special Test	2	<input type="radio"/>	Out of Tolerance >1x <2x
Repair/Adj/etc C.L.: 0	3	<input type="checkbox"/>	Calibration to MFG Specs	3	<input type="radio"/>	Out of Tolerance >2x <3x
Material Amount: 0	4	<input type="checkbox"/>	Clean	4	<input type="radio"/>	Out of Tolerance >3x <5x
Charge Number: 100853GSA	5	<input type="checkbox"/>	Limited Calibration	5	<input type="radio"/>	Out of Tolerance >5x
Cal Work Inst ID: SPEC	6	<input type="checkbox"/>	Functional Check	6	<input type="radio"/>	Out of Tolerance-Undetermined
Outside Vendor:	7	<input type="checkbox"/>	Performance Check	7	<input type="checkbox"/>	Inoperative
	8	<input type="checkbox"/>	Modify	8	<input type="checkbox"/>	Damaged
	9	<input type="checkbox"/>	Repair-needs Charge Level	9	<input type="checkbox"/>	Not Used
	10	<input type="checkbox"/>	Other	10	<input checked="" type="checkbox"/>	Not Determined
				11	<input type="checkbox"/>	Excessed
Calibrated By: Scott Lish	S#: 101141	Phone: 526-2761		12	<input type="checkbox"/>	Extension

## CALIBRATION STANDARDS USED

712330							

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

## LABORATORY TEMPERATURE AND HUMIDITY

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

## OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION

NOMINAL (STD)	UNITS	AS FOUND (UUT)	MFG. ACCURACY

## COMMENTS

Three runs were taken on the camera to assure repeatability. The test point was off the General Out, Positive Trigger coax connection. The uncertainty analysis was based on the uncertainty of the frequency counter and statistical analysis. See history folder for analysis.

**INL**  
**CALIBRATION INPUT DATA**

12/4/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 376450      Mfr: PHOTRON  
CAMERA      Serial #: 118114112

Model: APX

Noun Name: HIGH SPEED DIGITAL

Calibration Date: 8/14/2006

**ACTION CODE****AS FOUND**

Next Cal Due Date: 7/26/2007	1	<input type="checkbox"/>	Acceptance Test	1	<input checked="" type="radio"/>	In Tolerance
Charge Level: 8	2	<input checked="" type="checkbox"/>	Special Test	2	<input type="radio"/>	Out of Tolerance >1x <2x
Repair/Adj/etc C.L: 0	3	<input type="checkbox"/>	Calibration to MFG Specs	3	<input type="radio"/>	Out of Tolerance >2x <3x
Material Amount: 0	4	<input type="checkbox"/>	Clean	4	<input type="radio"/>	Out of Tolerance >3x <5x
Charge Number: 100853GSA	5	<input type="checkbox"/>	Limited Calibration	5	<input type="radio"/>	Out of Tolerance >5x
Cal Work Inst ID: SPEC	6	<input type="checkbox"/>	Functional Check	6	<input type="radio"/>	Out of Tolerance-Undetermined
Outside Vendor:	7	<input type="checkbox"/>	Performance Check	7	<input type="checkbox"/>	Inoperative
	8	<input type="checkbox"/>	Modify	8	<input type="checkbox"/>	Damaged
	9	<input type="checkbox"/>	Repair-needs Charge Level	9	<input type="checkbox"/>	Not Used
	10	<input type="checkbox"/>	Other	10	<input type="checkbox"/>	Not Determined
				11	<input type="checkbox"/>	Excessed
Calibrated By: Scott Lish      S#: 101141      Phone: 526-2761				12	<input type="checkbox"/>	Extension

**CALIBRATION STANDARDS USED**

700661							

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED  
VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

NOMINAL (STD)	OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION UNITS	AS FOUND (UUT)	MFG. ACCURACY

**COMMENTS**

Three runs were taken on the camera to assure repeatability. The test point was off the General Out, Positive Trigger coax connection. The uncertainty analysis was based on the uncertainty of the frequency counter and statistical analysis. See history folder for analysis.

**INL**  
**CALIBRATION INPUT DATA**

12/4/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 376450      Mfr: PHOTRON      Model: APX      Noun Name: HIGH SPEED DIGITAL  
CAMERA      Serial #: 118114112

Calibration Date: 7/18/2007

**ACTION CODE****AS FOUND**

Next Cal Due Date: 7/18/2008	1	<input type="checkbox"/>	Acceptance Test	1	<input checked="" type="radio"/>	In Tolerance
Charge Level: 8	2	<input type="checkbox"/>	Special Test	2	<input type="radio"/>	Out of Tolerance >1x <2x
Repair/Adj/etc C.L.: 0	3	<input checked="" type="checkbox"/>	Calibration to MFG Specs	3	<input type="radio"/>	Out of Tolerance >2x <3x
Material Amount: 0	4	<input type="checkbox"/>	Clean	4	<input type="radio"/>	Out of Tolerance >3x <5x
Charge Number: 100853GSA	5	<input type="checkbox"/>	Limited Calibration	5	<input type="radio"/>	Out of Tolerance >5x
Cal Work Inst ID: 6142	6	<input type="checkbox"/>	Functional Check	6	<input type="radio"/>	Out of Tolerance-Undetermined
Outside Vendor:	7	<input type="checkbox"/>	Performance Check	7	<input type="checkbox"/>	Inoperative
	8	<input type="checkbox"/>	Modify	8	<input type="checkbox"/>	Damaged
	9	<input type="checkbox"/>	Repair-needs Charge Level	9	<input type="checkbox"/>	Not Used
	10	<input type="checkbox"/>	Other	10	<input type="checkbox"/>	Not Determined
				11	<input type="checkbox"/>	Excessed
Calibrated By: Scott Lish      S#: 101141      Phone: 526-2761				12	<input type="checkbox"/>	Extension

**CALIBRATION STANDARDS USED**

700661							

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED  
VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

**OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION**

NOMINAL (STD)	UNITS	AS FOUND (UUT)	MFG. ACCURACY

**COMMENTS**

Three runs were taken on the camera to assure repeatability. The test point was off the General Out, Positive Trigger coax connection.

# **LOAD CELLS**

# **INL CALIBRATION INPUT DATA**

12/20/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 721515      Mfr: INTERFACE      Model: 500 LB      Noun Name: LOAD CELL      Serial #:  
143547A/1210AF

Calibration Date: 9/9/2003

**ACTION CODE****AS FOUND**

Next Cal Due Date: 6/9/2004

Charge Level: 12

Repair/Adj/etc C.L: 0

Material Amount: 0

Charge Number: 100348027

Cal Work Inst ID: 5748C

Outside Vendor:

- |    |                                     |                           |    |                                  |                               |
|----|-------------------------------------|---------------------------|----|----------------------------------|-------------------------------|
| 1  | <input type="checkbox"/>            | Acceptance Test           | 1  | <input checked="" type="radio"/> | In Tolerance                  |
| 2  | <input type="checkbox"/>            | Special Test              | 2  | <input type="radio"/>            | Out of Tolerance >1x <2x      |
| 3  | <input type="checkbox"/>            | Calibration to MFG Specs  | 3  | <input type="radio"/>            | Out of Tolerance >2x <3x      |
| 4  | <input type="checkbox"/>            | Clean                     | 4  | <input type="radio"/>            | Out of Tolerance >3x <5x      |
| 5  | <input checked="" type="checkbox"/> | Limited Calibration       | 5  | <input type="radio"/>            | Out of Tolerance >5x          |
| 6  | <input type="checkbox"/>            | Functional Check          | 6  | <input type="radio"/>            | Out of Tolerance-Undetermined |
| 7  | <input type="checkbox"/>            | Performance Check         | 7  | <input type="checkbox"/>         | Inoperative                   |
| 8  | <input type="checkbox"/>            | Modify                    | 8  | <input type="checkbox"/>         | Damaged                       |
| 9  | <input type="checkbox"/>            | Repair-needs Charge Level | 9  | <input type="checkbox"/>         | Not Used                      |
| 10 | <input type="checkbox"/>            | Other                     | 10 | <input type="checkbox"/>         | Not Determined                |
|    |                                     |                           | 11 | <input type="checkbox"/>         | Excessed                      |
|    |                                     |                           | 12 | <input type="checkbox"/>         | Extension                     |

Calibrated By: Stan Zohner      S#: 58146      Phone: 526-2761

**CALIBRATION STANDARDS USED**

721171	721190						

**STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED  
VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES**

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

NOMINAL (STD)	OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION		MFG. ACCURACY
	~ UNITS	AS FOUND (UUT)	

**COMMENTS**

721171 THRU 721190  
INITIAL CALIBRATION  
LIMITED: MUST BE USED WITH READOUT P# 721514 CH 1  
ECAL 2.174 ESCL 499.6

**INL  
CALIBRATION INPUT DATA**

12/20/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 721515      Mfr: INTERFACE      Model: 500 LB      Noun Name: LOAD CELL      Serial #:

Calibration Date: 1/22/2004 10:59:06 AM

**ACTION CODE****AS FOUND**

Next Cal Due Date:	1	<input type="checkbox"/>	Acceptance Test	1	<input type="radio"/>	In Tolerance
Charge Level: 2	2	<input type="checkbox"/>	Special Test	2	<input type="radio"/>	Out of Tolerance >1x <2x
Repair/Adj/etc C.L: 0	3	<input type="checkbox"/>	Calibration to MFG Specs	3	<input type="radio"/>	Out of Tolerance >2x <3x
Material Amount: 0	4	<input type="checkbox"/>	Clean	4	<input type="radio"/>	Out of Tolerance >3x <5x
Charge Number: 100664gsx	5	<input type="checkbox"/>	Limited Calibration	5	<input type="radio"/>	Out of Tolerance >5x
Cal Work Inst ID: 5748C	6	<input type="checkbox"/>	Functional Check	6	<input type="radio"/>	Out of Tolerance-Undetermined
Outside Vendor:	7	<input type="checkbox"/>	Performance Check	7	<input type="checkbox"/>	Inoperative
	8	<input type="checkbox"/>	Modify	8	<input checked="" type="checkbox"/>	Damaged
	9	<input type="checkbox"/>	Repair-needs Charge Level	9	<input type="checkbox"/>	Not Used
	10	<input checked="" type="checkbox"/>	Other	10	<input type="checkbox"/>	Not Determined
				11	<input type="checkbox"/>	Excessed
				12	<input type="checkbox"/>	Extension

Calibrated By: Stan Zohner

S#: 58146 Phone: 526-2761

**CALIBRATION STANDARDS USED**


**STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED  
VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES**

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-5.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

**OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION**

NOMINAL (STD)	UNITS	AS FOUND (UUT)	MFG. ACCURACY

**COMMENTS**

USER SAID WAS OVERANGED. UNIT IS NOT WORKING PROPERLY RECOMMEND EXCESS

**ITM weights under 500 pounds were initially weighed with load cell #721515 in January 2004.**

**Afterwards, that load cell was damaged prior to re-calibration so weights under 500 pounds were rechecked in February 2004 using load cell # 721882. The weights were found acceptable as marked.**

**INL Calibration Laboratory measurements in October 2007 also confirmed accuracy of weights weighing less than 500 pounds.**



# INL CALIBRATION INPUT DATA

12/20/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 721882      Mfr: INTERFACE      Model: 500 LB      Noun Name: LOAD CELL      Serial #: 147313A

Calibration Date: 2/19/2004 4:09:47 PM

**ACTION CODE****AS FOUND**

Next Cal Due Date: 11/19/2004

1 ☐ Acceptance Test1 ☒ In Tolerance

Charge Level: 10

2 ☐ Special Test2 ☐ Out of Tolerance >1x <2x

Repair/Adj/etc C.L: 0

3 ☐ Calibration to MFG Specs3 ☐ Out of Tolerance >2x <3x

Material Amount: 0

4 ☐ Clean4 ☐ Out of Tolerance >3x <5x

Charge Number: 100664gsx

5 ☒ Limited Calibration5 ☐ Out of Tolerance >5x

Cal Work Inst ID: 5748C

6 ☐ Functional Check6 ☐ Out of Tolerance-Undetermined

Outside Vendor:

7 ☐ Performance Check7 ☐ Inoperative8 ☐ Modify8 ☐ Damaged9 ☐ Repair-needs Charge Level9 ☐ Not Used10 ☐ Other10 ☐ Not Determined11 ☐ Excessed

Calibrated By: Stan Zohner

S#: 58146 Phone: 526-2761

12 ☐ Extension**CALIBRATION STANDARDS USED**

375257							

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION			
NOMINAL (STD)	UNITS	AS FOUND (UUT)	MFG. ACCURACY

**COMMENTS**

INITIAL CALIBRATION  
LIMITED: MUST BE USED WITH READOUT P# 721514  
MUST BE USED ON CHANNEL 1  
ECL 2.121 AND ESCL 500.0

# INL CALIBRATION INPUT DATA

12/20/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 721882      Mfr: INTERFACE      Model: 500 LB      Noun Name: LOAD CELL      Serial #: 147313A

Calibration Date: 3/10/2005

**ACTION CODE****AS FOUND**

Next Cal Due Date: 3/10/2006

Charge Level: 10

Repair/Adj/etc C.L: 0

Material Amount: 0

Charge Number: 100853GSA

Cal Work Inst ID: 5748F

Outside Vendor:

1 ☐

Acceptance Test

2 ☐

Special Test

3 ☐

Calibration to MFG Specs

4 ☐

Clean

5 ☒

Limited Calibration

6 ☐

Functional Check

7 ☐

Performance Check

8 ☐

Modify

9 ☐

Repair-needs Charge Level

10 ☐

Other

1 ☒ In Tolerance2 ☐ Out of Tolerance >1x <2x3 ☐ Out of Tolerance >2x <3x4 ☐ Out of Tolerance >3x <5x5 ☐ Out of Tolerance >5x6 ☐ Out of Tolerance-Undetermined7 ☐ Inoperative8 ☐ Damaged9 ☐ Not Used10 ☐ Not Determined11 ☐ Excessed12 ☐ Extension

Calibrated By: Stan Zohner

S#: 58146 Phone: 526-2761

**CALIBRATION STANDARDS USED**

350816	350815	722028					

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED  
VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

**OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION**

NOMINAL (STD)

UNITS

AS FOUND (UUT)

MFG. ACCURACY


**COMMENTS**

LIMITED: MUST BE USED WITH READOUT P# 721514  
MUST BE USED ON CHANNEL 1  
ECAL 2.121 AND ESCL 500.0

**INL  
CALIBRATION INPUT DATA**

12/20/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 721882	Mfr: INTERFACE	Model: 500 LB	Noun Name: LOAD CELL	Serial #: 147313A
-------------------	----------------	---------------	----------------------	-------------------

Calibration Date: 8/28/2006

**ACTION CODE****AS FOUND**

Next Cal Due Date: 8/28/2007

Charge Level: 10

Repair/Adj/etc C.L: 0

Material Amount: 0

Charge Number: 100853GSA

Cal Work Inst ID: 5748H

Outside Vendor:

1 ☐ Acceptance Test2 ☐ Special Test3 ☐ Calibration to MFG Specs4 ☐ Clean5 ☒ Limited Calibration6 ☐ Functional Check7 ☐ Performance Check8 ☐ Modify9 ☐ Repair-needs Charge Level10 ☐ Other1 ☒ In Tolerance2 ☐ Out of Tolerance >1x <2x3 ☐ Out of Tolerance >2x <3x4 ☐ Out of Tolerance >3x <5x5 ☐ Out of Tolerance >5x6 ☐ Out of Tolerance-Undetermined7 ☐ Inoperative8 ☐ Damaged9 ☐ Not Used10 ☐ Not Determined11 ☐ Exceeded12 ☐ Extension

Calibrated By: Stan Zohner

S#: 58146 Phone: 526-2761

**CALIBRATION STANDARDS USED**

375257							

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED  
VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-5.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

NOMINAL (STD)	OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION UNITS	AS FOUND (UUT)	MFG. ACCURACY

**COMMENTS**

LIMITED: MUST BE USED WITH READOUT P# 721514  
MUST BE USED ON CHANNEL 1  
ECAL 2.121 AND ESCL 500.0

# INL CALIBRATION INPUT DATA

12/20/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 721516      Mfr: INTERFACE      Model: 5000 LB (D & mV)      Noun Name: LOAD CELL      Serial #: 144177A/1210AF

Calibration Date: 9/8/2003

## ACTION CODE

## AS FOUND

Next Cal Due Date: 6/8/2004	1	<input type="checkbox"/>	Acceptance Test	1	<input checked="" type="radio"/>	In Tolerance
Charge Level: 12	2	<input type="checkbox"/>	Special Test	2	<input type="radio"/>	Out of Tolerance >1x <2x
Repair/Adj/etc C.L: 0	3	<input type="checkbox"/>	Calibration to MFG Specs	3	<input type="radio"/>	Out of Tolerance >2x <3x
Material Amount: 0	4	<input type="checkbox"/>	Clean	4	<input type="radio"/>	Out of Tolerance >3x <5x
Charge Number: 100348027	5	<input checked="" type="checkbox"/>	Limited Calibration	5	<input type="radio"/>	Out of Tolerance >5x
Cal Work Inst ID: 5748C	6	<input type="checkbox"/>	Functional Check	6	<input type="radio"/>	Out of Tolerance-Undetermined
Outside Vendor:	7	<input type="checkbox"/>	Performance Check	7	<input type="checkbox"/>	Inoperative
	8	<input type="checkbox"/>	Modify	8	<input type="checkbox"/>	Damaged
	9	<input type="checkbox"/>	Repair-needs Charge Level	9	<input type="checkbox"/>	Not Used
	10	<input type="checkbox"/>	Other	10	<input type="checkbox"/>	Not Determined
				11	<input type="checkbox"/>	Excessed
Calibrated By: Stan Zohner      S#: 58146      Phone: 526-2761				12	<input type="checkbox"/>	Extension

## CALIBRATION STANDARDS USED

708595	711804	714631					

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

## LABORATORY TEMPERATURE AND HUMIDITY

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

NOMINAL (STD)	OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION		MFG. ACCURACY
	UNITS	AS FOUND (UUT)	


## COMMENTS

INITIAL CALIBRATION  
LIMITED: MUST BE USED WITH READOUT P# 721514 CH 2  
ECAL 4.005 ESCL 4665

# INL CALIBRATION INPUT DATA

12/20/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 721516 Mfr: INTERFACE Model: 5000 LB (D & mV) Noun Name: LOAD CELL Serial  
 #: 144177A/1210AF

Calibration Date: 6/17/2004

## ACTION CODE

## AS FOUND

Next Cal Due Date: 3/17/2005	1	<input type="checkbox"/>	Acceptance Test	1	<input type="radio"/>	In Tolerance
Charge Level: 64	2	<input type="checkbox"/>	Special Test	2	<input checked="" type="radio"/>	Out of Tolerance >1x <2x
Repair/Adj/etc C.L: 0	3	<input type="checkbox"/>	Calibration to MFG Specs	3	<input type="radio"/>	Out of Tolerance >2x <3x
Material Amount: 0	4	<input type="checkbox"/>	Clean	4	<input type="radio"/>	Out of Tolerance >3x <5x
Charge Number: 100664GSA	5	<input checked="" type="checkbox"/>	Limited Calibration	5	<input type="radio"/>	Out of Tolerance >5x
Cal Work Inst ID: 5748C	6	<input type="checkbox"/>	Functional Check	6	<input type="radio"/>	Out of Tolerance-Undetermined
Outside Vendor:	7	<input type="checkbox"/>	Performance Check	7	<input type="checkbox"/>	Inoperative
	8	<input type="checkbox"/>	Modify	8	<input type="checkbox"/>	Damaged
	9	<input type="checkbox"/>	Repair-needs Charge Level	9	<input type="checkbox"/>	Not Used
	10	<input type="checkbox"/>	Other	10	<input type="checkbox"/>	Not Determined
				11	<input type="checkbox"/>	Excessed
Calibrated By: Stan Zohner S#: 58146 Phone: 526-2761				12	<input type="checkbox"/>	Extension

## CALIBRATION STANDARDS USED

715606	714631	715559	714644	709226			

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

## LABORATORY TEMPERATURE AND HUMIDITY

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

## OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION

NOMINAL (STD)	UNITS	AS FOUND (UUT)	MFG. ACCURACY
4492.6238	LBS/TENS/DISP	4500	+/- 6
4692.369	LBS/TENS/DISP	4700	+/- 6
4493.3956	LBS/COMP/DISP	4500	+/- 6
4692.9214	LBS/COMP/DISP	4700	+/- 6

## COMMENTS

UNIT WAS OUT OF TOLERANCE IN DISPLAY MODE. MADE ADJUSTMENTS AND RECALIBRATED.  
 INITIAL CALIBRATION USING mV OUT  
 LIMITED: MUST BE USED WITH DISPLAY P# 721514 CH 2. ECAL 4.041 ESCL 4770. IF NOT USED WITH ABOVE LISTED DISPLAY MAY BE USED IN mV OUTPUT MODE.  
 EXCITATION VOLTAGE: 10 VDC.  
 CALIBRATED AND THEN GENERATED NEW COEFFICIENTS.

**Notification Status:** Approved

**Area Point of Contact:** Gary D Roberts

**If evaluation assistance is desired, call Quality Engineering. If further information is desired regarding the out-of-tolerance condition, call the calibrator listed below.**

**Calibration Organization:** S&CL

**Calibration Remarks:** *(file attachments allowed)*

UNIT WAS OUT OF TOLERANCE. MADE ADJSUTMENTS AND RECALIBRATED.  
LIMITED: MUST BE USED WITH READOUT P# 721514 CH 2  
ECAL 4.0401 ESCL 4700

---

**Evaluation:**

**Evaluator:** Tommy E Rahl

**Evaluation:** *(file attachments allowed)*

All weighing activites performed with the 5k load cell were tension measurements at weights below 2000 lbs. Since the out-of- tolerances occurred in the weight range above 4000 lbs, the weight measurements are still valid and within acceptable tolerances.

**Selected Approver:** Dana K Morton

**NOTE:** This evaluation or a separate out-of-tolerance evaluation must be retained with organization quality assurance records. Quality requirements for this evaluation and quality record retention are based on MCP-2391 and PRD-101.

---

**Evaluation Approval:**

**Date:** 06/16/2004 **Approval:** Dana K Morton

**Approval Comments:** Agree with evaluator statement.

---

**Comments/Attachments**

Add any comments here and/or use the area for any attachment.

---

**Document History:**

Created by Stanley D Zohner on 06/16/2004 10:00:51 AM

Stanley D Zohner edited document 10:15 AM 06/16/2004

Notification sent: Stanley D Zohner on Wednesday, June 16, 2004.

Sent to user: Tommy E Rahl: 10:15 AM 06/16/2004

Tommy E Rahl edited document 11:01 AM 06/16/2004

Submitted for approval to Dana K Morton : Tommy E Rahl on Wednesday, June 16, 2004

Sent for approval to Dana K Morton: 11:02 AM 06/16/2004

Dana K Morton edited document 11:50 AM 06/16/2004

Evaluation Approval: Dana K Morton on Wednesday, June 16, 2004

---

**INL  
CALIBRATION INPUT DATA**

12/20/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 721516      Mfr: INTERFACE      Model: 5000 LB (D & mV)      Noun Name: LOAD CELL      Serial #: 144177A/1210AF

Calibration Date: 3/10/2005

**ACTION CODE****AS FOUND**

Next Cal Due Date: 3/10/2006	1	<input type="checkbox"/>	Acceptance Test	1	<input checked="" type="radio"/>	In Tolerance
Charge Level: 20	2	<input type="checkbox"/>	Special Test	2	<input type="radio"/>	Out of Tolerance >1x <2x
Repair/Adj/etc C.L: 0	3	<input type="checkbox"/>	Calibration to MFG Specs	3	<input type="radio"/>	Out of Tolerance >2x <3x
Material Amount: 0	4	<input type="checkbox"/>	Clean	4	<input type="radio"/>	Out of Tolerance >3x <5x
Charge Number: 100853GSA	5	<input checked="" type="checkbox"/>	Limited Calibration	5	<input type="radio"/>	Out of Tolerance >5x
Cal Work Inst ID: 5748F	6	<input type="checkbox"/>	Functional Check	6	<input type="radio"/>	Out of Tolerance-Undetermined
Outside Vendor:	7	<input type="checkbox"/>	Performance Check	7	<input type="checkbox"/>	Inoperative
	8	<input type="checkbox"/>	Modify	8	<input type="checkbox"/>	Damaged
	9	<input type="checkbox"/>	Repair-needs Charge Level	9	<input type="checkbox"/>	Not Used
	10	<input type="checkbox"/>	Other	10	<input type="checkbox"/>	Not Determined
				11	<input type="checkbox"/>	Excessed
				12	<input type="checkbox"/>	Extension

Calibrated By: Stan Zohner      S#: 58146      Phone: 526-2761

**CALIBRATION STANDARDS USED**

350816	350815	722028					

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-5.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

NOMINAL (STD)	OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION UNITS	AS FOUND (UUT)	MFG. ACCURACY

**COMMENTS**

LIMITED: MUST BE USED WITH DISPLAY P# 721514 CH 2. ECAL 4.041 ESCL 4700.  
 IF NOT USED WITH ABOVE LISTED DISPLAY MAY BE USED IN mV OUTPUT MODE.  
 EXCITATION VOLTAGE: 10 VDC.  
 TENSION COEFFICIENTS A: -0.4066477956 B: 116.5617946 C: -0.001512210655  
 COMPRESSION COEFFICIENTS A: -0.7305250315 B: -116.6423674 C: -0.002982122665



# INL CALIBRATION INPUT DATA

12/20/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 721516 Mfr: INTERFACE Model: 5000 LB (D & mV) Noun Name: LOAD CELL Serial  
 #: 144177A/1210AF

Calibration Date: 8/28/2006

## ACTION CODE

## AS FOUND

Next Cal Due Date: 8/28/2007

Charge Level: 20

Repair/Adj/etc C.L: 0

Material Amount: 0

Charge Number: 100853GSA

Cal Work Inst ID: 5748H

Outside Vendor:

- |    |                                     |                           |    |                                  |                               |
|----|-------------------------------------|---------------------------|----|----------------------------------|-------------------------------|
| 1  | <input type="checkbox"/>            | Acceptance Test           | 1  | <input checked="" type="radio"/> | In Tolerance                  |
| 2  | <input type="checkbox"/>            | Special Test              | 2  | <input type="radio"/>            | Out of Tolerance >1x <2x      |
| 3  | <input type="checkbox"/>            | Calibration to MFG Specs  | 3  | <input type="radio"/>            | Out of Tolerance >2x <3x      |
| 4  | <input type="checkbox"/>            | Clean                     | 4  | <input type="radio"/>            | Out of Tolerance >3x <5x      |
| 5  | <input checked="" type="checkbox"/> | Limited Calibration       | 5  | <input type="radio"/>            | Out of Tolerance >5x          |
| 6  | <input type="checkbox"/>            | Functional Check          | 6  | <input type="radio"/>            | Out of Tolerance-Undetermined |
| 7  | <input type="checkbox"/>            | Performance Check         | 7  | <input type="checkbox"/>         | Inoperative                   |
| 8  | <input type="checkbox"/>            | Modify                    | 8  | <input type="checkbox"/>         | Damaged                       |
| 9  | <input type="checkbox"/>            | Repair-needs Charge Level | 9  | <input type="checkbox"/>         | Not Used                      |
| 10 | <input type="checkbox"/>            | Other                     | 10 | <input type="checkbox"/>         | Not Determined                |
|    |                                     |                           | 11 | <input type="checkbox"/>         | Excessed                      |
|    |                                     |                           | 12 | <input type="checkbox"/>         | Extension                     |

Calibrated By: Stan Zohner S#: 58146 Phone: 526-2761

## CALIBRATION STANDARDS USED

350816	350815	723926					

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED  
 VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

## LABORATORY TEMPERATURE AND HUMIDITY

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

NOMINAL (STD)	OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION UNITS	AS FOUND (UUT)	MFG. ACCURACY

## COMMENTS

LIMITED: MUST BE USED WITH DISPLAY P# 721514 CH 2. ECAL 4.041 ESCL 4700.  
 IF NOT USED WITH ABOVE LISTED DISPLAY MAY BE USED IN mV OUTPUT MODE.  
 EXCITATION VOLTAGE: 10 VDC.  
 TENSION COEFFICIENTS A: 0.4066477956 B: 116.5617946 C: -0.001512210655  
 COMPRESSION COEFFICIENTS A: 0.7305250315 B: -116.6423674 C: -0.002982122665

# **TENSILE TEST MACHINE**



## **Calibration Issue of the Tinius-Olsen Compression Tension Test Machine - 267216**

Static load testing was performed using the Tinius-Olsen test machine on June 3, 2004. The calibration for that machine expired one day earlier on June 2, 2004. At the time it was thought that an extension to the calibration had been granted. That turned out to not be the case and because of that oversight, technically, the testing was done with a machine that had an expired calibration. The oversight was realized and the machine was calibrated on July 27, 2004, approximately 7-1/2 weeks past the calibration due date. Results of that calibration indicated that the machine was in tolerance with no adjustments necessary. Hence, it can be concluded that the machine was in tolerance on June 3, 2004.



### INL CALIBRATION INPUT DATA

11/30/2007

NAME: NATHAN DEWITT	BADGE: 901441	PH: 526-9118	AREA: CFA	BLDG: 602	RM:
---------------------	---------------	--------------	-----------	-----------	-----

ID Number: 267216      Mfr: TINIUS OLS      Model: 400K      Noun Name: COMPRESSION TENSION TEST  
MACHINE      Serial #: 152760

Calibration Date: 8/1/2005

**ACTION CODE**

**AS FOUND**

Next Cal Due Date: 9/27/2005

1 ☐ Acceptance Test

# 1 In Tolerance

Charge Level: 1

2 ☐ Special Test

2 ☐ Out of Tolerance  $>1x <2x$

Repair/Adj/etc C.L: 0

### 3 ☐ Calibration to MFG Specs

3 ☐ Out of Tolerance  $>2x < 3x$

Material Amount: 0

4 ☐ Clean

4 ☐ Out of Tolerance  $>3x <5x$

Charge Number: p51401008

5 ☐ Limited Calibration

5 C Out of Tolerance >5x

Cal Work Inst ID: 3547D

6 ☐ Functional Check

6 C Out of Tolerance-Undetermined

Outside Vendor:

7 ☐ Performance Check

7 ☐ Inoperative

8  **Modify**

8 ☐ Damaged

9 ☐ Repair-needs Charge Level

9 ☐ Not Used

10 ☒ Other

10 ☒ Not Determined

11 ☐ Excessed

Calibrated By: Larry Deming S#: 39571 Phone: 526-2761

## 12 Extension

### CALIBRATION STANDARDS USED


STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

### LABORATORY TEMPERATURE AND HUMIDITY

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

**Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.**

[illegible]

## COMMENTS

BASED ON THE PAST IN TOLERANCE CALIBRATIONS AND THE INHERENT STABILITY OF THIS MACHINE, THE CALIBRATION DUE DATE HAS BEEN EXTENDED 2 MONTHS. NEW DUE DATE: 9/27/2005

**INL  
CALIBRATION INPUT DATA**

11/30/2007

NAME: NATHAN DEWITT	BADGE: 901441	PH: 526-9118	AREA: CFA	BLDG: 602	RM:
---------------------	---------------	--------------	-----------	-----------	-----

ID Number: 267216      Mfr: TINIUS OLS      Model: 400K      Noun Name: COMPRESSION TENSION TEST

MACHINE      Serial #: 152760

Calibration Date: 9/15/2005

Next Cal Due Date: 9/15/2006

Charge Level: 16

Repair/Adj/etc C.L: 0

Material Amount: 0

Charge Number: P51401008

Cal Work Inst ID: 3547E

Outside Vendor:

**ACTION CODE**

**AS FOUND**

- |  |   |
|--|---|
| 1 <input type="checkbox"/> Acceptance Test                     | 1 <input checked="" type="radio"/> In Tolerance       |
| 2 <input type="checkbox"/> Special Test                        | 2 <input type="radio"/> Out of Tolerance >1x <2x      |
| 3 <input checked="" type="checkbox"/> Calibration to MFG Specs | 3 <input type="radio"/> Out of Tolerance >2x <3x      |
| 4 <input type="checkbox"/> Clean                               | 4 <input type="radio"/> Out of Tolerance >3x <5x      |
| 5 <input type="checkbox"/> Limited Calibration                 | 5 <input type="radio"/> Out of Tolerance >5x          |
| 6 <input type="checkbox"/> Functional Check                    | 6 <input type="radio"/> Out of Tolerance-Undetermined |
| 7 <input type="checkbox"/> Performance Check                   | 7 <input type="checkbox"/> Inoperative                |
| 8 <input type="checkbox"/> Modify                              | 8 <input type="checkbox"/> Damaged                    |
| 9 <input type="checkbox"/> Repair-needs Charge Level           | 9 <input type="checkbox"/> Not Used                   |
| 10 <input type="checkbox"/> Other                              | 10 <input type="checkbox"/> Not Determined            |
|  | 11 <input type="checkbox"/> Excessed                  |
|  | 12 <input type="checkbox"/> Extension                 |

Calibrated By: Stan Zohner      S#: 58146      Phone: 526-2761

**CALIBRATION STANDARDS USED**

350816	350815	701349	710406	281283			

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-5.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

**OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION**

NOMINAL (STD)	UNITS	AS FOUND (UUT)	MFG. ACCURACY

**COMMENTS**

# **INL CALIBRATION INPUT DATA**

11/30/2007

NAME: NATHAN DEWITT	BADGE: 901441	PH: 526-9118	AREA: CFA	BLDG: 602	RM:
---------------------	---------------	--------------	-----------	-----------	-----

ID Number: 267216      Mfr: TINUIUS OLS      Model: 400K      Noun Name: COMPRESSION TENSION TEST  
 MACHINE      Serial #: 152760

Calibration Date: 9/18/2006

**ACTION CODE****AS FOUND**

Next Cal Due Date: 9/18/2007

Charge Level: 16

Repair/Adj/etc C.L.: 0

Material Amount: 0

Charge Number: P51101ICS

Cal Work Inst ID: 3547F

Outside Vendor:

- |    |                                     |                           |    |                                  |                               |
|----|-------------------------------------|---------------------------|----|----------------------------------|-------------------------------|
| 1  | <input type="checkbox"/>            | Acceptance Test           | 1  | <input checked="" type="radio"/> | In Tolerance                  |
| 2  | <input type="checkbox"/>            | Special Test              | 2  | <input type="radio"/>            | Out of Tolerance >1x <2x      |
| 3  | <input checked="" type="checkbox"/> | Calibration to MFG Specs  | 3  | <input type="radio"/>            | Out of Tolerance >2x <3x      |
| 4  | <input type="checkbox"/>            | Clean                     | 4  | <input type="radio"/>            | Out of Tolerance >3x <5x      |
| 5  | <input type="checkbox"/>            | Limited Calibration       | 5  | <input type="radio"/>            | Out of Tolerance >5x          |
| 6  | <input type="checkbox"/>            | Functional Check          | 6  | <input type="radio"/>            | Out of Tolerance-Undetermined |
| 7  | <input type="checkbox"/>            | Performance Check         | 7  | <input type="checkbox"/>         | Inoperative                   |
| 8  | <input type="checkbox"/>            | Modify                    | 8  | <input type="checkbox"/>         | Damaged                       |
| 9  | <input type="checkbox"/>            | Repair-needs Charge Level | 9  | <input type="checkbox"/>         | Not Used                      |
| 10 | <input type="checkbox"/>            | Other                     | 10 | <input type="checkbox"/>         | Not Determined                |
|    |                                     |                           | 11 | <input type="checkbox"/>         | Excessed                      |
|    |                                     |                           | 12 | <input type="checkbox"/>         | Extension                     |

Calibrated By: Stan Zohmer      S#: 58146      Phone: 526-2761

**CALIBRATION STANDARDS USED**

350816	350815	701349	710406	281283			

**STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES**

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

**OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION****NOMINAL (STD)****UNITS****AS FOUND (UUT)****MFG. ACCURACY**


**COMMENTS**



**INL  
CALIBRATION INPUT DATA**

11/30/2007

NAME: NATHAN DEWITT	BADGE: 901441	PH: 526-9118	AREA: CFA	BLDG: 602	RM:
---------------------	---------------	--------------	-----------	-----------	-----

ID Number: 267216      Mfr: TINIUS OLS      Model: 400K      Noun Name: COMPRESSION TENSION TEST  
 MACHINE      Serial #: 152760

Calibration Date: 9/18/2007

**ACTION CODE**

**AS FOUND**

Next Cal Due Date: 9/18/2008	1	<input type="checkbox"/>	Acceptance Test	1	<input checked="" type="radio"/>	In Tolerance
Charge Level: 16	2	<input type="checkbox"/>	Special Test	2	<input type="radio"/>	Out of Tolerance >1x <2x
Repair/Adj/etc C.L: 0	3	<input checked="" type="checkbox"/>	Calibration to MFG Specs	3	<input type="radio"/>	Out of Tolerance >2x <3x
Material Amount: 0	4	<input type="checkbox"/>	Clean	4	<input type="radio"/>	Out of Tolerance >3x <5x
Charge Number: P51101SS9	5	<input type="checkbox"/>	Limited Calibration	5	<input type="radio"/>	Out of Tolerance >5x
Cal Work Inst ID: 3547J	6	<input type="checkbox"/>	Functional Check	6	<input type="radio"/>	Out of Tolerance-Undetermined
Outside Vendor:	7	<input type="checkbox"/>	Performance Check	7	<input type="checkbox"/>	Inoperative
	8	<input type="checkbox"/>	Modify	8	<input type="checkbox"/>	Damaged
	9	<input type="checkbox"/>	Repair-needs Charge Level	9	<input type="checkbox"/>	Not Used
	10	<input type="checkbox"/>	Other	10	<input type="checkbox"/>	Not Determined
				11	<input type="checkbox"/>	Excessed
Calibrated By: Stan Zohner      S#: 58146      Phone: 526-2761				12	<input type="checkbox"/>	Extension

**CALIBRATION STANDARDS USED**

350816	350815	701349	710406	281283			

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

NOMINAL (STD)	OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION UNITS	AS FOUND (UUT)	MFG. ACCURACY

COMMENTS

# **TEMPERATURE READOUTS**

**INL**  
**CALIBRATION INPUT DATA**

11/30/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 722806    Mfr: EXTECH    Model: EA15    Noun Name: TEMPERATURE READOUT    Serial #:  
040704878

Calibration Date: 5/11/2005

**ACTION CODE****AS FOUND**

Next Cal Due Date: 5/11/2006	1	<input type="checkbox"/>	Acceptance Test	1	<input checked="" type="radio"/>	In Tolerance
Charge Level: 4	2	<input type="checkbox"/>	Special Test	2	<input type="radio"/>	Out of Tolerance >1x <2x
Repair/Adj/etc C.L: 0	3	<input checked="" type="checkbox"/>	Calibration to MFG Specs	3	<input type="radio"/>	Out of Tolerance >2x <3x
Material Amount: 0	4	<input type="checkbox"/>	Clean	4	<input type="radio"/>	Out of Tolerance >3x <5x
Charge Number: 100853GSA	5	<input type="checkbox"/>	Limited Calibration	5	<input type="radio"/>	Out of Tolerance >5x
Cal Work Inst ID: 5113	6	<input type="checkbox"/>	Functional Check	6	<input type="radio"/>	Out of Tolerance-Undetermined
Outside Vendor:	7	<input type="checkbox"/>	Performance Check	7	<input type="checkbox"/>	Inoperative
	8	<input type="checkbox"/>	Modify	8	<input type="checkbox"/>	Damaged
	9	<input type="checkbox"/>	Repair-needs Charge Level	9	<input type="checkbox"/>	Not Used
	10	<input type="checkbox"/>	Other	10	<input type="checkbox"/>	Not Determined
				11	<input type="checkbox"/>	Excessed
Calibrated By: Brian Berls    S#: 102182    Phone: 526-2761				12	<input type="checkbox"/>	Extension

**CALIBRATION STANDARDS USED**

216578	701587	715356	700666	700392	700393	700394	700855
706129							

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-5.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

NOMINAL (STD)	OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION UNITS	AS FOUND (UUT)	MFG. ACCURACY

**COMMENTS**

Initial calibration.  
Receiving inspection test referral QA# 114093

**INL**  
**CALIBRATION INPUT DATA**

11/30/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 722806    Mfr: EXTECH    Model: EA15    Noun Name: TEMPERATURE READOUT    Serial #:

Calibration Date: 4/27/2006

**ACTION CODE****AS FOUND**

Next Cal Due Date: 4/27/2007

Charge Level: 4

Repair/Adj/etc C.L: 0

Material Amount: 0

Charge Number: 100853GSA

Cal Work Inst ID: 5113

Outside Vendor:

- |    |                                     |                           |    |                                  |                               |
|----|-------------------------------------|---------------------------|----|----------------------------------|-------------------------------|
| 1  | <input type="checkbox"/>            | Acceptance Test           | 1  | <input checked="" type="radio"/> | In Tolerance                  |
| 2  | <input type="checkbox"/>            | Special Test              | 2  | <input type="radio"/>            | Out of Tolerance >1x <2x      |
| 3  | <input checked="" type="checkbox"/> | Calibration to MFG Specs  | 3  | <input type="radio"/>            | Out of Tolerance >2x <3x      |
| 4  | <input type="checkbox"/>            | Clean                     | 4  | <input type="radio"/>            | Out of Tolerance >3x <5x      |
| 5  | <input type="checkbox"/>            | Limited Calibration       | 5  | <input type="radio"/>            | Out of Tolerance >5x          |
| 6  | <input type="checkbox"/>            | Functional Check          | 6  | <input type="radio"/>            | Out of Tolerance-Undetermined |
| 7  | <input type="checkbox"/>            | Performance Check         | 7  | <input type="checkbox"/>         | Inoperative                   |
| 8  | <input type="checkbox"/>            | Modify                    | 8  | <input type="checkbox"/>         | Damaged                       |
| 9  | <input type="checkbox"/>            | Repair-needs Charge Level | 9  | <input type="checkbox"/>         | Not Used                      |
| 10 | <input type="checkbox"/>            | Other                     | 10 | <input type="checkbox"/>         | Not Determined                |
|    |                                     |                           | 11 | <input type="checkbox"/>         | Excessed                      |
|    |                                     |                           | 12 | <input type="checkbox"/>         | Extension                     |

Calibrated By: Donnie Lindsay    S#: 105138    Phone: 526-2761

**CALIBRATION STANDARDS USED**

P020713	701587	700666	700392	700393	700394	700855	706129
715356							

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

**OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION****NOMINAL (STD)****UNITS****AS FOUND (UUT)****MFG. ACCURACY**


**COMMENTS**

S&amp;CL overcheck required.

**INL**  
**CALIBRATION INPUT DATA**

11/30/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 722806    Mfr: EXTECH    Model: EA15    Noun Name: TEMPERATURE READOUT    Serial #:

Calibration Date: 5/7/2007

**ACTION CODE****AS FOUND**

Next Cal Due Date: 5/7/2008	1	<input type="checkbox"/>	Acceptance Test	1	<input checked="" type="radio"/>	In Tolerance
Charge Level: 4	2	<input type="checkbox"/>	Special Test	2	<input type="radio"/>	Out of Tolerance >1x <2x
Repair/Adj/etc C.L: 1	3	<input checked="" type="checkbox"/>	Calibration to MFG Specs	3	<input type="radio"/>	Out of Tolerance >2x <3x
Material Amount: 0	4	<input type="checkbox"/>	Clean	4	<input type="radio"/>	Out of Tolerance >3x <5x
Charge Number: 100853GSA	5	<input type="checkbox"/>	Limited Calibration	5	<input type="radio"/>	Out of Tolerance >5x
Cal Work Inst ID: 5113	6	<input type="checkbox"/>	Functional Check	6	<input type="radio"/>	Out of Tolerance-Undetermined
Outside Vendor:	7	<input type="checkbox"/>	Performance Check	7	<input type="checkbox"/>	Inoperative
	8	<input type="checkbox"/>	Modify	8	<input type="checkbox"/>	Damaged
	9	<input type="checkbox"/>	Repair-needs Charge Level	9	<input type="checkbox"/>	Not Used
	10	<input type="checkbox"/>	Other	10	<input type="checkbox"/>	Not Determined
				11	<input type="checkbox"/>	Excessed
Calibrated By: Donnie Lindsay    S#: 105138    Phone: 526-2761				12	<input type="checkbox"/>	Extension

**CALIBRATION STANDARDS USED**

715356	701587	700666	700392	700393	700394	700856	700855
702971							

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

**OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION**

NOMINAL (STD)	UNITS	AS FOUND (UUT)	MFG. ACCURACY
1500 Deg F	Temp type K	1502.5 Deg F	+/- 2.2 Deg F
1900 Deg F	Temp type K	1903.1 Deg F	+/- 2.4 Deg F
1000 Deg C	Temp type K	1001.5 Deg C	+/- 1.2 Deg C

**COMMENTS**

UUT was out of tolerance on upper end of test range. Adjusted UUT to meet manufacturer's specifications and recalibrated unit.

Notification Status: Approved

**Area Point of Contact:** Gary D Roberts

**If evaluation assistance is desired, call Quality Engineering. If further information is desired regarding the out-of-tolerance condition, call the calibrator listed below.**

Cal Work Ins No.: 51130

Calibration Remarks: *(file attachments allowed)*

UUT was out of tolerance on upper end of test range. Adjusted UUT to meet manufacturer's specifications and recalibrated unit.

Evaluation Required?:

- ☒ REQUIRED  
☐ Not Required-New Equipment  
☐ Not Required-Other

---

**Evaluation:**

Evaluator: Dana K Morton

Evaluation: *(file attachments allowed)*

This test equipment was only used to measure temperatures well below 1000F. The out of tolerance values were for temperatures at 1500F or higher. Therefore, there is no adverse affect on any test data previously taken with this test equipment.

Selected Approver: Robert K Blandford

NOTE: This evaluation or a separate out-of-tolerance evaluation must be retained with organization quality assurance records. Quality requirements for this evaluation and quality record retention are based on PDD-13450 and LWP-13455.

---

**Evaluation Approval:**

Date: 05/10/2007 Approval: Robert K Blandford

Approval Comments: No Comment

---

**Comments/Attachments**

Add any comments here and/or use the area for any attachment.

---

**Document History:**

Created by Donovan M Lindsay on 05/07/2007 04:10:37 PM

Donovan M Lindsay edited document 04:14 PM 05/07/2007

Notification sent: Donovan M Lindsay on Monday, May 07, 2007.

Sent to user: Dana K Morton: 04:14 PM 05/07/2007

Dana K Morton edited document 11:13 AM 05/10/2007

Submitted for approval to Robert K Blandford : Dana K Morton on Thursday, May 10, 2007

Sent for approval to Robert K Blandford: 11:13 AM 05/10/2007

Evaluation Approval: Robert K Blandford on Thursday, May 10, 2007

---

**INL**  
**CALIBRATION INPUT DATA**

11/30/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 722807    Mfr: EXTECH    Model: EA15    Noun Name: TEMPERATURE READOUT    Serial #:

G037371

Calibration Date: 5/11/2005

**ACTION CODE****AS FOUND**

Next Cal Due Date: 5/11/2006	1	<input type="checkbox"/>	Acceptance Test	1	<input checked="" type="radio"/>	In Tolerance
Charge Level: 4	2	<input type="checkbox"/>	Special Test	2	<input type="radio"/>	Out of Tolerance >1x <2x
Repair/Adj/etc C.L: 0	3	<input checked="" type="checkbox"/>	Calibration to MFG Specs	3	<input type="radio"/>	Out of Tolerance >2x <3x
Material Amount: 0	4	<input type="checkbox"/>	Clean	4	<input type="radio"/>	Out of Tolerance >3x <5x
Charge Number: 100853GSA	5	<input type="checkbox"/>	Limited Calibration	5	<input type="radio"/>	Out of Tolerance >5x
Cal Work Inst ID: 5113	6	<input type="checkbox"/>	Functional Check	6	<input type="radio"/>	Out of Tolerance-Undetermined
Outside Vendor:	7	<input type="checkbox"/>	Performance Check	7	<input type="checkbox"/>	Inoperative
	8	<input type="checkbox"/>	Modify	8	<input type="checkbox"/>	Damaged
	9	<input type="checkbox"/>	Repair-needs Charge Level	9	<input type="checkbox"/>	Not Used
	10	<input type="checkbox"/>	Other	10	<input type="checkbox"/>	Not Determined
				11	<input type="checkbox"/>	Excessed
Calibrated By: Brian Berls				12	<input type="checkbox"/>	Extension

S#: 102182 Phone: 526-2761

**CALIBRATION STANDARDS USED**

216578	701587	700666	715356	700392	700393	700394	700855
706129							

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-5.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

**OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION****NOMINAL (STD)****UNITS****AS FOUND (UUT)****MFG. ACCURACY**


**COMMENTS**

Initial calibration.  
Receiving inspection test referral QA# 114093

11/30/2007



**INL**  
**CALIBRATION INPUT DATA**

11/30/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 722807    Mfr: EXTECH    Model: EA15    Noun Name: TEMPERATURE READOUT    Serial #: G037371

Calibration Date: 5/23/2005

**ACTION CODE****AS FOUND**

Next Cal Due Date: 5/23/2006	1	<input type="checkbox"/>	Acceptance Test	1	<input type="radio"/>	In Tolerance
Charge Level: 0	2	<input type="checkbox"/>	Special Test	2	<input type="radio"/>	Out of Tolerance >1x <2x
Repair/Adj/etc C.L: 2	3	<input checked="" type="checkbox"/>	Calibration to MFG Specs	3	<input checked="" type="radio"/>	Out of Tolerance >2x <3x
Material Amount: 0	4	<input type="checkbox"/>	Clean	4	<input type="radio"/>	Out of Tolerance >3x <5x
Charge Number: 100853GSA	5	<input type="checkbox"/>	Limited Calibration	5	<input type="radio"/>	Out of Tolerance >5x
Cal Work Inst ID: 5113	6	<input type="checkbox"/>	Functional Check	6	<input type="radio"/>	Out of Tolerance-Undetermined
Outside Vendor:	7	<input type="checkbox"/>	Performance Check	7	<input type="checkbox"/>	Inoperative
	8	<input type="checkbox"/>	Modify	8	<input type="checkbox"/>	Damaged
	9	<input type="checkbox"/>	Repair-needs Charge Level	9	<input type="checkbox"/>	Not Used
	10	<input type="checkbox"/>	Other	10	<input type="checkbox"/>	Not Determined
				11	<input type="checkbox"/>	Excessed
Calibrated By: Brian Berls	S#: 102182	Phone: 526-2761		12	<input type="checkbox"/>	Extension

**CALIBRATION STANDARDS USED**

216578	701587	700666	715356				

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

**OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION**

NOMINAL (STD)	UNITS	AS FOUND (UUT)	MFG. ACCURACY
Ch 1 input	-190 to 1900 F	+3.5 all points	0.05% +1.5

**COMMENTS**

Adjusted unit to meet manufacturers specifications and recalibrated unit.

**Temperature Readout**

**#722807**

**Not Used To Collect Data**

**Prior To**

**5/23/2005**

**INL**  
**CALIBRATION INPUT DATA**

11/30/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 722807    Mfr: EXTECH    Model: EA15    Noun Name: TEMPERATURE READOUT    Serial #:

G037371

Calibration Date: 4/27/2006

**ACTION CODE**

**AS FOUND**

Next Cal Due Date: 4/27/2007

Charge Level: 4

Repair/Adj/etc C.L: 0

Material Amount: 0

Charge Number: 100853GSA

Cal Work Inst ID: 5113

Outside Vendor:

- |    |                                     |                           |    |                                  |                               |
|----|-------------------------------------|---------------------------|----|----------------------------------|-------------------------------|
| 1  | <input type="checkbox"/>            | Acceptance Test           | 1  | <input checked="" type="radio"/> | In Tolerance                  |
| 2  | <input type="checkbox"/>            | Special Test              | 2  | <input type="radio"/>            | Out of Tolerance >1x <2x      |
| 3  | <input checked="" type="checkbox"/> | Calibration to MFG Specs  | 3  | <input type="radio"/>            | Out of Tolerance >2x <3x      |
| 4  | <input type="checkbox"/>            | Clean                     | 4  | <input type="radio"/>            | Out of Tolerance >3x <5x      |
| 5  | <input type="checkbox"/>            | Limited Calibration       | 5  | <input type="radio"/>            | Out of Tolerance >5x          |
| 6  | <input type="checkbox"/>            | Functional Check          | 6  | <input type="radio"/>            | Out of Tolerance-Undetermined |
| 7  | <input type="checkbox"/>            | Performance Check         | 7  | <input type="checkbox"/>         | Inoperative                   |
| 8  | <input type="checkbox"/>            | Modify                    | 8  | <input type="checkbox"/>         | Damaged                       |
| 9  | <input type="checkbox"/>            | Repair-needs Charge Level | 9  | <input type="checkbox"/>         | Not Used                      |
| 10 | <input type="checkbox"/>            | Other                     | 10 | <input type="checkbox"/>         | Not Determined                |
|    |                                     |                           | 11 | <input type="checkbox"/>         | Excessed                      |
|    |                                     |                           | 12 | <input type="checkbox"/>         | Extension                     |

Calibrated By: Donnie Lindsay    S#: 105138    Phone: 526-2761

**CALIBRATION STANDARDS USED**

715356	P020713	701587	700666	700392	700393	700394	700855
706129							

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

**OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION**

NOMINAL (STD)	UNITS	AS FOUND (UUT)	MFG. ACCURACY

**COMMENTS**

S&CL overcheck required.

INL  
CALIBRATION INPUT DATA

11/30/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 722807    Mfr: EXTECH    Model: EA15    Noun Name: TEMPERATURE READOUT    Serial #: G037371

Calibration Date: 5/7/2007

ACTION CODE

AS FOUND

Next Cal Due Date: 5/7/2008	1	<input type="checkbox"/>	Acceptance Test	1	<input checked="" type="radio"/>	In Tolerance
Charge Level: 4	2	<input type="checkbox"/>	Special Test	2	<input type="radio"/>	Out of Tolerance >1x <2x
Repair/Adj/etc C.L: 0	3	<input checked="" type="checkbox"/>	Calibration to MFG Specs	3	<input type="radio"/>	Out of Tolerance >2x <3x
Material Amount: 0	4	<input type="checkbox"/>	Clean	4	<input type="radio"/>	Out of Tolerance >3x <5x
Charge Number: 100853GSA	5	<input type="checkbox"/>	Limited Calibration	5	<input type="radio"/>	Out of Tolerance >5x
Cal Work Inst ID: 5113	6	<input type="checkbox"/>	Functional Check	6	<input type="radio"/>	Out of Tolerance-Undetermined
Outside Vendor:	7	<input type="checkbox"/>	Performance Check	7	<input type="checkbox"/>	Inoperative
	8	<input type="checkbox"/>	Modify	8	<input type="checkbox"/>	Damaged
	9	<input type="checkbox"/>	Repair-needs Charge Level	9	<input type="checkbox"/>	Not Used
	10	<input type="checkbox"/>	Other	10	<input type="checkbox"/>	Not Determined
				11	<input type="checkbox"/>	Excessed
Calibrated By: Donnie Lindsay    S#: 105138    Phone: 526-2761				12	<input type="checkbox"/>	Extension

CALIBRATION STANDARDS USED

715356	702971	700666	700392	700393	700394	700856	700855
701587							

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

LABORATORY TEMPERATURE AND HUMIDITY

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION

NOMINAL (STD)	UNITS	AS FOUND (UUT)	MFG. ACCURACY

COMMENTS

**INL  
CALIBRATION INPUT DATA**

11/30/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 723407    Mfr: EXTECH    Model: EA15    Noun Name: TEMPERATURE READOUT    Serial #: 40704910

			ACTION CODE		AS FOUND
Calibration Date: 11/28/2005	1	<input type="checkbox"/>	Acceptance Test	1	<input checked="" type="radio"/> In Tolerance
Next Cal Due Date: 11/28/2006	2	<input type="checkbox"/>	Special Test	2	<input type="radio"/> Out of Tolerance >1x <2x
Charge Level: 4	3	<input checked="" type="checkbox"/>	Calibration to MFG Specs	3	<input type="radio"/> Out of Tolerance >2x <3x
Repair/Adj/etc C.L: 0	4	<input type="checkbox"/>	Clean	4	<input type="radio"/> Out of Tolerance >3x <5x
Material Amount: 0	5	<input type="checkbox"/>	Limited Calibration	5	<input type="radio"/> Out of Tolerance >5x
Charge Number: 100853GSA	6	<input type="checkbox"/>	Functional Check	6	<input type="radio"/> Out of Tolerance-Undetermined
Cal Work Inst ID: 5113N	7	<input type="checkbox"/>	Performance Check	7	<input type="checkbox"/> Inoperative
Outside Vendor:	8	<input type="checkbox"/>	Modify	8	<input type="checkbox"/> Damaged
	9	<input type="checkbox"/>	Repair-needs Charge Level	9	<input type="checkbox"/> Not Used
	10	<input type="checkbox"/>	Other	10	<input type="checkbox"/> Not Determined
				11	<input type="checkbox"/> Excessed
				12	<input type="checkbox"/> Extension

Calibrated By: Chris Dudley    S#: 101142    Phone: 526-2761

**CALIBRATION STANDARDS USED**

714801	701587	714891	706129	700855	715356		

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

NOMINAL (STD)	OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION UNITS	AS FOUND (UUT)	MFG. ACCURACY

**COMMENTS**

INITIAL CALIBRATION  
RECEIVING INSPECTION TEST REFERRAL QA# 115845

**INL  
CALIBRATION INPUT DATA**

11/30/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 723407    Mfr: EXTECH    Model: EA15    Noun Name: TEMPERATURE READOUT    Serial #: 40704910

Calibration Date: 12/4/2006

**ACTION CODE****AS FOUND**

Next Cal Due Date: 12/4/2007	1	<input type="checkbox"/>	Acceptance Test	1	<input checked="" type="radio"/>	In Tolerance
Charge Level: 4	2	<input type="checkbox"/>	Special Test	2	<input type="radio"/>	Out of Tolerance >1x <2x
Repair/Adj/etc C.L.: 0	3	<input checked="" type="checkbox"/>	Calibration to MFG Specs	3	<input type="radio"/>	Out of Tolerance >2x <3x
Material Amount: 0	4	<input type="checkbox"/>	Clean	4	<input type="radio"/>	Out of Tolerance >3x <5x
Charge Number: 100853GSA	5	<input type="checkbox"/>	Limited Calibration	5	<input type="radio"/>	Out of Tolerance >5x
Cal Work Inst ID: 5113N	6	<input type="checkbox"/>	Functional Check	6	<input type="radio"/>	Out of Tolerance-Undetermined
Outside Vendor:	7	<input type="checkbox"/>	Performance Check	7	<input type="checkbox"/>	Inoperative
	8	<input type="checkbox"/>	Modify	8	<input type="checkbox"/>	Damaged
	9	<input type="checkbox"/>	Repair-needs Charge Level	9	<input type="checkbox"/>	Not Used
	10	<input type="checkbox"/>	Other	10	<input type="checkbox"/>	Not Determined
				11	<input type="checkbox"/>	Excessed
Calibrated By: Donnie Lindsay    S#: 105138    Phone: 526-2761				12	<input type="checkbox"/>	Extension

**CALIBRATION STANDARDS USED**

715356	701587	702971	700666	700392	700393	700394	700855
706129	385979						

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

**OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION**

NOMINAL (STD)	UNITS	AS FOUND (UUT)	MFG. ACCURACY

**COMMENTS**

S&amp;CL overcheck required.

**INL  
CALIBRATION INPUT DATA**

12/11/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 723407    Mfr: EXTECH    Model: EA15    Noun Name: TEMPERATURE READOUT    Serial #: 40704910

Calibration Date: 12/5/2007

**ACTION CODE****AS FOUND**

Next Cal Due Date: 12/5/2008	1	<input type="checkbox"/>	Acceptance Test	1	<input checked="" type="radio"/>	In Tolerance
Charge Level: 4	2	<input type="checkbox"/>	Special Test	2	<input type="radio"/>	Out of Tolerance >1x <2x
Repair/Adj/etc C.L: 0	3	<input checked="" type="checkbox"/>	Calibration to MFG Specs	3	<input type="radio"/>	Out of Tolerance >2x <3x
Material Amount: 0	4	<input type="checkbox"/>	Clean	4	<input type="radio"/>	Out of Tolerance >3x <5x
Charge Number: 100853GSA	5	<input type="checkbox"/>	Limited Calibration	5	<input type="radio"/>	Out of Tolerance >5x
Cal Work Inst ID: 5113T	6	<input type="checkbox"/>	Functional Check	6	<input type="radio"/>	Out of Tolerance-Undetermined
Outside Vendor:	7	<input type="checkbox"/>	Performance Check	7	<input type="checkbox"/>	Inoperative
	8	<input type="checkbox"/>	Modify	8	<input type="checkbox"/>	Damaged
	9	<input type="checkbox"/>	Repair-needs Charge Level	9	<input type="checkbox"/>	Not Used
	10	<input type="checkbox"/>	Other	10	<input type="checkbox"/>	Not Determined
				11	<input type="checkbox"/>	Excessed
Calibrated By: Donnie Lindsay    S#: 105138    Phone: 526-2761				12	<input type="checkbox"/>	Extension

**CALIBRATION STANDARDS USED**

714801	715356	304399	700855	700856			

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

NOMINAL (STD)	OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION UNITS	AS FOUND (UUT)	MFG. ACCURACY

**COMMENTS**

**INL**  
**CALIBRATION INPUT DATA**

11/30/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 723408    Mfr: EXTECH    Model: EA15    Noun Name: TEMPERATURE READOUT    Serial #:

Calibration Date: 11/28/2005

**ACTION CODE****AS FOUND**

Next Cal Due Date: 11/28/2006

Charge Level: 4

Repair/Adj/etc C.L: 0

Material Amount: 0

Charge Number: 100853GSA

Cal Work Inst ID: 5113N

Outside Vendor:

- |    |                                     |                           |    |                                  |                               |
|----|-------------------------------------|---------------------------|----|----------------------------------|-------------------------------|
| 1  | <input type="checkbox"/>            | Acceptance Test           | 1  | <input checked="" type="radio"/> | In Tolerance                  |
| 2  | <input type="checkbox"/>            | Special Test              | 2  | <input type="radio"/>            | Out of Tolerance >1x <2x      |
| 3  | <input checked="" type="checkbox"/> | Calibration to MFG Specs  | 3  | <input type="radio"/>            | Out of Tolerance >2x <3x      |
| 4  | <input type="checkbox"/>            | Clean                     | 4  | <input type="radio"/>            | Out of Tolerance >3x <5x      |
| 5  | <input type="checkbox"/>            | Limited Calibration       | 5  | <input type="radio"/>            | Out of Tolerance >5x          |
| 6  | <input type="checkbox"/>            | Functional Check          | 6  | <input type="radio"/>            | Out of Tolerance-Undetermined |
| 7  | <input type="checkbox"/>            | Performance Check         | 7  | <input type="checkbox"/>         | Inoperative                   |
| 8  | <input type="checkbox"/>            | Modify                    | 8  | <input type="checkbox"/>         | Damaged                       |
| 9  | <input type="checkbox"/>            | Repair-needs Charge Level | 9  | <input type="checkbox"/>         | Not Used                      |
| 10 | <input type="checkbox"/>            | Other                     | 10 | <input type="checkbox"/>         | Not Determined                |
|    |                                     |                           | 11 | <input type="checkbox"/>         | Excessed                      |
|    |                                     |                           | 12 | <input type="checkbox"/>         | Extension                     |

Calibrated By: Chris Dudley    S#: 101142    Phone: 526-2761

**CALIBRATION STANDARDS USED**

714801	701587	714891	706129	700855	715356		

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

**OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION**

NOMINAL (STD)	UNITS	AS FOUND (UUT)	MFG. ACCURACY

**COMMENTS**

INITIAL CALIBRATION  
RECEIVING INSPECTION TEST REFERRAL QA# 115845



**INL**  
**CALIBRATION INPUT DATA**

11/30/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 723408    Mfr: EXTECH    Model: EA15    Noun Name: TEMPERATURE READOUT    Serial #:

Calibration Date: 12/4/2006

Next Cal Due Date: 12/4/2007

Charge Level: 4

Repair/Adj/etc C.L: 0

Material Amount: 0

Charge Number: 100853GSA

Cal Work Inst ID: 5113N

Outside Vendor:

**ACTION CODE****AS FOUND**

- |    |                                     |                           |    |                                  |                               |
|----|-------------------------------------|---------------------------|----|----------------------------------|-------------------------------|
| 1  | <input type="checkbox"/>            | Acceptance Test           | 1  | <input checked="" type="radio"/> | In Tolerance                  |
| 2  | <input type="checkbox"/>            | Special Test              | 2  | <input type="radio"/>            | Out of Tolerance >1x <2x      |
| 3  | <input checked="" type="checkbox"/> | Calibration to MFG Specs  | 3  | <input type="radio"/>            | Out of Tolerance >2x <3x      |
| 4  | <input type="checkbox"/>            | Clean                     | 4  | <input type="radio"/>            | Out of Tolerance >3x <5x      |
| 5  | <input type="checkbox"/>            | Limited Calibration       | 5  | <input type="radio"/>            | Out of Tolerance >5x          |
| 6  | <input type="checkbox"/>            | Functional Check          | 6  | <input type="radio"/>            | Out of Tolerance-Undetermined |
| 7  | <input type="checkbox"/>            | Performance Check         | 7  | <input type="checkbox"/>         | Inoperative                   |
| 8  | <input type="checkbox"/>            | Modify                    | 8  | <input type="checkbox"/>         | Damaged                       |
| 9  | <input type="checkbox"/>            | Repair-needs Charge Level | 9  | <input type="checkbox"/>         | Not Used                      |
| 10 | <input type="checkbox"/>            | Other                     | 10 | <input type="checkbox"/>         | Not Determined                |
|    |                                     |                           | 11 | <input type="checkbox"/>         | Excessed                      |
|    |                                     |                           | 12 | <input type="checkbox"/>         | Extension                     |

Calibrated By: Donnie Lindsay    S#: 105138    Phone: 526-2761

**CALIBRATION STANDARDS USED**

715356	701587	702971	700666	700392	700393	700394	700855
385979	706129						

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

NOMINAL (STD)	OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION UNITS	AS FOUND (UUT)	MFG. ACCURACY

**COMMENTS**

S&amp;CL overcheck required.

**INL  
CALIBRATION INPUT DATA**

12/11/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 723408    Mfr: EXTECH    Model: EA15    Noun Name: TEMPERATURE READOUT    Serial #: 040704827

Calibration Date: 12/5/2007

**ACTION CODE****AS FOUND**

Next Cal Due Date: 12/5/2008	1	<input type="checkbox"/>	Acceptance Test	1	<input checked="" type="radio"/>	In Tolerance
Charge Level: 4	2	<input type="checkbox"/>	Special Test	2	<input type="radio"/>	Out of Tolerance >1x <2x
Repair/Adj/etc C.L: 0	3	<input checked="" type="checkbox"/>	Calibration to MFG Specs	3	<input type="radio"/>	Out of Tolerance >2x <3x
Material Amount: 0	4	<input type="checkbox"/>	Clean	4	<input type="radio"/>	Out of Tolerance >3x <5x
Charge Number: 100853GSA	5	<input type="checkbox"/>	Limited Calibration	5	<input type="radio"/>	Out of Tolerance >5x
Cal Work Inst ID: 5113T	6	<input type="checkbox"/>	Functional Check	6	<input type="radio"/>	Out of Tolerance-Undetermined
Outside Vendor:	7	<input type="checkbox"/>	Performance Check	7	<input type="checkbox"/>	Inoperative
	8	<input type="checkbox"/>	Modify	8	<input type="checkbox"/>	Damaged
	9	<input type="checkbox"/>	Repair-needs Charge Level	9	<input type="checkbox"/>	Not Used
	10	<input type="checkbox"/>	Other	10	<input type="checkbox"/>	Not Determined
				11	<input type="checkbox"/>	Excessed
Calibrated By: Donnie Lindsay    S#: 105138    Phone: 526-2761				12	<input type="checkbox"/>	Extension

**CALIBRATION STANDARDS USED**

714801	715356	304399	700855	700856			

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

**OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION**

NOMINAL (STD)	UNITS	AS FOUND (UUT)	MFG. ACCURACY

**COMMENTS**

**INL  
CALIBRATION INPUT DATA**

11/30/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 724265    Mfr: EXTECH    Model: EA15    Noun Name: TEMPERATURE READOUT    Serial #: 040406419

		ACTION CODE	AS FOUND
Calibration Date: 10/31/2006	1	<input type="checkbox"/> Acceptance Test	1 <input checked="" type="radio"/> In Tolerance
Next Cal Due Date: 10/31/2007	2	<input type="checkbox"/> Special Test	2 <input type="radio"/> Out of Tolerance >1x <2x
Charge Level: 4	3	<input checked="" type="checkbox"/> Calibration to MFG Specs	3 <input type="radio"/> Out of Tolerance >2x <3x
Repair/Adj/etc C.L: 0	4	<input type="checkbox"/> Clean	4 <input type="radio"/> Out of Tolerance >3x <5x
Material Amount: 0	5	<input type="checkbox"/> Limited Calibration	5 <input type="radio"/> Out of Tolerance >5x
Charge Number: 100853GSA	6	<input type="checkbox"/> Functional Check	6 <input type="radio"/> Out of Tolerance-Undetermined
Cal Work Inst ID: 5113N	7	<input type="checkbox"/> Performance Check	7 <input type="checkbox"/> Inoperative
Outside Vendor:	8	<input type="checkbox"/> Modify	8 <input type="checkbox"/> Damaged
	9	<input type="checkbox"/> Repair-needs Charge Level	9 <input type="checkbox"/> Not Used
	10	<input type="checkbox"/> Other	10 <input type="checkbox"/> Not Determined
			11 <input type="checkbox"/> Excessed
			12 <input type="checkbox"/> Extension

Calibrated By: Donnie Lindsay    S#: 105138    Phone: 526-2761

**CALIBRATION STANDARDS USED**

715356	700666	700392	700393	700394	706129	700855	702971
701587	385979						

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

NOMINAL (STD)	OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION		MFG. ACCURACY
	UNITS	AS FOUND (UUT)	

**COMMENTS**

S&CL overcheck required.  
INITIAL CALIBRATION  
TEST REFERRAL QA# 121542

**INL**  
**CALIBRATION INPUT DATA**

11/30/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 724265    Mfr: EXTECH    Model: EA15    Noun Name: TEMPERATURE READOUT    Serial #: 040406419

Calibration Date: 10/23/2007

**ACTION CODE****AS FOUND**

Next Cal Due Date: 10/23/2008

Charge Level: 4

Repair/Adj/etc C.L.: 0

Material Amount: 0

Charge Number: 100853GSA

Cal Work Inst ID: 5113S

Outside Vendor:

- |    |                                     |                           |    |                                  |                               |
|----|-------------------------------------|---------------------------|----|----------------------------------|-------------------------------|
| 1  | <input type="checkbox"/>            | Acceptance Test           | 1  | <input checked="" type="radio"/> | In Tolerance                  |
| 2  | <input type="checkbox"/>            | Special Test              | 2  | <input type="radio"/>            | Out of Tolerance >1x <2x      |
| 3  | <input checked="" type="checkbox"/> | Calibration to MFG Specs  | 3  | <input type="radio"/>            | Out of Tolerance >2x <3x      |
| 4  | <input type="checkbox"/>            | Clean                     | 4  | <input type="radio"/>            | Out of Tolerance >3x <5x      |
| 5  | <input type="checkbox"/>            | Limited Calibration       | 5  | <input type="radio"/>            | Out of Tolerance >5x          |
| 6  | <input type="checkbox"/>            | Functional Check          | 6  | <input type="radio"/>            | Out of Tolerance-Undetermined |
| 7  | <input type="checkbox"/>            | Performance Check         | 7  | <input type="checkbox"/>         | Inoperative                   |
| 8  | <input type="checkbox"/>            | Modify                    | 8  | <input type="checkbox"/>         | Damaged                       |
| 9  | <input type="checkbox"/>            | Repair-needs Charge Level | 9  | <input type="checkbox"/>         | Not Used                      |
| 10 | <input type="checkbox"/>            | Other                     | 10 | <input type="checkbox"/>         | Not Determined                |
|    |                                     |                           | 11 | <input type="checkbox"/>         | Excessed                      |
|    |                                     |                           | 12 | <input type="checkbox"/>         | Extension                     |

Calibrated By: Donnie Lindsay    S#: 105138    Phone: 526-2761

**CALIBRATION STANDARDS USED**

715356	701587	714891	700666	700392	700393	700394	700855
700856							

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

NOMINAL (STD)	OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION UNITS	AS FOUND (UUT)	MFG. ACCURACY

**COMMENTS**

# INL CALIBRATION INPUT DATA

11/30/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 724266    Mfr: EXTECH    Model: EA15    Noun Name: TEMPERATURE READOUT    Serial #: 050607560

			ACTION CODE		AS FOUND
Calibration Date: 10/31/2006					
Next Cal Due Date: 10/31/2007	1	<input type="checkbox"/>	Acceptance Test	1	<input checked="" type="radio"/> In Tolerance
Charge Level: 4	2	<input type="checkbox"/>	Special Test	2	<input type="radio"/> Out of Tolerance >1x <2x
Repair/Adj/etc C.L: 0	3	<input checked="" type="checkbox"/>	Calibration to MFG Specs	3	<input type="radio"/> Out of Tolerance >2x <3x
Material Amount: 0	4	<input type="checkbox"/>	Clean	4	<input type="radio"/> Out of Tolerance >3x <5x
Charge Number: 100853GSA	5	<input type="checkbox"/>	Limited Calibration	5	<input type="radio"/> Out of Tolerance >5x
Cal Work Inst ID: 5113N	6	<input type="checkbox"/>	Functional Check	6	<input type="radio"/> Out of Tolerance-Undetermined
Outside Vendor:	7	<input type="checkbox"/>	Performance Check	7	<input type="checkbox"/> Inoperative
	8	<input type="checkbox"/>	Modify	8	<input type="checkbox"/> Damaged
	9	<input type="checkbox"/>	Repair-needs Charge Level	9	<input type="checkbox"/> Not Used
	10	<input type="checkbox"/>	Other	10	<input type="checkbox"/> Not Determined
				11	<input type="checkbox"/> Excessed
Calibrated By: Donnie Lindsay    S#: 105138    Phone: 526-2761				12	<input type="checkbox"/> Extension

## CALIBRATION STANDARDS USED

715356	700666	700392	700393	700394	706129	700855	702971
701587	385979						

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

## LABORATORY TEMPERATURE AND HUMIDITY

Physical STD (106C)    20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)    23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)    20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)    23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)    20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:    23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

## OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION

NOMINAL (STD)	UNITS	AS FOUND (UUT)	MFG. ACCURACY

## COMMENTS

S&CL overcheck required.  
INITIAL CALIBRATION  
TEST REFERRAL QA# 121542

**INL**  
**CALIBRATION INPUT DATA**

11/30/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 724266    Mfr: EXTECH    Model: EA15    Noun Name: TEMPERATURE READOUT    Serial #:

Calibration Date: 2/28/2007

Next Cal Due Date: 2/28/2008

Charge Level: 4

Repair/Adj/etc C.L: 0.5

Material Amount: 0

Charge Number: 100853GSA

Cal Work Inst ID: 5113N

Outside Vendor:

**ACTION CODE****AS FOUND**

- |    |                                     |                           |    |                                  |                               |
|----|-------------------------------------|---------------------------|----|----------------------------------|-------------------------------|
| 1  | <input type="checkbox"/>            | Acceptance Test           | 1  | <input type="radio"/>            | In Tolerance                  |
| 2  | <input type="checkbox"/>            | Special Test              | 2  | <input type="radio"/>            | Out of Tolerance >1x <2x      |
| 3  | <input checked="" type="checkbox"/> | Calibration to MFG Specs  | 3  | <input type="radio"/>            | Out of Tolerance >2x <3x      |
| 4  | <input type="checkbox"/>            | Clean                     | 4  | <input type="radio"/>            | Out of Tolerance >3x <5x      |
| 5  | <input type="checkbox"/>            | Limited Calibration       | 5  | <input type="radio"/>            | Out of Tolerance >5x          |
| 6  | <input type="checkbox"/>            | Functional Check          | 6  | <input checked="" type="radio"/> | Out of Tolerance-Undetermined |
| 7  | <input type="checkbox"/>            | Performance Check         | 7  | <input type="checkbox"/>         | Inoperative                   |
| 8  | <input type="checkbox"/>            | Modify                    | 8  | <input type="checkbox"/>         | Damaged                       |
| 9  | <input type="checkbox"/>            | Repair-needs Charge Level | 9  | <input type="checkbox"/>         | Not Used                      |
| 10 | <input type="checkbox"/>            | Other                     | 10 | <input type="checkbox"/>         | Not Determined                |
|    |                                     |                           | 11 | <input type="checkbox"/>         | Excessed                      |
|    |                                     |                           | 12 | <input type="checkbox"/>         | Extension                     |

Calibrated By: Donnie Lindsay    S#: 105138    Phone: 526-2761

**CALIBRATION STANDARDS USED**

715356	700666	700392	700393	700394	700855	706129	702971
701587							

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-5, -3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

**OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION**

NOMINAL (STD)	UNITS	AS FOUND (UUT)	MFG. ACCURACY
-100.0 deg F	Type K	-290.4 deg F	+/- 1.5 deg F
-50.0 deg F	Type K	-157.0 deg F	+/- 1.5 deg F
0.0 deg F	Type K	-51.0 deg F	+/- 1.5 deg F
100.0 deg F	Type K	139.2 deg F	+/- 1.5 deg F
500.0 deg F	Type K	880.7 deg F	+/- 1.7 deg F
1000 deg F	Type K	1839.5 deg F	+/- 2.0 deg F

**COMMENTS**

Unit came in with a comment stating that it was not measuring properly. Found unit to be out of tolerance at all test points, on both channels, and checking all thermocouple types. Adjusted unit using manufacture's procedure to meet the manufacture's specifications, and re-calibrated the unit.  
S&CL overcheck required.

**Temperature Readout**

**#724266**

**Not Used To Collect Data**

**Prior To**

**2/28/2007**

**INL**  
**CALIBRATION INPUT DATA**

12/11/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 724266    Mfr: EXTECH    Model: EA15    Noun Name: TEMPERATURE READOUT    Serial #:

Calibration Date: 12/5/2007

Next Cal Due Date: 12/5/2008

Charge Level: 4

Repair/Adj/etc C.L.: 0

Material Amount: 0

Charge Number: 100853GSA

Cal Work Inst ID: 5113T

Outside Vendor:

**ACTION CODE****AS FOUND**

- |    |                                     |                           |    |                                  |                               |
|----|-------------------------------------|---------------------------|----|----------------------------------|-------------------------------|
| 1  | <input type="checkbox"/>            | Acceptance Test           | 1  | <input checked="" type="radio"/> | In Tolerance                  |
| 2  | <input type="checkbox"/>            | Special Test              | 2  | <input type="radio"/>            | Out of Tolerance >1x <2x      |
| 3  | <input checked="" type="checkbox"/> | Calibration to MFG Specs  | 3  | <input type="radio"/>            | Out of Tolerance >2x <3x      |
| 4  | <input type="checkbox"/>            | Clean                     | 4  | <input type="radio"/>            | Out of Tolerance >3x <5x      |
| 5  | <input type="checkbox"/>            | Limited Calibration       | 5  | <input type="radio"/>            | Out of Tolerance >5x          |
| 6  | <input type="checkbox"/>            | Functional Check          | 6  | <input type="radio"/>            | Out of Tolerance-Undetermined |
| 7  | <input type="checkbox"/>            | Performance Check         | 7  | <input type="checkbox"/>         | Inoperative                   |
| 8  | <input type="checkbox"/>            | Modify                    | 8  | <input type="checkbox"/>         | Damaged                       |
| 9  | <input type="checkbox"/>            | Repair-needs Charge Level | 9  | <input type="checkbox"/>         | Not Used                      |
| 10 | <input type="checkbox"/>            | Other                     | 10 | <input type="checkbox"/>         | Not Determined                |
|    |                                     |                           | 11 | <input type="checkbox"/>         | Excessed                      |
|    |                                     |                           | 12 | <input type="checkbox"/>         | Extension                     |

Calibrated By: Donnie Lindsay    S#: 105138    Phone: 526-2761

**CALIBRATION STANDARDS USED**

714801	715356	304399	700856	700855			

**STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES**

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

**OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION**

NOMINAL (STD)	UNITS	AS FOUND (UUT)	MFG. ACCURACY

**COMMENTS**



# **THERMOCOUPLES**

**INL**  
**CALIBRATION INPUT DATA**

11/30/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 722808      Mfr: EXTECH      Model: TYPE K      Noun Name: THERMOCOUPLE      Serial #:

Calibration Date: 5/5/2005

**ACTION CODE**

**AS FOUND**

Next Cal Due Date: 5/5/2006	1	<input type="checkbox"/>	Acceptance Test	1	<input checked="" type="radio"/>	In Tolerance
Charge Level: 4	2	<input type="checkbox"/>	Special Test	2	<input type="radio"/>	Out of Tolerance >1x <2x
Repair/Adj/etc C.L: 0	3	<input checked="" type="checkbox"/>	Calibration to MFG Specs	3	<input type="radio"/>	Out of Tolerance >2x <3x
Material Amount: 0	4	<input type="checkbox"/>	Clean	4	<input type="radio"/>	Out of Tolerance >3x <5x
Charge Number: 100853GSA	5	<input type="checkbox"/>	Limited Calibration	5	<input type="radio"/>	Out of Tolerance >5x
Cal Work Inst ID: 3052	6	<input type="checkbox"/>	Functional Check	6	<input type="radio"/>	Out of Tolerance-Undetermined
Outside Vendor:	7	<input type="checkbox"/>	Performance Check	7	<input type="checkbox"/>	Inoperative
	8	<input type="checkbox"/>	Modify	8	<input type="checkbox"/>	Damaged
	9	<input type="checkbox"/>	Repair-needs Charge Level	9	<input type="checkbox"/>	Not Used
	10	<input type="checkbox"/>	Other	10	<input type="checkbox"/>	Not Determined
				11	<input type="checkbox"/>	Excessed
Calibrated By: Chris Dudley				12	<input type="checkbox"/>	Extension

**CALIBRATION STANDARDS USED**

715424	717617	714272	714273				

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

**OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION**

NOMINAL (STD)

UNITS

AS FOUND (UUT)

MFG. ACCURACY


**COMMENTS**

INITIAL CALIBRATION TEST REFERRAL qa# 114094  
VERIFIED TO 220 DEG C.

**INL  
CALIBRATION INPUT DATA**

11/30/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 722808      Mfr: EXTECH      Model: TYPE K      Noun Name: THERMOCOUPLE      Serial #:

Calibration Date: 4/27/2006

**ACTION CODE****AS FOUND**

Next Cal Due Date: 4/27/2007	1	<input type="checkbox"/>	Acceptance Test	1	<input checked="" type="radio"/>	In Tolerance
Charge Level: 4	2	<input type="checkbox"/>	Special Test	2	<input type="radio"/>	Out of Tolerance >1x <2x
Repair/Adj/etc C.L: 0	3	<input checked="" type="checkbox"/>	Calibration to MFG Specs	3	<input type="radio"/>	Out of Tolerance >2x <3x
Material Amount: 0	4	<input type="checkbox"/>	Clean	4	<input type="radio"/>	Out of Tolerance >3x <5x
Charge Number: 100853GSA	5	<input type="checkbox"/>	Limited Calibration	5	<input type="radio"/>	Out of Tolerance >5x
Cal Work Inst ID: 3052V	6	<input type="checkbox"/>	Functional Check	6	<input type="radio"/>	Out of Tolerance-Undetermined
Outside Vendor:	7	<input type="checkbox"/>	Performance Check	7	<input type="checkbox"/>	Inoperative
	8	<input type="checkbox"/>	Modify	8	<input type="checkbox"/>	Damaged
	9	<input type="checkbox"/>	Repair-needs Charge Level	9	<input type="checkbox"/>	Not Used
	10	<input type="checkbox"/>	Other	10	<input type="checkbox"/>	Not Determined
				11	<input type="checkbox"/>	Excessed
Calibrated By: Brian Berls	S#:	102182	Phone: 526-2761	12	<input type="checkbox"/>	Extension

**CALIBRATION STANDARDS USED**

714272	714273	717615	715424				

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

NOMINAL (STD)	OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION UNITS	AS FOUND (UUT)	MFG. ACCURACY

**COMMENTS**

VERIFIED TO 220 DEG C.

# **INL CALIBRATION INPUT DATA**

11/30/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 722808      Mfr: EXTECH      Model: TYPE K      Noun Name: THERMOCOUPLE      Serial #:

Calibration Date: 5/8/2007

**ACTION CODE****AS FOUND**

Next Cal Due Date: 5/8/2008	1	<input type="checkbox"/>	Acceptance Test	1	<input checked="" type="radio"/>	In Tolerance
Charge Level: 4	2	<input type="checkbox"/>	Special Test	2	<input type="radio"/>	Out of Tolerance >1x <2x
Repair/Adj/etc C.L: 0	3	<input checked="" type="checkbox"/>	Calibration to MFG Specs	3	<input type="radio"/>	Out of Tolerance >2x <3x
Material Amount: 0	4	<input type="checkbox"/>	Clean	4	<input type="radio"/>	Out of Tolerance >3x <5x
Charge Number: 100853GSA	5	<input type="checkbox"/>	Limited Calibration	5	<input type="radio"/>	Out of Tolerance >5x
Cal Work Inst ID: 3052V	6	<input type="checkbox"/>	Functional Check	6	<input type="radio"/>	Out of Tolerance-Undetermined
Outside Vendor:	7	<input type="checkbox"/>	Performance Check	7	<input type="checkbox"/>	Inoperative
	8	<input type="checkbox"/>	Modify	8	<input type="checkbox"/>	Damaged
	9	<input type="checkbox"/>	Repair-needs Charge Level	9	<input type="checkbox"/>	Not Used
	10	<input type="checkbox"/>	Other	10	<input type="checkbox"/>	Not Determined
				11	<input type="checkbox"/>	Excessed
Calibrated By: Brian Berls      S#: 102182      Phone: 526-2761				12	<input type="checkbox"/>	Extension

**CALIBRATION STANDARDS USED**

715356	715424	712788	714272				

**STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES**

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-5.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

NOMINAL (STD)	OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION UNITS	AS FOUND (UUT)	MFG. ACCURACY

**COMMENTS**

**INL**  
**CALIBRATION INPUT DATA**

11/30/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 722809      Mfr: EXTECH      Model: TYPE K      Noun Name: THERMOCOUPLE      Serial #:

Calibration Date: 5/5/2005

**ACTION CODE**

**AS FOUND**

Next Cal Due Date: 5/5/2006	1	<input type="checkbox"/>	Acceptance Test	1	<input checked="" type="radio"/> In Tolerance
Charge Level: 4	2	<input type="checkbox"/>	Special Test	2	<input type="radio"/> Out of Tolerance >1x <2x
Repair/Adj/etc C.L: 0	3	<input checked="" type="checkbox"/>	Calibration to MFG Specs	3	<input type="radio"/> Out of Tolerance >2x <3x
Material Amount: 0	4	<input type="checkbox"/>	Clean	4	<input type="radio"/> Out of Tolerance >3x <5x
Charge Number: 100853GSA	5	<input type="checkbox"/>	Limited Calibration	5	<input type="radio"/> Out of Tolerance >5x
Cal Work Inst ID: 3052	6	<input type="checkbox"/>	Functional Check	6	<input type="radio"/> Out of Tolerance-Undetermined
Outside Vendor:	7	<input type="checkbox"/>	Performance Check	7	<input type="checkbox"/> Inoperative
	8	<input type="checkbox"/>	Modify	8	<input type="checkbox"/> Damaged
	9	<input type="checkbox"/>	Repair-needs Charge Level	9	<input type="checkbox"/> Not Used
	10	<input type="checkbox"/>	Other	10	<input type="checkbox"/> Not Determined
				11	<input type="checkbox"/> Excessed
Calibrated By: Chris Dudley      S#: 101142      Phone: 526-2761				12	<input type="checkbox"/> Extension

**CALIBRATION STANDARDS USED**

715424	714272	714273	717617				

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-5.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

**OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION**

NOMINAL (STD)	UNITS	AS FOUND (UUT)	MFG. ACCURACY
---------------	-------	----------------	---------------


**COMMENTS**

INITIAL CALIBRATION TEST REFERRAL qa# 114094  
VERIFIED TO 220 DEG C.

**INL  
CALIBRATION INPUT DATA**

11/30/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 722809      Mfr: EXTECH      Model: TYPE K      Noun Name: THERMOCOUPLE      Serial #:

Calibration Date: 4/27/2006

**ACTION CODE**

**AS FOUND**

Next Cal Due Date: 4/27/2007	1	<input type="checkbox"/>	Acceptance Test	1	<input checked="" type="radio"/>	In Tolerance
Charge Level: 4	2	<input type="checkbox"/>	Special Test	2	<input type="radio"/>	Out of Tolerance >1x <2x
Repair/Adj/etc C.L.: 0	3	<input checked="" type="checkbox"/>	Calibration to MFG Specs	3	<input type="radio"/>	Out of Tolerance >2x <3x
Material Amount: 0	4	<input type="checkbox"/>	Clean	4	<input type="radio"/>	Out of Tolerance >3x <5x
Charge Number: 100853GSA	5	<input type="checkbox"/>	Limited Calibration	5	<input type="radio"/>	Out of Tolerance >5x
Cal Work Inst ID: 3052V	6	<input type="checkbox"/>	Functional Check	6	<input type="radio"/>	Out of Tolerance-Undetermined
Outside Vendor:	7	<input type="checkbox"/>	Performance Check	7	<input type="checkbox"/>	Inoperative
	8	<input type="checkbox"/>	Modify	8	<input type="checkbox"/>	Damaged
	9	<input type="checkbox"/>	Repair-needs Charge Level	9	<input type="checkbox"/>	Not Used
	10	<input type="checkbox"/>	Other	10	<input type="checkbox"/>	Not Determined
				11	<input type="checkbox"/>	Excessed
Calibrated By: Brian Berls				12	<input type="checkbox"/>	Extension
S#: 102182			Phone: 526-2761			

**CALIBRATION STANDARDS USED**

714272	714273	717615	715424				

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

NOMINAL (STD)	OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION UNITS	AS FOUND (UUT)	MFG. ACCURACY

**COMMENTS**

VERIFIED TO 220 DEG C.

**INL**  
**CALIBRATION INPUT DATA**

11/30/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 722809    Mfr: EXTECH    Model: TYPE K    Noun Name: THERMOCOUPLE    Serial #:

Calibration Date: 5/8/2007

**ACTION CODE****AS FOUND**

Next Cal Due Date: 5/8/2008	1	<input type="checkbox"/>	Acceptance Test	1	<input checked="" type="radio"/>	In Tolerance
Charge Level: 4	2	<input type="checkbox"/>	Special Test	2	<input type="radio"/>	Out of Tolerance >1x <2x
Repair/Adj/etc C.L: 0	3	<input checked="" type="checkbox"/>	Calibration to MFG Specs	3	<input type="radio"/>	Out of Tolerance >2x <3x
Material Amount: 0	4	<input type="checkbox"/>	Clean	4	<input type="radio"/>	Out of Tolerance >3x <5x
Charge Number: 100853GSA	5	<input type="checkbox"/>	Limited Calibration	5	<input type="radio"/>	Out of Tolerance >5x
Cal Work Inst ID: 3052V	6	<input type="checkbox"/>	Functional Check	6	<input type="radio"/>	Out of Tolerance-Undetermined
Outside Vendor:	7	<input type="checkbox"/>	Performance Check	7	<input type="checkbox"/>	Inoperative
	8	<input type="checkbox"/>	Modify	8	<input type="checkbox"/>	Damaged
	9	<input type="checkbox"/>	Repair-needs Charge Level	9	<input type="checkbox"/>	Not Used
	10	<input type="checkbox"/>	Other	10	<input type="checkbox"/>	Not Determined
				11	<input type="checkbox"/>	Excessed
Calibrated By: Brian Berls    S#: 102182    Phone: 526-2761				12	<input type="checkbox"/>	Extension

**CALIBRATION STANDARDS USED**

715356	715424	712788	714272				

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
PhysDim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

NOMINAL (STD)	OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION		MFG. ACCURACY
	UNITS	AS FOUND (UUT)	

COMMENTS

**INL  
CALIBRATION INPUT DATA**

11/30/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 722810      Mfr: EXTECH      Model: TYPE K      Noun Name: THERMOCOUPLE      Serial #:

Calibration Date: 5/5/2005

**ACTION CODE****AS FOUND**

Next Cal Due Date: 5/5/2006

Charge Level: 4

Repair/Adj/etc C.L: 0

Material Amount: 0

Charge Number: 100853GSA

Cal Work Inst ID: 3052

Outside Vendor:

1 ☐

Acceptance Test

2 ☐

Special Test

3 ☒

Calibration to MFG Specs

4 ☐

Clean

5 ☐

Limited Calibration

6 ☐

Functional Check

7 ☐

Performance Check

8 ☐

Modify

9 ☐

Repair-needs Charge Level

10 ☐

Other

1 ☒ In Tolerance2 ☐ Out of Tolerance >1x <2x3 ☐ Out of Tolerance >2x <3x4 ☐ Out of Tolerance >3x <5x5 ☐ Out of Tolerance >5x6 ☐ Out of Tolerance-Undetermined7 ☐ Inoperative8 ☐ Damaged9 ☐ Not Used10 ☐ Not Determined11 ☐ Excessed12 ☐ Extension

Calibrated By: Chris Dudley

S#: 101142 Phone: 526-2761

**CALIBRATION STANDARDS USED**

715424	714272	714273	717617				

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)

20.0 ° +/-0.3 ° C (40-55% RH)

Electronic STD (106D)

23.0 ° +/-0.5 ° C (30-45% RH)

Dimensional STD (106B)

20.0 ° +/-0.25 ° C (30-45% RH)

Electronic CAL (Lab 112)

23.0 ° +/-1.0 ° C (20-50% RH)

Phys/Dim CAL (Lab 111)

20.0 ° +/-0.5 ° C (20-50% RH)

Remaining S&amp;CL calibration areas:

23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

**OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION**

NOMINAL (STD)

UNITS

AS FOUND (UUT)

MFG. ACCURACY


**COMMENTS**

INITIAL CALIBRATION TEST REFERRAL qa# 114094  
VERIFIED TO 220 DEG C.



**INL**  
**CALIBRATION INPUT DATA**

11/30/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 722810      Mfr: EXTECH      Model: TYPE K      Noun Name: THERMOCOUPLE      Serial #:

Calibration Date: 4/27/2006

**ACTION CODE****AS FOUND**

Next Cal Due Date: 4/27/2007	1	<input type="checkbox"/>	Acceptance Test	1	<input checked="" type="radio"/>	In Tolerance
Charge Level: 4	2	<input type="checkbox"/>	Special Test	2	<input type="radio"/>	Out of Tolerance >1x <2x
Repair/Adj/etc C.L.: 0	3	<input checked="" type="checkbox"/>	Calibration to MFG Specs	3	<input type="radio"/>	Out of Tolerance >2x <3x
Material Amount: 0	4	<input type="checkbox"/>	Clean	4	<input type="radio"/>	Out of Tolerance >3x <5x
Charge Number: 100853GSA	5	<input type="checkbox"/>	Limited Calibration	5	<input type="radio"/>	Out of Tolerance >5x
Cal Work Inst ID: 3052V	6	<input type="checkbox"/>	Functional Check	6	<input type="radio"/>	Out of Tolerance-Undetermined
Outside Vendor:	7	<input type="checkbox"/>	Performance Check	7	<input type="checkbox"/>	Inoperative
	8	<input type="checkbox"/>	Modify	8	<input type="checkbox"/>	Damaged
	9	<input type="checkbox"/>	Repair-needs Charge Level	9	<input type="checkbox"/>	Not Used
	10	<input type="checkbox"/>	Other	10	<input type="checkbox"/>	Not Determined
				11	<input type="checkbox"/>	Excessed
Calibrated By: Brian Berls				12	<input type="checkbox"/>	Extension

**CALIBRATION STANDARDS USED**

714272	714273	717615	715424				

**STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES**

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

**Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.**

NOMINAL (STD)	OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION UNITS	AS FOUND (UUT)	MFG. ACCURACY

**COMMENTS**

VERIFIED TO 220 DEG C.

**INL**  
**CALIBRATION INPUT DATA**

11/30/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 722810      Mfr: EXTECH      Model: TYPE K      Noun Name: THERMOCOUPLE      Serial #:

Calibration Date: 5/8/2007

**ACTION CODE****AS FOUND**

Next Cal Due Date: 5/8/2008

Charge Level: 4

Repair/Adj/etc C.L.: 0

Material Amount: 0

Charge Number: 100853GSA

Cal Work Inst ID: 3052V

Outside Vendor:

1 ☐2 ☐3 ☒4 ☐5 ☐6 ☐7 ☐8 ☐9 ☐10 ☐

Acceptance Test

Special Test

Calibration to MFG Specs

Clean

Limited Calibration

Functional Check

Performance Check

Modify

Repair-needs Charge Level

Other

1 ☒ In Tolerance2 ☐ Out of Tolerance >1x <2x3 ☐ Out of Tolerance >2x <3x4 ☐ Out of Tolerance >3x <5x5 ☐ Out of Tolerance >5x6 ☐ Out of Tolerance-Undetermined7 ☐ Inoperative8 ☐ Damaged9 ☐ Not Used10 ☐ Not Determined11 ☐ Exceeded12 ☐ Extension

Calibrated By: Brian Berls

S#: 102182 Phone: 526-2761

**CALIBRATION STANDARDS USED**

715356	715424	712788	714272				

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED  
VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

**OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION**

NOMINAL (STD)

UNITS

AS FOUND (UUT)

MFG. ACCURACY


COMMENTS

**INL**  
**CALIBRATION INPUT DATA**

11/30/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 722811    Mfr: EXTECH    Model: TYPE K    Noun Name: THERMOCOUPLE    Serial #:

Calibration Date: 5/5/2005

**ACTION CODE****AS FOUND**

Next Cal Due Date: 5/5/2006

Charge Level: 4

Repair/Adj/etc C.L: 0

Material Amount: 0

Charge Number: 100853GSA

Cal Work Inst ID: 3052

Outside Vendor:

1 ☐

Acceptance Test

2 ☐

Special Test

3 ☒

Calibration to MFG Specs

4 ☐

Clean

5 ☐

Limited Calibration

6 ☐

Functional Check

7 ☐

Performance Check

8 ☐

Modify

9 ☐

Repair-needs Charge Level

10 ☐

Other

1 ☒ In Tolerance2 ☐ Out of Tolerance >1x <2x3 ☐ Out of Tolerance >2x <3x4 ☐ Out of Tolerance >3x <5x5 ☐ Out of Tolerance >5x6 ☐ Out of Tolerance-Undetermined7 ☐ Inoperative8 ☐ Damaged9 ☐ Not Used10 ☐ Not Determined11 ☐ Exceeded12 ☐ Extension

Calibrated By: Chris Dudley    S#: 101142    Phone: 526-2761

**CALIBRATION STANDARDS USED**

715424	714272	714273	717617				

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-5.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

**OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION**

NOMINAL (STD)

UNITS

AS FOUND (UUT)

MFG. ACCURACY

\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

**COMMENTS**

INITIAL CALIBRATION TEST REFERRAL qa# 114094  
VERIFIED TO 220 DEG C.

**INL**  
**CALIBRATION INPUT DATA**

11/30/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 722811      Mfr: EXTECH      Model: TYPE K      Noun Name: THERMOCOUPLE      Serial #:

Calibration Date: 4/27/2006

**ACTION CODE****AS FOUND**

Next Cal Due Date: 4/27/2007

1 ☐

Acceptance Test

1 ☒ In Tolerance

Charge Level: 4

2 ☐

Special Test

2 ☐ Out of Tolerance >1x <2x

Repair/Adj/etc C.L.: 0

3 ☒

Calibration to MFG Specs

3 ☐ Out of Tolerance >2x <3x

Material Amount: 0

4 ☐

Clean

4 ☐ Out of Tolerance >3x <5x

Charge Number: 100853GSA

5 ☐

Limited Calibration

5 ☐ Out of Tolerance >5x

Cal Work Inst ID: 3052V

6 ☐

Functional Check

6 ☐ Out of Tolerance-Undetermined

Outside Vendor:

7 ☐

Performance Check

7 ☐ Inoperative8 ☐

Modify

8 ☐ Damaged9 ☐

Repair-needs Charge Level

9 ☐ Not Used10 ☐

Other

10 ☐ Not Determined11 ☐ Excessed

Calibrated By: Brian Berls

S#: 102182 Phone: 526-2761

12 ☐ Extension**CALIBRATION STANDARDS USED**

714272	714273	715424	717615				

**STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES**

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-5.0 ° C (20-50% RH)

**Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.**

**OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION****NOMINAL (STD)****UNITS****AS FOUND (IUT)****MFG. ACCURACY**


**COMMENTS**

VERIFIED TO 220 DEG C.

**INL**  
**CALIBRATION INPUT DATA**

11/30/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 722811      Mfr: EXTECH      Model: TYPE K      Noun Name: THERMOCOUPLE      Serial #:

Calibration Date: 5/8/2007

**ACTION CODE****AS FOUND**

Next Cal Due Date: 5/8/2008	1	<input type="checkbox"/>	Acceptance Test	1	<input checked="" type="radio"/>	In Tolerance
Charge Level: 4	2	<input type="checkbox"/>	Special Test	2	<input type="radio"/>	Out of Tolerance >1x <2x
Repair/Adj/etc C.L.: 0	3	<input checked="" type="checkbox"/>	Calibration to MFG Specs	3	<input type="radio"/>	Out of Tolerance >2x <3x
Material Amount: 0	4	<input type="checkbox"/>	Clean	4	<input type="radio"/>	Out of Tolerance >3x <5x
Charge Number: 100853GSA	5	<input type="checkbox"/>	Limited Calibration	5	<input type="radio"/>	Out of Tolerance >5x
Cal Work Inst ID: 3052V	6	<input type="checkbox"/>	Functional Check	6	<input type="radio"/>	Out of Tolerance-Undetermined
Outside Vendor:	7	<input type="checkbox"/>	Performance Check	7	<input type="checkbox"/>	Inoperative
	8	<input type="checkbox"/>	Modify	8	<input type="checkbox"/>	Damaged
	9	<input type="checkbox"/>	Repair-needs Charge Level	9	<input type="checkbox"/>	Not Used
	10	<input type="checkbox"/>	Other	10	<input type="checkbox"/>	Not Determined
				11	<input type="checkbox"/>	Excessed
Calibrated By: Brian Berls      S#: 102182      Phone: 526-2761				12	<input type="checkbox"/>	Extension

**CALIBRATION STANDARDS USED**

715424	715356	712788	714272				

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED  
VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

NOMINAL (STD)	OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION UNITS	AS FOUND (UUT)	MFG. ACCURACY

COMMENTS

**INL**  
**CALIBRATION INPUT DATA**

11/30/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 722812      Mfr: EXTECH      Model: TYPE K      Noun Name: THERMOCOUPLE      Serial #:

Calibration Date: 5/5/2005

**ACTION CODE****AS FOUND**

Next Cal Due Date: 5/5/2006	1	<input type="checkbox"/>	Acceptance Test	1	<input checked="" type="radio"/>	In Tolerance
Charge Level: 4	2	<input type="checkbox"/>	Special Test	2	<input type="radio"/>	Out of Tolerance >1x <2x
Repair/Adj/etc C.L.: 0	3	<input checked="" type="checkbox"/>	Calibration to MFG Specs	3	<input type="radio"/>	Out of Tolerance >2x <3x
Material Amount: 0	4	<input type="checkbox"/>	Clean	4	<input type="radio"/>	Out of Tolerance >3x <5x
Charge Number: 100853GSA	5	<input type="checkbox"/>	Limited Calibration	5	<input type="radio"/>	Out of Tolerance >5x
Cal Work Inst ID: 3052	6	<input type="checkbox"/>	Functional Check	6	<input type="radio"/>	Out of Tolerance-Undetermined
Outside Vendor:	7	<input type="checkbox"/>	Performance Check	7	<input type="checkbox"/>	Inoperative
	8	<input type="checkbox"/>	Modify	8	<input type="checkbox"/>	Damaged
	9	<input type="checkbox"/>	Repair-needs Charge Level	9	<input type="checkbox"/>	Not Used
	10	<input type="checkbox"/>	Other	10	<input type="checkbox"/>	Not Determined
				11	<input type="checkbox"/>	Excessed
				12	<input type="checkbox"/>	Extension

Calibrated By: Chris Dudley      S#: 101142      Phone: 526-2761

**CALIBRATION STANDARDS USED**

715424	714272	714273	717617				

**STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES**

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

NOMINAL (STD)	OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION UNITS	AS FOUND (UUT)	MFG. ACCURACY

**COMMENTS**

INITIAL CALIBRATION TEST REFERRAL qa# 114094  
VERIFIED TO 220 DEG C.

**INL**  
**CALIBRATION INPUT DATA**

11/30/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 722812      Mfr: EXTECH      Model: TYPE K      Noun Name: THERMOCOUPLE      Serial #:

Calibration Date: 4/27/2006

**ACTION CODE****AS FOUND**

Next Cal Due Date: 4/27/2007	1	<input type="checkbox"/>	Acceptance Test	1	<input checked="" type="radio"/>	In Tolerance
Charge Level: 4	2	<input type="checkbox"/>	Special Test	2	<input type="radio"/>	Out of Tolerance >1x <2x
Repair/Adj/etc C.L: 0	3	<input checked="" type="checkbox"/>	Calibration to MFG Specs	3	<input type="radio"/>	Out of Tolerance >2x <3x
Material Amount: 0	4	<input type="checkbox"/>	Clean	4	<input type="radio"/>	Out of Tolerance >3x <5x
Charge Number: 100853GSA	5	<input type="checkbox"/>	Limited Calibration	5	<input type="radio"/>	Out of Tolerance >5x
Cal Work Inst ID: 3052V	6	<input type="checkbox"/>	Functional Check	6	<input type="radio"/>	Out of Tolerance-Undetermined
Outside Vendor:	7	<input type="checkbox"/>	Performance Check	7	<input type="checkbox"/>	Inoperative
	8	<input type="checkbox"/>	Modify	8	<input type="checkbox"/>	Damaged
	9	<input type="checkbox"/>	Repair-needs Charge Level	9	<input type="checkbox"/>	Not Used
	10	<input type="checkbox"/>	Other	10	<input type="checkbox"/>	Not Determined
				11	<input type="checkbox"/>	Excessed
Calibrated By: Brian Berls	S#: 102182	Phone: 526-2761		12	<input type="checkbox"/>	Extension

**CALIBRATION STANDARDS USED**

714272	714273	717615	715424				

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-5.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

NOMINAL (STD)	OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION UNITS	AS FOUND (UUT)	MFG. ACCURACY

**COMMENTS**

VERIFIED TO 220 DEG C.

11/30/2007

**INL**  
**CALIBRATION INPUT DATA**

11/30/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 722812      Mfr: EXTECH      Model: TYPE K      Noun Name: THERMOCOUPLE      Serial #:

Calibration Date: 5/8/2007

**ACTION CODE**

**AS FOUND**

Next Cal Due Date: 5/8/2008	1	<input type="checkbox"/>	Acceptance Test	1	<input checked="" type="radio"/>	In Tolerance
Charge Level: 4	2	<input type="checkbox"/>	Special Test	2	<input type="radio"/>	Out of Tolerance >1x <2x
Repair/Adj/etc C.L.: 0	3	<input checked="" type="checkbox"/>	Calibration to MFG Specs	3	<input type="radio"/>	Out of Tolerance >2x <3x
Material Amount: 0	4	<input type="checkbox"/>	Clean	4	<input type="radio"/>	Out of Tolerance >3x <5x
Charge Number: 100853GSA	5	<input type="checkbox"/>	Limited Calibration	5	<input type="radio"/>	Out of Tolerance >5x
Cal Work Inst ID: 3052V	6	<input type="checkbox"/>	Functional Check	6	<input type="radio"/>	Out of Tolerance-Undetermined
Outside Vendor:	7	<input type="checkbox"/>	Performance Check	7	<input type="checkbox"/>	Inoperative
	8	<input type="checkbox"/>	Modify	8	<input type="checkbox"/>	Damaged
	9	<input type="checkbox"/>	Repair-needs Charge Level	9	<input type="checkbox"/>	Not Used
	10	<input type="checkbox"/>	Other	10	<input type="checkbox"/>	Not Determined
				11	<input type="checkbox"/>	Excessed
Calibrated By: Brian Berls      S#: 102182      Phone: 526-2761				12	<input type="checkbox"/>	Extension

**CALIBRATION STANDARDS USED**

715424	715356	712788	714272				

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

**OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION**

NOMINAL (STD)	UNITS	AS FOUND (UUT)	MFG. ACCURACY

**COMMENTS**



**INL**  
**CALIBRATION INPUT DATA**

11/30/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 722813      Mfr: EXTECH      Model: TYPE K      Noun Name: THERMOCOUPLE      Serial #:

Calibration Date: 5/5/2005

**ACTION CODE****AS FOUND**

Next Cal Due Date: 5/5/2006	1	<input type="checkbox"/>	Acceptance Test	1	<input checked="" type="radio"/>	In Tolerance
Charge Level: 4	2	<input type="checkbox"/>	Special Test	2	<input type="radio"/>	Out of Tolerance >1x <2x
Repair/Adj/etc C.L: 0	3	<input checked="" type="checkbox"/>	Calibration to MFG Specs	3	<input type="radio"/>	Out of Tolerance >2x <3x
Material Amount: 0	4	<input type="checkbox"/>	Clean	4	<input type="radio"/>	Out of Tolerance >3x <5x
Charge Number: 100853GSA	5	<input type="checkbox"/>	Limited Calibration	5	<input type="radio"/>	Out of Tolerance >5x
Cal Work Inst ID: 3052	6	<input type="checkbox"/>	Functional Check	6	<input type="radio"/>	Out of Tolerance-Undetermined
Outside Vendor:	7	<input type="checkbox"/>	Performance Check	7	<input type="checkbox"/>	Inoperative
	8	<input type="checkbox"/>	Modify	8	<input type="checkbox"/>	Damaged
	9	<input type="checkbox"/>	Repair-needs Charge Level	9	<input type="checkbox"/>	Not Used
	10	<input type="checkbox"/>	Other	10	<input type="checkbox"/>	Not Determined
				11	<input type="checkbox"/>	Excessed
Calibrated By: Chris Dudley      S#: 101142      Phone: 526-2761				12	<input type="checkbox"/>	Extension

**CALIBRATION STANDARDS USED**

715424	714272	714273	717617				

**STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES**

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

NOMINAL (STD)	OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION UNITS	AS FOUND (UUT)	MFG. ACCURACY

**COMMENTS**

INITIAL CALIBRATION TEST REFERRAL qa# 114094  
VERIFIED TO 220 DEG C.

**INL**  
**CALIBRATION INPUT DATA**

11/30/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 722813    Mfr: EXTECH    Model: TYPE K    Noun Name: THERMOCOUPLE    Serial #:

Calibration Date: 4/27/2006

**ACTION CODE****AS FOUND**

Next Cal Due Date: 4/27/2007	1	<input type="checkbox"/>	Acceptance Test	1	<input checked="" type="radio"/>	In Tolerance
Charge Level: 4	2	<input type="checkbox"/>	Special Test	2	<input type="radio"/>	Out of Tolerance >1x <2x
Repair/Adj/etc C.L: 0	3	<input checked="" type="checkbox"/>	Calibration to MFG Specs	3	<input type="radio"/>	Out of Tolerance >2x <3x
Material Amount: 0	4	<input type="checkbox"/>	Clean	4	<input type="radio"/>	Out of Tolerance >3x <5x
Charge Number: 100853GSA	5	<input type="checkbox"/>	Limited Calibration	5	<input type="radio"/>	Out of Tolerance >5x
Cal Work Inst ID: 3052V	6	<input type="checkbox"/>	Functional Check	6	<input type="radio"/>	Out of Tolerance-Undetermined
Outside Vendor:	7	<input type="checkbox"/>	Performance Check	7	<input type="checkbox"/>	Inoperative
	8	<input type="checkbox"/>	Modify	8	<input type="checkbox"/>	Damaged
	9	<input type="checkbox"/>	Repair-needs Charge Level	9	<input type="checkbox"/>	Not Used
	10	<input type="checkbox"/>	Other	10	<input type="checkbox"/>	Not Determined
				11	<input type="checkbox"/>	Excessed
				12	<input type="checkbox"/>	Extension

Calibrated By: Brian Berls    S#: 102182    Phone: 526-2761

**CALIBRATION STANDARDS USED**

714272	714273	717615	715424				

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED  
VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-5.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

**OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION**

NOMINAL (STD)	UNITS	AS FOUND (UUT)	MFG. ACCURACY

**COMMENTS**

VERIFIED TO 220 DEG C.

**INL  
CALIBRATION INPUT DATA**

11/30/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 722813      Mfr: EXTECH      Model: TYPE K      Noun Name: THERMOCOUPLE      Serial #:

Calibration Date: 5/8/2007

**ACTION CODE**

**AS FOUND**

Next Cal Due Date: 5/8/2008	1	<input type="checkbox"/>	Acceptance Test	1	<input checked="" type="radio"/>	In Tolerance
Charge Level: 4	2	<input type="checkbox"/>	Special Test	2	<input type="radio"/>	Out of Tolerance >1x <2x
Repair/Adj/etc C.L: 0	3	<input checked="" type="checkbox"/>	Calibration to MFG Specs	3	<input type="radio"/>	Out of Tolerance >2x <3x
Material Amount: 0	4	<input type="checkbox"/>	Clean	4	<input type="radio"/>	Out of Tolerance >3x <5x
Charge Number: 100853GSA	5	<input type="checkbox"/>	Limited Calibration	5	<input type="radio"/>	Out of Tolerance >5x
Cal Work Inst ID: 3052V	6	<input type="checkbox"/>	Functional Check	6	<input type="radio"/>	Out of Tolerance-Undetermined
Outside Vendor:	7	<input type="checkbox"/>	Performance Check	7	<input type="checkbox"/>	Inoperative
	8	<input type="checkbox"/>	Modify	8	<input type="checkbox"/>	Damaged
	9	<input type="checkbox"/>	Repair-needs Charge Level	9	<input type="checkbox"/>	Not Used
	10	<input type="checkbox"/>	Other	10	<input type="checkbox"/>	Not Determined
				11	<input type="checkbox"/>	Excessed
Calibrated By: Brian Berls				12	<input type="checkbox"/>	Extension

**CALIBRATION STANDARDS USED**

715424	715356	712788	714272				

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

**OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION**

NOMINAL (STD)

UNITS

AS FOUND (UUT)

MFG. ACCURACY


**COMMENTS**

# INL CALIBRATION INPUT DATA

11/30/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 722814    Mfr: EXTECH    Model: TYPE K    Noun Name: THERMOCOUPLE    Serial #:

Calibration Date: 5/5/2005

**ACTION CODE****AS FOUND**

Next Cal Due Date: 5/5/2006	1	<input type="checkbox"/>	Acceptance Test	1	<input checked="" type="radio"/>	In Tolerance
Charge Level: 4	2	<input type="checkbox"/>	Special Test	2	<input type="radio"/>	Out of Tolerance >1x <2x
Repair/Adj/etc C.L: 0	3	<input checked="" type="checkbox"/>	Calibration to MFG Specs	3	<input type="radio"/>	Out of Tolerance >2x <3x
Material Amount: 0	4	<input type="checkbox"/>	Clean	4	<input type="radio"/>	Out of Tolerance >3x <5x
Charge Number: 100853GSA	5	<input type="checkbox"/>	Limited Calibration	5	<input type="radio"/>	Out of Tolerance >5x
Cal Work Inst ID: 3052	6	<input type="checkbox"/>	Functional Check	6	<input type="radio"/>	Out of Tolerance-Undetermined
Outside Vendor:	7	<input type="checkbox"/>	Performance Check	7	<input type="checkbox"/>	Inoperative
	8	<input type="checkbox"/>	Modify	8	<input type="checkbox"/>	Damaged
	9	<input type="checkbox"/>	Repair-needs Charge Level	9	<input type="checkbox"/>	Not Used
	10	<input type="checkbox"/>	Other	10	<input type="checkbox"/>	Not Determined
				11	<input type="checkbox"/>	Excessed
				12	<input type="checkbox"/>	Extension

Calibrated By: Chris Dudley    S#: 101142    Phone: 526-2761

**CALIBRATION STANDARDS USED**

715424	714272	714273	717617				

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-5.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

NOMINAL (STD)	OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION UNITS	AS FOUND (UUT)	MFG. ACCURACY

**COMMENTS**

INITIAL CALIBRATION TEST REFERRAL qa# 114094  
VERIFIED TO 220 DEG C.

**INL  
CALIBRATION INPUT DATA**

11/30/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 722814      Mfr: EXTECH      Model: TYPE K      Noun Name: THERMOCOUPLE      Serial #:

Calibration Date: 4/27/2006

**ACTION CODE**

**AS FOUND**

Next Cal Due Date: 4/27/2007	1	<input type="checkbox"/>	Acceptance Test	1	<input checked="" type="radio"/>	In Tolerance
Charge Level: 4	2	<input type="checkbox"/>	Special Test	2	<input type="radio"/>	Out of Tolerance >1x <2x
Repair/Adj/etc C.L: 0	3	<input checked="" type="checkbox"/>	Calibration to MFG Specs	3	<input type="radio"/>	Out of Tolerance >2x <3x
Material Amount: 0	4	<input type="checkbox"/>	Clean	4	<input type="radio"/>	Out of Tolerance >3x <5x
Charge Number: 100853GSA	5	<input type="checkbox"/>	Limited Calibration	5	<input type="radio"/>	Out of Tolerance >5x
Cal Work Inst ID: 3052V	6	<input type="checkbox"/>	Functional Check	6	<input type="radio"/>	Out of Tolerance-Undetermined
Outside Vendor:	7	<input type="checkbox"/>	Performance Check	7	<input type="checkbox"/>	Inoperative
	8	<input type="checkbox"/>	Modify	8	<input type="checkbox"/>	Damaged
	9	<input type="checkbox"/>	Repair-needs Charge Level	9	<input type="checkbox"/>	Not Used
	10	<input type="checkbox"/>	Other	10	<input type="checkbox"/>	Not Determined
				11	<input type="checkbox"/>	Excessed
Calibrated By: Brian Berls      S#: 102182      Phone: 526-2761				12	<input type="checkbox"/>	Extension

**CALIBRATION STANDARDS USED**

714272	714273	717615	715424				

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

NOMINAL (STD)	OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION UNITS	AS FOUND (UUT)	MFG. ACCURACY

**COMMENTS**

VERIFIED TO 220 DEG C.

**INL**  
**CALIBRATION INPUT DATA**

11/30/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 722814	Mfr: EXTECH	Model: TYPE K	Noun Name: THERMOCOUPLE	Serial #:
Calibration Date: 5/8/2007	<b>ACTION CODE</b>		<b>AS FOUND</b>	
Next Cal Due Date: 5/8/2008	1 <input type="checkbox"/>	Acceptance Test	1 <input checked="" type="radio"/>	In Tolerance
Charge Level: 4	2 <input type="checkbox"/>	Special Test	2 <input type="radio"/>	Out of Tolerance >1x <2x
Repair/Adj/etc C.L: 0	3 <input checked="" type="checkbox"/>	Calibration to MFG Specs	3 <input type="radio"/>	Out of Tolerance >2x <3x
Material Amount: 0	4 <input type="checkbox"/>	Clean	4 <input type="radio"/>	Out of Tolerance >3x <5x
Charge Number: 100853GSA	5 <input type="checkbox"/>	Limited Calibration	5 <input type="radio"/>	Out of Tolerance >5x
Cal Work Inst ID: 3052V	6 <input type="checkbox"/>	Functional Check	6 <input type="radio"/>	Out of Tolerance-Undetermined
Outside Vendor:	7 <input type="checkbox"/>	Performance Check	7 <input type="checkbox"/>	Inoperative
	8 <input type="checkbox"/>	Modify	8 <input type="checkbox"/>	Damaged
	9 <input type="checkbox"/>	Repair-needs Charge Level	9 <input type="checkbox"/>	Not Used
	10 <input type="checkbox"/>	Other	10 <input type="checkbox"/>	Not Determined
			11 <input type="checkbox"/>	Excessed
Calibrated By: Brian Berls	S#: 102182	Phone: 526-2761	12 <input type="checkbox"/>	Extension

**CALIBRATION STANDARDS USED**

715424	715356	712788	714272				

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

NOMINAL (STD)	OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION UNITS	AS FOUND (UUT)	MFG. ACCURACY

COMMENTS

**INL  
CALIBRATION INPUT DATA**

11/30/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 722815    Mfr: EXTECH    Model: TYPE K    Noun Name: THERMOCOUPLE    Serial #:

Calibration Date: 5/5/2005

**ACTION CODE**

**AS FOUND**

Next Cal Due Date: 5/5/2006	1	<input type="checkbox"/>	Acceptance Test	1	<input checked="" type="radio"/>	In Tolerance
Charge Level: 4	2	<input type="checkbox"/>	Special Test	2	<input type="radio"/>	Out of Tolerance >1x <2x
Repair/Adj/etc C.L: 0	3	<input checked="" type="checkbox"/>	Calibration to MFG Specs	3	<input type="radio"/>	Out of Tolerance >2x <3x
Material Amount: 0	4	<input type="checkbox"/>	Clean	4	<input type="radio"/>	Out of Tolerance >3x <5x
Charge Number: 100853GSA	5	<input type="checkbox"/>	Limited Calibration	5	<input type="radio"/>	Out of Tolerance >5x
Cal Work Inst ID: 3052	6	<input type="checkbox"/>	Functional Check	6	<input type="radio"/>	Out of Tolerance-Undetermined
Outside Vendor:	7	<input type="checkbox"/>	Performance Check	7	<input type="checkbox"/>	Inoperative
	8	<input type="checkbox"/>	Modify	8	<input type="checkbox"/>	Damaged
	9	<input type="checkbox"/>	Repair-needs Charge Level	9	<input type="checkbox"/>	Not Used
	10	<input type="checkbox"/>	Other	10	<input type="checkbox"/>	Not Determined
				11	<input type="checkbox"/>	Excessed
				12	<input type="checkbox"/>	Extension

Calibrated By: Chris Dudley    S#: 101142    Phone: 526-2761

**CALIBRATION STANDARDS USED**

715424	714272	714273	717617				

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

NOMINAL (STD)	OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION		MFG. ACCURACY
	UNITS	AS FOUND (UUT)	

**COMMENTS**

INITIAL CALIBRATION TEST REFERRAL qa# 114094  
VERIFIED TO 220 DEG C.

**INL**  
**CALIBRATION INPUT DATA**

11/30/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 722815      Mfr: EXTECH      Model: TYPE K      Noun Name: THERMOCOUPLE      Serial #:

Calibration Date: 4/27/2006

**ACTION CODE****AS FOUND**

Next Cal Due Date: 4/27/2007	1	<input type="checkbox"/>	Acceptance Test	1	<input checked="" type="radio"/>	In Tolerance
Charge Level: 4	2	<input type="checkbox"/>	Special Test	2	<input type="radio"/>	Out of Tolerance >1x <2x
Repair/Adj/etc C.L: 0	3	<input checked="" type="checkbox"/>	Calibration to MFG Specs	3	<input type="radio"/>	Out of Tolerance >2x <3x
Material Amount: 0	4	<input type="checkbox"/>	Clean	4	<input type="radio"/>	Out of Tolerance >3x <5x
Charge Number: 100853GSA	5	<input type="checkbox"/>	Limited Calibration	5	<input type="radio"/>	Out of Tolerance >5x
Cal Work Inst ID: 3052V	6	<input type="checkbox"/>	Functional Check	6	<input type="radio"/>	Out of Tolerance-Undetermined
Outside Vendor:	7	<input type="checkbox"/>	Performance Check	7	<input type="checkbox"/>	Inoperative
	8	<input type="checkbox"/>	Modify	8	<input type="checkbox"/>	Damaged
	9	<input type="checkbox"/>	Repair-needs Charge Level	9	<input type="checkbox"/>	Not Used
	10	<input type="checkbox"/>	Other	10	<input type="checkbox"/>	Not Determined
				11	<input type="checkbox"/>	Excessed
Calibrated By: Brian Berls				12	<input type="checkbox"/>	Extension

**CALIBRATION STANDARDS USED**

714272	714273	717615	715424				

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

**OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION**

NOMINAL (STD)	UNITS	AS FOUND (UUT)	MFG. ACCURACY

**COMMENTS**

VERIFIED TO 220 DEG C.



**INL  
CALIBRATION INPUT DATA**

11/30/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 722815      Mfr: EXTECH      Model: TYPE K      Noun Name: THERMOCOUPLE      Serial #:

Calibration Date: 5/8/2007

**ACTION CODE**

**AS FOUND**

Next Cal Due Date: 5/8/2008	1	<input type="checkbox"/>	Acceptance Test	1	<input checked="" type="radio"/>	In Tolerance
Charge Level: 4	2	<input type="checkbox"/>	Special Test	2	<input type="radio"/>	Out of Tolerance >1x <2x
Repair/Adj/etc C.L: 0	3	<input checked="" type="checkbox"/>	Calibration to MFG Specs	3	<input type="radio"/>	Out of Tolerance >2x <3x
Material Amount: 0	4	<input type="checkbox"/>	Clean	4	<input type="radio"/>	Out of Tolerance >3x <5x
Charge Number: 100853GSA	5	<input type="checkbox"/>	Limited Calibration	5	<input type="radio"/>	Out of Tolerance >5x
Cal Work Inst ID: 3052V	6	<input type="checkbox"/>	Functional Check	6	<input type="radio"/>	Out of Tolerance-Undetermined
Outside Vendor:	7	<input type="checkbox"/>	Performance Check	7	<input type="checkbox"/>	Inoperative
	8	<input type="checkbox"/>	Modify	8	<input type="checkbox"/>	Damaged
	9	<input type="checkbox"/>	Repair-needs Charge Level	9	<input type="checkbox"/>	Not Used
	10	<input type="checkbox"/>	Other	10	<input type="checkbox"/>	Not Determined
				11	<input type="checkbox"/>	Excessed
Calibrated By: Brian Berls      S#: 102182      Phone: 526-2761				12	<input type="checkbox"/>	Extension

**CALIBRATION STANDARDS USED**

715424	715356	712788	714272				

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-5.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

NOMINAL (STD)	OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION UNITS	AS FOUND (UUT)	MFG. ACCURACY

COMMENTS

# INL CALIBRATION INPUT DATA

11/30/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 723409    Mfr: EXTECH    Model: TYPE K    Noun Name: THERMOCOUPLE    Serial #:

Calibration Date: 11/21/2005

## ACTION CODE

## AS FOUND

Next Cal Due Date: 5/21/2007	1	<input type="checkbox"/>	Acceptance Test	1	<input checked="" type="radio"/>	In Tolerance
Charge Level: 4	2	<input type="checkbox"/>	Special Test	2	<input type="radio"/>	Out of Tolerance >1x <2x
Repair/Adj/etc C.L: 0	3	<input checked="" type="checkbox"/>	Calibration to MFG Specs	3	<input type="radio"/>	Out of Tolerance >2x <3x
Material Amount: 0	4	<input type="checkbox"/>	Clean	4	<input type="radio"/>	Out of Tolerance >3x <5x
Charge Number: 100853GSA	5	<input type="checkbox"/>	Limited Calibration	5	<input type="radio"/>	Out of Tolerance >5x
Cal Work Inst ID: 3052U	6	<input type="checkbox"/>	Functional Check	6	<input type="radio"/>	Out of Tolerance-Undetermined
Outside Vendor:	7	<input type="checkbox"/>	Performance Check	7	<input type="checkbox"/>	Inoperative
	8	<input type="checkbox"/>	Modify	8	<input type="checkbox"/>	Damaged
	9	<input type="checkbox"/>	Repair-needs Charge Level	9	<input type="checkbox"/>	Not Used
	10	<input type="checkbox"/>	Other	10	<input type="checkbox"/>	Not Determined
				11	<input type="checkbox"/>	Excessed
				12	<input type="checkbox"/>	Extension

Calibrated By: Chris Dudley    S#: 101142    Phone: 526-2761

## CALIBRATION STANDARDS USED

370602	723247	722456	722915	717617			

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

## LABORATORY TEMPERATURE AND HUMIDITY

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

## OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION

NOMINAL (STD)

UNITS

AS FOUND (UUT)

MFG. ACCURACY


## COMMENTS

INITIAL CALIBRATION  
RECEIVING INSPECTION TEST REFERRAL QA# 115846

# INL CALIBRATION INPUT DATA

11/30/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 723409      Mfr: EXTECH      Model: TYPE K      Noun Name: THERMOCOUPLE      Serial #:

Calibration Date: 5/8/2007

**ACTION CODE****AS FOUND**

Next Cal Due Date: 11/8/2008	1	<input type="checkbox"/>	Acceptance Test	1	<input checked="" type="radio"/>	In Tolerance
Charge Level: 4	2	<input type="checkbox"/>	Special Test	2	<input type="radio"/>	Out of Tolerance >1x <2x
Repair/Adj/etc C.L: 0	3	<input checked="" type="checkbox"/>	Calibration to MFG Specs	3	<input type="radio"/>	Out of Tolerance >2x <3x
Material Amount: 0	4	<input type="checkbox"/>	Clean	4	<input type="radio"/>	Out of Tolerance >3x <5x
Charge Number: 100853GSA	5	<input type="checkbox"/>	Limited Calibration	5	<input type="radio"/>	Out of Tolerance >5x
Cal Work Inst ID: 3052V	6	<input type="checkbox"/>	Functional Check	6	<input type="radio"/>	Out of Tolerance-Undetermined
Outside Vendor:	7	<input type="checkbox"/>	Performance Check	7	<input type="checkbox"/>	Inoperative
	8	<input type="checkbox"/>	Modify	8	<input type="checkbox"/>	Damaged
	9	<input type="checkbox"/>	Repair-needs Charge Level	9	<input type="checkbox"/>	Not Used
	10	<input type="checkbox"/>	Other	10	<input type="checkbox"/>	Not Determined
				11	<input type="checkbox"/>	Excessed
Calibrated By: Brian Berls      S#: 102182      Phone: 526-2761				12	<input type="checkbox"/>	Extension

**CALIBRATION STANDARDS USED**

715424	712788	715356	714272				

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-5.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

NOMINAL (STD)	OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION		MFG. ACCURACY
	UNITS	AS FOUND (UUT)	


COMMENTS

**INL  
CALIBRATION INPUT DATA**

11/30/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 723410      Mfr: EXTECH      Model: TYPE K      Noun Name: THERMOCOUPLE      Serial #:

Calibration Date: 11/21/2005

**ACTION CODE**

**AS FOUND**

Next Cal Due Date: 5/21/2007	1	<input type="checkbox"/>	Acceptance Test	1	<input checked="" type="radio"/>	In Tolerance
Charge Level: 4	2	<input type="checkbox"/>	Special Test	2	<input type="radio"/>	Out of Tolerance >1x <2x
Repair/Adj/etc C.L.: 0	3	<input checked="" type="checkbox"/>	Calibration to MFG Specs	3	<input type="radio"/>	Out of Tolerance >2x <3x
Material Amount: 0	4	<input type="checkbox"/>	Clean	4	<input type="radio"/>	Out of Tolerance >3x <5x
Charge Number: 100853GSA	5	<input type="checkbox"/>	Limited Calibration	5	<input type="radio"/>	Out of Tolerance >5x
Cal Work Inst ID: 3052U	6	<input type="checkbox"/>	Functional Check	6	<input type="radio"/>	Out of Tolerance-Undetermined
Outside Vendor:	7	<input type="checkbox"/>	Performance Check	7	<input type="checkbox"/>	Inoperative
	8	<input type="checkbox"/>	Modify	8	<input type="checkbox"/>	Damaged
	9	<input type="checkbox"/>	Repair-needs Charge Level	9	<input type="checkbox"/>	Not Used
	10	<input type="checkbox"/>	Other	10	<input type="checkbox"/>	Not Determined
				11	<input type="checkbox"/>	Excessed
				12	<input type="checkbox"/>	Extension

Calibrated By: Chris Dudley

S#: 101142 Phone: 526-2761

**CALIBRATION STANDARDS USED**

370602	723247	722456	722915	717617			

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

NOMINAL (STD)	OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION UNITS	AS FOUND (UUT)	MFG. ACCURACY

**COMMENTS**

INITIAL CALIBRATION  
RECEIVING INSPECTION TEST REFERRAL QA# 115846

**INL**  
**CALIBRATION INPUT DATA**

11/30/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 723410    Mfr: EXTECH    Model: TYPE K    Noun Name: THERMOCOUPLE    Serial #:

Calibration Date: 5/8/2007

**ACTION CODE**

**AS FOUND**

Next Cal Due Date: 11/8/2008	1	<input type="checkbox"/>	Acceptance Test	1	<input checked="" type="radio"/>	In Tolerance
Charge Level: 4	2	<input type="checkbox"/>	Special Test	2	<input type="radio"/>	Out of Tolerance >1x <2x
Repair/Adj/etc C.L: 0	3	<input checked="" type="checkbox"/>	Calibration to MFG Specs	3	<input type="radio"/>	Out of Tolerance >2x <3x
Material Amount: 0	4	<input type="checkbox"/>	Clean	4	<input type="radio"/>	Out of Tolerance >3x <5x
Charge Number: 100853GSA	5	<input type="checkbox"/>	Limited Calibration	5	<input type="radio"/>	Out of Tolerance >5x
Cal Work Inst ID: 3052V	6	<input type="checkbox"/>	Functional Check	6	<input type="radio"/>	Out of Tolerance-Undetermined
Outside Vendor:	7	<input type="checkbox"/>	Performance Check	7	<input type="checkbox"/>	Inoperative
	8	<input type="checkbox"/>	Modify	8	<input type="checkbox"/>	Damaged
	9	<input type="checkbox"/>	Repair-needs Charge Level	9	<input type="checkbox"/>	Not Used
	10	<input type="checkbox"/>	Other	10	<input type="checkbox"/>	Not Determined
				11	<input type="checkbox"/>	Excessed
Calibrated By: Brian Berls    S#: 102182    Phone: 526-2761				12	<input type="checkbox"/>	Extension

**CALIBRATION STANDARDS USED**

715424	715356	712788	714272				

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-5.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

NOMINAL (STD)	OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION		MFG. ACCURACY
	UNITS	AS FOUND (UUT)	

COMMENTS

**INL**  
**CALIBRATION INPUT DATA**

11/30/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 723411      Mfr: EXTECH      Model: TYPE K      Noun Name: THERMOCOUPLE      Serial #:

Calibration Date: 11/21/2005

**ACTION CODE****AS FOUND**

Next Cal Due Date: 5/21/2007	1	<input type="checkbox"/>	Acceptance Test	1	<input checked="" type="radio"/>	In Tolerance
Charge Level: 4	2	<input type="checkbox"/>	Special Test	2	<input type="radio"/>	Out of Tolerance >1x <2x
Repair/Adj/etc C.L: 0	3	<input checked="" type="checkbox"/>	Calibration to MFG Specs	3	<input type="radio"/>	Out of Tolerance >2x <3x
Material Amount: 0	4	<input type="checkbox"/>	Clean	4	<input type="radio"/>	Out of Tolerance >3x <5x
Charge Number: 100853GSA	5	<input type="checkbox"/>	Limited Calibration	5	<input type="radio"/>	Out of Tolerance >5x
Cal Work Inst ID: 3052U	6	<input type="checkbox"/>	Functional Check	6	<input type="radio"/>	Out of Tolerance-Undetermined
Outside Vendor:	7	<input type="checkbox"/>	Performance Check	7	<input type="checkbox"/>	Inoperative
	8	<input type="checkbox"/>	Modify	8	<input type="checkbox"/>	Damaged
	9	<input type="checkbox"/>	Repair-needs Charge Level	9	<input type="checkbox"/>	Not Used
	10	<input type="checkbox"/>	Other	10	<input type="checkbox"/>	Not Determined
				11	<input type="checkbox"/>	Excessed
				12	<input type="checkbox"/>	Extension

Calibrated By: Chris Dudley      S#: 101142      Phone: 526-2761

**CALIBRATION STANDARDS USED**

370602	723247	722915	722456	717617			

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-5.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

NOMINAL (STD)	OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION UNITS	AS FOUND (UUT)	MFG. ACCURACY

**COMMENTS**

INITIAL CALIBRATION  
RECEIVING INSPECTION TEST REFERRAL QA# 115846

**INL  
CALIBRATION INPUT DATA**

11/30/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 723411      Mfr: EXTECH      Model: TYPE K      Noun Name: THERMOCOUPLE      Serial #:

Calibration Date: 5/8/2007

**ACTION CODE**

**AS FOUND**

Next Cal Due Date: 11/8/2008	1	<input type="checkbox"/>	Acceptance Test	1	<input checked="" type="radio"/>	In Tolerance
Charge Level: 4	2	<input type="checkbox"/>	Special Test	2	<input type="radio"/>	Out of Tolerance >1x <2x
Repair/Adj/etc C.L.: 0	3	<input checked="" type="checkbox"/>	Calibration to MFG Specs	3	<input type="radio"/>	Out of Tolerance >2x <3x
Material Amount: 0	4	<input type="checkbox"/>	Clean	4	<input type="radio"/>	Out of Tolerance >3x <5x
Charge Number: 100853GSA	5	<input type="checkbox"/>	Limited Calibration	5	<input type="radio"/>	Out of Tolerance >5x
Cal Work Inst ID: 3052V	6	<input type="checkbox"/>	Functional Check	6	<input type="radio"/>	Out of Tolerance-Undetermined
Outside Vendor:	7	<input type="checkbox"/>	Performance Check	7	<input type="checkbox"/>	Inoperative
	8	<input type="checkbox"/>	Modify	8	<input type="checkbox"/>	Damaged
	9	<input type="checkbox"/>	Repair-needs Charge Level	9	<input type="checkbox"/>	Not Used
	10	<input type="checkbox"/>	Other	10	<input type="checkbox"/>	Not Determined
				11	<input type="checkbox"/>	Excessed
Calibrated By: Brian Berls      S#: 102182      Phone: 526-2761				12	<input type="checkbox"/>	Extension

**CALIBRATION STANDARDS USED**

715424	715356	712788	714272				

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

NOMINAL (STD)	OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION UNITS	AS FOUND (UUT)	MFG. ACCURACY

**COMMENTS**

Replaced connector.

**INL  
CALIBRATION INPUT DATA**

11/30/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 723412    Mfr: EXTECH    Model: TYPE K    Noun Name: THERMOCOUPLE    Serial #:

Calibration Date: 11/21/2005

**ACTION CODE**

**AS FOUND**

Next Cal Due Date: 5/21/2007	1	<input type="checkbox"/>	Acceptance Test	1	<input checked="" type="radio"/>	In Tolerance
Charge Level: 4	2	<input type="checkbox"/>	Special Test	2	<input type="radio"/>	Out of Tolerance >1x <2x
Repair/Adj/etc C.L.: 0	3	<input checked="" type="checkbox"/>	Calibration to MFG Specs	3	<input type="radio"/>	Out of Tolerance >2x <3x
Material Amount: 0	4	<input type="checkbox"/>	Clean	4	<input type="radio"/>	Out of Tolerance >3x <5x
Charge Number: 100853GSA	5	<input type="checkbox"/>	Limited Calibration	5	<input type="radio"/>	Out of Tolerance >5x
Cal Work Inst ID: 3052U	6	<input type="checkbox"/>	Functional Check	6	<input type="radio"/>	Out of Tolerance-Undetermined
Outside Vendor:	7	<input type="checkbox"/>	Performance Check	7	<input type="checkbox"/>	Inoperative
	8	<input type="checkbox"/>	Modify	8	<input type="checkbox"/>	Damaged
	9	<input type="checkbox"/>	Repair-needs Charge Level	9	<input type="checkbox"/>	Not Used
	10	<input type="checkbox"/>	Other	10	<input type="checkbox"/>	Not Determined
				11	<input type="checkbox"/>	Excessed
Calibrated By: Chris Dudley    S#: 101142    Phone: 526-2761				12	<input type="checkbox"/>	Extension

**CALIBRATION STANDARDS USED**

370602	723247	722456	722915	717617			

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

**OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION**

NOMINAL (STD)	UNITS	AS FOUND (UUT)	MFG. ACCURACY
---------------	-------	----------------	---------------


**COMMENTS**

INITIAL CALIBRATION  
RECEIVING INSPECTION TEST REFERRAL QA# 115846



**INL**  
**CALIBRATION INPUT DATA**

11/30/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 723412    Mfr: EXTECH    Model: TYPE K    Noun Name: THERMOCOUPLE    Serial #:

Calibration Date: 4/27/2006

**ACTION CODE****AS FOUND**

Next Cal Due Date: 10/27/2007	1	<input type="checkbox"/>	Acceptance Test	1	<input checked="" type="radio"/>	In Tolerance
Charge Level: 4	2	<input type="checkbox"/>	Special Test	2	<input type="radio"/>	Out of Tolerance >1x <2x
Repair/Adj/etc C.L: 0	3	<input checked="" type="checkbox"/>	Calibration to MFG Specs	3	<input type="radio"/>	Out of Tolerance >2x <3x
Material Amount: 0	4	<input type="checkbox"/>	Clean	4	<input type="radio"/>	Out of Tolerance >3x <5x
Charge Number: 100853gsa	5	<input type="checkbox"/>	Limited Calibration	5	<input type="radio"/>	Out of Tolerance >5x
Cal Work Inst ID: 3052V	6	<input type="checkbox"/>	Functional Check	6	<input type="radio"/>	Out of Tolerance-Undetermined
Outside Vendor:	7	<input type="checkbox"/>	Performance Check	7	<input type="checkbox"/>	Inoperative
	8	<input type="checkbox"/>	Modify	8	<input type="checkbox"/>	Damaged
	9	<input type="checkbox"/>	Repair-needs Charge Level	9	<input type="checkbox"/>	Not Used
	10	<input type="checkbox"/>	Other	10	<input type="checkbox"/>	Not Determined
				11	<input type="checkbox"/>	Excessed
Calibrated By: Brian Berls    S#: 102182    Phone: 526-2761				12	<input type="checkbox"/>	Extension

**CALIBRATION STANDARDS USED**

714272	714273	717615	715424				

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED  
VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-5.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

NOMINAL (STD)	OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION UNITS	AS FOUND (UUT)	MFG. ACCURACY

**COMMENTS**

VERIFIED TO 220 DEG C.

**INL**  
**CALIBRATION INPUT DATA**

11/30/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 723412      Mfr: EXTECH      Model: TYPE K      Noun Name: THERMOCOUPLE      Serial #:

Calibration Date: 5/8/2007

**ACTION CODE****AS FOUND**

Next Cal Due Date: 11/8/2008	1	<input type="checkbox"/>	Acceptance Test	1	<input checked="" type="radio"/>	In Tolerance
Charge Level: 4	2	<input type="checkbox"/>	Special Test	2	<input type="radio"/>	Out of Tolerance >1x <2x
Repair/Adj/etc C.L: 0	3	<input checked="" type="checkbox"/>	Calibration to MFG Specs	3	<input type="radio"/>	Out of Tolerance >2x <3x
Material Amount: 0	4	<input type="checkbox"/>	Clean	4	<input type="radio"/>	Out of Tolerance >3x <5x
Charge Number: 100853GSA	5	<input type="checkbox"/>	Limited Calibration	5	<input type="radio"/>	Out of Tolerance >5x
Cal Work Inst ID: 3052V	6	<input type="checkbox"/>	Functional Check	6	<input type="radio"/>	Out of Tolerance-Undetermined
Outside Vendor:	7	<input type="checkbox"/>	Performance Check	7	<input type="checkbox"/>	Inoperative
	8	<input type="checkbox"/>	Modify	8	<input type="checkbox"/>	Damaged
	9	<input type="checkbox"/>	Repair-needs Charge Level	9	<input type="checkbox"/>	Not Used
	10	<input type="checkbox"/>	Other	10	<input type="checkbox"/>	Not Determined
				11	<input type="checkbox"/>	Excessed
Calibrated By: Brian Berls	S#:	102182	Phone: 526-2761	12	<input type="checkbox"/>	Extension

**CALIBRATION STANDARDS USED**

715424	715356	712788	714272				

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-5.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

NOMINAL (STD)	OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION UNITS	AS FOUND (UUT)	MFG. ACCURACY

**COMMENTS**

**INL  
CALIBRATION INPUT DATA**

11/30/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 723413      Mfr: EXTECH      Model: TYPE K      Noun Name: THERMOCOUPLE      Serial #:

Calibration Date: 11/21/2005

**ACTION CODE****AS FOUND**

Next Cal Due Date: 5/21/2007	1	<input type="checkbox"/>	Acceptance Test	1	<input checked="" type="radio"/>	In Tolerance
Charge Level: 4	2	<input type="checkbox"/>	Special Test	2	<input type="radio"/>	Out of Tolerance >1x <2x
Repair/Adj/etc C.L: 0	3	<input checked="" type="checkbox"/>	Calibration to MFG Specs	3	<input type="radio"/>	Out of Tolerance >2x <3x
Material Amount: 0	4	<input type="checkbox"/>	Clean	4	<input type="radio"/>	Out of Tolerance >3x <5x
Charge Number: 100853GSA	5	<input type="checkbox"/>	Limited Calibration	5	<input type="radio"/>	Out of Tolerance >5x
Cal Work Inst ID: 3052U	6	<input type="checkbox"/>	Functional Check	6	<input type="radio"/>	Out of Tolerance-Undetermined
Outside Vendor:	7	<input type="checkbox"/>	Performance Check	7	<input type="checkbox"/>	Inoperative
	8	<input type="checkbox"/>	Modify	8	<input type="checkbox"/>	Damaged
	9	<input type="checkbox"/>	Repair-needs Charge Level	9	<input type="checkbox"/>	Not Used
	10	<input type="checkbox"/>	Other	10	<input type="checkbox"/>	Not Determined
				11	<input type="checkbox"/>	Excessed
				12	<input type="checkbox"/>	Extension

Calibrated By: Chris Dudley

S#: 101142 Phone: 526-2761

**CALIBRATION STANDARDS USED**

370602	723247	722915	722456	717617			

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

NOMINAL (STD)	OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION		MFG. ACCURACY
	UNITS	AS FOUND (UUT)	

**COMMENTS**

INITIAL CALIBRATION  
RECEIVING INSPECTION TEST REFERRAL QA# 115846

**INL**  
**CALIBRATION INPUT DATA**

11/30/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 723413      Mfr: EXTECH      Model: TYPE K      Noun Name: THERMOCOUPLE      Serial #:

Calibration Date: 5/8/2007

**ACTION CODE**

**AS FOUND**

Next Cal Due Date: 11/8/2008

Charge Level: 4

Repair/Adj/etc C.L.: 0

Material Amount: 0

Charge Number: 100853GSA

Cal Work Inst ID: 3052V

Outside Vendor:

1 ☐

2 ☐

3 ☒

4 ☐

5 ☐

6 ☐

7 ☐

8 ☐

9 ☐

10 ☐

Acceptance Test

Special Test

Calibration to MFG Specs

Clean

Limited Calibration

Functional Check

Performance Check

Modify

Repair-needs Charge Level

Other

1 ☒ In Tolerance

2 ☐ Out of Tolerance >1x <2x

3 ☐ Out of Tolerance >2x <3x

4 ☐ Out of Tolerance >3x <5x

5 ☐ Out of Tolerance >5x

6 ☐ Out of Tolerance-Undetermined

7 ☐ Inoperative

8 ☐ Damaged

9 ☐ Not Used

10 ☐ Not Determined

11 ☐ Excessed

12 ☐ Extension

Calibrated By: Brian Berls

S#: 102182 Phone: 526-2761

**CALIBRATION STANDARDS USED**

715424	715356	712788	714272				

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)    20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)    23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)    20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)    23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)    20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:    23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

**OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION**

NOMINAL (STD)

UNITS

AS FOUND (UUT)

MFG. ACCURACY


COMMENTS

**INL  
CALIBRATION INPUT DATA**

11/30/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 723414    Mfr: EXTECH    Model: TYPE K    Noun Name: THERMOCOUPLE    Serial #:

Calibration Date: 11/21/2005

**ACTION CODE**

**AS FOUND**

Next Cal Due Date: 5/21/2007	1	<input type="checkbox"/>	Acceptance Test	1	<input checked="" type="radio"/>	In Tolerance
Charge Level: 4	2	<input type="checkbox"/>	Special Test	2	<input type="radio"/>	Out of Tolerance >1x <2x
Repair/Adj/etc C.L: 0	3	<input checked="" type="checkbox"/>	Calibration to MFG Specs	3	<input type="radio"/>	Out of Tolerance >2x <3x
Material Amount: 0	4	<input type="checkbox"/>	Clean	4	<input type="radio"/>	Out of Tolerance >3x <5x
Charge Number: 100853GSA	5	<input type="checkbox"/>	Limited Calibration	5	<input type="radio"/>	Out of Tolerance >5x
Cal Work Inst ID: 3052U	6	<input type="checkbox"/>	Functional Check	6	<input type="radio"/>	Out of Tolerance-Undetermined
Outside Vendor:	7	<input type="checkbox"/>	Performance Check	7	<input type="checkbox"/>	Inoperative
	8	<input type="checkbox"/>	Modify	8	<input type="checkbox"/>	Damaged
	9	<input type="checkbox"/>	Repair-needs Charge Level	9	<input type="checkbox"/>	Not Used
	10	<input type="checkbox"/>	Other	10	<input type="checkbox"/>	Not Determined
				11	<input type="checkbox"/>	Excessed
Calibrated By: Chris Dudley    S#: 101142    Phone: 526-2761				12	<input type="checkbox"/>	Extension

**CALIBRATION STANDARDS USED**

370602	723247	722456	722915	717617			

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

NOMINAL (STD)	OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION		MFG. ACCURACY
	UNITS	AS FOUND (UUT)	

**COMMENTS**

INITIAL CALIBRATION  
RECEIVING INSPECTION TEST REFERRAL QA# 115846

# INL CALIBRATION INPUT DATA

11/30/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 723414      Mfr: EXTECH      Model: TYPE K      Noun Name: THERMOCOUPLE      Serial #:

Calibration Date: 5/8/2007

**ACTION CODE****AS FOUND**

Next Cal Due Date: 11/8/2008

1 ☐

Acceptance Test

1 ☒ In Tolerance

Charge Level: 4

2 ☐

Special Test

2 ☐ Out of Tolerance >1x <2x

Repair/Adj/etc C.L: 0

3 ☒

Calibration to MFG Specs

3 ☐ Out of Tolerance >2x <3x

Material Amount: 0

4 ☐

Clean

4 ☐ Out of Tolerance >3x <5x

Charge Number: 100853GSA

5 ☐

Limited Calibration

5 ☐ Out of Tolerance >5x

Cal Work Inst ID: 3052V

6 ☐

Functional Check

6 ☐ Out of Tolerance-Undetermined

Outside Vendor:

7 ☐

Performance Check

7 ☐ Inoperative8 ☐

Modify

8 ☐ Damaged9 ☐

Repair-needs Charge Level

9 ☐ Not Used10 ☐

Other

10 ☐ Not Determined11 ☐ Excessed

Calibrated By: Brian Berls

S#: 102182 Phone: 526-2761

12 ☐ Extension**CALIBRATION STANDARDS USED**

715424	715356	712788	714272				

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED  
VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-5.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

**OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION****NOMINAL (STD)****UNITS****AS FOUND (UUT)****MFG. ACCURACY**


**COMMENTS**

**INL  
CALIBRATION INPUT DATA**

11/30/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 723415    Mfr: EXTECH    Model: TYPE K    Noun Name: THERMOCOUPLE    Serial #:

Calibration Date: 11/21/2005

**ACTION CODE**

**AS FOUND**

Next Cal Due Date: 5/21/2007	1	<input type="checkbox"/>	Acceptance Test	1	<input checked="" type="radio"/>	In Tolerance
Charge Level: 4	2	<input type="checkbox"/>	Special Test	2	<input type="radio"/>	Out of Tolerance >1x <2x
Repair/Adj/etc C.L: 0	3	<input checked="" type="checkbox"/>	Calibration to MFG Specs	3	<input type="radio"/>	Out of Tolerance >2x <3x
Material Amount: 0	4	<input type="checkbox"/>	Clean	4	<input type="radio"/>	Out of Tolerance >3x <5x
Charge Number: 100853GSA	5	<input type="checkbox"/>	Limited Calibration	5	<input type="radio"/>	Out of Tolerance >5x
Cal Work Inst ID: 3052U	6	<input type="checkbox"/>	Functional Check	6	<input type="radio"/>	Out of Tolerance-Undetermined
Outside Vendor:	7	<input type="checkbox"/>	Performance Check	7	<input type="checkbox"/>	Inoperative
	8	<input type="checkbox"/>	Modify	8	<input type="checkbox"/>	Damaged
	9	<input type="checkbox"/>	Repair-needs Charge Level	9	<input type="checkbox"/>	Not Used
	10	<input type="checkbox"/>	Other	10	<input type="checkbox"/>	Not Determined
				11	<input type="checkbox"/>	Excessed
Calibrated By: Chris Dudley    S#: 101142    Phone: 526-2761				12	<input type="checkbox"/>	Extension

**CALIBRATION STANDARDS USED**

370602	723247	722915	722456	717617			

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

**OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION**

NOMINAL (STD)

UNITS

AS FOUND (UUT)

MFG. ACCURACY


**COMMENTS**

INITIAL CALIBRATION  
RECEIVING INSPECTION TEST REFERRAL QA# 115846

**INL**  
**CALIBRATION INPUT DATA**

11/30/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 723415	Mfr: EXTECH	Model: TYPE K	Noun Name: THERMOCOUPLE	Serial #:
Calibration Date: 5/8/2007	<b>ACTION CODE</b>		<b>AS FOUND</b>	
Next Cal Due Date: 11/8/2008	1 <input type="checkbox"/>	Acceptance Test	1 <input checked="" type="radio"/>	In Tolerance
Charge Level: 4	2 <input type="checkbox"/>	Special Test	2 <input type="radio"/>	Out of Tolerance >1x <2x
Repair/Adj/etc C.L: 0	3 <input checked="" type="checkbox"/>	Calibration to MFG Specs	3 <input type="radio"/>	Out of Tolerance >2x <3x
Material Amount: 0	4 <input type="checkbox"/>	Clean	4 <input type="radio"/>	Out of Tolerance >3x <5x
Charge Number: 100853GSA	5 <input type="checkbox"/>	Limited Calibration	5 <input type="radio"/>	Out of Tolerance >5x
Cal Work Inst ID: 3052V	6 <input type="checkbox"/>	Functional Check	6 <input type="radio"/>	Out of Tolerance-Undetermined
Outside Vendor:	7 <input type="checkbox"/>	Performance Check	7 <input type="checkbox"/>	Inoperative
	8 <input type="checkbox"/>	Modify	8 <input type="checkbox"/>	Damaged
	9 <input type="checkbox"/>	Repair-needs Charge Level	9 <input type="checkbox"/>	Not Used
	10 <input type="checkbox"/>	Other	10 <input type="checkbox"/>	Not Determined
			11 <input type="checkbox"/>	Excessed
Calibrated By: Brian Berls	S#: 102182	Phone: 526-2761	12 <input type="checkbox"/>	Extension

**CALIBRATION STANDARDS USED**

715424	712788	715356	714272				

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-5.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

NOMINAL (STD)	OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION UNITS	AS FOUND (UUT)	MFG. ACCURACY

COMMENTS



**INL**  
**CALIBRATION INPUT DATA**

11/30/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 723416      Mfr: EXTECH      Model: TYPE K      Noun Name: THERMOCOUPLE      Serial #:

Calibration Date: 11/21/2005

**ACTION CODE****AS FOUND**

Next Cal Due Date: 5/21/2007	1	<input type="checkbox"/>	Acceptance Test	1	<input checked="" type="radio"/>	In Tolerance
Charge Level: 4	2	<input type="checkbox"/>	Special Test	2	<input type="radio"/>	Out of Tolerance >1x <2x
Repair/Adj/etc C.L: 0	3	<input checked="" type="checkbox"/>	Calibration to MFG Specs	3	<input type="radio"/>	Out of Tolerance >2x <3x
Material Amount: 0	4	<input type="checkbox"/>	Clean	4	<input type="radio"/>	Out of Tolerance >3x <5x
Charge Number: 100853GSA	5	<input type="checkbox"/>	Limited Calibration	5	<input type="radio"/>	Out of Tolerance >5x
Cal Work Inst ID: 3052U	6	<input type="checkbox"/>	Functional Check	6	<input type="radio"/>	Out of Tolerance-Undetermined
Outside Vendor:	7	<input type="checkbox"/>	Performance Check	7	<input type="checkbox"/>	Inoperative
	8	<input type="checkbox"/>	Modify	8	<input type="checkbox"/>	Damaged
	9	<input type="checkbox"/>	Repair-needs Charge Level	9	<input type="checkbox"/>	Not Used
	10	<input type="checkbox"/>	Other	10	<input type="checkbox"/>	Not Determined
				11	<input type="checkbox"/>	Excessed
Calibrated By: Chris Dudley	S#: 101142	Phone: 526-2761		12	<input type="checkbox"/>	Extension

**CALIBRATION STANDARDS USED**

370602	717617	723247	722915	722456			

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED  
VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

NOMINAL (STD)	OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION UNITS	AS FOUND (UUT)	MFG. ACCURACY

**COMMENTS**

INITIAL CALIBRATION  
RECEIVING INSPECTION TEST REFERRAL QA# 115846

**INL**  
**CALIBRATION INPUT DATA**

11/30/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 723416      Mfr: EXTECH      Model: TYPE K      Noun Name: THERMOCOUPLE      Serial #:

Calibration Date: 5/8/2007

**ACTION CODE****AS FOUND**

Next Cal Due Date: 11/8/2008	1	<input type="checkbox"/>	Acceptance Test	1	<input checked="" type="radio"/>	In Tolerance
Charge Level: 4	2	<input type="checkbox"/>	Special Test	2	<input type="radio"/>	Out of Tolerance >1x <2x
Repair/Adj/etc C.L: 0	3	<input checked="" type="checkbox"/>	Calibration to MFG Specs	3	<input type="radio"/>	Out of Tolerance >2x <3x
Material Amount: 0	4	<input type="checkbox"/>	Clean	4	<input type="radio"/>	Out of Tolerance >3x <5x
Charge Number: 100853GSA	5	<input type="checkbox"/>	Limited Calibration	5	<input type="radio"/>	Out of Tolerance >5x
Cal Work Inst ID: 3052V	6	<input type="checkbox"/>	Functional Check	6	<input type="radio"/>	Out of Tolerance-Undetermined
Outside Vendor:	7	<input type="checkbox"/>	Performance Check	7	<input type="checkbox"/>	Inoperative
	8	<input type="checkbox"/>	Modify	8	<input type="checkbox"/>	Damaged
	9	<input type="checkbox"/>	Repair-needs Charge Level	9	<input type="checkbox"/>	Not Used
	10	<input type="checkbox"/>	Other	10	<input type="checkbox"/>	Not Determined
				11	<input type="checkbox"/>	Excessed
				12	<input type="checkbox"/>	Extension

Calibrated By: Brian Berls      S#: 102182      Phone: 526-2761

**CALIBRATION STANDARDS USED**

715424	715356	712788	714272				

**STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES**

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

**OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION****NOMINAL (STD)****UNITS****AS FOUND (UUT)****MFG. ACCURACY**


**COMMENTS**

**INL  
CALIBRATION INPUT DATA**

11/30/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 724271      Mfr: EXTECH      Model: TYPE K      Noun Name: THERMOCOUPLE      Serial #:

Calibration Date: 11/16/2006

**ACTION CODE**

**AS FOUND**

Next Cal Due Date: 5/16/2008	1	<input type="checkbox"/>	Acceptance Test	1	<input checked="" type="radio"/>	In Tolerance
Charge Level: 4	2	<input type="checkbox"/>	Special Test	2	<input type="radio"/>	Out of Tolerance >1x <2x
Repair/Adj/etc C.L: 0	3	<input checked="" type="checkbox"/>	Calibration to MFG Specs	3	<input type="radio"/>	Out of Tolerance >2x <3x
Material Amount: 0	4	<input type="checkbox"/>	Clean	4	<input type="radio"/>	Out of Tolerance >3x <5x
Charge Number: 100853GSA	5	<input type="checkbox"/>	Limited Calibration	5	<input type="radio"/>	Out of Tolerance >5x
Cal Work Inst ID: 3052W	6	<input type="checkbox"/>	Functional Check	6	<input type="radio"/>	Out of Tolerance-Undetermined
Outside Vendor:	7	<input type="checkbox"/>	Performance Check	7	<input type="checkbox"/>	Inoperative
	8	<input type="checkbox"/>	Modify	8	<input type="checkbox"/>	Damaged
	9	<input type="checkbox"/>	Repair-needs Charge Level	9	<input type="checkbox"/>	Not Used
	10	<input type="checkbox"/>	Other	10	<input type="checkbox"/>	Not Determined
				11	<input type="checkbox"/>	Excessed
Calibrated By: Brian Berls	S#: 102182	Phone: 526-2761		12	<input type="checkbox"/>	Extension

**CALIBRATION STANDARDS USED**

715424	712788	714272	714273				

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

**OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION**

NOMINAL (STD)	UNITS	AS FOUND (UUT)	MFG. ACCURACY
---------------	-------	----------------	---------------


**COMMENTS**

S&CL discovered due date in ICAL data base does not match due date on unit, due to confusion between S&CL, Receiving inspection and user (Keith Morton). We are updating database and changing the due date from 5-2-08 to 5-16-08 and the calibration date from 11-2-06 to 11-16-06 to match the calibration label on the unit in the user's possession. Unit was visually inspected when received back from receiving inspection on 11-16-06. Upon performing evaluation, there is very minimal risk in changing the due date 14 days later to 5-16-08 to match the label on the unit. Brian Berls 3-22-07

**INL**  
**CALIBRATION INPUT DATA**

12/11/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 724271	Mfr: EXTECH	Model: TYPE K	Noun Name: THERMOCOUPLE	Serial #:
Calibration Date: 12/5/2007	<b>ACTION CODE</b>		<b>AS FOUND</b>	
Next Cal Due Date: 6/5/2009	1 <input type="checkbox"/>	Acceptance Test	1 <input checked="" type="radio"/>	In Tolerance
Charge Level: 4	2 <input type="checkbox"/>	Special Test	2 <input type="radio"/>	Out of Tolerance >1x <2x
Repair/Adj/etc C.L.: 0	3 <input checked="" type="checkbox"/>	Calibration to MFG Specs	3 <input type="radio"/>	Out of Tolerance >2x <3x
Material Amount: 0	4 <input type="checkbox"/>	Clean	4 <input type="radio"/>	Out of Tolerance >3x <5x
Charge Number: 100853GSA	5 <input type="checkbox"/>	Limited Calibration	5 <input type="radio"/>	Out of Tolerance >5x
Cal Work Inst ID: 3052HH	6 <input type="checkbox"/>	Functional Check	6 <input type="radio"/>	Out of Tolerance-Undetermined
Outside Vendor:	7 <input type="checkbox"/>	Performance Check	7 <input type="checkbox"/>	Inoperative
	8 <input type="checkbox"/>	Modify	8 <input type="checkbox"/>	Damaged
	9 <input type="checkbox"/>	Repair-needs Charge Level	9 <input type="checkbox"/>	Not Used
	10 <input type="checkbox"/>	Other	10 <input type="checkbox"/>	Not Determined
			11 <input type="checkbox"/>	Excessed
Calibrated By: Donnie Lindsay	S#: 105138	Phone: 526-2761	12 <input type="checkbox"/>	Extension

**CALIBRATION STANDARDS USED**

715424	715356	714273	717615				

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

NOMINAL (STD)	OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION UNITS	AS FOUND (UUT)	MFG. ACCURACY

**COMMENTS**

S&CL overcheck required.

# INL CALIBRATION INPUT DATA

11/30/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 724272      Mfr: EXTECH      Model: TYPE K      Noun Name: THERMOCOUPLE      Serial #:

Calibration Date: 11/16/2006

**ACTION CODE****AS FOUND**

Next Cal Due Date: 5/16/2008	1	<input type="checkbox"/>	Acceptance Test	1	<input checked="" type="radio"/>	In Tolerance
Charge Level: 4	2	<input type="checkbox"/>	Special Test	2	<input type="radio"/>	Out of Tolerance >1x <2x
Repair/Adj/etc C.L: 0	3	<input checked="" type="checkbox"/>	Calibration to MFG Specs	3	<input type="radio"/>	Out of Tolerance >2x <3x
Material Amount: 0	4	<input type="checkbox"/>	Clean	4	<input type="radio"/>	Out of Tolerance >3x <5x
Charge Number: 100853GSA	5	<input type="checkbox"/>	Limited Calibration	5	<input type="radio"/>	Out of Tolerance >5x
Cal Work Inst ID: 3052W	6	<input type="checkbox"/>	Functional Check	6	<input type="radio"/>	Out of Tolerance-Undetermined
Outside Vendor:	7	<input type="checkbox"/>	Performance Check	7	<input type="checkbox"/>	Inoperative
	8	<input type="checkbox"/>	Modify	8	<input type="checkbox"/>	Damaged
	9	<input type="checkbox"/>	Repair-needs Charge Level	9	<input type="checkbox"/>	Not Used
	10	<input type="checkbox"/>	Other	10	<input type="checkbox"/>	Not Determined
				11	<input type="checkbox"/>	Excessed
Calibrated By: Brian Berls      S#: 102182      Phone: 526-2761				12	<input type="checkbox"/>	Extension

**CALIBRATION STANDARDS USED**

715424	714272	712788	714273				

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

NOMINAL (STD)	OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION		MFG. ACCURACY
	UNITS	AS FOUND (UUT)	

**COMMENTS**

S&CL discovered due date in ICAL data base does not match due date on unit, due to confusion between S&CL, Receiving inspection and user (Keith Morton). We are updating database and changing the due date from 5-2-08 to 5-16-08 and the calibration date from 11-2-06 to 11-16-06 to match the calibration label on the unit in the user's possession. Unit was visually inspected when received back from receiving inspection on 11-16-06. Upon performing evaluation, there is very minimal risk in changing the due date 14 days later to 5-16-08 to match the label on the unit. Brian Berls 3-22-07

# INL CALIBRATION INPUT DATA

12/11/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 724272    Mfr: EXTECH    Model: TYPE K    Noun Name: THERMOCOUPLE    Serial #:

Calibration Date: 12/5/2007

**ACTION CODE****AS FOUND**

Next Cal Due Date: 6/5/2009	1	<input type="checkbox"/>	Acceptance Test	1	<input checked="" type="radio"/>	In Tolerance
Charge Level: 4	2	<input type="checkbox"/>	Special Test	2	<input type="radio"/>	Out of Tolerance >1x <2x
Repair/Adj/etc C.L: 0	3	<input checked="" type="checkbox"/>	Calibration to MFG Specs	3	<input type="radio"/>	Out of Tolerance >2x <3x
Material Amount: 0	4	<input type="checkbox"/>	Clean	4	<input type="radio"/>	Out of Tolerance >3x <5x
Charge Number: 100853GSA	5	<input type="checkbox"/>	Limited Calibration	5	<input type="radio"/>	Out of Tolerance >5x
Cal Work Inst ID: 3052HH	6	<input type="checkbox"/>	Functional Check	6	<input type="radio"/>	Out of Tolerance-Undetermined
Outside Vendor:	7	<input type="checkbox"/>	Performance Check	7	<input type="checkbox"/>	Inoperative
	8	<input type="checkbox"/>	Modify	8	<input type="checkbox"/>	Damaged
	9	<input type="checkbox"/>	Repair-needs Charge Level	9	<input type="checkbox"/>	Not Used
	10	<input type="checkbox"/>	Other	10	<input type="checkbox"/>	Not Determined
				11	<input type="checkbox"/>	Excessed
				12	<input type="checkbox"/>	Extension

Calibrated By: Donnie Lindsay    S#: 105138    Phone: 526-2761

**CALIBRATION STANDARDS USED**

715424	715356	714273	717615				

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-5.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

**OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION****NOMINAL (STD)****UNITS****AS FOUND (UUT)****MFG. ACCURACY**


**COMMENTS**

S&amp;CL overcheck required.

**INL  
CALIBRATION INPUT DATA**

11/30/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 724273    Mfr: EXTECH    Model: TYPE K    Noun Name: THERMOCOUPLE    Serial #:

Calibration Date: 11/16/2006

**ACTION CODE**

**AS FOUND**

Next Cal Due Date: 5/16/2008	1	<input type="checkbox"/>	Acceptance Test	1	<input checked="" type="radio"/>	In Tolerance
Charge Level: 4	2	<input type="checkbox"/>	Special Test	2	<input type="radio"/>	Out of Tolerance >1x <2x
Repair/Adj/etc C.L: 0	3	<input checked="" type="checkbox"/>	Calibration to MFG Specs	3	<input type="radio"/>	Out of Tolerance >2x <3x
Material Amount: 0	4	<input type="checkbox"/>	Clean	4	<input type="radio"/>	Out of Tolerance >3x <5x
Charge Number: 100853gsa	5	<input type="checkbox"/>	Limited Calibration	5	<input type="radio"/>	Out of Tolerance >5x
Cal Work Inst ID: 3052W	6	<input type="checkbox"/>	Functional Check	6	<input type="radio"/>	Out of Tolerance-Undetermined
Outside Vendor:	7	<input type="checkbox"/>	Performance Check	7	<input type="checkbox"/>	Inoperative
	8	<input type="checkbox"/>	Modify	8	<input type="checkbox"/>	Damaged
	9	<input type="checkbox"/>	Repair-needs Charge Level	9	<input type="checkbox"/>	Not Used
	10	<input type="checkbox"/>	Other	10	<input type="checkbox"/>	Not Determined
				11	<input type="checkbox"/>	Excessed
Calibrated By: Brian Berls    S#: 102182    Phone: 526-2761				12	<input type="checkbox"/>	Extension

**CALIBRATION STANDARDS USED**

715424	712788	714272	370602				

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-5.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

**OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION**

NOMINAL (STD)

UNITS

AS FOUND (UUT)

MFG. ACCURACY


COMMENTS

**INL**  
**CALIBRATION INPUT DATA**

12/11/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 724273    Mfr: EXTECH    Model: TYPE K    Noun Name: THERMOCOUPLE    Serial #:

Calibration Date: 12/5/2007

**ACTION CODE****AS FOUND**

Next Cal Due Date: 6/5/2009	1	<input type="checkbox"/>	Acceptance Test	1	<input checked="" type="radio"/>	In Tolerance
Charge Level: 4	2	<input type="checkbox"/>	Special Test	2	<input type="radio"/>	Out of Tolerance >1x <2x
Repair/Adj/etc C.L: 0	3	<input checked="" type="checkbox"/>	Calibration to MFG Specs	3	<input type="radio"/>	Out of Tolerance >2x <3x
Material Amount: 0	4	<input type="checkbox"/>	Clean	4	<input type="radio"/>	Out of Tolerance >3x <5x
Charge Number: 100853GSA	5	<input type="checkbox"/>	Limited Calibration	5	<input type="radio"/>	Out of Tolerance >5x
Cal Work Inst ID: 3052HH	6	<input type="checkbox"/>	Functional Check	6	<input type="radio"/>	Out of Tolerance-Undetermined
Outside Vendor:	7	<input type="checkbox"/>	Performance Check	7	<input type="checkbox"/>	Inoperative
	8	<input type="checkbox"/>	Modify	8	<input type="checkbox"/>	Damaged
	9	<input type="checkbox"/>	Repair-needs Charge Level	9	<input type="checkbox"/>	Not Used
	10	<input type="checkbox"/>	Other	10	<input type="checkbox"/>	Not Determined
				11	<input type="checkbox"/>	Excessed
Calibrated By: Donnie Lindsay    S#: 105138    Phone: 526-2761				12	<input type="checkbox"/>	Extension

**CALIBRATION STANDARDS USED**

715356	715424	714273	717615				

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

**OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION****NOMINAL (STD)****UNITS****AS FOUND (UUT)****MFG. ACCURACY**


**COMMENTS**

S&amp;CL overcheck required.



# INL CALIBRATION INPUT DATA

11/30/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 724274    Mfr: EXTECH    Model: TYPE K    Noun Name: THERMOCOUPLE    Serial #:

Calibration Date: 11/16/2006

## ACTION CODE

## AS FOUND

Next Cal Due Date: 5/16/2008	1	<input type="checkbox"/>	Acceptance Test	1	<input checked="" type="radio"/>	In Tolerance
Charge Level: 4	2	<input type="checkbox"/>	Special Test	2	<input type="radio"/>	Out of Tolerance >1x <2x
Repair/Adj/etc C.L: 0	3	<input checked="" type="checkbox"/>	Calibration to MFG Specs	3	<input type="radio"/>	Out of Tolerance >2x <3x
Material Amount: 0	4	<input type="checkbox"/>	Clean	4	<input type="radio"/>	Out of Tolerance >3x <5x
Charge Number: 100853gsa	5	<input type="checkbox"/>	Limited Calibration	5	<input type="radio"/>	Out of Tolerance >5x
Cal Work Inst ID: 3052W	6	<input type="checkbox"/>	Functional Check	6	<input type="radio"/>	Out of Tolerance-Undetermined
Outside Vendor:	7	<input type="checkbox"/>	Performance Check	7	<input type="checkbox"/>	Inoperative
	8	<input type="checkbox"/>	Modify	8	<input type="checkbox"/>	Damaged
	9	<input type="checkbox"/>	Repair-needs Charge Level	9	<input type="checkbox"/>	Not Used
	10	<input type="checkbox"/>	Other	10	<input type="checkbox"/>	Not Determined
				11	<input type="checkbox"/>	Excessed
				12	<input type="checkbox"/>	Extension

Calibrated By: Brian Berls    S#: 102182    Phone: 526-2761

## CALIBRATION STANDARDS USED

715424	712788	714272	370602				

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

## LABORATORY TEMPERATURE AND HUMIDITY

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-5.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

NOMINAL (STD)	OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION		MFG. ACCURACY
	UNITS	AS FOUND (UUT)	

## COMMENTS

**INL**  
**CALIBRATION INPUT DATA**

12/11/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 724274    Mfr: EXTECH    Model: TYPE K    Noun Name: THERMOCOUPLE    Serial #:

Calibration Date: 12/5/2007

**ACTION CODE****AS FOUND**

Next Cal Due Date: 6/5/2009	1	<input type="checkbox"/>	Acceptance Test	1	<input checked="" type="radio"/>	In Tolerance
Charge Level: 4	2	<input type="checkbox"/>	Special Test	2	<input type="radio"/>	Out of Tolerance >1x <2x
Repair/Adj/etc C.L: 0	3	<input checked="" type="checkbox"/>	Calibration to MFG Specs	3	<input type="radio"/>	Out of Tolerance >2x <3x
Material Amount: 0	4	<input type="checkbox"/>	Clean	4	<input type="radio"/>	Out of Tolerance >3x <5x
Charge Number: 100853GSA	5	<input type="checkbox"/>	Limited Calibration	5	<input type="radio"/>	Out of Tolerance >5x
Cal Work Inst ID: 3052HH	6	<input type="checkbox"/>	Functional Check	6	<input type="radio"/>	Out of Tolerance-Undetermined
Outside Vendor:	7	<input type="checkbox"/>	Performance Check	7	<input type="checkbox"/>	Inoperative
	8	<input type="checkbox"/>	Modify	8	<input type="checkbox"/>	Damaged
	9	<input type="checkbox"/>	Repair-needs Charge Level	9	<input type="checkbox"/>	Not Used
	10	<input type="checkbox"/>	Other	10	<input type="checkbox"/>	Not Determined
				11	<input type="checkbox"/>	Excessed
				12	<input type="checkbox"/>	Extension

Calibrated By: Donnie Lindsay    S#: 105138    Phone: 526-2761

**CALIBRATION STANDARDS USED**

715356	715424	714273	717615				

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

**OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION**

NOMINAL (STD)	UNITS	AS FOUND (UUT)	MFG. ACCURACY

**COMMENTS**

S&amp;CL overcheck required.

# **THERMOMETERS**

**INL**  
**CALIBRATION INPUT DATA**

11/29/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 721279      Mfr: RAYTEK      Model: MX6      Noun Name: INFRARED THERMOMETER      Serial #: 297914-101-0001

Calibration Date: 9/3/2003 5:10:04 PM

**ACTION CODE****AS FOUND**

Next Cal Due Date: 9/3/2004

Charge Level: 6

Repair/Adj/etc C.L: 0

Material Amount: 0

Charge Number: 100348027

Cal Work Inst ID:

Outside Vendor:

- |    |                                     |                           |    |                                  |                               |
|----|-------------------------------------|---------------------------|----|----------------------------------|-------------------------------|
| 1  | <input type="checkbox"/>            | Acceptance Test           | 1  | <input checked="" type="radio"/> | In Tolerance                  |
| 2  | <input type="checkbox"/>            | Special Test              | 2  | <input type="radio"/>            | Out of Tolerance >1x <2x      |
| 3  | <input type="checkbox"/>            | Calibration to MFG Specs  | 3  | <input type="radio"/>            | Out of Tolerance >2x <3x      |
| 4  | <input type="checkbox"/>            | Clean                     | 4  | <input type="radio"/>            | Out of Tolerance >3x <5x      |
| 5  | <input checked="" type="checkbox"/> | Limited Calibration       | 5  | <input type="radio"/>            | Out of Tolerance >5x          |
| 6  | <input type="checkbox"/>            | Functional Check          | 6  | <input type="radio"/>            | Out of Tolerance-Undetermined |
| 7  | <input type="checkbox"/>            | Performance Check         | 7  | <input type="checkbox"/>         | Inoperative                   |
| 8  | <input type="checkbox"/>            | Modify                    | 8  | <input type="checkbox"/>         | Damaged                       |
| 9  | <input type="checkbox"/>            | Repair-needs Charge Level | 9  | <input type="checkbox"/>         | Not Used                      |
| 10 | <input type="checkbox"/>            | Other                     | 10 | <input type="checkbox"/>         | Not Determined                |
|    |                                     |                           | 11 | <input type="checkbox"/>         | Excessed                      |
|    |                                     |                           | 12 | <input type="checkbox"/>         | Extension                     |

Calibrated By: Chris Dudley

S#: 101142 Phone: 526-2761

**CALIBRATION STANDARDS USED**

719721		715424					

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

NOMINAL (STD)	OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION		MFG. ACCURACY
	UNITS	AS FOUND (UUT)	

**COMMENTS**

INITIAL CALIBRATION  
S&CL OVERCHECK REQUIRED  
LIMITED: -30 TO 420 DEG. C.

**INL**  
**CALIBRATION INPUT DATA**

11/29/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 721281      Mfr: RAYTEK      Model: TYPE K      Noun Name: THERMOCOUPLE PROBE      Serial #:

Calibration Date: 9/3/2003 5:16:06 PM

**ACTION CODE****AS FOUND**

Next Cal Due Date: 9/3/2004

Charge Level: 4

Repair/Adj/etc C.L: 0

Material Amount: 0

Charge Number: 100348027

Cal Work Inst ID: 30520

Outside Vendor:

- |    |                                     |                           |    |                                  |                               |
|----|-------------------------------------|---------------------------|----|----------------------------------|-------------------------------|
| 1  | <input type="checkbox"/>            | Acceptance Test           | 1  | <input checked="" type="radio"/> | In Tolerance                  |
| 2  | <input type="checkbox"/>            | Special Test              | 2  | <input type="radio"/>            | Out of Tolerance >1x <2x      |
| 3  | <input type="checkbox"/>            | Calibration to MFG Specs  | 3  | <input type="radio"/>            | Out of Tolerance >2x <3x      |
| 4  | <input type="checkbox"/>            | Clean                     | 4  | <input type="radio"/>            | Out of Tolerance >3x <5x      |
| 5  | <input checked="" type="checkbox"/> | Limited Calibration       | 5  | <input type="radio"/>            | Out of Tolerance >5x          |
| 6  | <input type="checkbox"/>            | Functional Check          | 6  | <input type="radio"/>            | Out of Tolerance-Undetermined |
| 7  | <input type="checkbox"/>            | Performance Check         | 7  | <input type="checkbox"/>         | Inoperative                   |
| 8  | <input type="checkbox"/>            | Modify                    | 8  | <input type="checkbox"/>         | Damaged                       |
| 9  | <input type="checkbox"/>            | Repair-needs Charge Level | 9  | <input type="checkbox"/>         | Not Used                      |
| 10 | <input type="checkbox"/>            | Other                     | 10 | <input type="checkbox"/>         | Not Determined                |
|    |                                     |                           | 11 | <input type="checkbox"/>         | Excessed                      |
|    |                                     |                           | 12 | <input type="checkbox"/>         | Extension                     |

Calibrated By: Chris Dudley

S#: 101142 Phone: 526-2761

**CALIBRATION STANDARDS USED**

715424	370602	717617	717614	714272			

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

**OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION****NOMINAL (STD)****UNITS****AS FOUND (UUT)****MFG. ACCURACY**


**COMMENTS**

INITIAL CALIBRATION  
S&CL OVERCHECK REQUIRED  
LIMITED: 0 TO 220 DEG. C.

# **INL CALIBRATION INPUT DATA**

11/29/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: WZD1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 721279    Mfr: RAYTEK    Model: MX6    Noun Name: INFRARED THERMOMETER    Serial #: 297914-101-0001

Calibration Date: 2/7/2005

**ACTION CODE****AS FOUND**

Next Cal Due Date: 2/7/2006

Charge Level: 6

Repair/Adj/etc C.L: 0

Material Amount: 0

Charge Number: 100853GSA

Cal Work Inst ID: 3124E

Outside Vendor:

- |    |                                     |                           |    |                                  |                               |
|----|-------------------------------------|---------------------------|----|----------------------------------|-------------------------------|
| 1  | <input type="checkbox"/>            | Acceptance Test           | 1  | <input checked="" type="radio"/> | In Tolerance                  |
| 2  | <input type="checkbox"/>            | Special Test              | 2  | <input type="radio"/>            | Out of Tolerance >1x <2x      |
| 3  | <input type="checkbox"/>            | Calibration to MFG Specs  | 3  | <input type="radio"/>            | Out of Tolerance >2x <3x      |
| 4  | <input type="checkbox"/>            | Clean                     | 4  | <input type="radio"/>            | Out of Tolerance >3x <5x      |
| 5  | <input checked="" type="checkbox"/> | Limited Calibration       | 5  | <input type="radio"/>            | Out of Tolerance >5x          |
| 6  | <input type="checkbox"/>            | Functional Check          | 6  | <input type="radio"/>            | Out of Tolerance-Undetermined |
| 7  | <input type="checkbox"/>            | Performance Check         | 7  | <input type="checkbox"/>         | Inoperative                   |
| 8  | <input type="checkbox"/>            | Modify                    | 8  | <input type="checkbox"/>         | Damaged                       |
| 9  | <input type="checkbox"/>            | Repair-needs Charge Level | 9  | <input type="checkbox"/>         | Not Used                      |
| 10 | <input type="checkbox"/>            | Other                     | 10 | <input type="checkbox"/>         | Not Determined                |
|    |                                     |                           | 11 | <input type="checkbox"/>         | Excessed                      |
|    |                                     |                           | 12 | <input type="checkbox"/>         | Extension                     |

Calibrated By: Chris Dudley    S#: 101142    Phone: 526-2761

**CALIBRATION STANDARDS USED**

370602	722456						

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

**OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION****NOMINAL (STD)****UNITS****AS FOUND (UUT)****MFG. ACCURACY**


**COMMENTS**

LIMITED: -30 TO 420 DEG. C.

**INL  
CALIBRATION INPUT DATA**

11/29/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 721281      Mfr: RAYTEK      Model: TYPE K      Noun Name: THERMOCOUPLE PROBE      Serial #:

Calibration Date: 2/7/2005	ACTION CODE		AS FOUND
Next Cal Due Date: 2/7/2006	1	<input type="checkbox"/> Acceptance Test	1 <input checked="" type="radio"/> In Tolerance
Charge Level: 4	2	<input type="checkbox"/> Special Test	2 <input type="radio"/> Out of Tolerance >1x <2x
Repair/Adj/etc C.L: 0	3	<input checked="" type="checkbox"/> Calibration to MFG Specs	3 <input type="radio"/> Out of Tolerance >2x <3x
Material Amount: 0	4	<input type="checkbox"/> Clean	4 <input type="radio"/> Out of Tolerance >3x <5x
Charge Number: 100853GSA	5	<input checked="" type="checkbox"/> Limited Calibration	5 <input type="radio"/> Out of Tolerance >5x
Cal Work Inst ID: 3052S	6	<input type="checkbox"/> Functional Check	6 <input type="radio"/> Out of Tolerance-Undetermined
Outside Vendor:	7	<input type="checkbox"/> Performance Check	7 <input type="checkbox"/> Inoperative
	8	<input type="checkbox"/> Modify	8 <input type="checkbox"/> Damaged
	9	<input type="checkbox"/> Repair-needs Charge Level	9 <input type="checkbox"/> Not Used
	10	<input type="checkbox"/> Other	10 <input type="checkbox"/> Not Determined
			11 <input type="checkbox"/> Excessed
			12 <input type="checkbox"/> Extension

Calibrated By: Chris Dudley      S#: 101142      Phone: 526-2761

**CALIBRATION STANDARDS USED**

715424	712788	714273	714272				

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

NOMINAL (STD)	OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION		MFG. ACCURACY
	UNITS	AS FOUND (UUT)	

**COMMENTS**

LIMITED: 0 TO 220 DEG. C.

**INL  
CALIBRATION INPUT DATA**

11/29/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 721279    Mfr: RAYTEK    Model: MX6    Noun Name: INFRARED THERMOMETER    Serial #: 297914-101-0001

Calibration Date: 1/25/2006

**ACTION CODE****AS FOUND**

Next Cal Due Date: 1/25/2007	1	<input type="checkbox"/>	Acceptance Test	1	<input checked="" type="radio"/>	In Tolerance
Charge Level: 6	2	<input type="checkbox"/>	Special Test	2	<input type="radio"/>	Out of Tolerance >1x <2x
Repair/Adj/etc C.L.: 0	3	<input type="checkbox"/>	Calibration to MFG Specs	3	<input type="radio"/>	Out of Tolerance >2x <3x
Material Amount: 0	4	<input type="checkbox"/>	Clean	4	<input type="radio"/>	Out of Tolerance >3x <5x
Charge Number: 100853GSA	5	<input checked="" type="checkbox"/>	Limited Calibration	5	<input type="radio"/>	Out of Tolerance >5x
Cal Work Inst ID: 31241	6	<input type="checkbox"/>	Functional Check	6	<input type="radio"/>	Out of Tolerance-Undetermined
Outside Vendor:	7	<input type="checkbox"/>	Performance Check	7	<input type="checkbox"/>	Inoperative
	8	<input type="checkbox"/>	Modify	8	<input type="checkbox"/>	Damaged
	9	<input type="checkbox"/>	Repair-needs Charge Level	9	<input type="checkbox"/>	Not Used
	10	<input type="checkbox"/>	Other	10	<input type="checkbox"/>	Not Determined
				11	<input type="checkbox"/>	Excessed
Calibrated By: Brian Berls				12	<input type="checkbox"/>	Extension

**CALIBRATION STANDARDS USED**

370602	717617						

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-5.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

**OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION****NOMINAL (STD)****UNITS****AS FOUND (UUT)****MFG. ACCURACY**


**COMMENTS**

LIMITED: -30 TO 420 DEG. C.



**INL  
CALIBRATION INPUT DATA**

11/29/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 721281      Mfr: RAYTEK      Model: TYPE K      Noun Name: THERMOCOUPLE PROBE      Serial #:

Calibration Date: 1/25/2006

**ACTION CODE**

**AS FOUND**

Next Cal Due Date: 1/25/2007

Charge Level: 4

Repair/Adj/etc C.L: 0

Material Amount: 0

Charge Number: 100853GSA

Cal Work Inst ID: 3052V

Outside Vendor:

- |    |                                     |                           |
|----|-------------------------------------|---------------------------|
| 1  | <input type="checkbox"/>            | Acceptance Test           |
| 2  | <input type="checkbox"/>            | Special Test              |
| 3  | <input type="checkbox"/>            | Calibration to MFG Specs  |
| 4  | <input type="checkbox"/>            | Clean                     |
| 5  | <input checked="" type="checkbox"/> | Limited Calibration       |
| 6  | <input type="checkbox"/>            | Functional Check          |
| 7  | <input type="checkbox"/>            | Performance Check         |
| 8  | <input type="checkbox"/>            | Modify                    |
| 9  | <input type="checkbox"/>            | Repair-needs Charge Level |
| 10 | <input type="checkbox"/>            | Other                     |

- |    |                                  |                               |
|----|----------------------------------|-------------------------------|
| 1  | <input checked="" type="radio"/> | In Tolerance                  |
| 2  | <input type="radio"/>            | Out of Tolerance >1x <2x      |
| 3  | <input type="radio"/>            | Out of Tolerance >2x <3x      |
| 4  | <input type="radio"/>            | Out of Tolerance >3x <5x      |
| 5  | <input type="radio"/>            | Out of Tolerance >5x          |
| 6  | <input type="radio"/>            | Out of Tolerance-Undetermined |
| 7  | <input type="checkbox"/>         | Inoperative                   |
| 8  | <input type="checkbox"/>         | Damaged                       |
| 9  | <input type="checkbox"/>         | Not Used                      |
| 10 | <input type="checkbox"/>         | Not Determined                |
| 11 | <input type="checkbox"/>         | Excessed                      |
| 12 | <input type="checkbox"/>         | Extension                     |

Calibrated By: Brian Berls      S#: 102182      Phone: 526-2761

**CALIBRATION STANDARDS USED**

370602	717617						

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

NOMINAL (STD)	OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION UNITS	AS FOUND (UUT)	MFG. ACCURACY
---------------	---	----------------	---------------


**COMMENTS**

LIMITED: 0 TO 220 DEG. C.

**INL  
CALIBRATION INPUT DATA**

11/29/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 721279    Mfr: RAYTEK    Model: MX6    Noun Name: INFRARED THERMOMETER    Serial #: 297914-101-0001

Calibration Date: 1/18/2007

**ACTION CODE****AS FOUND**

Next Cal Due Date: 1/18/2008	1	<input type="checkbox"/>	Acceptance Test	1	<input checked="" type="radio"/>	In Tolerance
Charge Level: 6	2	<input type="checkbox"/>	Special Test	2	<input type="radio"/>	Out of Tolerance >1x <2x
Repair/Adj/etc C.L.: 0	3	<input type="checkbox"/>	Calibration to MFG Specs	3	<input type="radio"/>	Out of Tolerance >2x <3x
Material Amount: 0	4	<input type="checkbox"/>	Clean	4	<input type="radio"/>	Out of Tolerance >3x <5x
Charge Number: 100853GSA	5	<input checked="" type="checkbox"/>	Limited Calibration	5	<input type="radio"/>	Out of Tolerance >5x
Cal Work Inst ID: 3124J	6	<input type="checkbox"/>	Functional Check	6	<input type="radio"/>	Out of Tolerance-Undetermined
Outside Vendor:	7	<input type="checkbox"/>	Performance Check	7	<input type="checkbox"/>	Inoperative
	8	<input type="checkbox"/>	Modify	8	<input type="checkbox"/>	Damaged
	9	<input type="checkbox"/>	Repair-needs Charge Level	9	<input type="checkbox"/>	Not Used
	10	<input type="checkbox"/>	Other	10	<input type="checkbox"/>	Not Determined
				11	<input type="checkbox"/>	Excessed
Calibrated By: Brian Berls    S#: 102182    Phone: 526-2761				12	<input type="checkbox"/>	Extension

**CALIBRATION STANDARDS USED**

722456	370602	715424	712788				

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

NOMINAL (STD)	OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION UNITS	AS FOUND (UUT)	MFG. ACCURACY

**COMMENTS**

LIMITED: -30 TO 420 DEG. C.

**INL**  
**CALIBRATION INPUT DATA**

11/29/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 721281    Mfr: RAYTEK    Model: TYPE K    Noun Name: THERMOCOUPLE PROBE    Serial #:

Calibration Date: 1/18/2007

**ACTION CODE****AS FOUND**

Next Cal Due Date: 1/18/2008	1	<input type="checkbox"/>	Acceptance Test	1	<input checked="" type="radio"/>	In Tolerance
Charge Level: 4	2	<input type="checkbox"/>	Special Test	2	<input type="radio"/>	Out of Tolerance >1x <2x
Repair/Adj/etc C.L.: 0	3	<input type="checkbox"/>	Calibration to MFG Specs	3	<input type="radio"/>	Out of Tolerance >2x <3x
Material Amount: 0	4	<input type="checkbox"/>	Clean	4	<input type="radio"/>	Out of Tolerance >3x <5x
Charge Number: 100853GSA	5	<input checked="" type="checkbox"/>	Limited Calibration	5	<input type="radio"/>	Out of Tolerance >5x
Cal Work Inst ID: 3052V	6	<input type="checkbox"/>	Functional Check	6	<input type="radio"/>	Out of Tolerance-Undetermined
Outside Vendor:	7	<input type="checkbox"/>	Performance Check	7	<input type="checkbox"/>	Inoperative
	8	<input type="checkbox"/>	Modify	8	<input type="checkbox"/>	Damaged
	9	<input type="checkbox"/>	Repair-needs Charge Level	9	<input type="checkbox"/>	Not Used
	10	<input type="checkbox"/>	Other	10	<input type="checkbox"/>	Not Determined
				11	<input type="checkbox"/>	Excessed
Calibrated By: Brian Berls				12	<input type="checkbox"/>	Extension

**CALIBRATION STANDARDS USED**

722456	370602	715424	712788	714272	714273		

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-5.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

NOMINAL (STD)	OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION UNITS	AS FOUND (UUT)	MFG. ACCURACY

**COMMENTS**

LIMITED: 0 TO 220 DEG. C.

JNL  
CALIBRATION INPUT DATA

12/18/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 721279    Mfr: RAYTEK    Model: MX6    Noun Name: INFRARED THERMOMETER    Serial #: 297914-101-0001

			ACTION CODE			AS FOUND
Calibration Date: 12/18/2007	1	<input type="checkbox"/>	Acceptance Test	1	<input checked="" type="radio"/>	In Tolerance
Next Cal Due Date: 12/18/2008	2	<input type="checkbox"/>	Special Test	2	<input type="checkbox"/>	Out of Tolerance >1x <2x
Charge Level: 6	3	<input type="checkbox"/>	Calibration to MFG Specs	3	<input type="checkbox"/>	Out of Tolerance >2x <3x
Repair/Adj/etc C.L.: 0	4	<input type="checkbox"/>	Clean	4	<input type="checkbox"/>	Out of Tolerance >3x <5x
Material Amount: 0	5	<input checked="" type="checkbox"/>	Limited Calibration	5	<input type="checkbox"/>	Out of Tolerance >5x
Charge Number: 100853GSA	6	<input type="checkbox"/>	Functional Check	6	<input type="checkbox"/>	Out of Tolerance-Undetermined
Cal Work Inst ID: 3124N	7	<input type="checkbox"/>	Performance Check	7	<input type="checkbox"/>	Inoperative
Outside Vendor:	8	<input type="checkbox"/>	Modify	8	<input type="checkbox"/>	Damaged
	9	<input type="checkbox"/>	Repair-needs Charge Level	9	<input type="checkbox"/>	Not Used
	10	<input type="checkbox"/>	Other	10	<input type="checkbox"/>	Not Determined
				11	<input type="checkbox"/>	Excessed
Calibrated By: Brian Berls    S#: 102182    Phone: 526-2761				12	<input type="checkbox"/>	Extension

**CALIBRATION STANDARDS USED**

370602	722456						

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

NOMINAL (STD)	OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION UNITS	AS FOUND (UUT)	MFG. ACCURACY

COMMENTS

LIMITED: -30 TO 420 DEG. C.

**INL**  
**CALIBRATION INPUT DATA**

12/18/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 721281      Mfr: RAYTEK      Model: TYPE K      Noun Name: THERMOCOUPLE PROBE      Serial #:

Calibration Date: 12/18/2007

**ACTION CODE**

**AS FOUND**

Next Cal Due Date: 12/18/2008

Charge Level: 4

Repair/Adj/etc C.L.: 0

Material Amount: 0

Charge Number: 100853GSA

Cal Work Inst ID: 3052HH

Outside Vendor:

1 ☐

2 ☐

3 ☐

4 ☐

5 ☒

6 ☐

7 ☐

8 ☐

9 ☐

10 ☐

Acceptance Test

Special Test

Calibration to MFG Specs

Clean

Limited Calibration

Functional Check

Performance Check

Modify

Repair-needs Charge Level

Other

1 ☒ In Tolerance

2 ☐ Out of Tolerance >1x <2x

3 ☐ Out of Tolerance >2x <3x

4 ☐ Out of Tolerance >3x <5x

5 ☐ Out of Tolerance >5x

6 ☐ Out of Tolerance-Undetermined

7 ☐ Inoperative

8 ☐ Damaged

9 ☐ Not Used

10 ☐ Not Determined

11 ☐ Excessed

12 ☐ Extension

Calibrated By: Brian Berls

S#: 102182 Phone: 526-2761

**CALIBRATION STANDARDS USED**

715424	714272	717615					

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-5.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

**OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION**

NOMINAL (STD)

UNITS

AS FOUND (UUT)

MFG. ACCURACY

\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

COMMENTS

LIMITED: 0 TO 220 DEG. C.

# **WEIGHTS**

**INL**  
**CALIBRATION INPUT DATA**

11/30/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 725131    Mfr: BEA    Model: 1080.88 LB    Noun Name: WEIGHT    Serial #: WEIGHT A

Calibration Date: 10/10/2007

**ACTION CODE**

**AS FOUND**

Next Cal Due Date: 4/10/2009	1	<input type="checkbox"/>	Acceptance Test	1	<input checked="" type="radio"/>	In Tolerance
Charge Level: 8	2	<input type="checkbox"/>	Special Test	2	<input type="radio"/>	Out of Tolerance >1x <2x
Repair/Adj/etc C.L: 0	3	<input checked="" type="checkbox"/>	Calibration to MFG Specs	3	<input type="radio"/>	Out of Tolerance >2x <3x
Material Amount: 0	4	<input type="checkbox"/>	Clean	4	<input type="radio"/>	Out of Tolerance >3x <5x
Charge Number: 100853GSA	5	<input type="checkbox"/>	Limited Calibration	5	<input type="radio"/>	Out of Tolerance >5x
Cal Work Inst ID: 3014P	6	<input type="checkbox"/>	Functional Check	6	<input type="radio"/>	Out of Tolerance-Undetermined
Outside Vendor:	7	<input type="checkbox"/>	Performance Check	7	<input type="checkbox"/>	Inoperative
	8	<input type="checkbox"/>	Modify	8	<input type="checkbox"/>	Damaged
	9	<input type="checkbox"/>	Repair-needs Charge Level	9	<input type="checkbox"/>	Not Used
	10	<input type="checkbox"/>	Other	10	<input type="checkbox"/>	Not Determined
				11	<input type="checkbox"/>	Excessed
				12	<input type="checkbox"/>	Extension

Calibrated By: Steve Palmer    S#: 56710    Phone: 526-2761

**CALIBRATION STANDARDS USED**

344008							

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

NOMINAL (STD)	OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION		MFG. ACCURACY
	UNITS	AS FOUND (UUT)	

**COMMENTS**

Nominal: 1080.882 lb

Tolerance: +/- 3.00 lb

AS-found data: 1080.882 lb

Uncertainty of standards used: +/- 52.0 g (+/- 0.115 lb)

Note: Nominal was determined by the initial calibration value. All following calibrations will be compared to the nominal (initial calibration value).

**INL  
CALIBRATION INPUT DATA**

11/30/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 725132    Mfr: BEA    Model: 534.34 LB    Noun Name: WEIGHT    Serial #: WEIGHT B

Calibration Date: 10/10/2007

**ACTION CODE****AS FOUND**

Next Cal Due Date: 4/10/2009

Charge Level: 6

Repair/Adj/etc C.L: 0

Material Amount: 0

Charge Number: 100853GSA

Cal Work Inst ID: 3014P

Outside Vendor:

1 ☐ Acceptance Test2 ☐ Special Test3 ☒ Calibration to MFG Specs4 ☐ Clean5 ☐ Limited Calibration6 ☐ Functional Check7 ☐ Performance Check8 ☐ Modify9 ☐ Repair-needs Charge Level10 ☐ Other1 ☒ In Tolerance2 ☐ Out of Tolerance >1x <2x3 ☐ Out of Tolerance >2x <3x4 ☐ Out of Tolerance >3x <5x5 ☐ Out of Tolerance >5x6 ☐ Out of Tolerance-Undetermined7 ☐ Inoperative8 ☐ Damaged9 ☐ Not Used10 ☐ Not Determined11 ☐ Excessed12 ☐ Extension

Calibrated By: Steve Palmer    S#: 56710    Phone: 526-2761

**CALIBRATION STANDARDS USED**

344008							

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

NOMINAL (STD)	OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION		MFG. ACCURACY
	UNITS	AS FOUND (UUT)	

**COMMENTS**

Nominal: 534.335 lb

Tolerance: +/- 2.00 lb

AS-found data: 534.335 lb

Uncertainty of standards used: +/- 52.0 g (+/- 0.115 lb)

Note: Nominal was determined by the initial calibration value. All following calibrations will be compared to the nominal (initial calibration value).



# INL CALIBRATION INPUT DATA

11/30/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 725133      Mfr: BEA      Model: 102.168 LB      Noun Name: WEIGHT      Serial #: WEIGHT C

Calibration Date: 10/10/2007

**ACTION CODE****AS FOUND**

Next Cal Due Date: 4/10/2009	1	<input type="checkbox"/>	Acceptance Test	1	<input checked="" type="radio"/>	In Tolerance
Charge Level: 4	2	<input type="checkbox"/>	Special Test	2	<input type="radio"/>	Out of Tolerance >1x <2x
Repair/Adj/etc C.L: 0	3	<input checked="" type="checkbox"/>	Calibration to MFG Specs	3	<input type="radio"/>	Out of Tolerance >2x <3x
Material Amount: 0	4	<input type="checkbox"/>	Clean	4	<input type="radio"/>	Out of Tolerance >3x <5x
Charge Number: 100853GSA	5	<input type="checkbox"/>	Limited Calibration	5	<input type="radio"/>	Out of Tolerance >5x
Cal Work Inst ID: 3014P	6	<input type="checkbox"/>	Functional Check	6	<input type="radio"/>	Out of Tolerance-Undetermined
Outside Vendor:	7	<input type="checkbox"/>	Performance Check	7	<input type="checkbox"/>	Inoperative
	8	<input type="checkbox"/>	Modify	8	<input type="checkbox"/>	Damaged
	9	<input type="checkbox"/>	Repair-needs Charge Level	9	<input type="checkbox"/>	Not Used
	10	<input type="checkbox"/>	Other	10	<input type="checkbox"/>	Not Determined
				11	<input type="checkbox"/>	Excessed
Calibrated By: Steve Palmer      S#: 56710      Phone: 526-2761				12	<input type="checkbox"/>	Extension

**CALIBRATION STANDARDS USED**

344008							

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

NOMINAL (STD)	OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION		MFG. ACCURACY
	UNITS	AS FOUND (UUT)	

**COMMENTS**

Nominal: 102.168 lb

Tolerance: +/- 1.50 lb

AS-found data: 102.168 lb

Uncertainty of standards used: +/- 52.0 g (+/- 0.115 lb)

Note: Nominal was determined by the initial calibration value. All following calibrations will be compared to the nominal (initial calibration value).

**INL  
CALIBRATION INPUT DATA**

11/30/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 725134    Mfr: BEA    Model: 83.35 LB    Noun Name: WEIGHT    Serial #: WEIGHT 1

Calibration Date: 10/10/2007

**ACTION CODE****AS FOUND**

Next Cal Due Date: 4/10/2009

Charge Level: 2

Repair/Adj/etc C.L.: 0

Material Amount: 0

Charge Number: 100853GSA

Cal Work Inst ID: 3015P

Outside Vendor:

- |    |                                     |                           |    |                                  |                               |
|----|-------------------------------------|---------------------------|----|----------------------------------|-------------------------------|
| 1  | <input type="checkbox"/>            | Acceptance Test           | 1  | <input checked="" type="radio"/> | In Tolerance                  |
| 2  | <input type="checkbox"/>            | Special Test              | 2  | <input type="radio"/>            | Out of Tolerance >1x <2x      |
| 3  | <input checked="" type="checkbox"/> | Calibration to MFG Specs  | 3  | <input type="radio"/>            | Out of Tolerance >2x <3x      |
| 4  | <input type="checkbox"/>            | Clean                     | 4  | <input type="radio"/>            | Out of Tolerance >3x <5x      |
| 5  | <input type="checkbox"/>            | Limited Calibration       | 5  | <input type="radio"/>            | Out of Tolerance >5x          |
| 6  | <input type="checkbox"/>            | Functional Check          | 6  | <input type="radio"/>            | Out of Tolerance-Undetermined |
| 7  | <input type="checkbox"/>            | Performance Check         | 7  | <input type="checkbox"/>         | Inoperative                   |
| 8  | <input type="checkbox"/>            | Modify                    | 8  | <input type="checkbox"/>         | Damaged                       |
| 9  | <input type="checkbox"/>            | Repair-needs Charge Level | 9  | <input type="checkbox"/>         | Not Used                      |
| 10 | <input type="checkbox"/>            | Other                     | 10 | <input type="checkbox"/>         | Not Determined                |
|    |                                     |                           | 11 | <input type="checkbox"/>         | Excessed                      |
|    |                                     |                           | 12 | <input type="checkbox"/>         | Extension                     |

Calibrated By: Steve Palmer    S#: 56710    Phone: 526-2761

**CALIBRATION STANDARDS USED**

718069							

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED  
VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-5.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

NOMINAL (STD)	OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION UNITS	AS FOUND (UUT)	MFG. ACCURACY

**COMMENTS**

Weight #1

Nominal: 83.388 lb

Tolerance: +/- 1.50 lb

As-found data: 83.444 lb

Uncertainty of Standards Used: +/- 0.050 lb

Note: Nominal value was determined by the average weight from the initial calibration values for the (83 lb to 84 lb) weights.

**INL  
CALIBRATION INPUT DATA**

11/30/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 725135      Mfr: BEA      Model: 83.35 LB      Noun Name: WEIGHT      Serial #: WEIGHT 2

Calibration Date: 10/10/2007

**ACTION CODE****AS FOUND**

Next Cal Due Date: 4/10/2009	1	<input type="checkbox"/>	Acceptance Test	1	<input checked="" type="radio"/>	In Tolerance
Charge Level: 2	2	<input type="checkbox"/>	Special Test	2	<input type="radio"/>	Out of Tolerance >1x <2x
Repair/Adj/etc C.L.: 0	3	<input checked="" type="checkbox"/>	Calibration to MFG Specs	3	<input type="radio"/>	Out of Tolerance >2x <3x
Material Amount: 0	4	<input type="checkbox"/>	Clean	4	<input type="radio"/>	Out of Tolerance >3x <5x
Charge Number: 100853GSA	5	<input type="checkbox"/>	Limited Calibration	5	<input type="radio"/>	Out of Tolerance >5x
Cal Work Inst ID: 3014P	6	<input type="checkbox"/>	Functional Check	6	<input type="radio"/>	Out of Tolerance-Undetermined
Outside Vendor:	7	<input type="checkbox"/>	Performance Check	7	<input type="checkbox"/>	Inoperative
	8	<input type="checkbox"/>	Modify	8	<input type="checkbox"/>	Damaged
	9	<input type="checkbox"/>	Repair-needs Charge Level	9	<input type="checkbox"/>	Not Used
	10	<input type="checkbox"/>	Other	10	<input type="checkbox"/>	Not Determined
				11	<input type="checkbox"/>	Excessed
				12	<input type="checkbox"/>	Extension

Calibrated By: Steve Palmer      S#: 56710      Phone: 526-2761

**CALIBRATION STANDARDS USED**

718069							

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED  
VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-5.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

NOMINAL (STD)	OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION UNITS	AS FOUND (UUT)	MFG. ACCURACY

**COMMENTS**

Weight #2  
 Nominal: 83.388 lb  
 Tolerance: +/- 1.50 lb  
 As-found data: 83.400 lb  
 Uncertainty of Standards Used: +/- 0.050 lb  
 Note: Nominal value was determined by the average weight from the initial calibration values for the (83 lb to 84 lb) weights.

# INL CALIBRATION INPUT DATA

11/30/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 725136      Mfr: BEA      Model: 83.35 LB      Noun Name: WEIGHT      Serial #: WEIGHT 3

Calibration Date: 10/10/2007

**ACTION CODE****AS FOUND**

Next Cal Due Date: 4/10/2009	1	<input type="checkbox"/>	Acceptance Test	1	<input checked="" type="radio"/>	In Tolerance
Charge Level: 2	2	<input type="checkbox"/>	Special Test	2	<input type="radio"/>	Out of Tolerance >1x <2x
Repair/Adj/etc C.L: 0	3	<input checked="" type="checkbox"/>	Calibration to MFG Specs	3	<input type="radio"/>	Out of Tolerance >2x <3x
Material Amount: 0	4	<input type="checkbox"/>	Clean	4	<input type="radio"/>	Out of Tolerance >3x <5x
Charge Number: 100853GSA	5	<input type="checkbox"/>	Limited Calibration	5	<input type="radio"/>	Out of Tolerance >5x
Cal Work Inst ID: 3014P	6	<input type="checkbox"/>	Functional Check	6	<input type="radio"/>	Out of Tolerance-Undetermined
Outside Vendor:	7	<input type="checkbox"/>	Performance Check	7	<input type="checkbox"/>	Inoperative
	8	<input type="checkbox"/>	Modify	8	<input type="checkbox"/>	Damaged
	9	<input type="checkbox"/>	Repair-needs Charge Level	9	<input type="checkbox"/>	Not Used
	10	<input type="checkbox"/>	Other	10	<input type="checkbox"/>	Not Determined
				11	<input type="checkbox"/>	Excessed
				12	<input type="checkbox"/>	Extension

Calibrated By: Steve Palmer      S#: 56710      Phone: 526-2761

**CALIBRATION STANDARDS USED**

718069							

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-5.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

NOMINAL (STD)	OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION UNITS	AS FOUND (UUT)	MFG. ACCURACY

**COMMENTS**

Weight #3  
 Nominal: 83.388 lb  
 Tolerance: +/- 1.50 lb  
 As-found data: 83.308 lb  
 Uncertainty of Standards Used: +/- 0.050 lb  
 Note: Nominal value was determined by the average weight from the initial calibration values for the (83 lb to 84 lb) weights.

**INL  
CALIBRATION INPUT DATA**

11/30/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 725137    Mfr: BEA    Model: 83.35 LB    Noun Name: WEIGHT    Serial #: WEIGHT 4

Calibration Date: 10/10/2007

**ACTION CODE**

**AS FOUND**

Next Cal Due Date: 4/10/2009	1	<input type="checkbox"/>	Acceptance Test	1	<input checked="" type="radio"/>	In Tolerance
Charge Level: 2	2	<input type="checkbox"/>	Special Test	2	<input type="radio"/>	Out of Tolerance >1x <2x
Repair/Adj/etc C.L.: 0	3	<input checked="" type="checkbox"/>	Calibration to MFG Specs	3	<input type="radio"/>	Out of Tolerance >2x <3x
Material Amount: 0	4	<input type="checkbox"/>	Clean	4	<input type="radio"/>	Out of Tolerance >3x <5x
Charge Number: 100853GSA	5	<input type="checkbox"/>	Limited Calibration	5	<input type="radio"/>	Out of Tolerance >5x
Cal Work Inst ID: 3014P	6	<input type="checkbox"/>	Functional Check	6	<input type="radio"/>	Out of Tolerance-Undetermined
Outside Vendor:	7	<input type="checkbox"/>	Performance Check	7	<input type="checkbox"/>	Inoperative
	8	<input type="checkbox"/>	Modify	8	<input type="checkbox"/>	Damaged
	9	<input type="checkbox"/>	Repair-needs Charge Level	9	<input type="checkbox"/>	Not Used
	10	<input type="checkbox"/>	Other	10	<input type="checkbox"/>	Not Determined
				11	<input type="checkbox"/>	Excessed
Calibrated By: Steve Palmer				12	<input type="checkbox"/>	Extension

**CALIBRATION STANDARDS USED**

718069							

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-5.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

NOMINAL (STD)	OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION UNITS	AS FOUND (UUT)	MFG. ACCURACY

**COMMENTS**

Weight #4  
 Nominal: 83.388 lb  
 Tolerance: +/- 1.50 lb  
 As-found data: 83.468 lb  
 Accuracy of Standards Used: +/- 0.050 lb  
 Note: Nominal value was determined by the average weight from the initial calibration values for the (83 lb to 84 lb) weights.

**INL**  
**CALIBRATION INPUT DATA**

11/30/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 725138    Mfr: BEA    Model: 83.35 LB    Noun Name: WEIGHT    Serial #: WEIGHT 5

Calibration Date: 10/10/2007

**ACTION CODE****AS FOUND**

Next Cal Due Date: 4/10/2009

Charge Level: 2

Repair/Adj/etc C.L: 0

Material Amount: 0

Charge Number: 100853GSA

Cal Work Inst ID: 3014P

Outside Vendor:

1 ☐ Acceptance Test2 ☐ Special Test3 ☒ Calibration to MFG Specs4 ☐ Clean5 ☐ Limited Calibration6 ☐ Functional Check7 ☐ Performance Check8 ☐ Modify9 ☐ Repair-needs Charge Level10 ☐ Other1 ☒ In Tolerance2 ☐ Out of Tolerance >1x <2x3 ☐ Out of Tolerance >2x <3x4 ☐ Out of Tolerance >3x <5x5 ☐ Out of Tolerance >5x6 ☐ Out of Tolerance-Undetermined7 ☐ Inoperative8 ☐ Damaged9 ☐ Not Used10 ☐ Not Determined11 ☐ Excessed12 ☐ Extension

Calibrated By: Steve Palmer    S#: 56710    Phone: 526-2761

**CALIBRATION STANDARDS USED**

718069							

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED  
VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/- 0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/- 0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/- 0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/- 1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/- 0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/- 3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

**OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION****NOMINAL (STD)****UNITS****AS FOUND (UUT)****MFG. ACCURACY**


**COMMENTS**

Weight #5

Nominal: 83.388 lb

Tolerance: +/- 1.50 lb

As-found data: 83.432 lb

Uncertainty of Standards Used: +/- 0.050 lb

Note: Nominal value was determined by the average weight from the initial calibration values for the (83 lb to 84 lb) weights.

**INL  
CALIBRATION INPUT DATA**

11/30/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 725139    Mfr: BEA    Model: 83.35 LB    Noun Name: WEIGHT    Serial #: WEIGHT 6

Calibration Date: 10/10/2007

**ACTION CODE**

**AS FOUND**

Next Cal Due Date: 4/10/2009	1	<input type="checkbox"/>	Acceptance Test	1	<input checked="" type="radio"/>	In Tolerance
Charge Level: 2	2	<input type="checkbox"/>	Special Test	2	<input type="radio"/>	Out of Tolerance >1x <2x
Repair/Adj/etc C.L: 0	3	<input checked="" type="checkbox"/>	Calibration to MFG Specs	3	<input type="radio"/>	Out of Tolerance >2x <3x
Material Amount: 0	4	<input type="checkbox"/>	Clean	4	<input type="radio"/>	Out of Tolerance >3x <5x
Charge Number: 100853GSA	5	<input type="checkbox"/>	Limited Calibration	5	<input type="radio"/>	Out of Tolerance >5x
Cal Work Inst ID: 3014P	6	<input type="checkbox"/>	Functional Check	6	<input type="radio"/>	Out of Tolerance-Undetermined
Outside Vendor:	7	<input type="checkbox"/>	Performance Check	7	<input type="checkbox"/>	Inoperative
	8	<input type="checkbox"/>	Modify	8	<input type="checkbox"/>	Damaged
	9	<input type="checkbox"/>	Repair-needs Charge Level	9	<input type="checkbox"/>	Not Used
	10	<input type="checkbox"/>	Other	10	<input type="checkbox"/>	Not Determined
				11	<input type="checkbox"/>	Excessed
Calibrated By: Steve Palmer    S#: 56710    Phone: 526-2761				12	<input type="checkbox"/>	Extension

**CALIBRATION STANDARDS USED**

718069							

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

NOMINAL (STD)	OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION		MFG. ACCURACY
	UNITS	AS FOUND (UUT)	

**COMMENTS**

Weight #6  
 Nominal: 83.388 lb  
 Tolerance: +/- 1.50 lb  
 As-found data: 83.278 lb  
 Uncertainty of Standards Used: +/- 0.050 lb  
 Note: Nominal value was determined by the average weight from the initial calibration values for the (83 lb to 84 lb) weights.

**INL**  
**CALIBRATION INPUT DATA**

11/30/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 725140    Mfr: BEA    Model: 4.92 LB SHACKLE    Noun Name: WEIGHT    Serial #: CFA-S-05-1-12

Calibration Date: 10/10/2007

**ACTION CODE****AS FOUND**

Next Cal Due Date: 4/10/2009

Charge Level: 2

Repair/Adj/etc C.L: 0

Material Amount: 0

Charge Number: 100853GSA

Cal Work Inst ID: 3014P

Outside Vendor:

- |    |                                     |                           |
|----|-------------------------------------|---------------------------|
| 1  | <input type="checkbox"/>            | Acceptance Test           |
| 2  | <input type="checkbox"/>            | Special Test              |
| 3  | <input checked="" type="checkbox"/> | Calibration to MFG Specs  |
| 4  | <input type="checkbox"/>            | Clean                     |
| 5  | <input type="checkbox"/>            | Limited Calibration       |
| 6  | <input type="checkbox"/>            | Functional Check          |
| 7  | <input type="checkbox"/>            | Performance Check         |
| 8  | <input type="checkbox"/>            | Modify                    |
| 9  | <input type="checkbox"/>            | Repair-needs Charge Level |
| 10 | <input type="checkbox"/>            | Other                     |

- |    |                                  |                               |
|----|----------------------------------|-------------------------------|
| 1  | <input checked="" type="radio"/> | In Tolerance                  |
| 2  | <input type="radio"/>            | Out of Tolerance >1x <2x      |
| 3  | <input type="radio"/>            | Out of Tolerance >2x <3x      |
| 4  | <input type="radio"/>            | Out of Tolerance >3x <5x      |
| 5  | <input type="radio"/>            | Out of Tolerance >5x          |
| 6  | <input type="radio"/>            | Out of Tolerance-Undetermined |
| 7  | <input type="checkbox"/>         | Inoperative                   |
| 8  | <input type="checkbox"/>         | Damaged                       |
| 9  | <input type="checkbox"/>         | Not Used                      |
| 10 | <input type="checkbox"/>         | Not Determined                |
| 11 | <input type="checkbox"/>         | Excessed                      |
| 12 | <input type="checkbox"/>         | Extension                     |

Calibrated By: Steve Palmer    S#: 56710    Phone: 526-2761

**CALIBRATION STANDARDS USED**

328934							

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED  
VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

**OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION****NOMINAL (STD)****UNITS****AS FOUND (UUT)****MFG. ACCURACY**


**COMMENTS**

Nominal: 4.920 lb  
Tolerance: +/- 0.50 lb  
AS-found data: 4.9095344 lb  
Uncertainty of standards used: +/- 0.722 g (+/- 0.00159 lb)  
Note: Nominal value was determined by the average weight from the initial calibration values for the (5 lb) shackles.



**INL**  
**CALIBRATION INPUT DATA**

11/30/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 725141      Mfr: BEA      Model: 4.92 LB SHACKLE      Noun Name: WEIGHT      Serial #: CFA-S-05-1-7

Calibration Date: 10/10/2007

**ACTION CODE****AS FOUND**

Next Cal Due Date: 4/10/2009	1	<input type="checkbox"/>	Acceptance Test	1	<input checked="" type="radio"/>	In Tolerance
Charge Level: 2	2	<input type="checkbox"/>	Special Test	2	<input type="radio"/>	Out of Tolerance >1x <2x
Repair/Adj/etc C.L: 0	3	<input checked="" type="checkbox"/>	Calibration to MFG Specs	3	<input type="radio"/>	Out of Tolerance >2x <3x
Material Amount: 0	4	<input type="checkbox"/>	Clean	4	<input type="radio"/>	Out of Tolerance >3x <5x
Charge Number: 100853GSA	5	<input type="checkbox"/>	Limited Calibration	5	<input type="radio"/>	Out of Tolerance >5x
Cal Work Inst ID: 3014P	6	<input type="checkbox"/>	Functional Check	6	<input type="radio"/>	Out of Tolerance-Undetermined
Outside Vendor:	7	<input type="checkbox"/>	Performance Check	7	<input type="checkbox"/>	Inoperative
	8	<input type="checkbox"/>	Modify	8	<input type="checkbox"/>	Damaged
	9	<input type="checkbox"/>	Repair-needs Charge Level	9	<input type="checkbox"/>	Not Used
	10	<input type="checkbox"/>	Other	10	<input type="checkbox"/>	Not Determined
				11	<input type="checkbox"/>	Excessed
Calibrated By: Steve Palmer				12	<input type="checkbox"/>	Extension

**CALIBRATION STANDARDS USED**

328934							

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-5.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

**OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION**

NOMINAL (STD)	UNITS	AS FOUND (UUT)	MFG. ACCURACY

**COMMENTS**

Nominal: 4.920 lb  
Tolerance: +/- 0.50 lb  
AS-found data: 4.92899021b  
Uncertainty of standards used: +/- 0.722 g (+/- 0.00159 lb)  
Note: Nominal value was determined by the average weight from the initial calibration values for the (5 lb) shackles.

# INL CALIBRATION INPUT DATA

11/30/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 725142    Mfr: BEA    Model: 0.8 LB LINK    Noun Name: WEIGHT    Serial #: CITRIC-ML-02-2

Calibration Date: 10/10/2007	ACTION CODE	AS FOUND
Next Cal Due Date: 4/10/2009	1 <input type="checkbox"/> Acceptance Test	1 <input checked="" type="radio"/> In Tolerance
Charge Level: 2	2 <input type="checkbox"/> Special Test	2 <input type="radio"/> Out of Tolerance >1x <2x
Repair/Adj/etc C.L: 0	3 <input checked="" type="checkbox"/> Calibration to MFG Specs	3 <input type="radio"/> Out of Tolerance >2x <3x
Material Amount: 0	4 <input type="checkbox"/> Clean	4 <input type="radio"/> Out of Tolerance >3x <5x
Charge Number: 100853GSA	5 <input type="checkbox"/> Limited Calibration	5 <input type="radio"/> Out of Tolerance >5x
Cal Work Inst ID: 3014P	6 <input type="checkbox"/> Functional Check	6 <input type="radio"/> Out of Tolerance-Undetermined
Outside Vendor:	7 <input type="checkbox"/> Performance Check	7 <input type="checkbox"/> Inoperative
	8 <input type="checkbox"/> Modify	8 <input type="checkbox"/> Damaged
	9 <input type="checkbox"/> Repair-needs Charge Level	9 <input type="checkbox"/> Not Used
	10 <input type="checkbox"/> Other	10 <input type="checkbox"/> Not Determined
		11 <input type="checkbox"/> Excessed
Calibrated By: Steve Palmer    S#: 56710    Phone: 526-2761		12 <input type="checkbox"/> Extension

## CALIBRATION STANDARDS USED

328934							

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

### LABORATORY TEMPERATURE AND HUMIDITY

Physical STD (106C)	20.0 ° +/-0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/-0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/-0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/-1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/-0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/-3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

NOMINAL (STD)	OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION		MFG. ACCURACY
	UNITS	AS FOUND (UUT)	

### COMMENTS

Nominal: 0.8 lb  
Tolerance: +/- 0.50 lb  
As-found data: 0.809 lb  
Uncertainty of Standards Used: +/- 0.722 g (+/- 0.00159 lb)  
Note: Nominal value was determined by the average weight from the initial calibration values for the (1 lb) links.

# INL CALIBRATION INPUT DATA

11/30/2007

NAME: DANA KEITH MORTON	BADGE: 35698	PH: 526-1274	AREA: STC	BLDG: EROB	RM: W2D1
-------------------------	--------------	--------------	-----------	------------	----------

ID Number: 725143    Mfr: BEA    Model: 0.8 LB LINK    Noun Name: WEIGHT    Serial #: CITRIC-ML-07-1/2-2

Calibration Date: 10/10/2007

**ACTION CODE****AS FOUND**

Next Cal Due Date: 4/10/2009	1	<input type="checkbox"/>	Acceptance Test	1	<input checked="" type="radio"/>	In Tolerance
Charge Level: 2	2	<input type="checkbox"/>	Special Test	2	<input type="radio"/>	Out of Tolerance >1x <2x
Repair/Adj/etc C.L: 0	3	<input checked="" type="checkbox"/>	Calibration to MFG Specs	3	<input type="radio"/>	Out of Tolerance >2x <3x
Material Amount: 0	4	<input type="checkbox"/>	Clean	4	<input type="radio"/>	Out of Tolerance >3x <5x
Charge Number: 100853GSA	5	<input type="checkbox"/>	Limited Calibration	5	<input type="radio"/>	Out of Tolerance >5x
Cal Work Inst ID: 3014P	6	<input type="checkbox"/>	Functional Check	6	<input type="radio"/>	Out of Tolerance-Undetermined
Outside Vendor:	7	<input type="checkbox"/>	Performance Check	7	<input type="checkbox"/>	Inoperative
	8	<input type="checkbox"/>	Modify	8	<input type="checkbox"/>	Damaged
	9	<input type="checkbox"/>	Repair-needs Charge Level	9	<input type="checkbox"/>	Not Used
	10	<input type="checkbox"/>	Other	10	<input type="checkbox"/>	Not Determined
				11	<input type="checkbox"/>	Excessed
Calibrated By: Steve Palmer    S#: 56710    Phone: 526-2761				12	<input type="checkbox"/>	Extension

**CALIBRATION STANDARDS USED**

328934							

STANDARDS USED ARE TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY DERIVED FROM ACCEPTED VALUES FOR NATURAL PHYSICAL CONSTANTS, OR DERIVED FROM THE RATIO TYPE OF SELF CALIBRATION TECHNIQUES

**LABORATORY TEMPERATURE AND HUMIDITY**

Physical STD (106C)	20.0 ° +/- 0.3 ° C (40-55% RH)	Electronic STD (106D)	23.0 ° +/- 0.5 ° C (30-45% RH)
Dimensional STD (106B)	20.0 ° +/- 0.25 ° C (30-45% RH)	Electronic CAL (Lab 112)	23.0 ° +/- 1.0 ° C (20-50% RH)
Phys/Dim CAL (Lab 111)	20.0 ° +/- 0.5 ° C (20-50% RH)	Remaining S&CL calibration areas:	23.0 ° +/- 3.0 ° C (20-50% RH)

Manufacturer's environmental specifications are evaluated for conformance when calibrations are performed outside the above stated conditions.

**OUT OF TOLERANCE CONDITIONS FOUND DURING CALIBRATION**

NOMINAL (STD)	UNITS	AS FOUND (UUT)	MFG. ACCURACY

**COMMENTS**

Nominal: 0.846 lb  
Tolerance: +/- 0.50 lb  
As-found data: 0.883 lb  
Uncertainty of Standards Used: +/- 0.722 g (+/- 0.00159 lb)  
Note: Nominal value was determined by the average weight from the initial calibration values for the (1 lb) links.