

TRANSMISSION VERIFICATION REPORT

TIME : 11/25/2009 07:45  
NAME : USNRC RIII  
FAX : 6308299782  
TEL :  
SER.# : 000A7J925774

DATE, TIME 11/25 07:44  
FAX NO./NAME 84178310195  
DURATION 00:00:41  
PAGE(S) 03  
RESULT OK  
MODE STANDARD  
ECM

NRC FORM 386 (RIII)  
(1-2004)



UNITED STATES  
NUCLEAR REGULATORY COMMISSION  
REGION III  
2443 Warrenville Road, Suite 210  
Lisle, Illinois 60532-4352

TELEFAX TRANSMITTAL

DATE: 11/25/2009 NUMBER OF PAGES: 3  
(including this page)

SEND TO: Larry Albaugh, CNMT

LOCATION: Nuclear Diagnosis, Inc.

FAX NUMBER: 417 - 831 - 0195  **VERIFY BY CALLING SENDER**

FROM: Jose Macatangay *JM*  
(SENDER)

TELEPHONE NUMBER: 630 - 829 - 9892 FAX NUMBER: 630 - 515 - 1078

If you do not receive the complete fax transmittal, please contact the sender as soon as possible at the telephone number provided above.

MESSAGE  
Larry,



UNITED STATES  
NUCLEAR REGULATORY COMMISSION  
REGION III  
2443 Warrenville Road, Suite 210  
Lisle, Illinois 60532-4352

TELEFAX TRANSMITTAL

DATE: 11/25/2009 NUMBER OF PAGES: 3  
(including this page)

SEND TO: Larry Albaugh, CNMT

LOCATION: Nuclear Diagnosis, Inc.

FAX NUMBER: 417 - 831 - 0195  **VERIFY BY CALLING SENDER**

FROM: Jose Macatangay *AM*  
(SENDER)

TELEPHONE NUMBER: 630 - 829 - 9892 FAX NUMBER: 630 - 515 - 1078

If you do not receive the complete fax transmittal, please contact the sender as soon as possible at the telephone number provided above.

MESSAGE

*Larry,  
Per our phone conversation on 11/24/2009. Also, please include another point-of-contact for this request.*

*Thank you.  
Jose Macatangay*

NOTICE

This message is intended only for the use of the individual or entity to which it is addressed and may contain information that is privileged, confidential, or exempt from disclosure under applicable law. If the reader of this message is not the intended recipient or the employee responsible for delivering the message to the intended recipient, you are hereby notified that any dissemination, distribution or copying of this communication is strictly prohibited. If you have received this communication in error, please notify the sender immediately by telephone and return the original to the above address, by U.S. Mail. Thank you.

**JOSE D. MACATANGAY**  
**MATERIALS LICENSING BRANCH**  
**UNITED STATES NUCLEAR REGULATORY COMMISSION**  
REGION III  
2443 WARRENVILLE ROAD STE 210  
LISLE, ILLINOIS 60532-4352  
OFFICE: (630)-829-9892 FAX: (630) 515-1078

<b>CONVERSATION RECORD</b>	TIME	DATE
<b>ACTUALLY FAXED?</b>	<b>Yes</b>	<b>11/25/2009</b>
NAME OF PERSON(S) CONTACTED	ORGANIZATION	TELEPHONE NO.
Larry Albaugh, CNMT	Nuclear Diagnosis, Inc.	O: 417-831-6520 F: 417-831-0195
SUBJECT		
License No.: 24-17561-01	Control No.: 318503	
SUMMARY		

We have reviewed your amendment request dated September 9, 2009, and find that we need additional information as follows:

1. We cannot authorize licensees to release the areas (excepting areas covered by 10 CFR 35.13(e) and 35.14(b)(4)) or locations of use from licenses for unrestricted use (even by other members of their staff) until we have received and reviewed a copy of the results of decommissioning and close-out surveys for the facilities.

The final status survey should consist of exposure rate measurements to show that all sources of radioactive material have been removed, and contamination checks of areas where radioactive materials were used or stored.

Please submit the following information with your close-out survey (refer to 10 CFR 30.36(j)(2)):

- a. Diagrams of each facility with exposure rate survey and wipe test results keyed to specific locations, as appropriate.
- b. The name of the person performing the survey.
- c. The date the survey was performed.
- d. The instrument(s) used for exposure rate measurements and for analysis of the wipes.
- e. Background readings and each instruments' efficiency or correction factor.
- f. The date(s) that the survey instruments were last calibrated.
- g. The action levels for both exposure rate measurements and wipe tests. Include the identity of areas exceeding these levels, corrective actions taken and results of corrective actions taken.

2. Also, per our phone conversation on November 24, 2009, you indicated that sealed sources were stored in the two locations of use to be released. In order to release as unrestricted areas of use, please provide the most current leak test results for your sealed sources at those two locations of use.

We will be unable to continue processing your request until we receive this information. In accordance with 10 CFR 2.390 of the NRC's "Rules of Practice," a copy of this letter will be available electronically in the NRC Public Document Room or from the Publicly Available Record (PARS) component of NRC's document system (ADAMS) accessible from the NRC Website at <http://www.nrc.gov/reading-rem/adams.html>.

---

**ACTION REQUIRED**

Please submit a **signed written response** within 7-days or contact me to arrange an alternate response date. Be sure to reference control number 318503 to facilitate correct processing of your response.

If we do not receive a **signed written response** within 7-days, please note that we may void this request in order to enable you to prepare a quality response without time constraints. This would be done without prejudice to the resubmission of your request at a later date. Upon receipt of your response we will resume our review. Address your written response to my attention at the above address.

Please note that a "Void" is an administrative procedure that puts your amendment request "On Hold" (takes it out of our active casework database) until you reactivate it via submission of a written response.

Upon receipt of your response we will resume our review.

**PLEASE DIRECT ANY QUESTIONS YOU MAY HAVE TO ME AT (630) 829-9892.**

---

NAME OF PERSON DOCUMENTING CONVERSATION	SIGNATURE	DATE
Jose Macatangay		November 25, 2009

---