



December 7, 2009

United States NRC (Region III)
Attn: Jose Macatangay
2443 Warrenville Road, Suite 210
Lisle, Illinois 60532-4352

RE: Control Number 318568 (License No. 21-32687-01)

Mr. Macatangay:

Surveying Solutions, Inc. recently received a request for additional information from your office related to Control Number 318568. In response to your request, we offer the following:

Question 1.) Confirm that you want to remove 561 W. M-61, Standish, MI from your license.

Answer 1.) That is correct; please remove 561 W. M-61, Standish, MI from our license.

Question 2.) Please send the most recent leak test certificates for the portable gauges that were stored at the 561 W. M-61 location.

Answer 2.) Leak test certificates for both portable gauges stored at the 561 W. M-61 location are attached to this letter.

Question 3.) Please verify that no leaking portable gauges have been stored at the 561 W. M-61 location.

Answer 3.) Surveying Solutions, Inc. verifies that there were no leaking portable gauges stored at the 561 W. M-61 location.

This is also confirmation of our new address which is **4471 M-61, Standish, MI 48658**. If you have any questions, please feel free to contact me at (989) 846-6601.

Sincerely,

SURVEYING SOLUTIONS, INC.

Adam R. Ball, P.E.
Project Engineer
aball@ssi-mi.com



Troxler Electronic Laboratories, Inc.

3008 Cornwallis Rd., P.O. Box 12057
Research Triangle Park, NC 27709
Tel: (877) 876-9537 Fax: (866) 391-2759
License: NC 032-0182-1

Cust ID: 7365

TROXLER ELECTRONIC LABS
RTP, NC 27709

LEAK TEST CERTIFICATE

DEVICE:

Model: 3440 **Serial No:** 60986

SEALED SOURCES:

Serial No.	Measure Date	Nuclide	GBq	mCi
77-7928	03/12/08	CS-137	0.296	8
78-4921	04/07/08	AM-241:BE	1.48	40

LEAK TEST ANALYSIS:

Sample collected on: 05/21/2009

Sample analyzed on: 05/27/2009 at 8:13:00 AM

Analyzed by: T. Pjura

	ALPHA	BETA-GAMMA
Conversion factor (cpm/Bq)	1.24E+01	2.03E+01
Background measurement (cpm)	1	26
Sample measurement (cpm)	0	29
Activity (Bq)	< MDA	< MDA
Min. Detectable Activity (Bq)	5.3E-01	1.3E+00

This certifies that the leak test results are:

Less than 185 Bq (0.005 uCi)

Greater than 185 Bq (0.005 uCi)



Troxler Electronic Laboratories, Inc.

3008 Cornwallis Rd., P.O. Box 12057
Research Triangle Park, NC 27709
Tel: (877) 876-9537 Fax: (866) 391-2759
License: NC 032-0182-1

BRIAN BARLETT
SURVERYING SOLUTION INC.
317 SOUTH MAIN STREET
STANDISH, MI 48658

LEAK TEST CERTIFICATE

DEVICE:

Model: 3440 **Serial No:** 60943

SEALED SOURCES:

Serial No.	Measure Date	Nuclide	GBq	mCi
77-7909	3/12/2008	CS-137	0.296	8
78-4904	12/19/2007	AM-241:BE	1.48	40

LEAK TEST ANALYSIS:

Sample collected on: 05/01/2009
Sample analyzed on: 05/06/2009 at 8:46:00 AM
Analyzed by: Terri Pjura

	ALPHA	BETA-GAMMA
Conversion factor (cpm/Bq)	1.24E+01	2.03E+01
Background measurement (cpm)	1	26
Sample measurement (cpm)	0	31
Activity (Bq)	< MDA	< MDA
Min. Detectable Activity (Bq)	5.3E-01	1.3E+00

This certifies that the leak test results are:

Less than 185 Bq (0.005 uCi) **Greater than 185 Bq (0.005 uCi)**

SURVEYING SOLUTIONS, INC.

P.O. BOX 1022
STANDISH, MI 48658-1022



United States NRC (Region III)
Attn: Jose Macatangay
2443 Warrenville Road, Suite 210
Lisle, Illinois 60532-4352

60532+4352

