



REGION I
475 ALLENDALE ROAD
KING OF PRUSSIA, PENNSYLVANIA 19406-1415

December 9, 2009

Docket No. 03034704
Control No. 143962

License No. 52-25430-01

David Rhoe
Radiation Safety Officer
CRMI
Paseo de la Fuente
D-4 Calle Tivoli
San Juan, PR 00926-6459

SUBJECT: CRMI, VOIDANCE OF APPLICATION FOR LICENSE AMENDMENT, CONTROL
NO. 143962

Dear Mr. Rhoe:

This concerns the subject application for an amendment to your service provider license (52-25430-01). We have voided your application for your service provider license because the license does not authorize the possession of cobalt-57 and will not be reviewed as part of your request to dispose of Co-57 by decay in storage for your medical license (52-25430-03). This action is taken without prejudice to the submission of your request for decay in storage of Co-57 for your medical license. In addition, because the issue of using the Australian and New Zealand Society of Nuclear Medicine standards for calibration of dose calibrators was addressed in Christian Einberg's letter dated May 27, 2009, we will not be acting on your request to perform dose calibrator linearity tests annually.

Sincerely,

Original signed by Tara L. Weidner

Tara L. Weidner
Health Physicist
Medical Branch
Division of Nuclear Materials Safety

DOCUMENT NAME: G:\WordDocs\Current\Lic Cvr Letter\L52-25430-01.143962.doc

SUNSI Review Complete: TWeidner

After declaring this document "An Official Agency Record" it will be released to the Public.

To receive a copy of this document, indicate in the box: "C" = Copy w/o attach/encl "E" = Copy w/ attach/encl "N" = No copy

OFFICE	DNMS/RI	N	DNMS/RI		DNMS/RI			
NAME	TWeidner/TLW							
DATE	12.9.09							

OFFICIAL RECORD COPY