

December 04, 2009

Medical Imaging Center
6 Northwestern Drive – Suite 102
Bloomfield, CT 06002-3463

ATTENTION RADIATION SAFETY OFFICER

This letter confirms the return for disposal of the following returned source:

Co-57 Flood Source
Serial Number: 1230-052
Calibration: 10mCi A/O 06/01/2007.

This source was returned on December 4, 2009 to Eckert & Ziegler Isotope Products,
1800 No. Keystone Street, Burbank, CA 91504.


Wipe Test Result:

Counter Efficiency: 0.93
Bkg cpm: 70
Net cpm: 0
Net dpm: 0

Passed with <0.005 uCi removable contamination.

Thank you for choosing Cardinal Health Nuclear Pharmacy Services for your
radiopharmaceutical needs.

Sincerely,



Peter Sposato
Pharmacy Manager

December 04, 2009

Medical Imaging Center
6 Northwestern Drive – Suite 102
Bloomfield, CT 06002-3463

ATTENTION RADIATION SAFETY OFFICER

This letter confirms the return for disposal of the following returned source:

Co-57 Spot Markers
Serial Number: BM03-100 x 2, B2988-23
Calibration: 98 uCi A/O 8/17/04 x 2, 65 uCi A/O 10/1/00
Cs137 Vial source
Serial Number: 3560983A-29
Calibration: 204 uCi A/O 9/19/83


This source was returned on December 4, 2009 to Eckert & Ziegler Isotope Products,
1800 No. Keystone Street, Burbank, CA 91504.

Wipe Test Result:

Counter Efficiency: 0.93
Bkg cpm: 70
Net cpm: 0
Net dpm: 0

Passed with <0.005 uCi removable contamination.
Thank you for choosing Cardinal Health Nuclear Pharmacy Services for your
radiopharmaceutical needs.

Sincerely,



Peter Sposato
Pharmacy Manager