

SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

1. LICENSEE/LOCATION INSPECTED:

N Gupta, m.d.
1660 Kingway Court
Trenton, NJ 08618
REPORT NUMBER(S)

2009/001

2. NRC/REGIONAL OFFICE

U.S. Nuclear Regulatory Commission
Region III
2443 Warrenville Road
Suite 210
Lisle, Illinois 60532-4351

3. DOCKET NUMBER(S)

03034742

4. LICENSEE NUMBER(S)

21-32096-01

5. DATE(S) OF INSPECTION

11/17/09

LICENSEE:

The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The inspection findings are as follows:

- 1. Based on the inspection findings, no violations were identified.
- 2. Previous violation(s) closed.
- 3. The violation(s), specifically described to you by the inspector as non-cited violations, are not being cited because they were self-identified, non-repetitive, and corrective action was or is being taken, and the remaining criteria in the NRC Enforcement Policy, NUREG-1600, to exercise discretion, were satisfied.

_____ Non-Cited Violation(s) was/were discussed involving the following requirement(s) and Corrective Action(s):

- 4. During this inspection certain of your activities, as described below and/or attached, were in violation of NRC requirements and are being cited. This form is a NOTICE OF VIOLATION, which may be subject to posting in accordance with 10 CFR 19.11.
(Violations and Corrective Actions)

Licensee's Statement of Corrective Actions for Item 4, above.

I hereby state that, within 30 days, the actions described by me to the inspector will be taken to correct the violations identified. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201 (corrective steps already taken, corrective steps which will be taken, date when full compliance will be achieved). I understand that no further written response to NRC will be required, unless specifically requested.

Title	Printed Name	Signature	Date
LICENSEE'S REPRESENTATIVE			
NRC INSPECTOR	<i>G. Park</i>	<i>[Signature]</i>	<i>11/17/09</i>

845

SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

1. LICENSEE N. Gupta, M.D. P.C. REPORT NUMBER(S) 2009-001		2. NRC/REGIONAL OFFICE Region III 2443 Warrenville Road, Suite 210 Lisle, IL 60532	
3. DOCKET NUMBER(S) 030-34742	4. LICENSE NUMBER(S) 21-32096-01	5. DATE(S) OF INSPECTION 11/17/2009	
6. INSPECTION PROCEDURES USED 87130		7. INSPECTION FOCUS AREAS 03.01-03.07	

SUPPLEMENTAL INSPECTION INFORMATION

1. PROGRAM CODE(S) 2201	2. PRIORITY 5	3. LICENSEE CONTACT Jack E. Juni, M.D.	4. TELEPHONE NUMBER 734.676.8400
----------------------------	------------------	---	-------------------------------------

Main Office Inspection Next Inspection Date: 11/2014

Field Office _____

Temporary Job Site Inspection _____

PROGRAM SCOPE

This program performs approximately four nuclear cardiology procedures daily utilizing primarily Tc-99m with occasional thallium-201. Two technologists currently perform all procedures. Generators are not received and all material is obtained from an area nuclear pharmacy in the form of unit doses. An outside consultant (MPC) performs quarterly program compliance audits. The licensee is operational Monday-Friday, 8:30am-5:00pm.

Performance Observations

Interviews conducted with available staff revealed an adequate level of understanding of emergency and material handling procedures and techniques. Dose calibrator consistency checks, daily surveys, waste disposal and injection techniques were successfully described or observed.

Licensed material was observed adequately secured during the review and was not readily accessible to members of the general public. Licensed material is stored in a lockable cabinet which houses the hot-lab area of the clinic.

Independent measurements taken indicated no abnormal readings. Personal dosimetry records reviewed indicated no abnormal reading.

8/5