



# SOUTHWEST RESEARCH INSTITUTE®

6220 Culebra Road, P.O. Drawer 28510  
Institute Quality Systems  
Institute Calibration Laboratory  
Phone: 210-522-5215 Fax 210-522-4834



Calibration Laboratory  
Certificate #0972-01

## Certificate of Calibration

**Cost Center / Customer:** DIV20 / DON BANNON

**Mail Stop:** B51

**Manufacturer/Model:** KEITHLEY / 614

**Description:** ELECTROMETER

**Serial Number:** 0704934

**Asset Number:** 007088

**Procedure:** KEITHLEY 614 - 21 MAR 09

**Work Order:** 303091229

**Date Issued:** 17-Nov-2009

**Date Calibrated:** 16-Nov-2009

**\* Date Due :** 16-Nov-2010

**\*\* Results:** FOUND-LEFT

**Temperature:** 74.0 °F

**Humidity:** 40 %RH

**Barometer:** N/A

This certificate documents traceability to the National Institute of Standards and Technology (NIST) and the International System of Units (SI). The Laboratory quality system conforms to ISO/IEC 17025, 2005, ANSI/NCCL Z540-1-1994 and relevant requirements of the ISO 9000-2000 standard. This certificate shall not be reproduced, except in full, without the written approval of the Southwest Research Institute Calibration Laboratory. This certificate shall not be used to claim product endorsement by Southwest Research Institute, American Association for Laboratory Accreditation (A2LA) or any agency of the U. S. Government. Results of this calibration relate only to the instrument described above at the time of calibration and does not imply any long term stability of the instrument.

\*Determined by the customer, does not imply the instrument will remain within tolerance as any number of factors may cause an out-of-tolerance condition before this date. \*\*Data type found in this certificate or attached measurement report must be interpreted as: Found-left - adjustment and/or repair was not performed, As-found - data is before unit is adjusted and/or repaired, As-left - data is after adjusted and/or repaired was performed. The customer has sole responsibility for determination of in-/out-of-tolerance or compliance/noncompliance.

Measurement uncertainty calculated in accordance with the method described in the ISO "Guide to the Expression of Uncertainty in Measurement" (GUM), for a confidence level of approximately 95 percent using a coverage factor of  $k=2$ .

**Remarks:** nC not calibrated.

### Standards Used

<u>Asset #</u>	<u>Manufacturer</u>	<u>Model</u>	<u>Description</u>	<u>Cal Date</u>	<u>Due Date</u>
000101	BIDDLE	72-6346-1	DECADE RESISTOR	24-Aug-2009	24-Aug-2010
000182	FLUKE	5700A/EP	CALIBRATOR	2-Nov-2009	2-Feb-2010
000201	FLUKE	5725A	AMPLIFIER	2-Nov-2009	2-Feb-2010
009829	ESI	SR1050-10M	RESISTANCE TRANSFER STANDARD	14-May-2009	14-May-2010

Walt Hill

Laboratory Manager

Bob Trollinger

Metrology Technician

Southwest Research Institute  
Calibration Laboratory  
Measurement Report

Work Order:	303091229	Mfr.	KEITHLEY	Technician	blt
Asset No.	007088	Model	614	Type Data:	Found-left
Serial No.	0704934	Type.	Electrometer	Cal Date.	16-Nov-09
Remarks: nC not calibrated per customer.					

Function/Range	Test Point	TI Reading	Difference	+/-Test Limits	+/-Uncertainty	Result	%Limit
Zero	Volts	Volts	Volts	Volts	Volts		
0.2 V	0.19000	0.19002	0.00002	0.00017	0.000010	Pass	12%
	-0.19000	-0.18998	0.00002	0.00017	0.000010	Pass	12%
2 Volt	1.9000	1.9001	0.0001	0.0016	0.00010	Pass	6%
	-1.9000	-1.9001	-0.0001	0.0016	0.00010	Pass	6%
20 Volt	19.000	19.001	0.001	0.016	0.0010	Pass	6%
	-19.000	-19.001	-0.001	0.016	0.0010	Pass	6%
DC Amps	uAmps	uAmps	uAmps	uAmps	uAmps		
200 uAmp	190.0	190.1	0.1	0.7	0.055	Pass	14%
	nAmps	nAmps	nAmps	nAmps	nAmps		
200 nAmp	190.0	190.2	0.2	1.0	0.055	Pass	20%
	pAmps	pAmps	pAmps	pAmps	pAmps		
2000 pAmp	1900	1890	-10	29	0.50	Pass	34%
Resistance	kohm	kohm	kohm	kohm	kohm		
20 kohm	19.00	18.98	-0.02	0.11	0.0058	Pass	18%
200 kohm	190.0	189.8	-0.2	1.1	0.058	Pass	18%
	Mohm	Mohm	Mohm	Mohm	Mohm		
20 Mohm	10.00	9.98	-0.02	0.10	0.0058	Pass	20%
	Gohm	Gohm	Gohm	Gohm	Gohm		
20 Gohm	9.99	9.99	0.00	0.04	0.0058	Pass	2%

END OF REPORT