

SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

1. LICENSEE/LOCATION INSPECTED: Cameron Memorial Community Hospital 416 East Maumee Street Angola, IN 46703		2. NRC/REGIONAL OFFICE U.S. Nuclear Regulatory Commission Region III 2443 Warrenville Road Suite 210 Lisle, Illinois 60532-4351	
REPORT NUMBER(S) 2010-001			
3. DOCKET NUMBER(S) 030-37870	4. LICENSEE NUMBER(S) 13-32723-01	5. DATE(S) OF INSPECTION JANUARY 4 2010	

LICENSEE:
The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The inspection findings are as follows:

- 1. Based on the inspection findings, no violations were identified.
- 2. Previous violation(s) closed.
- 3. The violation(s), specifically described to you by the inspector as non-cited violations, are not being cited because they were self-identified, non-repetitive, and corrective action was or is being taken, and the remaining criteria in the NRC Enforcement Policy, NUREG-1600, to exercise discretion, were satisfied.

_____ Non-Cited Violation(s) was/were discussed involving the following requirement(s) and Corrective Action(s):

4. During this inspection certain of your activities, as described below and/or attached, were in violation of NRC requirements and are being cited. This form is a NOTICE OF VIOLATION, which may be subject to posting in accordance with 10 CFR 19.11.
(Violations and Corrective Actions)

Licensee's Statement of Corrective Actions for Item 4, above.

I hereby state that, within 30 days, the actions described by me to the inspector will be taken to correct the violations identified. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201 (corrective steps already taken, corrective steps which will be taken, date when full compliance will be achieved). I understand that no further written response to NRC will be required, unless specifically requested.

Title	Printed Name	Signature	Date
LICENSEE'S REPRESENTATIVE			
NRC INSPECTOR	S. J. Mulay		1/4/10

22/5

**SAFETY INSPECTION REPORT
AND COMPLIANCE INSPECTION**

1. LICENSEE Cameron Memorial Community Hospital		2. NRC/REGIONAL OFFICE Region III 2443 Warrenville Road, Suite 210 Lisle, IL 60532	
REPORT NUMBER(S) 2010-001			
3. DOCKET NUMBER(S) 030-37870	4. LICENSE NUMBER(S) 13-32723-01	5. DATE(S) OF INSPECTION January 04, 2010	
6. INSPECTION PROCEDURES USED 87130		7. INSPECTION FOCUS AREAS 03.01-03.07	

SUPPLEMENTAL INSPECTION INFORMATION

1. PROGRAM CODE(S) 2121	2. PRIORITY G5	3. LICENSEE CONTACT Michael Kinzer, MD., RSO	4. TELEPHONE NUMBER 260-665-2141
<input checked="" type="checkbox"/> Main Office Inspection		Next Inspection Date: January 2015	
<input type="checkbox"/> Field Office			
<input type="checkbox"/> Temporary Job Site Inspection			

PROGRAM SCOPE

This active facility performs about 20 diagnostic PET/CT imaging studies monthly utilizing fluorine-18 (FDG) only. The facility is staffed by two full-time technologists, one for PET imaging and one for CT studies. Unit doses are received from an area nuclear pharmacy. The licensee is operational Monday-Friday from 7:30 am-3:30 pm, however, the department is closed when patients are not scheduled.

Performance Observations

Interviews conducted with available staff revealed an adequate level of understanding of emergency and material handling procedures and techniques. Package surveys and wipes, dose calibrator constancy checks, area surveys and injection techniques were demonstrated and described with no regulatory issues identified. Compliance audits are performed at least quarterly by a consultant which appears to adequately oversee program activities.

Although not specifically reviewed, according to statements made by the licensee's consulting physicist, personal dosimetry results were maintained within ALARA limits.

Licensed material was observed well secured and/or under surveillance in the hot-lab area and was not readily accessible to members of the general public. Independent measurements taken in restricted and unrestricted areas of the hot-lab and imaging areas did not indicate readings above expected. Side-by-side readings with the licensee's survey instrument revealed comparable results. The licensee's sealed sources were verified present as compared to the most recent inventory.

JRS