



ECS MID-ATLANTIC, LLC

"Setting the Standard for Service"

Geotechnical • Construction Materials • Environmental • Facilities

January 13, 2010

Regional Administrator
Division of Nuclear Material Safety
Nuclear Materials Safety Branch
U.S. Nuclear Regulatory Commission, Region I
475 Allendale Road
King of Prussia, PA 19406-1415

Q-2

03037455

NRC license # 19-31242-01

144369
Control # ~~144869~~ *new*

Humboldt 5001 Serial # 3016 that was in temporary job site storage at Wal-Mart Store #1948-Wierton, WV Job Site, 243 Three Springs Drive Road, Wierton, WV. Was regularly leak tested. No signs of any leakage was noted. Attached is a copy of the leak tests while in temporary job site storage.

Sincerely,

Edward Dorsey, RSO/Field Service Manager

ECS Mid-Atlantic, LLC
5112 Pegasus Court, Suite S
Frederick, MD 21704
(301)668-4303 [Office]
(301)668-3519 [Fax]
(240)674-3272 [Mobile]

RECEIVED
REGION I
2010 JAN 19 PM 1:24

144369
NRC/RCM MATERIALS-002

27



North East Technical Services, Inc.

75 Aileron Ct., Suite 4

Westminster, MD 21157

Ph: 410.751.5090 Fax: 410.751.5091

ECS Mid-Atlantic, LLC - Frederick

5112 Pegasus Ct., Suite S

Frederick, MD 21704

ATTN: Ed Dorsey

Shipping Address: 5112 Pegasus Ct., Suite S

Frederick, MD 21704

LEAK TEST CERTIFICATE

MD Materials License # MD-13-020-01

This certifies that leak test analysis was conducted on the sample with the following information. The results shown below accurately represent the level of removeable contamination.

Gauge Model	5001EZ	Gauge S/N	H3016	Leak Test Date	3/2/2009
	Source		Reading in microCuries		
	5053GO		0.00005402		
	NJ03205		0.00000141		

Note: 0.005 microCuries (185 Bq) or greater is considered a leaking source. The source(s) tested above may remain in use.

Reviewed by: Douglas Sims

Date: MAR 04 2009

#4



North East Technical Services, Inc.
75 Aileron Ct., Suite 4
Westminster, MD 21157
Ph: 410.751.5090 Fax: 410.751.5091

ECS Mid-Atlantic, LLC - Frederick
5112 Pegasus Ct., Suite S
Frederick, MD 21704
ATTN: Ed Dorsey

Shipping Address: 5112 Pegasus Ct., Suite S
Frederick, MD 21704

LEAK TEST CERTIFICATE

MD Materials License # MD-13-020-01

This certifies that leak test analysis was conducted on the sample with the following information. The results shown below accurately represent the level of removeable contamination.

Gauge Model	5001EZ	Gauge S/N	H3016	Leak Test Date	8/22/2009
	Source		Reading in microCuries		
	8053GO		0.00006689		
	NJ03205		0.00000		

Note: 0.005 microCuries (185 Bq) or greater is considered a leaking source. The source(s) tested above may remain in use.

Reviewed by:

Douglas Simon

Date:

AUG 26 2009

Wednesday, August 26, 2009