

SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

1. LICENSEE/LOCATION INSPECTED: <i>Nesa & Associates</i> <i>23840 Dequindre Rd</i> <i>Warren, MI 48091</i> REPORT NUMBER(S) <i>2010</i> <i>2009/001</i>	2. NRC/REGIONAL OFFICE U.S. Nuclear Regulatory Commission Region III 2443 Warrenville Road Suite 210 Lisle, Illinois 60532-4351
--	--

3. DOCKET NUMBER(S) <i>030-38034</i>	4. LICENSEE NUMBER(S) <i>21-32747-01</i>	5. DATE(S) OF INSPECTION <i>1/22/10</i>
---	---	--

LICENSEE:
 The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The inspection findings are as follows:

- 1. Based on the inspection findings, no violations were identified.
- 2. Previous violation(s) closed.
- 3. The violation(s), specifically described to you by the inspector as non-cited violations, are not being cited because they were self-identified, non-repetitive, and corrective action was or is being taken, and the remaining criteria in the NRC Enforcement Policy, NUREG-1600, to exercise discretion, were satisfied.

_____ Non-Cited Violation(s) was/were discussed involving the following requirement(s) and Corrective Action(s):

4. During this inspection certain of your activities, as described below and/or attached, were in violation of NRC requirements and are being cited. This form is a NOTICE OF VIOLATION, which may be subject to posting in accordance with 10 CFR 19.11.
 (Violations and Corrective Actions)

Licensee's Statement of Corrective Actions for Item 4, above.

I hereby state that, within 30 days, the actions described by me to the inspector will be taken to correct the violations identified. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201 (corrective steps already taken, corrective steps which will be taken, date when full compliance will be achieved). I understand that no further written response to NRC will be required, unless specifically requested.

Title	Printed Name	Signature	Date
LICENSEE'S REPRESENTATIVE			
NRC INSPECTOR	<i>E. L. Kulzick</i>	<i>E. L. Kulzick</i>	<i>1/22/10</i>

JRS

SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

1. LICENSEE NESAs & Associates REPORT NUMBER(S) 2010-001		2. NRC/REGIONAL OFFICE Region III 2443 Warrenville Road, Suite 210 Lisle, IL 60532	
3. DOCKET NUMBER(S) 030-38034	4. LICENSE NUMBER(S) 21-32747-01	5. DATE(S) OF INSPECTION Jan. 22, 2010	
6. INSPECTION PROCEDURES USED 87124		7. INSPECTION FOCUS AREAS 0.01-0.07	

SUPPLEMENTAL INSPECTION INFORMATION

1. PROGRAM CODE(S) 3121	2. PRIORITY 5	3. LICENSEE CONTACT Jeshua Hansen	4. TELEPHONE NUMBER 586.427.5525
----------------------------	------------------	--------------------------------------	-------------------------------------

Main Office Inspection Next Inspection Date: 1/2015

Field Office

Temporary Job Site Inspection

PROGRAM SCOPE

The licensee was a small engineering /consulting company. The office possessed one moisture/density gauge which was used at least weekly for highway projects during the construction season. Two individuals were authorized to use the gauge. The licensee did not perform any service or maintenance activities on its gauges. These activities were performed by the manufacturer.

Performance Observations

During the inspection, the inspectors toured the facilities and performed radiation surveys of the storage area, examined the labeling on the gauge units, noted the locks on the gauge unit, reviewed selected records, and interviewed selected staff. At the time of the inspection, none of the units were in use. The inspector verified that the gauge was securely stored and locked within the storage area. The inspector examined the labeling on the gauge and transport case, and reviewed select records. The inspector verified documents pertaining to shipping papers, operating/emergency procedures, and NRC regulations.