

entire by
LICENSE CONTROL FORM

applied revised

Applicant Tenneco Chemicals, Inc. Heyden Division 290 River Drive Garfield, New Jersey 07026	License No. 29-09898-01	Amend No. 1	Expiration Date <i>April 30, 1968</i>	Control No. 76131
	Date Received 03/30/66	Date Issued APR 21 1966	PRUS 614	ASG REE 76131

() Letter () Application Dated: *March 29/66*

<u>BYPRODUCT</u>	<u>CHEMICAL OR PHYSICAL FORM</u>	<u>POSS. LIMIT</u>
<i>H₂ Hydrogen?</i>	<p><i>A foil contained in detector cells (Wilkins Model 02-104)</i></p> <p><i>B. To be used with Wilkins Instrument and Research, Inc. Model 600 gas chromatography device, with with Model 550-B over oven, for sample analysis</i></p>	<p><i>A Not to exceed 250 milliwatts per cell</i></p>

5
 7 ~~7~~ I. B. Wakeman
 21A
 21B
 16 *app - March 29, 1966*

Mail To: <i>H. R. Johnson</i>	Date Mailed: <i>APR 21 1966</i>	Reviewer: <i>REB</i>	Date Completed: <i>4/20/66</i>
----------------------------------	------------------------------------	-------------------------	-----------------------------------

April 30, 1966

DATE RECEIVED MAR 2 1964	EXPIRATION DATE April 25, 1966	ISSUE DATE	REVIEWER BHC
CONTROL NO. 57829	REFERENCE NO.	LICENSE NO. 29-9898-1 (D66)	AMENDMENT NO. NEW
ISOTOPE H ³	FORM A. Titanium trioxide foil contained in Wilkins Instrument + Research, Inc., Model 02-104 detector cell B. C. D. E. F. G. H.	POSSESSION LIMIT A. 250 mc B. C. D. E. F. G. H.	

AUTHORIZED USE

A. To be used in Wilkins Instrument + Research Corp. Model A-600-B, A-600-C, or A-550-B gas chromatography units.

REMARKS: Letters, Phone Calls, Visits, Exemptions, Etc. (Use reverse side if necessary)

- (5)
- (7) allow E. Klein, Daniel Tylicke, Irving B. Wakeman
H³ foil shall not be removed from ~~the~~ detector cells by the licensee
- (21)
- (16) app. ltr. Feb 25, 1964

MAIL TO: H R Johnson	DATE MAILED APR 29 1964	REVIEWER BWC RKB	DATE 4/29/64 4/29/64
-------------------------	----------------------------	------------------------	----------------------------

ACKNOWLEDGED