

JOSE D. MACATANGAY
MATERIALS LICENSING BRANCH
UNITED STATES NUCLEAR REGULATORY COMMISSION
REGION III
2443 WARRENVILLE ROAD STE 210
LISLE, ILLINOIS 60532-4352
OFFICE: (630)-829-9841 FAX: (630) 515-1078

CONVERSATION RECORD

|TIME

|DATE

ACTUALLY FAXED?

No – telephone only

01/20/2010

NAME OF PERSON(S) CONTACTED

David Simmonds, CNMT
(designated contact on request)

ORGANIZATION

Cardiology Consultants of East MI

TELEPHONE NO.

O: 810-664-4870

SUBJECT

License No.: 21-32035-01

Control No.: 318704

SUMMARY

We have reviewed your amendment request dated November 24, 2009, and need additional information as follows:

1. Please clarify what your action levels are for your wipe results.
2. Please clarify if the submitted results are net dpm or raw dpm counts.

Mr. Simmonds' Response:

Mr. Simmonds indicated that the wipe test action level is 2000 dpm/100 cm². He also stated that the wipe results are net dpm.

NAME OF PERSON DOCUMENTING CONVERSATION

SIGNATURE

DATE

Jose Macatangay



January 20, 2010