

JOSE D. MACATANGAY
MATERIALS LICENSING BRANCH
UNITED STATES NUCLEAR REGULATORY COMMISSION
REGION III
2443 WARRENVILLE ROAD STE 210
LISLE, ILLINOIS 60532-4352
OFFICE: (630)-829-9841 FAX: (630) 515-1078

CONVERSATION RECORD

|TIME

|DATE

ACTUALLY FAXED?

No – telephone only

01/14/2010

NAME OF PERSON(S) CONTACTED

Ray A. Carlson, Consultant

ORGANIZATION

Millennium Diagnostic Center

TELEPHONE NO.

O: 734-455-4730

C: 734-395-7361

SUBJECT

License No.: 21-32035-01

Control No.: 318704

SUMMARY

We have reviewed your facsimile amendment request sent December 19, 2009, and need additional information as follows:

1. Please submit another annotate diagram of the new location. Your submitted diagram is not legible. Please reference NUREG 1556, Vol. 9, Rev.2 Facility Diagrams, i.e. hot lab, storage, etc.
2. Will PET radionuclides be used at the new location?
3. What will be used at the new location – 35.100/200/300?

Mr. Carlson's Response:


Mr. Carlson indicated that no PET radionuclides or 35.300 materials are to be used at the new location. He agreed to submit a response by 01/20/2010.

NAME OF PERSON DOCUMENTING CONVERSATION

SIGNATURE

DATE

Jose Macatangay



January 14, 2010

J. J. J. J. 1/21/10