JOSE D. MACATANGAY
MATERIALS LICENSING BRANCH
UNITED STATES NUCLEAR REGULATORY COMMISSION
REGION III
2443 WARRENVILLE ROAD STE 210
LISLE, ILLINOIS 60532-4352
OFFICE: (630)-829-9841 FAX: (630) 515-1078

CONVERSATION RECORD	TIME	DATE
ACTUALLY FAXED?	No – telephone only	01/13/2010

NAME OF PERSON(S) CONTACTED	ORGANIZATION	TELEPHONE NO.
Sharon Dedo, RSO	Northstar Health System	O: 906-265-0463

SUBJECT
License No.: 21-18586-01 Control No.: 318667

SUMMARY
We have reviewed your amendment request received December 1, 2009 (dated November 1, 2009), and need additional information as follows:

1. Please clarify/explain how you derived at the background count and DPM (column 4) on your wipe test results.

I left message for Ms. Dedo to contact me on 01/13/2010 12PM Central.

Spoke with Ms. Dedo at 1PM Central:

Ms. Dedo's Response:
Ms. Dedo indicated that her background count was actually over 300 dpm and not sure why it indicated "0"; she also noted that the 4th column is "NET DPM". She will be sending response, via facsimile, to these clarifications either today or tomorrow (01/13 or 01/14).

I asked her to ensure that the facsimile is dated and signed.

NAME OF PERSON DOCUMENTING CONVERSATION	SIGNATURE	DATE
Jose Macatangay	Jose Macatangay	January 13, 2010