

SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

1. LICENSEE/LOCATION INSPECTED: West Shore Cardiology Consultants, P.C. 2525 Roberts Muskegon, MI 49444 REPORT NUMBER(S)	2. NRC/REGIONAL OFFICE U.S. Nuclear Regulatory Commission Region III 2443 Warrenville Road Suite 210 Lisle, Illinois 60532-4351
---	--

3. DOCKET NUMBER(S) 030-29788	4. LICENSEE NUMBER(S) 21-24873-01	5. DATE(S) OF INSPECTION <i>February 11, 2010</i>
--	--	---

LICENSEE:

The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The inspection findings are as follows:

- 1. Based on the inspection findings, no violations were identified.
- 2. Previous violation(s) closed.
- 3. The violation(s), specifically described to you by the inspector as non-cited violations, are not being cited because they were self-identified, non-repetitive, and corrective action was or is being taken, and the remaining criteria in the NRC Enforcement Policy, NUREG-1600, to exercise discretion, were satisfied.

_____ Non-Cited Violation(s) was/were discussed involving the following requirement(s) and Corrective Action(s):

- 4. During this inspection certain of your activities, as described below and/or attached, were in violation of NRC requirements and are being cited. This form is a NOTICE OF VIOLATION, which may be subject to posting in accordance with 10 CFR 19.11.
(Violations and Corrective Actions)

Licensee's Statement of Corrective Actions for Item 4, above.

I hereby state that, within 30 days, the actions described by me to the inspector will be taken to correct the violations identified. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201 (corrective steps already taken, corrective steps which will be taken, date when full compliance will be achieved). I understand that no further written response to NRC will be required, unless specifically requested.

Title	Printed Name	Signature	Date
LICENSEE'S REPRESENTATIVE			
NRC INSPECTOR	Robert P. Hays		2/11/10

RP

SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

1. LICENSEE West Shore Cardiology Consultants, P.C.		2. NRC/REGIONAL OFFICE Region III 2443 Warrenville Road, Suite 210 Lisle, IL 60532	
REPORT NUMBER(S) 2010-01			
3. DOCKET NUMBER(S) 03029788	4. LICENSE NUMBER(S) 21-24873-01	5. DATE(S) OF INSPECTION February 11, 2010	
6. INSPECTION PROCEDURES USED 87130 (10/24/02)		7. INSPECTION FOCUS AREAS 03.01-03.07	

SUPPLEMENTAL INSPECTION INFORMATION

1. PROGRAM CODE(S) 02201	2. PRIORITY 5	3. LICENSEE CONTACT M. Meengs, MD, RSO	4. TELEPHONE NUMBER 231/739-9427
------------------------------------	-------------------------	--	--

Main Office Inspection Next Inspection Date: **February 2015**

Field Office _____

Temporary Job Site Inspection _____

PROGRAM SCOPE

The licensee was a medical clinic authorized by the license to use any byproduct material as needed permitted by 10 CFR 35.100 and 10 CFR 35.200 at the location specified on the license. The clinic was staffed with three nuclear medicine technologists (NMTs) who routinely conducted an average of 8-14 patient procedures per day, Monday thru Friday, each week. The licensee received pre-calibrated unit doses of Myoview® for rest and stress cardiac tests as ordered from a local GE nuclear pharmacy. All waste was held for decay-in-storage (DIS) or returned to the nuclear pharmacy as a limited quantity shipment. No change in RSO since the previous inspection. The inspector performed independent and confirmatory radiation measurements which indicated results consistent with licensee survey records and postings.

Performance Observations

Interviews with licensee personnel indicated adequate knowledge of radiation safety concepts and procedures. The licensee's NMTs demonstrated/discussed: (1) unit dose prep and safe use; (2) package check-in and return procedures; (3) wipe test counting; (4) dose calibrator tests; (5) security of license materials; (6) quarterly radiation safety program reviews; (7) surveys; (8) sealed source inventory; (9) dosimetry (< 10% of annual regulatory limits); and (10) contamination events-(low activity contamination identified on treadmill and documented).