



59th and Arsenal

Building C

2 Campbell Plaza

St. Louis, MO 63139

TEL 314 644 3134

FAX 314 644 3135

14 YEARS OF PROFESSIONAL SERVICE

February 25, 2010

Mr. Jose Macatangy
US NRC Region III
2443 Warrenville Road

Re: Control # 318521

Dear Mr. Macatangy,

Per our conversation, here are the corrections of the unconformities on Nuclear Material License 24-32205-01 that you faxed to us on 2-17-10. Our responses are as follow:

Item #1

Our legal name is Tech Services Incorporated, as is stated on license.

Item #2

Our mailing address has changed from 2 Campbell Plaza/59th & Arsenal, Building C Saint Louis, MO 63139 to 5850 Arsenal Street Saint Louis, MO 63139.

Item #3

We would like to officially remove the follow storage locations:

- A) 2 Campbell Plaza/59th & Arsenal, Building C Saint Louis, MO 63139, in which we never had a leaking source, and we have attached the most recent leak tests for the gauges stored there.
- B) 1600 Genessee Street, Suite 960 Kansas City, Missouri 64102, in which we never had a leaking source. We have attached the most recent leak tests.

Our current storage locations are:

- a) 5911 Southwest Avenue Saint Louis, MO 63139
- b) 1322 Adam Street Kansas City KS 66103

Item#4

Nuclear gauges will be used at temporary jobsites and returned to the storage location after usage.

Item #5

The maximum possession limit per radionuclide is:

- 99mCi of Cs-137
- 396mCi of Am-241



Item #6

We are currently using Humboldt Model 5001 Series Gauges

Item #7

Please see attachments

Item #8

Please see attachments

Mr. Macatangy, I hope I have addressed all of NRC's concerns. Please contact me if you have additional concerns.

**Sincerely,
TSi Engineering Inc.**

**Carlos Villarreal, RSO
Laboratory Supervisor**



Two Campbell Plaza
Bldg. C
59th & Arsenal
St. Louis, MO 63139
314-644-3134 314-644-3135 Fax

Facsimile Transmittal Sheet

To JOSE MACATANGY Company: NRC

Fax #: 630-515-1078 Date: 2-26-10

From: CARLOS VILLARRETA Pages including cover: 4

cc:

Project Description:

Subject: RE: CONTROL #: 318521

MATERIAL LICENSE #: 24-32205-01

WE HAVE ATTACHED THE CORRECTIONS FOR
ITEMS #2 AND #3

PLEASE CALL WITH ANY QUESTION

CARLOS V.

314-645-0703

WARNING

The information contained in this communication is confidential, may constitute non-public information, and is intended only for the use of the addressee. Unauthorized use, disclosure or copying is strictly prohibited and may be unlawful. If you have received this communication in error, please immediately notify us at (314) 644-3134 and return the original message to us at the address above via the U.S. Postal Service.

Thank you