



UNITED STATES
NUCLEAR REGULATORY COMMISSION
REGION IV
612 EAST LAMAR BLVD, SUITE 400
ARLINGTON, TEXAS 76011-4125

March 2, 2010

West Valley Medical Center, Inc.
dba West Valley Medical Center
ATTN: Terri Steele, CNMT, RT(N)
Radiation Safety Officer
1717 Arlington Avenue
Caldwell, ID 83605

SUBJECT: LICENSE AMENDMENT AND NOTIFICATION PER 10 CFR 35.14(a)

Please find enclosed Amendment No. 19 to NRC License No. 11-27087-01 **adding Dr. Codina as an authorized user for any byproduct material permitted by 10 CFR 35.200 as requested.** An environmental assessment for this action is not required, since this action is categorically excluded under 10 CFR 51.22(c)(14)(iv). You should review the enclosed document carefully and be sure that you understand all conditions. If there are any errors or if you have any questions, contact me at 817-276-6590.

NRC expects licensees to conduct their programs with meticulous attention to detail and a high standard of compliance. Because of the serious consequences to employees and the public that can result from failure to comply with NRC requirements, you must conduct your radiation safety program according to the conditions of your NRC license, representations made in your license application, and NRC regulations. In particular, note that you must:

1. Operate by NRC regulations 10 CFR Part 19, "Notices, Instructions and Reports to Workers: Inspection and Investigations," 10 CFR Part 20, "Standards for Protection Against Radiation," and other applicable regulations.
2. Notify NRC in writing of any change in mailing address.
3. By 10 CFR 30.36(d) and/or license condition, notify NRC, promptly, in writing, and request termination of the license:
 - a. When you decide to terminate all activities involving materials authorized under the license whether at the entire site or any separate building or outdoor area;
 - b. If you decide not to acquire or possess and use authorized material; or
 - c. When no principal activities under the license have been conducted for a period of 24 months.
4. In accordance with 10 CFR 35.14, notify the NRC no later than 30 days after:
 - a. The date that the licensee permits an individual to work as an authorized user, an authorized nuclear pharmacist or an authorized medical physicist under 10 CFR 35.13(b)(1) through (b)(4);
 - b. An authorized user, an authorized nuclear pharmacist, a Radiation Safety Officer, or an authorized medical physicist permanently discontinues duties under the license or has a name change;

- c. The licensee's mailing address changes;
 - d. The licensee's name changes, but the name change does not constitute a transfer of control of the license as described in 10 CFR 30.34(b); or
 - e. The licensee has added to or changed the areas of use identified in the application or on the license where byproduct material is used in accordance with either 35.100 or 35.200.
5. Request and obtain a license amendment before you:
- a. Change Radiation Safety Officers;
 - b. Order byproduct material in excess of the amount, radionuclide or form authorized on the license;
 - c. Add or change the areas or address(es) of use identified in the license application or on the license, except for areas of use where byproduct material is used only in accordance with either 10 CFR 35.100 or 35.200; or
 - d. Change the name or ownership of your organization.
6. Submit a complete renewal application or termination request at least 30 days before the expiration date on your license. You will receive a reminder notice approximately 90 days before the expiration date. Possession of radioactive material after your license expires is a violation of NRC regulations.

In addition, please note that NRC Form 313 requires the applicant, by signature, to verify that the applicant understands that all statements contained in the application are true and correct to the best of the applicant's knowledge. The signatory for the application should be the licensee or certifying official rather than a consultant. Since the NRC also accepts a letter requesting amendment of an NRC license, the signatory for such a request should also be the licensee or certifying official rather than a consultant.

NRC will periodically inspect your radiation safety program. Failure to conduct your program according to NRC regulations, license conditions, and representations made in your license application and supplemental correspondence with NRC may result in enforcement action against you. This could include issuance of a notice of violation; imposition of a civil penalty; or an order suspending, modifying, or revoking your license as specified in the NRC Enforcement Policy. The NRC Enforcement Policy is available on the following internet address:

<http://www.nrc.gov/reading-rm/doc-collections/enforcement/>.

NRC no longer publishes the NRC Rules and Regulations loose leaf supplements. However, an electronic version of the NRC's regulations is available on the NRC Web site at www.nrc.gov. Additional information regarding medical uses of radioactive materials may be obtained on the NRC Web site at: <http://www.nrc.gov/materials/miau/med-use-toolkit.html>. This site also provides the updated Training and Experience NRC Form 313A series of forms and guidance, as well as information on the revised regulations for naturally-occurring and accelerator-produced radioactive materials (NARM).

West Valley Medical Center, Inc.
dba West Valley Medical Center

-3-

In accordance with 10 CFR 2.390 of the NRC's "Rules of Practice," a copy of this letter and its enclosure will be available electronically for public inspection in the NRC Public Document Room or from the NRC's document system (ADAMS). ADAMS is accessible from the NRC Web site at <http://www.nrc.gov/reading-rm/adams.html>.

Thank you for your cooperation.

Sincerely,

/RA/

Michelle Simmons, Health Physicist
Nuclear Materials Safety Branch B

Docket: 030-32242
License: 11-27087-01
Control: 472558

Enclosures: As stated