

SOUTHWEST RESEARCH INSTITUTE®

6220 Culebra Road, P.O. Drawer 28510
Institute Quality Systems
Institute Calibration Laboratory
Phone: 210-522-5215 Fax 210-522-4834



Calibration Laboratory
Certificate #0972-01

Certificate of Calibration

Cost Center / Customer: DIV20 / DON BANNON

Mail Stop: B51

Manufacturer/Model: TROEMNER / 200G

Description: WEIGHT, CLASS 1

Serial Number: 63813

Asset Number: 004199

Procedure: WEIGHTS - 10 DEC 07

Work Order: 303091650

Date Issued: 10-Dec-2009

Date Calibrated: 10-Dec-2009

*** Date Due :** 10-Dec-2010

**** Results:** FOUND-LEFT

Temperature: 68.0 °F

Humidity: 41 %RH

Barometer: 14.39 psia

This certificate documents traceability to the National Institute of Standards and Technology (NIST) and the International System of Units (SI). The Laboratory quality system conforms to ISO/IEC 17025, 2005, ANSI/NCCL Z540-1-1994 and relevant requirements of the ISO 9000-2000 standard. This certificate shall not be reproduced, except in full, without the written approval of the Southwest Research Institute Calibration Laboratory. This certificate shall not be used to claim product endorsement by Southwest Research Institute, American Association for Laboratory Accreditation (A2LA) or any agency of the U. S. Government. Results of this calibration relate only to the instrument described above at the time of calibration and does not imply any long term stability of the instrument.

*Determined by the customer, does not imply the instrument will remain within tolerance as any number of factors may cause an out-of-tolerance condition before this date. **Data type found in this certificate or attached measurement report must be interpreted as: Found-left - adjustment and/or repair was not performed, As-found - data is before unit is adjusted and/or repaired, As-left - data is after adjusted and/or repaired was performed. The customer has sole responsibility for determination of in-/out-of-tolerance or compliance/noncompliance.

Measurement uncertainty calculated in accordance with the method described in the ISO "Guide to the Expression of Uncertainty in Measurement" (GUM), for a confidence level of approximately 95 percent using a coverage factor of $k=2$.


Remarks: THE CONVENTIONAL MASS VALUE OF THE 200 g WGT IS: 200.000 16 g UNC +/-0.053 mg. CLASS 1 TOL +/-0.5 mg.

Standards Used

<u>Asset #</u>	<u>Manufacturer</u>	<u>Model</u>	<u>Description</u>	<u>Cal Date</u>	<u>Due Date</u>
007103	TROEMNER	200 G	WEIGHT, CLASS E1	24-Sep-2009	24-Sep-2010
012068	SARTORIUS	CC310	MASS COMPARATOR	29-Jan-2009	29-Jan-2010


Walt Hill

Laboratory Manager


Carlos Mendoza

Metrology Technician

Southwest Research Institute
Calibration Laboratory
Double Substitution Worksheet
per SOP 4, December 2003

Asset Number: 004199
Date of Calibration: 12/10/2009
Accuracy Class: 1

Nominal	Reading	Unit
200.0		g
1. S	0.00000	g
2. X	0.00026	g
3. X	0.00025	g
4. S	0.00010	g
▲	0.000205	g
<p>▲ = d_x</p> <p>$d_x = 0.000205 \text{ g}$</p> <p>$C_s = 199.999952 \text{ g}$</p> <p>$C_x = C_s + d_x$</p> <p>$C_x = 200.000157 \text{ g}$</p>		