

U.S. NUCLEAR REGULATORY COMMISSION

MATERIALS LICENSE

Pursuant to the Atomic Energy Act of 1954, as amended, the Energy Reorganization Act of 1974 (Public Law 93-438), and Title 10, Code of Federal Regulations, Chapter I, Parts 30, 31, 32, 33, 34, 35, 36, 39, 40, and 70, and in reliance on statements and representations heretofore made by the licensee, a license is hereby issued authorizing the licensee to receive, acquire, possess, and transfer byproduct, source, and special nuclear material designated below; to use such material for the purpose(s) and at the place(s) designated below; to deliver or transfer such material to persons authorized to receive it in accordance with the regulations of the applicable Part(s). This license shall be deemed to contain the conditions specified in Section 183 of the Atomic Energy Act of 1954, as amended, and is subject to all applicable rules, regulations, and orders of the Nuclear Regulatory Commission now or hereafter in effect and to any conditions specified below.

<p style="text-align: center;">Licensee</p> <p>1. Fischer Technology, Incorporated</p> <p>2. 750 Marshall Phelps Road Windsor, Connecticut 06095</p>	<p>In accordance with the application dated February 5, 2010,</p> <p>3. License number 06-19165-01 is amended in its entirety to read as follows:</p> <hr/> <p>4. Expiration date November 30, 2011</p> <hr/> <p>5. Docket No. 030-17073 Reference No.</p>
--	--

<p>6. Byproduct, source, and/or special nuclear material</p> <p>A. Carbon 14</p> <p>B. Strontium 90</p> <p>C. Promethium 147</p> <p>D. Thallium 204</p>	<p>7. Chemical and/or physical form</p> <p>A. Sealed Source (Helmut Fischer GmbH Model 604-082 or 600-493)</p> <p>B. Sealed Source (Helmut Fischer GmbH Model C07 Series or Isotope Products Model TCB-1)</p> <p>C. Sealed Source (Helmut Fischer GmbH Model C07 Series or Isotope Products Model TCB-1)</p> <p>D. Sealed Source (Helmut Fischer GmbH Model C07 Series or Isotope Products Model TCB-1)</p>	<p>8. Maximum amount that licensee may possess at any one time under this license</p> <p>A. No single source to exceed the maximum activity specified in the certificate of registration issued by the U.S. Nuclear Regulatory Commission or an Agreement State and 20 millicuries total</p> <p>B. 25 microcuries per source and 2 millicuries total</p> <p>C. 900 microcuries per source and 180 millicuries total</p> <p>D. 150 microcuries per source and 30 millicuries total</p>
---	---	---

**MATERIALS LICENSE
SUPPLEMENTARY SHEET**License Number
06-19165-01Docket or Reference Number
030-17073

Amendment No. 7

9. Authorized use:

A. through D. For possession incident to service for other persons as defined in 10 CFR 20.1003 for:

- (1) Installation, initial radiation surveys, relocation, removal from service, dismantling, alignment, replacement, disposal of the sealed source and non-routine maintenance or repair of components related to the radiological safety of Fischer Technology Fischerscope Beta Models 870, 2045, and 2060 gauge devices that have been registered either with the U.S. Nuclear Regulatory Commission under 10 CFR 32.210 or with an Agreement State.
- (2) Instruction and training of individuals in the use of Fischer Technology Fischerscope Beta Models 870, 2045, and 2060 devices, including demonstration of backscatter thickness gauges.
- (3) Leak test sample collection and analysis.

A. through D. Distribution of Fischer Technology Fischerscope Beta Models 870, 2045, and 2060 backscatter thickness gauges to persons authorized to receive the licensed material pursuant to the terms and conditions of a specific license issued by the U.S. Nuclear Regulatory Commission or an Agreement State.

CONDITIONS

10. Licensed material may be used or stored only at the licensee's facilities located at 750 Marshall Phelps Road Windsor Connecticut, and at temporary job sites of the licensee anywhere in the United States where the U.S. Nuclear Regulatory Commission maintains jurisdiction for regulating the use of licensed material.
11.
 - A. Licensed material shall be used by, or under the supervision of, Helmut Fischer, Robert Christensen, or Raymond Moncevicus.
 - B. The Radiation Safety Officer for this license is Raymond Moncevicus.
12.
 - A. Sealed sources shall be tested for leakage and/or contamination at intervals not to exceed six months or at the intervals specified in the certificate of registration issued by the U.S. Nuclear Regulatory Commission under 10 CFR 32.210 or under equivalent regulations of an Agreement State.
 - B. Each sealed source fabricated by the licensee shall be inspected and tested for construction defects, leakage, and contamination prior to any use or transfer as a sealed source.
 - C. In the absence of a certificate from a transferor indicating that a leak test has been made within the intervals specified in the certificate of registration issued by the U.S. Nuclear Regulatory Commission under 10 CFR 32.210 or under equivalent regulations of an Agreement State, prior to the transfer, a sealed source received from another person shall not be put into use until tested and the test results received.

**MATERIALS LICENSE
SUPPLEMENTARY SHEET**License Number
06-19165-01Docket or Reference Number
030-17073

Amendment No. 7

- D. Sealed sources need not be tested if they contain only hydrogen-3; or they contain only a radioactive gas; or the half-life of the isotope is 30 days or less; or they contain not more than 100 microcuries of beta- and/or gamma-emitting material or not more than 10 microcuries of alpha-emitting material.
- E. Sealed sources need not be tested if they are in storage and are not being used; however, when they are removed from storage for use or transferred to another person and have not been tested within the required leak test interval, they shall be tested before use or transfer. No sealed source shall be stored for a period of more than 10 years without being tested for leakage and/or contamination.
- F. The leak test shall be capable of detecting the presence of 0.005 microcurie (185 becquerels) of radioactive material on the test sample. If the test reveals the presence of 0.005 microcurie (185 becquerels) or more of removable contamination, a report shall be filed with the U.S. Nuclear Regulatory Commission in accordance with 10 CFR 30.50(c)(2), and the source shall be removed immediately from service and decontaminated, repaired, or disposed of in accordance with Commission regulations.
- G. Tests for leakage and/or contamination, including leak test sample collection and analysis, shall be performed by the licensee or by other persons specifically licensed by the U.S. Nuclear Regulatory Commission or an Agreement State to perform such services.
- H. Records of leak test results shall be kept in units of microcuries and shall be maintained for 5 years.
13. The licensee shall conduct a physical inventory every six months, or at other intervals approved by the U.S. Nuclear Regulatory Commission, to account for all sources and/or devices received and possessed under the license. Records of inventories shall be maintained for 5 years from the date of each inventory and shall include the radionuclides, quantities, manufacturer's name and model numbers, and the date of the inventory.
14. Sealed sources or detector cells containing licensed material shall not be opened or sources removed from source holders by the licensee.
15. This license does not authorize distribution to persons licensed pursuant to 10 CFR 32.72 or 32.74; to persons exempt from licensing; or to general licensees.
16. The licensee is authorized to transport licensed material in accordance with the provisions of 10 CFR Part 71, "Packaging and Transportation of Radioactive Material."

**MATERIALS LICENSE
SUPPLEMENTARY SHEET**License Number
06-19165-01Docket or Reference Number
030-17073

Amendment No. 7

17. Except as specifically provided otherwise in this license, the licensee shall conduct its program in accordance with the statements, representations, and procedures contained in the documents, including any enclosures, listed below. The U.S. Nuclear Regulatory Commission's regulations shall govern unless the statements, representations, and procedures in the licensee's application and correspondence are more restrictive than the regulations.

A. Application dated November 15, 2001 (ML013250198)



For the U.S. Nuclear Regulatory Commission

Date March 3, 2010

By

Original signed by Dennis R. Lawyer

Dennis R. Lawyer
Commercial and R&D Branch
Division of Nuclear Materials Safety
Region I
King of Prussia, Pennsylvania 19406